

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property Organization  
International Bureau



(10) International Publication Number  
**WO 2020/009930 A1**

(43) International Publication Date  
09 January 2020 (09.01.2020)

- (51) International Patent Classification:  
G01T 1/02 (2006.01) G01T 1/08 (2006.01)  
G01T 1/06 (2006.01)
- (21) International Application Number:  
PCT/US2019/039811
- (22) International Filing Date:  
28 June 2019 (28.06.2019)
- (25) Filing Language: English
- (26) Publication Language: English
- (30) Priority Data:  
62/694,291 05 July 2018 (05.07.2018) US
- (71) Applicant: **ISP INVESTMENTS LLC** [US/US]; 1011 Centre Road, Suite 315, Wilmington, Delaware 19805 (US).
- (72) Inventors: **CROHN, Rob**; 38 Douglas Dr., Towaco, New Jersey 07082 (US). **HOOD, David K.**; 29 Concord Lane, Basking Ridge, New Jersey 07920 (US).

(74) Agent: **DAVIS, William J.** et al.; Ashland LLC, 1005 U.S. 202/206, Bridgewater, New Jersey 08807 (US).

(81) Designated States (unless otherwise indicated, for every kind of national protection available): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BN, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DJ, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IR, IS, JO, JP, KE, KG, KH, KN, KP, KR, KW, KZ, LA, LC, LK, LR, LS, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PA, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SA, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.

(84) Designated States (unless otherwise indicated, for every kind of regional protection available): ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, ST, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM,

(54) Title: A NEW DOSIMETRY DEVICE FOR QUANTIFICATION OF RADIATION





Fig. 1

(57) Abstract: Provided herein is a dosimetry device for quantifying the dosage of radiation emitted from a radiation source, the device comprising: (i) a radiation dose indicator; (ii) an optical means to capture the color change; and (iii) a software means to compare the optical density of the dose indicator as compared to a predetermined calibration curve. Also provided herein is a method of quantifying the dosage of radiation emitted from the radiation source. Further provided herein is use of said dosimetry device in various medical, food and industrial applications.



WO 2020/009930 A1

TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW,  
KM, ML, MR, NE, SN, TD, TG).

**Published:**

— *with international search report (Art. 21(3))*

## **A NEW DOSIMETRY DEVICE FOR QUANTIFICATION OF RADIATION**

### **FIELD OF THE INVENTION**

[0001] The present application relates to a dosimetry device comprising a radiation sensitive composition, an optical means and a software means to quantify the dosage of radiation emitted from the radiation source. The present application further relates to use of said dosimetry device in various medical, food and industrial applications.

### **BACKGROUND OF THE INVENTION**

[0002] Quantification of radiation emitted from various sources is an important function and finds several applications in medical, research, food storage, transportation of sensitive goods, and industrial operations. A dosimeter is one such device used to indicate or measure exposure to ionizing radiation. It is a solid object either available as a plate, or in any other shape that can be easily viewed and sometimes shows visual transformation of color without the use of a spectrophotometer. Several types of dosimeters such as Thermoluminescence Dosimeters (TLD), Optically Stimulated Luminescence (OSL), Radio Luminescence Glass (RLG), X-ray film, and Track Etch are currently available in the market. Typically, these are used for measuring and monitoring both medical and industrial radiations such as X-rays, gamma rays, high speed electrons, etc.

[0003] Color changing / developing self-indicating Instant Radiation Alert Dosimeters (SIRAD) for monitoring low dose, e.g., are commercially marketed by JP Laboratories Inc., New Jersey under trademark SIRAD ®. Dosimeter SIRAD® has a sensing strip made of polymeric material such as colorless solid monomers of diacetylenes. They usually form red or blue colored polymers / plastics when irradiated with high energy radiation such as X-ray, gamma ray, electrons, and neutrons. As exposure to radiation increases, the color of the sensing strip comprising diacetylenes intensifies proportional to the dose.

[0004] US Patent No. 8,242,464 [Assigned to Gordhanbhai N. Patel/JP Labs] discloses Identification Personal Dosimeter comprising self indicating radiation sensor, reader allowing the user to estimate the dose instantly.

[0005] US Patent No. 10,060786 [Assigned to L'Oreal] discloses personal Ultra-Violet (UV) radiation measurement system comprising UV measuring device, terminal device to capture the UV radiation and receive the specific information about user's risk level of UV exposure.

[0006] RadSure® irradiation indicator is one kind of dosimetry device, marketed by Ashland Specialty Ingredients Inc. providing positive, visual verification of irradiation. When attached to blood products, RAD-SURE® indicators show whether the blood products have been irradiated or not. Before a blood product and its attached indicator are irradiated, the indicator reads "NOT IRRADIATED". After irradiation, the word "NOT" in the indicator window is obscured and the indicator reads "IRRADIATED". There are a few more commercially available indicators under the tradenames RADTAG® and ONPOINT® as shown in Figure 1.

[0007] Many modifications have been made to the indicators in means of identifying amount of irradiation, reducing contamination, safe-guarding temperature, and visual identification means to alert the user about risk of radiation. However, there is no accurate mechanism to measure actual amount of radiation emitted from the radiation source.

[0008] Therefore, there exists a need in the art for a dosimetry device which can quantify the actual radiation (X-rays or gamma rays) exposure at the location of the user and provide detailed information regarding the user.

[0009] Surprisingly, our dosimetry device quantifies the actual radiation emitted from the radiation source with high precision.

### **SUMMARY OF THE INVENTION**

[0010] In one aspect, provided herein is a dosimetry device for quantifying the dosage of radiation emitted from a radiation source, the device comprising: (i) a radiation dose indicator comprising radiation sensitive composition to measure the radiation and visually represent the amount of radiation emitted as color change; (ii) an optical means to capture the color change of the dose indicator after exposure to radiation; and (iv) a software means to compare the optical density of the dose indicator as compared to a predetermined calibration curve developed using percentage optical density versus cumulative radiation dosage to quantify the dosage of radiation emitted from the radiation source.

[0011] In another aspect, provided herein is use of a dosimetry device comprising: (i) a radiation dose indicator comprising radiation sensitive composition to measure the radiation emitted from a radiation source and visually represent the amount of radiation emitted as color change; (ii) an optical means to capture the color change of the dose indicator after exposure to radiation; and (iii) a software means to compare the optical density of the dose indicator to a predetermined calibration curve developed using percentage optical density versus

cumulative radiation dosage to quantify the dosage of radiation emitted from the radiation source; in sterilization of surfaces and solutions, medical imaging, medical or industrial equipment quality assurance testing, UV light measurement, food processing and storage, transportation of radiation sensitive materials, or blood storage.

[0012] In yet another aspect, provided herein is a method of quantifying the dosage of radiation emitted from a radiation source, comprising the steps of: (i) exposing a dosimetry device to irradiation, said dosimetry device comprising radiation sensitive indicator which measures the radiation and visually represents the amount of radiation emitted as color change; (ii) capturing the color change of the dose indicator after exposure to radiation using an optical means; and; (iii) comparing the optical density of dose indicator to a predetermined calibration curve developed using percentage optical density versus cumulative radiation dosage to quantify the dosage of radiation emitted from the source.

[0013] In yet another aspect, provided herein is a dosimetry device as shown in Figures 1 to 7.

**BRIEF DESCRIPTION OF DRAWINGS**

[0014] The invention will be better understood from the following detailed description when read in conjunction with the accompanying drawings. It is emphasized that, according to common practice, the various features of the drawing are not to scale. On the contrary, the dimensions of the various features are arbitrarily expanded or reduced for clarity. The following figures are included in the drawing:

[0015] Fig. 1 is a pictorial view of RadTag® (above) and Onpoint® (below) indicators

[0016] Fig. 2 is a pictorial view of new Rad-Sure®

[0017] Fig.3 is a pictorial view of product test assembly for Gafchromic® film #1 before (left) and after (right) exposure to 10Gy in X-ray cabinet

[0018] Fig. 4 is a graphical view of comparative optical density measurements for film vs. X-ray exposure level.

[0019] Fig. 5 is a graphical view of Gafchromic® film #1 response to varying degree of X-ray radiation exposure (without flash)

[0020] Fig.6 is a graphical view of Gafchromic® film #1 RGB data from iPhone captured images.

[0021] Fig. 7 is a flow diagram of radiation dosimetry device.

## DETAILED DESCRIPTION OF THE INVENTION

[0022] The following detailed description is intended to be representative only and not limiting as to the described dosimeter enabled with improved and accurate radiation quantification functionality. Many variations can be derived by one skilled in the art which are included within the scope of the present invention. The following detailed discussion of the various alternative and preferred embodiments will illustrate the general principles of the invention.

[0023] As utilized in accordance with the disclosure, the following terms, unless otherwise indicated, shall be understood to have the following meanings.

[0024] All percentages, parts, proportions and ratios as used herein, are by weight of the total composition, unless otherwise specified. All such weights as they pertain to listed ingredients are based on the active level and, therefore; do not include solvents or by-products that may be included in commercially available materials, unless otherwise specified.

[0025] As used herein, the words “comprising” (and any form of comprising, such as “comprise” and “comprises”), “having” (and any form of having, such as “have” and “has”), “including” (and any form of including, such as “includes” and “include”) or “containing” (and any form of containing, such as “contains” and “contained”.

[0026] All publications, articles, papers, patents, patent publications, and other references cited herein are hereby incorporated herein in their entireties for all purposes to the extent consistent with the disclosure herein.

[0027] The term “Color Space” as used herein generally refers to a color model (or color system) and is an abstract mathematical model which simply describes the range of colors as tuples or numbers, typically as 3 or 4 values or color components (e.g. RGB). Each color in the system is represented as dot. When defining a color space, the usual reference standard is the CIELAB or CIEXYZ color spaces, which were specifically designed to encompass all colors the average human eyes can see.

[0028] The term “Gafchromic® film” as used herein generally refers to a radiochromic dosimetry film designed for the quantitative measurement of absorbed dose of high-energy photons. Key technical features include: (i) dynamic dose range: 10 Gy to 1000 Gy; (ii)

develops in real time without post-exposure treatment; (iii) energy-dependence: minimal response difference from 100keV into the MV range; (iv) near tissue equivalent; (v) high spatial resolution – can resolve features to 5 $\mu$ m, or less; (vi) active coating exposed for detection of low energy photon and electron; (vii) proprietary new technology incorporating a marker dye in the active layer: enables non-uniformity correction by using triple-channel dosimetry and decreases UV/light sensitivity; and (viii) stable at temperatures up to 60°C.

[0029] The term "ionizing radiation" as used herein generally refers to radiation with a level of energy that is high enough to cause atoms to lose electrons and become charged or ionized. Ionizing radiation may be in the form of high energy particles, like an alpha and beta particles, protons and neutrons, or in the form of electromagnetic waves, like gamma rays or X-rays. High energy particles and electromagnetic waves are released from the nuclei of radioactive atoms that are decaying or may be created by causing accelerated electrons to strike a metal target.

[0030] The term "radiation detection medium" as used herein generally refers to a medium that undergoes a detectable change when exposed to radiation. The change may be immediately visible or may require a development process.

[0031] The term "radiochromic film" or film as used herein generally refers to film that changes color and produces a visible image when exposed to ionizing radiation, but experiences insignificant change from exposure to visible light or other forms of non-ionizing radiation. This film requires no chemical or physical processing.


[0032] The term "Rad-Sure®" as used herein, is a blood irradiation indicator that provides positive visual verification of irradiation at the minimum specified dose. Rad-Sure® indicator is available in two types: Gamma and X-Ray. Gamma is compatible with Cesium-137 or Cobalt-60 radiation sources and X-Ray is compatible with X-ray irradiators that utilize X-rays generated from 160kVp sources that are filtered through 0.38 mm of copper, or 150kVp sources that are filtered through 1 mm of aluminum. Manufactured from Gafchromic® film, the world's highest resolution dosimeter, Rad-Sure® is the standard for blood irradiation indicators for over 25 years.

[0033] In one embodiment, the present application provides a dosimetry device comprising: (i) a radiation dose indicator comprising radiation sensitive composition to measure the radiation and visually represent the amount of radiation emitted from the radiation source as


color change; (ii) an optical means to capture the color change of the dose indicator after exposure to radiation; and (iii) a software means to compare the optical density of the dose indicator to a predetermined calibration curve developed using percentage optical density versus cumulative radiation dosage to quantify the dosage of radiation emitted from the source.

[0034] Accordingly, the radiation dose indicator comprises a radiation sensitive composition which measures the radiation and indicates the change in radiation. The radiation sensitive composition is selected from radiation sensitive film, a radiation sensitive patch, or any other device which can detect radiation emitted from the radiation source.

[0035] Accordingly, the radiation sensitive composition comprises a radiation sensitive film such as radiochromic film, a film that instantly changes color upon exposure to ionizing radiation and needs no chemical processing is a radiochromic film. These films are widely recognized for their large exceptional spatial resolution reaching to a level of at least 0.025 *mm*. Another advantage of radiochromic film is its tissue equivalence that the adsorbed radiation dose is truly reflection of the dose adsorbed by tissue. Materials used in the sensing strip of radiochromic film are a unique class of compounds called diacetylenes with a general formula:



wherein  $R^1$  and  $R^{11}$  are substituent groups. Diacetylenes are colorless solid monomers. They usually form red or blue colored polymers with a general formula:



where  $n$  is number of monomer units, when irradiated with high energy radiation such as X-ray, gamma ray, electrons, and neutrons. As exposure increases, the color of the sensing strip comprising diacetylenes intensifies proportional to the dose.

[0036] Accordingly, commercially available radiochromic film models are manufactured under the commercial name Gafchromic ® by Ashland Specialty Ingredients Inc. Radiochromic film employed in the current application can be selected from HD-V2, EBT3, EBT-XD, MD-V3, RTQA2, XR-QA2, XR-CT2, XR-M2, and XR-RV3.

[0037] Accordingly, the radiochromic film has a radiation dose range selected from about 0.1 to about 10,000 Gray (Gy) units per second.

[0038] Accordingly, the radiochromic film is Gafchromic® film having an active component which is micro-particulate, radiation sensitive monomer that is dispersed in a polymer matrix and coated onto a polyester film base. When the active monomeric component is exposed to ionizing radiation, a polymerization reaction is initiated, resulting in the production of a dye polymer. Since this polymer is by nature, a dye, the exposure produces coloration within the film.

[0039] In another embodiment, the color indicator can be provided with a dye or pigment or combinations thereof that functions as a marker. Accordingly, the dose indicator is activated once radiation is emitted from the radiation source and this change is visually shown as color change.

[0040] Accordingly, suitable pigment materials are described in Hunger's "Industrial Organic Pigments," Holes "Dictionary of Pigments," and Leach and Pierce's "Printing Ink Manual."

[0041] Examples of yellow colored, organic and inorganic, pigments include C.I. Pigment Yellow 1, C.I. Pigment Yellow 74, azo pigments such as C.I. Pigment 12 and C.I. Pigment Yellow 17 and the like.

[0042] Examples of black colored pigments include carbon black, titanium black, aniline black, and the like.

[0043] Examples of white colored pigments include basic lead carbonate, zinc oxide, barium sulfate, titanium oxide, silver white, strontium titanate, and the like.

[0044] Examples of red colored pigments include naphthol red (C.I. Pigment Red 2), C.I. Pigment Red 3, C.I. Pigment Red 176 and C.I. Pigment Red 23 and the like.

[0045] Examples of green colored pigments include phthalocyanine green (C.I. Pigment Green 7), C.I. Pigment Green 36, and C.I. Pigment Green 1 and the like.

[0046] Examples of blue colored pigments include phthalocyanine blue (C.I. Pigment Blue 15:3), C.I. Pigment Blue 15:6, and C.I. Pigment Blue 16 and the like.

[0047] Examples of blue colored dyes include methylene blue, Acid Blue 1, Basic Blue 1 and C.I. Solvent Blue 7 and the like.

[0048] Examples of red colored dyes include Acid Red 18, Basic Red 1 and C.I. Solvent Red 8 and the like.

[0049] Examples of green colored dyes include Acid Green 1 and Basic Green 1 and the like.

[0050] Examples of black colored dyes include C.I. Solvent Black 5 and the like.

[0051] In another embodiment, the optical means provided in the dosimetry device captures the color change which occurred due to emission of radiation from the source. Optical means is a camera selected from group comprising a smart phone camera, a high definition camera, a magnifying camera, a densitometer, an image scanner, a video camera, a TV camera, an optical imaging device and the like.

[0052] Accordingly, the camera can be a smart phone camera such as iPhone from Apple Inc. The smart phone can comprise a Central Processing Unit (CPU), I/O interface, and a network controller, frequency modulation, Bluetooth combo chip. Further, the CPU can be implemented as multiple processors working in collaboration such as cloud computing environment to perform the instructions of the dosimetry device.

[0053] In another embodiment, the software means provided in the dosimetry device converts the captured color change into numerical color data and compares with predetermined data using quantification algorithms.

[0054] Accordingly, the captured color is in the form of pixels. The color intensity is measured in the form of "color space". A color space is a useful method for users to understand the color capabilities of a specific digital device or file. It represents what a camera can see, a monitor can display, or a printer can print, and the like. There are a variety of color spaces, such as RGB, CMY, HSV, HIS. Each color is represented as a dot. All the dots are identified and converted to specific set of numbers and measured in triples or tuples, typically as 3 or 4 color components.

[0055] Accordingly, RGB (R=Red, G=Green, B=Blue) is a kind of color space which uses red, green and blue to elaborate color model. An RGB color space can be simply interpreted as "all possible colors" which can be made from three colors for red, green and blue. In such conception, each pixel of an image is assigned a range of 0 to 255 intensity values of RGB components. That is to say, using only these three colors, there can be 16,777,216 colors on the screen by different mixing ratios. Currently, RGB color spaces are sRGB and Adobe RGB.

The sRGB was developed by both HP and Microsoft in 1977, where "S" can be interpreted as "standard"; while Adobe RGB is a standard color gamut from Adobe. For these two commonly-used color spaces, Adobe RGB offers a wider color space comparing to sRGB, and it contains CMYK color gamut (sRGB does not have it). As a result, Adobe RGB has richer color layers and lower color saturation.

[0056] Accordingly, the radiation emitted from the radiation source is captured as color change, color intensity is identified and converted into numbers using RGB software. Preferably, the software program is "Get RGB" from the Imaging Technology Group (ITG) at Beckman Institute at the University of Illinois-Urbana Champaign. It is a free, open source software for 2D image analysis; software program developed based on quantification algorithms.

[0057] In another embodiment, the color data is compared to the predetermined calibration curves of color versus radiation dose response data. The calibration curve is a plot of measured signal (transmission or optical density) versus the absorbed dose. Hence calibration curve can also be termed as an "absorbed dose sensitivity" curve, since the slope at any absorbed dose point is the film absorbed dose sensitivity for that absorbed dose. In general, a calibration curve in terms of optical density will have a non-linear very low dose region, a linear or nearly linear low-to medium dose region, and a non-linear high dose region. The linear region is useful for the quantization of film sensitivity in terms of net optical density per unit absorbed dose. Various optical density measurement data for various radiochromic films have been illustrated in Figure 4. Calibration curves for a radiation detection medium are often prepared by exposing one or more areas of the detection medium to different and known amounts of radiation using a linear accelerator or a similar device capable of generating a range of known dose levels. Another method frequently used is to expose the detection means to a continuously varying level of doses. This can be done by interposing a wedge of material with continuously varying thickness between the radiation source and the detection medium. Alternatively, the radiation sensitive medium may be sandwiched between two blocks and positioned so that the medium is in a plane parallel to the beam when it is exposed. In this configuration the dose applied to the radiation sensitive medium decreases continuously with depth below the top surface of the blocks. This type of exposure is often referred to as a depth-dose exposure. Typically, calibration curves are generated by measuring the response of the radiation sensitive medium for different dose levels. In the instance of radiation films, it is common to measure the light transmission or optical density of the medium for numerous different radiation dosage levels.

Color change images measured as density-red values are cropped into software program selected from "Get RGB" and various calibration curves for each lot of dosimetry film are constructed. Hence a library of calibration curves is made available to the user.

[0058] In another embodiment, emission of radiation such as X-rays, gamma rays, a nuclear detonation, or any other a radiation source, the self indicating radiation sensor of the device develops color, preferably blue or red, instantly. The color intensifies as the dose increases thereby providing the wearer and medical personnel instantaneous information on cumulative radiation exposure. The color intensity of the sensing strip increases with increasing dose. Dose can be estimated with an accuracy which is better than 20% with color reference chart and better than 10% using a calibration plot of optical density versus dose or CD camera.

[0059] In another embodiment, the test assemblies are subjected to optical density measurements. A cell phone camera is employed to measure color change and is captured in the form of images in HDR mode or with flash mode. The density-red data exhibits a nearly linear response for the radiochromic films in the radiation range of interest (25 Gy). Surprisingly, it is found that, the red linear response of radiochromic film would be a suitable candidate for developing a dosimetry application, aligning the red optical density experimental result to the X-ray dosage through a linear-type mathematical equation. Figure 3 and Figure 4 relate to product test assemble before and after exposure to 20 Gy and 10 Gy radiation doses respectively.

[0060] Accordingly, the present application provides an accurate method of monitoring the radiation dose using a new dosimetry device expressed as linear response characteristic of optical density-red of radiation dose sensor with Gafchromic® film radiation within the 0.01 Gray (Gy) per second to about 2000 Gy (Gray) per second dose range using photographic densitometer.

[0061] Accordingly, the test assemblies were subjected to optical density measurements. Using X-Rite 301T photographic densitometer equipped with a reflection head assembly. The results will be measured as dv (density-visual), dr (density-red), dg (density-green) and db (density-blue). As the color changes from blue to red based on the amount of radiation passing through the test subject to radiation, there is consistent increase in the density-red values as measured using a densitometer.

[0062] In yet another important embodiment, the present invention provides an accurate method of monitoring the radiation dose using a new dosimetry device expressed as linear response characteristic of optical density-red of radiation dose sensor with Gafchromic® film radiation within the 0.01 Gray (Gy) per second to about 2000 (Gray) Gy per second dose range using smartphone selected from Apple iPhone. The cropped film images are provided in Figure 6.

[0063] In another embodiment, the present application provides a method of quantifying the dosage of radiation emitted from a radiation source, comprising the steps of: (i) exposing the dosimetry device to irradiation, said dosimetry device comprising radiation sensitive indicator which measures the radiation and visually represent the amount of radiation emitted as color change; (ii) capturing the color change of the dose indicator after exposure to radiation using an optical means; and (iii) comparing the optical density of dose indicator to a predetermined calibration curve developed using percentage optical density versus cumulative radiation dosage to quantify the dosage of radiation emitted from the source.

[0064] In another embodiment, dosimetry data received from the software means is uploaded into the cloud enabling access to a plurality of information networks and software management tools.

[0065] In another embodiment, the present application provides a dosimetry device as shown in the figures. Inventive dosimetry device with dose sensor is provided in Figure 2.

[0066] In another embodiment, the general work flow of the dosimetry device is as shown in Figure 7.

[0067] In another embodiment, the present application provides use of a dosimetry device comprising: a radiation dose indicator comprising a radiation sensitive composition to measure the radiation and visually represent the amount of radiation emitted as color change; (ii) an optical means to capture the color change of the dose indicator after exposure to radiation; and (iii) a software means to compare the optical density of the dose indicator to a predetermined calibration curve developed using percentage optical density versus cumulative radiation dosage to quantify the dosage of radiation emitted from the radiation source; wherein the dosimetry device is used in sterilization of surfaces and solutions, medical imaging, medical or industrial equipment quality assurance testing, UV light measurement, food processing and storage, transportation of radiation sensitive materials, or blood storage.

[0068] In another embodiment, the dosimetry device of the present application finds various uses in blood bags, their method of storage and use. Blood products are irradiated using photons to diminish the risk of transfusion-associated graft vs. host disease (TA-GVHD). The desired effect of irradiating the blood is to inhibit lymphocyte function and therefore to prevent GVHD while not causing damage to platelets and other blood fractions. The UK guidelines for blood irradiation state that all parts of the blood pack should receive at least 2500 cGy. The minimum dosage requirement for blood product are stated as: *“The dose of irradiation delivered (for blood processing) should be 2500 cGy (25 Gy) targeted to the central portion of the container and 1500 cGy (15 Gy) should be the minimum dose at any other point.”* From a regulatory perspective, the EU requires a dosimeter function for such an indicator product to possess a CE mark. For accuracy, all the EU regulation states is the following: per article 15.1 devices with measuring function should be designed and manufactured such way as to provide sufficient accuracy, precision, and stability for intended purpose. Limit of accuracy should be indicated by manufacturer. There are no specifics for degree of accuracy however, there are a couple of ISO standards that may help: ISO 51929:2013 blood irradiation dosimetry; ISO 52628:2013 standard practice for dosimeter in radiation processing.

[0069] Accordingly, the dosimetry device of the present application allows user to process the blood product, take a picture of indicator window after blood processing, software application will then process the picture and determine the dosage applied to the blood product. Stability is measured accurately until the shelf life is reached.

[0070] The current dosimetry device will continue with the same function as originally devised for without any complexities in using and training required for the end-users. The new indicator will provide data related to actual radiation level each blood bag received during processing. The inventive indicator shall enable users to develop new procedures, data and information with regards to their blood treatment processes.

[0071] Within the scope, the current radiation level indicator has the potential for image capture / retention, wireless data interfacing, dosimetry to simplify determination of proper blood treatment, result file sharing directly to hospital / treatment center IT systems.

[0072] The following, non-limiting examples illustrate important aspects of the present invention and do not in any way limit the scope of the invention.



## EXAMPLES

### [0073] Example 1:

#### [0074] **Comparative data of Radiochromic Films**

[0075] To demonstrate the potential for a dosimeter function, a variety of existing radiochromic film products were evaluated as a function of X-ray radiation exposure. The films or products evaluated are presented in Table 1.

[0076] Table 1: Products screened for Dosimetry Potential

<b>Product</b>	<b>Dose Range (Gy)</b>
Film #3	70
Film #2	145
Film #1	10 to 1000

### [0077] Example 2:

#### [0078] **Exposure to X-Ray Machine**

[0079] The films or products presented in Table 1 were exposed at 10, 20, 30, 40, 50 and 60 Gy in a Pantak® Unipolar Series 2 HF 160 X-ray machine. The X-ray cabinet was marked XRAD-160. This apparatus was equipped with a CNMC K602 precision dosimeter to determine the actual dosage applied to the films in Gray (Gy) units. In a separate series of film exposures, Gafchromic® film assemblies, mounted in similar fashion, were also tested. An example of this assembly is presented in Figure 3.

### [0080] Example 3:

#### [0081] **Optical Density Measurements**

[0082] The films or products were also evaluated for optical density. The equipment employed for measuring the optical density was the X-Rite 310T photographic densitometer equipped with a reflection head assembly (model 310-06). This equipment reported four measurement results: dv (density-visual), dr (density-red), dg (density-green) and db (density-blue). Table 2 presents the results of densitometer tests.

[0083] Table 2: Densitometry results for products screened for response to different X-ray exposure levels

Product	dv	dr	dg	db	Gy
Film #3 70 Gy	0.816	0.405	2.639	2.944	0
	1.586	1.235	2.653	2.860	10
	1.842	1.430	2.615	2.808	20
	1.983	1.497	2.816	2.999	30
	2.047	1.538	2.824	3.048	40
	2.065	1.549	2.754	3.052	50
	2.148	1.655	2.833	3.003	60
Film #2 145 Gy	0.834	0.454	2.557	2.874	0
	1.600	1.256	2.636	2.847	10
	1.857	1.429	2.726	2.923	20
	1.994	1.528	2.757	2.955	30
	2.024	1.518	2.734	2.937	40
	2.066	1.589	2.652	2.907	50
	2.135	1.655	2.755	2.924	60
Film #1	0.2375	0.201	0.301167	0.8035	0
	0.332	0.336	0.346	0.811	10
	0.401	0.415	0.359	0.777	20
	0.455	0.507	0.424	0.849	30
	0.485	0.549	0.409	0.778	40
	0.558	0.643	0.476	0.839	50
	0.599	0.696	0.501	0.855	60

dv-(density-violet) dr-(density-red) dg-(density-green) db-(density-blue)

[0084] In analyzing the densitometer results, the dr (density-red) data exhibits a linear response for Gafchromic® film #1 in this exposure range. A comparative plot for the dr results was presented in Figure 4. The red linear response of film #1 suggests that this film technology would be a suitable candidate for developing a dosimetry application, aligning the red optical density experimental result to the X-ray dosage through a linear-type mathematical equation.

[0085] **Example 4:**

[0086] **Radiation measured as color change and captured in the form of images**

[0087] A cell phone camera was employed to measure the color change in the radiochromic film. an Apple iPhone 5S (Model A1533) equipped with software version 11.2.1(15C153) was

employed to take pictures, in HDR mode or without flash (which removes the HDR mode). The cropped radiochromic film images from this exercise are presented in Figure 5.

[0088] Using the iPhone, pictures in HDR mode were captured. These images were cropped and saved in .jpg file format. To convert the color information in these images into RGB format, a software program called “Get RGB,” from the Imaging Technology Group (ITG) at the Beckman Institute at the University of Illinois-Urbana Champaign was employed. The results (via point selection, not averaged) show that there was minimal variation in this raw data and conversion of these raw data into RGB format is presented in Table 3.

[0089] Table 3: RGB data for jpg images in Figure 6

<b>Film #1</b>	<b>R</b>	<b>G</b>	<b>B</b>	<b>Gy</b>
	179	162	84	0
	158	163	96	10
	143	159	94	20
	125	152	97	30
	102	132	82	40
	100	137	86	50
	97	136	89	60

[0090] The red data from Table 3 exhibited a nearly linear response for the radiochromic film #1 in the radiation range of interest. This effect is more easily visualized in the graph presented in Figure 5. From the results presented in Figure 6, a linear calibration curve relating color to dose appears feasible.

[0091] Although the present invention is shown and described with respect to certain aspects, it is obvious that various modifications will become apparent to those skilled in the art upon reading and understanding the specification and the appended claims. The present invention includes all Such improvements and modifications and is limited only by the scope of the claims.

## We Claim:

1. A dosimetry device for quantifying the dosage of radiation emitted from a radiation source, the device comprising:
  - (i) a radiation dose indicator comprising a radiation sensitive composition to measure the radiation and visually represent the amount of radiation emitted as color change;
  - (ii) an optical means to capture the color change of the dose indicator after exposure to radiation; and
  - (iii) a software means to compare the optical density of the dose indicator as compared to a predetermined calibration curve developed using percentage optical density versus cumulative radiation dosage to quantify the dosage of radiation emitted from the radiation source.
2. The dosimetry device of claim 1, wherein the radiation sensitive composition is a radiation sensitive film prepared from polyacetylene, lithium, sodium, potassium or zinc salt of polyacetylene.
3. The dosimetry device of claim 2, wherein the radiation sensitive film has a radiation dose range of about 0.01 Gray (Gy) per second to about 10,000 Gray (Gy) per second.
4. The dosimetry device of claim 2, wherein the radiation sensitive film is a radiochromic film.
5. The dosimetry device of claim 1, wherein the radiation sensitive composition comprises a marker.
6. The dosimetry device of claim 5, wherein the marker is a dye or pigment.
7. The dosimetry device of claim 1, wherein the optical means to capture the color change of the indicator is selected from the group consisting of a smart phone camera, a high definition camera, a magnifying camera, a magnifying microscope, a densitometer, an image scanner, a video camera, a TV camera or an optical imaging device.
8. The dosimetry device of claim 1, wherein the software comprises a software program based on quantification algorithms to characterize the color change in the form of

color spaces, convert the identified color spaces into numerical dosimetry data set and compare the data to pre-determined color versus radiation dose response data.

9. The dosimetry device of claim 8, wherein the dosimetry data is up-loaded into the cloud enabling access to a plurality of information networks and software management tools.

10. The dosimetry device of claim 1, wherein the dosimetry device is regulatory compliant and used in safe storage of blood in blood bags.

11. Use of a dosimetry device comprising: (i) a radiation dose indicator comprising a radiation sensitive composition to measure the radiation emitted from a radiation source and visually represent the amount of radiation emitted as color change; (ii) an optical means to capture the color change of the dose indicator after exposure to radiation; and (iii) a software means to compare the optical density of the dose indicator to a predetermined calibration curve developed using percentage optical density versus cumulative radiation dosage to quantify the dosage of radiation emitted from the radiation source; wherein said dosimetry device is used in sterilization of surfaces and solutions, medical imaging, medical or industrial equipment quality assurance testing, UV light measurement, food processing and storage, transportation of radiation sensitive materials, or blood storage.

12. A method of quantifying the dosage of radiation emitted from a radiation source, comprising the steps of:

- (i) exposing a dosimetry device to irradiation, said dosimetry device comprising radiation sensitive indicator which measures the radiation and visually represents the amount of radiation emitted as color change;
- (ii) capturing the color change of the dose indicator after exposure to radiation using an optical means; and
- (iii) comparing the optical density of dose indicator to a predetermined calibration curve developed using percentage optical density versus cumulative radiation dosage to quantify the dosage of radiation emitted from the radiation source.

13. The method of claim 12, wherein the predetermined calibration curve is optical density-red of radiochromic film having radiation dose range of about 0.01 Gray (Gy) per second to about 2000 Gray (Gy) per second, and the optical means is a photographic densitometer.

14. The method of claim 12, wherein the predetermined calibration curve is optical density-red of radiochromic film having radiation dose range of about 0.01 Gray (Gy) per second to about 2000 Gray (Gy) per second, and the optical means is a smart phone.

15. A dosimetry device as shown in the Figures 1 to 7.





Fig. 1




Fig. 2

3/5




Fig. 3




Fig. 4

4/5




Fig. 5




Fig. 6




Fig. 7

INTERNATIONAL SEARCH REPORT

International application No.

PCT/US19/39811

A. CLASSIFICATION OF SUBJECT MATTER

IPC - G01T 1/02, 1/06, 1/08 (2019.01)

CPC - G01T 1/02, 1/06, 7/005

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

See Search History document

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

See Search History document

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)

See Search History document

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X --- Y --- A	US 8,604,415 B2 (MICKE, A et al.) 10 December 2013; column 4, line 61 through column 5, line 1, column 5, line 66 through column 6, line 9, column 8, lines 21-24, 38-43, column 24, lines 1-5, 17-21	1, 2, 4-6, 12 --- 10 --- 13, 14
X --- Y --- A	US 2008/0023647 A1 (PATEL, G) 31 January 2008; paragraphs [0049], [0065], [0068]-[0069], [0075], [0104]	1-3, 7, 11, 12 --- 8, 9 --- 13, 14
Y --- A	WO 2004/017095 A2 (PATEL, G) 26 February 2004; paragraphs [0071], [0081], figure 1	1-3, 7-9, 11, 12 --- 13, 14
Y	US 9,833,214 B2 (FUJIFILM CORPORATION) 05 December 2017; figure 18, column 18, lines 26-30, column 20, lines 61-63	8, 9
Y	US 8,340,249 B2 (KRAUS, T et al.) 25 December 2012; column 1, lines 10-12 and lines 49-58	10
A	US 5,637,876 A (DONAHUE, J et al.) 10 June 1997; column 3, lines 13-19	1-14
A	WO 2017/054008 A1 (MASSACHUSETTS INSTITUTE OF TECHNOLOGY) 30 March 2017; paragraph [0102]	1-14

Further documents are listed in the continuation of Box C.

See patent family annex.

* Special categories of cited documents:	"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention
"A" document defining the general state of the art which is not considered to be of particular relevance	"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone
"D" document cited by the applicant in the international application	"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art
"E" earlier application or patent but published on or after the international filing date	"&" document member of the same patent family
"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)	
"O" document referring to an oral disclosure, use, exhibition or other means	
"P" document published prior to the international filing date but later than the priority date claimed	

Date of the actual completion of the international search

03 September 2019 (03.09.2019)

Date of mailing of the international search report

23 SEP 2019

Name and mailing address of the ISA/US  
Mail Stop PCT, Attn: ISA/US, Commissioner for Patents  
P.O. Box 1450, Alexandria, Virginia 22313-1450  
Facsimile No. 571-273-8300

Authorized officer

Shane Thomas

Telephone No. PCT Helpdesk: 571-272-4300

INTERNATIONAL SEARCH REPORT

International application No.

PCT/US19/39811

**Box No. II Observations where certain claims were found unsearchable (Continuation of item 2 of first sheet)**

This international search report has not been established in respect of certain claims under Article 17(2)(a) for the following reasons:

1.  Claims Nos.:  
because they relate to subject matter not required to be searched by this Authority, namely:
  
2.  Claims Nos.: 15  
because they relate to parts of the international application that do not comply with the prescribed requirements to such an extent that no meaningful international search can be carried out, specifically:  
Claim 15 is an improper unsearchable omnibus type claim because it refers to a "device as shown in Figures 1 to 7".
  
3.  Claims Nos.:  
because they are dependent claims and are not drafted in accordance with the second and third sentences of Rule 6.4(a).

**Box No. III Observations where unity of invention is lacking (Continuation of item 3 of first sheet)**

This International Searching Authority found multiple inventions in this international application, as follows:

1.  As all required additional search fees were timely paid by the applicant, this international search report covers all searchable claims.
2.  As all searchable claims could be searched without effort justifying additional fees, this Authority did not invite payment of additional fees.
3.  As only some of the required additional search fees were timely paid by the applicant, this international search report covers only those claims for which fees were paid, specifically claims Nos.:
  
4.  No required additional search fees were timely paid by the applicant. Consequently, this international search report is restricted to the invention first mentioned in the claims; it is covered by claims Nos.:

- Remark on Protest**
- The additional search fees were accompanied by the applicant's protest and, where applicable, the payment of a protest fee.
  - The additional search fees were accompanied by the applicant's protest but the applicable protest fee was not paid within the time limit specified in the invitation.
  - No protest accompanied the payment of additional search fees.