

(51) International Patent Classification:
G01N 35/02 (2006.01)

(21) International Application Number:

PCT/US2012/033667

(22) International Filing Date:

13 April 2012 (13.04.2012)

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

61/476,175	15 April 2011 (15.04.2011)	US
61/476,167	15 April 2011 (15.04.2011)	US

(71) Applicant (for all designated States except US): **BEC-TON, DICKINSON AND COMPANY** [US/US]; 1 Bec-ton Drive, Franklin Lakes, NJ 07417 (US).

(72) Inventors; and

(75) Inventors/Applicants (for US only): **GUBATAYAO, Thomas, Catalino** [US/US]; 34808 Farragut Street, West-land, MI 48186 (US). **HANDIQUE, Kalyan** [IN/US]; 5570 Hampshire Lane, Ypsilanti, MI 48197 (US).**GANESAN, Karthik** [US/US]; 2484 Dundee Drive, Ann Arbor, MI 48103 (US). **DRUMMOND, Daniel, M.** [US/US]; 36101 Curtis Road, Livonia, MI 48152 (US).(74) Agent: **MALLON, Joseph, J.**; Knobbe Martens Olson & Bear LLP, 2040 Main Street, 14th Floor, Irvine, CA 92614 (US).

(81) Designated States (unless otherwise indicated, for every kind of national protection available): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IS, JP, KE, KG, KM, KN, KP, KR, KZ, LA, LC, LK, LR, LS, LT, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.

(84) Designated States (unless otherwise indicated, for every kind of regional protection available): ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, MD, RU,

[Continued on next page]

(54) Title: SCANNING REAL-TIME MICROFLUIDIC THERMO-CYCLER AND METHODS FOR SYNCHRONIZED THER-MOCYCLING AND SCANNING OPTICAL DETECTION




FIG. 1A




FIG. 1B

(57) Abstract: Systems and methods for performing simultaneous nucleic acid amplification and detection. The systems and methods comprise meth-ods for managing a plurality of protocols in conjunction with directing a sensor array across each of a plurality of reaction chambers. In certain em-bodiments, the protocols comprise thermocycling profiles and the methods may introduce offsets and duration extensions into the thermocycling pro-files to achieve more efficient detection behavior.



TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, ML, MR, NE, SN, TD, TG).

Declarations under Rule 4.17:

— *of inventorship (Rule 4.17(iv))*

Published:

— *with international search report (Art. 21(3))*

SCANNING REAL-TIME MICROFLUIDIC THERMOCYCLER AND METHODS FOR SYNCHRONIZED THERMOCYCLING AND SCANNING OPTICAL DETECTION

CROSS-REFERENCE TO RELATED APPLICATIONS

[0001] This application claims the benefit under 35 U.S.C. Section 119(e) of co-pending U.S. Provisional Patent Application Serial No. 61/476,175, filed on April 15, 2011, entitled “**SOFTWARE CONTROL PROCESS TO SYNCHRONIZE THERMOCYCLING AND SCANNING OPTICAL DETECTION**” and co-pending U.S. Provisional Patent Application Serial No. 61/476,167, filed on April 15, 2011, entitled “**6-COLOR SCANNING REAL-TIME MICROFLUIDIC THERMOCYCLER**” which applications are incorporated by reference herein in their entirety.

TECHNICAL FIELD

[0002] The systems and methods disclosed herein relate generally to the automated execution of nucleic acid amplification assays, such as Polymerase Chain Reaction (PCR), and in some instances real-time PCR, in a plurality of micro-fluidic reaction chambers in a microfluidic cartridge. The system may subsequently detect target nucleic acids, *e.g.*, target amplicons, within each of the reaction chambers.

BACKGROUND OF THE INVENTION

[0003] The medical diagnostics industry is a critical element of today's healthcare infrastructure. At present, however, *in vitro* diagnostic analyses, no matter how routine, have become a bottleneck in patient care. There are several reasons for this. First, many diagnostic analyses can only be done with highly specialized equipment that is both expensive and only operable by trained clinicians. Such equipment may be found in only a few locations---often just one in any given urban area. This requires hospitals to send out samples for analyses to these locations, thereby incurring shipping costs and transportation delays, and possibly even sample loss or mishandling. Second, the equipment in question is typically not available “on-demand” but instead runs in batches, thereby delaying the processing time for many samples as they must wait for a machine to reach capacity before they can be run.

[0004] Understanding that diagnostic assays on biological samples may break down into several key steps, it is often desirable to automate one or more steps. For example, a biological sample, such as those obtained from a patient, can be used in nucleic acid amplification assays, in order to amplify a target nucleic acid (*e.g.*, DNA, RNA, or the like) of

interest. Once amplified, the presence of a target nucleic acid, or amplification product of a target nucleic acid (*e.g.*, a target amplicon) reactor can be detected, wherein the presence of a target nucleic acid and/or target amplicon is used to identify and/or quantify the presence of a target (*e.g.*, a target microorganism or the like). Often, nucleic acid amplification assays involve multiple steps, which can include nucleic acid extraction, nucleic acid amplification, and detection. It is desirable to automate certain steps of these processes.

[0005] There is a need for a method and apparatus for carrying out molecular diagnostic assays on multiple samples in parallel, with or without amplification of target nucleic acids, and detection on a prepared biological samples. The system may be configured for high throughput, and operation in a commercial reference laboratory or at the point of care, thereby eliminating the need to send the sample out to a specialized facility.

SUMMARY OF THE INVENTION

[0006] The embodiments disclosed herein relate to methods and devices for the simultaneous testing of multiple samples. Certain embodiments contemplate an apparatus for performing real-time nucleic acid amplification and detection. The apparatus can include a detector head comprising a plurality of photodetector and light source pairs. The detector head can be mounted on a rail, wherein the detector and light source pairs are aligned in a first row and a second row. The apparatus can include a receptacle for a microfluidic cartridge that has a plurality of independent reaction chambers aligned in adjacent columns of a first row and a second row. The apparatus can also include an aperture plate that is configured to be positioned over the microfluidic cartridge when the cartridge is present in the receptacle. The aperture plate can include a plurality of apertures that are each aligned over each of the plurality of reaction chambers when the receptacle is holding the microfluidic cartridge. The detector head can be located over the aperture plate, and be moveable along the rail, such that each of the plurality of photodetector and light source pairs in the first row can be positioned over each aperture in the first row of the aperture plate, and each of the plurality of photodetector and light source pairs in the second row can be positioned over each aperture in the second row of the aperture plate.

[0007] In some embodiments, the apparatus also includes a second detector head that has a plurality of photodetector and light source pairs aligned into a first row and a second row. The second detector head can be mounted on the rail. The apparatus can also include a second receptacle for a microfluidic cartridge including a plurality of independent reaction chambers aligned in adjacent columns of a first row and a second row. The apparatus can also include a second aperture plate configured to be positioned over the second microfluidic cartridge when the second cartridge is present in the second receptacle, and which can include a plurality of

apertures that are each aligned over each of the plurality of reaction chambers of the second microfluidic cartridge when the second receptacle is holding the second microfluidic cartridge. The second detector head can be located over the aperture plate, and can be moveable along the rail such that each of the plurality of photodetector and light source pairs in the first row of the second detector head can be positioned over each aperture in the first row of the second aperture plate, and each of the plurality of photodetector and light source pairs in the second row of the second detector head can be positioned over each aperture in the second row of the second aperture plate.

[0008] In some embodiments, the photodetector and light source pairs can include at least six different photodetector and light source pairs operating in six different wavelengths. In some embodiments, the six different wavelengths comprise a light source emitting a green colored light, a light source emitting a yellow colored light, a light source emitting an orange colored light, a light source emitting a red colored light, and a light source emitting a crimson colored light. In some embodiments, the detector head includes at least N rows of photodetector and light source pairs, and the detector is configured to move to at least $M + N - 1$ positions over an aperture plate comprising M rows of apertures.

[0009] In some embodiments, the aperture plate comprises steel, aluminum, nickel, or a combination thereof. In some embodiments, the aperture plate can have a thickness of approximately .25 inches. In some embodiments, at least part of the aperture plate is electrochemically oxidized to be darker than when the aperture plate is not electrochemically oxidized. In some embodiments, the aperture plate provides substantially uniform pressure across the area of the microfluidic cartridge, when the cartridge is present within the receptacle. In some embodiments, the aperture plate comprises at least one of aluminum, zinc or nickel, the aperture plate further comprising a colorant.

[0010] In some embodiments, the apparatus further comprises a heater plate, wherein the heater plate is positioned underneath the microfluidic cartridge when a cartridge is present in the receptacle. In some embodiments the heater plate comprises at least one of glass or quartz. In some embodiments, the aperture plate provides substantially uniform pressure across the area of the microfluidic cartridge when a cartridge is present within the receptacle. The substantially uniform pressure can facilitate substantially uniform thermal contact between the microfluidic reaction chambers and the heater plate. As such, in some embodiments, the aperture plate provide uniform pressure that can ensure that each of the plurality of reaction chambers or reactors in the microfluidic cartridge are in uniformly thermal contact or communication with a respective a plurality of heating elements located within the heater plate.

[0011] In some embodiments, the apparatus further comprises a photodetector, the photodetector located over the aperture plate, wherein the micro-fluidic chamber is configured to receive light at a glancing angle from a light source relative to the photodetector. In some embodiments, the heater plate comprises a plurality of heating elements, wherein each of the plurality of heating elements is positioned such that when the microfluidic cartridge is present in the receptacle, the plurality of heating elements are in thermal connection with each of the plurality of reaction chambers, respectively.

[0012] Certain embodiments contemplate a method implemented on one or more computer processors for optimizing protocols, such as polymerase chain reaction (PCR) protocols or the like, for simultaneously performing a plurality of thermal cycling reactions, wherein each thermal cycling reaction comprises one or more detection steps, and wherein the thermal cycling reactions are performed in a plurality of reactors. The method can include the steps of determining or providing or accessing a detection cycle time for each of the plurality of reactors; receiving or accessing a protocol step, the step associated with a step duration, the step comprising a time for detection; and determining a first adjustment to the step such that the step duration is a multiple of the detection cycle time.

[0013] In some embodiments the method further comprises determining a second adjustment to the step, wherein the time for detection is a multiple of the detection cycle time when the step is adjusted by the first adjustment and by the second adjustment. In some embodiments the method further comprises determining a starting offset adjustment based on a position of a reaction chamber associated with the protocol. In some embodiments, the detection cycle time comprises the amount of time required for a detector head to perform a predetermined plurality of detections for a reactor. In some embodiments, the detection cycle time includes a time required for movement of the detector head to each of a plurality of reactors and movement of the detector head to the start position. In some embodiments, the method further comprises initiating the protocol.

[0014] Certain embodiments contemplate a non-transitory computer-readable medium comprising instructions, the instructions configured to cause one or more processors to perform the following steps: determining or providing or accessing a detection cycle time; receiving or accessing a protocol step, wherein the step is associated with a step duration, and the wherein step includes a time for detection; and determining a first adjustment to the step such that the step duration is a multiple of the detection cycle time.

[0015] In some embodiments, the protocol step is associated with a protocol from a plurality of protocols. Each of the plurality of protocols can be associated with at least one of a plurality of thermal cycling reactions, such as polymerase chain reaction (PCR) protocols,

wherein each thermal cycling reaction comprises one or more detection steps, and wherein the determining a first adjustment is based at least in part on a timing of one or more detection steps associated with the thermal cycling reactions of at least two or more of the plurality of protocols when the two or more of the plurality of protocols are simultaneously run. In some embodiments, the method also includes the step of determining a second adjustment to the step, wherein the time for detection is a multiple of the detection cycle time when the step is adjusted by the first adjustment and by the second adjustment. In some embodiments, the method also includes the step of determining a starting offset adjustment based on a position of a reaction chamber associated with the protocol. In some embodiments, the detection cycle time includes the amount of time required for a detector head to perform a predetermined plurality of detections for a reaction chamber. In some embodiments, the detection cycle time also includes a time required for movement of the detector head to each of a plurality of reaction chamber detection positions and movement of the detector head to a start position. In some embodiments, the method further comprises initiating the protocol.

[0016] Certain embodiments contemplate a system for optimizing protocols for a plurality of reaction chambers. The system can include a processor configured to perform the following: determining or providing or accessing a detection cycle time; receiving or accessing a protocol step, wherein the step can be associated with a step duration, and wherein the step includes a time for detection; and determining a first adjustment to the step such that the step duration is a multiple of the detection cycle time.

[0017] In some embodiments, the protocol step is associated with a protocol from a plurality of protocols. Each of the plurality of protocols can be associated with at least one of a plurality of thermal cycling reactions, such as a polymerase chain reaction (PCR) protocol, wherein each thermal cycling reaction comprises one or more detection steps, and wherein the determining a first adjustment is based at least in part on a timing of one or more detection steps associated with the thermal cycling reactions of at least two or more of the plurality of protocols when the two or more of the plurality of protocols are simultaneously run. In some embodiments, the processor is also configured to determine a second adjustment to the step, wherein the time for detection is a multiple of the detection cycle time when the step is adjusted by the first adjustment and by the second adjustment. In some embodiments, the processor is also configured to determine a starting offset adjustment based on a position of a reaction chamber associated with the protocol. In some embodiments, the detection cycle time includes the amount of time required for a detector head to perform a predetermined plurality of detections for a reaction chamber. In some embodiments, the detection cycle time also includes a time required for movement of the detector head to each of a plurality of reaction chamber

detection positions and movement of the detector head to the start position. In some embodiments, the processor is further configured to initiate the protocol.

[0018] Certain embodiments contemplate a method for simultaneously performing real-time PCR in a plurality of PCR reaction chambers, comprising: (a) providing a scan time sufficient for a detector assembly to perform a scan cycle during which it can scan each of the plurality of PCR reaction chambers for at least one detectable signal and become ready to repeat the scan; (b) providing a reaction protocol for each of the PCR reaction chambers that includes multiple cycles, each cycle comprising a cycle time that includes at least one heating step, at least one cooling step, and at least one temperature plateau that includes a reading cycle period during which the detector assembly is to scan the reaction chamber for at least one detectable signal; (c) determining, using a processor, whether the cycle time for that reaction chamber is the same as or an integer multiple of the scan time, and if not, adjusting the scan time or the cycle time so that the cycle time is the same as or an integer multiple of the scan time; (d) performing at least steps (b) and (c) for the reaction protocol for each of the plurality of PCR reaction chambers so that the cycle time for each reaction protocol is the same as or an integer multiple of the scan time; and (e) under direction of a processor, performing real time PCR on each of the reaction chambers using the reaction protocol for each of the reaction chambers, including performing multiple scan cycles with the detector assembly, wherein each PCR reaction chamber is scanned by the detector assembly during each reading cycle period for that reaction chamber.

[0019] In some embodiments the method further comprises phase adjusting the cycle time of the reaction protocol for at least one of the reaction chambers. In some embodiments, at least one said reaction protocol is different from another said reaction protocol. In some embodiments, at least one cycle time in one reaction protocol is different from the cycle time in another reaction protocol. .

BRIEF DESCRIPTION OF THE DRAWINGS

[0020] Figure 1A is a front plan view of a diagnostic apparatus as used in certain of the embodiments.

[0021] Figure 1B is a top perspective view of the diagnostic apparatus of Figure 1A showing certain of the apparatus' internal components.

[0022] Figure 2 illustrates an interior view of the diagnostic apparatus of Figures 1A and 1B.

[0023] Figure 3A illustrates a top-plan view of one possible microfluidic arrangement within certain embodiments of a microfluidic cartridge as described herein.

[0024] Figure 3B illustrates the layout of a heater substrate in relation to the reaction chamber of certain of the embodiments.

[0025] Figure 4A illustrates an exterior view of the optical module including the detector head of certain of the embodiments described herein.

[0026] Figure 4B illustrates a view of the optical module of Figure 4A with a side cover removed.

[0027] Figure 4C illustrates a bottom view of the optical module of Figure 4A.

[0028] Figure 5 illustrates a detector head used within the optical module of certain of the embodiments along the line 13 of Figure 4B.

[0029] Figure 6 depicts the layout of the light sources and optical detectors as used in certain of the embodiments of the detector head disclosed herein.

[0030] Figure 7 is a graph of the fluorescence versus time of using real time PCR of target nucleic acids performed in an apparatus of certain embodiments as described herein.

[0031] Figure 8 is an abstract depiction of certain of the chamber, aperture, and heating layers found in certain of the embodiments as described herein.

[0032] Figures 9A-H illustrate various perspectives of one embodiment of the aperture plate.

[0033] Figure 10 illustrates various dimensions of the perspectives of the aperture plate of Figures 9A-H.

[0034] Figure 11 is plot of a portion of a thermal profile for a possible protocol implemented in certain of the embodiments.

[0035] Figure 12 is a flow diagram depicting a process for determining protocol durations, offsets, and detection times, so as to optimize and regiment detector efficiency.

[0036] Figure 13 illustrates a portion of a user interface for selecting durations of certain protocol steps and substeps and determination of the accompanying intra-cycle adjustment.

[0037] Figure 14 is a plot of a thermal profile comprising an inter-cycle adjustment.

[0038] Figures 15A-C plot a plurality of thermal profiles for a plurality of protocols implemented in certain of the embodiments. Figures 15A and 15B illustrate the character of the protocol profiles prior to the starting offset adjustment. Figure 15C illustrates the plurality of protocol profiles relative to one another after applying the starting offset adjustments.

[0039] Figure 16 is plot of a thermal profile under active cooling as implemented in certain of the embodiments.

DETAILED DESCRIPTION

[0040] Certain of the present embodiments contemplate an apparatus, referred to herein as a thermocycler, which may consistently heat and analyze microfluidic chambers. Polynucleotide amplification, such as by real-time PCR, can be performed within the microfluidic chambers. In some embodiments, the thermocycler can be configured to perform individual thermocycling and detection protocols in a plurality of microfluidic reaction chambers within a microfluidic cartridge. The thermocycling can be used to amplify nucleic acids, *e.g.*, DNA, RNA or the like, *e.g.*, by real-time PCR or other nucleic acid amplification protocols described herein, within the microfluidic reaction chambers. The thermocycler may comprise a detector head, comprising a plurality of detector pairs, *e.g.*, six or more detector head pairs, wherein each detector pair comprises a light-emitting source, *e.g.*, an LED or the like, and a cognate photodiode. In some embodiments, each individual detector pair is configured to generate and detect light emitted from a fluorescent moiety, *e.g.*, a fluorescent probe, to indicate the presence of a target polynucleotide.

[0041] As used herein, the term “microfluidic” refers to volumes of less than 1 ml, preferably less than 0.9 ml, *e.g.*, 0.8 ml, 0.7 ml, 0.6 ml, 0.5 ml, 0.4 ml, 0.3 ml, 0.2 ml, 0.1 ml, 90 μ l, 80 μ l, 70 μ l, 60 μ l, 50 μ l, 40 μ l, 30 μ l, 20 μ l, 10 μ l, 5 μ l, 4 μ l, 3 μ l, 2 μ l, 1 μ l, or less, or any amount in between. It is to be understood that, unless specifically made clear to the contrary, where the term PCR is used herein, any variant of PCR including but not limited to real-time and quantitative PCR, and any other form of polynucleotide amplification is intended to be encompassed.

[0042] The detection process used in the assay may also be multiplexed to permit multiple concurrent measurements on multiple reactions concurrently. In some embodiments, these measurements may be taken from separate reaction chambers. Certain of these embodiments perform a plurality of PCR reactions simultaneously in a single PCR reaction chamber, *e.g.*, multiplex PCR. A PCR protocol may comprise guidelines for performing the successive annealing and denaturing of the polynucleotides in the reaction chamber prior to detection. Such guidelines, comprising a time profile for heating the chamber, may be referred to as a “protocol”. Certain of the disclosed embodiments facilitate consistent heating and/or cooling across a plurality of reaction chambers performing PCR, while facilitating detection using a sensor array. In certain embodiments, the apparatus may comprise an aperture plate which facilitates consistent heating and cooling of the reaction chambers by applying pressure to a cartridge containing a plurality of PCR reaction chambers. Certain details and methods for processing polynucleotides may be found in *e.g.*, U.S. Patent Application Publication 2009-

0131650 and U.S. Patent Application Publication 2010-0009351, incorporated herein by reference.

[0043] The skilled artisan will appreciate that the embodiments disclosed herein are useful for various types of nucleic acid amplification reactions. For example, methods of nucleic acid amplification in connection with the embodiments disclosed herein can include, but are not limited to: polymerase chain reaction (PCR), strand displacement amplification (SDA), for example multiple displacement amplification (MDA), loop-mediated isothermal amplification (LAMP), ligase chain reaction (LCR), immuno-amplification, and a variety of transcription-based amplification procedures, including transcription-mediated amplification (TMA), nucleic acid sequence based amplification (NASBA), self-sustained sequence replication (3SR), and rolling circle amplification. See, e.g., Mullis, "Process for Amplifying, Detecting, and/or Cloning Nucleic Acid Sequences," U.S. Pat. No. 4,683,195; Walker, "Strand Displacement Amplification," U.S. Pat. No. 5,455,166; Dean et al, "Multiple displacement amplification," U.S. Pat. No. 6,977,148; Notomi et al., "Process for Synthesizing Nucleic Acid," U.S. Pat. No. 6,410,278; Landegren et al. U.S. Pat. No. 4,988,617 "Method of detecting a nucleotide change in nucleic acids"; Birkenmeyer, "Amplification of Target Nucleic Acids Using Gap Filling Ligase Chain Reaction," U.S. Pat. No. 5,427,930; Cashman, "Blocked-Polymerase Polynucleotide Immunoassay Method and Kit," U.S. Pat. No. 5,849,478; Kacian et al., "Nucleic Acid Sequence Amplification Methods," U.S. Pat. No. 5,399,491; Malek et al., "Enhanced Nucleic Acid Amplification Process," U.S. Pat. No. 5,130,238; Lizardi et al., *BioTechnology*, 6:1197 (1988); Lizardi et al., U.S. Pat. No. 5,854,033 "Rolling circle replication reporter systems."

[0044] In some embodiments disclosed herein, the target nucleic acid, *e.g.*, target amplicon, can be detected using an oligonucleotide probe. Preferably, the probes include one or more detectable moieties that can be detected by the systems disclosed herein. The skilled artisan will appreciate that several probe technologies are useful in the embodiments described herein. By way of example, the embodiments disclosed herein can be used with TAQMAN® probes, molecular beacon probes, SCORPION™ probes, and the like.

[0045] TaqMan® assays are homogenous assays for detecting polynucleotides (see U.S. Pat. No. 5,723,591). In TAQMAN® assays, two PCR primers flank a central TAQMAN® probe oligonucleotide. The probe oligonucleotide contains a fluorophore and quencher. During the polymerization step of the PCR process, the 5' nuclease activity of the polymerase cleaves the probe oligonucleotide, causing the fluorophore moiety to become physically separated from the quencher, which increases fluorescence emission. As more PCR product is created, the intensity of emission at the novel wavelength increases.

[0046] Molecular beacons are an alternative to TAQMAN® probes for the detection of polynucleotides, and are described in, *e.g.*, U.S. Pat. Nos. 6,277,607; 6,150,097; and 6,037,130. Molecular beacons are oligonucleotide hairpins which undergo a conformational change upon binding to a perfectly matched template. The conformational change of the oligonucleotide increases the physical distance between a fluorophore moiety and a quencher moiety present on the oligonucleotide. This increase in physical distance causes the effect of the quencher to be diminished, thus increasing the signal derived from the fluorophore.

[0047] The adjacent probes method amplifies the target sequence by polymerase chain reaction in the presence of two nucleic acid probes that hybridize to adjacent regions of the target sequence, one of the probes being labeled with an acceptor fluorophore and the other probe labeled with a donor fluorophore of a fluorescence energy transfer pair. Upon hybridization of the two probes with the target sequence, the donor fluorophore interacts with the acceptor fluorophore to generate a detectable signal. The sample is then excited with light at a wavelength absorbed by the donor fluorophore and the fluorescent emission from the fluorescence energy transfer pair is detected for the determination of that target amount. U.S. Pat. No. 6,174,670 discloses such methods.

[0048] Sunrise primers utilize a hairpin structure similar to molecular beacons, but attached to a target binding sequence which serves as a primer. When the primer's complementary strand is synthesized, the hairpin structure is disrupted, thereby eliminating quenching. These primers detect amplified product and do not require the use of a polymerase with a 5' exonuclease activity. Sunrise primers are described by Nazarenko et al. (Nucleic Acids Res. 25:2516-21 (1997) and in U.S. Pat. No. 5,866,336.

[0049] SCORPION™ probes combine a primer with an added hairpin structure, similar to Sunrise primers. However, the hairpin structure of SCORPION™ probes is not opened by synthesis of the complementary strand, but by hybridization of part of the hairpin structure with a portion of the target which is downstream from the portion which hybridizes to the primer.

[0050] DzyNA-PCR involves a primer containing the antisense sequence of a DNAzyme, an oligonucleotide capable of cleaving specific RNA phosphodiester bonds. The primer binds to a target sequence and drives an amplification reaction producing an amplicon which contains the active DNAzyme. The active DNAzyme then cleaves a generic reporter substrate in the reaction mixture. The reporter substrate contains a fluorophore-quencher pair, and cleavage of the substrate produces a fluorescence signal which increases with the amplification of the target sequence. DNAzy-PCR is described in Todd et al., Clin. Chem. 46:625-30 (2000), and in U.S. Pat. No. 6,140,055.

[0051] Fiandaca et al. describes a fluorogenic method for PCR analysis utilizing a quencher-labeled peptide nucleic acid (Q-PNA) probe and a fluorophore-labeled oligonucleotide primer. Fiandaca et al. *Genome Research*. 11:609-613 (2001). The Q-PNA hybridizes to a tag sequence at the 5' end of the primer.

[0052] Li et al. describes a double stranded probe having a quencher and fluorophore on opposite oligonucleotide strands. Li et al. *Nucleic Acids Research*. 30(2): e5, 1-9 (2002). When not bound to the target, the strands hybridize to each other and the probe is quenched. However, when a target is present at least one strand hybridizes to the target resulting in a fluorescent signal.

[0053] Fluorophore labels and moieties useful in the embodiments disclosed herein include, but are not limited to, dyes of the fluorescein family, the carboxyrhodamine family, the cyanine family, and the rhodamine family. Other families of dyes that can be used in the invention include, e.g., polyhalofluorescein-family dyes, hexachlorofluorescein-family dyes, coumarin-family dyes, oxazine-family dyes, thiazine-family dyes, squaraine-family dyes, chelated lanthanide-family dyes, the family of dyes available under the trade designation Alexa Fluor J, from Molecular Probes, and the family of dyes available under the trade designation Bodipy J, from Invitrogen (Carlsbad, Calif.). Dyes of the fluorescein family include, e.g., 6-carboxyfluorescein (FAM), 2',4',1,4,-tetrachlorofluorescein (TET), 2',4',5',7',1,4-hexachlorofluorescein (HEX), 2',7'-dimethoxy-4',5'-dichloro-6-carboxyrhodamine (JOE), 2'-chloro-5'-fluoro-7',8'-fused phenyl-1,4-dichloro-6-carboxyfluorescein (NED), 2'-chloro-7'-phenyl-1,4-dichloro-6-carboxyfluorescein (VIC), 6-carboxy-X-rhodamine (ROX), and 2',4',5',7'-tetrachloro-5-carboxy-fluorescein (ZOE). Dyes of the carboxyrhodamine family include tetramethyl-6-carboxyrhodamine (TAMRA), tetrapropano-6-carboxyrhodamine (ROX), Texas Red, R110, and R6G. Dyes of the cyanine family include Cy2, Cy3, Cy3.5, Cy5, Cy5.5, and Cy7. Fluorophores are readily available commercially from, for instance, Perkin-Elmer (Foster City, Calif.), Molecular Probes, Inc. (Eugene, Oreg.), and Amersham GE Healthcare (Piscataway, N.J.).

[0054] As discussed above, in some embodiments, the probes useful in the embodiments disclosed herein can comprise a quencher. Quenchers may be fluorescent quenchers or non-fluorescent quenchers. Fluorescent quenchers include, but are not limited to, TAMRA, ROX, DABCYL, DABSYL, cyanine dyes including nitrothiazole blue (NTB), anthraquinone, malachite green, nitrothiazole, and nitroimidazole compounds. Exemplary non-fluorescent quenchers that dissipate energy absorbed from a fluorophore include those available under the trade designation Black Hole™ from Biosearch Technologies, Inc. (Novato, Calif.), those available under the trade designation Eclipse™. Dark, from Epoch Biosciences (Bothell,

Wash.), those available under the trade designation Qx1J, from Anaspec, Inc. (San Jose, Calif.), and those available under the trade designation Iowa Black™ from Integrated DNA Technologies (Coralville, Iowa).

[0055] In some embodiments discussed above, a fluorophore and a quencher are used together, and may be on the same or different oligonucleotides. When paired together, a fluorophore and fluorescent quencher can be referred to as a donor fluorophore and acceptor fluorophore, respectively. A number of convenient fluorophore/quencher pairs are known in the art (see, for example, Glazer et al, Current Opinion in Biotechnology, 1997;8:94-102; Tyagi et al., 1998, Nat. Biotechnol., 16:49-53) and are readily available commercially from, for instance, Molecular Probes (Junction City, Oreg.), and Applied Biosystems (Foster City, Calif.). Examples of donor fluorophores that can be used with various acceptor fluorophores include, but are not limited to, fluorescein, Lucifer Yellow, B-phycoerythrin, 9-acridineisothiocyanate, Lucifer Yellow VS, 4-acetamido-4'-isothio-cyanatostilbene-2,2'-disulfonic acid, 7-diethylamino-3-(4'-isothiocyanatophenyl)-4-methylcoumarin, succinimdy 1-pyrenebutyrate, and 4-acetamido-4'-isothiocyanatostilbene-2,2'-disulfonic acid derivatives. Acceptor fluorophores typically depend upon the donor fluorophore used. Examples of acceptor fluorophores include, but are not limited to, LC-Red 640, LC-Red 705, Cy5, Cy5.5, Lissamine rhodamine B sulfonyl chloride, tetramethyl rhodamine isothiocyanate, rhodamine x isothiocyanate, erythrosine isothiocyanate, fluorescein, diethylenetriamine pentaacetate or other chelates of Lanthanide ions (e.g., Europium, or Terbium). Donor and acceptor fluorophores are readily available commercially from, for instance, Molecular Probes or Sigma Chemical Co. (St. Louis, Mo.). Fluorophore/quencher pairs useful in the compositions and methods disclosed herein are well-known in the art, and can be found, e.g., described in S. Marras, "Selection of Fluorophore and Quencher Pairs for Fluorescent Nucleic Acid Hybridization Probes" available at the world wide web site molecular-beacons.org/download/marras,mmb06%28335%293.pdf (as of April 11, 2012).

[0056] The detection process used in the assays disclosed herein advantageously permits multiple concurrent measurements of multiple detectable moieties, e.g., a plurality of probes containing different detectable moieties, etc. In some embodiments, these measurements may be taken from separate reaction chambers within a microfluidic cartridge, e.g., comprising a chamber layer (the chamber layer referring herein to that portion of the microfluidic cartridge containing the reaction chambers). Certain of these embodiments perform a plurality of amplification reactions simultaneously in a single reaction chamber, e.g., multiplex PCR. A PCR protocol may comprise guidelines for performing the successive annealing and denaturing of the polynucleotides in the reaction chamber prior to detection. In certain embodiments, the

apparatus is configured to facilitate consistent heating and/or cooling across a plurality of reaction chambers to perform nucleic acid amplification, and to facilitate detection of target amplicons in individual reaction chambers, *e.g.*, by detecting fluorescent emissions, using a sensor array.

[0057] In certain embodiments, the apparatus may comprise an aperture plate which facilitates consistent heating and cooling of the reaction chambers by applying pressure to a cartridge containing a plurality of reaction chambers via multiple, independent optical pairs. The aperture plate is preferably configured to enable and facilitate the generation and detection of fluorescent signals from probes within multiple, independent reaction chambers. In some embodiments, the aperture plate is configured such that there is an individual aperture (or windows), positioned over each of the individual reaction chambers in the microfluidic cartridge.

Diagnostic Apparatus

[0058] Figures 1A and 1B show a diagnostic apparatus 10 of certain of the present embodiments. In the embodiment illustrated in Figure 1A, the diagnostic apparatus includes an apparatus housing 30. The housing 30 may ensure a controlled environment for processing of the microfluidic samples and for preventing undesirable light from entering the detection space. The housing 30 may comprise a cover 16 which includes a handle 14 and a translucent window 12. The cover 16 may be brought down to close the opening in the front of the diagnostic apparatus 10 when the diagnostic apparatus 10 is in operation.

[0059] As seen in the embodiments of Figures 1A and 1B, the diagnostic apparatus 10 may house two specimen racks 24a, 24b in the front portion of the diagnostic apparatus 10. The skilled artisan will appreciate, however, that the depiction of the diagnostic apparatus in Figures 1A and 1B is exemplary only, and that in some embodiments, the apparatus can be configured to house more than two specimen racks, *e.g.*, three, four, five, six, seven, eight, nine, ten, or more specimen racks. Preferably, the apparatus is configured to house the same number of specimen racks, *e.g.*, two, as microfluidic cartridges.

[0060] In some embodiments, each specimen rack 24a, 24b may include multiple holders 26. The holders 26 may include receptacles for holding diagnostic reagents, such as reagents for nucleic acid amplification, *e.g.*, PCR reagents or the like. The racks 24 may also include specimen tubes (not shown) and mixing tubes (not shown) for preparing diagnostic-ready samples, such as amplification-ready samples. The apparatus may prepare the desired reagents in the racks 24a, 24b using the dispenser 400. Further description of various fluid

dispensers may be found in e.g., U.S. Patent Application Publication 2009-0130719 and U.S. Patent Application Publication 2009-0155123, incorporated herein by reference.

[0061] In some embodiments, the reaction chambers within the microfluidic cartridge(s) includes one or more reagents, buffers, etc., used in the nucleic amplification assay. For example, in some embodiments, the reaction chambers of the microfluidic cartridge can include, e.g., amplification primers, probes, nucleotides, enzymes such as polymerase, buffering agents, or the like. By way of example, in some embodiments, the reaction chambers can include lyophilized reagents, to which processed biological sample (e.g., a solution of extracted nucleic acids) is added. The prepared fluids may then be transferred to a microfluidic cartridge and be inserted into heater/optical modules 500a, 500b for processing and analysis.

[0062] Figure 1A is a front plan view of the diagnostic apparatus 10 of certain of the embodiments. As seen in Figure 1A, the diagnostic apparatus 10 can include a fluid dispenser 400, mounted on a lateral rail 20. The lateral rail 20 may be part of a motor-driven gantry 18, which may also include a fore-aft rail 22 (not shown). The fore-aft rail 22 may be connected to the lateral rail 20 and mounted perpendicularly to the lateral rail 20 in the diagnostic apparatus 10.

[0063] Figure 1A further illustrates the cover 28 over the heater/optical modules 500a, 500b. Receiving trays 520a and 520b may be located beneath or within the housing of the heater/optical modules 500a, 500b. Receiving tray 520a is illustrated in an open position, making it available to receive a microfluidic cartridge 200. Receiving tray 520b is illustrated in a closed position. Closing the tray not only places the reagents in the appropriate position for processing, but also further protects the interior of the heater/optical modules from receiving any unwanted stray light. Were stray light introduced into the detection area, the system may identify erroneous fluorescent levels derived from light which is not emitted from the reaction chamber.

[0064] Figure 1B is a perspective view of the diagnostic apparatus 10 showing certain of the internal components found in certain of the embodiments. To better illustrate certain features, the apparatus housing 30, the cover 16, and the heater/optical cover 28 found in Figure 1A have been removed from view in Figure 1B. Shown in Figure 1B is the gantry 18, including the lateral rail 20 fore-aft rail 22. The fluid dispenser 400 may be mounted on the lateral rail 20 and may slide laterally along the long lateral rail 20. The lateral rail 20 may be connected to the fore-aft rail 22 which may move in the fore-aft direction. In this manner the fluid dispenser 400 is available to move in the X, Y direction throughout the diagnostic device 10. As described below, the fluid dispenser 400 may also able to move up and down in the z-

plane on the lateral rail 20, thereby giving the dispenser 400 the ability to move in three directional degrees throughout the diagnostic device 10.

[0065] Also shown in Figure 1B are the heater/optical modules 500a, 500b with the cover 28 of the heater/optical modules of Figure 1A removed. The receiving trays 520a and 520b are depicted in the open position and are each holding cartridges 200. In some embodiments, the receiving trays may each include a heater substrate 600 (not shown) beneath each of the microfluidic cartridges 200. The heater/optical modules 500a, 500b may also each include a detector head 700 described in greater detail below.

[0066] As will be described in more detail below, the diagnostic apparatus 10 may be capable of conducting real-time diagnostics on one or more samples. The sample to be tested may first be placed in a specimen tube (not shown) on the rack 24a or 24b. Diagnostic reagents may be located in the holders 26 on the rack 24a inside the diagnostic apparatus 10. The fluid dispenser 400 may mix and prepare the sample for diagnostic testing and may then deliver the prepared sample to the microfluidic cartridge 200 for thermal cycling and analyte detection in the heater/optical modules 500a, 500b. Alternatively, the fluid dispenser 400 may deliver nucleic acid samples to the reaction chambers of the microfluidic cartridge, wherein the reaction chambers of the microfluidic cartridge already contain reagents for an amplification reaction.

[0067] FIGURE 2 illustrates an interior view of the diagnostic apparatus 10, showing the rack 24a holding a number of sample tubes 32 and reagent holders 26, and a cartridge 200 situated in the receiving tray 520a. The receiving tray 520a is in an open position extending from the heater/optical module 500a which has the cover 28 attached. The receiving tray 520b is in a closed position. Advantageously, in some embodiments the receiving trays 520a, b may allow easy placement of the microfluidic cartridge 200, by a user or by an auto-loading device. Such a design may also accommodate multiplexed pipetting of samples using the robotic fluid dispenser 400.

Receiving Tray

[0068] As illustrated in Figure 2, the recessed bay 524 can be a portion of the receiving tray 520 that is configured to selectively receive the microfluidic cartridge 200. For example, the recessed bay 524 and the microfluidic cartridge 200 can have an edge 526 which is complementary in shape so that the microfluidic cartridge 200 is selectively received in, e.g., a single orientation. For example, the microfluidic cartridge 200 can have a registration member 202 that fits into a complementary feature of the bay. The registration member 202 can be, for example, a cut-out on an edge of the cartridge 200 (as shown in Figure 3A) or one or more notches that are made on one or more of the sides. The skilled artisan will readily appreciate

that complementarity between the cartridge and the receiving bay can be easily achieved using other suitable arrangements, *e.g.*, a post or protrusion that fits within an aperture. By selectively receiving the cartridge 200, the recessed bay 524 can help a user to place the cartridge 200 so that the optical module 502 can properly operate on the cartridge 200. In this way, error-free alignment of the cartridges 200 can be achieved.

[0069] The receiving tray 520 may be aligned so that various components of the apparatus that can operate on the microfluidic cartridge 200 (such as, heat sources, detectors, force members, and the like) are positioned to properly operate on the microfluidic cartridge 200 while the cartridge 200 is received in the recessed bay 524 of the receiving tray 520. For example, contact heat sources on the heater substrate 600 may be positioned in the recessed bay 524 such that the heat sources can be thermally coupled to distinct locations on the microfluidic cartridge 200 that is received in the receiving tray 520.

Microfluidic Cartridge

[0070] Certain embodiments contemplate a microfluidic cartridge configured to carry out amplification, such as by PCR, of one or more polynucleotides from one or more samples. By cartridge is meant a unit that may be disposable, or reusable in whole or in part, and that may be configured to be used in conjunction with some other apparatus that has been suitably and complementarily configured to receive and operate on (such as deliver energy to) the cartridge.

[0071] By microfluidic, as used herein, is meant that volumes of sample, and/or reagent, and/or amplified polynucleotide are from about 0.1 μ l to about 999 μ l, such as from 1-100 μ l, or from 2-25 μ l, as defined above. Similarly, as applied to a cartridge, the term microfluidic means that various components and channels of the cartridge, as further described herein, are configured to accept, and/or retain, and/or facilitate passage of microfluidic volumes of sample, reagent, or amplified polynucleotide. Certain embodiments herein can also function with nanoliter volumes (in the range of 10-500 nanoliters, such as 100 nanoliters).

[0072] Figure 3A is a top plan view of a microfluidic cartridge 200. The cartridge 200 may comprise a plurality of sample lanes 1706a-c. The lanes may lead to PCR chambers 1703 located on “left” and a “right” sides (*i.e.*, rows) of the cartridge. As indicted in Figure 3a, the lanes may provide inlet ports 1705 in a convenient location near the user. However, the lanes to which the ports are connected may then take independent paths to separate chambers 1703a-c. In the embodiment of Fig 3a, for example, the first lane 1706a is in communication with the first chamber 1703a of the left side, the second lane 1706b is in communication with the first chamber of the right side 1703b, the third lane 1706c is in communication with the

second chamber 1703c of the left side, etc. Each of the microfluidic lanes may also comprise microfluidic valves 1702, 1704, microfluidic gates, and microfluidic channels. These gates and valves may be configured, *e.g.*, by thermal actuation, to facilitate timed release and controlled diffusion of certain fluids within the lanes 1706 of cartridge 200. The cartridge of this embodiment may comprise venting holes 1701 which prevent air from blocking fluid passage within the cartridge. Further description of various cartridge components, such as valves, may be found in *e.g.*, U.S. Patent Application Publication 2009-0130719, incorporated herein by reference.

[0073] The microfluidic cartridge 200 may include a registration member 202, for example, a cutout, which corresponds to a complementary edge in the recessed bay 524 of the receiving tray 520a,b of the heater/optical modules 500a, 500b. The registration member 202 and the complementary edge 526 may allow for secure and correct placement of the microfluidic cartridge 200 in the receiving tray 520a, b.

[0074] In various embodiments, the components of a microfluidic networks in the sample lanes 1706 of the cartridge 200 may be heated by thermally coupling them with the heaters in a heater substrate 600. The heater substrate 600 may be configured to heat a sample mixture comprising amplification reagents and an amplification-ready polynucleotide sample and cause it to undergo thermal cycling conditions suitable for creating amplicons from the amplification-ready sample. The heater substrate 600 may be located on the cartridge 200 in some embodiments or in the recessed bay 524.

[0075] The microfluidic network in each lane may be configured to carry out nucleic acid amplification, such as by PCR, on an amplification-ready sample, such as one containing nucleic acid extracted from a sample. An amplification-ready sample may comprise a mixture of amplification reagents and the extracted polynucleotide sample. The mixture may be suitable for subjecting to thermal cycling conditions to create amplicons from the extracted polynucleotide sample. For example, an amplification-ready sample, such as a PCR-ready sample, may include a PCR reagent mixture comprising a polymerase enzyme, a positive control nucleic acid, a fluorogenic hybridization probe selective for at least a portion of the positive control nucleic acid and a plurality of nucleotides, and at least one probe that is selective for a target polynucleotide sequence. The microfluidic network may be configured to couple heat from an external heat source with the mixture comprising the PCR reagent and the extracted polynucleotide sample under thermal cycling conditions suitable for creating PCR amplicons from the extracted polynucleotide sample.

[0076] In various embodiments, the reagent mixture may comprise fluorescent or other optically-detectable labels for the detection of the generation of a desired amplicon. In

some embodiments, multiple sets of primers and multiple labels can be used in a multiplex assay format, *e.g.*, multiplexed PCR, where each of a plurality of different amplicons can be detected in a single reaction chamber, if present. For example, one assay chamber could include template nucleic acids from a test sample, positive control template nucleic acids, one or more primer pairs for the amplification of specific target sequences, one or more probes for the detection of target amplicons, and one or more primer pairs and a probe for the detection of positive control amplicons. Additionally, the skilled artisan will appreciate that in some embodiments, the microfluidic cartridge accommodates a negative control polynucleotide that will not produce an amplicon with primer pairs used to amplify target or positive control sequences.

[0077] In certain of the illustrated embodiments, the chambers 1703a-c respectively associated with each lane 1706a-c of a multi-lane cartridge 200 may perform independent amplification reactions. The results of the reactions for the first column of chambers (1703a, 1703b) for the first two lanes (1706a, 1706b) may then be simultaneously and independently measured using a detector head which comprises a “left” and a “right” light source-photodetector pair. That is each chamber 1703a-b of each lane 1706a-b may receive light from a separate light source and be observed by a separate photodetector simultaneously. In this manner, a variety of combinations of reactions may be performed in the cartridge efficiently. For example, in some embodiments, a plurality of amplification assays for the detection of a plurality target nucleic acids can be performed in one lane, a positive control and a negative control in two other lanes; or one or more amplification assays for the detection of one or more target nucleic acids, respectively, in combination with an internal positive control in one lane, with a negative control in a separate lane. In one particular embodiment, 2, 3, 4, 5, 6, or more assays are multiplexed in a single lane, with at least that number of fluorescently distinct fluorophores in the reaction chamber.

[0078] A microfluidic cartridge 200 may be constructed from a number of layers. Accordingly, one aspect of the present technology relates to a micro fluidic cartridge that comprises a first, second, third, fourth, and fifth layers wherein one or more layers define a plurality of microfluidic networks, each network having various components configured to carry out PCR on a sample in which the presence or absence of one or more polynucleotides is to be determined. In another embodiment, the microfluidic cartridge 200 can comprise a plurality of lanes, each including a reaction chamber, etched or molded in a single plane, such as in a molded plastic substrate, with each lane being closed by a cover layer, such as an adhesive plastic film layer. Embodiments with 8, 10, 12, 14, 16, 18, 20, 22, 24, 28, 30, or more lanes per cartridge are contemplated. For example, one suitable design is a single cartridge 200 having 24 reaction chambers, arranged in two rows of 12 reaction chambers, optionally having relatively

aligned inlet ports. Further description of various cartridges and their components may be found in e.g., U.S. Patent Application Publication 2008-0182301 and U.S. Patent Application Publication 2009-0130719, incorporated herein by reference.

Heater Substrate

[0079] Shown in Figure 3B is a top plan view of certain embodiments of the heater substrate 600. Any type of heater can be used, including resistive, Peltier, or moving-fluid heaters, with either passive or active cooling contemplated. One of many possible embodiments includes a plurality of resistive heaters in thermal contact with each reaction chamber, preferably also including one or more temperature sensors. Because resistive heaters also exhibit some thermistor effect, i.e., their resistance changes with temperature, the resistive heaters themselves can double as temperature sensors, allowing precise temperature control of each reaction chamber while simplifying the product design. Although the heaters can be controlled in concert with each other, in some embodiments each reaction chamber can have one or more individual heaters in thermal contact therewith, such that the heaters are separately controllable and each reaction chamber can be heated and allowed to cool independently of the other reaction chambers. This allows different assays to be performed simultaneously in each of a plurality of reaction chambers. One particular resistive heater assembly for use with an individual reaction chamber is shown in Figure 3B. In the embodiment shown in Figure 3B, any combination of a top sensor heater/sensor 1604, a bottom heater/sensor 1603, a side heater/sensor 1601 and a center heater/sensor 1602 may be used to heat the reaction chamber located above. For ease of comprehension, an outline of the PCR chamber 1703 of certain of the embodiments is overlaid on the heater substrate. In certain embodiments, the heaters in the heater substrate 600 may be contact heaters. Such contact heaters may comprise (for example) a resistive heater (or network thereof), a radiator, a fluidic heat exchanger and a Peltier device. The contact heat source may be configured in the recessed bay 524 to be thermally coupled to one or more distinct locations of the microfluidic cartridge 200 received in the receiving tray 520a, b whereby the distinct locations are selectively heated. The contact heat sources may each be configured in the heater substrate 600 to be independently thermally coupled to a different distinct location in a microfluidic cartridge 200 received in the receiving tray 520a,b whereby the distinct locations are independently heated. The contact heat sources can advantageously be configured to be in direct physical contact with distinct locations of a microfluidic cartridge 200 received in the receiving tray 520a,b. In various embodiments, each contact source heater may be configured to heat a distinct location having an average diameter in 2 dimensions from about 1 millimeter (mm) to about 15 mm (typically about 1 mm to about 10 mm), or a distinct location having a

surface area of between about 1 mm about 225 mm (in some embodiments between about 1 mm and about 100 mm, or in some embodiments between about 5 mm and about 50 mm).

[0080] The heater substrate 600 may be organized into “lanes” 1605a, b paralleling the structure of the lanes 1706a-c of the cartridge 200. In some embodiments, the heater substrate 600 may include 24 heater lanes 1605a, 1605b corresponding to the sample lanes 1706 of cartridge 200. When the microfluidic cartridge 200 is placed in the recessed bay 524 of the receiving tray 520a,b, the components of the cartridge 200 may be aligned adjacent to, and above, the corresponding heaters in the heater substrate 600. When the microfluidic cartridge 200 is placed in the recessed bay 524, the heaters may be in physical contact with the respective components. In some embodiments the heaters remain thermally coupled to their respective components, e.g., through one or more intermediate layers or materials, though not in direct physical contact. Further description of lanes may be found e.g., in U.S. Patent Application Publication 2009-0130719, herein incorporated by reference.

[0081] In some embodiments, multiple heaters may be configured to simultaneously and uniformly activate to heat their respective adjacent cartridge components of the microfluidic network in the microfluidic cartridge 200. Each heater may be independently controlled by a processor and/or control circuitry used in conjunction with the apparatus described herein. Generally, the heating of microfluidic components (gates, valves, chambers, etc.) in the microfluidic cartridge 200, is controlled by passing currents through suitably configured micro-fabricated heaters. Under control of suitable circuitry, the lanes 1706 of a multi-lane cartridge can then be heated independently, and thereby controlled independently, of one another. Furthermore, as is described in more detail below, the individual heaters 1601-1604 can be heated independently, and thereby controlled independently, of one another. This can lead to a greater energy efficiency and control of the apparatus, because not all heaters are heating at the same time, and a given heater is receiving current for only that fraction of the time when it is required to heat.

[0082] The heater substrate 600 may also include one or more heat sensors. In order to reduce the number of sensor or heaters required to control the heaters in a heater lanes 1605a, 1605b, the heaters may be used to sense temperature as well as heat, and thereby obviate the need to have a separate dedicated sensor for each heater. For example, the impedance and/or resistance of some materials change with the surrounding temperature. Accordingly, the resistance of heater/sensors 1601-1604 may be used as an indication of temperature when the sensors are not being actively heated.

[0083] In some embodiments, the heaters in the heater substrate 600 may be designed to have sufficient wattage to allow the heaters to be grouped in series or in parallel to

reduce the number of electronically-controllable elements, thereby reducing the burden on the associated electronic circuitry. Heaters that are grouped together in this manner would be operated under synchronized and substantially simultaneous control.

[0084] In some embodiments, the reaction chamber heaters on opposite sides of the second stage heaters can be grouped and configured to operate under synchronized control. For example, in some embodiments, the PCR/amplification heaters 1601-1602 can be grouped and configured to operate under synchronized control. Alternative groupings and configurations can be applied to other heater groups of the PCR/amplification heaters 1601-1604. The PCR/amplification heaters 1601-1604 may be configured to operate individually and independently or they can be configured to operate in groups of two (pairs), three (thirds), four, five or six.

[0085] In some embodiments, the heating may be controlled by periodically turning the current on and off to a respective heater with varying pulse width modulation (PWM), wherein pulse width modulation refers to the on-time/off-time ratio for the current. The current can be supplied by connecting a micro fabricated heater to a high voltage source (for example, 30V), which can be gated by the PWM signal. In some embodiments, the device may include 48 PWM signal generators. In some embodiments there will be two PWM signal generators associated with each reaction chamber. Operation of a PWM generator may include generating a signal with a chosen, programmable period (the end count) and granularity. For instance, the signal can be 4000 us (micro-seconds) with a granularity of 1 us, in which case the PWM generator can maintain a counter beginning at zero and advancing in increments of 1 us until it reaches 4000 us, when it returns to zero. Thus, the amount of heat produced can be adjusted by adjusting the end count. A high end count corresponds to a greater length of time during which the micro fabricated heater receives current and therefore a greater amount of heat produced.

[0086] In various embodiments, the operation of a PWM generator may also include a programmable start count in addition to the aforementioned end count and granularity. In such embodiments, multiple PWM generators can produce signals that can be selectively non-overlapping (e.g., by multiplexing the on-time of the various heaters) such that the current capacity of the high voltage power is not exceeded.

[0087] Multiple heaters can be controlled by different PWM signal generators with varying start and end counts. The heaters can be divided into banks, whereby a bank defines a group of heaters of the same start count. Control of heating elements, and cooling elements, if present, in certain embodiments is discussed in further detail below.

Optical Module

[0088] Figures 4A-C illustrate the heater/optical module 500 of the detection apparatus 10 found in certain embodiments. The heater/optical module 500 may comprise an optical module 502 and a receiving tray 520 or a portion of the receiving tray. Figure 4A shows one embodiment of the enclosed optical module 502 having a motor 504 externally attached thereto for driving movement of detector head 700. The detector head 700 may be housed inside the optical module 502. Figure 4A illustrates the receiving tray 520 coupled to a bottom side 506 of the optical module 502. The receiving tray 520 may receive a cartridge 200 comprising samples upon which detection is to be performed. After receiving the samples, the receiving tray 520 may be moved (*e.g.*, mechanically or manually) on rails 522 to a position underneath the optical module 502. In some embodiments, described in greater detail below, the receiving tray may comprise an auto-loading device, which automatically aligns the cartridge once positioned beneath the optical module 502. In some embodiments, a recessed bay 524 of the receiving tray 520 may contain a heater substrate 600. In some embodiments, the receiving tray may subsequently be raised to place the cartridge in contact with the optical module 502, such as in contact with an aperture plate 540 at the base of the optical module 502.

[0089] Figure 4B illustrates an embodiment of the optical module 502 with a front panel 508 removed to show the interior of the optical module 502. Shown in Figure 4B is the detector head 700. As described in detail below, movement of the detector head 700 may be driven by the motor 504 to move laterally across the interior of the optical module 502 to provide optical scanning and detection on the cartridge 200 when the cartridge 200 is positioned below the optical module 502 in the receiving tray 520. Shown in Figure 4B is an aperture plate 540, positioned on the bottom side 506 of the optical module 502.

[0090] Figure 4C provides a bottom plan view of the optical module 502. Shown in Figure 4C is the aperture plate 540 and a normalizer plate 546 attached to the bottom of the 506 of the optical module 502. The normalizer plate may be used to calibrate the light source – photodetector pairs of the detector head. The normalizer plate 546 preferably comprises a one or more components having known, standardized optical characteristics, and is configured to calibrate, standardize, or confirm proper operation of the detector head 700 and associated circuitry. The normalizer plate 546 may extend into the optical module and the detector head 700 may be positioned over the normalizer plate. In some embodiments, prior to the start of cartridge optical measurements the detector head 700 is calibrated using the known properties of the normalizer plate 546. If the detector head 700 is not working properly, corrective action may be taken, such as including an offset in the measurements or notifying the user of the error. In some embodiments, the normalizer plate may be made of optically-transparent material such as polycarbonate mixed with a highly fluorescent dye, or other standardized chromophore or

fluorophore. In one embodiment, the normalizer plate includes a standardized chromophore or fluorophore for each channel or color in the detector head 700.

[0091] As shown in Figure 4C, the aperture plate 540 contains apertures 557. The dimensions of apertures 557 are such that the detector's light sources and photodetectors may have access to (optically excite or view) the contents in cartridge 200's reaction chambers when the detector is moved to a plurality of positions within optical module 502. That is, when a light source-photodetector pair of the detector is located in a position over a particular aperture light may travel from the light source and reach the chamber reactor through the aperture 557. The fluorescing reagents in the reaction chamber may then be visible to the photodetector via the aperture 557.

Detector Head

[0092] Figure 5 shows a cross-section of the detector head 700 taken along line 13 of Figure 4B. The detector head 700 may be configured to optically excite and/or monitor fluorescence emitted in connection with detection of from one or more polynucleotides present in the reaction chambers 1703. Note that a positive result (presence of a target amplicons) may be indicated by increased fluorescence or decreased fluorescence, depending on assay design. For example, when the assay involves a fluorophore and a quencher, the quencher may quench fluorescence when the target is present, or in other assay designs, when the target is absent. The system may comprise, for example, a plurality of detector pairs, *e.g.*, 3, 4, 5, 6, 7, 8, 9, 10, 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 23, 24, or more, such as the detector pair 726. Each detector pair 726 can be comprised of a light source 726a, such as a light-emitting diode (LED), and a corresponding light detector 726b, such as a photodiode. The light source 726a may selectively emit light in an absorption band of a fluorescent probe. The light detector 726b may selectively detect light in an emission band of the fluorescent probe, wherein the fluorescent probe corresponds to a polynucleotide probe or a fragment thereof. In certain embodiments the light source 726a may comprise a bandpass-filtered diode that selectively emits light in the absorption band of the fluorescent probe. Light detector 726b may comprise a bandpass filtered photodiode that selectively detects light in the emission band of a fluorescent moiety, *e.g.*, emission from a fluorescent probe. In certain embodiments, a filter 726a1, such as a bandpass filter may be applied to the light source 726a's light. The light from the light source 726a passes through a filter before passing through the sample in the micro-fluidic channel (300g deep in certain embodiments). In certain embodiments, the optical path-length for the light from the reaction chamber to the light detector 726b may be very small. The incident light from light source 726a generates fluorescence in the reaction chamber. Light from the reaction chamber

then travels to the light detector 726b. Certain embodiments seek to mitigate any undesired light from entering the detector and thereby adversely affecting the light signal from the reaction chamber.

[0093] In some embodiments, each one of the plurality of detector pairs may be arranged along the length of the detector head 700 in rows. That is, behind the pairs 726 and 727 illustrated in Figure 5 may be another column of pairs in a similar or same orientation. For the sake of explanation, a collection of cartridges or detector pairs along the length of the cartridge are referred to as a “row” and those along the width as a “column”. Thus, the vertical direction in Figures 3A and 6 indicates a “column” and the horizontal direction a “row”. Certain embodiments contemplate six or more columns of such detector pairs. In these embodiments, there would be 12 detector pairs in total (two rows of six) with two detector pairs per column, permitting 12 separate and simultaneous detections.

[0094] Each light source, such as for example light source 726a, may be configured to produce light of a wavelength specific to a specific fluorescent moiety associated with, *e.g.*, a probe, contained in the reaction chambers. Each light detector, such as for example 726b, may be configured to detect the light emitted from the fluorescent probes associated with the light produced by the light emitter in the detector pair. The detector pairs may be configured to independently detect a plurality of fluorescent moieties, *e.g.*, different fluorescent probes, having different fluorescent emission spectra, wherein in each reaction chamber, emission from each fluorescent probe is indicative of the presence or absence of one particular target polynucleotide or a fragment thereof. Although folded light paths can be used, one embodiment utilizes a detector and emitter pair where each is in direct optical contact with the reaction chamber, preferably simultaneously in such contact. Optionally, the detector and emitter of a pair are aligned with the reaction chamber along lines that substantially intersect at an acute angle at the reaction chamber. The angle can be, for example, between about 5 and 70 degrees, preferably between about 8 and 60 degrees, more preferably between about 10 and 50 degrees.

[0095] In some embodiments, the detector head includes two rows of photodetector and light source pairs that correspond to two rows of reaction chambers of microfluidic cartridges, when present in the apparatus. For example, the detector head can include a first or top row of six photodetector and light source pairs, and a second, or bottom row of photodetector and light source pairs, that are configured to query first and second rows of reaction chambers within a microfluidic cartridge, respectively.

[0096] Figure 6 illustrates one possible photodetector and light source layout implemented in certain embodiments of the detector. The first column comprises ROX light emitters 201a, 201b and corresponding detectors 207a, 207b. The second column comprises

HRM light emitters 201c, 201d and corresponding detectors 207c, 207d. The third column comprises CY5 light emitters 201e, 201f and corresponding detectors 207e, 207f. The fourth column comprises FAM light emitters 201g, 201h and corresponding detectors 207g, 207h. The fifth column comprises Q705 light emitters 201i, 201j and corresponding detectors 207i, 207j. The sixth column comprises VIC light emitters 201k, 201l and corresponding detectors 207k, 207l. In some instances, the detectors and emitters are selected with reference to particular fluorophores to be used in an assay. In the embodiment illustrated in Figure 6, the first or top row detector and light source pairs comprises a plurality of photodetector and light source pairs, e.g. emitters 201a, 201c, 201e, 201g, 201i, and 201k and detectors 207a, 207c, 207e, 207g, 207i, and 207k. The second or bottom row detector and light source pairs comprises a plurality of photodetector and light source pairs, e.g. emitters 201b, 201d, 201f, 201h, 201j, and 201l and detectors 207b, 207d, 207f, 207h, 207j, and 207l. A summary of the properties of exemplary emitters and detectors is shown in Table 1 below.

TABLE 1

Color (Spec)	Dye (Assay)	Wavelength Name (Ex/Em)	Software	CT#
Green	FAM	470/510	FAM	4
Yellow	TET, VIC	530/555	VIC	6
Orange	Texas Red, ROX	585/610	Cal Red / ROX	1
Red	Cy5	625/660	Cy5	3
Crimson	Cy5.5	680/715	Cy5.5	5
ultraviolet	null	ultraviolet	HRM	2

[0097] The exemplary arrangement of photodetectors and light sources depicted in Figure 6 can inhibit cross-talk between detection columns. That is, the wavelength range for each emitter detector pair may be selected so as to possess a minimal overlap with its neighboring emitter-detector pairs. Thus, for example, where CT# refers to the column of a particular emitter-detector pair in a 6-column detector head, Ex is the excitation wavelength of a fluorophore, and Em is the emission wavelength, it will be apparent that adjacent emission wavelengths are not adjacent to each other in the detector head. That the row HRM's dye is null merely indicates that a variety of dyes, not required for this particular example, may be used. In some embodiments, HRM refers to a "High Resolution Melt" and a corresponding light source for this photodetector may comprise an LED operating in the ultraviolet spectrum. One will recognize that the columns may be arranged in alternative variations and that alternative selections of light emitting sources and detectors may be substituted for those shown.

[0098] The light-emitter and photodetector pairs of each column may be calibrated using the normalizer plate. After calibration, the detector head may be moved to a position such that a first column of light-emitter and photodetector pairs is located over a first group of lanes

such that each light-emitter and photodetector pair has access to a reaction chamber of the lanes. Detection of the reaction chambers in the first group of lanes will then be performed using the first column of emitters/detectors. Then, the detector head may be moved to a second position such that the first column is over a second group of lanes and the second column is over the first group of lanes. Detection of the reaction chambers in the second group of lanes will then be performed using the first column of emitters/detectors and detection of the reaction chambers in the first group of lanes will then be performed using the second column of emitters/detectors. The process may continue until each column has passed over each lane. Thus, for N columns of detectors and M columns of chambers, the detector will perform detections at least $M + N - 1$ positions. For example, in the embodiments of Figure 11 there are 6 columns. For a cartridge comprising 12 lanes, the detector would need to move between at least 17 positions (18 if the calibration position is considered).

[0099] Figure 7 depicts the final results after operation of certain of the embodiments. Plotted are the detected fluorescent levels for each light emitter – photodetector pair 801-805 over time for a single reaction chamber (or reactor) associated with a single lane. After a sufficient number of iterations (approximately 30 in this example) of the annealing and denaturing protocol, the detectors identify increasing levels of fluorescence within the reactor.

Chamber Plate

[0100] Certain of the present embodiments relate to the plating surrounding and including the chamber layer. Particularly, certain embodiments contemplate the manufacture of an aperture layer comprising characteristics that advantageously facilitate consistent results across trials of the heating/detection module, as discussed in further detail below.

[0101] Figure 8 illustrates the plating arrangement found in certain embodiments of the scanning thermocycler's optical module and the associated receiving tray and cartridge. When the cartridge is brought within proximity of the aperture layer 540 of the optical module 500a, the thermal layer 600, chamber layer 200 (which may comprise a chamber substrate), and aperture layer 540 may be situated as depicted in the embodiment of Figure 8. As discussed above, the chamber layer 200 may comprise a plurality of reaction chambers 1703a-d, which may be located so as to be thermally controlled separately from one another, or in groups. Thermal layer 600 may comprise a plurality of thermal units 1605a, 1605b, 1605c. Figure 8 is a simplified, abstract diagram of the above description, and certain features of the microfluidic pathway are not shown. In certain embodiments the thermal units may be both mechanically and thermally disconnected from one another (as illustrated by their physical separation in Figure 4). However, in other embodiments, the thermal units may be each placed within a same

substrate material, but spaced such that they remain thermally disconnected, as discussed above. Thus, it is possible for the thermal units to be thermally separated, but not mechanically separated.

[0102] In this manner, each thermal unit may be associated with one or more reaction chambers 1703a-d, separately from the remaining reaction chambers. In agreement with the protocol specified for each reaction chamber, the thermal units may successively heat and/or cool their corresponding chamber appropriately. For example, thermal unit 1605c may cool and/or heat chamber 1703a such that the temperature of chamber 1703a is substantially independent of the cooling and thermal state of the chamber 1703a. While heating may be accomplished by running current through a microfluidic or electronic circuit, cooling may be “passive” in that only convection between the microfluidic chamber and is used to reduce the chamber’s temperature. The thermal units 1605a, 1605b, 1605c may be controlled using a closed loop control system.

[0103] In some embodiments, aperture plate 540 may be located over the chamber layer 200 and can provide pressure to chamber layer 200 to facilitate heating and cooling of the microfluidic cartridge, *e.g.*, the chamber layer, by thermal layer 600. The aperture plate can include a plurality of apertures 557a-d to facilitate each photodetector’s 726b observation of an individual reaction chambers 1703a-d. In the absence of aperture plate 540, and depending on the configuration of the thermal layer 600 and chamber layer 200, chamber layer 200 may “warp” and/or be sufficiently flexible that the thermal communication between chambers and the respective thermal units is inconsistent. Inconsistent heating and cooling can lead to less accurate execution of the protocols and less precise and accurate results. As described above, significant warping may restrict the optical head from lateral movement. Thus, the thickness of the aperture plate must be appropriately selected to facilitate a proper light path between each reaction chamber and the light sources and photodetectors while still ensuring proper heating and cooling of the chamber layer. If the aperture layer is too thick, the distance from the photodetector 726b to the chamber may be too great, undesirably attenuating the fluorescence reading from the reaction chamber. In addition to increasing the distance to the reaction chamber, an aperture layer 540 which is too thick or too heavy will place too much pressure on the reaction chamber, causing convection to be too great. Conversely, if the aperture layer 540 is too thin it may not prevent the chamber layer 200 from bending and warping, and the aperture layer 540 may bend and warp itself. Warping of apertures 557a-d or the chambers 1703a-d may deflect light from the light source 726a and prevent accurate readings by photodetector 726b.

[0104] Accordingly, the embodiments described herein provide aperture layers that advantageously avoid the drawbacks described above. In certain embodiments, the aperture

layer 540 is made, at least in part, of steel. In these embodiments, steel provides the appropriate strength, density and resistance to deflection desired for operation. Furthermore, the steel may provide low self-fluorescence and is therefore less likely to adversely affect the reading of photodetector 726b. The steel may also be electrochemically treated to diminish its self-fluorescence and thereby be less likely to adversely affect the reading of the photodetector. In certain embodiments, the aperture layer may instead comprise black nickel (Ni), i.e. Ni with a colorant added to it to reduce self-fluorescence. Certain embodiments contemplate combinations of these different materials and electrochemical treatments. In certain embodiments, the aperture layer 540 is made of aluminum and when secured by the adjoining support panels 500, 506, and 546, provide the appropriate strength. The aluminum may be electrochemically plated with an anodic oxide finish, *e.g.*, with a black colorant added to reduce self-fluorescence.

[0105] The illumination optics may be designed so that the excitation light falling on the reaction chamber, or reactor, is incident along an area that is similar to the shape of the reactor. As the reactor may be long and narrow, the illumination spot may also be long and narrow, i.e., extended, as well. Thus the shape of apertures 557a-d may be designed with consideration both to the dimensions of the reaction chamber underneath, as well as to the relative positions of the corresponding light emitter and photodetector. The length of the spot may be adjusted by altering a number of factors, including: the diameter of the bore where the photodetector 726b is placed (the tube that holds the filter and lens may have an aperturing effect); the distance of the photodetector 726b from the PCR reactor; and the use of proper lens in photodetector 726b.

Force Member

[0106] In certain embodiments, the receiving tray 520 places the chamber layer 200 in proximity to the thermal layer 600 or aperture layer 540, but does not mechanically couple and/or thereby place the layers in contact with one another. In this manner, the chamber layer 200 may be thermally, but not mechanically, coupled to the thermal layer 600. In other embodiments, the receiving tray places the thermal layer 600 in both mechanical and thermal contact with the chamber layer 200 and the chamber layer in mechanical contact with the aperture layer 540. In various embodiments, the apparatus may include one or more force members (not shown) that are configured to apply pressure to the receiving tray 520 in order to thermally couple the heat sources to the microfluidic cartridge 200 positioned in the receiving tray 520. The application of pressure may be important to ensure consistent thermal contact between the heater substrate and the reaction chambers, gates, and valves, etc., in the

microfluidic cartridge 200. When the receiving tray 520 is in a closed position, thereby being positioned under the aperture plate 540 of the optical module 502, the force member, such as a motor assembly, below the receiving tray 520 may begin traveling upwards towards the optical module 502, thereby bringing the receiving tray 520 closer to the optical module 502. As the receiving tray 520 travels upwards towards the optical module 502, the cartridge 200 may begin to come in contact with a bottom surface of the aperture plate 540. The cartridge 200 may continue traveling upward until sufficient pressure is received on the cartridge 200. As discussed above, the aperture plate 540 may apply an equal pressure across all points of the top of the cartridge 200 and thus, presses the cartridge 200 against the heater substrate 600 with uniform pressure. As discussed, the aperture layer may be selected to possess properties which facilitate this operation. For example, the material selection of the aperture plate 540 may provide very little deflection of the cartridge 200, when pressed against it.

[0107] The application of uniform pressure of the cartridge 200 against the heater substrate 600 may allow for uniform heating for each of the components of the cartridge when desirable. Although uniform pressure and contact may be obtained between the heaters in the heater substrate 600 and the components (valves, gates, chambers, etc.) of the microfluidic networks in the cartridge 200, the heaters are not necessarily activated simultaneously, as discussed above. In certain embodiments, application of even pressure does not necessarily result in equal heating of different components of the cartridge 200. In some embodiments, both the activation of a specific heater in the heater substrate 600 along with the pressure applied by the aperture plate 540 to the cartridge 200 activate a particular component of cartridge 200.

[0108] Figures 9A-H are diagrams of the dimensions of one possible embodiment of the aperture plate. In this embodiment, a chemical conversion coat may be applied to adjust the reflective properties of the aperture layer. Some portions 9002 may be selected not to receive the chemical conversion coat. The coating may be applied to the surface of plate 540 or deposited throughout its material. In some embodiments, the material of the plate 540 may comprise steel. In other embodiments, the plate 540 may comprise aluminum. In yet other embodiments, the material of the plate 540 may comprise nickel. In some embodiments, the material of the plate can be a combination of two or more materials, including for example, aluminum, nickel, or steel.

[0109] In the embodiment shown in Figures 9A-H the dimensions of the plate have been selected to meet the constraints regarding chamber pressure and optical path to the detector pairs discussed above. The material thickness of the plate 540 starts out at 0.3125 inches and is machined down to the desired thickness. As indicated, much of the plate comprises a thickness of approximately 0.25 inches. However this thickness may vary, for example, the thickness over

the aperture openings 557 may be 0.19 inches. As discussed above, the aperture opening thickness facilitates an unimpeded optical path between the photodetector and light source to the contents of the reaction chamber.

[0110] In general the dimensions of the aperture plate 540 are selected such that in combination with the properties of the materials constituting the aperture plate 540, the plate 540 provides sufficient pressure to the underlying chamber plate to facilitate proper heating and cooling as well as sufficient rigidity to prevent warping or deformation of the chamber plate. Such deformation may result in obstructions to the light source and photodetector optical path to the reaction chamber. Simultaneously, the dimensions of the plate should not impose an unfavorable distance from the reaction chamber of the chamber layer to the light-source and photodetector pair through the apertures 557. Neither should the aperture plate's dimensions 540 obstruct the optical path from the light-source and photodetector pair to the contents of the chamber reactor.

[0111] In some embodiments the normalizer plate 546 may be attached to the aperture plate by inserting screws at positions 9001 or other fixation means through an aperture. In other embodiments, these positions may facilitate broader calibration techniques via the apertures over the normalizer plates than with regard to the remaining apertures.

[0112] Figure 10 illustrates various dimensions of the perspectives of the aperture plate of Figures 9A-H. As discussed above, in this embodiment, a chemical conversion coat may be first applied to prevent the base materials, *e.g.*, aluminum, nickel or steel, from oxidation while also providing enhancing electrical grounding for proper electronics operation. Only surfaces which may be exposed to the optical operation are then selectively coated with black anodization.

Diagnostic Analysis Consistency

[0113] Certain of the present embodiments contemplate methods for ensuring consistent diagnostic analyses across trials within the same heater/detector and across different heater/detectors. Particularly, embodiments of a system and process for determining the duration and offsets for a plurality of PCR protocols so as to synchronize detection therebetween are disclosed. Additionally, methods for adjusting the reactor cooling time to ensure more consistent results are discussed.

[0114] Figure 11 is a temperature profile for a reaction chamber undergoing a particular protocol 2000. As illustrated above, the system in operation may comprise many different protocols of many different durations operating simultaneously in different reaction chambers. The protocol 2000 involves a plurality of identical heating/cooling cycles, where

each cycle comprises denaturing plateaus 2000B and annealing plateaus 2000D where the temperature is maintained constant for a period of time. These cycles may be preceded by a non-periodic cycle of the protocol, such as an incubation period. In certain embodiments, the protocol may be specified as a collection of temperatures and periods of time. That is, the protocol may initially specify only that the chamber is to be held at 95°C for the duration B and then held at 61°C for the duration D. Certain embodiments contemplate grouping these segments into “steps” and “substeps” to facilitate user and automated control. For example, the heating and cooling cycle 2000B and D may be referred to as a “step” with the duration B at 95°C and the duration D at 61°C referred to as “substeps”. In certain embodiments, a user may specify the durations of the substeps. In other embodiments, these durations may be retrieved from a database. Typically, these times are established either from a standard protocol or by user input, sometimes using an established “recipe” of temperatures and plateau times. In addition to these substeps, the protocol’s temperature profile will also comprise transitions, such as transition 2000A from 61°C to 95°C and transition 2000C from 95°C to 61°C. The duration of these transitions may be a consequence of the materials and environment about the reaction chamber and the nature of the heating elements employed.

[0115] In certain embodiments the thermal trajectory for both heating and cooling may be determined for the entirety of the reaction prior to the start of the run. In some systems, the contour of temperature versus time is monitored and adjusted throughout the reaction in order to minimize transition temperatures, and taking into account the variations in efficiencies of different heating elements. In other words, some systems utilize feedback control loops to drive to a target temperature, wherein the actual contour of the temperature time relationship can vary from cycle to cycle. Such adjustments can result in different overall reaction times, and, more importantly, different overall reaction efficiencies. Accordingly, in some embodiments, the systems and methods described herein advantageously provide systems wherein the contour of the temperature versus time relationship of the complete reaction for each independent reaction chamber (or group of chambers) is predetermined set prior to the start of the run. Not only does this advantageously allow for synchronization of the multiple detection steps across a plurality of different reactors, but it also enables for stricter control over parameters that minimize differences in reaction efficiencies that may arise as a result of different temperature/time contours. In some embodiments, the systems and methods provided herein provide for the report of errors at the end of a reaction if the measured temperature is different from the expected value when a run is completed.

[0116] At various points in the protocol temperature profile 2000, the user or recipe may specify that a detection occur. For example, for some protocols a detection may be

requested at the end of segment 2000D. Were detections arbitrarily specified in each protocol, the detector head would need to travel between positions in an inefficient manner and may even find it impossible to perform detections at the requested times. That is, were each of a plurality of protocols to be initiated simultaneously and run in parallel simultaneously across each of the reaction chambers in the cartridge, it would be very inefficient for the detector to meet each protocol's detection requests. Particularly, once calibration was complete the detector would need to first travel to positions suitable to perform detections for each light source-detector pair in its array for the first profile. By the time the detector finished, however, each of the remaining protocols would be entering a period when detection is not to be performed. There will therefore be a "dead time" period when the detector cannot perform any detections and must instead simply sit idle waiting for the opportunity to perform the next detection. This "dead time" is inefficient and unnecessarily prolongs the diagnostic process. Furthermore, where successive detections are to be performed, the "dead time" may generate irregular and aperiodic detections of the same chamber, possibly introducing inconsistent readings.

[0117] Certain of the present embodiments contemplate automated adjustments to portions of the profile 2000 to facilitate efficient detection across multiple protocols. This may be accomplished by allowing the user to edit, or the system may edit automatically, the length of segment 2000B or 2000D.

[0118] It should be understood that so long as at least a minimum plateau time occurs, some minor extension of plateau times can be accommodated in most amplification protocols. This flexibility is utilized to all efficient accommodation of different assays being performed simultaneously, while performing real-time monitoring of amplification by reading the various assays using a scanning detector head.

[0119] If detection were to be performed during segment 2000B, for example, the system or the user may extend the duration of segment 2000B as necessary to accommodate detector head movement and to coordinate the reading of a plurality of assays being performed simultaneously. The duration of segments 2000A and 2000C may be calculated using a predetermined standard cooling rate from the preceding temperatures and incorporated into the analysis. Some embodiments do not allow the user to edit these segments and they are instead accounted for by the system internally.

[0120] In certain embodiments, the protocol adjustments determined by the system may comprise at least three separate forms. The first adjustment may comprise an "intra-cycle adjustment" wherein plateaus such as 2000B and 2000D of the protocol are extended such that the entire step cycle 2000A-D achieves a desired duration, in some instances an integer multiple of a detection cycle time. This adjustment is described with respect to Figure 13. Once the

intra-cycle adjustment is complete, the system may then perform an “inter-cycle adjustment”. An inter-cycle adjustment may ensure that detection events within each cycle occur at integer multiples of a desired duration apart from one another (such as an integer multiple of the detection cycle time) between the cycles. These adjustments are discussed with regard to Figure 14. The third adjustment may comprise a “starting offset adjustment” which may depend only on the lane used for protocol execution. These adjustments are discussed with respect to Figures 15A-C.

Protocol Adjustment Overview

[0121] Figure 12 depicts a flow diagram of a process 4000 used in certain of the disclosed embodiments to determine an appropriate solution for the detector detection times and protocol profiles. Process 4000 may be implemented in software, hardware, or a firmware combination of the two. For example, the process may be implemented in any of an FPGA, a microcontroller, or software running on a computer processor. Portions of the process may be performed by a general purpose processor, such as a microcontroller, while other portions may be performed by dedicated hardware, software, or firmware systems. The process begins 4001 by determining a detection cycle time (or using a predetermined detection cycle time, e.g., already in memory) for the system 4002. The detection cycle time may comprise the time that is required for the detector to move to each of the detection positions (detection with each of the emitter/detector pairs in a detection head in each of the six columns of Figure 6), perform all necessary detections, and return to an initial position. Optionally, the user or system may be allowed to make adjustments to the detection procedure so as to modify the detection cycle time. For example, the user may wish to only perform detection using a subset of the detectors. In some embodiments the detection cycle time is approximately 10 seconds, when the embodiment comprises six columns of detector pairs and all six columns are used.

[0122] In some embodiments, the process may first determine a plurality of “intra-cycle adjustments” for one or more of the protocols 4003. As discussed below with respect to Figure 13, the durations for a step or substep may comprise the time to perform a particular step or substep within the protocol. The cycle times may be determined by a combination of user specifications and system identified constraints. In certain embodiments, the system will require the plurality of cycle times to be integer multiples of the detection cycle time. “Intra-cycle adjustments” may be introduced to satisfy this constraint. For example, if the detection cycle time were 12.2 seconds, the cycle times for a protocol step may be 22.4, 33.6, 44.8, or any other $N * 12.2$ duration, where N is an integer greater than 0. In some embodiments, it is only necessary to impose this constraint when a detection is to be performed within the cycle.

[0123] Thus, intra-cycle adjustments ensure that the cycle of the protocol is an integer multiple of the detection cycle time. However, a detection may be requested at any point within a cycle. If the detection cycle time is 10 seconds, then the very earliest that a detection may be performed is at 10 seconds after the protocol initiates. Detections may then be performed at integer multiples after that time (20, 30, 40 seconds, etc.).

[0124] Thus, a further adjustment, an “inter-cycle” adjustment 4004, may then be determined to ensure that the requested detection occurs at the appropriate time. These “inter-cycle adjustments” may be incorporated into the protocol as additional delays between protocol steps or substeps. Phrased differently, a PCR protocol once subjected to “intra-cycle” adjustments may comprise “valid” cycle steps. The PCR protocol may then be generated by chaining together each of the steps and adding transitions from step to step. The “inter-cycle adjustments” 4004 ensure that the detection times occur at the desired integer multiples of the detection cycle time after the cycles have been chained together.

[0125] For example, for a system having a detection cycle time of 10 seconds a protocol may comprise a step having its first detection at 18 seconds into a cycle. The cycle duration (the duration of the entire step) may last for 30 seconds (perhaps after an “intra-cycle” adjustment). Thus, while the cycle time as a whole is properly aligned with the 10 second detection cycle time ($3 \times 10 = 30$ seconds) the first detection is itself not properly aligned with the detection (18 second is not a multiple of 10 seconds). The system will add 2 seconds of “inter-cycle” adjustment to the very first detection run so that the first detection occurs 20 seconds after the start of the protocol. This may be done by extending the previous step's final hold temperature for an additional 2 seconds via a “padding adjustment”. If there is no previous step, the system would insert a 2 second hold at ambient temperature to the beginning of the first run of the cycle. Thus, if the system begins operation at T_0 , the first detection will occur at $T_0 + 20$ seconds, the second detection at $T_0 + 50$ seconds, and so forth.

[0126] Because of the inter and intra-cycle adjustments, the protocol is now in a form such that detections will only be requested at times convenient for the detector head to move to the reaction chamber performing the protocol. Were all protocols performed in reaction chambers located in the first column of the cartridge (and sufficient number of detectors present in the detector head) intra and inter-cycle adjustments alone would suffice to properly modify the protocol for efficient detection (a first column here referring to a column of lanes such as lanes 1706a and 1706b with associated chambers 1703a and 1703b in Figure 3A). However, because the protocols operate in different columns of the cartridge it is further necessary to offset the protocol's initiation to compensate for the detector head's delay in reaching the chamber location.

[0127] Thus “starting adjustment offsets” are added to the protocol based on the location of the chamber in which the protocol is performed. These “starting adjustment offsets” 4005 are described in greater detail with respect to Figures 15A-C. In some embodiments, these adjustments are made at run time and rely solely on the location of the lane of execution. For example, no adjustment may be necessary for a protocol running in lanes located in a first column of the chamber, so a protocol run in these lanes’ chambers will have a delayed start time of +0 seconds. Each subsequent column of lanes gains a delay of 400 milliseconds for its distance from the first column, due to the time required for the detections (two detections at 100 milliseconds each, performed asynchronously in this embodiment) and the detector motor movement (200 milliseconds). In this example, with a detection cycle time of 10 seconds, the first possible detection for each column of lanes is as follows: column 1 has its first detection at 10 seconds, column 2 has its first detection at 10.4 seconds, column 3 has its first detection at 10.8 seconds, etc. By delaying the start of a properly aligned protocol by the necessary time for a particular lane, the expected alignment can be maintained. While this particular example assumes that the protocol is already aligned (from adjustments 4003 and 4004), the skilled artisan will readily appreciate that other embodiments may determine the offsets anticipating future adjustments.

[0128] Although described in the order of steps 4003, 4005, and 4004, one will readily recognize that these steps may be arranged into any other suitable order, and neither need the system perform each step successively. In some embodiments, however, such as that described above, it may be necessary to perform inter-cycle adjustments after performing intra-cycle adjustments, as the inter-cycle adjustment depends on the intra-cycle modification. In contrast, the starting-offset adjustment 4005 may not depend on any previous determination. That is, in some embodiments the starting offset 4005 need be determined only once at run time, whereas the intra-cycle adjustments 4003 and inter-cycle adjustments 4004 may be performed for each cycle step in the protocols.

[0129] In some embodiments, once the protocol times have been properly adjusted, the process may then initiate the protocols 4006. In some embodiments a processor may simply place the offsets in a memory location for retrieval by a separate dedicated component of the system which itself initiates each protocol.

Intra-Cycle Adjustment

[0130] “Intra-cycle adjustments” comprise adjustments to step or substep intervals, as may have been specified by a user or received from a database, so that the step as a whole is an integer multiple of a predetermined duration. With reference to Figure 13 in certain

embodiments the user may specify certain features of the protocol profile, such as the desired times for a protocol substep, using a user interface, or graphical user interface (GUI). In some embodiments, the system may then validate the user's selection. After calculating segment lengths for each of the substeps, the system software will validate the step cycle time and indicate if any adjustments are necessary. In some embodiments a "valid" step cycle time is a step cycle time that is an integer multiple of the detection cycle time. If a step cycle time is not valid, the user may be prompted to make adjustments for that step 5003b. If no adjustment is necessary, the user may be notified that the step is aligned properly 5003a.

[0131] In the example of Figure 13, the user has required an incubation step 5001 comprising a single substep of 900 seconds. The user has requested that the step 5001 occur only once 5010 and therefore comprises a single step cycle. In this example, the detection cycle comprises 10 seconds. The user has not specified that any detection is to be performed and accordingly the step is valid, since the step will not require that the detector head's position be adjusted. When no detection is requested the system records the requested time intervals, for future offset considerations, but may not impose any constraint that the time interval be a multiple of the detection time (though the duration may be considered in determining a subsequent inter-cycle adjustment). If, however, in this example the user had requested detection during this step, the step would still be valid if no other delays are incurred, as 900 seconds is a multiple of the 10 seconds detection cycle. In either event, in the illustrated embodiment, the system has determined that this step entry is valid.

[0132] In the example illustrated in Figure 13, the step PCR 5002 comprises two substeps, a first substep where the chamber is to be held at 95°C and another substep where the chamber is to be held at 61°C. The user has requested that 45 cycles of this step be performed 5011. The user has requested that the first substep last 2 seconds and that the second substep last 10.2 seconds for a total of 12.2 seconds. As discussed above with respect to Figure 7, the system may have also calculated the transition time from 95°C to 61°C and added this duration to the user requests. In this example, heating from 61°C to 95°C requires 4.25 seconds and cooling from 95°C to 61°C in 7.05 seconds. These values may be stored internally in the system's memory or determined dynamically based on the user inputs. Finally, in some embodiments, when a detection is requested for a substep 5004, as the user has requested here, the system adds an additional delay to the hold time for that substep. In this example, that delay is 2.2 seconds, which accounts for the minimal time required to allow the detector to move and detect with each of six columns of light emitter-photodetector pairs in the detector head. That is, in this example, each color detection requires 200 milliseconds of exposure time and 200

milliseconds to move the motor between columns (5 transitions * 200ms + 6 detections * 200ms = 2.2 seconds).

[0133] Thus, the total duration for the step as a whole is:

[0134] 4.25 (heat) + 2.0 (denature) + 7.05 (cool) + 10.2 (anneal) + 2.2 (detection)= 25.7 seconds.

[0135] As 25.7 seconds is not a multiple of the 10 second detection time, adjustment will be necessary. As indicated 5003b, the system informs the user that they may either remove 5.7 seconds from the step duration or add an additional 4.3 seconds to achieve a multiple of the detection cycle time (i.e., a multiple of 10 seconds). These “intra-cycle step adjustments” will be incorporated into the protocol after the user’s selection.

[0136] One will recognize that the system may consider a plurality of other factors not indicated in this example when providing the user with an adjustment range. For example, additional delays to motor movement or incubation preparation may be factored in to the system’s analysis.

Inter-Cycle Adjustments

[0137] As mentioned above, “inter-cycle adjustments” comprise adjustments to the first cycle of a substep so as to create a delay between cycle steps. “Inter-cycle adjustments” may depend on the timing of the preceding steps and the end temperature of the immediately preceding step (if one exists). With reference to Figure 14 the “inter-cycle adjustment” 6005 determined to achieve a proper detection time occurrence, will be described.

[0138] In some embodiments the adjustment 6005 is determined by first determining the time required to heat or cool the temperature from the end of the previous step to the first substep temperature of the next step. If any additional time is necessary for alignment, the temperature from the end of the previous step may be maintained for this time. An example of alignment between the end temperature of a hold step at 75°C to the first substep temperature of 95°C is shown in Figure 14. The temperature is ramped at 8°C/s from 75°C to 95°C from points 6001b to 6001c, with the remaining time required for alignment spent holding at 75°C after the end of the previous step, from points 6001a to 6001b. This period may be referred to as an “inter-cycle adjustment”. To achieve continuance of detection alignment between steps, it may be necessary to shift (or delay) the start of a step cycle after the end of the previous step by this “inter-cycle adjustment”. The time required to heat or cool the temperature from the end of the previous step to the first substep temperature of the next step may then be calculated. If any

additional time is necessary for alignment, the temperature from the end of the previous step is maintained for the time of the “inter-cycle adjustment”. In some embodiments, the system may factor in these considerations when receiving user input via GUI 5000 and incorporate them into the proposed variance 5003b.

Starting Offset Adjustments

[0139] Figure 15A illustrates the beginning cycles of two separate protocol profiles 3001 and 3005. In each of these protocols the inter and intra-cycle adjustments may have been performed, but the starting offset has yet to be applied. In this example, profile 3001 includes a step with a cycle time of 30 seconds (the interval from time 0 to time 30). A time for detection 3020a, or a detection request occurs 30 seconds. Note that pursuant to the inter-cycle adjustments discussed above, a small delay may have been included in the protocol 3001 just prior to the first heat ramp for alignment of the first detection 3020a. As discussed above, the inter-cycle and intra-cycle adjustments facilitate detection requests being made at integer multiples of the detection cycle time. Here, for a detection cycle time of 10 seconds, the requests 3020a and 3020b occur at the integer multiples 30 and 60 seconds.

[0140] The second protocol 3005 includes a different profile from 3001. The profile 3005 comprises an initialization step lasting from 0 to 30 seconds. The profile 3005 is then followed by a plurality of 50 second cycles, with the first detection at 40 seconds. These cycles represent a 3-Temperature PCR, which includes a denature at a high temperature, the anneal and detection at the low temperature, and then an extension at a middle temperature. As before, the first initialization cycle may include a small inter-cycle delay at the beginning for alignment. One will recognize that this inter-cycle delay may be inserted at a variety of positions about the initialization step to ensure detection alignment.

[0141] Figure 15B illustrates multiple instances of the two protocols from Figure 15A. Were it possible to perform detections across all lanes in all chamber columns simultaneously, the profiles illustrated in Figure 15B would be suitable. However, due to time delay required for the detector head to scan across a chamber column with each of its columns of detector pairs, it is necessary to offset each of the protocols 3001-3007 based on the location of the chamber in which they are executed. For simplicity, each of the protocols 3001-3007 is presumed to be run in a neighboring column. If protocol 3001 is run in the first chamber column, then protocol 3002 is run in the second, 3003 in the third, 3004 in the fourth, etc.

[0142] Figure 15C illustrates the execution of the protocols 3001-3007 with the “starting offset adjustments” 3010-3015 introduced to ensure detection alignment. The starting offsets demonstrate the movement of the detector across the lanes of the cartridge and the

synchronization of that movement with the detections required by each of the executing protocols. Thus, protocol 3001 will request detection, using the first detector head column at request 3020a. When the detector head moves to align the second detector head column with the chamber of protocol 3001, the first detector head column will be arranged over the chamber of 3002 which, advantageously, is now also requesting a detection 3021a. Subsequently, the process continues with the first column of the detector head now reading protocol 3003 at request 3022a, the second column reading 3002, and the third reading 3001. One will recognize that the skew illustrated in Figure 15C is not to scale (in some embodiments the skew may be on the order of ~400 milliseconds), and has been illustrated as shown for only for purposes of explanation.

[0143] Thus, with properly selected “starting adjustments” the system can ensure consistent detection times across each of the reactors. As illustrated in Figure 15C, orderly and efficient detections are made along lines 3007a-d when the determined solution is implemented by the detector system. Thus, detection for a particular reactor will occur at the same time from cycle to cycle. The details for one embodiment for determining these solutions will be described in greater detail with regard to Figure 16.

Active Cooling

[0144] In certain of the embodiments while heating of the reactor chamber is active, that is, heaters are actively applied to the chamber, cooling of the reactor chamber may passive, where convection alone is used to cool the reactor contents. In order to further provide for consistent diagnostic performance, certain of the embodiments contemplate active participation in the reactor’s cooling process to ensure consistent behavior. Figure 16 illustrates a thermal profile 7001 comprising a cooling component. The profile 7001 comprises a rise time 7006, a plateau 7005, and a cooling period 7002/7003.

[0145] The ambient temperature in the location where the heating/detection unit is located may not be the same. That is, a system operating in southern Arizona may not be subjected to the same ambient temperatures as a system operating in northern Alaska. Thus, in the hottest ambient temperature in which the system is expected to be operated, the profile 7001 may have a cooling curve 7003. In a cooler environment, the cooling profile 7002 may instead result. To compensate for the difference, certain embodiments contemplate monitoring the reactor cooling profile via the temperature sensors, possibly those discussed with regard to Figure 3b. When deviations from the maximum profile 7003 are detected, sufficient heating may be applied so that the profile 7002 instead follows the profile 7003. In some embodiments heat may be applied periodically at times 7004a-c, whereas the heat may be applied

continuously in other embodiments. In this manner, consistent profiles may be achieved regardless of the thermocycler's geographic location or operating ambient temperature.

[0146] Certain of these embodiments apply Newton's law of cooling to determine when to apply the heaters:

$$T(t) = T_a + (T(0) - T_a)e^{-rt}$$

[0147] Where: $T(t)$ is the temperature at time t , $T(0)$ is the initial temperature, T_a is the ambient temperature parameter, r is the decay constant parameter, and t is time. In some embodiments 50.2 degrees Celsius and 0.098 may be used as the ambient temperature parameter and decay constant parameter, respectively. In this embodiment, the ambient temperature parameter is selected to be higher than any expected ambient operating temperature, thus allowing full control over the cooling cycle by applying at least some small amount of heat during each cooling cycle, regardless of ambient temperature, in order to match the actual cooling to the cooling curve of the maximal profile 7003 in each instance.

[0148] As used herein, an "input" can be, for example, data received from a keyboard, rollerball, mouse, voice recognition system or other device capable of transmitting information from a user to a computer. The input device can also be a touch screen associated with the display, in which case the user responds to prompts on the display by touching the screen. The user may enter textual information through the input device such as the keyboard or the touch-screen.

[0149] The invention is operational with numerous other general purpose or special purpose computing system environments or configurations. Examples of well-known computing systems, environments, and/or configurations that may be suitable for use with the invention include, but are not limited to, microcontrollers, personal computers, server computers, handheld or laptop devices, multiprocessor systems, microprocessor-based systems, programmable consumer electronics, network PCs, minicomputers, mainframe computers, distributed computing environments that include any of the above systems or devices.

[0150] As used herein, "instructions" refer to computer-implemented steps for processing information in the system. Instructions can be implemented in software, firmware or hardware and include any type of programmed step undertaken by components of the system.

[0151] A "microprocessor" or "processor" may be any conventional general purpose single- or multi-core microprocessor such as a Pentium® processor, Intel® Core™, a 8051 processor, a MIPS® processor, or an ALPHA® processor. In addition, the microprocessor may be any conventional special purpose microprocessor such as a digital signal processor or a graphics processor. A "processor" may also refer to, but is not limited to, microcontrollers, field programmable gate arrays (FPGAs), application-specific integrated circuits (ASICs),

complex programmable logic devices (CPLDs), programmable logic arrays (PLAs), microprocessors, or other similar processing devices.

[0152] The system is comprised of various modules as discussed in detail below. As can be appreciated by one of ordinary skill in the art, each of the modules comprises various sub-routines, procedures, definitional statements and macros. Each of the modules are typically separately compiled and linked into a single executable program. Therefore, the following description of each of the modules is used for convenience to describe the functionality of the preferred system. Thus, the processes that are undergone by each of the modules may be arbitrarily redistributed to one of the other modules, combined together in a single module, or made available in, for example, a shareable dynamic link library.

[0153] Certain embodiments of the system may be used in connection with various operating systems such as SNOW LEOPARD®, iOS®, LINUX, UNIX or MICROSOFT WINDOWS®, or any other suitable operating system.

[0154] Certain embodiments of the system may be written in any conventional programming language such as assembly, C, C++, BASIC, Pascal, or Java, and run under a conventional operating system, or the like, or any other suitable programming language..

[0155] In addition, the modules or instructions may be stored onto one or more programmable storage devices, such as FLASH drives, CD-ROMs, hard disks, and DVDs. One embodiment includes a programmable storage device having instructions stored thereon.

[0156] While the above processes and methods are described above as including certain steps and are described in a particular order, it should be recognized that these processes and methods may include additional steps or may omit some of the steps described. Further, each of the steps of the processes does not necessarily need to be performed in the order it is described.

[0157] While the above description has shown, described, and pointed out novel features of the invention as applied to various embodiments, it will be understood that various omissions, substitutions, and changes in the form and details of the system or process illustrated may be made by those skilled in the art without departing from the spirit of the invention. As will be recognized, the present invention may be embodied within a form that does not provide all of the features and benefits set forth herein, as some features may be used or practiced separately from others.

[0158] The steps of a method or algorithm described in connection with the embodiments disclosed herein may be embodied directly in hardware, in a software module executed by a processor, or in a combination of the two. A software module may reside in RAM memory, flash memory, ROM memory, EPROM memory, EEPROM memory, registers, hard

disk, a removable disk, a CD-ROM, or any other form of storage medium known in the art. An exemplary storage medium may be coupled to the processor such the processor can read information from, and write information to, the storage medium. In the alternative, the storage medium may be integral to the processor. The processor and the storage medium may reside in an ASIC. The ASIC may reside in a user terminal. In the alternative, the processor and the storage medium may reside as discrete components in a user terminal.

WHAT IS CLAIMED IS:

1. A apparatus for performing real-time nucleic acid amplification and detection, comprising:

a detector head comprising a plurality of photodetector and light source pairs, said detector head being mounted on a rail, wherein the detector and light source pairs are aligned into a first row and a second row;

a receptacle for a microfluidic cartridge comprising a plurality of independent reaction chambers aligned in adjacent columns of a first row and a second row; and

an aperture plate, the aperture plate configured to be positioned over the microfluidic cartridge when the cartridge is present in the receptacle, the detector head located over the aperture plate,

wherein the aperture plate comprises a plurality of apertures that are each aligned over each of the plurality of reaction chambers when the receptacle is holding the microfluidic cartridge,

wherein the detector head is moveable along the rail, such that each of the plurality of photodetector and light source pairs in the first row can be positioned over each aperture in the first row of the aperture plate, and each of the plurality of photodetector and light source pairs in the second row can be positioned over each aperture in the second row of the aperture plate.

2. The apparatus of Claim 1, further comprising

a second detector head comprising a plurality of photodetector and light source pairs, said second detector head being mounted on the rail, wherein the second detector and light source pairs are aligned into a first row and a second row;

a second receptacle for a microfluidic cartridge comprising a plurality of independent reaction chambers aligned in adjacent columns of a first row and a second row; and

a second aperture plate, the second aperture plate configured to be positioned over the second microfluidic cartridge when the second cartridge is present in the second receptacle, the second detector head located over the aperture plate,

wherein the second aperture plate comprises a plurality of apertures that are each aligned over each of the plurality of reaction chambers of the second microfluidic cartridge when the second receptacle is holding the second microfluidic cartridge,

wherein the second detector head is moveable along the rail, such that each of the plurality of photodetector and light source pairs in the first row of the second detector head can be positioned over each aperture in the first row of the second aperture plate, and each of the plurality of photodetector and light source pairs in the second row of the second detector head can be positioned over each aperture in the second row of the second aperture plate.

3. The apparatus of Claim 1, wherein the photodetector and light source pairs comprise at least six different photodetector and light source pairs operating in six different wavelengths.

4. The apparatus of Claim 3, wherein the six different wavelengths comprise a light source emitting a green colored light, a light source emitting a yellow colored light, a light source emitting an orange colored light, a light source emitting a red colored light, and a light source emitting a crimson colored light.

5. The apparatus of Claim 1, the detector head comprises at least N rows of photodetector and light source pairs, and the detector is configured to move to at least $M + N - 1$ positions over an aperture plate comprising M rows of apertures.

6. The apparatus of Claim 1, wherein the aperture plate comprises steel.

7. The apparatus of Claim 6, wherein the aperture plate comprises a thickness of approximately .25 inches

8. The apparatus of Claim 1, wherein at least part of the aperture plate is electrochemically oxidized to be darker than when the aperture plate is not electrochemically oxidized.

9. The apparatus of Claim 1, wherein the aperture plate provides substantially uniform pressure across the area of the microfluidic cartridge, when the cartridge is present within the receptacle.

10. The apparatus of Claim 1, further comprising a heater plate, wherein the heater plate is positioned underneath the microfluidic cartridge, if present in the receptacle

11. The apparatus of Claim 10, wherein the aperture plate provides substantially uniform pressure substantially across the area of the microfluidic cartridge when the cartridge is present within the receptacle, and wherein the substantially uniform pressure facilitates substantially uniform thermal contact between the reaction chambers and the heater plate.

12. The apparatus of Claim 1, wherein the aperture plate comprises at least one of aluminum, zinc or nickel, the aperture plate further comprising a colorant.

13. The apparatus of Claim 1, wherein the reaction chamber is configured to receive light at a glancing angle from the light source relative to the photodetector.

14. The apparatus of Claim 12, wherein the heater plate comprises a plurality of heating elements, wherein each of the plurality of heating elements is positioned such that when the microfluidic cartridge is present in the receptacle, the plurality of heating elements are in thermal communication with each of the plurality of reaction chambers, respectively.

15. A method implemented on one or more computer processors for optimizing protocols for simultaneously performing a plurality of thermal cycling reactions, wherein each thermal cycling reaction comprises one or more detection steps, and wherein said thermal cycling reactions are performed in a plurality of reactors, the method comprising:

determining or providing or accessing a detection cycle time for each of said plurality of reactors;

receiving or accessing a protocol step, the step associated with a step duration, the step comprising a time for detection; and

determining a first adjustment to the step such that the step duration is a multiple of the detection cycle time.

16. The method of Claim 15, further comprising determining a second adjustment to the step, wherein the time for detection is a multiple of the detection cycle time when the step is adjusted by the first adjustment and by the second adjustment.

17. The method of Claim 15, further comprising determining a starting offset adjustment based on a position of a reaction chamber associated with the protocol.

18. The method of Claim 15, wherein the detection cycle time comprises the amount of time required for a detector head to perform a predetermined plurality of detections for a reactor.

19. The method of Claim 18, wherein the detection cycle time further comprises a time required for movement of the detector head to each of a plurality of reactors and movement of the detector head to the start position.

20. The method of Claim 15, wherein the protocol comprises a polymerase chain reaction (PCR) protocol.

21. The method of Claim 15, further comprising initiating the protocol.

22. A non-transitory computer-readable medium comprising instructions, the instructions configured to cause one or more processors to perform the following method:

determining or providing or accessing a detection cycle time;

receiving or accessing a protocol step, the step associated with a step duration, the step comprising a time for detection;

determining a first adjustment to the step such that the step duration is a multiple of the detection cycle time.

23. The non-transitory computer-readable medium of Claim 22, wherein the protocol step is associated with a protocol from a plurality of protocols, each of the plurality of protocols associated with at least one of a plurality of thermal cycling reactions, wherein each thermal cycling reaction comprises one or more detection steps, and wherein the determining a first adjustment is based at least in part on a timing of one or more detection steps associated with the thermal cycling reactions of at least two or more of the plurality of protocols when the two or more of the plurality of protocols are simultaneously run.

24. The non-transitory computer-readable medium of Claim 22, the method further comprising determining a second adjustment to the step, wherein the time for detection is a multiple of the detection cycle time when the step is adjusted by the first adjustment and by the second adjustment.

25. The non-transitory computer-readable medium of Claim 22, the method further comprising determining a starting offset adjustment based on a position of a reaction chamber associated with the protocol.

26. The non-transitory computer-readable medium of Claim 22, wherein the detection cycle time comprises the amount of time required for a detector head to perform a predetermined plurality of detections for a reaction chamber.

27. The non-transitory computer-readable medium of Claim 26, wherein the detection cycle time further comprises a time required for movement of the detector head to each of a plurality of reaction chamber detection positions and movement of the detector head to a start position.

28. The non-transitory computer-readable medium of Claim 22, wherein the protocol comprises a polymerase chain reaction (PCR) protocol.

29. The non-transitory computer-readable medium of Claim 22, the method further comprising initiating the protocol.

30. A system for optimizing protocols for a plurality of reaction chambers, the system comprising:

a processor configured to perform the following:

determining or providing or accessing a detection cycle time;

receiving or accessing a protocol step, the step associated with a step duration, the step comprising a time for detection; and

determining a first adjustment to the step such that the step duration is a multiple of the detection cycle time.

31. The system of Claim 30, wherein the protocol step is associated with a protocol from a plurality of protocols, each of the plurality of protocols associated with at least one of a plurality of thermal cycling reactions, wherein each thermal cycling reaction comprises one or more detection steps, and wherein the determining a first adjustment is based at least in part on a timing of one or more detection steps associated with the thermal cycling reactions of at least two or more of the plurality of protocols when the two or more of the plurality of protocols are simultaneously run.

32. The system of Claim 30, wherein the processor is further configured to determine a second adjustment to the step, wherein the time for detection is a multiple of the detection cycle time when the step is adjusted by the first adjustment and by the second adjustment.

33. The system of Claim 30, wherein the processor is further configured to determine a starting offset adjustment based on a position of a reaction chamber associated with the protocol.

34. The system of Claim 30, wherein the detection cycle time comprises the amount of time required for a detector head to perform a predetermined plurality of detections for a reaction chamber.

35. The system of Claim 34, wherein the detection cycle time further comprises a time required for movement of the detector head to each of a plurality of reaction chamber detection positions and movement of the detector head to the start position.

36. The system of Claim 30, wherein the protocol comprises a polymerase chain reaction (PCR) protocol.

37. The system of Claim 30, the processor is further configured to initiate the protocol.

38. A method for simultaneously performing real-time PCR in a plurality of PCR reaction chambers, comprising:

(a) providing a scan time sufficient for a detector assembly to perform a scan cycle during which it can scan each of the plurality of PCR reaction chambers for at least one detectable signal and become ready to repeat the scan;

(b) providing a reaction protocol for each of the PCR reaction chambers that includes multiple cycles, each cycle comprising a cycle time that includes at least one heating step, at least one cooling step, and at least one temperature plateau that includes a reading cycle period during which the detector assembly is to scan the reaction chamber for at least one detectable signal;

(c) determining, using a processor, whether the cycle time for that reaction chamber is the same as or an integer multiple of the scan time, and if not, adjusting the scan time or the cycle time so that the cycle time is the same as or an integer multiple of the scan time;

(d) performing at least steps (b) and (c) for the reaction protocol for each of the plurality of PCR reaction chambers so that the cycle time for each reaction protocol is the same as or an integer multiple of the scan time; and

(e) under direction of a processor, performing real time PCR on each of the reaction chambers using the reaction protocol for each of the reaction chambers, including performing multiple scan cycles with the detector assembly, wherein each PCR reaction chamber is scanned by the detector assembly during each reading cycle period for that reaction chamber.

39. The method of Claim 38, further comprising phase adjusting the cycle time of the reaction protocol for at least one of the reaction chambers.

40. The method of Claim 38, wherein at least one said reaction protocol is different from another said reaction protocol.

41. The method of Claim 40, wherein at least one cycle time in one reaction protocol is different from the cycle time in another reaction protocol.




FIG. 1A




FIG. 1B





FIG. 2

4/29






FIG. 3B

6/29




FIG. 4A

7/29




FIG. 4B




FIG. 4C




FIG. 5

10/29




FIG. 6




FIG. 7

12/29




FIG. 8




FIG. 9A




FIG. 9B




FIG. 9C




FIG. 9E




FIG. 9D




FIG. 9F




FIG. 9H




FIG. 9G




FIG. 10A




FIG. 10B




FIG. 100*




FIG. 10E




FIG. 10D




FIG. 10F




FIG. 10G

FIG. 10H




FIG. 11

23/29




FIG. 12

5001

5002

5011

5000

5010

5004

Test Details

Incubate

Cycles1

Profile TypeHold

Type	Time (s)	Temp (C)	Detect
	900.0	95.0	☐

PCR

Cycles45

Profile Type2 - Temperature

Type	Time (s)	Temp (C)	Detect
	2.0	95.0	☐
	10.2	61.0	☒

☐ This step is valid.

☐ This step is not valid.
You must add 4.3 seconds
or remove 5.7 seconds.

5003a

5003b

FIG. 13




FIG. 14




FIG. 15A




FIG. 15B




FIG. 15C




FIG. 16

INTERNATIONAL SEARCH REPORT

International application No
PCT/US2012/033667

A. CLASSIFICATION OF SUBJECT MATTER
INV. G01N35/02
ADD.

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)
G01N B01L C12Q

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)

EPO-Internal, WPI Data

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	US 2009/130745 A1 (WILLIAMS JEFF [US] ET AL) 21 May 2009 (2009-05-21) paragraphs [0002], [0483], [0505], [0516] - [0518]; figures 78,79,82,83 -----	1-41
X	US 2008/149840 A1 (HANDIQUE KALYAN [US] ET AL) 26 June 2008 (2008-06-26) paragraphs [0199] - [0202]; figures 26-30 -----	1-14
A	WO 2010/118541 A1 (BIOCARTIS SA [CH]; KOLESNYCHENKO ALEKSEY [NL]; DE VRIES JORRIT E [NL];) 21 October 2010 (2010-10-21) abstract; figures 1,3,5 page 14, line 14 - page 15, line 11 page 15, lines 23-26 ----- -/--	1-41



Further documents are listed in the continuation of Box C.



See patent family annex.

* Special categories of cited documents :

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier application or patent but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art

"&" document member of the same patent family

Date of the actual completion of the international search

6 July 2012

Date of mailing of the international search report

12/07/2012

Name and mailing address of the ISA/

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+31-70) 340-2040,
Fax: (+31-70) 340-3016

Authorized officer

van Lith, Joris

INTERNATIONAL SEARCH REPORT

International application No
PCT/US2012/033667

C(Continuation). DOCUMENTS CONSIDERED TO BE RELEVANT		
Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	US 2009/189089 A1 (BEDINGHAM WILLIAM [US] ET AL) 30 July 2009 (2009-07-30) abstract; figure 1 paragraphs [0039] - [0041], [0049] - [0051], [0056] -----	1-41
A	US 6 236 456 B1 (GIEBELER ROBERT [US] ET AL) 22 May 2001 (2001-05-22) column 27, line 27 - column 30, line 34 -----	17
A	WO 2008/030914 A2 (APPLERA CORP [US]; TAN LIM HI [SG]; NGUI JEW KWEE [SG]; SHIN HOU SIU []) 13 March 2008 (2008-03-13) paragraph [0131]; figure 27 -----	1-41

INTERNATIONAL SEARCH REPORT

International application No.
PCT/US2012/033667

Box No. II Observations where certain claims were found unsearchable (Continuation of item 2 of first sheet)

This international search report has not been established in respect of certain claims under Article 17(2)(a) for the following reasons:

1. ☐ Claims Nos.:
because they relate to subject matter not required to be searched by this Authority, namely:
2. ☐ Claims Nos.:
because they relate to parts of the international application that do not comply with the prescribed requirements to such an extent that no meaningful international search can be carried out, specifically:
3. ☐ Claims Nos.:
because they are dependent claims and are not drafted in accordance with the second and third sentences of Rule 6.4(a).

Box No. III Observations where unity of invention is lacking (Continuation of item 3 of first sheet)

This International Searching Authority found multiple inventions in this international application, as follows:

see additional sheet

1. ☐ As all required additional search fees were timely paid by the applicant, this international search report covers all searchable claims.
2. ☒ As all searchable claims could be searched without effort justifying an additional fees, this Authority did not invite payment of additional fees.
3. ☐ As only some of the required additional search fees were timely paid by the applicant, this international search report covers only those claims for which fees were paid, specifically claims Nos.:
4. ☐ No required additional search fees were timely paid by the applicant. Consequently, this international search report is restricted to the invention first mentioned in the claims; it is covered by claims Nos.:

Remark on Protest

- ☐ The additional search fees were accompanied by the applicant's protest and, where applicable, the payment of a protest fee.
- ☐ The additional search fees were accompanied by the applicant's protest but the applicable protest fee was not paid within the time limit specified in the invitation.
- ☐ No protest accompanied the payment of additional search fees.

FURTHER INFORMATION CONTINUED FROM PCT/ISA/ 210

This International Searching Authority found multiple (groups of) inventions in this international application, as follows:

1. claims: 1-14

an arrangement for detecting the PCR reactants

2. claims: 15-41

a method for optimizing protocols for PCR

INTERNATIONAL SEARCH REPORT

Information on patent family members

International application No

PCT/US2012/033667

Patent document cited in search report			Publication date		Patent family member(s)		Publication date	
US 2009130745	A1	21-05-2009	US	2009130745	A1		21-05-2009	
			US	2012171759	A1		05-07-2012	

US 2008149840	A1	26-06-2008	US	2008149840	A1		26-06-2008	
			WO	2008060604	A2		22-05-2008	

WO 2010118541	A1	21-10-2010	AU	2010237532	A1		18-08-2011	
			CA	2752760	A1		21-10-2010	
			CN	102341710	A		01-02-2012	
			EP	2419743	A1		22-02-2012	
			KR	20120012468	A		10-02-2012	
			US	2012033210	A1		09-02-2012	
			WO	2010118541	A1		21-10-2010	

US 2009189089	A1	30-07-2009	AU	2006266464	A1		11-01-2007	
			BR	PI0613798	A2		15-02-2011	
			CA	2614151	A1		11-01-2007	
			CN	101218496	A		09-07-2008	
			EP	1902306	A1		26-03-2008	
			JP	4908507	B2		04-04-2012	
			JP	2009500621	A		08-01-2009	
			KR	20080041617	A		13-05-2008	
			US	2007009383	A1		11-01-2007	
			US	2009189089	A1		30-07-2009	
			WO	2007005076	A1		11-01-2007	
			ZA	200801187	A		25-03-2009	

US 6236456	B1	22-05-2001	EP	0981043	A1		23-02-2000	
			US	6236456	B1		22-05-2001	

WO 2008030914	A2	13-03-2008	US	2008274511	A1		06-11-2008	
			WO	2008030914	A2		13-03-2008	
