

(51) International Patent Classification:
A61B 5/103 (2006.01)(21) International Application Number:
PCT/IB20 11/052269(22) International Filing Date:
25 May 2011 (25.05.2011)

(25) Filing Language: English

(26) Publication Language: English

(30) Priority Data:
PCT/IB2010/05321 14 July 2010 (14.07.2010) IB(71) Applicant (for all designated States except US): **ECOLE
POLYTECHNIQUE FEDERALE DE LAUSANNE
(EPFL)** [CH/CH]; EPFL-SRI Station 10, CH-1015 Lau-
sanne (CH).

(72) Inventors; and

(75) Inventors/Applicants (for US only): **MARIANI, Benoit**
[FR/CH]; CH-1020 Renens (CH). **AMINIAN, Kamiar**
[CH/CH]; Chemin du Gregnolet 5, CH-1814 La Tour-de-
Peilz (CH).(74) Agent: **ROLAND, Andre**; c/o ANDRE ROLAND S.A.,
P.O. Box 1255, CH-1001 Lausanne (CH).(81) Designated States (unless otherwise indicated, for every
kind of national protection available): AE, AG, AL, AM,
AO, AT, AU, AZ, BA, BB, BG, BH, BR, BW, BY, BZ,
CA, CH, CL, CN, CO, CR, CU, CZ, DE, DK, DM, DO,
DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT,
HN, HR, HU, ID, IL, IN, IS, JP, KE, KG, KM, KN, KP,
KR, KZ, LA, LC, LK, LR, LS, LT, LU, LY, MA, MD,
ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI,
NO, NZ, OM, PE, PG, PH, PL, PT, RO, RS, RU, SC, SD,
SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR,
TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.(84) Designated States (unless otherwise indicated, for every
kind of regional protection available): ARIPO (BW, GH,
GM, KE, LR, LS, MW, MZ, NA, SD, SL, SZ, TZ, UG,
ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, MD, RU, TJ,
TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK,
EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU,
LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK,
SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ,
GW, ML, MR, NE, SN, TD, TG).**Published:**

— with international search report (Art. 21(3))

(54) Title: SYSTEM AND METHOD FOR 3D GAIT ASSESSMENT




Figure 11

(57) **Abstract:** The invention relates to a system and a method for assessment of walking and running gait in human. The method is preferably based on the fusion of a portable device featuring inertial sensors and several new dedicated signal processing algorithms: the detection of specific temporal events and parameters, 5 optimized fusion and de-drifted integration of inertial signals, automatic and online virtual alignment of sensors module, 3D foot kinematics estimation, a kinematic model for automatic online heel and toe position estimation, and finally the extraction of relevant and clinically meaningful outcome parameters. Advantageously including at least one wireless inertial module attached to foot, the system provides common spatio-temporal parameters (gait cycle time, stride length, and stride velocity), with the 10 advantage of being able to work in unconstrained condition such as during turning or running. It furthermore may provide original parameters for each gait cycle, both temporal (load, foot-flat and push duration) and spatial (foot clearance and turning angle), and their inter-cycles variability. The system and method according to the invention allows the assessment of various aspects of gait which have shown recently to be of premium importance in research and clinical field, including foot clearance, 15 turns, gait initiation and termination, running, or gait variability. The system may be light weight, easy to wear and use, and suitable for any application requiring objective and quantitative evaluation of gait without heavy laboratory settings.



System and method for 3D gait assessment

1 Field of invention

5 The present invention relates to the assessment of gait (walking or running) through the use of at least one sensor fixed to the foot

2 State of the art

In clinical setting, gait and mobility is commonly evaluated using questionnaire, observation or simple functional performance assessments (Tinetti, 1986; Podsiadlo and Richardson, 1991). These evaluations
10 do not require sophisticated equipments and have the advantage of being easy to perform by trained evaluators. However, they are often subjective and dependant on the experience of evaluator. Furthermore, these measures do not allow evaluating specific spatio-temporal gait parameters that have been associated with frequent geriatric syndromes, such as falls, dementia, or frailty (Hausdorff et al., 2001; Kressig et al., 2004; Seematter-Bagnoud et al., 2009). Generally, spatio-temporal gait analysis
15 requires dedicated laboratories with complex systems such as optical motion capture. Recently, ambulatory devices have overcome some of these limitations by using body-worn sensors measuring and analyzing gait kinematics. Unlike standard optical motion capture that requires a dedicated working volume, body worn sensors can be linked to a light data-logger carried by the subject performing his activities outside the lab with minimal hindrance. Nevertheless, recorded data require appropriate
20 algorithms to compute relevant parameters for clinical use (Aminian, 2006).

Most common gait parameters, such as stride length or gait cycle time, can be obtained from the analysis of foot kinematics. Systems based on Micro-Electro-Mechanical Systems (MEMS) gyroscopes and accelerometers suffer from measurement errors and integration drifts, which limits position and orientation assessment during long-term measurements. However, by placing sensors on foot, drift can be
25 corrected periodically by assuming null velocity of foot during foot-flat period of stance (Curey et al., 2004). Using this hypothesis, Sabatini et al. (2005) proposed a 2-Dimensional (2D) analysis method with periodic linear drift correction at each stance, and Bamberg et al. (2008) used a similar approach with wireless hardware. However, both studies were restricted to analysis in sagittal plane. Subsequently, Sabatini (2005) used a 3-Dimensional (3D) approach using quaternion for foot orientation and position.
30 Veltink et al. (2003) suggested a method for 3D foot kinematics estimation using ambulatory device for

drop-foot stimulator with drift and azimuth resetting at each step. Using additional force sensors, Schepers et al. (2007) applied similar algorithms, focusing on foot placement in forward and lateral directions. Yet, these previous studies were limited to few subjects and the proposed methodologies were not evaluated against any reference instrumentation or only in "optimal" conditions, i.e. during straight walking. Some other studies have been published to track position wearing additional magnetometers (Yun et al., 2007) and/or GPS (Foxlin, 2005), but results remain essentially qualitative and were not validated for use in clinical field.

3 General description of the invention

The present invention relates to a method and a wearable system based on inertial sensors and dedicated algorithms for precise and accurate assessment of 3D gait spatio-temporal parameters.

In one preferred embodiment of the invention the method is based on temporal parameters detection coupled to an optimized fusion of inertial signals.

The invention offers the opportunity to assess 3D gait features outside lab.

Foot clearance and turning angle are two parameters which may be (separately or simultaneously) used in the present invention.

Foot clearance, defined as the foot's height during the swing phase, appears to be an important gait parameter that should be related to the risk of falling. Contrary to other gait parameters, there is an unambiguous mechanism that links impaired foot clearance to falls. During walking, insufficient or fluctuations in foot clearance could lead directly to tripping, a major cause of fall in older people. Foot clearance encapsulates several aspects, notably toe and heel clearance, and their 3D trajectory in space during gait. The present invention allow extracting minimal and maximal values of toe and heel clearance that are particularly relevant for clinical gait evaluation.

Turning angle is defined as the relative change in azimuth (i.e. the projection of orientation in ground plane (XY)) between the beginning and the end of gait cycle. Previous studies have shown the computation of 3D Orientation of body segments from body-fixed sensors, nevertheless, this information is difficult to interpret from a clinical point of view. Turning Angle allows quantifying the amount of turning of the user at each stride with a single metric. It notably allows interpreting gait variability due to extrinsic factors such as a change in walking direction of a user wearing the system.

The invention also advantageously takes stance analysis into account. In clinical gait evaluation, stance phase is defined as the period of time where the foot is in contact with the ground. In normal gait, stance phase last approximately 60% of the total gait cycle, with a period of double limb stance followed by single limb stance (also referred as single support), and again double limb stance (Sutherland et al 1988).
5 (Winter et al. 1990) has characterized healthy elderly gait changes by a longer flat-footed landing. It shows that the sub-part of stance phase where the foot is almost completely still on the ground (so called "foot-flat") is particularly relevant in clinical gait assessment.

If most of the studies have proven the suitability of ambulatory measurements with the obvious advantage that information can be reliably derived on a large amount of data collected in daily condition, there are
10 still some important limits to the existing methods. In most of them, stance phase is considered a single block without any subdivision from heel-strike to Toe-off, making it impossible to assess aspects such as foot-flat duration. A technical and a clinical objective are also reached with the present invention:

- As technical objective, it is aimed at extracting inner-stance phases, notably foot-flat, from the precise detection of stance temporal events on foot inertial signals. Those suitable inertial features
15 are hypothesized for each temporal event could be found in agreement with force measurements on a three-segment foot model.
- As clinical objective, it is aimed at introducing new temporal parameters in gait assessment as potential outcome tools for clinical evaluations. It is assumed that healthy and patient populations can be discriminated by different stance strategies and foot-flat duration.

20 So the present invention also offers a precise and objective method to detect the events of stance phase and extract temporal metrics to address both said technical and clinical objectives.

The invention also offers the opportunity to design a method providing objective parameters during running, in order to support or reject the hypothesis that sport activity, and particularly running, should be
25 prohibited for patient after hip replacement. To this effect the method consists in adapting the existing algorithm to extract running phases and kinematics by using only shoe-mounted sensors. In addition muscular activity is recorded via electromyography (EMG). EMG and signals from MEMS accelerometers and gyroscopes may be recorded by two synchronized Physilog™ system. The full 3D motion of runner's foot and EMG may then be studied in order to extract precise and reliable parameters
30 that can be used to evaluate running performance, limp and coordination.

4 Detailed description of the invention

The invention will be better understood below by way of non-limitative examples and the following figures:

Figure 1 - S-Sense module with compliant foam attached with Velcro to hind part of shoe

5 **Figure 2** - "Physilog 3" Foot-worn sensors featuring 6DIMU and its fixation to user's shoes on forefoot

Figure 3 - The inertial signals and the temporal events for one typical gait cycle and corresponding events. Inertial signals re scaled to be depicted together in one graph.

Figure 4 - Temporal phases

10 **Figure 4** - Block Diagram of 3D Gait Analysis Algorithm

Figure 5 - De-drifted integration of vertical acceleration (A) to obtain vertical velocity (V) using linear function versus p-chip interpolation function (PIF)

Figure 6 - 3D Gait parameters estimation from 3D foot position (P) and azimuth (Θ): Stride Length (SL), Stride Velocity (SV), Foot clearance (FC) and Turning Angle (TA).

15 **Figure 7** - Sensor relative position to heel and toes

Figure 8 - Heel and Toe clearance estimation obtained from inertial sensors and kinematic model

Figure 9 - Heel (*) and Toe (*) trajectory estimated by foot worn sensor system during a typical gait cycle

20 **Figure 10** - Sub-components of the invention and their interactions

4.1 Foot-worn sensors

A wireless 6 Dimensional-Inertial Measurement Unit (6D-IMU) referred as "S-Sense" has been designed (Van de Molengraft et al., 2009). S-Sense module is a small ($57 \times 41 \times 19.5 \text{ mm}^3$) and low power ($18.5 \text{ mA} @ 3.6 \text{ V}$) stand-alone unit integrating microcontroller, radio transmitter, memory, three-axis accelerometer (ADXL, Analog Device, range 3g), three-axis gyroscope (ADXRS, Analog Device, roll, yaw with 300 deg/s range, pitch with 800 deg/s range), and batteries, and can also feature datalogger recording on flash memory cards such as SD-card. S-Sense modules can be fixed on shoes at hind foot position using a compliant foam structure and double sided Velcro straps (figure 1).

Raw sensor data was low-pass filtered at 17Hz, sampled on 12bits at 200Hz, and wirelessly transmitted in real time to a PC using "S-Base" receiver plugged in USB. In another embodiment of the invention, sample rate can be lowered to 100 Hz and/or sampled with higher resolution. Signals from two S-Senses were synchronized by considering the absolute real time clock sent by the base station to each module at the start of recording. Raw data were preliminary processed to extrapolate some missing data due to wireless data loss or sensor's output saturation (Van de Molengraft et al., 2009). Data from the two feet were finally converted to physical units (g or $^\circ/\text{s}$) using in-field calibration method (Ferraris et al., 1995).

More generally, the invention can use any equivalent sensor measuring 3D accelerations and 3D angular velocities on foot. For example, it can be a 6D-IMU module directly integrated in shoe. Another new wireless 3D gait analysis system was designed, the "Physilog 3" (figure 2), which is also small and lightweight and more conveniently worn on the front foot using shape-memory foam and Velcro bands, and contain an equivalent 6D-IMU.

In addition, the module might contain or be synchronized with additional sensors such as GPS receiver, force sensors, magnetometers, optical range sensors or EMG electrodes, providing extrinsic information.

4.2 Temporal events

During normal walking gait, stance phase is the period between initial contact, referred as Heel-Strike (HS), and terminal contact, referred as Toe-Off (TO), with the ground. In addition, this period encapsulates the instant where toes touch the ground, referred as Toe-strike (TS), and the instant where the heel leave the ground, referred as Heel-Off (HO). We call the successive events of HS, TS, HO, and TO the "Inner-stance phase events". The two negative peaks on pitch angular velocity (Ω_p) are known to be approximate estimates of Heel-strike and Toe-off events (Aminian, 2002). Those features have shown to be robust on a wide range of healthy and pathologic populations (Salarian, 2004) and were used to

distinguish them (Aminian, 2004). Then, we determined a time window between these two peaks to find other features to detect HS and TO based on the norm of the accelerometer signal ($\|A_N\|$) and the norm of the gyroscope signal, i.e., angular velocity, ($\|\Omega\|$) where the Euclidian norm of a vector $X = [x_1, x_2, x_3]$ is defined as $\|X\| = \sqrt{x_1^2 + x_2^2 + x_3^2}$. Regarding the period between TS and HO events, it is characterized by a lower amount of movement since the ground applies a mechanical constraint to the foot, and it is so-called foot-flat period. Consequently, TS and HO features are detected using a threshold on derivative of angular velocity norm ($\|\dot{\Omega}\|$), on pitch angular velocity (Ω_p), and the absolute value of the Jerk, indicated by the derivative of accelerometer signal's norm ($\|\dot{A}\|$). For all inertial signals, the use of the norm of 3D signals allows being independent of sensor placement on the foot, making it more repeatable and no specific calibration to align them with anatomical frames (Cappozzo, 1995) is required whereas the use of only pitch signal allow to use a single sensor configuration. Those entire hypotheses for detecting temporal events are illustrated in figure 3.

During running gait, foot kinematics can be slightly modified. Other robust features and thresholds were adapted consequently to detect temporal events.

Furthermore, temporal event detection also provides the static periods where no movement is sensed (typically when signal variations are below a defined threshold), referred as motion-less for those which occurs during stance phase. Finally, midswing event (MS) detected from positive peak of pitch angular velocity during swing provides a robust hypothesis for detecting gait cycle in any condition.

4.3 Temporal parameters

Based on the detected temporal events, meaningful metrics (e.g. parameters) for clinical gait analysis can be defined (figure 4). Thereby, stance phase was defined as follow:

$$Stance = t(TO) - t(HS)$$

Where $t()$ is the occurrence instant of each event. Subsequently, the three periods composing the stance phase were defined as follows:

$$Load = t(TS) - t(HS)$$

$$Foot-flat = t(HO) - t(TS)$$

$$Push = t(TO) - t(HO)$$

Furthermore, Swing time, gait cycle time, can also be calculated as follow:

$$Swing = t(HS^{+1}) - t(TO)$$

$$Gait\ cycle = t(HS^{+1}) - t(HS) \text{ (or } t(TO^{+1}) - t(TO) \text{ etc. ...)}$$

In case the system is mounted on two feet, double support parameters can also be calculated using classical definition (Aminian, 2002). Absolute metrics are calculated in milliseconds, and relative metrics are calculated as percentage of the stance time or gait cycle time.

4.4 3D Virtual Alignment

- 5 When foot-worn sensors are fixed on subject's foot, their relative orientation in space is unknown. 3D virtual alignment method consists in finding the initial orientation of the sensor, represented equivalently by a 3x3 Matrix or a quaternion or an axis angle. Initial 3D orientation of module is obtained by using 3D acceleration (a_n) as inclination during static periods (provided by temporal events for example), and azimuth was set at which maximized the variance of angular velocity signal around pitch axis of foot.
- 10 This original method has the great advantage of not requiring any functional calibration or precise positioning of the sensor module on subjects foot.

4.5 3D foot kinematics

- During each gait cycle n , 3D orientation (R_n), velocity (V_n), and trajectory (P_n) of foot were estimated from inertial signals. Practically, this involves the temporal detection of cycles, the knowledge of initial conditions of position and orientation, the gravity cancellation of measured acceleration, and the de-
- 15 drifted integration of g-free acceleration. Moreover, kinematics measured by sensors in xyz should be expressed in XYZ to be compared with reference. Figure 5 illustrates the main algorithmic steps.

- Initial conditions were updated for each cycle n at t_{ff_n} , where the foot was considered motion-less. Initial 3D orientation of sensor module (R_{0n}) was obtained by using 3D acceleration (a_n) as inclination (i.e. by aligning z axis with Z), and azimuth was set at the value derived from the orientation at last sample (N) of previous step (R^N)). It means the system can work with any terrain inclination, i.e. that the invention can also detect the initial conditions during walking in slopes, thus making it possible to analyze 3D foot kinematics in such situations. Gravity cancellation was achieved by aligning the accelerometers' axes (xyz) with fixed frame (XYZ) and subtracting gravity vector. From initial orientation R_{0n} , the orientation
- 25 of the foot relative to fixed frame ($R_n(i)$) was updated at each time frame ($i=1, 2, \dots, N$) by a quaternion-based time integration of angular velocity vector Ω_n between two successive foot-flats ($t_{ff_{n-1}}$, t_{ff_n}) (Sabatini, 2005; Favre et al., 2008). At each time frame i of cycle n , using measured accelerations ($a_n(i)$), gravity-free component of acceleration in fixed frame ($A_n(i)$) can be summarized by (1).

30
$$A_n(i) = a_n(i) * R_n(i) - g, \quad \text{where } g = (0, 0, I) \quad (1)$$

Single and double-integration of gravity-free acceleration (A_n) allowed obtaining 3D velocity and position of foot at each gait cycle n . By assuming that foot velocity is null at each tff_n (Curey et al., 2004), estimation of velocity (V_n) was obtained by trapezoidal integration of A_n and Position (P_n) was finally deduced by simple trapezoidal integration of velocity (V_n).

4.6 Drift Modeling

Integration step which is performed at 2.5 is prone to drifting errors, due to various sources such as electronic noise or sensors non-linear behaviors. So in practice, to obtain acceptable performance for estimating 3D foot kinematics, system drift needs to be corrected. This can be done using a classic linear de-drifting at each gait cycle between two motion-less period. In a preferred embodiment of the invention, the drift is removed by subtracting a sigmoid-like curve modeled based on a p-chip interpolation function (Carlson and Fritsch, 1985). The p-chip interpolation function (PIF), is defined between the value of $A_{n-1}(tff_{n-1})$ and $A_n(tff_n)$, (figure 6). As it is illustrated in figure 6, it provides a better estimation of drift in the particular case of gait since it is proportional to the quantity of movement, thus allowing improvement of accuracy and precision of 3D foot kinematics.

4.7 Spatial parameters

From the 3D foot kinematics, in addition to the overall 3D foot trajectory, the following gait parameters were extracted at each cycle n for both reference system and Foot-worn sensors using (2), (3), (4) and (5), where N represent the last sample of cycle n :

Stride length (SL) was defined as the distance measured between two successive foot-flat positions of the foot. This calculation is valid for curved and turning path as well (Huxham et al., 2006).

$$SL_n = |P_n(N) - P_n(1)| \quad (2)$$

Foot clearance (FC) was defined as the maximal foot height during swing phase relative to the height at foot-flat:

$$FC_n = \max(P_n(1), P_n(2), \dots, P_n(N)) - P_n(1) \quad (3)$$

Stride velocity (SV) was considered as the mean value of foot velocity in ground plane (XY) during each gait cycle:

$$SV_n = \text{mean}(V_{n|XY}(1), V_{n|XY}(2), \dots, V_{n|XY}(N)) \quad (4)$$

4.8 Turning

Turning Angle (TA) was defined as the relative change in azimuth (i.e. the projection of orientation in ground plane (XY)) between the beginning and the end of gait cycle.

$$TA_{\eta} = \theta_{\eta}(N) - \theta_{\eta}(I) \quad \text{where} \quad \theta_n = R_{nXY} \quad (5)$$

Extracted spatial parameters and turning are illustrated in figure 7.

4.9 Position modeling

Foot clearance provided by 2.7 gives general information which is dependent to sensor positioning on foot. Typically, a bigger FC is measured if sensor is on the heel compare to the case where sensor is on the foot. In order to be independent of sensor positioning, we design a method to automatically model sensor relative position to heel and toe, based on 3D foot kinematics and biomechanical assumptions. The relative position of sensor module in foot sagittal frame to the toe and heel of the subject can be represented by 3 variables (figure 8).

By combining the position of sensor (P), the knowledge of foot orientation (R) and shoe size and assuming that at Toe-off (TO) (respectively heel-strike (HS)), toe's (respectively heel's) vertical position is 0, $\{a, b, c\}$ during gait for each cycle n were computed by solving the following analytical equations:

$$\begin{cases} P_{nz}(HS) + b - b * R_{nz}(HS) - a * R_{nx}(HS) = 0 \\ P_{nz}(TO) + b - b * R_{nz}(TO) + c * R_{nx}(TO) = 0 \\ a + c = ShoeSize \end{cases}$$

4.10 Foot clearance

Knowing sensor trajectory (P_{sensor}) and orientation (R) and relative position to heel and toe ($\{a, b, c\}$), heel clearance (HC) and toe clearance (TC) can be estimated by the following trigonometric relations:

$$HC = P_{sensor} + b - b * R_z - a * R_x$$

$$TC = P_{sensor} + b - b * R_z + c * R_x$$

In addition to FC, parameters such as Minimal Toe Clearance (MTC) can be extracted from heel and toe clearance at each gait cycle as illustrated in fig.9.

Other parameters such as MaxHC, MinHC, MaxTC or other statistical measures can be extracted from clearance curves according to figure 10.

4.11 Variability

Since subjects are not always performing pure straight walking, direct variability of gait observed can be due to the turning at the end of the pathway. So in order to focus on the assessment of the 'intrinsic

dynamics' of continuous, normal walking, we need to ensure that the analysis is not influenced by those atypical strides outliers. Detection and correction of outliers in gait parameters series consists of the following steps:

- detect the gait cycles during turning, i.e. when TA is above a threshold obtained empirically
- 5 - replace turning gait cycle parameter with its median value during straight walking or simply remove it from the analysis.
- apply statistical method such as three-sigma rule to the new series in order to remove outlier related to other origins such as data loss or walking breaks (*facultative*)

To further quantify the stride-to-stride fluctuations in walking, there are various tools including commonly used linear parameters as well as non-linear methods. Parameters time series can be Foot clearance, Stride Length, Stride Velocity, Gait Cycle Time, or any other spatio-temporal parameter provided by the invention and previously described methods.

15 4.11.1 Linear Parameters

TABLE I
NOTATIONS

Symbol	Quantity
s	gait parameter time series (can be Foot clearance, Stride Length, Stride Velocity, Gait Cycle Time)
ms	mean of s
md	mean of the first derivative of
σ_s^2	variance of s
σ_s	standard deviation of s
σ_d	standard deviation of the first derivative of s
σ_{dd}	standard deviation of the second derivative of s
•	Coefficient of variation :

$$CV_s (\%) = \frac{\sigma_s}{m_s} \times 100$$

- Burstiness parameter:

$$B_s = \frac{\sigma_s - m_s}{\sigma_s + m_s}$$

- Median Absolute Deviation (MAD):

$$MAD_s = \text{median}(|s - \text{median}(s)|)$$

- 5
 - Standard deviation of the first derivative gait/stride time series:

$$\sigma_d \quad (\text{SD1})$$

- Interquartile range of the second derivative gait/stride time series: iqr
- Hjorth Complexity parameter:

$$c_H = \frac{\sigma_{\dot{\sigma}_d} \sigma_s^2}{(\sigma_d)^2}$$

- 10
 - Signal Permutation (Turns) Counts (STC)

In a given time series a data sample can be identified as a 'signal permutation/turn' (Note that 'turn' is not related to walking turn/outliers!) if it satisfies the following two criteria: 1) it represents an alteration of direction in the signal, i.e., a change in the sign of the derivative and 2) the difference (absolute value) between its amplitude and that of the preceding sample should be greater than a specific threshold. The number of signal permutation/turns in a time series represents the degree of signal variability.

4.11.2 Non-linear parameters: stride-to-stride variability analysis

The CV and related linear variability parameters quantify the magnitude of stride-to-stride variability but are not sensitive to changes in the ordering of the stride times or the dynamics. Randomly reordering a time series will not affect the magnitude of the variability but may dramatically alter the dynamic properties. To quantify how the dynamics fluctuate over time during the walk, fractal DFA analysis and symbolic entropy measures are applied to the stride time series.

4.12 3D Gait Assessment

In a preferred embodiment of the invention all previously discussed parameters are taken into consideration for the 3D Gait assessment. The system may provide objective evaluation of walking and running gait performance of a subject through original parameters such as foot clearance, foot-flat duration etc. ..., in any sort of walking situation or test. The diagram of figure 11 gives an overview of the interactions between the subcomponents of a system including all those parameters.

5 Final comments

The invention differs from the prior art in that it uses a least one original parameter (either temporal or spatial) that can be measured when performing any gait activity. It allows assessment during straight and curved trajectory, during outdoor locomotion, in ramp, stairs or even during running. A new drift compensation method renders the system more robust for precise and accurate gait assessment despite errors due to the sensors. These extracted parameters show promising preliminary discriminative performance, as it was possible to distinguish young and elderly subjects. The system according to the present invention was used successfully in more than 600 elderly subjects. It may be used for various purposes such as clinical gait evaluation, performance assessment in athletes, functional tests in patient with gait impairments, treatment follow-up, etc... For other application such as long-term tracking or clinical research however, it could require to be coupled with additional sensors such as magnetometers, GPS, EMG electrodes etc. ...

The method according to the invention can be applied with sensor worn on any foot position.

Compared to other inertial-based gait analysis system (Aminian et al., 2002; Salarian et al., 2004; Sabatini et al., 2005; Schepers et al., 2007), similar or slightly better accuracy and precision was obtained for SL and SV. The method also provides stride-to-stride variability of gait, with the advantage of being able to extract outliers due to turning or other extrinsic variation that can be measured from the system. In controlled environments, previous studies showed significant associations between gait variability and various syndromes associated with aging, such as frailty (Seematter-Bagnoud et al., 2009), and fear of falling (Rochat et al., 2010).

The method according to the invention allows the analysis of curved trajectories, and provides new parameters such as TA and FC, which were not provided by any previous inertial-based system. Actually, TA is an important outcome to evaluate gait in Parkinson disease (Zampieri et al.) and FC, which was shown to be the most discriminative parameters between young and elderly subjects in our study, could also be an important new gait parameter to estimate risk of fall in elderly (Begg et al., 2007; Lai et al., 2008).

Finally, the system according to the invention is lightweight and can be worn directly on user's casual shoes or barefoot, thus minimizing intrusiveness and interference with normal gait conditions. It could also be directly integrated in the foot-wear. The system can be used as an objective tool in many applications requiring gait evaluation in real conditions without usual constraints of limited space due to

5 laboratory settings.

The invention is of course not limited to the examples discussed previously.

1. References

- Aminian, K., Najafi, B., Bula, C. J., Leyvraz, P. F., Robert, P., 2002. Spatio-temporal parameters of gait measured by an ambulatory system using miniature gyroscopes. *Journal of Biomechanics* 35, 689-699.
- 5 Aminian, K., 2006. Monitoring Human Movement with Body-Fixed Sensors and its Clinical Applications. Invited chapter in "Computational intelligence for movement sciences : neural networks and other emerging techniques". Idea Group Pub., p. 101-138.
- Bamberg, S. J. M., Benbasat, A. Y., Scarborough, D. M., Krebs, D. E., Paradiso, J. A., 2008. Gait analysis using a shoe-integrated wireless sensor system. *IEEE Transactions on Information Technology in*
- 10 *Biomedicine* 12, 413-423.
- Begg, R., Best, R., Dell'Oro, L., Taylor, S., 2007. Minimum foot clearance during walking: Strategies for the minimisation of trip-related falls. *Gait and Posture* 25, 191-198.
- Bland, J. M., Altman, D. G., 1986. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1, 307-310.
- 15 Carlson, R. E., Fritsch, F. N., 1985. MONOTONE PIECEWISE BICUBIC INTERPOLATION. *SIAM Journal on Numerical Analysis* 22, 386-400.
- Crapo, R. O., Casaburi, R., Coates, A. L., Enright, P. L., MacIntyre, N. R., McKay, R. T., Johnson, D., Wanger, J. S., Zeballos, R. J., Bittner, V., Mottram, C , 2002. ATS statement: Guidelines for the six-minute walk test. *American Journal of Respiratory and Critical Care Medicine* 166, 111-117.
- 20 Curey, R. K., Ash, M. E., Thielman, L. O., Barker, C. H., 2004. Proposed IEEE inertial systems terminology standard and other inertial sensor standards. In *Proceedings of the Record - IEEE PLANS, Position Location and Navigation Symposium*, 83-90.
- Favre, J., Jolles, B. M., Aissaoui, R., Aminian, K., 2008. Ambulatory measurement of 3D knee joint angle. *Journal of Biomechanics* 41, 1029-1035.
- 25 Ferraris, F., Grimaldi, U., Parvis, M., 1995. Procedure for effortless in-field calibration of three-axis rate gyros and accelerometers. *Sensors and Materials* 7, 311-330.
- Foxlin, E., 2005. Pedestrian tracking with shoe-mounted inertial sensors. *IEEE Computer Graphics and Applications* 25, 38-46.
- Giansanti, D., 2006. Does centripetal acceleration affect trunk flexion monitoring by means of
- 30 accelerometers? *Physiological Measurement* 27, 999-1008.
- Hausdorff, J. M., Rios, D. A., Edelberg, H. K., 2001. Gait variability and fall risk in community-living older adults: A 1-year prospective study. *Archives of Physical Medicine and Rehabilitation* 82, 1050-1056.

- Huxham, F., Gong, J., Baker, R., Morris, M., Iansek, R., 2006. Defining spatial parameters for non-linear walking. *Gait and Posture* 23, 159-163.
- Kressig, R. W., Gregor, R. J., Oliver, A., Waddell, D., Smith, W., O'Grady, M., Curns, A. T., Kutner, M., Wolf, S. L., 2004. Temporal and spatial features of gait in older adults transitioning to frailty. *Gait and Posture* 20, 30-35.
- Kuipers, J. B., 1999. Quaternions and rotation sequences. Princeton Univ. Press, p.
- Lai, D. T. H., Begg, R. K., Taylor, S., Palaniswami, M., 2008. Detection of tripping gait patterns in the elderly using autoregressive features and support vector machines. *Journal of Biomechanics* 41, 1762-1772.
- Luinge, H. J., 2002. Inertial sensing of human movement. PhD. thesis, Twente University Press, Enschede, p.
- Menz, H. B., Latt, M. D., Tiedemann, A., Kwan, M. M. S., Lord, S. R., 2004. Reliability of the GAITRite® walkway system for the quantification of temporo-spatial parameters of gait in young and older people. *Gait and Posture* 20, 20-25.
- Podsiadlo, D., Richardson, S., 1991. The timed 'Up and Go': A test of basic functional mobility for frail elderly persons. *Journal of the American Geriatrics Society* 39, 142-148.
- Sabatini, A. M., 2005. Quaternion-based strap-down integration method for applications of inertial sensing to gait analysis. *Medical and Biological Engineering and Computing* 43, 94-101.
- Sabatini, A. M., Martelloni, C., Scapellato, S., Cavallo, F., 2005. Assessment of walking features from foot inertial sensing. *IEEE Transactions on Biomedical Engineering* 52, 486-494.
- Salarian, A., Russmann, H., Vingerhoets, F. J. G., Dehollain, C., Blanc, Y., Burkhard, P. R., Aminian, K., 2004. Gait assessment in Parkinson's disease: Toward an ambulatory system for long-term monitoring. *IEEE Transactions on Biomedical Engineering* 51, 1434-1443.
- Schepers, H. M., Koopman, H. F. J. M., Veltink, P. H., 2007. Ambulatory assessment of ankle and foot dynamics. *IEEE Transactions on Biomedical Engineering* 54, 895-902.
- Seematter-Bagnoud, L., Santos-Eggimann, B., Rochat, S., Martin, E., Karmaniola, A., Aminian, K., Piot-Ziegler, C., Bula, C. J., 2009. Vulnerability in high-functioning persons aged 65 to 70 years: The importance of the fear factor. *Aging Clin Exp Res*
- Tegner, Y., Lysholm, J., Lysholm, M., Gillquist, J., 1986. A performance test to monitor rehabilitation and evaluate anterior cruciate ligament injuries. *American Journal of Sports Medicine* 14, 156-159.
- Tinetti, M. E., 1986. Performance-orientated assessment of mobility problems in elderly patients. *Journal of the American Geriatrics Society* 34, 119-126.

- Van de Molengraft, J., Nimmala, S., Mariani, B., Aminian, K., Penders, J., 2009. Wireless 6D inertial measurement platform for ambulatory gait monitoring. In Proceedings of the Proceedings of the 6th international workshop on Wearable, Micro and Nanosystems for Personalised Health, Oslo, Norway, 63-64.
- 5 Veltink, P. H., Slycke, P., Hemssems, J., Buschman, R., Bultstra, G., Hermens, H., 2003. Three dimensional inertial sensing of foot movements for automatic tuning of a two-channel implantable drop-foot stimulator. *Medical Engineering and Physics* 25, 21-28.
- Von Eye, A., Mun, E. Y., 2006. Analyzing rater agreement: Manifest variable methods. *Applied Psychological Measurement* 30, 154-156.
- 10 Winter, D. A., Patla, A. E., Frank, J. S., Walt, S. E., 1990. Biomechanical walking pattern changes in the fit and healthy elderly. *Physical Therapy* 70, 340-347.
- Yun, X., Bachmann, E. R., Moore Iv, H., Calusdian, J., 2007. Self-contained position tracking of human movement using small inertial/magnetic sensor modules. In Proceedings of the Proceedings - IEEE International Conference on Robotics and Automation, 2526-2533.
- 15 Rochat, S., Bula, C.J., Martin, E., Seematter-Bagnoud, L., Karmaniola, A., Aminian, K., Piot-Ziegler, C , Santos-Eggimann, B., 2010. What is the Relationship Between Fear of Falling and Gait in Well-Functioning Older Persons Aged 65 to 70 Years? *Archives of Physical Medicine and Rehabilitation*, 91 (6), pp. 879-884.
- Zampieri, C., Salarian, A., Carlson-Kuhta, P., Aminian, K., Nutt, J. G., Horak, F. B., 2010. The
20 instrumented timed up and go test: Potential outcome measure for disease modifying therapies in Parkinson's disease. *Journal of Neurology, Neurosurgery and Psychiatry* 81, 171-176.

Claims

1. Method and dedicated algorithm based on sensor fusion and foot kinematics, for the 3D gait assessment using at least one inertial sensor fixed to the foot or integrated on a shoe; said method taking into consideration at least one of the following foot parameters obtained from said sensor(s) :
 - turning
 - spatial
 - clearance
 - temporal
2. Method according to claim 1 comprising the variability measurement of said parameters.
3. Method according to claim 1 or 2 taking into consideration at least said turning and said clearance parameters.
4. Method according to one of the previous claims wherein said temporal parameters are derived from temporal events of gait detected by said sensor(s), said temporal events being based on mathematical transformation of angular velocity and acceleration signals and including active or static periods, heel-strike, toe-strike, motion-less, heel-off, toe-off and midswing.
5. Method according to one of the previous claims 1 to 3 wherein said temporal parameters are derived from temporal events of gait detected by said sensor(s), said temporal events being based on at least one of the following parameters : load, foot-flat, push, swing and stance duration, as well as double-support when both feet features a sensing module.
6. Method according to one of the previous claims furthermore comprising the determination of the 3D foot kinematics (position and orientation) during human gait locomotion, including straight walking, turning, walking on uneven terrain, stairs, ramp, running, or any combination of those activities.

7. Method according to one of the previous claims wherein said spatial parameters are extracted at each gait cycle from temporal events and foot kinematics, including 3D foot trajectory, stride length, stride velocity, and foot clearance.
8. Method according to one of the previous claims furthermore using a kinematic model providing position of heel and toe during gait from combination of temporal events and the trajectory of another point on foot.
9. Method according to one of the previous claims furthermore quantifying said clearance parameter from spatial trajectories of sensor, heel or toe during gait at each cycle, including minimum and maximum of toe and heel clearance.
10. Method according to one of the previous claims furthermore quantifying said turning parameter during gait from spatial orientation of the foot and temporal events, including turning angle.
11. Method according to one of the previous claims furthermore taking into consideration the stride-to-stride variability of a gait parameters time series with removing of turning outliers.
12. A wearable system for the 3D gait assessment comprising one sensor adapted to be fixed to a foot or integrated on a shoe; said system furthermore being adapted to measure at least one of the following foot parameters :
 - turning
 - spatial
 - clearance
 - temporal
13. System according to claim 12 comprising a 6 degree of freedom inertial sensor.
14. A method using a system as defined in claim 12 or 13 and comprising a 3D alignment of said system relative to user's gait frame.
15. A method according to claim 4 based on function generation between said temporal events for modeling and compensating inertial sensors drift.




Figure 1




Figure 2




Figure 3




Figure 4




Figure 5




Figure 6




Figure 7




Figure 8




Figure 9




Figure 10




Figure 11

INTERNATIONAL SEARCH REPORT

International application No
PCT/IB2011/Q52269

A. CLASSIFICATION OF SUBJECT MATTER
INV. A61B5/103
ADD.

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)
A61B

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

EPO-Internal , WPI Data

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X, P	<p>MARIANI B ET AL: "3D gait assessment in young and elderly subjects using foot-worn inertial sensors" , JOURNAL OF BIOMECHANICS, PERGAMON PRESS, NEW YORK, NY, US, vol . 43, no. 15, 16 November 2010 (2010-11-16) , pages 2999-3006, XP027472961, ISSN: 0021-9290, DOI : 10. 1016/J .JBIOMECH.2010.07.003 [retrieved on 2010-07-24] the whole document</p> <p>----- -/--</p>	1-15



Further documents are listed in the continuation of Box C.



See patent family annex.

* Special categories of cited documents :

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier document but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art.

"&" document member of the same patent family

Date of the actual completion of the international search

12 October 2011

Date of mailing of the international search report

18/10/2011

Name and mailing address of the ISA/

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+31-70) 340-2040,
Fax: (+31-70) 340-3016

Authorized officer

Trachterna, Morten

INTERNATIONAL SEARCH REPORT

International application No
PCT/IB2011/052269

C(Continuation). DOCUMENTS CONSIDERED TO BE RELEVANT		
Category	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	IRIS TIEN ET AL: "Results of Using a Wireless Inertial Measuring System to Quantify Gait Motions in Control Subjects", IEEE TRANSACTIONS ON INFORMATION TECHNOLOGY IN BIOMEDICINE, IEEE SERVICE CENTER, LOS ALAMITOS, CA, US, vol. 14, no. 4, 1 July 2010 (2010-07-01), pages 904-915, XP011345721, ISSN: 1089-7771, DOI: 10.1109/TITB.2009.2021650	1,2,4-15
Y	abstract * IV. Initial control subject results p. 909-912 *	3
X	----- LAI D T H ET AL: "Measuring toe clearance using a wireless inertial sensing device", INTELLIGENT SENSORS, SENSOR NETWORKS AND INFORMATION PROCESSING, 2008. ISSNIP 2008. INTERNATIONAL CONFERENCE ON, IEEE, PISCATAWAY, NJ, USA, 15 December 2008 (2008-12-15), pages 375-380, XP031412593, ISBN: 978-1-4244-3822-8	1,2,4-15
Y	abstract * II. Device components and design p. 376-377 * * III. Experimental method and results p. 377-378 *	3
X	----- Benoit Mariani ET AL: "3D gait analysis using wearable 6D IMU on shoe", Proceedings of the 19th Conference International Society for Posture and Gait Research, 1 January 2009 (2009-01-01), pages 47-48, XP55009113, Retrieved from the Internet: URL: http://www.smi1ingproject.eu/pdf/ISPGR09_010.pdf [retrieved on 2011-10-10] the whole document ----- -/--	1-15

INTERNATIONAL SEARCH REPORT

International application No
PCT/IB2011/Q52269

C(Continuation). DOCUMENTS CONSIDERED TO BE RELEVANT		
Category	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	<p>Jef Van De Mol engraft ET AL: "Wi reless 6D inertial measurement platform for ambulatory gait monitoring" , Proceedings of the 6th international workshop on Wearable, Micro and Nanosystems for Personal ised Heal th, 1 June 2009 (2009-06-01) , pages 63-67, XP55009112, Retrieved from the Internet: URL: http://www.smi1ingproject.eu/pdf/pHealth2Q09.pdf [retrieved on 2011-10-10] abstract * III. Embedded real -time wal king phase detection p. 65-66 *</p> <p>-----</p>	1,2,4-15
X	<p>US 6 122 960 A (HUTCHINGS LAWRENCE J [US] ET AL) 26 September 2000 (2000-09-26) column 7, line 23 - column 8, line 55 column 9, line 64 - column 14, line 18; figure 5</p> <p>-----</p>	1-15
A	<p>H Stol ze ET AL: "Comparative analysi s of the gait di sorder of normal pressure hydrocephalus and Parkinson 's di sease. ", J Neurol Neurosurg Psychi atry, 1 March 2001 (2001-03-01) , pages 289-297, XP55009206, DOI : 10.1136/jnnp. 70.3.289 Retrieved from the Internet: URL: http://www.ncbi.nlm.nih.gov/pmc/articles/PMC1737236/pdf/v070p00289.pdf [retrieved on 2011-10-11] abstract * Results (Study 1) p. 291-293 *</p> <p>-----</p>	3

INTERNATIONAL SEARCH REPORT

Information on patent Tamilly members

International application No

PCT/IB2011/Q52269

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 6122960	A	26-09-2000	NONE
