

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property Organization

International Bureau



(10) International Publication Number

WO 2018/156962 A1

(43) International Publication Date
30 August 2018 (30.08.2018)(10) International Publication Number
WO 2018/156962 A1

(51) International Patent Classification:

A61B 17/12 (2006.01)

(72) Inventors: LUBOCK, Paul; 8 Argonaut, Suite 100, Aliso Viejo, California 92656 (US). QUICK, Richard; 8 Argonaut, Suite 100, Aliso Viejo, California 92656 (US). ROSENBLUTH, Robert; 8 Argonaut, Suite 100, Aliso Viejo, California 92656 (US). COX, Brian J.; 8 Argonaut, Suite 100, Aliso Viejo, California 92656 (US).

(21) International Application Number:

PCT/US2018/019532

(74) Agent: KLEIN, Howard J. et al.; 16755 Von Karman Ave., #275, Irvine, California 92606 (US).

(22) International Filing Date:

23 February 2018 (23.02.2018)

(81) Designated States (unless otherwise indicated, for every kind of national protection available): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BN, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DJ, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IR, IS, JO, JP, KE, KG, KH, KN, KP, KR, KW, KZ, LA, LC, LK, LR, LS, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PA, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SA,

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

62/463,498 24 February 2017 (24.02.2017) US
62/507,641 17 May 2017 (17.05.2017) US

(71) Applicant: INCEPTUS MEDICAL LLC [US/US]; 8 Argonaut, Suite 100, Aliso Viejo, California 92656 (US).

(54) Title: VASCULAR OCCLUSION DEVICES AND METHODS




FIG. 3

(57) **Abstract:** A vascular occlusion device includes a braided filament mesh structure defining a longitudinal axis. The mesh structure has a relaxed configuration in which it has an axial array of radially-extending occlusion regions, each of which has a proximal side and a distal side meeting at a peripheral edge, the sides of each occlusion region forming a first angle relative to the longitudinal axis. Each occlusion region is axially separated from the adjacent occlusion region by a reduced-diameter connecting region. The mesh structure is radially compressible to a compressed state in which it is deployed intravascularly to a target site through a catheter. Upon deployment, the device radially expands to a constrained configuration in which the peripheral edges of the occlusion regions engage the vascular wall, and the sides of the occlusion regions form a second angle relative to the longitudinal axis that is smaller than the first angle.



SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.

(84) Designated States (*unless otherwise indicated, for every kind of regional protection available*): ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, ST, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, KM, ML, MR, NE, SN, TD, TG).

Declarations under Rule 4.17:

- *as to applicant's entitlement to apply for and be granted a patent (Rule 4.17(ii))*
- *as to the applicant's entitlement to claim the priority of the earlier application (Rule 4.17(iii))*

Published:

- *with international search report (Art. 21(3))*
- *before the expiration of the time limit for amending the claims and to be republished in the event of receipt of amendments (Rule 48.2(h))*

VASCULAR OCCLUSION DEVICES AND METHODS

CROSS-REFERENCE TO RELATED APPLICATIONS

[0001] This application claims priority under 35 U.S.C. §119(e) from US Provisional Application No. 62/463,498; filed February 24, 2017, and from US Provisional Application No. 62/507,641; filed May 17, 2017. The disclosures of both provisional applications, to the extent they are not inconsistent with the disclosure herein, are expressly incorporated by reference.

FEDERALLY SPONSORED RESEARCH OR DEVELOPMENT

[0002] Not applicable.

BACKGROUND

[0003] This disclosure relates to devices for closing a passageway in a body. The passageway may be a natural defect, one having occurred due to disease or trauma, or a healthy, but undesirable passageway. The subject matter of this disclosure includes methods for closing an opening in tissue, a body lumen, blood vessel, a hollow organ, or other body cavity. The disclosure also relates to occlusion devices and methods for minimally invasive implantation, in particular, percutaneous implantation, of expandable braided occlusion devices for closing passageways in the cardiovascular system.

[0004] The cardiovascular system is part of the larger circulatory system, which circulates fluids throughout the body. The circulatory system includes both the cardiovascular system and the lymphatic system. The cardiovascular system moves blood throughout the body, and the lymphatic system moves lymph, which is a clear fluid that is similar to the plasma in blood. The cardiovascular system consists of the heart and blood vessels (arteries, veins, and capillaries). It delivers oxygen and nutrients to the tissues and carries waste products to the organs responsible for elimination. The arteries carry blood from the heart to the rest of the body, and the veins carry blood back to the heart.

[0005] Initially open, and subsequently minimally invasive, surgical procedures were used to occlude vessels and lumens, particularly when symptoms were significant and drug therapy inadequate. These techniques, although often successful, were hampered by less than desirable morbidity and mortality rates, and many patients were excluded due to the severity and risks associated with these procedures. In clinical practice today, occlusion devices are generally

delivered, in a collapsed state, percutaneously with a catheter through leg or arm vessels to the target vascular site or defect under fluoroscopic or ultrasonic guidance. Upon placement at the target site, the devices are allowed to expand, *in situ*, to an expanded state for implantation. Reduced morbidity and mortality risks have been observed using these percutaneously-delivered occlusion devices. Nevertheless, these devices may still suffer from potential drawbacks, including navigation difficulty through the catheter when collapsed, maintenance of device delivery flexibility sufficient to reliably navigate blood vessels through small diameter introducers, insufficient sealing of the vessel or defect, inadequate fixation of the device, subsequent reopening of initially occluded vessels, and inadequate provision for natural tissue ingrowth and healing following implantation. These drawbacks and others provide the impetus for the subject matter of this disclosure.

SUMMARY

[0006] Broadly, this disclosure relates to a medical device which, in some aspects, comprises an expandable mesh for at least partially closing a vascular defect (e.g., a fistula), occluding or blocking leakage around a previously-implanted vascular device, or closure of a healthy, but undesirable blood vessel or other bodily lumen.

[0007] In some aspects, a medical device in accordance with this disclosure comprises an expandable tubular mesh structure for at least partially closing a target site (e.g., a target vascular site), wherein the expandable mesh is fabricated from a self-expanding filamentous braid (i.e., made of metallic wire or non-metallic fibers), formed into a linear array of flexible, reconfigurable disc-like mesh occlusion regions providing between 4 to 10 mesh layers per centimeter when the device is in an unconstrained, radially expanded configuration. When the device is deployed in the target site, the mesh occlusion regions are reconfigured within a linear distance that provides a short, focal occlusion between about 0.5 cm and 2.5 cm as measured between the most distal and most proximal occlusion regions.

[0008] In some aspects, a medical device, as described herein, comprises an expandable tubular mesh structure for at least partially closing a structural vascular defect, wherein the expandable mesh structure is fabricated from self-expanding braid containing any of the materials described herein, resulting in a pore size of between about 0.05 mm and 0.50 mm, such

as between about 0.10 mm and 0.30 mm. Alternatively, pore sizes in the range of about 0.10 mm to 2.0 mm may be employed. In some embodiments, the pore size may be in the range of 0.20 mm to 0.75 mm.

[0009] In some aspects, a medical device, as described herein, comprises an expandable tubular mesh structure for at least partially closing a bodily lumen, such as a blood vessel, fistula, aneurysm, a patent ductus arteriosus, the left atrial appendage of the heart, leakage paths around implants (e.g. heart valves), or other vascular embolization targets, wherein the expandable mesh is fabricated from a tubular, self-expanding or shape-memory braided mesh structure.

[0010] In some aspects, a method is described for embolizing a target luminal (e.g., vascular) site by providing an expandable mesh structure that at least partially occludes the site, wherein the device is over-sized diametrically, with respect to the site, by between about 10% and 150%. In some embodiments, the amount of over-sizing may be between about 25% and 75%.

[0011] In some aspects, a method is described for treating a vascular condition by providing an expandable mesh structure that at least partially occludes a target vascular site, such as an undesirable blood vessel, fistula, aneurysm, patent ductus arteriosus, the left atrial appendage of the heart, leakage paths around implants (e.g. heart valves), and other vascular embolization targets.

[0012] In some aspects, a method is described for treating a vascular site by providing an expandable tubular mesh structure that at least partially occludes the site, wherein the expandable mesh structure is fabricated from a self-expanding mesh of braided fibers, filaments, or wires containing at least one of the following materials: nickel-titanium alloys (e.g. Nitinol), stainless steel, alloys of cobalt-chrome, polyester, PTFE, ePTFE, TFE, polypropylene, nylon, TPE, PGA, PGLA, PLA, polyethylene (PE), high-density PE (HDPE), and ultra-high molecular weight PE (UHMWPE) (including oriented strands of UHMWPE commercially available under the tradenames "DYNEEMA" and "SPECTRA").

[0013] In some aspects, a method is described for embolizing a target vascular site, wherein the method includes delivering to the target site an expandable mesh occlusion device,

and allowing the device to expand to at least partially occlude the target site, wherein the expandable mesh is fabricated from a self-expanding braided mesh with fiber diameters ranging between about 0.013 mm and 0.080 mm.

[0014] In some aspects, a method is described for embolizing a blood vessel, wherein the method includes delivering to a target site in the vessel an expandable mesh occlusion device, and allowing the device to expand so that it at least partially occludes the vessel, wherein the expandable mesh is fabricated from a self-expanding braided mesh that allows the device to expand from a compressed state having an essentially linear or tubular configuration, to an expanded or relaxed state in which it assumes a radially-expanded state comprising a plurality of axially-spaced, radially-expanded occlusion regions, i.e., a linear array of radially-expanded disc-like occlusion regions. In some embodiments, at least a portion of each of the occlusion regions forms an angle optimally between about 80° and 90° relative to the longitudinal axis of the device when the device is not constrained by the interior wall of the bodily lumen in which the device is deployed. This angle is decreased when the device is deployed and constrained by the endoluminal wall. The occlusion regions may then have an axially-flattened state such that each disc forms a pair of conjoined cones sharing a common base and defining a peripheral edge.

[0015] As used herein, the term "occlusion device," unless otherwise limited, shall be construed to encompass any device that is implantable in a bodily lumen for the purpose of at least partially obstructing the lumen, either by itself, or as an adjunct for embolization or tissue ingrowth. The term "bodily lumen" shall be construed as encompassing blood vessels and any other organ or structure that has a lumen or passageway that may provide a target site for occlusion.

BRIEF DESCRIPTION OF THE DRAWINGS

[0016] Fig. 1 is a diagrammatic view of a portion of a human cardiovascular system, showing a typical arrangement for the intravascular delivery of an implant, such as an occlusion device in accordance with this disclosure.

[0017] Fig. 2 is a simplified view of a device for the intravascular delivery of an occlusion device in accordance with this disclosure, showing the device being deployed intravascularly.

[0018] Fig. 3 is an elevational view of an occlusion device in accordance with an embodiment of this disclosure.

[0019] Fig. 4 is an elevation view of the occlusion device of Fig. 3, showing the device being delivered intravascularly to a target site.

[0020] Fig. 5 is a simplified view of a portion of an apparatus for braiding the filament from which an occlusion device in accordance with aspects of this disclosure is made.

[0021] Fig. 6 is a simplified view of the filamentous structure of an occlusion device in accordance with aspects of this disclosure.

[0022] Fig. 7 shows the process of flattening a filament used in the manufacture of an occlusion device in accordance with aspects of this disclosure.

[0023] Fig. 8 is an enlarged, detailed view of the filamentous structure of an occlusion device in accordance with aspects of this disclosure.

[0024] Fig. 9 is an elevation view of an occlusion device in accordance with an embodiment of this disclosure, showing the device deployed at a vascular site.

[0025] Fig. 10 shows an embodiment of a heat treatment fixture for manufacturing an occlusion device in accordance with aspects of this disclosure.

[0026] Fig. 11 shows two envelopes of braid angles for braids of 0.019 mm diameter NiTi wire with mandrels sized from 3.0 mm to 5.5 mm.

[0027] Fig. 12 shows two envelopes of braid angles for braids of 0.025 mm diameter NiTi wire with mandrels sized from 5.5 mm to 8.0 mm.

DETAILED DESCRIPTION

[0028] Methods and apparatus are presented for the occlusion of bodily lumens, particularly vascular sites and vessels, by minimally invasive implantation, including percutaneous implantation of an expandable occlusion device. In some embodiments, the device may comprise a plurality of radially-expanded occlusion regions formed from a single tubular mesh structure, such as a braid of fine filaments (i.e., fibers or wires). In the treatment of vascular sites, the device may be delivered to the site through an elongate tubular delivery device such as a catheter or cannula, which may be inserted intravascularly through a blood vessel (such as the femoral vein or radial artery) and advanced to the target site. Typically, the device delivery and

deployment procedure would be conducted under external imaging means such as fluoroscopy, x-ray, MRI or the like. Radiopaque markers may be incorporated into the occlusion device and/or the delivery device to help provide additional visibility under image guidance. Marker materials may include, for example, tungsten, tantalum, platinum, palladium, gold, iridium or other suitable materials.

[0029] Occlusion devices in accordance with this disclosure are configured for intravascular delivery to a target site in a bodily lumen, particularly a vascular target site, by known techniques and procedures for intravascular deployment of vascular implants. In accordance with typical procedures of this type, a guide wire 100 capable of traversing the circulatory system is introduced into the patient through a vascular access site or puncture 102, as shown, for example, in Fig. 1. For example, an endoluminal vascular access site may be formed in a femoral artery or vein, iliac artery or vein or jugular artery or vein of a patient. This methodology of positioning a guide wire is known to physicians skilled in the art of interventional therapy. Once the guide wire 100 is positioned, the vascular access site 102 is typically dilated to permit entry of a delivery catheter 20 therethrough. A protective sheath (not shown) may be advanced to protect the vascular structure. Thereafter, the guide wire 100 may be introduced into the patient through the vascular access site 102 and advanced to the site or target of embolization. The delivery catheter 20 is then advanced over the guidewire 100 until its distal end is situated adjacent the target site. Optionally, the delivery catheter 20 may be preceded by an outer catheter or "introducer" 22, as shown in Fig. 2, through which the delivery catheter 20 is inserted. The introducer 22 may optionally be formed with a tapered distal end portion to assist in navigation through vessels; alternatively, a flexible or removable dilator (not shown) may be used, as is well-known. The delivery catheter 20 likewise can have a tapered distal end portion.

[0030] Referring still to Fig. 2, an occlusion device 10, in accordance with this disclosure, is structured, as described more fully below, so that it may be resiliently compressed from a relaxed or expanded state to a compressed state for delivery, which is how the device is shown in Fig. 2. In the compressed state, the device 10 may be substantially linear in configuration; that is, it resembles a cylinder or tube. The device 10 is connected to the distal end of a pusher element 24 by a connection element (not shown) having a detachment mechanism, which may

be of any suitable type known in the art. The pusher element 24, typically in the form of, or comprising, a flexible wire, coil, multi-stranded helical wire tube or rod, is introduced into the proximal end of the delivery catheter 20 with the occlusion device 10 constrained in its compressed configuration, in which it is maintained by the constraint imposed by the inner wall surface of the delivery catheter 20. With the device 10 thus retained in a compressed state in the delivery catheter 20, it is advanced through the delivery catheter 20 by means of the pusher element 24 until it is at or near the target site in a bodily lumen, such as a blood vessel V (Fig. 2). At this point, the delivery catheter 20 is pulled proximally, freeing the occlusion device 10 from the constraint of the catheter 20, thereby allowing the occlusion device 10 to expand radially within the lumen so as to engage the endoluminal wall of the blood vessel V, as further described below, for deployment and implantation at the target site. When the device 10 is deployed at the target site, it is detached from the pusher element 24 by operation of the detachment mechanism.

[0031] If an outer catheter or introducer 22 is used (as shown, for example, in Fig. 2), the introducer 22 is advanced intravascularly until its distal end near the target site. The delivery catheter 20 is advanced distally through the introducer 22 until the distal end of the delivery catheter 20 is located at the target site. The device 10 is inserted into the proximal end of the delivery catheter 20 and pushed, by the pusher element 24, along the length of the delivery catheter up to the distal end. The delivery catheter 20 is retracted, and the device 10 is urged (e.g., by the pusher element 24) out of the distal end of the delivery catheter 20, thereby allowing the entire occlusive device 10 to expand to its functional size in an appropriate position for engagement with the walls of the target vessel or lumen at the target site. Once the device 10 is properly positioned, it is detached by the user from the pusher element 24. In some embodiments, the detachment is accomplished with minimal (preferably zero or close to zero) force being imparted to the device by the detachment mechanism to minimize movement of the device. In some embodiments, the occlusion device 10 is attached to the pusher element 24 by a tether (not shown), and the force applied needed for detachment of the tether is preferably less than 1 newton, such as, for example, 0.2 - 0.8 newton. Once the device has been detached, the delivery system may be withdrawn from the patient's vasculature or patient's body.

[0032] Specific embodiments of the occlusion devices disclosed herein are designed for delivery through a small intravascular delivery catheter 20 that preferably has an interior lumen diameter generally less than about 0.7 mm, known in the art as a “microcatheter.” For the device to be deliverable through the microcatheter, the collapsed profile or diameter of the device must be at least slightly less than the catheter lumen. In addition, the device must have sufficient flexibility, and sufficiently little outward radial force, to allow it to track through the microcatheter without excessive friction or resistance. For a device made of braided filaments, these delivery constraints for microcatheter deliverability limit the number and size of the filaments from which the braided device is made. In addition, the number of filaments and their diameters dictate the pore structure of the device, which directly impacts how rapidly the device can occlude a vessel or lumen. The filaments and their resultant geometry in the formed structure of the device must provide sufficient strength and radial stiffness to provide for device stability when implanted to prevent distal migration and/or embolization downstream of the intended occlusion site. Accordingly, this disclosure describes the unique inter-relationships of the structural parameters of the device which allow the braided mesh structure occlusion devices described herein to achieve the desired delivery and performance characteristics.

[0033] An occlusion device in accordance with this disclosure advantageously has three inter-related performance criteria: It must be deliverable to the target site; it must remain at the target site (i.e., it does not migrate and/or embolize downstream from the target site); and it must be occlusive within the target site.

[0034] Preferred deliverability characteristics are as follows:

[0035] Compressed Diameter (D_p) of the device is given by the equation:

[0036] $D_p = \sqrt{N_w} * 1.14 * W/\eta$, where N_w is the number of filaments forming the device, W is the diameter of each filament, and η is the packing density of the constituent filaments.

[0037] The packing density for the filaments is the sum of the cross-sectional areas of each wire relative to the area inside the perimeter of the packed wires Collapsed tubular braided structures of round filaments have been shown to have a packing density of about 0.75.

[0038] To be efficiently deliverable through a microcatheter of inside diameter D_c , the device must meet the following condition:

[0039] $(D_p + 2 * T_h) < D_c$, where T_h is the wall thickness of the device hub, as described below.

[0040] The friction of the device passing through the microcatheter is proportional to S_r and to N_F , where S_r is the radial stiffness of each surface feature of the braided wire portion the device, and N_F is the number of surface features of the radially-expanded occlusion portions of the device, as described below.

[0041] The radial stiffness S_r , in turn, is proportional to N_w and W^4 , and it is positively correlated to α and $1/R_E$, where α is the braid or pick angle (as described below), and R_E is the edge radius of each radially-expanded occlusion region, as described below.

[0042] It is also assumed that the braid or pick angle $\alpha = F(W, N_w, D_b)$, where D_b is the diameter of the mandrel on which the mesh structure is formed, as explained below.

[0043] The change in braid angle α for a change in braid length, B may be expressed as:

[0044] $\alpha_n = \cos^{-1}((B_n/B_i) \cos\alpha_i)$, where i indicates the initial value and n is a subsequent value

[0045] These parameters drive all of the resulting device performance parameters. The goal is to use as many wires of the largest diameter possible to construct a vascular occlusion device that is occlusive and stable, while being deliverable through a microcatheter sized appropriately for the anatomical structures of concern in a particular patient.

[0046] In some embodiments, to achieve the balance of the key parameters, a tubular braided wire mesh structure is formed on a mandrel using a low-tension braiding system as described herein, and the braiding parameters are set to obtain a braid or pick angle α in a particular range for a given wire diameter and mandrel diameter. In some embodiments, the braid angle will be within a preferred envelope a shown in Figs. 11 and 12. Fig. 11 shows the two preferred envelopes (A and B) of braid angles for braids of 0.019 mm (0.00075") diameter NiTi wire for various mandrels sized from 3.0 mm to 5.5 mm. Fig. 12 shows the two preferred envelopes (C and D) of braid angles for braids of 0.025 mm (0.001") diameter NiTi wire with various mandrels sized from 5.5 mm to 8.0 mm. After initial forming, the braid may be secondarily formed, typically by heat treatment described herein, into a final form. In one embodiment, a braided comprising a plurality of wires with an average diameter of between about 0.017 mm

and 0.021 mm, formed into a tubular mesh structure having a braid angle that lies within envelope A, can then be formed into an occlusion device with a plurality of disc-like occlusion regions, wherein the device will have both the desirable radial stiffness and the necessary deliverability through a microcatheter. In another embodiment, a braided mesh comprising a plurality of wires, each with an average diameter of between about 0.022 mm and 0.028 mm, formed into a tubular braided mesh structure having a braid angle that lies within envelope C, can then be formed into an occlusion device with a plurality of radially-expanded, disc-like occlusion regions, wherein the device will have both the desirable radial stiffness and the necessary deliverability through a catheter of a given internal diameter.

[0047] As shown in Fig. 3, an occlusion device 10 in accordance with an exemplary embodiment of this disclosure comprises a tubular braided structure, which has, in a relaxed or expanded state, a plurality of radially-expanded and thus radially extending occlusion regions 12, which, in some embodiments, are disc-like in configuration, as mentioned above. The occlusion regions 12 are axially spaced along the length of the device; that is, the device comprises a linear array of such occlusion regions 12. Each radially extending occlusion region 12 (hereinafter "disc," solely for the sake of brevity, and not in any limiting way) is separated from the nearest adjacent radially extending disc 12 by a tubular braided connecting region or core section 14 of reduced radius relative to the occlusion regions 12. The core sections 14 separate each adjacent pair of discs 12 by a core length that is advantageously less than about 3.0 mm, such as, for example, between about 0.5 mm and 2.5 mm.

[0048] As shown in Fig. 3, the device 10 defines a longitudinal axis X between a distal end (on the left side of the drawing) and a proximal end (on the right side of the drawing). The braided mesh structure is gathered together at each of the proximal and distal ends, where the filaments of the mesh are terminated in and retained by a proximal hub 26 and a distal hub 28, respectively. An attachment fixture 29 may advantageously be fixed to, or be integral with, the proximal hub 26, for detachable connection to a pusher element 24 by a detachment mechanism, which may advantageously include a detachable tether (not shown), as mentioned above.

[0049] Each disc 12 may advantageously be configured with a rounded peripheral edge or "apex" 16, the apices 16 of successive discs 12 being separated by a distance that may be

referred to as a “pitch,” as with screw threads. In some embodiments, the final, as-formed occlusive device 10 in its relaxed state may define a plurality of discs 12 with a spacing or pitch (apex to apex) distance, as shown in Fig. 3, between about 0.5 mm to 5 mm, such as, for example, between about 0.7 mm and 4.5 mm.

[0050] Referring still to Fig. 3, in the relaxed or radially-expanded state of the device 10, each disc includes a proximal or trailing side 17 and a distal or leading side 19, meeting at the peripheral edge or apex 16. A first disc angle θ_1 is defined between either the proximal side 17 or the distal side 19 of each radially-extending disc 12 and the longitudinal axis X. When the device 10 is in its relaxed or expanded state, the angle θ_1 may advantageously be between 70° and 90°, although smaller angles (e.g., not greater than 85°) may be satisfactory in some embodiments, depending on the application. The disc angle decreases to a second disc angle θ_2 that is smaller than the first disc angle θ_1 when the device is partially constrained owing to the engagement of the peripheral edges 16 of the discs 12 with an endoluminal wall upon deployment of the device in a bodily lumen, as shown in Fig. 4. The second disc angle θ_2 may be as small as 25°, although it may be beneficial for the second disc angle θ_2 to be up to about 70°, as a larger angle makes the total occlusion length or “landing zone” smaller or more compact, and thus allows the physician to keep the occlusion more focal at desired location. In addition, the more compact the device and the closer the radially extending discs are to each other, the more efficiently the device may work, thereby achieving occlusion more rapidly due to subsequent layers of filament acting on the flow of blood before the flow has returned to laminar flow after being disrupted and slowed by an initial mesh layer.

[0051] The effective size of the braid may be reduced as blood passes through each layer of mesh. The pore size of the initial layer ranges between 50 microns and 500 microns. Effective pore sizes may be reduced with each subsequent layer.

[0052] The radially-extending discs 12 can have a variety of shapes with various radial cross-sections, including, but not limited to, circular (as shown), ovoid, rectangular, diamond, or arbitrarily-shaped. In addition, the discs 12 can have generally the same diameter (as shown), or different diameters. They may have similar thicknesses and spacing between them, or varied thicknesses and spacing between them. In some embodiments, the disc diameters are more than

about five times their thickness at the thickest point when unconstrained. The occlusion device 10 may comprise between 2 and 8 discs in some embodiments, such as, for example, 3 to 5 discs. Each disc 12 may comprise two layers (mesh structures), that generally conform to the cross-section of the vessel or embolization site, thus spanning the lumen with a microporous flow disruptor. In some embodiments, narrow and tightly spaced discs may provide for a large number of mesh layers in close proximity, which may offer improved hemostasis. In some embodiments, when unconstrained, the occlusion device may comprise from 6 and 12 layers of mesh (each disc comprising two layers) in a linear distance of less than about 10 mm, and in less than about 25 mm in other embodiments. In some embodiments, the ratio of the linear distance between the discs and the largest disc diameter may be between about 0.7 and 2. In some embodiments, the device provides at least about 4 mesh layers per centimeter and in some embodiments, it provides from 4 to 10 mesh layers per centimeter. Optionally, the smaller diameter portions of the braid that extend from the proximal and distal discs, and that form the core portions 14 of braid between discs 12, may have a substantially uniform cylindrical shape; alternatively, the core portions 14 may have different diameters.

[0053] A short, more focal occlusion can be beneficial by keeping the occlusion zone away from branch vessels in which occlusion is not desired. In some embodiments, the device may provide a high density of mesh layers that each define a second disc angle θ_2 relative to the longitudinal axis X of the occlusion device 10 that is between about 25° and 70°, preferably between 40° and 60°, when the device assumes its partially constrained configuration upon deployment at the target site, as noted above. In some embodiments the second disc angle θ_2 (in the partially constrained state upon deployment) may be smaller than the first disc angle θ_1 (in the relaxed or expanded state) by about 20° to 60° implanted state.

[0054] When the device 10 is deployed (implanted) at the target site, each of the discs 12 may create short, substantially circular contact regions between the peripheral edge 16 and the vessel wall, as shown in Fig. 4. Accordingly, the device 10 may provide a plurality of short and closely spaced substantially circular contact regions. In some embodiments, for example, there may be between 3 and 7 such contact regions within a linear distance of between about 1 cm and 3 cm within the blood vessel, preferably less than about 2.5 cm. In some embodiments, for

example, there may be 2-4 contact regions per cm. The substantially circular contact regions are defined by the thickness of the peripheral edge 16 of each disc 12, typically less than about 2 mm, and preferably less than about 1 mm, corresponding to very short lengths along the longitudinal axis X. When the device 10 is implanted in a bodily lumen, as shown in Fig. 4, an occlusion zone Z in the lumen may be defined as the length of the lumen between the most distal and the most proximal occlusion discs 12 (as measured between their respective peripheral edges 16). Because of the small size of the contact regions (as defined above) relative to the length of the occlusion zone Z, when the device 10 is implanted, only a relatively small portion of the surface area of the occlusion zone Z is contacted by the device. In some embodiments, for example, the surface area of contact is less than about 25% of the total surface area of the occlusion zone Z. The total occlusion zone provided by the device 10 within a vessel is thus shorter and more focal than has heretofore been the norm, thereby yielding the aforementioned advantage over devices that provide relatively long occlusion zones and large vessel contact areas.

[0055] One or both of the sides of any or all of the discs may be concave or convex. In some embodiments, as shown in Fig. 9, one or more discs 12' may comprise a concavity 31 formed into a cup-like shape that faces downstream, so that blood flow and pressure in the vascular site to be embolized tends to increase the outward force against the vessel that may enhance the friction force holding the device in place, and therefore improve its stability and reduce the risk of downstream movement of the device. Alternatively, the discs may have a simple planar structure or have a small angulation so that they are thicker at the base than at the apex 16, as shown in Figure 3. Disc and core portion diameters, shapes and dimensions vary to cover the wide range of defect sizes and shapes. In some embodiments the disc diameter may range between 2 mm and 12 mm, and the core portion diameter may range between 0.25 mm and 3 mm.

[0056] As shown in Fig. 3, in some embodiments, the peripheral edges or apices 16 of the radially-extending discs 12 of the occlusion device may advantageously have an edge radius R_E of between about 0.10 mm and 0.40 mm. The core portions 14 between adjacent discs may advantageously define a core radius of R_C of between about 0.25 mm and 1.0 mm. These specific

ranges of disc edge and core radii for the occlusion device may provide a heretofore unknown desirable balance of deliverability and maintenance of the close spacing of the discs. Smaller radii increase the force required for delivery through a microcatheter that may be necessary for treatment of tortuous and small vessel targets. Conversely, larger disc edge radii, in particular, may result in a longer implant and less focal occlusion when implanted. A shorter device and more focal occlusion are generally desirable to the user (physician). Further, larger spacing of the discs may result in longer occlusion times.

[0057] In any of the embodiments described herein, the device components may comprise a mesh of wires, filaments, threads, sutures, fibers or the like (herein called "filaments 30") that have been configured to form a porous fabric or structure. The filaments 30 may be constructed using metals, polymers, composites, and/or biologic materials. Polymer materials may also include polymers such as polyester, polypropylene, nylon, PTFE, ePTFE, TFE, PET, TPE, PGA, PGLA, or PLA. Other suitable materials known in the art of elastic implants may be used. Metallic materials may include, but are not limited to, nickel-titanium alloys (e.g. Nitinol), platinum, cobalt-chrome alloys such as Elgiloy, stainless steel, tungsten, or titanium.

[0058] In some cases, the specific construction of a drawn filled tube wire or filament may be important to enhance the external visualization and/or maintain desired performance characteristics of the occlusion device. More specifically, it may be important to balance the stiffness, elasticity and radiopacity of the composition. In particular, for drawn filled tube filament embodiments that include an internal wire of ductile radiopaque material, such as platinum, and an outer tube of an elastic or super-elastic material such as NiTi, it can be necessary (or at least advantageous) to balance the ratio of the percent cross-sectional area of the internal wire with regard to the overall cross-sectional area of the filament. Such a ratio may be referred to as a "fill ratio." If an embodiment includes too little radiopaque internal tube material relative to the external tube material, there may not be sufficient radiopacity and visibility. On the other hand, if an embodiment includes too much internal wire material with respect to the elastic external tube, the mechanical properties of the ductile radiopaque material may overwhelm the elastic properties of the outer tube material, and the filaments may be prone to taking a set after compression, etc., resulting in permanent deformation. For some embodiments, a desired

composite or drawn filled tube filament may be constructed with a fill ratio of cross-sectional area of internal fill wire to cross-sectional area of the entire composite filament of between about 10% and about 50%, more specifically between about 20% and about 40%, and even more specifically, between about 25% and about 35%.

[0059] In some embodiments, the devices for treatment of a patient's vasculature may have at least about 25% composite filaments relative to the total number of filaments, and in some embodiments such devices may have at least about 40% composite filaments relative to a total number of filaments in the device. For example, a first subset of elongate resilient filaments may comprise composite materials, each having a combination of highly radiopaque material and a high strength material, and a second subset of elongate resilient filaments may comprise mostly a high strength material. For example, the highly radiopaque material may comprise platinum, platinum alloy such as 90% platinum/10% iridium, or gold, or tantalum. The high strength material may comprise NiTi. While composite wires may provide enhanced visualization and/or mechanical characteristics, they may, in some configurations, have reduced tensile strength in comparison to NiTi wires of a similar diameter. In other configurations, depending on their diameter, the composite wires may increase the collapsed or compressed profile of the devices. Therefore, it may be beneficial to minimize the number of composite wires. Lower percentages of composite wires may not be sufficiently visible with current imaging equipment, particularly in neurovascular applications where the imaging is done through the skull. In addition, too many composite wires (or composite wires with extremely high fill ratios) may result in devices with excessive artifact on CT or MRI imaging. The described ratios and amounts of highly radiopaque material provide a unique situation for neurovascular implants where the periphery of the device is just visible under transcranial fluoroscopy, but the device imaged area is not completely obliterated (i.e., due to an artifact), as it is with conventional embolic coils that are made substantially out of platinum or platinum alloys. One manner of achieving the desired degree of radiopacity is by selecting a particular combination of fill ratio of the composite filaments and the percent of composite filaments in relation to the total number of wires. The total percent of radiopaque platinum, for example, in the braided structure may be between about 15% and 30%.

[0060] In any of the braid filaments described herein, the cross-section may be circular, ovoid, rectangular or square. In some embodiments, the filaments 30 may be formed initially with a circular cross-section, and then rolled or partially flattened to form a flattened filament 30' having a generally rectangular cross-section with rounded corners, as shown in Figure 7. This configuration may provide improved ability to braid the desired configuration of braid angle and pore structure.

[0061] Specifically, for a given filament thickness, the larger the “pick length” (distance between filament crossings), the longer the beam length relative to the filament displacement, and the lower the stress in the filament induced by the displacement. Also, for a given pick length, the smaller the filament thickness, the larger the pick length, and the longer the beam length relative to the filament displacement, the lower the stress in the filament induced by the displacement. If a circular filament is subsequently flattened along its axis, forming two parallel planes symmetric about the filament axis, then the filament thickness will be reduced without reducing the cross-section and thus the tensile strength of the filament.

[0062] Flattening the filament requires less beam displacement to form the braid. Therefore, one can expect smaller pick length for a given filament tension. Accordingly, reducing pick length reduces the “pore” size of the weave. Also, the flattened filament is “wider” than the diameter of the circular filament, and as such, occupies more of the pore surface area for any equal pick length, further reducing pore size.

[0063] Bending stiffness is $E*I$, where E is the tensile modulus, and I is the polar moment of inertia of the cross-section normal to the bending axis. E is the same for both flattened (elliptical) and circular filaments. The value of I of the flattened filament is 22% that of the circular filament of the same cross-sectional area. Therefore, the flattened filament is only 22% as stiff. Under the same tensile load, the pick length would be reduced based on this factor alone. Flattening a circular filament will reduce the braided pore size due to smaller bending displacement, wider cross-section, and reduced bending stiffness.

[0064] In some embodiments, the device may comprise a braided mesh of fine filaments. In some embodiments, the braided mesh may be formed over a mandrel 40, as is known in the art of tubular braid manufacturing and shown in Figure 5. The mesh is formed with a braid angle

α relative the longitudinal axis Y of the mandrel 40. The braid angle α may be controlled by various means known in the art of filament braiding. The braids for the mesh components may have a generally constant braid angle α over the length of a component, or the braid angle α may be varied to provide different zones of pore size and radial stiffness. The “pore size” is defined as the largest circle that may be inscribed inside the diamond formed by the crossed filaments, as indicated by the dashed circle P in Figure 8.

[0065] In some embodiments, as shown in Fig. 6, substantially non-structural filamentary or fibrous members 32 may be attached or interwoven into the structural filaments 30 of a portion of the device to increase a resistance to the flow of blood through the mesh structure or enhance the formation of thrombus and/or healing of the tissue around the device. In some embodiments, one or more fibers or yarns 32 may be interwoven or interlaced into the braided mesh, as shown in FIG. 6. The non-structural fibers 32 may be interlaced into the braided mesh during the braiding process or after. The non-structural fibers 32, which may be microfibers or any other suitable fibers, may be polymeric and may beneficially be more thrombogenic than the mesh filaments. The non-structural fibers 32 may include, but are not limited to, any of the materials described herein.

[0066] In some embodiments, the tubular braided filament mesh structure can be formed by a braiding machine. It may be advantageous to utilize a braiding machine that does not employ bobbins or other filament spooling and tensioning mechanisms, typical of many conventional braiders, as they make braiding of very fine filaments and mixing of different filament diameters more difficult. Low tension braiding machines can allow less filament breakage and are more amenable to mixed diameter filament braid. In some embodiments, the braided filament mesh structure can be braided using methods or devices described in one or more of: U.S. Patent No. 8,833,224, entitled “BRAIDING MECHANISM AND METHOD OF USE”; U.S. Patent No. 8,826,791, entitled “BRAIDING MECHANISM AND METHOD OF USE”; U.S. Patent No. 8,261,648, entitled “BRAIDING MECHANISM AND METHOD OF USE”; U.S. Patent No. 8,820,207, entitled “BRAIDING MECHANISM AND METHOD OF USE”; and U.S. Patent Publication No. 2014/0318354, entitled “BRAIDING MECHANISM AND METHOD OF USE”. The disclosures of all of the aforementioned patents are hereby incorporated by reference herein.

[0067] The tubular braided mesh may then be further shaped using a heat setting process. As is known in the art of heat setting nitinol wires, a fixture, mandrel or mold may be used to hold the braided tubular structure in its desired configuration, then subjected to an appropriate heat treatment, such that the resilient filaments of the braided tubular member assume or are otherwise shape-set to the outer contour of the mandrel or mold. A conventional heat treatment fixture 50 is shown in Figure 10, in which a mesh device or component 10 is held. The fixture 50 is configured to hold the device or component in a desired shape, and the device or component is heated while so held to about 475-525° C. for about 5-30 minutes to shape-set the structure. The heat setting process may be performed in an oven or fluidized bed, as is well-known. In some embodiments, the heat setting process may be done in an inert gas environment. In some embodiments, a system for heat treating a device comprises a chamber configured to contain bath media, a container within the chamber and configured to hold the device, an air inlet gate fluidly upstream of the chamber and configured to be coupled to a gas source to flow gas into the chamber to fluidize the bath media, and a heating element between the air inlet gate and the chamber. The bath media may advantageously and optionally include sand. The gas source may comprise air, nitrogen, hydrogen, carbon monoxide, or any combination thereof. The system may further comprise thermal sensors electrically connected to a temperature regulator. In some embodiments, a system for heat treating a device comprises a chamber configured to contain bath media, and a container within the chamber and configured to hold the device. Gas flow into the chamber is provided so as to fluidize the bath media.

[0068] In some embodiments, the device can be formed at least in part from a cylindrical braid of elastic filaments. Thus, the braid may be radially constrained without plastic deformation and will self-expand on release of the radial constraint. Such a braid of elastic filaments is herein referred to as a “self-expanding braid.” In accordance with the present disclosure, braids of generally smaller filaments than current occlusion devices are preferably used. In some embodiments, the thickness of the braid filaments would be less than about 0.10 mm. In some embodiments, the braid may be fabricated from wires with diameters ranging from about 0.015 mm to about 0.080 mm.

[0069] In some embodiments, the diameter of the braid filaments can be less than about 0.5 mm. In other embodiments, the filament diameter may range from about 0.01 mm to about 0.40 mm. In some embodiments, braid filaments of varying diameters may be combined in the mesh to impart different characteristics including, e.g., stiffness, elasticity, structure, radial force, pore size, occlusive ability, and/or other features. In some embodiments, larger, structural filaments may be fabricated from wires with diameters ranging from about 0.025 mm to about 0.25 mm. In some embodiments, the thickness of the smaller braid filaments would be between about 0.01 mm about 0.05 mm. The ratio of the number of small filaments to the number of large filaments may be between about 2 and 20, such as, for example, between 4 and 12. An exemplary embodiment comprising filaments of two different sizes (cross-sectional areas) is shown, for example, in Fig. 8, in which an array of larger filaments 30L is integrated into an array of smaller filaments 30S.

[0070] Prior to treatment, a device size is chosen based on the size and morphology of the site to be embolized. In general, the device is generally selected so that the diameter of its occlusion regions in the expanded state is larger than the diameter or largest dimension of the site, and, thus over-sized, provides a residual radial force, lending stability to the device. When deployed in a vascular site in an over-sized and at least partially compressed state, the device will assume a constrained configuration, in which the shape of the discs 12 will be axially elongated as compared to the expanded or relaxed state of the device, thereby decreasing the disc angle relative to the longitudinal axis of the device, as described above. The device may be over-sized diametrically relative to the target vessel diameter by between about 10% and 100% in some embodiments, and preferably between about 25% and 75%. When deployed in an over-sized and thus partially compressed state, the discs assume a constrained configuration, forming a plurality of mesh layers (i.e., two for each disc 12), in which at least a portion of the external surface of each disc 12 forms the aforementioned second disc angle θ_2 with respect to the longitudinal axis X of the device (which will generally correspond with the local longitudinal axis of the vascular site, e.g. blood vessel). As mentioned above, the second disc angle θ_2 will typically be measurably smaller than 70°, such as, for example, between about 40° and 60° relative to the longitudinal axis X of the device. In some embodiments, the device 10 has a plurality of layers

packed or configured within a linear distance that provides a short, focal occlusion zone Z of less than about 2.5 cm in length, such as, for example, between about 0.5 cm and 2.0 cm, as measured between the apices 16 of the most distal and most proximal discs 12 when the occlusion device is implanted, as shown in Figure 4.

[0071] The shape and porosity of the device work synergistically to provide vascular occlusion and a biocompatible scaffold to promote new tissue ingrowth. These functions can be influenced by the “pore size” or “weave density” generally described as the pick count (number of interlacings) per unit length of the material. The devices in accordance with the present disclosure provide generally higher wire counts than current occlusion devices and thus smaller pore sizes, yielding improved occlusion performance, and thus may obviate the need for polymer fabric components that can increase thromboembolic risk due to clot formation, while also increasing the collapsed profile or diameter of the device. Pore sizes in the range of about 0.05 mm to 0.50 mm may be utilized. In some embodiments, the pore size may be in the range of 0.10 mm to 0.25 mm. The relatively higher wire counts provide a large number of filament or wire radial traverses of the target vessel that is being occluded. Since each disc provides two filament radial traverses from the central core portion to the apex of the disc, the total number of wire radial traverses in a device is equal to the number of filaments or wire times the number of discs times 2. In some embodiments, each radially extending occlusion region or disc defines two radial traverses for each wire, wherein the total number of radial traverses for the device is between 720 and 2000.

[0072] In any of the embodiments herein, the device 10 may contain at least one tine, barb, hook, pin or anchor (hereinafter called “barbs”, not shown) that may be incorporated into the structure to help provide additional fixation of the device to the vessel wall or other tissue. The length of the barbs may be from about 0.25 to 3 mm, and preferably between about 0.5 to 2 mm.

[0073] Optionally, the occlusion device may be constructed to provide the elution or delivery of one or more beneficial drugs and/or other bioactive substances into the blood or the surrounding tissue. Optionally, the device may be coated with various polymers to enhance its performance, fixation, and/or biocompatibility. Optionally, the device may incorporate cells

and/or other biologic material to promote fixation, sealing, reduction of leaks or healing. In any of the above embodiments, the device may include a drug or bioactive agent to enhance the performance of the device and/or healing of the target site and nearby tissue. In some embodiments, at least a portion of the device may include: an antiplatelet agent, including but not limited to aspirin; glycoprotein IIb/IIIa receptor inhibitors (including one or more of abciximab, eptifibatide, tirofiban, lamifiban, fradafiban, cromafiban, toxifiban, XV454, lefradafiban, klerval, lotrafiban, orbofiban, and xemilofiban), dipyridamole, apo-dipyridamole, persantine, prostacyclin, ticlopidine, clopidogrel, cromafiban, cilostazol, and/or nitric oxide. In any of the above embodiments, the device may include an anticoagulant, such as heparin, low molecular weight heparin, hirudin, warfarin, bivalirudin, hirudin, argatroban, forskolin, ximelagatran, vaprost, prostacyclin and prostacyclin analogues, dextran, synthetic antithrombin, Vasoflux, argatroban, efegatran, tick anticoagulant peptide, Ppack, HMG-CoA reductase inhibitors, and thromboxane A2 receptor inhibitors. Device embodiments herein may include a surface treatment or coating on a portion, side surface, or all surfaces, wherein the surface treatment or coating promotes or inhibits thrombosis, clotting, healing or other embolization performance measures. The surface treatment or coating may be a synthetic, biologic or combination thereof. For some embodiments, at least a portion of a surface of the device may have a surface treatment or coating made of a biodegradable or bioresorbable material such as a polylactide, polyglycolide or a copolymer thereof. Another surface treatment or coating material which may enhance the embolization performance of a device includes a polysaccharide, such as an alginate-based material. Some coating embodiments may include extracellular matrix proteins such as ECM proteins. One example of such a coating may be Finale Prohealing coating which is commercially available from Surmodics Inc., Eden Prairie, Minn. Another exemplary coating may be Polyzene-F, which is commercially available from CeloNovo BioSciences, Inc., Newnan, Ga. In some embodiments, the coatings may be applied with a thickness that is less than about 25% of a transverse dimension of the filaments.

[0074] The terms “formed”, “preformed” and “fabricated” may include the use of molds or tools that are designed to impart a shape, geometry, bend, curve, slit, serration, scallop, void, hole in the elastic, super-elastic, or shape memory material or materials used in the components

of the device. These molds or tools may impart such features at prescribed temperatures or heat treatments.

[0075] An exemplary low force detachment system may comprise a mechanism for deploying an occlusion device in accordance with this disclosure as an implant in a bodily lumen, wherein the implant is delivered to a luminal site through a catheter by means of a pusher implement having a distal end to which a proximal end of the implant is detachably connected by a filamentous tether (not shown). The tether extends from a first end, through a “fenestration feature” (e.g., an eyelet or a loop) on the proximal end of the device, to a second end. The first end of the tether is connected to a tether take-up assembly in a tether control device, and the second end of the tether is releasably captured in the tether control device by a tether retention assembly that is operable to release the second end when the device is deployed. Upon release of the second end by the retention assembly, the take-up assembly withdraws the tether through the fenestration feature until the tether is free of the fenestration feature. The system may preferably detach the device instantaneously or nearly so, that is, in less than about 1 second. The system may preferably detach the device by mechanical means, but without the transmission of torque in or through the delivery catheter or pusher. In some embodiments, the system may comprise an interlocking loops detachment system where a user can disengage an occlusion device through a single action on the retraction device that preferably, only acts on the implant pusher guidewire during a retraction phase of movement, and allows for resetting of the retraction device for retraction without reinsertion of the guidewire. In one embodiment, the system may comprise an implant comprising a proximal end and a distal end with a first loop attached at its proximal end; a sheath comprising a proximal end and a distal end with a second loop attached near its distal end; a guidewire slidably disposed within the sheath having a proximal end and distal end; wherein the guidewire, when engaged through the first and second loops provides a releasable connection of the implant to the sheath that can be released by retraction of the guidewire; a handle handpiece comprising a retraction mechanism adapted to receive the proximal ends of the sheath and guidewire; the retraction mechanism comprising a finger actuator, a ratcheting device, a first spring member and second spring member; the ratcheting device comprising a rotatable lever, a gripping insert, an anvil, a guide arm and a guide

ramp; wherein the first spring member biases the rotatable lever against the guidewire and the second spring member provides a biasing force against retraction of the finger actuator; wherein the gripping insert, when actuated by the finger actuator, acts on the guidewire to retract the guidewire within the sheath and release the implant.

[0076] For some embodiments, the detachment of the device from the delivery apparatus of the delivery system may be effected by the delivery of energy (e.g., electrical current, heat, radiofrequency, ultrasound, vibrational, or laser) to a junction or release mechanism between the device and the delivery apparatus.

[0077] Disclosed herein is a detailed description of various illustrated embodiments of the disclosed subject matter. This description is not to be taken in a limiting sense, but is made merely for the purpose of illustration of the general principles of such subject matter. Further features and advantages of the subject matter of this disclosure will become apparent to those of skill in the art in view of the description of embodiments disclosed, when considered together with the attached drawings.

[0078] Although several embodiments of the subject matter of this disclosure have been described in terms of particular embodiments and applications, one of ordinary skill in the art, in light of this teaching, can generate additional embodiments and modifications without departing from the spirit of or exceeding the scope of the disclosure. Accordingly, it is to be understood that the drawings and descriptions herein are proffered by way of example to facilitate comprehension of the disclosure and should not be construed to limit the scope thereof.

CLAIMS

1. A device for occluding a bodily lumen, comprising:

 a mesh structure comprising a plurality of braided filaments, the mesh structure defining a longitudinal axis extending between a proximal end and a distal end, the filaments of the mesh structure being fixed in a proximal hub at the proximal end and in a distal hub at the distal end;

 wherein the mesh structure is radially compressible from a fully-expanded relaxed configuration to a compressed configuration, and is resiliently expandable from the compressed configuration to a partially-expanded, constrained configuration when deployed in the lumen;

 wherein the mesh structure in the relaxed configuration includes a plurality of radially extending occlusion regions spaced along the longitudinal axis, each of the occlusion regions having a side defining an angle of 70° to 85° relative to the longitudinal axis of the structure; and

 wherein, in the constrained configuration of the mesh structure, each radially extending occlusion region defines two radial traverses of the lumen for each filament, wherein the total number of radial traverses is between 720 and 2000.

2. The device of claim 1, wherein the mesh structure is a single layer continuous tubular braided structure.

3. The device of claim 1, wherein each radially extending occlusion region has a first diameter and is separated from the nearest adjacent radially extending occlusion region by a connecting portion of the mesh structure having a second diameter less than the first diameter.

4. The device of claim 1, wherein the mesh structure in the relaxed configuration comprises at least three radially extending occlusion regions.

5. The device of claim 1, wherein the mesh structure defines an axial length along the longitudinal axis, and wherein the occlusion regions provide between 4 and 10 mesh layers per centimeter of the axial length.

6. The device of claim 1, wherein the angle when the mesh structure is in the constrained configuration is smaller than the angle when the mesh structure is in the relaxed configuration.

7. The device of claim 1, wherein each of the occlusion regions has a peripheral edge at which each of the occlusion regions has a thickness of not more than about 2 mm.

8. The device of claim 7, wherein the device has an axial length along the longitudinal axis, and wherein the device comprises between three and seven radially extending occlusion regions, whereby the device includes between three and seven peripheral edges per centimeter of the axial length.

9. A device for occluding a bodily lumen, comprising:

 a mesh structure defining an axial length along a longitudinal axis between a proximal end and a distal end, the mesh structure comprising a braided filament mesh extending between the proximal end and the distal end, the braided filament mesh comprising a plurality of braided filaments terminating at the proximal end in a proximal hub and at the distal end in a distal hub;

 wherein the mesh structure has a radially compressed configuration and a radially expanded relaxed configuration;

 wherein the mesh structure in the relaxed configuration has a plurality of radially extending occlusion regions having a first diameter spaced from each other along the longitudinal axis, each radially extending occlusion region being axially separated from the nearest adjacent radially extending occlusion region by a connecting region having a second diameter less than the first diameter, each radially extending occlusion region having a proximal side and a distal side meeting at a peripheral edge; and

wherein the occlusion regions provide between 4 and 10 mesh layers per centimeter of the axial length.

10. The device of claim 9, wherein the mesh structure is a single layer continuous braided structure.

11. The device of claim 9, wherein the mesh structure in the relaxed configuration has at least three radially extending occlusion regions.

12. The device of claim 9, wherein at least one of the proximal side and the distal side defines an angle of between 70° and 85° relative to the longitudinal axis in the relaxed configuration.

13. The device of claim 9, wherein the mesh structure assumes a partially-expanded constrained configuration within the bodily lumen, whereby, in the constrained configuration, each radially extending occlusion region defines two radial traverses of the lumen for each filament, wherein the total number of radial traverses is between 720 and 2000.

14. The device of claim 9, wherein the mesh structure assumes a partially-expanded constrained configuration within the bodily lumen, and wherein the angle when the mesh structure is in the constrained configuration is smaller than the angle when the mesh structure is in the relaxed configuration.

15. The device of claim 9, wherein each of the occlusion regions has a peripheral edge at which each of the occlusion regions has a thickness of not more than about 2 mm.

16. The device of claim 15, wherein the device comprises between three and seven radially extending occlusion regions, whereby the device includes between two and four peripheral edges per centimeter of the axial length.

17. A device for occluding a bodily lumen defined by an endoluminal wall, the device comprising:

a mesh structure having a longitudinal axis between a proximal end and a distal end, the mesh structure comprising a braided filamentous mesh extending between the proximal end and the distal end, the mesh comprising a plurality of braided filaments, each of which terminates at the proximal end in a proximal hub and at the distal end in a distal hub;

wherein the mesh structure has a radially compressed configuration and a relaxed configuration;

wherein the mesh structure in the radially compressed configuration has an outside diameter not exceeding 0.7 mm for delivery to a target site in the bodily lumen; and

wherein the mesh structure in the relaxed configuration has a plurality of radially extending occlusion regions, wherein each radially extending occlusion region has a proximal side and a distal side that meet at a peripheral edge, wherein at least one of the proximal side and the distal side forms a first angle relative to the longitudinal axis of the mesh structure when in the relaxed configuration, and a second angle less than the first angle when the mesh structure is radially constrained by contact between the peripheral edge and the endoluminal wall.

18. The device of claim 17, wherein the mesh structure is a single layer continuous braided structure.

19. The device of claim 17, wherein the mesh structure in the relaxed configuration has at least three radially-extending occlusion regions.

20. The device of Claim 17, wherein the first angle is between about 70° and 85°, and the second angle is less than about 70°.

21. The device of claim 17, wherein the mesh structure assumes a partially-expanded constrained configuration when in the bodily lumen, whereby, in the constrained configuration,

each radially extending occlusion region defines two radial traverses of the lumen for each filament, wherein the total number of radial traverses is between 720 and 2000.

22. The device of claim 17, wherein the mesh structure defines an axial length along the longitudinal axis, and wherein the occlusion regions provide between 4 and 10 mesh layers per centimeter of the axial length.

23. The device of claim 17, wherein the peripheral edge of each of the occlusion regions has a thickness of not more than about 2 mm.

24. The device of claim 23, wherein the device has an axial length along the longitudinal axis, and wherein the device comprises between three and seven radially extending occlusion regions, whereby the device includes between three and seven peripheral edges per centimeter of the axial length.

25. A method for occluding a blood vessel, comprising the steps of:

(a) providing a catheter dimensioned to be introduced intravascularly to a target vascular site; the catheter having a lumen between a distal open end and a proximal open end;

(b) providing an occlusion device having a longitudinal axis between a proximal device end and a distal device end, the occlusion device comprising a braided filament mesh extending between the proximal device end and the distal device end, the filaments terminating at the proximal device end in a proximal hub and at the distal device end in a distal hub, wherein the occlusion device has a relaxed configuration in which the occlusion device has a plurality of radially-extending occlusion regions, wherein each radially-extending occlusion region has a proximal side and a distal side that meet at a peripheral edge;

(c) advancing the catheter intravascularly to position its distal open end in or near the target vascular site;

(d) advancing the occlusion device, in a radially-compressed configuration, distally through the lumen of the catheter until the occlusion device emerges from the open distal end of the catheter for deployment in the target vascular site;

(e) allowing the occlusion device to expand radially at the target vascular site until the peripheral edge of each radially-extending occlusion region engages an endoluminal surface of the target vascular site, wherein when the occlusion device is deployed, each radially extending occlusion region defines two radial traverses of the lumen for each filament, wherein the total number of radial filament traverses is between 720 and 2000; and

(f) maintaining the occlusion device in the target vascular site at least until the target vascular site is embolized to occlude blood flow through the target vascular site.

26. The method of claim 25, wherein the occlusion device is a single layer continuous braided structure.

27. The method of claim 25, wherein the braided filament mesh has a pick angle of at least 40°.


28. The method of claim 25, wherein the occlusion device has a maximum relaxed configuration diameter that is over-sized relative to the diameter of the lumen of the catheter by between about 10% and 150%.

29. The method of claim 25, wherein the occlusion device defines an axial length along the longitudinal axis, and wherein the occlusion regions provide between 4 and 10 mesh layers per centimeter of the axial length.

30. The method of claim 25, wherein the occlusion device, when deployed in the bodily lumen, defines an occlusion zone in the lumen between the peripheral edge of a most distal one of the occlusion regions and the peripheral edge of a most proximal one of the occlusion regions, the occlusion zone defining a surface area within the lumen, wherein the peripheral edges of the

occlusion regions contact less than about 25% of the surface area defined by the occlusion zone.

1/10

**FIG. 1**

2/10




FIG. 2

3/10




FIG. 3

4/10







FIG. 4



5/10

**FIG. 5**

6/10

**FIG. 6****FIG. 7**

7/10

**FIG. 8****FIG. 9**

8/10




FIG. 10

9/10




FIG. 11

10/10




FIG. 12

INTERNATIONAL SEARCH REPORT

International application No.
PCT/US2018/019532

Box No. II Observations where certain claims were found unsearchable (Continuation of item 2 of first sheet)

This international search report has not been established in respect of certain claims under Article 17(2)(a) for the following reasons:

1. Claims Nos.: 25-30
because they relate to subject matter not required to be searched by this Authority, namely:
Claims 25-30 pertain to methods for treatment of the human body and thus relate to a subject-matter which this International Searching Authority is not required, under PCT Article 17(2)(a)(i) and PCT Rule 39.1(iv), to search.
2. Claims Nos.:
because they relate to parts of the international application that do not comply with the prescribed requirements to such an extent that no meaningful international search can be carried out, specifically:
3. Claims Nos.:
because they are dependent claims and are not drafted in accordance with the second and third sentences of Rule 6.4(a).

Box No. III Observations where unity of invention is lacking (Continuation of item 3 of first sheet)

This International Searching Authority found multiple inventions in this international application, as follows:

1. As all required additional search fees were timely paid by the applicant, this international search report covers all searchable claims.
2. As all searchable claims could be searched without effort justifying an additional fees, this Authority did not invite payment of any additional fees.
3. As only some of the required additional search fees were timely paid by the applicant, this international search report covers only those claims for which fees were paid, specifically claims Nos.:
4. No required additional search fees were timely paid by the applicant. Consequently, this international search report is restricted to the invention first mentioned in the claims; it is covered by claims Nos.:

Remark on Protest

The additional search fees were accompanied by the applicant's protest and, where applicable, the payment of a protest fee.

The additional search fees were accompanied by the applicant's protest but the applicable protest fee was not paid within the time limit specified in the invitation.

No protest accompanied the payment of additional search fees.

INTERNATIONAL SEARCH REPORT

International application No.
PCT/US2018/019532

A. CLASSIFICATION OF SUBJECT MATTER

A61B 17/12(2006.01)i

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)
A61B 17/12; A61M 29/00; A61B 17/221; A61B 17/22; A61D 1/02; A61M 29/02Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched
Korean utility models and applications for utility models
Japanese utility models and applications for utility modelsElectronic data base consulted during the international search (name of data base and, where practicable, search terms used)
eKOMPASS(KIPO internal) & keywords: occlusion, expanded, compressed, mesh, braid, filaments

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	US 2014-0324091 A1 (INCEPTUS MEDICAL, LLC) 30 October 2014 See paragraphs [0034]–[0068]; claims 1–20; figures 4–9D.	1–24
X	US 5827304 A (HART, CHARLES C.) 27 October 1998 See column 4, lines 14–56; claims 1–18; figures 1A–9B.	1–24
X	US 8715316 B1 (INSERA THERAPEUTICS, INC.) 06 May 2014 See column 95, line 49–column 96, line 17; column 179, line 20–column 181, line 25; claims 1–20; figures 2A–9D, 10H–10I, 29A–29C.	1–24
A	US 2004-0215230 A1 (FRAZIER, ANDREW G. C. et al.) 28 October 2004 See the whole document.	1–24
A	US 2015-0223829 A1 (MEDINA MEDICAL, INC.) 13 August 2015 See the whole document.	1–24

 Further documents are listed in the continuation of Box C. See patent family annex.

* Special categories of cited documents:
 "A" document defining the general state of the art which is not considered to be of particular relevance
 "E" earlier application or patent but published on or after the international filing date
 "L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)
 "O" document referring to an oral disclosure, use, exhibition or other means
 "P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention
 "X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone
 "Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art
 "&" document member of the same patent family

Date of the actual completion of the international search
27 June 2018 (27.06.2018)Date of mailing of the international search report
27 June 2018 (27.06.2018)Name and mailing address of the ISA/KR
International Application Division
Korean Intellectual Property Office
189 Cheongsa-ro, Seo-gu, Daejeon, 35208, Republic of Korea
Facsimile No. +82-42-481-8578

Authorized officer

CHOI, Sang Won

Telephone No. +82-42-481-8291



INTERNATIONAL SEARCH REPORT

Information on patent family members

International application No.

PCT/US2018/019532

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 2014-0324091 A1	30/10/2014	EP 2897536 A1 EP 3019097 A1 US 2014-0155908 A1 US 2015-0018859 A1 US 2015-0018860 A1 US 2015-0238207 A1 US 2015-0305756 A1 US 2015-0374391 A1 US 2016-0008014 A1 US 2016-0113666 A1 US 2016-0143721 A1 US 2016-0262790 A1 US 2016-0287276 A1 US 2017-0105745 A1 US 2017-0325839 A1 US 2017-189041 A1 US 8784434 B2 US 8968330 B2 US 9259237 B2 US 9408620 B2 US 9717519 B2 WO 2014-047650 A1 WO 2014-081892 A1 WO 2015-006782 A1 WO 2015-061365 A1	29/07/2015 18/05/2016 05/06/2014 15/01/2015 15/01/2015 27/08/2015 29/10/2015 31/12/2015 14/01/2016 28/04/2016 26/05/2016 15/09/2016 06/10/2016 20/04/2017 16/11/2017 06/07/2017 22/07/2014 03/03/2015 16/02/2016 09/08/2016 01/08/2017 27/03/2014 30/05/2014 15/01/2015 30/04/2015
US 5827304 A	27/10/1998	WO 97-17889 A1	22/05/1997
US 8715316 B1	06/05/2014	US 2014-0260928 A1 US 2014-0265030 A1 US 2014-0265096 A1 US 2014-0277082 A1 US 2014-0277083 A1 US 2014-0277098 A1 US 2015-0028005 A1 US 2015-0032120 A1 US 2015-0032121 A1 US 2015-0032122 A1 US 2015-0032146 A1 US 2015-0032147 A1 US 2016-0045212 A1 US 2018-0085136 A1 US 8715314 B1 US 8715317 B1 US 8721677 B1 US 8728116 B1 US 8728117 B1 US 8733618 B1 US 8735777 B1	18/09/2014 18/09/2014 18/09/2014 18/09/2014 18/09/2014 18/09/2014 29/01/2015 29/01/2015 29/01/2015 29/01/2015 29/01/2015 29/01/2015 18/02/2016 29/03/2018 06/05/2014 06/05/2014 13/05/2014 20/05/2014 20/05/2014 27/05/2014 27/05/2014

INTERNATIONAL SEARCH REPORT

Information on patent family members

International application No.

PCT/US2018/019532

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
		US 8784446 B1	22/07/2014
		US 8789452 B1	29/07/2014
		US 8790365 B1	29/07/2014
		US 8795330 B1	05/08/2014
		US 8803030 B1	12/08/2014
		US 8813625 B1	26/08/2014
		US 8816247 B1	26/08/2014
		US 8828045 B1	09/09/2014
		US 8845678 B1	30/09/2014
		US 8845679 B1	30/09/2014
		US 8852227 B1	07/10/2014
		US 8859934 B1	14/10/2014
		US 8863631 B1	21/10/2014
		US 8866049 B1	21/10/2014
		US 8869670 B1	28/10/2014
		US 8870901 B1	28/10/2014
		US 8870910 B1	28/10/2014
		US 8872068 B1	28/10/2014
		US 8882797 B2	11/11/2014
		US 8904914 B2	09/12/2014
		US 8910555 B2	16/12/2014
		US 8932320 B1	13/01/2015
		US 8932321 B1	13/01/2015
		US 9179995 B2	10/11/2015
		US 9833251 B2	05/12/2017
US 2004-0215230 A1	28/10/2004	EP 1620020 A2	01/02/2006
		EP 1620020 B1	08/09/2010
		US 7597704 B2	06/10/2009
		WO 2004-096060 A2	11/11/2004
		WO 2004-096060 A3	20/01/2005
US 2015-0223829 A1	13/08/2015	EP 2539012 A1	02/01/2013
		EP 2539012 A4	27/04/2016
		EP 2539012 B1	24/01/2018
		US 2011-0213403 A1	01/09/2011
		US 2015-0196744 A1	16/07/2015
		US 9211396 B2	15/12/2015
		US 9931495 B2	03/04/2018
		WO 2011-106426 A1	01/09/2011