



(86) Date de dépôt PCT/PCT Filing Date: 2009/09/01
(87) Date publication PCT/PCT Publication Date: 2010/03/11
(85) Entrée phase nationale/National Entry: 2011/03/01
(86) N° demande PCT/PCT Application No.: CA 2009/001212
(87) N° publication PCT/PCT Publication No.: 2010/025547
(30) Priorité/Priority: 2008/09/02 (US61/093,667)

(51) Cl.Int./Int.Cl. *G01N 27/327* (2006.01),
C12Q 1/00 (2006.01), *C12Q 1/68* (2006.01),
C40B 60/00 (2006.01), *G01N 27/30* (2006.01),
G01N 27/403 (2006.01), *G01N 27/416* (2006.01),
G01N 33/53 (2006.01)
(71) Demandeur/Applicant:
THE GOVERNING COUNCIL OF THE UNIVERSITY OF
TORONTO, CA
(72) Inventeurs/Inventors:
TAFT, BRADFORD, US;
SOLEYMANI, LEYLA, CA;
FANG, ZHICHAO, CA;
SARGENT, EDWARD, CA; ...

(54) Titre : MICROELECTRODES NANOSTRUCTUREES ET DISPOSITIFS DE BIODETECTION LES COMPRENANT
(54) Title: NANOSTRUCTURED MICROELECTRODES AND BIOSENSING DEVICES INCORPORATING THE SAME




FIG. 2

(57) Abrégé/Abstract:

Nanostructured microelectrodes and biosensing devices incorporating the same are disclosed herein.

(72) Inventeurs(suite)/Inventors(continued): KELLEY, SHANA, CA

(74) Agent: SMART & BIGGAR

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property Organization
International Bureau



(43) International Publication Date
11 March 2010 (11.03.2010)

(10) International Publication Number
WO 2010/025547 A1

(51) International Patent Classification:

G01N 27/327 (2006.01) *G01N 33/53* (2006.01)
G01N 27/30 (2006.01) *C12Q 1/68* (2006.01)
G01N 27/403 (2006.01) *C12Q 1/00* (2006.01)
G01N 27/416 (2006.01) *C40B 60/00* (2006.01)

(21) International Application Number:

PCT/CA2009/001212

(22) International Filing Date:

1 September 2009 (01.09.2009)

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

61/093,667 2 September 2008 (02.09.2008) US

(71) Applicant (for all designated States except US): **THE GOVERNING COUNCIL OF THE UNIVERSITY OF TORONTO** [CA/CA]; 27 King's College Circle, Toronto, Ontario M5S 1A1 (CA).

(72) Inventors; and

(75) Inventors/Applicants (for US only): **SOLEYMANI, Leyla** [CA/CA]; Unit 702, 20 Collier Street, Toronto, Ontario M4W 3Y4 (CA). **FANG, Zhichao** [CN/CA]; Room

312, 666 Ontario Street, Toronto, Ontario M4X 1N1 (CA). **KELLEY, Shana** [US/CA]; 421 Walmer Road, Toronto, Ontario M5P 2X9 (CA). **SARGENT, Edward** [CA/CA]; 421 Walmer Road, Toronto, Ontario M5P 2X9 (CA). **TAFT, Bradford** [US/US]; Unit 419, 650 Delancey Street, San Francisco, California 94107 (US).


(74) Agents: **OGILVY RENAULT LLP/S.E.N.C.R.L., s.r.l.** et al.; Suite 1600, 1981 McGill College Avenue, Montréal, Québec H3A 2Y3 (CA).

(81) Designated States (unless otherwise indicated, for every kind of national protection available): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IS, JP, KE, KG, KM, KN, KP, KR, KZ, LA, LC, LK, LR, LS, LT, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PE, PG, PH, PL, PT, RO, RS, RU, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.

(84) Designated States (unless otherwise indicated, for every kind of regional protection available): ARIPO (BW, GH, GM, KE, LS, MW, MZ, NA, SD, SL, SZ, TZ, UG, ZM,

[Continued on next page]

(54) Title: NANOSTRUCTURED MICROELECTRODES AND BIOSENSING DEVICES INCORPORATING THE SAME



(57) Abstract: Nanostructured microelectrodes and biosensing devices incorporating the same are disclosed herein.

FIG. 2

WO 2010/025547 A1



ZW), Eurasian (AM, AZ, BY, KG, KZ, MD, RU, TJ, **Published:**
TM), European (AT, BE, BG, CH, CY, CZ, DE, DK, EE, — *with international search report (Art. 21(3))*
ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV,
MC, MK, MT, NL, NO, PL, PT, RO, SE, SI, SK, SM,
TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW,
ML, MR, NE, SN, TD, TG).

NANOSTRUCTURED MICROELECTRODES AND BIOSENSING DEVICES INCORPORATING THE SAME

RELATED APPLICATION

This application claims priority to U.S. Provisional Application No. 61/093,667,
5 filed on September 5, 2008, entitled Nanostructured Microelectrodes and Biosensing
Devices Incorporating the Same, which is incorporated by reference herein, in its entirety.

BACKGROUND OF THE INVENTION

Genomic analysis is revolutionizing early disease diagnosis and dramatically
enhancing patient care (McGuire et al. *Science* 317 :1687, Srinivas et al., *Lancet Oncol.* 2
10 :698). Microarrays (Drmanac et al., *Science* 260:1649, Hacia et al., *Nat. Genet.* 14:441) and
polymerase chain reaction (PCR)-based techniques (Saiki et al., *Science* 230:1350) have, as
tools, helped to spearhead this revolution, enabling the discovery and the initial
development of assays for patient testing (Morris et al., *Curr. Opin. Oncol.* 19:547).
However, spreading the reach of the genomics revolution to the patient bedside demands
15 cost effective tools for individual biomarker profiling assessed relative to a posited disease
state. Specifically, tools enabling routine patient care preferably would be simpler, more
portable, and less expensive than PCR-based methods, yet should retain a high degree of
selectivity and sensitivity.

Biomarker analysis based on electronic readout has long been cited as a promising
20 approach that would enable a new family of chip-based devices with appropriate cost and
sensitivity for medical testing (Drummond et al., *Nat. Biotechnol.* 21:1192, Katz et al.,
Electroanalysis 15: 913). The sensitivity of electronic readout is in principle sufficient to
allow direct detection of small numbers of analyte molecules with simple instrumentation.
However, despite tremendous advances in this area as well as related fields working
25 towards new diagnostics (Clack et al., *Nat. Biotechnol.* 26:825, Geiss et al., *Nat.*
Biotechnol. 26 :317, Hahm et al., *Nano Lett.* 4 :51, Munge et al., *Anal. Chem.* 77 :4662,
Nicewarner-Pena et al., *Science* 294 :137, Park et al., *Science* 295 :1503, Sinensky et al.,
Nat. Nano. 2 :653, Steemers et al., *Nat. Biotechnol.* 18:91, Xiao et al., *J. Am. Chem. Soc.*
129:11896, Zhang et al., *Nat. Nano.* 1:214, Zhang et al., *Anal. Chem.* 76 :4093, Yi et al.,
30 *Biosens. Bioelectron.* 20:1320, Ke et al., *Science* 319:180, Armani et al., *Science* 317:783),

current multiplexed chips have yet to achieve direct electronic detection of biomarkers in cellular and clinical samples. The challenges that have limited the implementation of such devices primarily stem from the difficulty of obtaining very low detection limits in the presence of high background noise levels present when complex biological samples are assayed, and the challenge of generating multiplexed systems that are highly sensitive and specific.

The miniaturization of electrochemical systems continues to be a major focus in analytical and bioanalytical chemistry (Matysik, *Miniaturization of Electroanalytical Systems* (Springer-Verlag, 2002)), as the attainment of enhanced sensitivity may be enabled with systems possessing micro-to nano- scale dimensions (Szamocki et al., *A. Anal. Chem.* 2007, 79, 533-539). A great deal of work has been carried out with electrodes with dimensions on the micrometer or sub-micrometer scale. These systems offer many advantages over conventional macroelectrodes such as faster double-layer charging, reduced ohmic loss, high mass-transport rates, and high current density (Bond et al. *Anal. Chimi. Acta* 1989m 216, 177-230, Heinze, *Angew. Chem. Int. Ed.* 2003, 32, 1268-1288). Indeed, such electrodes have become well-established tools in a wide range of analytical applications (Bard, *Electrochemical Methods: Fundamentals and Applications* (Wiley, New York, 2001), Reimers, *Chem. Rev.* 2007, 107, 590-600, Zosic, *Handbook of electrochemistry* (Elsevier, 2007)). Working with nanoscale electrodes, however, is significantly more challenging, as fabrication is typically labour-intensive, insufficiently reproducible, and the currents obtained from such structures are typically difficult to measure accurately.

The use of nanowire electrodes for ultrasensitive nucleic acids and protein detection has been investigated (Gasparac et al. *J Am Chem Soc* 126:12270). The use of this electrode platform enables the electrochemical detection of picomolar levels of analytes, a level of sensitivity that is not possible using macroscale materials. Although it has been reported that nanowires are able to detect attomolar levels of analytes, this actually corresponds to picomolar levels when dealing with the volumes typically used for analysis. It has also been demonstrated that nanoparticle-modified electrodes may exhibit several advantages over conventional macroelectrodes such as enhancement of mass transport, catalysis, high effective surface area and control over electrode microenvironment (Katz et al.

Electroanalysis 2004, 16, 19-44, Welch et al. *Anal. Bioanal. Chem.* 2006, 384, 601-619). Manufacturing arrays of nanowire electrodes, however, is non-trivial.

Boron doped diamond microelectrodes modified by electrodeposition of platinum nanoparticles have been used for the oxidative determination of As(III) at levels below 1
5 ppb (Hrapovic et al. *Anal. Chem.* 2007, 79, 500-507). However, this type of electrode cannot be incorporated into an array-based format for multiplexed experiments.

The analysis of panels of nucleic acid or protein biomarkers offers valuable diagnostic and prognostic information for clinical decision making. Existing methods that offer the specificity and sensitivity to profile clinical samples are typically costly, slow and
10 serial. There is thus a need for an ultrasensitive device for detecting biomarkers in a multiplexed fashion.

SUMMARY OF THE INVENTION

In one aspect, the invention features nanostructured microelectrodes (NMEs). NMEs are electrodes, which are nanotextured and thus have an increased surface area.
15 Preferred NMEs are comprised of a noble metal, (e.g. gold, platinum, palladium, silver, osmium, indium, rhodium, ruthenium); alloys of noble metals (e.g. gold-palladium, silver-platinum, etc.); conducting polymers (e.g. polypyrrole (PPY)); non-noble metals (e.g. copper, nickel, aluminum, tin, titanium, indium, tungsten, platinum); metal oxides (e.g. zinc oxide, tin oxide, nickel oxide, indium tin oxide, titanium oxide, nitrogen-doped titanium
20 oxide (TiOxNy); metal silicides (nickel silicide, platinum silicide); metal nitrides (titanium nitride (TiN), tungsten nitride (WN) or tantalum nitride (TaN)), carbon (nanotubes, fibers, graphene and amorphous) or combinations of any of the above. NMEs of the above-described materials are highly conductive and form strong bonds with probes (e.g. nucleic acids and peptides). Preferred NMEs have a height in the range of about 0.5 to about 100
25 microns (μm), for example in the range of about 5 to about 20 microns (e.g. 10 microns) ; a diameter in the range of about 1 to about 10 microns; and have nanoscale morphology (e.g. are nanostructured on a length scale of about 1 to about 300 nanometers and more preferably in the range of about 10 to about 20 nanometers). NMEs can be any of a variety of shapes, including hemispherical, irregular (e.g. spiky), cyclical (wire-like) or fractal (e.g.
30 dendritic). The surface of an NME may be further coated with a material, which maintains the electrode's high conductivity, but facilitates binding with a probe. For example,

nitrogen containing NMEs (e.g. TiN, WN or TaN) can bind with an amine functional group of the probe. Similarly, silicon/silica chemistry as part of the NME can bind with a silane or siloxane group on the probe.

In another aspect, the invention features an NME associated with a probe. In one embodiment, the probe is a nucleic acid (e.g. a ribonucleic acid (RNA), deoxyribonucleic acid (DNA) or analog thereof, including, for example, a peptide nucleic acid (PNA), which contains a backbone comprised of N-(2-aminoethyl)-glycine units linked by peptides rather than deoxyribose or ribose, peptide nucleic acids, locked nucleic acids, or phosphorodiamidate morpholino oligomers. Under appropriate conditions, the probe can hybridize to a complementary nucleic acid to provide an indication of the presence of the nucleic acid in the sample. In another embodiment, the probe is a peptide or protein (e.g. antibody) that is able to bind to or otherwise interact with a biomarker target (e.g. receptor or ligand) to provide an indication of the presence of the ligand or receptor in the sample. The probe may include a functional group (e.g., thiol, dithiol, amine, carboxylic acid) that facilitates binding with an NME. Probes may also contain other features, such as longitudinal spacers, double-stranded and/or single-stranded regions, polyT linkers, double stranded duplexes as rigid linkers and PEG spacers.

In a further aspect, the invention features a plurality of NMEs arrayed on a substrate. Preferred substrates are comprised of a semiconductor material, such as silicon, silica, quartz, germanium, gallium arsenide, silicon carbide and indium compounds (e.g. indium arsenide, indium, antimonide and indium phosphide), selenium sulfide, ceramic, glass, plastic, polycarbonate or other polymer or combinations of any of the above. Substrates may optionally include a passivation layer, which is comprised of a material, which offers high resistance and maintains a small active surface area. Examples of appropriate materials include: silicon dioxide, silicon nitride, nitrogen doped silicon oxide (SiOxNy) or paralyene. In certain embodiments, the plurality of NMEs arrayed on the substrate include probes in conjunction with monolayer spacers, which minimize probe density, thereby maximizing complexation efficiency. Preferred monolayer spacers have an affinity to metal and can be comprised, for example, of a thiol alcohol, such as mercaptohexanol, alkanethiols, cysteine, cystamine, thiol-amines, aromatic thiols (e.g. benzene thiol, dithiol), phosphonic acids or phosphinic acids.

Another aspect features biosensing devices, such as integrated circuits, comprising, for example, a substrate; an electrically conductive lead on the substrate; an insulating or passivation layer covering the lead, the insulating layer having an aperture exposing a portion of the lead; and a nanostructured microelectrode in electrical communication with
5 the exposed portion of the lead, the microelectrode being adapted to generate a charge in response to a biomolecular stimulus (e.g. nucleic acid hybridization or protein-to-protein binding).

In still another aspect, the invention features methods for manufacturing NMEs. The use of electrodeposition to grow nanostructured microelectrodes from an NME seed
10 allows the sizes and morphologies of these structures to be finely controlled, and versatile fabrication of electrodes composed of one or a variety of substances. NMEs may be prepared on a biosensing device, such as a chip-based format, such that a series of NMEs may be made on a single chip to enable multiplexed experiments. This NME system may be particularly useful and versatile, allowing adjusting of several parameters, including: the
15 microscale control of the NME size and shape, the nanoscale control of NME nanotexturing, and selection of the NME material.

Yet another aspect features methods for manufacturing biosensing devices having nanostructured microelectrodes. For example, the methods can comprise the steps of providing a substrate and an electrically conductive lead on the substrate, the lead being
20 covered by an insulating layer; etching an aperture in the insulating layer to expose a portion of the lead; and electrodepositing an electrically conductive material on the exposed portion of the lead to form a nanostructured microelectrode as described above.

In another aspect, there is provided a biosensing cartridge comprising: a sample chamber for containing a biological sample; a biosensing chamber for containing a
25 biosensing device as described above and carrying out a biosensing process.

In yet another aspect, there is provided a biosensing workstation comprising: a cartridge holder for holding a biosensing cartridge as described above; an instrument tip for accessing the biosensing cartridge; a selection mechanism for selecting a biosensing process to be carried out; a processor adapted to carry out the biosensing process using the
30 biosensing cartridge and to determine results of the biosensing process from electronic

signals generated from the biosensing cartridge; and a display for displaying the results of the biosensing process.

A further aspect features methods for carrying out a biosensing process using probe containing nanostructured microelectrodes incorporated into a device as described above; 5 biasing the microelectrode relative to a reference electrode; measuring a reference charge or reference current flow between the microelectrode and the reference electrode; exposing the microelectrode to a biomolecular stimulus (e.g hybridization between a nucleic acid probe with a complementary nucleic acid or binding between a peptide probe and a binding partner present in a biological sample); measuring a charge or current flow generated at the 10 microelectrode in response to the biomolecular stimulus; and determining the amount of biomolecular stimulus present by comparing the measured charge or measured current flow against the reference charge or reference current flow.

NMEs are versatile, robust and easy to work with. In addition, they can be manufactured using existing silicon CMOS foundry fabrication procedures for top-metal 15 fabrication, or simple extrapolations thereof, such as electroless deposition or electrodeposition onto top-metal layers from a CMOS foundry, allowing the manufacture of NMEs to be easily integrated into existing manufacturing facilities. In addition, NMEs are able to consistently attach to probe molecules. Further, NMEs promote ready accessibility of target molecules such that, when a target molecule that is complementary to the NME- 20 attached probe molecule enters into proximity with that probe, hybridization or protein-to-protein binding occurs with high probability. NMEs are further compatible with the performance of electrocatalytic electrochemistry employed in the read-out of the hybridization event.

Other features and advantages of the inventions disclosed herein will become 25 apparent from the following detailed description and claims.

BRIEF DESCRIPTION OF DRAWINGS

FIG. 1(A) is a schematic of an NME biosensing device; FIG. 1(B) is a schematic cross-sectional view of an aperture for forming an NME; and FIG. 1(C) is a schematic view of the formation of an NME in an aperture;

30 FIG. 2 is a schematic of an NME biosensing device;

FIG. 3 is a schematic of an NME with probes, further showing the presence of spacers in the probe monolayer and between the electrode and the probe;

FIG. 4 are SEM images of NMEs with increasing degrees of nanostructuring;

FIG. 5 illustrates steps involved in the sensing of specific sequences using an NME;

5 FIG. 6 is a cross-sectional view of an integrated circuit having an NME;

FIG. 7 is a circuit diagram of a circuit that may be used with an NME;

FIG. 8 is an illustration of a biosensing device having an array of NMEs;

FIG. 9 is an illustration of a biosensing device having different NMEs;

FIG. 10 is a schematic illustration of a biosensing cartridge; and

10 FIG. 11 is a schematic illustration of a biosensing workstation.

DETAILED DESCRIPTION OF THE INVENTION

The present disclosure describes nanostructured microelectrodes (NMEs), which may be used in a biosensing device, such as a biosensing chip.

NMEs

15 FIG. 1A shows a schematic diagram of an exemplary device incorporating NMEs. In the example shown, the device is a chip having an array of eight leads. In this example, the NME is formed on gold leads that taper to a width of 5 microns. The lead is provided on a substrate of Si and SiO₂, although other suitable substrate materials may be used. On top of the lead, an insulating layer, such as SiO₂, is deposited to electrically insulate and
20 passivate the lead. An aperture, in this case a 500 nm hole, is created in the insulating layer to expose a portion of the lead.

FIG. 1B is a side view of the portion of the chip where the aperture is located, showing the layers of the device. Most commonly known photolithography techniques may be suitable for creating an aperture, such as 100 nm to 1 μm diameter apertures. This is
25 typically achievable in existing fabrication facilities with high robustness and reproducibility. Given that only this exposed surface is electrochemically active,

electrodeposition (Menke et al. *Nature Mater.* 2006, 5, 914-919) may be used to grow an NME within this space.

FIG. 1C is a cross-sectional schematic of an example of NME deposition, using Pd for the NME. This process will be described in greater detail below.

5 Reference is now made to FIG. 2, which also illustrates the formation of an NME on a chip. As in FIG. 1A, small electrodes are provided *in situ* on a chip, with the position and electrical contacting of the NMEs defined photolithographically. As in Figure 1, this chip is an 8-fold multiplexed passive chip. On a silicon substrate, a ~350 nm thick gold layer is patterned, using conventional photolithography techniques, to connect eight 5- μ m-wide Au
10 leads to large metal pads for connection to off-chip instrumentation. A pinhole-free insulating SiO₂ layer is deposited and patterned to create ~500 nm openings (e.g., by etching) at the end of each of the Au leads, to expose a section of the lead. A metal NME is then plated in this opening using electrodeposition.

 The NME may comprise different conductive materials. Some examples of NMEs
15 have been formed as follows, although variations are possible and will be described in greater detail below: Branched fractal Pd NMEs were deposited in an aqueous solution containing 5 mM of H₂PdCl₄ and 0.5 M of HClO₄ using DC potential amperametry at -250 mV for 15 s. Hemispherical Pd NMEs with nanoscale roughness were deposited in an aqueous solution containing 5 mM of H₂PdCl₄ and 0.5 M of HCl using DC potential
20 amperametry at -100 mV for 300 s. Smooth hemispherical Pd NMEs were fabricated in the same solution at 0 mV for 300 s. Au NMEs were fabricated in a gold bath containing 0.01 M solution of HAuCl₄ and 0.5 M H₂SO₄ at -100mV for 40 seconds. Pt NMEs were fabricated in a platinum bath containing 5 mM solution of H₂PtCl₆ and 0.5 M H₂SO₄ at -100 mV for 500 s. NME size and morphology can be controlled by varying the metal salt
25 concentration, type and concentration of the supporting electrolyte, and electrodeposition potential and duration.

 FIG. 3 is a schematic of an NME with probes with spacers in the probe monolayer and between the electrode and probe. A chemical solution containing a metal cation can be brought into communication with the surface of the NME and a reference electrode. The
30 reference electrode may be an NME or a conventional electrode on same lead. An electrical bias can be provided between the NME and the reference electrode. The chemical solution

can then be removed and the electrodes washed. A solution containing the probe molecule can then be brought into communication with the surface of the NME. The probe molecule may be modified or functionalized so that it binds to the surface of the NME. For example, the probe molecule may be functionalized with a thiol, amine, or carboxylate group.

5 NMEs with increasing nanostructuring are shown in FIG. 4. Unexpectedly, variation of electrodeposition conditions allowed growth of extremely smooth hemispherical microelectrodes (left); highly branched nanoscale fractal structures (right); or hemispheres with nanoscale roughness (center). The structure on the left was made with HCl as a supporting electrolyte with an applied potential of 0 mV. The center structures
10 were also made with HCl as a supporting electrolyte but with an applied potential of -100 mV. The structure on the right was made with HClO₄ as a supporting electrolyte and an applied potential of -250 mV. The scale bar on the figure corresponds to 5 μ m unless otherwise indicated.

FIG. 5 illustrates steps involved in the sensing of specific sequences (Lapierre et al.,
15 *Anal. Chem.* 75:6327, Ratilainen et al., *Biochemistry* 39:7781, Tomlins et al., *Science* 310:644) In this example, Pd NMEs are first modified with thiol-derivatized probe sequences, and then target sequences are hybridized. The presence of the target is then transduced using an electrocatalytic reporter system. Electrocatalysis provides electronic amplification, or gain, facilitating high-sensitivity readout: hundreds of electrons may result
20 from each biomolecular complexation event. The approach used herein relies on the primary electron acceptor Ru(NH₃)₆³⁺, which is electrostatically attracted to the electrode surfaces at levels that are correlated with the amount of bound nucleic acid. The inclusion of Fe(CN)₆³⁻ during electrochemical readout serves to regenerate the Ru(III) substrate, as the Fe(III) species is even easier to reduce, but it is electrostatically repelled from the
25 electrode and thus only undergoes chemical reduction by Ru(II). This method is also label-free and does not require the sample to be processed in any way.

The biosensing device may be provided in the form of a chip, such as an integrated circuit (IC) chip. In general, an IC incorporating the NME may have a substrate with an electrically conductive lead that is covered by an insulating layer. The insulating layer has
30 an aperture that exposes a portion of the lead, and the NME is provided at the exposed portion of the lead. The NME is responsive to a biomolecular stimulus. In particular, the NME may be functionalized with probe molecules that undergo a hybridization reaction

with a target biomolecule (e.g., a nucleic acid sequence), resulting in a charge generated at the NME. The IC also has a charge storage (e.g., a capacitor or a battery) in electrical communication with the lead to store this generated charge. In typical usage, the NME may be exposed to a sample for a known time duration or an integration period, and the charge
5 stored over that time would then be indicative of the presence and/or amount of the target biomolecule.

The stored charge may be communicated to a computing device for analysis, or may be displayed (e.g., through a digital display component) for direct reading of the charge stored after the integration period.

10 Such an IC may be manufactured using common IC manufacturing equipment, allowing this device to be easily manufactured and to be less costly than other forms of biosensing microelectrodes. The materials used may be those already commonly used in IC manufacturing. For example, the substrate may be made from silicon, quartz, glass, ceramics, silica, sapphire, gallium arsenide, or other materials currently used for ICs. The
15 substrates or supports can incorporate conductive material to serve as an electrode. Conductive supports with a gold surface may also be used. The supports usually comprise a flat (planar) surface, or at least a structure in which the probes or p to be interrogated are in approximately the same plane. The support can be an electrode, or can be attached to an electrode.

20 The lead may be made of Au, Al, W, TiN, polysilicon or other commonly used lead materials. The IC may include a transistor, such as a field-effect transistor (FET) including n-type silicon channel FETs and p-type silicon channel FETs, or a bipolar transistor including n-p-n bipolar junction transistors and p-n-p bipolar junction transistors.

The IC may be provided with, immersed in or otherwise exposed to an
25 electrocatalytic solution in chemical and electrical communication with the NME. This may assist in charge generation in the NME.

Reference is now made to FIG. 6. This figure shows a cross-section of an integrated circuit suited to sensing the presence of biomolecules in a biological sample. The substrate (1) is a conventional semiconductor device substrate such as silicon. The channel of a
30 transistor (2), a gate oxide (3), and a polysilicon gate electrode (4) are shown to illustrate the use of conventional CMOS electronics to form the integrated circuit's transistors. A

metal (5) is used to contact the gate electrode. A passivation oxide (6) separates the silicon transistor levels below from the top surface of the chip above. A series of metal vias (7) and interconnects provide selective paths for electrical communication between the transistor layer and the top electrode(s). A substantially planar top surface is a heterogeneous combination of top electrodes (9) and top insulating material (8). The figure illustrates an NME (10) provided on the electrode, for example using the methods described above. The figure illustrates probe biomolecules (11) such as thiol-terminated nucleic acids that are displayed for efficient hybridization with complementary target molecules. An electrocatalytic solution (12) may be employed to provide catalytic read-out of hybridization with the biomolecules (11). Electrical potentials are conveyed, and currents flowed, in a continuous fashion from the NME (10) through the electrical contacts (9) (7) (5) (4) down to the electronic circuitry that resides beneath.

Reference is now made to FIG. 7, showing a circuit diagram of an example circuit that may be used with the disclosed NME. In this example, the circuit may provide the following functions: biasing of the probe-functionalized NME; integration of the current flowing through the NME into a charge store having a known charge-storage capacity; read-out of the voltage on the charge-store; and selection of the charge-store or NME of interest when a two-dimensional array of stores and electrodes is provided in the context of a highly multiplexed array chip.

The components of the example circuit are now described. A bias voltage is provided at V_{bias} ; a typical choice of bias may be in the range of about 0.1 - 2.8 V. A bias voltage V_{biasD} is provided at the source-follower drain; a typical choice may be in the range of about 0.1 - 2.8 V. A bias voltage V_{biasR} is provided at the reset node; a typical choice may be an adjustable value between about -2 V and 2.8 V. The signal voltages are V_{src} which may typically be in the range of about 1.5 - 2.5 V. The column voltage V_{col} may be in the range of about 1.5 V - 0.5 V. Timing control signals include that for row select (e.g., range may be about 0-2.8 V) and for reset (e.g., range may be about 0-4 V). The transistors may be the reset transistor $\text{Tr}(\text{Reset})$, the read-out buffer transistor $\text{TR}(\text{source-follower})$, and the row-select transistor $\text{TR}(\text{row-select})$.

The above biases, signal voltages and timing control signals are examples only and other values may be used. These biases, voltages and signals may be selected or adjusted to suit certain applications or manufacturing conditions, as is commonly known in the art. In

this example, the probe-functionalized NME may include a NME functionalized using a thiolated nucleic acid probe, for example a probe as described above. V_{src} is applied to the probe-functionalized electrode and V_{biasR} is applied to a second electrode, which may be a NME or any other common electrode, in electrical communication with the electrocatalytic solution. This results in voltage difference between the probe-functionalized electrode and the electrocatalytic solution. A current may thus flow as a consequence of this potential difference. The amount of current flowing may be typically dependent on the amount of hybridization on the probe-functionalized NME, that is the current may be indicative of the amount of target thus detected by the NME.

The operation of the example circuit is now described. In order to capture the current flowing, I_{sense} , the reset transistor is turned on by setting the node 'reset' high enough (e.g., up to 4V, which may be through an on-chip charge-pump or regulator circuit as commonly known in the art) so that node 11 will be charged to a voltage equal to V_{bias} (node 6) which may be typically set to the supply rail: e.g., 2.8V. This is the reset phase. Once this 'reset' operation is completed, Node 5 may be set to 0V to turn off the reset transistor $Tr(Reset)$ (1). In doing so, charge injection and parasitic capacitive feedthrough effects will cause node 11, now becoming a floating node, to drop by approximately 300 mV. Therefore after the 'reset' operation the actual 'reset' voltage value at node 11 is approximately 2.5V. At this time, the current I_{sense} flowing is dependent on the voltage applied (i.e., $V_{src} - V_{biasR}$). With V_{biasR} being able to be set arbitrarily to any voltage level from -2V to 2.8V, the applied potential difference may be adjusted. The current I_{sense} discharges the parasitic capacitance at the V_{src} node (11) and its voltage level drops at a rate dependent on the value of the parasitic capacitance at V_{src} (node 11) as well as the flowing I_{sense} during the integration time. After a specific integration time, the resulting integrated voltage at node 11 will be read out through transistors $TR(source-follower)$ and $TR(row-select)$, the source follower buffer transistor and the row-select transistor, by setting the node SEL (8) to a high level (2.8V).

The charge store that is discharged at node 11 may comprise parasitic capacitance of one or more of the transistors that are in electrical communication with the electrode of the pixel region at V_{src} (node 11). The electrode at V_{src} (node 11) may be in electrical communication with the gate of a transistor, such as $Tr(source-follower)$ 2, which provides a parasitic capacitance. In one example embodiment, the charge store may be provided at

least in part by a parasitic capacitance between the gate and drain of the source follower transistor, $Tr(\text{source-follower})$ 2, and a parasitic capacitance between the source and substrate of the reset transistor, $Tr(\text{Reset})$ 1. These are parasitic capacitances between the structures on the semiconductor substrate (e.g., the poly, n-well and substrate) on which or
5 in which the pixel circuit is formed. In an example embodiment, these parasitic capacitances may be in the range of about 1-2 femtoFarads or more generally in the range of about 0.5 to 3 femtoFarads or any range subsumed therein. The contacts to the probe-functionalized NME may be formed in different layers above the regions of the semiconductor substrate used to form the transistors. In an alternate embodiment, the
10 polarity of the bias may be reversed and the parasitic capacitance at V_{src} may be charged instead of discharged during the integration period.

Reference is now made to FIG. 8, which shows a top view of an example IC that has a multiplexed array of individually-addressable probe-functionalized NMEs. By individually addressable, it is meant that each NME may be individually electrically
15 accessed, such that the current or charge generated by each NME may be individually measured. In this example, the NMEs are arrayed in a row-column fashion. There are n rows and m columns for a total of $m \times n$ independent NMEs. If there were only a single row or column of NME, then it may not be necessary to have row/column address circuitry. However, when large total numbers of NMEs are desired on a single device, it may be more
20 efficient to array them in a two-dimensional grid or similar, and thus independent electrical access to each NME may be useful. This may be efficiently achieved using the circuitry illustrated in FIG. 7. In this approach, the charge associated with the current flowing through each NME is integrated into a charge storage, such as a capacitor; and a voltage proportional to the stored charge may be read out for the NME in a particular row by setting
25 the node SEL (8) to a high level and monitoring the voltage on that column V_{col} .

The figure illustrates that, for each column, there may exist a time-dependent signal (whose time-dependence may be determined by the clocking of the row-address circuitry) which, in some embodiments, may be fed, in cases with the aid of electronic buffering or amplification, into an analog-to-digital converter. The analog-to-digital converter may
30 accept signals having a pre-determined voltage swing (such as 0-1 V, typically) and, for each input channel, may carry out a quantization operation in which a digital representation of the analog level in that signal is estimated. The output of the A/D converter is a digital

stream which combines parallelism (e.g., multiple parallel wires, each corresponding to a significant figure in the binary representation of the values) and serial timing (e.g., a timed representation of sequential data elements corresponding, for example, to different probe-functionalized NMEs).

5 Reference is now made to FIG. 9, showing three adjacent NMEs along a single row, in a configuration that may be provided on a biosensing device or IC as described above. These three NMEs are read using three different columns j , $j+1$, and $j+2$. This figure illustrates a number of features with respect to the differences among NMEs.

10 NMEs $E_{i,j}$ and $E_{i,j+1}$ may be functionalized both with the same class of probes (e.g. thiolfunctionalized PNA), but the sequences may be different. That is, each NME may be functionalized with similar probes that have different target biomolecules. In this example, electrodes $E_{i,j}$ and $E_{i,j+1}$ are response to different sequences present in the sample under study. In sum, the use of different functionalizations enables sensing of biomolecules within a single class, but having a different sequence, conformation, or functionality.

15 NMEs $E_{i,j}$, $E_{i,j+1}$, and $E_{i,j+2}$ are shown having different morphologies and/or sizes, and different degrees of nanostructuring. As discussed above, different morphologies and/or sizes may provide both different limits of detection, and different dynamic ranges, in detecting of target molecules. By incorporating NMEs having different morphologies and/or degrees of nanostructuring onto one device, it may be possible to expand the
20 dynamic range of target concentrations that may be sensed using a single device. In sum, the use of different NME morphologies, sizes and/or nanostructurings may enable sensing of a wider range of concentrations of a given target species than would otherwise be achieved if only one morphology/nanostructuring were provided on a biosensing device.

25 NMEs $E_{i,j+1}$ and $E_{i,j+2}$ are also depicted as being functionalized using different classes of probe molecules. For example, $E_{i,j+1}$ may be functionalized using a nucleic acid such as PNA, and $E_{i,j+2}$ may be functionalized using antibodies which attach to the electrode. In sum, the use of different classes of probe molecules may enable sensing of different classes of target biomolecules, for example ranging from DNA to RNA to micro-RNA to proteins, using a single biosensing device.

30

Biosensing cartridge and workstation

The biosensing device as described above may be incorporated into a biosensing cartridge. Such a cartridge may contain chambers for sample processing such as disruption and resin or bead-based nucleic acid purification, as well as a chamber for the biosensing device. The cartridge may be self-contained, for example all necessary reagents may be contained in the lid of the cartridge. The cartridge may be reusable, or may be disposable. A disposable cartridge may minimize the risk of cross-contamination between samples.

The cartridge may be used in a biosensing workstation for coordinating and carrying out the biosensing process. Components of the workstation may include sample holders, instrument tips such as pipettors for manipulation of the sample, a sample identification module, a selection mechanism for selecting a test to be carried out, an electronic display for indicating the results of a biosensing test and a processor for managing these components and carrying out the selected tests. The workstation may hold a number of different cartridges at one time (e.g., ten or more). The workstation may allow random access to the cartridges - that is, independent tests may be run at any time on any cartridge in the workstation. The workstation may have disposable instrument tips, which would be the only part of the workstation that comes into direct contact with the sample and reagents. Disposable tips, together with disposable cartridges, may minimize the risk of cross-contamination between samples tested in the workstation.

In general, a biosensing cartridge may have a first chamber for containing the sample to be tested, and a second chamber containing the biosensing device as described above. There may be addition chambers to perform other actions on the sample, such as purification, and/or subdivision (e.g., through chemical, mechanical or vibrational means). Some processing and disruption of the sample may be carried out in the first chamber itself. The sample may be introduced from the first to the second chamber for detection by the biosensing device in the second chamber. There may be pre-set time interval from activation of the device or start of the test to the introduction of the sample into the second chamber. This time interval may allow the biosensing device to be suitably biased or otherwise readied for the test. The charge or current flow generated in the biosensing device may be measured after an integration period, as described above.

Reference is now made to FIG. 10, showing an example of a biosensing cartridge and the steps of using such a cartridge. In this example, the cartridge has three chambers, a sample chamber containing the sample, a purifying chamber for purifying the sample, and a biosensing for performing the biosensing. As shown in this example, the lid of the cartridge is provided with capsules containing reagents for each chamber, and through which an instrument tip may be inserted. This allows the cartridge to be self-contained, already containing the reagents suitable to carry out the biosensing operation and tailored to the particular probes and/or target biomolecules of the biosensing device being used. In this example, the sample chamber has one capsule, for accessing the sample. The purifying chamber has three capsules - two containing a washing reagent, and one containing an elution buffer. The biosensing chamber has three capsules - each containing an electrochemical mix containing analytes for the biosensing device. The instrument tip may be inserted sequentially in each capsule, in order to carry out the biosensing operation. For example, the capsule on the sample chamber may contain a lysis buffer containing chemical denaturants (e.g., urea or formamide); the capsules on the purifying chamber may have two capsules containing a wash buffer and the elution buffer may be a standard buffer, such as one containing low levels of sodium, chloride and tris salt; the capsules on the biosensing chamber may contain redox reporter groups such as ruthenium hexamine, ferricyanide and a buffer containing sodium, phosphate, chloride and magnesium.

In use, the sample is first extracted from the sample chamber through its single capsule. The sample is then introduced into the purifying chamber, where it is washed twice as the instrument tip is introduced through the two washing capsules and the elution buffer is introduced. Through this process, the sample may be prepared for biosensing by the biosensing device. For example, in the case of a nucleic acid sample, the process in the purifying chamber may isolate the RNA or DNA in the sample. The purified sample is then introduced into the biosensing chamber, through each electrochemical capsule. The biosensing device in the biosensing chamber may then detect any target biomolecules present in the sample, and the generated current or charge may be measured. Where the cartridge is used in a workstation, the processor in the workstation may read this generated current or charge and determine the presence of the target biomolecule based on this reading.

Reference is now made to FIG. 11, showing an example of a biosensing workstation. This work station includes a bar code reader, allowing identification of samples using unique bar codes provided on each cartridge. The workstation has a selection mechanism, in this example a touch screen that allows the selection of a particular test to be carried out. There is also a waste container for disposing any wastes generated by the biosensing process. The processor of the workstation may be connected to an external computing device, such as another workstation, for further analysis. This connection may be through a wireless network. The workstation may be relatively small (e.g., a footprint of 1.5 x 1 ft), allowing convenience and ease of use.

10 Methods of Use

Methods for using NMEs and devices comprising the same is now described. A device may be provided with the NME already functionalized with a probe molecule, or the probe molecule may be bound to the NME when preparing the device for use. The device is then biased for use, for example by adding an electrocatalytic reporter and waiting a certain time interval. In addition to the NME, there may be a reference electrode, which may or may not be an NME, in contact with the electrocatalytic reporter but not in contact with the sample. The current flow or voltage bias generated over this time interval between the NME and the reference electrode may be measured and recorded as the reference point. The NME is then exposed to a sample of interest, and the current flow or charge generated over a certain time interval (also referred to as the integration period) may be measured. By comparing the difference in current flow or charge between the exposure time interval and the biasing time interval, the concentration, binding and/or amount of target biomolecule in the sample may be determined.

Devices comprising NMEs, as described herein, may be used in conjunction with appropriate probes to detect the presence or absence of particular biomarkers in a sample. A "sample" or "biological sample" as herein refers to any natural (e.g. plant, animal, algal, bacterial or viral) or synthetic material containing DNA, RNA and/or proteins, including, for example, clinical samples, such as tissues, cell cultures or fluids isolated from an individual (including without limitation blood, plasma, serum, cerebrospinal fluid, lymph, tears, urine, saliva, mucus, synovial fluid, cerebrospinal fluid and tissue sections) environment (e.g., water, food or air samples). Biological samples may be further processed via a variety of means, including lysis (electrical, mechanical and chemical), electrophoresis, enzymatic

digestion. Most often, the sample has been removed from an organism, but the term "biological sample" can also refer to cells or tissue analyzed in vivo, i.e., without removal. Typically, a "biological sample" will contain cells, but the term can also refer to non-cellular biological material, such as non-cellular fractions of blood, saliva, or urine. "A biological sample" further refers to a medium, such as a nutrient broth or gel in which an organism has been propagated, which contains cellular components, such as proteins or nucleic acid molecules.

Probes for use with the instant described NMEs may be comprised of nucleic acids. A "nucleic acid probe" refers to a nucleic acid (e.g. a ribonucleic acid (RNA), deoxyribonucleic acid (DNA) or an analog thereof, including, for example, a peptide nucleic acid (PNA), which contains a backbone comprised of N-(2-aminoethyl)-glycine units linked by peptides rather than deoxyribose or ribose linked by phosphodiesterase linkages) capable of binding to a target nucleic acid of complementary sequence through one or more types of chemical bonds, usually through complementary base pairing, usually through hydrogen bond formation. As used herein, a nucleic acid probe may include natural (i.e., A, G, C, or T) or modified on bases (7-deazaguanosine, inosine, etc.) or on sugar moiety. In addition, the bases in a probe can be joined by a linkage other than a phosphodiester bond, so long as it does not interfere with hybridization. It will be understood by one of skill in the art that probes may bind target sequences lacking complete complementarity with the probe sequence depending upon the stringency of the hybridization conditions. By assaying for the presence or absence of the probe, one can detect the presence or absence of the select sequence or subsequence. Methods for detecting target nucleic acids using nucleic acid probes are described, for example, in U.S. Patent No. 7,361,470 entitled "Electrocatalytic Nucleic Acid Hybridization Detection." and US 2005/0084881 of the same name.

"Hybridization" refers to any process by which a strand of nucleic acid binds with a complementary strand through base pairing. "Hybridization conditions" refer to standard conditions under which nucleic acid molecules are used to identify similar nucleic acid molecules. Such standard conditions are disclosed, for example, in Sambrook et al., *Molecular Cloning: A Laboratory Manual*, Cold Spring Harbor Labs Press, 1989. Sambrook et al., *ibid.*, is incorporated by reference herein in its entirety (see specifically, pages 9.31-9.62). In addition, formulae to calculate the appropriate hybridization and wash conditions to

achieve hybridization permitting varying degrees of mismatch of nucleotides are disclosed, for example, in Meinkoth et al., 1984, *Anal. Biochem.* 138, 267-284; Meinkoth et al., *ibid.*, is incorporated by reference herein in its entirety. Non-limiting examples of hybridization conditions include low stringency hybridization conditions, moderate stringency
5 hybridization conditions and high stringency hybridization conditions.

In another embodiment, the probe is a peptide (comprised of, for example, 4-40 amino acids) or proteins (e.g. antibody) that is able to bind to or otherwise interact with a biomarker target (e.g. receptor or ligand) to provide an indication of the presence of the ligand or receptor in the sample. Methods for detecting analytes using peptide or protein
10 probes are described, for example in International patent application WO 2007/094805 (PCT/US2006/013771) entitled "Method for Electrocatalytic Protein Detection."

Probes may include a functional group (e.g., thiol, dithiol, amine, carboxylic acid) that facilitates binding with an NME. Probes may also contain other features, such as longitudinal spacers, double-stranded and/or single-stranded regions, polyT linkers, double
15 stranded duplexes as rigid linkers and PEG spacers.

As described above, the surface nanostructure of the NME may be controlled, and may influence the sensitivity and/or efficiency of a device having the NME. In Example 1, the influence of surface nanostructure on the detection efficiency for nucleic acids was investigated. Two different types of NMEs were compared - a more finely nanostructured
20 NME obtained with a low deposition potential was compared to a more coarsely textured one obtained with a higher deposition potential.

While concentrations as low as 1 pM could be detected with the more finely nanostructured NME obtained with a low deposition potential, the detection limit was increased to 10 pM for the more coarsely textured one obtained with a higher deposition
25 potential.. These results demonstrate that increased nanostructuring contribute to more sensitive biosensing capabilities in an electrode platform. This analysis revealed that the more finely structured NMEs showed greater responsiveness to sub-nM concentrations of target sequences.

The 10 aM sensitivity observed here with the disclosed NMEs and electrocatalytic
30 reporter system provides a low detection limit for a label- and PCR-free sensor; the detection limit corresponds to the detection of < 100 copies of the target sequence. While

the measurement of 60-1000 copies of target sequence has been achieved previously with electrochemical detectors exploiting multi-step catalytic readout (Munge et al., *Anal. Chem.* 77 :4662, Nicewarner-Pena et al., *Science* 294:137, Park et al., *Science* 295 :1503, Sinensky et al., *Nat. Nano.* 2 :653, Steemers et al., *Nat. Biotechnol.* 18:91, Xiao et al., *J. Am. Chem. Soc.* 129:11896, Zhang et al., *Nat. Nano.* 1:214, Zhang et al., *Anal. Chem.* 76:4093), the disclosed device provides this measurement on a chip-based platform with single-step readout.

Example 2 describes use of a multiplexed electrode platform, as described herein, to directly read a panel of cancer biomarkers in clinically-relevant samples using electronic signals. The system combines nanotextured electrodes with rapid catalytic readout to achieve a long-standing goal: the multiplexed analysis of cancer biomarkers using an inexpensive and practical platform.

Example 3 describes the use of an NME based chip to detect microRNA, one of the most challenging detection targets. Electronic readout of microRNA profiles offers a rapid, yet highly accurate, method to directly assay RNA samples for specific sequences without the need for target amplification.

Although the provided examples are directed to the detection of cancer biomarkers, other applications may be possible for the NME device, which may involve detecting DNA, RNA and/or proteins. Examples include profiling of breast cancer genes (e.g., by detecting RNA markers); profiling of leukemia-related genes (e.g., by detecting RNA markers); profiling of cytochrome P450 mutations that affect drug metabolism (e.g. Warfarin) (e.g., by detecting DNA and RNA markers); profiling of mutations associated with genetic disease (e.g. Cystic fibrosis) (e.g., by detecting DNA markers); detection and typing of viruses (e.g. HPV and HIV) (e.g., by detecting DNA and RNA markers); detection of cancer-related proteins using an electrochemical immunoassay format (e.g. prostate specific antigen (PSA)) (e.g., by detecting protein markers); and detection of micro RNAs to identify cancer. Biosensing devices incorporating these NMEs may be adapted to detect for these other biomolecules by binding suitable probes to the NME and/or by selecting a suitable electrocatalytic reaction to be sensed, as commonly known in the art.

It would be understood by a person skilled in the art that variations are possible without departing from the present disclosure. All examples and embodiments described are

provided for the purpose of illustration only, and are not intended to be limiting. All references mentioned herein are hereby incorporated by reference in their entirety.

Example 1. Parameters for Manufacturing NMEs

In this example, Pd was used as an electrode material. To investigate the time-
5 dependence of the electrodeposition, the structures of the Pd NMEs being electrodeposited were monitored as a function of time. Time-dependent electrodeposition experiments were performed at -100 mV using 0.5 M HCl as supporting electrolyte. Pd structures were formed for (a) 25, (b) 50, (c) 125, (d) 250, and (e) 500 s. After 50 s, structures with average
10 diameters of 1.3 μm and heights of 0.5 μm were observed, and after 500 s the Pd electrodes were typically 8 μm and 5 μm in diameter and height. The smaller structures made with shorter deposition times typically exhibited depressions in the center of the microelectrodes, which may indicate that nucleation occurs preferentially at the boundary of the aperture.

Another controllable parameter that may influence the final structure of the NME is the deposition potential. Specifically, the size and surface morphology of the NMEs may be
15 controlled in this way. Dendritic fractals are phenomena generally observed in non-equilibrium growth such as the growth of snowflakes, the aggregation of soot particles, and the solidification of metals. Such fractal structures are also obtained by nonequilibrium electrodeposition of metals and used as model systems for the study of branching and fractal growth processes (Fleury, *Nature* 1997, 390, 145-148). It is generally believed that
20 the morphology of crystals heavily depends on the “distance” of their formation conditions from the thermodynamic equilibrium: near-equilibrium conditions lead to polyhedral crystals surrounded with thermodynamically stable crystal faces, but increase of this “distance” makes the growing fronts of crystals with flat surfaces instable to form dendrites (Fukami et al. *J. Phys. Chem. C* 2007, 111, 1150-1160). In the case of electrodeposition of
25 metals, such “distance” may be tuned continuously and reversibly by simply changing the deposition potential and more negative potential can exert higher driving force and thus increase the “distance” from the equilibrium for electrocrystallization. Thus, electrodeposition may be controlled spatially and kinetically to produce arrayed NMEs with varied well-defined morphologies.

30 Pd structures were formed at (a) 0 mV, (b) -100 mV, (c) -250 mV, and (d) -400 mV for 250s with the use of 0.5 M HCl as supporting electrolyte. More negative deposition

potentials were found to typically lead to larger, but less compact, microstructures. At a deposition potential of 0 mV, a cake-shaped structure of 3.5 μm diameter and 0.7 μm height was obtained. When the applied potential was changed to -100 mV, a rougher microstructure was obtained that was also larger in size (average diameter = 5 μm and height = 2.5 μm). The nanotexturing obtained is an irregular aggregate of very small nanoparticles. When a more negative potential of -250 mV was applied, a dendritic fractal microstructure was obtained and its diameter and height were found to further increase 11 and 6 μm , respectively. If the potential was increased further (e.g. to -400 mV), the microelectrode structure became more open and the structure is no longer continuous.

10 The electrochemical behaviour of the Pd NMEs formed as a function of potential was studied by cyclic voltammetry (CV). The electrochemical response of the NMEs was monitored in a solution containing 3 mM $\text{Ru}(\text{NH}_3)_6^{3+}$ and 0.09 M sodium phosphate, at a scan rate of 100 mV/s. As expected, steady-state voltammograms are observed for each electrode, consistent with the microscale dimensions of the electrodes. For the electrodes made using deposition potentials of 0, -100, or -250 mV, the currents observed are well-correlated with the size of the electrode. That is, the greater the diameter of the NME (i.e., formed at a larger applied potential), the greater the response current observed. However, for structures made at -400 mV, currents observed were lower than expected based on the size of the microelectrode, indicating that the discontinuity of the electrode structure may lead to poor electrical connectivity and loss of working area.

25 Thus, moderate deposition potentials appear to provide the most pronounced nanotexturing while maintaining the integrity of the resultant NME, as small nanoparticles are formed on the surface of the microelectrodes. It appears that providing too large of a driving potential for the deposition reaction accelerates the kinetics to a point where metal nanoparticles are formed without strong connectivity to the core of the NME.

30 NME morphology may also be controlled via electrolyte effects. Pd NMEs were formed at -100 mV for 250s using (a) 0.5M H_2SO_4 and (b) 0.5M HClO_4 as the supporting electrolyte. These structures were formed under the same conditions as described above, where HCl was used as a supporting electrolyte. The structures formed in H_2SO_4 and HClO_4 were significantly larger than those formed in HCl, and interestingly, all three displayed different types of nanostructuring. NMEs made in HClO_4 showed the finest nanostructuring, with features as small as 10-20 nm present. In HCl, the electrode was more

compact, and the nanostructuring was on the order of 100 nm. The coarsest nanostructuring was obtained in H₂SO₄, where the particles comprising the electrode were larger than 200 nm. These observations indicate that the NME morphology may also be controlled by varying the supporting electrolyte used for electrodeposition.

5 Typically, the dendritic structures for the NME depend on the conditions during manufacture, including concentrations of the electrodeposition solution, choice of the metal to be electrodeposited, and the applied potential during electrodeposition. These parameters are readily controllable. For example, it may be desirable to control the concentration and purity of the reagents used in electrodeposition to within 5%. The choice of metal is simple
10 to control as long as purity of the reagents is high, simply by obtaining the correct material. The potential during electrodeposition may be readily controlled to within a few mV, which is sufficient for controlling the size and morphology of the resultant NME.

Example 2. Direct Profiling of Prostate Cancer Biomarkers in Tumor Tissue Using a Multiplexed Nanostructured Microelectrode Integrated Circuit

15 **MATERIALS AND METHODS**

Chip fabrication. The chips were fabricated at the Canadian Photonics Fabrication Center. 3" silicon wafers were passivated using a thick layer of thermally grown silicon dioxide. A 350 nm gold layer was deposited on the chip using electron-beam assisted gold evaporation. The gold film was patterned using standard photolithography and a lift-off
20 process. A 500 nm layer of insulating silicon dioxide was deposited using chemical vapor deposition. 500 nm apertures were imprinted on the electrodes using standard photolithography, and 2 mm x 2 mm bond pads were exposed using standard photolithography.

Fabrication of nanostructured microelectrodes. Chips were cleaned by rinsing in
25 acetone, IPA, and DI water for 30 s and dried with a flow of nitrogen. All electrodeposition was performed at room temperature with a Bioanalytical Systems Epsilon potentiostat with a three-electrode system featuring an Ag/AgCl reference electrode and a platinum wire auxiliary electrode. 500 nm apertures on the fabricated electrodes were used as the working electrode and were contacted using the exposed bond pads. Platinum NMEs were fabricated
30 in a platinum bath containing 5 mM solution of H₂PtCl₆ and 0.5 M HClO₄ at -250 mV for 10 s using DC potential amperometry.

Preparation and purification of oligonucleotides. All synthetic oligonucleotides were stringently purified by reversed-phase HPLC. The following probe and target sequences were used in experiments. Seq. P1. Type III fusion probe (PNA): NH₂-Cys-Gly-ATA AGG CTT CCT GCC GCG CT-CONH₂ (SEQ ID NO. 1), Seq. P2. Type I fusion probe (PNA): NH₂-Cys-Gly-CTG GAA TAA CCT GCC GCG CT-CONH₂ (SEQ ID NO. 2), Seq. P3. Type VI fusion probe (PNA): NH₂-Cys-Gly-ATA AGG CTT CTG AGT TCA AA-CONH₂ (SEQ ID NO. 3), Seq. T1 (Type III TMPRSS2:ERG fusion DNA target): 5'AGC GCG GCA GGA AGC CTT AT3' (SEQ ID NO. 4), Seq. T2 (WT TMPRSS2 DNA target): 5'AGC GCG GCA GGT CAT 10 ATT GA3' (SEQ ID NO. 5), Seq. T3 (WT ERG DNA target): 5'TCA TAT CAA GGA AGC CTT AT3' (SEQ ID NO. 6), Seq. T4 (noncomplementary DNA target): 5'TTT TTT TTT TTT TTT TTT TT3' (SEQ ID NO. 7). Oligonucleotides were quantitated by measuring absorbance at 260 nm and ext. coefficients calculated using: <http://www.idtdna.com/analyzer/Applications/OligoAnalyzer/>.

Modification of NMEs with PNA probes. A solution containing 500 nM thiolated single stranded PNA, 25 mM sodium phosphate (pH 7), and 25 mM sodium chloride was heated at 50°C for 10 minutes. A suitable amount of 10 mM MCH was then added to make the final MCH concentration of 100 nM. 0.5-10 μ L (depending on the degree of multiplexing) of this mixture was deposited on the NMEs in a dark humidity chamber overnight at 4°C. The NMEs were rinsed in 25 mM sodium phosphate (pH 7), and 25 mM NaCl buffer before measurement.

Electrochemical measurements. Electrochemical signals were measured in solutions containing 10 μ M Ru(NH₃)₆³⁺, 25 mM sodium phosphate (pH 7), 25 mM sodium chloride, and 4 mM Fe(CN)₆³⁻. Differential pulse voltammetry (DPV) signals before and after hybridization were measured using a potential step of 5 mV, pulse amplitude of 50 mV, pulse width of 50 ms, and a pulse period of 100 ms. Cyclic voltammetry signals before and after hybridization were collected with a scan rate of 100 mV/s. Limiting reductive current (I) was quantified by subtracting the background at 0 mV from the cathodic current at -300 mV in a cyclic voltammetry signal. Signal changes corresponding to hybridization were calculated as follows: $\Delta I = (I_{ds} - I_{ss}) / I_{ss} \times 100\%$ (ss = before hybridization, ds = after hybridization).

Hybridization protocol. Hybridization solutions typically contained target sequences in 25 mM sodium phosphate (pH 7), and 25 mM NaCl. Electrodes were incubated at 37°C in humidity chamber in dark for 60 minutes and were washed extensively with buffer before electrochemical analysis.

5 **Isolation of mRNA.** The mRNAs were extracted from cell lines and patient tissue samples with the Dynabeads mRNA Direct Kit (Invitrogen). Two typical prostate cancer tissue samples were obtained from radical prostatectomies collected by from the Cooperative Human Tissue Network. The tissue was stored at -85°C until tumor-rich tissue was selected for mRNA extraction. The concentrations of mRNA targets were measured by
10 NanoDrop ND-1000 of Thermo Fisher Scientific (USA). All of the fusion sequences were confirmed by RT-PCR and direct sequencing.

Kinetic measurements of DNA hybridization at NMEs. PNA (seq. 2)-modified NMEs were prepared as described above. Rinsed NMEs were immersed in a solution containing 10 μ M Ru(NH₃)₆³⁺, 4 mM Fe(CN)₆³⁻, 100 fM DNA target (seq.4 to 7), 25
15 mM sodium phosphate (pH 7), and 25 mM NaCl. The electrocatalytic CV signals were obtained as described above. All measurements were performed at 37°C.

RESULTS AND DISCUSSIONS

 We sought to generate a nanomaterial-based platform for ultrasensitive bioanalysis that is i) highly robust and straightforward to fabricate; ii) multiplexed and scalable; and iii)
20 sensitive and specific when presented with heterogeneous biological samples. To satisfy requirements i) and ii) we required a means of achieving reproducible placement of each individual sensing element using a scalable protocol. To address requirement iii), we sought to incorporate nanoscale features into our sensing array. The production of arrayed nanostructured sensing elements, however, can be labor-intensive and prone to low
25 reproducibility. Electron-beam lithography provides the needed control over nanoscale features and their placement; however, it is a serial technique not presently suited to low-cost, high-volume chip production. Our approach was instead to use cost-effective conventional photolithography to position and address our electrodes; and then find a means to bring about, with a high degree of reproducibility, the nanostructuring of these
30 microelectrodes.

We constructed an 8-fold multiplexed chip by patterning a 350 nm thick gold layer on a silicon chip to create eight 5- μ m-wide Au wires attached to large metal pads that would serve as external contacts. SiO₂ was then deposited as a passivating layer and patterned to create apertures with 500 nm diameters at the end of each of the Au wires.

5 These openings were created to serve as individual templates for controlled, local growth of nanostructures. We then used palladium electrodeposition to deposit metal in the patterned apertures. We found that we were able to regulate the size of the nanostructures by varying the deposition time. We were readily able to confine the diameter of the structures to the ultramicroelectrode regime ($<10\ \mu$). Under conditions enabling rapid metal deposition, the

10 surfaces of the microelectrodes displayed a high level of nanostructuring, with feature sizes of approximately 20 nm. These structures displayed ideal microelectrode behavior, exhibiting low capacitive currents and high steady-state plateau currents .

In order to make these nanostructured microelectrodes (NMEs) functional as nucleic acids biosensors, we modified them with thiolated peptide-nucleic acids (PNA) probes.

15 The use of PNA as a probe molecule has been shown previously to increase the sensitivity of biosensing assays and is particularly advantageous in electrochemical assays because it produces lowered background currents. To transduce nucleic acids hybridization into an electrical signal, we employed an electrocatalytic reporter system previously developed by our laboratory. (Lapierre, M. A. et al., Anal. Chem. 2003, 75. 6327-6333). This reporter

20 system relies on the accumulation of Ru(NH₃)₆³⁺ at electrode surfaces when polyanionic species like nucleic acids bind, and the catalysis of the reduction of Ru(III) via the inclusion of Fe(CN)₆³⁻, which regenerates Ru(III) and allows multiple reductions per metal center. When PNA-modified NMEs were challenged with a complementary sequence, detectable signal changes could be clearly detected through the femtomolar concentration range.

25 Negligible signal changes were observed with completely non-complementary sequences.

The cancer biomarkers selected for analysis on this platform are a group of gene fusions specific to prostate cancer. These fusions, resulting from a chromosomal translocation that joins the ERG and TMPRSS2 genes, were recently discovered and appear in at least 50% of prostate tumours. Furthermore, there are ~ 20 sequence types that feature

30 different fusion sites, and the exact type of fusion present in a tumour appears to correlate with its aggressiveness and metastatic potential. These sequences are therefore not only promising diagnostic markers, but are also factors with prognostic value.

To determine whether the NME sensors could discriminate gene fusion sequences from the wild-type sequences that would be half-complementary, a sensor modified with a probe complementary to the splice site of the Type III fusion was challenged with: (1) the fusion target (seq. T1), (2) the sequence corresponding to the wild-type TMPRSS2 gene (seq. T2), and (3) a sequence corresponding to the wild-type ERG gene (seq. T3). A completely non-complementary control was also assayed (seq. T4). With a hybridization time of 60 minutes, large signal increases were observed with the fully complementary target, while a much lower signal change was seen with the TMPRSS2 target. The ERG target produced an even lower signal change, and that observed with the non-complementary sequence was negligible. The TMPRSS2 target binds to the portion of the probe located at the end of the sequence not attached to the electrode, while the ERG target binds to the portion of the probe located at the end tethered to the electrode surface. The different signal levels observed indicate that the most accessible side of the probe is better able to bind incoming target molecules, while hybridization with the more buried part of the sequence is inefficient.

To determine whether the hybridization of the different targets required the full 60 minute time period originally tested for accurate readout, the electrocatalytic signals were monitored at a variety of intervals within the window originally tested. Interestingly, the rise of the signals is very fast, with significant current changes observed within 2 minutes. Over the total 60-minute period, however, the signals for the half-complementary and non-complementary sequences fall noticeably; with 20-50% of the 2-minute signal vanishing by 60 minutes. It appears that for sequences that are not fully complementary, some non-specific binding occurs in the first few minutes of exposure of the NME sensor to the target solution, but these complexes do not remain stable and do not remain immobilized on the electrode. Thus, while non-complementary sequences can be discriminated from complementary sequences with short hybridization times, longer times increase the differential signal changes, and thus the degree of specificity.

The performance of these nanostructured microelectrodes as nucleic acids detectors indicated that the patterned structures were indeed sensitive and specific when used under appropriate hybridization conditions. We therefore sought to prove that multiplexed chip-based NMEs could be used to assay cancer biomarkers presented in heterogeneous biological samples. To explore this capability, cell extracts and tumour samples from

prostate cancer patients were assessed to determine whether the sensitivity and specificity of the system was robust enough for clinical testing.

To determine whether we could detect prostate-cancer associated gene fusions using the NME chip, we first analyzed mRNA isolated from two prostate cancer cell lines: VCaP and DU145. The former cell line is type III fusion positive, and the latter is fusion negative. No appreciable signal changes occurred when 10 ng of mRNA from the cell line that lacks this sequence were incubated with a NME displaying a probe complementary to the type III fusion (seq. P1), while large signal increases were observed in the presence of 10 ng mRNA from the cell line that does contain the type III fusion. In addition, the modification of NMEs with a probe complementary to a different fusion (seq. P2) did not yield a significant signal with positive mRNA sample. The detection of the fused gene is therefore highly specific. These results are significant, as efficiency in the use of sample (10 ng) and the total time required for analysis (less than 1.5 hours) significantly improve upon other detection methods like fluorescence in situ hybridization (FISH) and sequencing.

The ultimate application of the NME chip is the direct, multiplexed analysis of a panel of cancer biomarkers in relevant patient samples. To test the performance of our device for this type of application, we analyzed a panel of mRNA samples collected from cell lines and clinical tumor samples for a series of gene fusions. We obtained a group of samples that would allow the detection of the three most common types of prostate-cancer gene fusions: type I, type III, and type VI. Different clinical outcomes are associated with these sequences, with type III fusions being the most common but correlating with low cancer recurrence rates, whereas type I and VI fusions are correlated with aggressive cancers with high levels of recurrence. It is therefore of great interest to be able to differentiate these fusions in tumours, and a method that would permit their presence or absence to be assessed quickly and straightforwardly would be of value in their further study and validation as diagnostic biomarkers.

Probes complementary to each of the three fusions were deposited on their respective electrodes on NME chips, and 5 different mRNA samples were profiled for the presence of different gene fusions in a multiplexed format. Three cell lines were tested: VCap (type III positive),²⁸ NCI-H660 (type III and VI positive)³⁰, and DU145 (fusion negative).²⁸ In addition, two tumour samples (tissues collected by radical prostatectomies) were tested, one that was positive for the type I fusion, and one that was positive for the

type III fusion, as confirmed by conventional sequencing. In each case, all experiments took less than 2 hours and required only 10 ng of mRNA. By analyzing the electrochemical signals collected at NMEs displaying different probes, we ascertained the identity of fused genes present in each sample. For example, in the patient sample containing the type I fusion (as verified by sequencing), the current values observed at each probe-modified NME decreased in the following order: I >>>> III > VI. In the patient sample containing the type III fusion, the electronic signals again pointed to the correct identity of the fusion with probe III >>>> I > VI. These results, and those obtained with DU145, VCaP, and H660 cellular RNA, where electronic profiling correctly called the absence or presence of gene fusions, indicate that NME chips are able to profile these important biomarkers in complex samples and to distinguish biomarker profiles associated with different clinical outcomes.

The detection platform described here is not only specific, sensitive, and robust, it is also practical and scalable. The reproducible fabrication method we chose is amenable to the production of probe-modified chips using the same photolithographic technologies in widespread use in consumer electronics microchip fabrication; and only simple, inexpensive instrumentation is needed for readout. Microfluidics are not required for automated analysis, as hybridization can be performed and read out in a single reaction vessel. This system represents an attractive alternative to PCR-based methods that are sensitive but difficult to automate in a clinical setting.

In sum, the new multiplexed electrode platform we describe here is the first to read directly a panel of cancer biomarkers in clinically-relevant samples using electronic signals. The array enabling these measurements features microelectrodes that possess controllable and versatile nanotexturing essential for sensitivity. The system combines these nanotextured electrodes with rapid catalytic readout to achieve a long-standing goal: the multiplexed analysis of cancer biomarkers using an inexpensive and practical platform.

Example 3. Direct, Electronic MicroRNA Detection Reveals Differential Expression Profiles in 30 Minutes.

MATERIALS AND METHODS

Materials. 6-mercapto-1-hexanol (97% MCH), hexamine ruthenium chloride (99.9+%), potassium ferricyanide (99%), and palladium (II) chloride(99.9+%) were purchased from Sigma-Aldrich Canada Ltd (Oakville, ON). perchloric acid (70%), acetone (ACS grade) and isopropyl alcohol (IPA, ACS grade) were obtained from EMD (Gibbstown, NJ). Thiolated PNA oligomers were obtained from Biosynthesis Inc (Lewisville, TX) with HPLC purified grade. PNA probes a Cys-Gly dipeptide at their N-terminus. Gly acts as a spacer, while Cys provides free thiol for immobilization on the electrode surface. Synthetic microRNAs (5' end phosphorylated and HPLC purified) were obtained from Eurofins MWG Operon (Huntsville, AL). All PNA and RNA sequences are shown in table S1 provided in the supporting information.

Chip fabrication. The chips were fabricated at the Canadian Photonics Fabrication Center. 3" silicon wafers were passivated using a thick layer of thermally grown silicon dioxide. A 350 nm gold layer was deposited on the chip using electron-beam assisted gold evaporation. The gold film was patterned using standard photolithography and a lift-off process. A 500 nm layer of insulating silicon dioxide was deposited using chemical vapor deposition. 500 nm apertures were imprinted on the electrodes using standard photolithography, and 2 mm x 2 mm bond pads were exposed using standard photolithography.

Fabrication of nanostructured microelectrodes. Chips were cleaned by rinsing in acetone, IPA, and DI water for 30 s and dried with a flow of nitrogen. All electrodeposition was performed at room temperature with a Bioanalytical Systems Epsilon potentiostat with a three-electrode system featuring an Ag/AgCl reference electrode and a platinum wire auxiliary electrode. 500 nm apertures on the fabricated electrodes were used as the working electrode and were contacted using the exposed bond pads. A 2 mm portion of the chip was immersed into the plating bath containing 5 mM palladium (II) chloride and 0.5 M perchloric acid, and incubated for about 5 min prior to electroplating. The bond pads were kept free from solution. Pd NMEs were fabricated using DC potential amperometry at an applied potential of -100 mV for 6 s.

Modification of NMEs with PNA probes. Single-stranded thiolated PNA probes were dissolved in a buffer solution (pH 7) containing 25 mM sodium phosphate and 25 mM

sodium chloride at a concentration of 500 nM. The solution was then heated at 50°C for 10 minutes to fully dissolve the PNA molecules. A suitable amount of 10 mM MCH was then added to make the final MCH concentration of 100 nM. 10 μ L of this mixture was quickly deposited on a chip displaying Pd NMEs using a manual micropipettor. This PNA probe solution covered chip was then incubated in a dark humidity chamber overnight at 4°C. The probe-modified Pd NMEs were vigorously rinsed with the above buffer solution before measurements. For multiplexed experiments, chips with eight individually addressable leads were used.

Target hybridization. Hybridization solutions contained various concentrations of targets in 25 mM sodium phosphate (pH 7.0) and 25 mM NaCl. Pd NMEs were incubated with 10 μ L of target solution at 37°C in a humidity chamber for 30 mins to allow the immobilized probe molecules to hybridize with target molecules. The chip was then cooled and washed vigorously with buffer before the electrochemical analysis.

Electrochemical measurements. Electrochemical measurements were performed with an electrochemical analyzer (BASi, West Lafayette, USA) in a solution containing 10 mM $\text{Ru}(\text{NH}_3)_6^{3+}$, 4 mM $\text{Fe}(\text{CN})_6^{3-}$, 25 mM sodium phosphate (pH 7.0) and 25 mM NaCl. Cyclic voltammetry (CV) was conducted before and after the addition of target solutions at a scan rate of 100 mV/s. Differential pulse voltammetry (DPV) was performed at a potential step of 5 mV, pulse amplitude of 50 mV, pulse width of 50 ms and a pulse period of 100 ms. Cyclic voltammetry signals before and after hybridization were collected with a scan rate of 100 mV/s. Limiting reductive current (I) was quantified by subtracting the background at 0 mV from the cathodic current at -300 mV in a cyclic voltammetry signal. Signal changes corresponding to hybridization were calculated as follows: $\Delta I = (I_{ds} - I_{ss})/I_{ss} \times 100$ (ss = before hybridization, ds = after hybridization). The detection limit was determined as the first concentration where background (noncomplementary ΔI) subtracted signal was 2 times higher than the standard deviation of 10 fM non-complementary control sample.

SEM Imaging. HITACHI S-3400 SEM (Hitachi High Technologies America, Inc., Pleasanton, CA) was employed to study the morphology and dimension of the electroplated NMEs. The chip was affixed on a stainless steel SEM stub using double-sided adhesive black carbon tape. The SEM image was acquired using the secondary electron mode at 20 kV.

RNA extraction for PCR analyses and amplification protocol. Total RNA was extracted from cell lines with mirVana kit (Ambion). The quality of samples was assessed by RT-PCR analysis of the endogenous control RNU44 using Applied Biosystems TaqMan® microRNA Assay. This assay includes a reverse transcription (RT) step using the

5 TaqMan® MicroRNA Reverse Transcription Kit (Applied Biosystems, CA, USA) wherein a stemloop RT primer specifically hybridizes to a mir molecule and is then reverse transcribed with a MultiScribe reverse transcriptase. Briefly, the reverse transcription mix includes 50 nM stem-loop RT primers, 1x RT buffer, 0.25 mM each of dNTPs, 3.33 U/ μ L MultiScribe reverse transcriptase, and 0.25 U/ μ L RNase inhibitor. The 7.5 μ L reaction was

10 then incubated in an Applied Biosystems 7900 Thermocycler for 30 minutes at 16°C, 30 minutes at 42°C, 5 minutes at 85°C and then held at 4°C. The RT products were subsequently amplified with sequence-specific primers (hsa-mir-21 primer 4373090, and hsa-mir-205 primer 4373093 from Applied Biosystems) using the Applied Biosystems 7900 HT Real-Time PCR system. The 10 μ L PCR mix contains 0.67 μ L RT product, 1x4

15 TaqMan® Universal PCR Master Mix, 0.2 μ M TaqMan® probe, 1.5 μ M forward primer and 0.7 μ M reverse primer. The reactions were incubated in a 384-well plate at 95°C for 10 minutes followed by 40 cycles of 95°C for 15 seconds and 60°C for 1 minute.

Cloning and expression of pre-miRNA-21. cDNA was produced by reverse transcribing 500 ng total RNA from FaDu cells using High-Capacity cDNA Reverse

20 Transcription Kit (Applied Biosystems) according to manufacturer's instructions. 50 ng of cDNA were used as template for PCR amplification of the premiRNA-21 stem loop with 1 μ M of the primers MIR-F (CCT ACC ATC GTG ACA TCT CCA TGG) and MIR-R (ATG AGA ACA TTG GAT ATG GAT GGT). The conditions for the PCR were: 95°C for 2 min, followed by 40 cycles of 95°C for 1 min, 55°C for 1 min, and 72°C for 1 min, and a final

25 extension step at 72°C for 10 min. The PCR product was gel purified (Gel extraction kit, Qiagen) and cloned in pCR4-TOPO vector (Invitrogen). Colonies were picked and grown in LB medium. Plasmid DNA was purified with Plasmid Mini Kit (Qiagen), and checked for the correct product, orientation, and absence of undesired mutations with sequencing (ACGT corporation). Plasmid DNA was digested with NotI which does not generate 3'

30 protruding ends that produce high background vector RNA during in vitro transcription. 1 μ g of linearized plasmid DNA was used as a template for in vitro transcription with AmpliScribe™ T3 High Yield Transcription Kit (Epicentre Biotechnologies). After the completion of reverse transcription, DNase I was added to digest the template DNA and

RNA was purified with PureLink Micro to Midi Total RNA Purification Kit (Invitrogen). Concentration of RNA was determined by reading the absorbance at 260 nm.

RESULTS AND DISCUSSIONS

We endeavored to develop a new method for microRNA profiling that would
5 feature the convenience of array-based analysis, but would augment the power of such
multiplexing with the exceptional sensitivity required to assay small biological samples for
low abundance microRNAs. Given that conventional fluorescence-based methods are
insufficiently sensitive to monitor hybridization of small numbers of molecules to surface -
bound probe sequences with simple instrumentation, we instead pursued an approach that
10 employed electronic readout.

In order to provide a platform for electronic microRNA detection, a multiplexed
chip was prepared that featured an electrode pattern generated by photolithography. This
chip was made using a silicon wafer as a base, and a pattern of gold was deposited on its
surface to provide a multiplexed set of leads and external contacts. A layer of SiO was
15 deposited on top of the gold to passivate the metal, and then in the final fabrication step,
500 nanometer apertures were opened on the end of each lead to expose gold. To generate
protruding microelectrodes, palladium was electrodeposited in the apertures. The
electrodeposition step was engineered to produce highly nanostructured microelectrodes
(NMEs). Previous studies have indicated that nanostructured sensing elements can present
20 biomolecular probes more efficiently than bulk materials and facilitate surface
complexation reactions, but this advantage has never been exploited for direct biological
profiling.

To test the electronic chip for sensitivity and specificity in microRNA detection, Pd
NMEs were modified with PNA probes and exposed to RNA for hybridization.
25 Complexation was assayed using a redox reporter system, previously shown to exhibit
femtomolar sensitivity when used in conjunction with nanostructured electrodes and PNA
probes. (R. Gasparac, et. al., J. Am. Chem. Soc. 2004, 126, 12270; Z. Fang, S. O. Kelley,
Anal. Chem. 2009, 81, 612; M. A. Lapierre, et al., Anal. Chem. 2003, 75, 6327; M.A.
Lapierre-Devlin, et al., Nano Lett. 2005, 5, 1051). This reporter system relies on the
30 accumulation of Ru(III) when nucleic acids hybridize at an electrode surface, and the
signals obtained from this reporter are amplified by the inclusion of ferricyanide, which can

regenerate Ru(III) chemically after its electrochemical reduction. Titrations of the miR-21 sequence showed detectable signal changes relative to non-complementary control sequences when solutions containing as little as 10 aM of the target were exposed to the chip-based NMEs. This corresponds to 10 molecules per microliter of sample. The very
5 high level of sensitivity is accompanied by a limited dynamic range of only 10^2 , but for the detection of microRNAs, this tradeoff is merited given the low abundance of these sequences.

Two crucial additional sensing criteria are specifically demanded in microRNA detection. First, closely related sequences - different by as few as one base - must be
10 accurately distinguished. Second, sequence appendages like those found in mature vs. precursor microRNAs, must be discriminated. We sought to challenge our system with each of these requirements. We investigated first the specificity of the assay for mature microRNA sequences. This was conducted by analyzing signal changes observed when the chip was exposed to solutions containing either the full-length, double-stranded, precursor
15 form of miR-21, or the significantly shorter, single-stranded, mature miR-21 sequence. The signal obtained for the hairpin precursor structure approached background levels, while a robust signal change was observed for mature miR-21.

We evaluated the sensitivity of the detection approach to point mutations by monitoring the response of probe-modified sensing elements to two closely related
20 sequences, miR-26a and miR-26b. Probes complementary to each sequence were arrayed on the chip, and the response of these elements to the complementary sequences was monitored. The signal obtained when miR-26a was introduced was approximately 4 times for the fully matched miR-26a probe over the mismatched miR-26b probe, and similarly, the signal obtained when miR-26b was introduced was approximately 4.5 -fold higher for
25 the fully matched miR-26b probe over its mismatched counterpart probe. These results indicate that this multiplexed chip can successfully discriminate closely related microRNA sequences.

Deriving a "fingerprint" of microRNA expression from cell lines representing a particular tumour type relative to normal cells has been previously shown to be a powerful
30 approach to identify microRNAs that can serve as biomarkers in patients. Having confirmed the specificity and sensitivity of the chip towards microRNA targets, we then tested it using RNA samples extracted from human normal cells and those derived from human head and neck squamous cancer cell lines grown in culture. For example, total

RNA extracted from the human hypopharyngeal squamous cancer FaDu cell line and a normal oral epithelial cell line was titrated onto a nanostructured microelectrode displaying a probe complementary to miR -205. A positive signal was obtained with as little as 5 ng of RNA derived from the FaDu cells, while normal epithelial cells did not produce any
5 signal change with up to 20 ng of RNA. This indicates that the signal response corresponds to a unique marker present at significantly higher levels in the cancer cell lines.

We profiled two different microRNAs, miR-21 and miR-205, and also included a control RNA, RNU-44 in a panel of total RNA samples. We employed three different head and neck squamous cancer cell lines, and compared the response of the microelectrode chip
10 to these total RNA samples relative to RNA isolated from normal oral epithelial cells. As expected, RNU -44 levels, as judged by the electrochemical response measured for each total RNA sample exposed to a sensing element modified with a complementary probe, remained constant in all four cell lines. However, miR-21 and miR-205 signals were both significantly elevated in the cancer cell lines. Indeed, the levels of these microRNAs were
15 judged to be present at > 100-fold higher levels in the cancer cell lines relative to the normal epithelial cells. The over-expression of these targets was confirmed using conventional quantitative PCR (see supporting information). Both miR -21 and miR-205 have been previously observed to be elevated in primary human head and neck squamous carcinomas, indicating a significant potential for these micro-RNAs to serve as diagnostic biomarkers
20 for this malignancy.

In conclusion, the microRNA detection chip described here offers the sensitivity and specificity for the analysis of a novel class of nucleic acids biomarkers representing one of the most challenging detection targets. Electronic readout of microRNA profiles offers a rapid - yet highly accurate - method to directly assay RNA samples for specific sequences,
25 and the lack of labeling or amplification renders this approach to be extremely straightforward and efficient, features not attainable with other PCR or hybridization-based approaches.

CLAIMS:

1. A nanostructured microelectrode.
2. The microelectrode of claim 1, which is comprised of a material selected from the group consisting of: a noble metal, an alloy of a noble metal, a conducting
5 polymer, a metal oxide, a metal silicide, a metal nitride, carbon or a combination of any of the same..
3. The microelectrode of claim 1, which is of a height in the range of about 0.5 to about 100 microns.
4. The microelectrode of claim 1 wherein the length of the nanostructure is in
10 the range of about 1 to about 300 nanometers.
5. The microelectrode of claim 1, wherein the length of the nanostructure is in the range of about 10 to about 20 nanometers.
6. The microelectrode of claim 1, wherein the microelectrode has an approximate diameter in the range of about 1 to about 10 microns.
- 15 7. The microelectrode of claim 1, wherein the nanostructure is spiky or fractal.
8. The microelectrode of claim 1 which further comprises is a probe selected from the group consisting of: thiolated probe molecules, nucleic acids, peptide nucleic acids, locked nucleic acids, or phosphorodiamidate morpholino oligomers terminated with thiols, nucleic acids, peptide nucleic acids, locked nucleic acids, or phosphorodiamidate
20 morpholino oligomers terminated with amines, nucleic acids, peptide nucleic acids, locked nucleic acids, or phosphorodiamidate morpholino oligomers terminated with carboxylates, proteins or peptides functionalized with suitable tethering molecules and antibodies functionalized with suitable tethering molecules.
9. A biosensing device comprising a substrate; an electrically conductive lead
25 on the substrate; an insulating layer covering the lead, the insulating layer having an aperture exposing a portion of the lead; and a nanostructured microelectrode, the microelectrode being in electrical communication with the exposed portion of the lead, the microelectrode being adapted to generate a charge in response to a biomolecular stimulus.

10. The biosensing device of claim 9, wherein the lead comprises a material selected from the group consisting of: Au, Al, W, TiN, and polysilicon.

11. The biosensing device of claim 9, wherein the substrate comprises a material selected from the group consisting of: silicon, silica, quartz, glass, sapphire, gallium
5 arsenide, germanium, silicon carbide, indium compounds, selenium sulfide, ceramic, plastic, polycarbonate and other polymer or combinations of any of the above,.

12. The biosensing device of claim 9, wherein the insulating layer is comprised of a material selected from the group consisting of: silicon dioxide, silicon nitride, nitrogen doped silicon oxide, and parylene or combinations of any of the above.

10 13. The biosensing device of claim 9, wherein the aperture has an approximate diameter in the range of about 100 nm to about 10 μ m.

14. The biosensing device of claim 9, further comprising a charge storage in electrical communication with the lead for storing charge from the microelectrode, wherein the charge on the charge storage at the end of an integration period is indicative of the
15 biomolecular stimulus.

15. The biosensing device of claim 9, further comprising a display component for displaying a charge stored in the charge storage.

16. The biosensing device of claim 9, further comprising a transistor.

17. The biosensing device of claim 16 wherein the transistor is a field-effect
20 transistor (FET).

18. The biosensing device of claim 17 wherein the FET is selected from the group consisting of: an n-type silicon channel, and a p-type silicon channel.

19. The biosensing device of claim 16 wherein the transistor is a bipolar transistor.

25 20. The biosensing device of claim 19, wherein the bipolar transistor is selected from the group consisting of: an n-p-n bipolar junction transistor, and a p-n-p bipolar junction transistor.

21. The biosensing device of claim 9, comprising a plurality of leads, each having a respective microelectrode.

22. The biosensing device of claim 21, comprising at least two different probe molecules provided on the microelectrodes.

5 23. The biosensing device of claim 9, wherein the microelectrodes are provided in an array, and each microelectrode is individually electronically accessible.

24. A method for manufacturing a biosensing device, the method comprising: providing a substrate and an electrically conductive lead on the substrate, the lead being covered by an insulating layer; etching an aperture in the insulating layer to expose a
10 portion of the lead; and electrodepositing a conductive material on the exposed portion of the lead to form a nanostructured microelectrode .

25. The method of claim 24 wherein the electrodepositing comprises controlling an applied potential in order to achieve a desired dendritic structure.

26. The method of claim 25, wherein the applied potential is in the range of
15 about 0 mV to about -250 mV.

27. The method of claim 24, wherein the electrodepositing is performed for a time duration in the range of about 25 s to about 500 s.

28. The method of claim 24, wherein the electrodepositing is performed using an electrolyte selected from the group consisting of: H₂SO₄, HCl, and HClO₄.

20 29. The method of claim 24, wherein the etched region is defined photolithographically.

30. The method of claim 24 further comprising functionalizing the microelectrode by binding a probe molecule for a target biomolecule to the surface of the microelectrode.

25 31. The method of claim 30, wherein the probe molecule is functionalized with a thiol, amine or carboxylate group.

32. A method of carrying out a biosensing process using a nanostructured microelectrode comprising:

providing the device of claim 9;

biasing the microelectrode relative to a reference electrode;

5 measuring a reference charge or reference current flow between the microelectrode and the reference electrode;

exposing the microelectrode to a biomolecular stimulus;

measuring a charge or current flow generated at the microelectrode in response to the biomolecular stimulus; and

10 determining the amount of biomolecular stimulus present by comparing the measured charge or measured current flow against the reference charge or reference current flow.

33. A biosensing cartridge comprising: a sample chamber for containing a biological sample; a biosensing chamber for containing a biosensing device in accordance
15 with claim 9.

34. The cartridge of claim 33 wherein the biosensing chamber is adapted to subdivide the biological sample through chemical, mechanical or vibrational means.

35. The cartridge of claim 33 which further comprises a purifying chamber for purifying or isolating the sample.

20 36. The cartridge of claim 33 which further comprises a lid having capsules, the capsules containing reagents for carrying out the biosensing process.

37. The cartridge of claim 33 which is disposable.

38. A biosensing workstation comprising: a cartridge holder for holding a biosensing cartridge in accordance with claim 33; an instrument tip for accessing the
25 biosensing cartridge; a selection mechanism for selecting a biosensing process to be carried out; a processor adapted to carry out the biosensing process using the biosensing cartridge

and to determine results of the biosensing process from electronic signals generated from the biosensing cartridge; and a display for displaying the results of the biosensing process.

39. The workstation of claim 37, wherein there is a plurality of cartridge holders for holding a plurality of biosensing cartridges.

5 40. The workstation of claim 37, wherein the instrument tip is disposable.

41. The workstation of claim 37, further comprising a bar code reader for uniquely identifying the biosensing cartridge.

42. The workstataion of claim 37, further comprising a waste chamber for containing waste products from the biosensing process.

10 43. The workstation of claim 37, wherein the workstation is capable of wireless communication through a wireless network.

1/28




FIG. 1A

2/28




FIG. 1B

3/28





FIG. 1C

4/28

**FIG. 2**





FIG. 2B

6/28

**FIG. 3**

Increasing nanostructuring




FIG. 4

8/28





FIG. 5

9/28

**FIG. 6**

10/28




FIG. 6B




FIG. 7

12/28




FIG. 8




FIG. 9

14/28




FIG. 10

15/28




FIG. 11

16/28




FIG. 12

17/28




FIG. 13

18/28





FIG. 14

19/28

tree



sphere




FIG. 15

20/28




FIG. 16

21/28




FIG. 17

22/28




FIG. 18

23/28




FIG. 19

24/28




FIG. 20

25/28





FIG. 21

26/28

**FIG. 22A**

27/28

(continued)




FIG. 22B




FIG. 23





FIG. 2