



(11)

EP 2 952 170 B1

(12)

EUROPEAN PATENT SPECIFICATION

(45) Date of publication and mention
of the grant of the patent:
08.11.2017 Bulletin 2017/45

(51) Int Cl.:
A61H 31/00 (2006.01)

(21) Application number: **14171457.6**

(22) Date of filing: **06.06.2014**

(54) Mechanical CPR device with automatic suction cup attachment

Mechanische CPR-Vorrichtung mit automatischer Saugnapfbefestigung

Dispositif de réanimation cardiorespiratoire mécanique comportant un dispositif automatique de fixation à ventouse

(84) Designated Contracting States:
**AL AT BE BG CH CY CZ DE DK EE ES FI FR GB
GR HR HU IE IS IT LI LT LU LV MC MK MT NL NO
PL PT RO RS SE SI SK SM TR**

(43) Date of publication of application:
09.12.2015 Bulletin 2015/50

(60) Divisional application:
17167256.1 / 3 231 409

(73) Proprietor: **Physio-Control, Inc.
Redmond, WA 98073 (US)**

(72) Inventors:
• **Nilsson, Lars Anders
232 51 Akarp (SE)**
• **Härdig, Bjarne Madsen
224 67 Lund (SE)**

(74) Representative: **Hughes, Andrea Michelle
Dehns
St Bride's House
10 Salisbury Square
London EC4Y 8JD (GB)**

(56) References cited:
WO-A1-2009/136831 WO-A1-2013/030700

EP 2 952 170 B1

Note: Within nine months of the publication of the mention of the grant of the European patent in the European Patent Bulletin, any person may give notice to the European Patent Office of opposition to that patent, in accordance with the Implementing Regulations. Notice of opposition shall not be deemed to have been filed until the opposition fee has been paid. (Art. 99(1) European Patent Convention).

Description

BACKGROUND

[0001] Cardiopulmonary resuscitation (CPR) is a medical procedure performed on patients to maintain some level of circulatory and respiratory functions when patients otherwise have limited or no circulatory and respiratory functions. CPR is generally not a procedure that restarts circulatory and respiratory functions, but can be effective to preserve enough circulatory and respiratory functions for a patient to survive until the patient's own circulatory and respiratory functions are restored. CPR typically includes frequent torso compressions that usually are performed by pushing on or around the patient's sternum while the patient is lying on the patient's back. For example, torso compressions can be performed as at a rate of about 100 compressions per minute and at a depth of about 5 cm per compression for an adult patient. The frequency and depth of compressions can vary based on a number of factors, such as valid CPR guidelines.

[0002] Mechanical CPR has several advantages over manual CPR. A person performing CPR, such as a medical first-responder, must exert considerable physical effort to maintain proper compression timing and depth. Over time, fatigue can set in and compressions can become less consistent and less effective. The person performing CPR must also divert mental attention to performing manual CPR properly and may not be able to focus on other tasks that could help the patient. For example, a person performing CPR at a rate of 100 compressions per minute would likely not be able to simultaneously prepare a defibrillator for use to attempt to restart the patient's heart. Mechanical compression devices can be used with CPR to perform compressions that would otherwise be done manually. Mechanical compression devices can provide advantages such as providing constant, proper compressions for sustained lengths of time without fatiguing, freeing medical personnel to perform other tasks besides CPR compressions, and being usable in smaller spaces than would be required by a person performing CPR compressions. WO2009/136831 discloses a mechanical CPR device, comprising: a piston having a suction cup attached to an end thereof; a driving component configured to extend the piston toward a patient's torso and to retract the piston away from the patient's torso; and a controller configured to determine a reference position by controlling the driving component to at least: extend the piston until a first position at which the suction cup comes into contact with the patient's torso, further extend the piston to cause air to be forced out from an area between the suction cup and the patient's torso until a first threshold is reached, retract the piston until the suction cup is at the first position, wherein the controller is further configured to perform mechanical CPR by controlling the driving component to at least: compress the patient's torso by extending the piston to

a depth and retracting the piston from the depth.

SUMMARY

[0003] According to the invention, a device as claimed in claim 1 is provided.

[0004] The mechanical CPR device includes a piston having a suction cup attached to an end thereof, a driving component configured to extend the piston toward a patient's torso and retract the piston away from the patient's torso, and a controller. The controller determines a reference position by at least controlling the driving component to extend the piston until a first position at which the suction cup comes into contact with the patient's torso, further extend the piston to cause air to be forced out from an area between the suction cup and the patient's torso until a first threshold is reached, retract the piston until the suction cup is at the first position, further retract the piston from the first position until a second threshold is exceeded, and extend the piston to a second point at which the second threshold is no longer exceeded, the reference position being based at least in part on the second point. The controller performs mechanical CPR by controlling the driving component to compress the patient's torso by extending the piston from the reference position to a depth and retracting the piston from the depth to the reference position, and actively decompress the patient's torso by retracting the piston from the reference position to a height above the reference position. As used in this context, to actively decompress, an external force is applied to the patient's torso to decompress the torso above the torso's natural resting position and/or above a reference position that is above the torso's natural resting position, as opposed to merely discontinuing the externally applied force and allowing the torso to expand by the natural resiliency of the torso. In one embodiment, the torso can be lifted up to 10% beyond the torso's natural resting position to actively expand the patient's torso during decompression.

[0005] In some examples, the controller can be configured to compress the patient's torso and actively decompress the patient's torso in a cycle based on a frequency. The frequency can be a predetermined frequency or a frequency entered by a user into the mechanical CPR device. The depth can be a predetermined depth, a depth entered by a user into the mechanical CPR device, or a depth based on a force used to compress the patient's torso. The height can be a predetermined height, a height entered by a user into the mechanical CPR device, or a height based on a force used to actively decompress the patient's torso.

[0006] In other examples, the first threshold can be a force threshold and the mechanical CPR device can also include a force sensor to sense the force applied by the piston to cause air to be forced out from an area between the suction cup and the patient's torso. The second threshold can be a force threshold and the mechanical CPR device can also include a spring activation sensor

configured to signal when the piston has been extended to exceed the second threshold. The spring activation sensor can also stop signaling when the piston has been extended to the second point at which the second threshold is no longer exceeded. One or both of the first and second thresholds can be a pressure threshold, and the mechanical CPR device can further include a pressure sensor configured to sense pressure in the area between the suction cup and the patient's torso. The controller can determine the reference position in response to the mechanical CPR device receiving a user input. The controller can also determine the reference position a predetermined number of times before performing mechanical CPR.

[0007] According to the invention, a suction cup on the end of a piston of a mechanical CPR device is automatically attached to a patient's torso. The mechanical CPR device extends the piston until a first position at which the suction cup comes into contact with the patient's torso. The piston is further extended to cause air to be forced out from an area between the suction cup and the patient's torso until a first threshold is reached. The piston is retracted until the suction cup is at the first position. The piston is further retracted from the first position until a second threshold is exceeded. The piston is then extended to a second point at which the second threshold is no longer exceeded.

[0008] In one example, each of the first and second thresholds is at least one of a force threshold, a distance threshold, or a pressure threshold. The mechanical CPR device can include a spring activation sensor to signal when the piston has been extracted to exceed the threshold. The spring activation sensor can stop signaling when the piston has been extended to the second point at which the threshold is no longer exceeded.

[0009] Mechanical CPR is performed by the mechanical CPR device. The mechanical CPR device automatically attaches a suction cup of the mechanical CPR device to a patient's torso, automatically determines a reference position of the suction cup, extends the piston from the reference position to a particular depth below the reference position, retracts the piston from the particular depth to a particular height above the reference position, and extends the piston from the particular height to the reference position.

[0010] Extending the piston from the reference position to the particular depth and retracting the piston from the particular depth to the reference position causes compression of the patient's torso, and retracting the piston from the reference position to the particular height causes active decompression of the patient's torso. The compression of the patient's torso and the active decompression of the patient's torso can be performed a number of times in a cycle. The cycle can be performed based on a frequency, where the frequency is either a predetermined frequency or a frequency entered by a user into the mechanical CPR device. The particular depth can be a predetermined depth, a depth entered by a user into

the mechanical CPR device, or a depth based on a force used to compress the patient's torso. The particular height can be a predetermined height, a height entered by a user into the mechanical CPR device, or a height based on a force used to actively decompress the patient's torso.

BRIEF DESCRIPTION OF THE DRAWINGS

[0011] Throughout the drawings, reference numbers may be re-used to indicate correspondence between referenced elements. The drawings are provided to illustrate example embodiments described herein and are not intended to limit the scope of the disclosure.

Figures 1A and 1B depict an isometric view and a side view, respectively, of one embodiment of a mechanical CPR device.

Figures 2A, 2B, 2C, 2D, 2E, and 2F depict a system and a method for automatically attaching a suction cup of the mechanical CPR device to a patient's torso and automatically determining a reference position for a piston with respect to a patient's torso.

Figure 3 depicts an example of a method of automatically attaching a suction cup of a mechanical CPR device to a patient's torso and of automatically determining a reference position for a piston with respect to a patient's torso.

Figures 4A, 4B, 4C, 4D, and 4E depict a system and method of performing one cycle of mechanical CPR that includes both compression and active decompression.

Figure 5 depicts an example of a method of performing one cycle of mechanical CPR that includes both compression and active decompression.

Figures 6A, 6B, and 6C depict different wave forms representing positions of a piston with respect to a reference position during compression and decompression of a patient's torso.

DETAILED DESCRIPTION OF ILLUSTRATIVE EMBODIMENTS

[0012] Mechanical CPR compression devices can provide many advantages over manual CPR compressions. Mechanical CPR compression devices can include a back plate that is placed behind the back of the patient and a compression device located above the patient's sternum area. The compression device can be connected to the back plate on both sides of the patient. When the compression device pushes against the area around the patient's sternum, the back plate provides resistance that allows the compression device to compress the patient's torso. Such mechanical CPR compression devices surround the user's torso, such as in the case of a mechanical CPR device with a back plate behind the patient's back, a compression device above the patient's sternum, and legs along both sides of the user's torso.

[0013] Figures 1A and 1B depict an isometric view and a side view, respectively, of one embodiment of a mechanical CPR device 100. The mechanical CPR device 100 includes a lower portion 110 and an upper portion 120. The lower portion 110 can be in the form of a back plate that can be placed under the back of a patient. The upper portion 120 can have a main portion 121 and two legs 122 and 123. Each of the legs 122 and 123 can be releasably connected to one of the sides of the lower portion 110. Items that are releasably connected are easily disconnected by a user, such as connections that can snap in and snap out, connections that do not require the use of tools to disconnect, quick-release connections (e.g., push button release, quarter-turn fastener release, lever release, etc.), and the like. Items are not releasably connected if they are connected by more permanent fasteners, such as rivets, screws, bolts, and the like. In the embodiment shown in Figures 1A and 1B, the legs 122 and 123 are rotatably attached to the main portion 121 about axes 124 and 125, respectively. However, in other embodiments, the legs 122 and 123 can also be fixed with respect to the main portion 121.

[0014] The main portion 121 can include a piston 126 with an end 127. The end 127 can be blunt, contoured, or otherwise configured to interact with a patient's torso. The end 127 can also have a suction cup that can temporarily attach to a patient's torso. The main portion 121 can include other components. For example, the main portion 121 can include a driving component, such as a motor or actuator, that can extend and retract the piston 126. The main portion 121 can include a power source, such as a rechargeable battery, that can provide power for the driving component. The main portion 121 can also include a controller that can control the movement of the piston 126 by controlling the driving component. In one embodiment, the controller can include a processor and memory, and the memory stores instructions that can be executed by the processor. The instructions can include instructions for controlling the piston 126 by controlling the driving component. The main portion 121 can also include one or more sensors that can provide inputs to the controller. The one or more sensors can include one or more of a force sensor to sense a force exerted by the piston 126, a spring sensor to sense a displacement of the piston 126, a current sensor to sense an amount of current drawn by the driving component, or any other type of sensor. The main portion 121 can also include one or more user input mechanisms, such as buttons, keys, displays, and the like. A user can input information to adjust the operation of the mechanical CPR device 100, such as a depth of compressions, a frequency of compressions, a maximum exertion force by the piston 126, and the like.

[0015] In addition to the mechanical CPR device 100, Figure 1B also depicts a cross section of a patient's torso 130 with the patient's back against the lower portion 110 and the patient's chest facing the piston 126. While in the configuration depicted in Figure 1B, the piston can

be extended to the patient's torso 130, compress the patient's torso 130, and retract from the patient's torso. This process, wherein the piston 126 compresses the patient's torso 130 and is then retracted from the patient's torso, can be performed repeatedly to mechanically perform CPR.

[0016] Figures 2A to 2F depict embodiments of a system and a method for automatically attaching a suction cup of the mechanical CPR device 220 to a patient's torso 210 and automatically determining a reference position for a piston with respect to a patient's torso 210. Figures 2A to 2F depict a portion of a mechanical CPR device 220 that includes a piston 221. The end of the piston 221 includes a suction cup 222. The depictions in Figures 2A to 2F show cross sectional views of the mechanical CPR device 220, the piston 221, and the suction cup 222. The mechanical CPR device 220 could also include other components that are not depicted in Figures 2A to 2F, such as a back plate, legs to couple the depicted portions of the mechanical CPR device 220 to the back plate, and the like.

[0017] In Figure 2A, the piston is fully retracted into the mechanical CPR device 220. In this position the suction cup 222 is not in contact with the patient's torso 210. From the position depicted in Figure 2A, the piston 221 can be extended until the piston 221 is in the position depicted in Figure 2B where the suction cup 222 is in contact with the patient's torso 210. The piston 221 can be extended by a driving component, such as a motor or an actuator, in the mechanical CPR device 220. A controller in the mechanical CPR device 220 can also control the driving component.

[0018] From the position depicted in Figure 2B, the piston 221 can be further extended toward the patient's torso 210 until a threshold is reached so that air is forced out from the lower side of the suction cup 222, such as in the position depicted in Figure 2C. In one example, the threshold can be a force threshold and the controller in the mechanical CPR device 220 can measure the force exerted by the piston 221 as the air is forced out from the lower side of the suction cup 222. Once the force exerted on the by the piston 221 reaches the force threshold, the controller can stop the piston 221 from being extended any further. In another example, the threshold can be a distance threshold and the controller in the mechanical CPR device 220 can measure the distance travelled by the piston 221 as the air is forced out from the lower side of the suction cup 222. Once the distance travelled by the piston 221 reaches the distance threshold, the controller can stop the piston 221 from being extended any further. In yet another example, the threshold can be a pressure threshold and a pressure sensor can sense the pressure in the area between the suction cup 222 and the patient's torso 210. As the air is forced out from the lower side of the suction cup 222, the pressure sensor will sense a reduction in pressure. Once the pressure reaches the pressure threshold, the controller in the mechanical CPR device 220 can stop the piston 221 from

being extended any further. In any of these examples, the patient's torso 210 may be compressed to some extent as the piston 221 is extended, such as in the depiction in Figure 2C. At the point depicted in Figure 2C, the suction cup 222 is attached to the patient's torso 210 from the vacuum created by the air forced out of the lower side of the suction cup 222.

[0019] From the position depicted in Figure 2C, the piston 221 can be retracted to the position depicted in Figure 2D where the suction cup 222 originally came into contact with the patient's torso 210. From the position depicted in Figure 2D, the piston 221 can be further retracted until the point depicted in Figure 2E where the piston 221 reaches a second threshold. The second threshold can be a force threshold, such as a force exerted when pulling up on the patient's torso 210. This second threshold can be measured by a spring activation sensor or other force sensor. For example, the piston 221 can be retracted until the spring activation sensor is activated and then the driving component can stop retracting the piston 221. From the position depicted in Figure 2E, the piston 221 can be slowly extended back toward the patient's torso 210 until the location depicted in Figure 2F where the piston 221 no longer exceeds the second threshold. At this position, the location of the suction cup 222 can define a reference position 230 for the piston 221.

[0020] As described in greater detail below, the reference position 230 is a position from which the depth of CPR compressions and the height of CPR decompressions are measured. Defining and using reference position 230 as a position from which to measure the depth of CPR compressions and the height of CPR decompressions can help to avoid unintended injury to a patient. For example, a manual CPR device can be placed on a patient's torso and a user can manually push or pull on the manual CPR device to cause compressions or decompressions. However, the user of the manual CPR device does not have any reference position from which to measure the depth of compressions or the height of decompressions. Without a reference position, the user can cause additional injuries to the patient. For example, if the user pushes the manual CPR device down too far into the patient's chest during a compression, the compression might break one or more of the patient's ribs. When one or more of the patient's ribs are broken, it may be easier to compress the patient's chest and a subsequent compression by user of the manual CPR device can cause even more of the patient's ribs to be broken, and injury to the patient's internal organs. In contrast, establishing reference position 230 with respect to the patient's torso 210 can prevent CPR compressions from extending too deep. Moreover, even if one injury does occur (e.g., the breaking of a patient's rib), the reference position 230 will not change and the likelihood that a subsequent compression will cause even further injury can be reduced.

[0021] In addition to merely using a reference position 230, establishing a proper location for the reference po-

sition 230 can also help to avoid unintended injury to a patient. Retracting the piston 221 until the second threshold is exceeded and then extending the piston 221 until the second threshold is no longer exceeded (as shown in Figures 2E and 2F) can establish a reference position 230 that is closer to the natural resting position of the patient's torso 210. Using a reference position 230 that is close to the natural resting position of the patient's torso 210 can help to avoid unintended injury to a patient. For example, if the point at which the suction cup attached to the patient's torso 210 (i.e., the point shown in Figure 2C) was used as reference position, the reference position could be too low. If CPR compressions were measured from this low reference position, the depth of CPR compressions could cause injury such as breaking the patient's ribs, bruising the patient's internal organs, and the like. In another example, if the point at which the second threshold is exceeded (i.e., the point shown in Figure 2E) was used as a reference position, the reference position could be too high. In such a case, the CPR compressions would not extend low enough to properly compress the patient's torso 210 and decompression of the patient's torso 210 would go higher than desired and possibly cause damage from overstretching.

[0022] Using a reference position can also be beneficial in circumstances where the patient is not located in a stable or a flat position. For example, if a patient is being transported, such as on a stretcher or an ambulance, the patient may be jostled around or otherwise not in a stable position. However, if the mechanical CPR device is moving with the patient (e.g., if mechanical CPR is being performed in an ambulance while the patient is being transported), the reference position of the piston or suction cup can remain relatively fixed with respect to the patient and the mechanical CPR device can avoid over-compression and over-decompression. Thus, the benefits of avoiding unintended injury could still be realized if the patient is otherwise moving. In another example, the patient can be located in a position that is not flat, such as if the patient is being transported down stairs or the patient is on rough terrain. In these cases, if the mechanical CPR device is located with the patient in the same non-flat position, the reference position used by the mechanical CPR device would reflect the patient's non-flat position and the mechanical CPR device could avoid over-compression and over-decompression. A user performing manual CPR under such conditions may have difficulty in maintaining a desired compression depth and/or decompression height.

[0023] A number of other benefits can be realized using the process depicted in Figures 2A to 2F. One example is that the process depicted in Figures 2A to 2F can be used to automatically attach the suction cup 222 to the patient's torso 210. A controller in the mechanical CPR device 220 can control the movements of the piston 221 to perform the entire process depicted in Figures 2A to 2F. In this way, the suction cup 222 can be attached to the patient's torso 210 without manual intervention by a

user of the mechanical CPR device 220. A user can initiate the process of depicted in Figures 2A to 2F, such as by pressing a particular button or key on the mechanical CPR device 220. However, once the process depicted in Figures 2A to 2F is initiated, the mechanical CPR device 220 can automatically attach the suction cup 222 to the patient's torso. The process depicted in Figures 2A to 2F can be repeated a number of times, such as three times, to ensure that the suction cup 222 is attached to the patient's chest and/or to ensure that the reference position 230 was determined correctly. In one embodiment, the reference position 230 can be determined more than one time and the average measurement of the location of the reference position 230 can be used as a reference for CPR compressions and CPR decompressions. Another example is that the process depicted in Figures 2A to 2F can define a reference position 230 of the piston 221 with respect to the patient's torso 210. The reference position 230 of the piston 221 with respect to the patient's torso 210 may vary from patient to patient as different patients may have torsos of different sizes.

[0024] Figure 3 depicts an example of a method 300 of automatically attaching a suction cup of a mechanical CPR device to a patient's torso and of automatically determining a reference position for a piston with respect to a patient's torso. At block 301, a piston can be extended until a suction cup on the end of the piston makes contact with a patient's torso. At block 302, the piston can be further extended until a first threshold is reached. The first threshold can be a threshold amount of force exerted by the piston on the patient's torso. In this instance, the first threshold can be an amount of force that will force air out from the lower side of the suction cup to create a vacuum between the suction cup and the patient's torso. The first threshold can also be a distance threshold relating to the distance travelled by the piston, a pressure threshold relating to the pressure between the suction cup and the patient's torso, or any other type of threshold.

[0025] At block 303, the piston can be retracted beyond the point at which the suction cup first contacted the patient's torso until a second threshold is passed. The second threshold can be a force threshold that is passed when the force used to perform the active decompression is greater than the second threshold. The second threshold can also be a distance threshold relating to the distance travelled by the piston, a pressure threshold relating to the pressure between the suction cup and the patient's torso, or any other type of threshold. The point at which the second threshold has been passed can be signaled by a spring activation sensor. Retracting the piston in this way ensures that the suction cup is properly attached to the patient's torso. At block 304, the piston can be extended back toward the patient's torso until the point that the second threshold is no longer exceeded. In the case where a spring activation sensor is used, the spring activation sensor signal can cease once the piston no longer exceeds the second threshold. At block 305, the piston can be stopped and the location of the piston at

that point can be defined as a reference position. At this point, the suction cup is attached to the patient's torso and the reference position can be used during mechanical CPR for compression and active decompression.

[0026] The method 300 depicted in Figure 3 can be performed by a mechanical CPR device. A controller in the mechanical CPR device can be configured to perform each of the steps depicted in method 300. The mechanical CPR device can include executable instructions that, when executed by the mechanical CPR device, cause the mechanical CPR device to performing the method 300.

[0027] Figures 4A to 4E depict a system and method of performing one cycle of mechanical CPR that includes both compression and active decompression. Depicted in Figures 4A to 4E are a patient's torso 410 and a portion of a mechanical CPR device 420. The mechanical CPR device 420 includes a piston 421 and a suction cup 422 on the end of the piston 421. At the point depicted in Figure 4A, the suction cup 422 is attached to the patient's torso 410. The suction cup 422 could have been automatically attached to the patient's torso 410 using a method, such as the one depicted in Figures 2A to 2F or in Figure 3. Also at the point depicted in Figure 4A, the suction cup 422 is located at a reference position 430. The reference position 430 could have been automatically determined using a method, such as the one depicted in Figures 2A to 2F or in Figure 3. While the reference position 430 can be set in automatically by the mechanical CPR device 420, the reference position 430 can be set in in a number of other ways. For example, the reference position 430 can be set in manually by a user, reference position 430 can be set in by a user manually adjusting the reference position 430 after an initial automatic or manual setting of the reference position 430.

[0028] From the point depicted in Figure 4A, the piston can be extended to compress the patient's torso 410 until it reaches the point depicted in Figure 4B. In Figure 4B, a portion of the patient's torso 410 has been compressed to a depth 431. The depth 431 can be a predetermined depth, a depth entered by a user into a user interface of the mechanical CPR device 420, a depth based on the force required to compress the patient's torso 410, or any other depth. From the point depicted in Figure 4B, the piston can be retracted until the point shown in Figure 4C where the suction cup 422 is back at the reference position 430.

[0029] From the point depicted in Figure 4C, the piston can be retracted to actively decompress the patient's torso 410 until it reaches the point depicted in Figure 4C. In Figure 4B, a portion of the patient's torso 410 has been actively decompressed to a height 432. The height 432 can be a predetermined height, a height entered by a user into a user interface of the mechanical CPR device 420, a height based on the force required to actively decompress the patient's torso 410, or any other depth. From the point depicted in Figure 4D, the piston can be extended until the point shown in Figure 4E where the

suction cup 422 is back at the reference position 430.

[0030] The cycle of compression and decompression depicted in Figures 4A to 4E can be repeated any number of times as part of a mechanical CPR process. The active decompression that is part of the cycle can increase the effectiveness of the mechanical CPR. For example, adding active decompression to the mechanical CPR method can improve the venous return flow of blood back to the heart which can improve the patient's cardiac output. The frequency with which the cycle is repeated can be a pre-determined frequency, a frequency entered by a user into a user interface of the mechanical CPR device 420, or any other frequency.

[0031] In the cycle of compression and decompression depicted in Figures 4A to 4E, the piston 421 does not need to stop at each of the positions depicted in Figures 4A to 4E. For example, when the piston 421 is at the location depicted in Figure 4B (i.e., where the suction cup 422 has been extended to the depth 431 below the reference position 430), the piston 421 can be retracted without interruption from that position to the position depicted in Figure 4D (i.e., where the suction cup 422 has been retracted to the height 432 above the reference position 430). During this movement, the piston 421 will pass through the position shown in Figure 4C (i.e., where the suction cup 422 is at the reference position 430), but not stop at the position shown in Figure 4C. In this way, the position of the piston 421 during repeated cycle of compression and decompression can be represented by a square wave, by a sine wave, or by any other wave pattern. In the square wave example, the piston 421 can extend to the depth 431 below the reference position 430, wait at the depth 431 below the reference position 430 for a time, retract from the depth 431 below the reference position 430 to the height 432 above the reference position 430, wait for a time, extend to the depth 431 below the reference position 430, and so forth.

[0032] Figures 6A to 6C depict different wave forms representing positions of a piston 601 with respect to a reference position 603 during compression and decompression of a patient's torso 602. In the example depicted in Figure 6A, a chart 610 plots the position of a suction cup of the piston 601 over time. The chart 610 also depicts a height of the reference position 603. At time 601, the suction cup of the piston 601 is at the reference point. The piston can be extended to compress the patient's torso 602, as shown by downward slope 612 on the chart, until the suction cup reaches a particular depth below reference position 603. The position of the suction cup at the particular depth below reference position 603 is shown at point 613. The piston can remain in the position shown at point 613 for a time and then be retracted, as shown by upward slope 614, to decompress the patient's torso 602 until the suction cup reaches a particular height above reference position 603. The position of the suction cup at the particular height above reference position 603 is shown at point 615. The piston can remain in the position shown at point 615 for a time and then be extended,

as shown by downward slope 616, to compress the patient's torso 602 until the suction cup again reaches the particular depth below the reference position 603. The position of the suction cup at the particular depth below reference position 603 is again shown at point 617. As shown in this particular embodiment, the speeds at which the piston is extended and retracted can be different. For example, the downward slope 616 is steeper than the upward slope 614, indicating that speed of extending the piston during downward slope 616 is greater than the speed of retracting the piston during upward slope 614. This scenario may allow for blood to be drawn slowly into the patient's heart during decompression and then quickly pumped out of the heart during compression. Other speeds and differences in speeds are possible.

[0033] Figure 6B depicts a chart 620 representing heights of a suction cup on a piston with respect to reference position 603. At point 621, the suction cup can be located at the reference point. From there, a series of cycles of compression and decompression proceed. The compressions are made as the piston is extended until the suction cup reaches a particular depth below the reference position 603, as shown by the position at point 622. The decompressions are made as the piston is retracted until the suction cup reaches a particular height above the reference position 603, as shown by the position at point 623. In this particular embodiment, the cycles are performed at different frequencies. For example, cycles on the left side of the chart 620, such as cycle 624, are performed at a first frequency, and cycles on the right side of the chart 620, such as cycle 625, are performed at a second frequency. In this particular example the second frequency is higher than the first frequency. A low frequency in the first part of the chart 620 and a high frequency in the second part of the chart 620 may aid in preventing reperfusion injury or other injuries. In addition, the mechanical CPR machine may pause for a period, such as the rest period 626 depicted in Figure 6B, between different frequencies of operation. Other patterns of frequencies are possible and can be predetermined frequencies or user-entered frequencies. In one embodiment, the mechanical CPR device can perform compressions for a time without first defining a reference position and then rest for a time. During the rest time, the mechanical CPR device can define a reference position. After the rest time, the mechanical CPR device can perform compressions and active decompressions using the defined reference position to measure depth and height.

[0034] Figure 6C depicts a chart 630 representing heights of a suction cup on a piston with respect to reference position 603. At point 631, the suction cup can be located at the reference point. From there, a series of cycles of compression and decompression proceed. The compressions are made as the piston is extended until the suction cup reaches a particular depth below the reference position 603, as shown by the position at point 632. The decompressions are made as the piston is retracted until the suction cup reaches a particular height

above the reference position 603, as shown by the position at point 633. In chart 630, different duty cycles for compression and decompression are depicted. A duty cycle is a percentage of one period during which a particular characteristic is true. For example, the compression duty cycle can be measured as the percentage of one period during which the patient's torso is compressed. In Figure 6C, chart 630 depicts that compressions are held for a first period of time 634 and decompressions are performed for a second period of time 635. One full cycle or period takes a third period of time 636. The compression duty cycle is the percentage the third period of time 636 taken up by the first period of time 634 and the decompression duty cycle is the percentage the third period of time 636 taken up by the second period of time 635. In the example of Figure 6C, the compression duty cycle is a lower percentage than the decompression duty cycle because the first period of time 634 is less than the second period of time 635. Any duty cycle for the wave form (i.e., either the compression duty cycle or the decompression duty cycle) can be a predetermined duty cycle or a user-entered duty cycle.

[0035] Figure 5 depicts an example of a method 500 of performing one cycle of mechanical CPR that includes both compression and active decompression. At block 501, a suction cup of a mechanical CPR device can be automatically attached to a patient's torso. At block 502, a reference position of the suction cup can be determined. As described above with respect to the methods depicted in Figures 2A to 2F or in Figure 3, automatically attaching a suction cup of a mechanical CPR device and determining a reference position of the suction cup can be performed in the same process. Both automatically attaching a suction cup of a mechanical CPR device and determining a reference position of the suction cup can be performed by the mechanical CPR device.

[0036] At block 503, the piston can be extended until the suction cup is depressed a certain depth from the reference position. Extending the piston in this manner will cause the suction cup to compress the patient's torso. The depth can be a predetermined depth, a depth entered by a user into a user interface of the mechanical CPR device, a depth based on the force required to compress the patient's torso, or any other depth. At block 504, the piston can be retracted until the suction cup is returned to the reference position. At that point, the patient's torso is no longer in compression.

[0037] At block 505, the piston can be retracted until the suction cup is withdrawn a certain height from the reference position. Retracting the piston in this manner will cause the suction cup to actively decompress the patient's torso. The height can be a predetermined height, a height entered by a user into a user interface of the mechanical CPR device, a height based on the force required to actively decompress the patient's torso, or any other height. At block 506, the piston can be extended again until the suction cup is returned to the reference position. At that point, the patient's torso is no

longer in active decompression.

[0038] When performing the method 500 depicted in Figure 5, the piston does not need to stop moving after each of the steps described in method 500. For example, while block 504 indicates that the piston is retracted until the suction cup is at the reference position and block 505 indicates that the piston is further retracted until the suction cup is at the height above the reference position, the piston does not need to stop at the reference position. The piston can continuously move from the position at which the suction cup is at the depth below the reference position until the suction cup is at the height above the reference position. In another example, while block 506 indicates that the piston is extended until the suction cup is at the reference position, the piston can continue to move until the suction cup is at the depth below the reference position to start another cycle.

[0039] In any of the above examples, a suction cup can become disengaged from the patient's torso during CPR. The disengagement can be measured in a number of ways, such as by a pressure sensor configured to measure the pressure below the suction cup, a sensor that measures the force used during decompression, and the like. In such a case, the mechanical CPR device can automatically reattach the suction cup to the patient's torso and/or provide an alert (e.g., audio alert via a speaker, visual alert via a warning light, etc). The suction cup can be reattached to the patient's torso using the same method that it was originally attached to the suction cup, such as using the process depicted in Figures 2A to 2F. The mechanical CPR device can store data about a disengagement event in memory for later analysis. After reattachment, the mechanical CPR device can also modify its operation, such as by changing the compression and decompression waveform, changing the amount of force used to extend and retract the piston, changing the speed at which the piston is extend and/or retracted, etc.

[0040] In a number of embodiments discussed here, a suction cup has been described on the end of a piston. The suction cup can attach to a patient's torso so that, among other benefits, active decompression is possible.

[0041] Conditional language used herein, such as, among others, "can," "could," "might," "may," "e.g.," and the like, unless specifically stated otherwise, or otherwise understood within the context as used, is generally intended to convey that certain examples include, while other examples do not include, certain features, elements, and/or steps. The terms "comprising," "including," "having," and the like are synonymous and are used inclusively, in an open-ended fashion, and do not exclude additional elements, features, acts, operations, and so forth. Also, the term "or" is used in its inclusive sense (and not in its exclusive sense) so that when used, for example, to connect a list of elements, the term "or" means one, some, or all of the elements in the list.

[0042] Each of the processes, methods and algorithms described in the preceding sections may be embodied in, and fully or partially automated by, code modules ex-

ecuted by one or more computers or computer processors. The code modules may be stored on any type of non-transitory computer-readable medium or computer storage device, such as hard drives, solid state memory, optical disc and/or the like. The processes and algorithms may be implemented partially or wholly in application-specific circuitry. The results of the disclosed processes and process steps may be stored, persistently or otherwise, in any type of non-transitory computer storage such as, e.g., volatile or non-volatile storage.

[0043] It will also be appreciated that various items are illustrated as being stored in memory or on storage while being used, and that these items or portions of thereof may be transferred between memory and other storage devices for purposes of memory management and data integrity. Alternatively, in other embodiments some or all of the software modules and/or systems may execute in memory on another device and communicate with the illustrated computing systems via inter-computer communication. Furthermore, in some embodiments, some or all of the systems and/or modules may be implemented or provided in other ways, such as at least partially in firmware and/or hardware, including, but not limited to, one or more application-specific integrated circuits (ASICs), standard integrated circuits, controllers (e.g., by executing appropriate instructions, and including microcontrollers and/or embedded controllers), field-programmable gate arrays (FPGAs), complex programmable logic devices (CPLDs), etc. Some or all of the modules, systems and data structures may also be stored (e.g., as software instructions or structured data) on a computer-readable medium, such as a hard disk, a memory, a network or a portable media article to be read by an appropriate drive or via an appropriate connection. Such computer program products may also take other forms in other embodiments. Accordingly, the present invention may be practiced with other computer system configurations.

[0044] While certain examples or illustrative examples have been described, these examples have been presented by way of example only, and are not intended to limit the scope of the invention disclosed herein. Indeed, the systems described herein may be embodied in a variety of other forms. The scope of the invention is defined by the accompanying claims.

Claims

1. A mechanical CPR device, comprising:

a piston (221) having a suction cup (222) attached at an end thereof;
a driving component configured to, in use, extend the piston toward a patient's torso and retract the piston away from the patient's torso; and
a controller configured to determine a reference

position (230) by controlling the driving component to at least:

extend the piston until a first position at which the suction cup comes into contact with the patient's torso,
further extend the piston to cause air to be forced out from an area between the suction cup and the patient's torso until a first threshold is reached; and
retract the piston until the suction cup is at the first position, further retract the piston from the first position until a second threshold is exceeded, and

extend the piston to a point at which the second threshold is no longer exceeded, the reference position being based at least in part on the point; wherein the controller is further configured to perform mechanical CPR by controlling the driving component to at least:

compress the patient's torso by extending the piston from the reference position to a depth and retracting the piston from the depth to the reference position; and
actively decompress the patient's torso by retracting the piston from the reference position to a height above the reference position, whereby the patient's torso is decompressed above the torso's natural resting position and above the reference position; and wherein the reference position is the position from which the depth of CPR compressions and the height of decompressions are measured.

2. The mechanical CPR device of claim 1, wherein the controller is configured to compress the patient's torso and actively decompress the patient's torso in a cycle based on a frequency and a duty cycle; and wherein the frequency is at least one of a predetermined frequency or a frequency entered into the mechanical CPR device, and the duty cycle is at least one of a predetermined duty cycle or a duty cycle entered into the mechanical CPR device.

3. The mechanical CPR device of claim 1, wherein the depth is at least one of a predetermined depth, a depth entered into the mechanical CPR device, or a depth based on a force used to compress the patient's torso.

4. The mechanical CPR device of claim 1, wherein the height is at least one of a predetermined height, a height entered into the mechanical CPR device, or a height based on a force used to actively decompress the patient's torso.

5. The mechanical CPR device of claim 1, wherein the first threshold is a force threshold, the mechanical CPR device further comprising:

a force sensor configured to sense the force applied by the piston to cause air to be forced out from an area between the suction cup and the patient's torso.

6. The mechanical CPR device of claim 1, wherein the second threshold is a force threshold, the mechanical CPR device further comprising:

a sensor configured to signal when the piston has been extended to exceed the second threshold, wherein the sensor is further configured to stop signaling when the piston has been extended to the point at which the second threshold is no longer exceeded.

7. The mechanical CPR device of claim 1, wherein one or both of the first and second thresholds is a pressure threshold, the mechanical CPR device further comprising:

a pressure sensor configured to sense pressure in the area between the suction cup and the patient's torso.

8. The mechanical CPR device of claim 1, wherein the controller is configured to determine the reference position in response to the mechanical CPR device receiving a user input.

9. The mechanical CPR device of claim 1, wherein the controller is configured to determine the reference position a predetermined number of times before performing mechanical CPR.

Patentansprüche

1. Mechanische CPR-Vorrichtung mit einem Kolben (221) mit einer Saugkappe (222), die an einem Ende davon angebracht ist; einer Antriebskomponente, die eingerichtet ist, um bei der Verwendung den Kolben in Richtung des Torsos eines Patienten auszufahren bzw. zu bewegen und den Kolben weg von dem Torso des Patienten zurückzufahren bzw. zu bewegen; und einer Steuerung, die eingerichtet ist, um eine Referenzposition durch Steuerung der Antriebskomponente zu bestimmen, um wenigstens:

den Kolben auszufahren bis zu einer ersten Position, in der die Saugkappe in Kontakt mit dem Torso des Patienten kommt, den Kolben weiter auszufahren, um zu bewir-

ken, dass Luft aus einem Bereich zwischen der Saugkappe und dem Torso des Patienten herausgedrückt wird, bis eine erste Schwelle erreicht wird; und

den Kolben zurückzufahren, bis die Saugkappe in der ersten Position ist, den Kolben weiter zurückzufahren von der ersten Position, bis eine zweite Schwelle überschritten ist, und

den Kolben zu einem Punkt auszufahren, an dem die zweite Schwelle nicht länger überschritten ist, wobei die Referenzposition wenigstens teilweise auf dem Punkt basiert;

wobei die Steuereinrichtung ferner eingerichtet ist, um eine mechanische CPR durch Steuerung der Antriebskomponente durchzuführen, um wenigstens:

den Torso des Patienten zu komprimieren durch Ausfahren des Kolbens von der Referenzposition zu einer Tiefe, und den Kolben von der Tiefe zurück zu der Referenzposition zu verfahren; und

aktiv den Torso des Patienten zu dekomprimieren durch Zurückfahren des Kolbens von der Referenzposition zu einer Höhe oberhalb der Referenzposition, wodurch der Torso des Patienten oberhalb der natürlichen Ruheposition des Torsos und oberhalb der Referenzposition dekomprimiert wird, und wobei die Referenzposition eine Position ist, von welcher aus die Tiefe von CPR-Kompressionen und die Höhe von Dekompressionen gemessen wird.

2. Mechanische CPR-Vorrichtung nach Anspruch 1, wobei die Steuereinrichtung eingerichtet ist, um den Torso des Patienten zu komprimieren und den Torso des Patienten aktiv zu dekomprimieren in einem Zyklus basierend auf einer Frequenz und einem Arbeitszyklus bzw. einer Einschaltdauer, und wobei die Frequenz eine vorbestimmte Frequenz und/oder eine in die mechanische CPR-Vorrichtung eingegebene Frequenz ist, und der Arbeitszyklus ein vorbestimmter Arbeitszyklus und/oder ein in die mechanische CPR-Vorrichtung eingegebener Arbeitszyklus ist.

3. Mechanische CPR-Vorrichtung nach Anspruch 1, wobei die Tiefe eine vorbestimmte Tiefe und/oder eine Tiefe, die in die mechanische CPR-Vorrichtung eingegeben wird, und/oder eine Tiefe basierend auf einer Kraft, die zum komprimieren des Torsos des Patienten verwendet wird, ist.

4. Mechanische CPR-Vorrichtung nach Anspruch 1, wobei die Höhe ein vorbestimmte Höhe und/oder eine Höhe, die in die mechanische CPR-Vorrichtung

eingegeben wird, und/oder eine Höhe basierend auf einer Kraft, die zum aktiven Dekomprimieren des Torsos des Patienten verwendet wird, ist.

5. Mechanische CPR-Vorrichtung nach Anspruch 1, wobei die erste Schwelle eine Kraftschwelle ist, wobei die mechanische CPR-Vorrichtung ferner aufweist:

einen Kraftsensor, der eingerichtet ist zur Erfassung der Kraft, die durch den Kolben aufgebracht wird, um Luft aus dem Bereich zwischen der Saugkappe und dem Torso des Patienten zu drängen.

6. Mechanische CPR-Vorrichtung nach Anspruch 1, wobei die zweite Schwelle eine Kraftschwelle ist und die mechanische CPR-Vorrichtung ferner aufweist:

einen Sensor, der eingerichtet ist, um zu signalisieren, wenn der Kolben ausgefahren ist, und die zweite Schwelle überschreitet, wobei der Sensor ferner eingerichtet ist, um die Signalisierung zu stoppen, wenn der Kolben zu dem Punkt ausgefahren ist, an dem die zweite Schwelle nicht länger überschritten ist.

7. Mechanische CPR-Vorrichtung nach Anspruch 1, wobei eine oder beide der ersten und zweiten Schwellen eine Druckschwelle ist bzw. sind, und die mechanische CPR-Vorrichtung ferner aufweist:

einen Drucksensor, der eingerichtet ist, um einen Druck in dem Bereich zwischen der Saugkappe und dem Torso des Patienten zu erfassen.

8. Mechanische CPR-Vorrichtung nach Anspruch 1, bei der die Steuerung eingerichtet ist, um die Referenzposition in Reaktion darauf, dass die mechanische CPR-Vorrichtung eine Benutzereingabe erhält, zu bestimmen.

9. Mechanische CPR-Vorrichtung nach Anspruch 1, bei der die Steuerung eingerichtet ist, um die Referenzposition eine vorbestimmte Anzahl von Malen zu bestimmen, bevor eine mechanische CPR durchgeführt wird.

Revendications

1. Dispositif de réanimation cardiopulmonaire (CPR) mécanique, comportant :

un piston (221) ayant une ventouse (222) fixée à une extrémité de celui-ci ;
un composant d'entraînement configuré pour,

en utilisation, faire sortir le piston vers le torse d'un patient et faire rentrer le piston en l'éloignant du torse du patient ; et
un contrôleur configuré pour déterminer une position de référence (230) en commandant le composant d'entraînement pour au moins :

faire sortir le piston jusqu'à une première position dans laquelle la ventouse vient en contact avec le torse du patient,
faire davantage sortir le piston pour faire sortir l'air d'une zone entre la ventouse et le torse du patient jusqu'à ce qu'un premier seuil soit atteint ; et
faire rentrer le piston jusqu'à ce que la ventouse soit dans la première position,

faire davantage rentrer le piston de la première position jusqu'à ce qu'un second seuil soit dépassé, et
faire sortir le piston jusqu'à un point où le second seuil n'est plus dépassé, la position de référence étant basée au moins en partie sur ce point ;
dans lequel le contrôleur est en outre configuré pour effectuer une CPR mécanique en commandant le composant d'entraînement pour au moins :

compresser le torse du patient en faisant sortir le piston à partir de la position de référence jusqu'à une certaine profondeur et en faisant rentrer le piston à partir de la profondeur jusqu'à la position de référence ; et
décompresser activement le torse du patient en faisant rentrer le piston depuis la position de référence jusqu'à une hauteur au-dessus de la position de référence, moyennant quoi le torse du patient est décompressé au-dessus de la position de repos naturelle du torse et au-dessus de la position de référence ; et dans lequel la position de référence est la position à partir de laquelle la profondeur des compressions de PCR et la hauteur des décompressions sont mesurées.

2. Dispositif de PCR mécanique selon la revendication 1, dans lequel le contrôleur est configuré pour compresser le torse du patient et décompresser activement le torse du patient dans un cycle basé sur une fréquence et un cycle opératoire ; et dans lequel la fréquence est au moins un élément parmi une fréquence prédéterminée et une fréquence entrée dans le dispositif de PCR mécanique, et le cycle opératoire est au moins un élément parmi un cycle opératoire prédéterminé ou un cycle opératoire entré dans le dispositif de PCR mécanique.

3. Dispositif de PCR mécanique selon la revendication 1, dans lequel la profondeur est au moins un élément parmi une profondeur prédéterminée, une profondeur entrée dans le dispositif de PCR mécanique, ou une profondeur basée sur une force utilisée pour comprimer le torse du patient. 5

4. Dispositif de PCR mécanique selon la revendication 1, dans lequel la hauteur est au moins un élément parmi une hauteur prédéterminée, une hauteur entrée dans le dispositif de PCR mécanique, ou une hauteur basée sur une force utilisée pour décompresser activement le torse du patient. 10

5. Dispositif de PCR mécanique selon la revendication 1, dans lequel le premier seuil est un seuil de force, le dispositif de PCR mécanique comportant en outre : 15
 - un capteur de force configuré pour détecter la force appliquée par le piston pour faire sortir l'air d'une zone entre la ventouse et le torse du patient. 20

6. Dispositif de PCR mécanique selon la revendication 1, dans lequel le second seuil est un seuil de force, le dispositif de PCR mécanique comportant en outre : 25
 - un capteur configuré pour une signalisation lorsque le piston a été sorti pour dépasser le second seuil, dans lequel le capteur est en outre configuré pour arrêter la signalisation lorsque le piston a été sorti jusqu'au point où le second seuil n'est plus dépassé. 30 35

7. Dispositif de PCR mécanique selon la revendication 1, dans lequel un seuil parmi les premier et second seuils, ou les deux, est un seuil de pression, le dispositif de PCR mécanique comportant en outre : 40
 - un capteur de pression configuré pour détecter une pression dans la zone entre la ventouse et le torse du patient. 45

8. Dispositif de PCR mécanique selon la revendication 1, dans lequel le contrôleur est configuré pour déterminer la position de référence en réponse au dispositif de PCR mécanique recevant une entrée utilisateur. 50

9. Dispositif de PCR mécanique selon la revendication 1, dans lequel le contrôleur est configuré pour déterminer la position de référence un nombre de fois prédéterminé avant d'effectuer une PCR mécanique. 55




FIGURE 1A




FIGURE 1B




FIGURE 2A




FIGURE 2B




FIGURE 2C




FIGURE 2D




FIGURE 2E




FIGURE 2F




FIGURE 3




FIGURE 4A





FIGURE 4B



V
FIGURE 4C




FIGURE 4D




FIGURE 4E




FIGURE 5




FIGURE 6A




FIGURE 6B




FIGURE 6C

REFERENCES CITED IN THE DESCRIPTION

This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.

Patent documents cited in the description

- WO 2009136831 A [0002]