



US006679625B2

(12) **United States Patent**
Cordes et al.

(10) **Patent No.:** **US 6,679,625 B2**
(45) **Date of Patent:** **Jan. 20, 2004**

(54) **SCANNING HEAT FLOW PROBE**

(75) Inventors: **Steven Alan Cordes**, Yorktown Heights, NY (US); **David R. DiMilia**, Wappingers Falls, NY (US); **James Patrick Doyle**, Bronx, NY (US); **Matthew James Farinelli**, Bronx, NY (US); **Snigdha Ghoshal**, Austin, TX (US); **Uttam Shyamalindu Ghoshal**, Austin, TX (US); **Chandler Todd McDowell**, Austin, TX (US); **Li Shi**, Austin, TX (US)

(73) Assignee: **International Business Machines Corporation**, Armonk, NY (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 7 days.

(21) Appl. No.: **10/022,162**

(22) Filed: **Dec. 17, 2001**

(65) **Prior Publication Data**

US 2003/0112844 A1 Jun. 19, 2003

(51) **Int. Cl.**⁷ **G01K 17/00**; G01K 3/08

(52) **U.S. Cl.** **374/29**; 374/113; 374/57;
374/43

(58) **Field of Search** 374/179, 43, 29,
374/113, 45, 46, 47, 48, 49, 50, 51, 52

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,830,515 A * 5/1989 Cortes 374/208

4,960,975 A * 10/1990 Weinbrecht 219/241
5,388,323 A * 2/1995 Hopson et al. 29/595
5,838,005 A * 11/1998 Majumdar et al. 250/306
5,929,438 A * 7/1999 Suzuki et al. 250/306
5,969,238 A * 10/1999 Fischer 73/105
5,986,261 A * 11/1999 Lewis et al. 250/306
6,123,675 A * 9/2000 Kreizman et al. 600/549
6,487,515 B1 * 11/2002 Ghoshal 702/136

FOREIGN PATENT DOCUMENTS

JP 2001004455 A * 1/2001 G01K/7/18

* cited by examiner

Primary Examiner—Diego Gutierrez


Assistant Examiner—Lydia M. De Jesús

(74) *Attorney, Agent, or Firm*—Duke W. Yee; Casimer K. Salys; Patrick C. R. Holmes

(57) **ABSTRACT**

A scanning heat flow probe for making quantitative measurements of heat flow through a device under test is provided. In one embodiment the scanning heat flow probe includes an electric current conductor in a cantilever beam connected to a probe tip and coupled to two voltmeter leads. The probe also includes two thermocouple junctions in the cantilever beam electrically isolated from the electric current conductor and the two voltmeter leads. Heat flow is derived quantitatively using only voltage and current measurements. In other forms, the invention relates to the calibration of scanning heat flow probes through a method involving interconnected probes, and relates to the minimization of heat flow measurement uncertainty by probe structure design practices.

8 Claims, 11 Drawing Sheets








FIG. 3A





FIG. 3B






FIG. 4A




FIG. 5A




FIG. 4B








FIG. 5B






FIG. 4E




FIG. 4F




FIG. 5E








FIG. 5F






FIG. 4I




FIG. 4J




FIG. 5I




FIG. 5J




FIG. 4K





FIG. 4L




FIG. 5L




FIG. 4M




FIG. 5M




FIG. 4N

FIG. 5N





FIG. 6

SCANNING HEAT FLOW PROBE

BACKGROUND OF THE INVENTION

1. Technical Field

The present invention is directed to an apparatus for measuring heat flow and methods related to the fabrication and calibration of such apparatus.

2. Description of Related Art

One of the major difficulties in developing novel thin film thermoelectric materials lies in obtaining consistent and accurate measurement of their thermal and electrical properties. Traditional methods cannot be easily extended to microscopic characterization because of increased electrical and thermal parasitic losses associated with the probes used to perform the measurements. Additionally, the poor structural stability of some of the novel materials being investigated makes using traditional probe methods unworkable.

For example, in the case of measurements using a probe, such as the "ZT-meter," the time-scales of the transients become short and introduce errors in the electrical measurements. Scanning thermoelectric microscopy (STEM) based on atomic force microscope (AFM) probes are capable of performing measurements of thermal and electrical properties of thermoelectric materials at these small scales. However, STEM based on AFM probes still have several limitations. For example, present probes only give qualitative measurements of heat flow, which only allows one to determine whether there is more or less heat flow with one material versus another. Therefore, it would be desirable to have a scanning heat flow probe that allows quantitative measurements of heat flow to be made.

SUMMARY OF THE INVENTION

The present invention provides a scanning heat flow probe for making quantitative measurements of heat flow through a device under test. In one embodiment the scanning heat flow probe includes an electric current conductor in a cantilever beam connected to a probe tip and coupled to two voltmeter leads. The probe also includes two thermocouple junctions in the cantilever beam electrically isolated from the electric current conductor and the two voltmeter leads. Heat flow is derived quantitatively using only voltage and current measurements. In other forms, the invention relates to the calibration of scanning heat flow probes through a method involving interconnected probes, and relates to the minimization of heat flow measurement uncertainty by probe structure design practices.

BRIEF DESCRIPTION OF THE DRAWINGS

The novel features believed characteristic of the invention are set forth in the appended claims. The invention itself, however, as well as a preferred mode of use, further objectives and advantages thereof, will best be understood by reference to the following detailed description of an illustrative embodiment when read in conjunction with the accompanying drawings, wherein:

FIG. 1 is an exemplary cross-sectional view of the scanning thermoelectric microscopy probe in accordance with the present invention;

FIG. 1A depicts a schematic diagram of an exploded view of the components of the probe cantilever in accordance with the present invention;

FIG. 2 depicts a schematic diagram illustrating heat flow through a probe to a sample in accordance with the present invention;

FIG. 2A depicts a schematic diagram illustrating heat flow through the probe;

FIGS. 3A–3B depict schematic diagrams of the calibration procedure in accordance with the present invention;

FIGS. 4A–4N are exemplary cross sections illustrating a process of fabricating the scanning heat flow probe in accordance with the present invention;

FIGS. 5A–5N are exemplary top views illustrating a process of fabricating the scanning heat flow probe in accordance with the present invention; and

FIG. 6 depicts an example of a graph of fractional heat flow uncertainty versus probe length in accordance with the present invention.

DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENT

The present invention provides a method and apparatus for measuring and characterizing the thermal and electrical properties of the materials. The invention as embodied herein makes use of temperature and voltage sensors with a thermal probe to quantitatively measure heat flow through a tip of the probe. Also included is an electric conductor connected to the probe tip to allow for electrical stimulation of the sample.

Referring now to FIG. 1, an diagram illustrating an exemplary scanning heat flow probe **100** is depicted in accordance with the present invention. The probe shown in FIG. 1 is used to measure the thermal properties of materials in a manner described in detail hereafter.

The probe **100** includes a probe body **150**, a cantilever structure **110**, a first temperature sensing lead **106**, a second temperature sensing lead **108**, a first voltage sensing lead **101**, a second voltage sensing lead **112**, a current lead **160**, a thermistor **118**, and a probe tip **104**. A heater/cooler **116**, such as a thermoelectric heater/cooler (TEC), may also be thermally coupled to probe **100** in order to control the background temperature of the probe **100**. The leads **106**, **108**, and **114** create two thermocouples at points **102** and **122** in cantilever **110** which are used to calibrate probe **100** in a manner to be described hereafter, to permit quantitative measurement of the heat flow through the probe tip **104** into sample **202** as in FIG. 2.

Referring to FIG. 1A, electrical current flows through probe **100** via current conductor **160**. Voltage readings are taken at points **102** and **122** of current conductor **160** through leads **101** and **112**. The current conductor **160** is separated from the thermocouple leads **106**, **108**, and **114** by an electrical insulator **190**, such as, for example, silicon nitride. The distance between the proximal thermocouple at **122** and the distal thermocouple at **102** is typically between 100 and 900 microns. The distal thermocouple at **102** measures the temperature near the tip **104** using leads **108** and **114**. A differential thermocouple formed from the proximal thermocouple at **122**, and the distal thermocouple at **102** using leads **108** and **106**, measures the temperature drop along the beam. In one embodiment, cantilever **110** has a width of approximately 8 microns. In one embodiment, the probe tip **104** may be constructed from tungsten. A TEC heater/cooler **116** may be thermally coupled to the scanning heat flow probe **100** to bias the temperature at **122**. By varying the temperature at **122**, the heat flow through the probe can be adjusted to a desired value, including zero.

In some preferred embodiments, the probe includes a radiation shield surrounding but not contacting the probe cantilever **110** from points **122** to **102** to mitigate radiation

effects. Also, the probe and sample should be used in vacuum to mitigate gaseous convection/conduction effects.

While the probe structure shown in FIG. 1 shows a sharp probe tip, the probe tip may be of any shape desirable. For example, the cone-shaped probe tip may be very narrow or very wide in diameter, may have any value interior angle at the tip, and the like.

Calibration

In order to measure heat flow accurately, the probe **100** must be calibrated. In designing and calibrating the probe **100**, one goal is the minimization of error in the probe in order for the probe to provide accurate measurements of heat flow in a sample. In order to do this, the error in the heat flow, Q , through the probe should be minimized as much as possible.

As depicted in FIG. 2A, the heat flow through the probe cantilever **110** may be characterized by two components: a component, Q_j , due to joule heating from current flowing through the electrical conductor **160** and a component, Q_m , due to the temperature gradient across the probe cantilever **110** between points **122** and **102**.

The joule heating component, Q_j , may be measured from the voltage across the current conductor **160** with leads **101** and **112**, and a measurement of the current through the current conductor **160**. For purposes of calibration, it is accepted that one half of the heat for a uniform structure, Q_j , generated by joule heating flows toward the distal end of the probe **100** and one half flows toward the proximal end of the probe **100**.

Thus, heat flow, Q , through the probe **100** into the sample is given by the following relation:

$$Q = Q_m + \frac{1}{2}Q_j$$

Since Q_m is related to the temperature gradient across the probe, ΔT , and to the thermal resistivity, R_{th} , by the following equation:

$$Q_m = \Delta T / R_{th}$$

and since the heat flow due to joule heating is given by the relationships described above, the heat flow Q is given by the following relationship:

$$Q = \Delta T / R_{th} + \frac{1}{2}IV_e$$

where V_e is the voltage across the current conductor **160** between points **122** and **102** in the cantilever **110**. The temperature across gradient may be expressed as V_{th}/α where V_{th} is the difference between the voltage across the distal thermocouple (V_{TCd}) and the voltage across the proximal thermocouple (V_{TCp}) and α is the Seebeck coefficient across the junction of the differential thermocouple. Thus, Q may be represented by the following relationship:

$$Q = V_{th}/(\alpha R_{th}) + \frac{1}{2}IV_e$$

Therefore, the uncertainty, σQ , in the heat flow may be expressed as:

$$(\sigma Q)^2 = Q_m^2 [(\sigma V_{th}/V_{th})^2 + (\sigma \alpha R_{th}/\alpha R_{th})^2] + (\frac{1}{2}Q_j)^2 [(\sigma V_e/V_e)^2 + (\sigma I/I)^2]$$

This introduces four error terms: the error in the thermocouple voltage measurement, σV_{th} , the error in the calibration of αR_{th} , $\sigma \alpha R_{th}$, the error in the current conductor voltage measurement σV_e , and the error in the measurement of the current through the current conductor, σI . However, three of these terms are determinable from the tolerances provided with the commercially available instruments uti-

lized to make the voltage and current measurements. This leaves only the uncertainty in αR_{th} to be determined.

In order to determine the uncertainty in αR_{th} , the calibration method of the present invention utilizes two scanning heat flow probes as depicted schematically in FIGS. 3A–3B. In FIG. 3A, two scanning heat flow probes **301** and **302** are oriented such that their tips touch. The probes are shown oriented linearly in FIGS. 3A–3B, however, the particular orientation between the two probes is not important. In step **1**, a temperature gradient ΔT is created across the two probes by, for example, heating one probe with a heater. The temperature gradient induces a heat flow, Q , from one probe **301** to the other probe **302** as depicted in FIG. 3A. The voltages V_{tha1} and V_{thb1} of the two differential thermocouples in probes **301** and **302** are measured. Since the heat flow through the two probes is equal:

$$\alpha R_{tha}/\alpha R_{thb} = V_{tha1}/V_{thb1}$$

The second step is depicted in FIG. 3B. In FIG. 3B, the same two probes are used though they need not be in the exact same configuration. A current, I , is passed from probe **301** into probe **302** with voltage measurements made at points **310–313** using voltage leads **101** and **112** of each probe **301**, **302**. It is assumed that one half of the heat produced by joule heating in probe **301** passes into probe **302** and that one half of the heat produced by joule heating in probe **302** passes into probe **301**. There is also joule heating produced by the current I passing from probe **301** to probe **302** in the region between points **311** and **312**. However, since this section includes portions of both probes, no assumptions are made about how much of this heat passes through probe **301** and how much passes through probe **302**. Therefore, a fraction, fQ_c , of the heat, Q_c , produced by joule heating in region **c** passes into probe **302** and the remaining fraction of Q_c (i.e. $(1-f)Q_c$) passes into probe **301**. Furthermore, since the current flowing through the two probes is equal, the following relation may be determined:

$$V_{thb2}/\alpha R_{thb} + (-V_{tha2}/\alpha R_{tha}) = Q_c + \frac{1}{2}Q_a + \frac{1}{2}Q_b$$

Therefore, since there are now two equations, αR_{th} can be isolated for one of the probes. Thus, the following equation may be obtained for probe **a**:

$$\alpha R_{tha} = ((V_{tha1}/V_{thb1})V_{thb2} - V_{tha2}) / (I(V_e + \frac{1}{2}V_a + \frac{1}{2}V_b))$$

Thus, the uncertainty in αR_{tha} (i.e. $\sigma \alpha R_{tha}$) depends only upon measurable quantities (e.g., I and V). Thus, the uncertainty in the heat flow through the probe depends entirely on measurable quantities.

Probe Formation Process

FIGS. 4A–4N are exemplary cross sections illustrating a process of fabricating the scanning heat flow probe **100** and FIGS. 5A–5N are exemplary corresponding top views in the process of fabricating the scanning heat flow probe **100**. The scanning heat flow probe **100** is comprised of a number of different layers of material. The particular materials described hereafter with reference to the exemplary embodiment are meant to be for illustrative purposes and other materials having similar properties may be used in replacement or in addition to the materials described herein without departing from the spirit and scope of the present invention.

The formation of the scanning heat flow probe **100** will now be described with reference to FIGS. 4A–4N and 5A–5N. The mechanisms used to create the various layers of the probe, such as deposition and etching, are generally known in the art of semiconductor chip manufacture. However, these mechanisms have not previously been used to create the structure herein described.

The process starts with a silicon wafer **402** sandwiched between two polished silicon nitride layers **404** and **406**. The top-side silicon nitride layer **404** which will in part form beam **191** (FIG. 1A) is selectively etched, for example by a reactive ion etch (RIE), to remove a portion of the silicon nitride as depicted in FIG. 4B for front-back alignment marks **460**. For ease of illustration, the alignment marks **460** are not depicted in the next few sets of diagrams.

Next, using two separate masks, nickel **410** and chromium **408** leads are deposited onto silicon nitride **404** as depicted in FIGS. 4C and 5C. Preferably, the chromium leads **408** are approximately 60 nm thick and the nickel **410** leads are approximately 80 nm thick over 3 nm of chromium. A layer of silicon nitride **412** is then deposited over the chromium **408** and nickel **410** layers as well as over the remaining original silicon nitride layer **404** as depicted in FIGS. 4D and 5D by, for example, a PECVD process.

The resulting structure is then masked and a reactive ion etch is used to pattern the silicon nitride cantilever as depicted in FIGS. 4E and 5E. Next a tungsten (W) layer **414** is sputtered onto the exposed surface of the structure, preferably to a thickness of 2 microns resulting in a structure as depicted in FIGS. 4F and 5F. Next a layer of thick photoresist **416**, such as, for example, AZ or Shipley, is formed over the structure by spinning resulting in the structure as depicted in FIGS. 4G and 5G. Preferably the photoresist **416** is spun to a thickness of 3 microns. The photoresist **416** is then patterned into the tip shape as depicted in FIGS. 4H and 5H.

Next, a wet etch is performed on the W layer **414** undercutting the photoresist pattern **416** resulting in a structure as depicted in FIGS. 4I and 5I. Next, the photoresist **416** is stripped resulting in a structure as depicted in FIGS. 4J and 5J. The structure is then masked and the PECVD silicon nitride **412** is etched over the nickel and chromium contacts **418**, **420** resulting in the structure depicted in FIGS. 4K and 5K. The structure is then masked again and the gold contacts **422** and **424** are formed to the tungsten, nickel, chromium and resistor lines **426-428** as depicted in FIGS. 4L and 5L. Next, another mask is utilized to pattern the backside silicon nitride **406**. RIE results in the structure as depicted in FIGS. 4M and 5M. After a final bulk silicon etch in KOH, the final resulting structure is formed as depicted in FIGS. 4N and 5N.

The probe created using the process described above can be used for making measurements in many different applications. The probe may be used to measure thermoelectric properties of nano-scale structures, profiling of silicon dopants of semiconductor materials, characterizing giant magneto-resistive heads, and the like. The present invention is not limited to any one application of the probe and is intended to cover all possible applications to which the probe may be made. Those of ordinary skill in the art will appreciate that the probe of the present invention is preferably utilized along with a computing system in which the calibration and computations described above and hereafter are performed. The probe is used to provide measured quantities which are then processed by the computing system to calibrate the probe and generate values for the heat flow properties of the materials under test.

Method for Optimizing Design of Probe

In designing the probe, a method for determining a substantially optimum design for the probe in accordance with the present invention may be used. This method determines, for example, the optimum length of the probe cantilever, given the operational parameters under which the probe is intended to be used as well as the values of other

parameters already chosen. This method of the current invention is a method for exploring the design space of a scanning heat flow probe and selecting the design parameters that minimize the errors in probe measurement, specifically the uncertainty in heat flow through the probe.

Method overview:

An equation describing the heat, Q, flowing through the probe is formed. This equation is based upon

1 Design parameters (values, no uncertainties)

o1 material properties/composition

§1 electrical resistance

§2 thermal conductivity

§3 Seebeck coefficient

o2 Geometries

§1 Beam width/length

§2 Metal width/thickness

§3 Thermocouple width/thickness

o3 Configuration

§1 Implied in the thermal and electrical networks

2 Measurement environment (values, no uncertainties)

o1 Heat flow, Q, through the probe

o2 Current through the probe

3 Calibration conditions (values, no uncertainty)

o1 Two step/two probe method

o2 Calibration temperature difference

o3 Calibration current

o4 Contact resistance

4 Measurement equipment (uncertainties)

o1 Manufacturer stated uncertainty in the measurement for the given operational point

All the stated values, while not exact to the usage condition, are to place the probe in a certain operational range. The sole source of uncertainty is from the measurement equipment. Variation in the operational conditions of the probe will have an effect on the optimal probe design parameters, but this is a second order effect. Large deviations in operating conditions will substantially change the optimal design parameters and lead to degraded performance. This leads to the use of different probes for different operational conditions, just as a multi-meter changes ranges to improve performance.

Any probe will not be used in the exact operational range that determined the probes design parameters, but this is not cause for the performance of the probe to become suspect. When an actual measurement is made, the instrument measurements are entered into the heat flow equation yielding a heat flow based on measured values, no design parameters or constants. The calibration parameters are derived from measured values so they do not invalidate the previous statement. The uncertainty of the heat flow is also based on measured values and the manufacturers stated uncertainties of the measurement equipment.

So probe design and use come in two phases. In the design phase every parameter is chosen or derived from chosen values. The design parameters are varied, namely length, and the design parameters are chosen such that error in the measurement of the heat flow is minimized. In the second phase all design parameters are forgotten, and the heat flow and uncertainty are based on measured values and the manufacturers stated uncertainty for the equipment.

Equations

The following list of equations may be simplified to a single equation expressing the uncertainty on the measurement of the heat flow through the probe. Any parameter can be varied, though length of the probe is most common, and

a curve of uncertainty vs. that parameter can be formed. From this curve an optimal range for that design parameter can be chosen. An example of a graph of length of probe versus uncertainty in accordance with the present invention is depicted in FIG. 6. Good results for probe performance may be obtained using a cantilever beam having a length falling in region B. The increase in uncertainty for lengths smaller than this in region A is due to instrumentation uncertainty since such lengths result in smaller temperature gradients thereby becoming almost unmeasurable by the instrumentation. The increase in uncertainty for lengths in region C result from increase joule heating because of the increased length of the cantilever beam.

While this method is not sophisticated like a global error minimization it turns out that most of the parameters should be maximized or minimized individually and there is no system level trade off. For the length of the beam there is a system level trade off.

A longer beam produces a large temperature drop for a given heat flow making the thermocouple voltage easier to measure. This is true up to a point. Over some range the voltmeter has a fairly flat % error versus voltage. Increasing the voltage no longer decrease the uncertainty in heat flow.

A longer beam also has more resistance in the current conductor. This causes more heat to be generated. When the heat generated in the current conductor is greater than the heat that should be delivered to the object under test, the probe must have a counter heat flow (the measured heat flow) that reduces the heat flow to the object. This subtraction of two large values causes a large uncertainty on the result. It is for this reason that the beam length has an optimal value for a given operation condition.

In the equations that follow, a couple functions are used. The function Xspec gives the uncertainty in a measurement for a given instrument and measured value. The function mat_table gives the requested material property for a given substance.

L = 900e-6 [m]	distance between 122 and 102
I = 100e-6 [A]	current through 160
sI = Xspec(I, current source);	uncertainty in the current above
Q = 2e-6 [W]	Heat flow through the probe
a = 30e-6 [V/K]	Seebeck coefficient of the thermocouple material pair
SiNx_width = 8e-6 [m]	beam 191 width
SiNx_thick = 1.9e-6 [m]	beam 191 thickness
ni_width = 1e-6 [m]	Nickel lead 114 of thermocouple
ni_thick = 80e-9 [m]	
cr_width = 1e-6 [m]	Chromium leads 106 & 108 of thermocouple
cr_thick = 60e-9 [m]	
I_width = 2e-6 [m]	Current conductor 160
I_thick = 200e-9 [m]	
V_width = 1e-6 [m]	Voltage sense wires 110 & 112
V_thick = I_thick;	
pho_e = mat_table('W', 'electrical_resistivity');	Current conductor
Re = pho_e.*L./(V_width.*V_thick);	
Ve = I.*Re;	
sVe = Xspec(Ve, voltmeter);	
pho_th = mat_table('W', 'thermal_conductivity');	
Rth_SiNx =	
L./(mat_table('SiNx', 'thermal_conductivity').*SiNx_width.*SiNx_thick);	
Rth_tc = L./(. . .	

-continued

```

mat_table('Ni', 'thermal_conductivity').*ni_width.*ni_
thick + . . .
5 mat_table('Cr', 'thermal_conductivity').*cr_width.*cr_
thick);
Rth_Vsense = L./(pho_th.*V_width.*V_thick);
Rth_Wire = L./(pho_th.*I_width.*I_thick);
Rth = L./(. . .
10 mat_table('SiNx', 'thermal_conductivity').*SiNx_width.
.*SiNx_thick + . . .
mat_table('Ni', 'thermal_conductivity').*ni_width.*ni_
thick + . . .
mat_table('Cr', 'thermal_conductivity').*cr_width.*cr_
thick + . . .
pho_th.*V_width.*V_thick + . . .
15 pho_th.*I_width.*I_thick);
Rthp = 1./(1./Rth_SiNx + 1./Rth_tc + 1./Rth_Vsense +
1./Rth_Wire);
Vth = a.*(Q-1/2*I.^2.*Re).*Rth;
sVth = Xspec(Vth, voltmeter);
% Calibration portion
aRth = a.*Rth;
20 Ical = 100e-6*ones(n_X, n_Y);
sIcal = Xspec(Ical, current source);
dV1 = Re.*Ical;
sdV1 = Xspec(dV1, voltmeter);
dV2 = Re.*Ical;
sdV2 = Xspec(dV2, voltmeter);
25 Rcontact = 100 [Ohms]
dVc = Ical.*Rcontact;
sdVc = Xspec(dVc, voltmeter);
sW = W.*sqrt(. . .
(sI./I).^2 + (sdVc./dVc).^2);
30 Qcal = Ical.*(dVc + 0.5*(dV1 + dV2));
sQcal = Ical.*sqrt(. . .
(Qcal.*sIcal./Ical).^2 + . . .
(sdVc).^2 + . . .
(sdV1/2).^2 + . . .
(sdV2/2).^2);
QdTcal = 2e-6 [W]
35 Vth1T = aRth.*QdTcal;
sVth1T = Xspec(Vth1T, voltmeter);
Vth2T = aRth.*QdTcal;
sVth2T = Xspec(Vth2T, voltmeter);
Vth1I = aRth.*Qcal/2;
sVth1I = Xspec(Vth1I, voltmeter);
40 Vth2I = aRth.*Qcal/2;
sVth2I = Xspec(Vth2I, voltmeter);
aRthp = (Vth1T.*Vth2I./Vth2T + Vth1I)./Qcal;
% all together
saRth = sqrt(. . .
(Vth2I.*sVth1T./(Qcal.*Vth2T)).^2 + . . .
(Vth1T.*Vth2I.*sVth2T./(Qcal.*Vth2T).^2).^2 +
45 . . .
(Vth1T.*sVth2I./(Qcal.*Vth2T)).^2 + . . .
(sVth1I./Qcal).^2 + . . .
((Vth1T.*Vth2I./Vth2T +
Vth1I).*sQcal./Qcal.^2).^2);
% End Calibration portion
50 Qj = I.^2.*Re;
Qm = Q - 1/2*Qj;
sQ = sqrt(. . .
(Qm).^2.*(sVth./Vth).^2 +
(saRth./aRth).^2 + . . .
(1/2*Qj).^2.*(sVe./Ve).^2 + (sI./I).^2);
55

```

The description of the present invention has been presented for purposes of illustration and description, and is not intended to be exhaustive or limited to the invention in the form disclosed. Many modifications and variations will be apparent to those of ordinary skill in the art. The embodiment was chosen and described in order to best explain the principles of the invention, the practical application, and to enable others of ordinary skill in the art to understand the invention for various embodiments with various modifications as are suited to the particular use contemplated.

What is claimed is:

- 1. A scanning heat flow probe, comprising:
 an electric current conductor in a cantilever structure
 connecting a probe tip at one end to a probe body at an
 opposite end;
 first and second voltage sense leads coupled to the electric
 conductor at first and second sense points; and
 two thermocouple junctions in the cantilever structure and
 operatively positioned proximate the first and second
 sense points.
- 2. The scanning heat flow probe as recited in claim 1,
 further comprising:
 a dielectric layer situated between the electric current
 conductor and the two thermocouple junctions in the
 cantilever structure.
- 3. The scanning heat flow probe as recited in claim 2,
 wherein the dielectric layer comprises silicon nitride.
- 4. The scanning heat flow probe as recited in claim 2,
 wherein the cantilever structure, the electric current
 conductor, the voltage sense leads, the dielectric layer, and
 the two thermocouple junctions define a heat flow path to the
 probe tip.
- 5. The scanning heat flow probe as recited in claim 1,
 wherein the two thermocouple junctions are coupled to first,
 second, and third temperature responsive sense leads.
- 6. The scanning heat flow probe as recited in claim 1,
 wherein the two thermocouple junctions comprise nickel
 and chromium connections.

- 7. The scanning heat flow probe as recited in claim 1,
 wherein the first and second voltage sense lead couplings
 provide a measure of voltage drop along the electrical
 current conductor between first and second sense points.
- 8. A scanning heat flow probe, comprising:
 a probe body;
 a probe tip;
 a cantilever structure connecting the probe body to the
 probe tip;
 a electric current lead in the cantilever structure coupled
 to the probe tip and to the probe body with voltage
 measurement leads coupled to the electric current lead
 at a proximal point and a distal point;
 a first thermocouple junction positioned in the cantilever
 at the proximal point and electrically isolated from the
 electric current lead and the voltage measurement
 leads; and
 a second thermocouple junction positioned in the canti-
 lever at the distal point and electrically isolated from
 the electric current lead and the voltage measurement
 leads;
 wherein the thermocouple junction responses at the proxi-
 mate and distal points, voltages measured at the proxi-
 mal and distal points, and current measured through the
 electric current lead are related to the heat flow prop-
 erties of the cantilever structure.

* * * * *