



(51) International Patent Classification:

G01N 21/31 (2006.01) G01J 3/00 (2006.01)
G01N 21/47 (2006.01) G01N 33/02 (2006.01)
G01N 21/49 (2006.01) G01N 33/12 (2006.01)
A61B 5/00 (2006.01) G01N 33/18 (2006.01)

(21) International Application Number:

PCT/NL2017/050086

(22) International Filing Date:

14 February 2017 (14.02.2017)

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

2016273 16 February 2016 (16.02.2016) NL
2017700 2 November 2016 (02.11.2016) NL

(71) Applicant: **STICHTING HET NEDERLANDS KANKER INSTITUUT-ANTONI VAN LEEUWEN-HOEK ZIEKENHUIS** [NL/NL]; 121, Plesmanlaan, 1066 CX AMSTERDAM (NL).

(72) Inventors: **RUERS, Theodoor Jacques Marie**; c/o 121, Plesmanlaan, 1066 CX Amsterdam (NL). **KHO, Esther**; c/o 121, Plesmanlaan, 1066 CX Amsterdam (NL). **STER-**

ENBORG, Henricus Josephus Cornelus Maria; c/o 121, Plesmanlaan, 1066 CX Amsterdam (NL).

(74) Agent: **MERTENS, H.V.**; P.O. Box 3241, 2280 GE Rijswijk (NL).

(81) Designated States (unless otherwise indicated, for every kind of national protection available): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BN, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DJ, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IR, IS, JP, KE, KG, KH, KN, KP, KR, KW, KZ, LA, LC, LK, LR, LS, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PA, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SA, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.

(84) Designated States (unless otherwise indicated, for every kind of regional protection available): ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, ST, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK,

[Continued on next page]

(54) Title: METHOD, APPARATUS AND SOFTWARE FOR DETECTION AND LOCALIZATION OF HIDDEN DEFECTS IN OPTICALLY DIFFUSE MEDIA





FIG. 3

(57) Abstract: In a method and apparatus (400), a property of an optically diffuse medium comprising a first optical absorber having a first concentration and a second optical absorber having a second concentration is determined. A surface area (406) of the medium is imaged at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber. A reflectance spectrum of the medium at the surface area at the multiple wavelengths is determined. A derivative of the determined reflectance spectrum around the isosbestic wavelength is determined. From the derivative, a concentration ratio of the first concentration and the second concentration is estimated.



SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, KM, ML, MR, NE, SN, TD, TG). **Published:**

— *with international search report (Art. 21(3))*

Method, apparatus and software for detection and localization of hidden defects in optically diffuse media

5

FIELD OF THE INVENTION

The invention relates to optical imaging of highly scattering media. In particular, this invention relates to optical imaging of optically diffuse biological materials, or media, for detecting and locating damage, disease and/or (defects in) anatomic structure. This damage, disease and (defects in) anatomic structure may be hidden, i.e. poorly visible or invisible to the human eye.

BACKGROUND OF THE INVENTION

15

Visual inspection is the oldest technique available to mankind to assess the quality of biological materials such as fruits, vegetables, meat, fish and live human tissue. Moreover, until recently, visual inspection has been the only technique available to mankind to assess the health status of said biological materials from the outer surface thereof.

20

Recently, camera based optical imaging has made its entry into the medical field. The power of optical imaging lies in the phenomenon that light penetrates up to centimeters into many biological materials. In most of these materials light is scattered strongly and becomes diffuse. In addition, very specific wavelengths of light are absorbed by specific tissue components.

25

Together, absorption and scattering determine the amount and spectral distribution of the light that is reflected. Hence, diffuse optical images do not only give information on the shape and the surface of the material imaged, but more importantly on what is below the surface.

30

Depending on the wavelength of the light, sampling depths in the order of, or up to, a centimeter, cm, are feasible.

In multispectral or hyperspectral imaging, optical images are acquired using dedicated camera systems that detect light in different wavelength bands. The images taken by these cameras are often processed by various image processing methods to obtain information on the status of biochemical composition of the imaged object. Examples of materials subjected

35

to diffuse optical imaging range from raw foods to assess their freshness, crops to assess their status, to live human tissues to detect and locate remaining cancer cells during and after surgery.

- 5 Although light scattering is the very phenomenon that enables diffuse optical imaging, it also complicates quantitatively accurate imaging, and it limits the sharpness of the images taken.

When imaging a homogenous medium, the diffuse light captured in the form of diffuse reflection gives an excellent representation of the optical properties of the tissue. If images
10 are taken at sufficient wavelengths, then there are ample mathematical models available that allow spectral unmixing, i.e. the extraction of the different concentrations of the tissue components.

These methods vary from complicated diffusion theory based spectral analyses based on a
15 lot of prior knowledge on light transport and the tissue components to purely mathematical techniques such as spectral unmixing, SU, or machine learning, ML. Often, more simple approaches are used such as ratioing the diffuse reflectance at two strategically chosen wavelengths.

20 Preferentially, however, diffuse optical imaging is used in cases where a defect is not spread over the tissue evenly, and its location is not known. In fact, the purpose of imaging usually is to detect and locate a defect. In such a case the diffuse nature of the reflected light limits the sharpness of the image, and thereby decreases the accuracy of detection of small defects.

25 In diffuse media, the sampling depth and imaging resolution are inversely related: in the case of a decreased scattering coefficient the diffuse light detected has sampled deeper into the tissue, generating an image with a lower spatial resolution, while a sample with a larger scattering coefficient samples more superficially and may have a higher spatial resolution.

30 A case of inhomogeneously distributed optical properties is not rare, since most biological materials are very inhomogeneous. A possible defect can be considered as an additional inhomogeneity. For diffuse light this has serious consequences as the diffuse light will encounter different optical properties at different locations. Because these optical properties are wavelength dependent and this wavelength dependence is different for different tissue
35 compositions, the diffuse light will have different distributions at different wavelengths.

As a consequence, different wavelengths will sample different volumes. This will compromise any spectral unmixing approach. This phenomenon is very well known in the field of diffuse optical spectroscopy.

5 In many cases, simple ratios of images taken at strategically chosen wavelengths can give excellent information on the ratio between different absorbing components. These numbers can then be used to characterize the tissue.

10 As an example, a fat-to-water ratio image has been proposed to be of particular interest, in particular for a diagnosis of breast cancer. Such a ratio would be accurate in homogenous tissue. Due to the inhomogeneous nature of the distributions of these tissue components, the ratios calculated have been documented to be in error by as much as 50%.

15 For a medium, the diffuse reflection R_d can be related to the absorption coefficient by the relation:

$$R_d \sim e^{-\mu_a(l)} \tag{1.1}$$

20 where μ_a stands for the absorption coefficient and (l) for the average path length of the detected photons.

When taking two wavelengths λ_1 and λ_2 , a ratio X can be calculated from the natural logs of the diffuse reflections:

$$25 \quad X = \frac{(I_2)\mu_a(\lambda_2)}{(I_1)\mu_a(\lambda_1)} = \frac{\ln(R_d(\lambda_2))}{\ln(R_d(\lambda_1))} \tag{1.2}$$

Assuming the presence of two absorbers at molar concentrations C_1 and C_2 , then:

$$30 \quad \mu_a(\lambda) = C_1\mu_{a,1}(\lambda) + C_2\mu_{a,2}(\lambda) \tag{1.3}$$

where $\mu_{a,m}(\lambda)$ stands for the molar absorption coefficient of absorber with index m .

Introducing the concentration ratio:

$$35 \quad \psi = \frac{C_1}{C_1 + C_2} \tag{1.4}$$

so that (1.3) can be written as:

$$\mu_a(\lambda) = (C_1 + C_2) \left(\psi \mu_{a,1}(\lambda) + (1 - \psi) \mu_{a,2}(\lambda) \right) \quad (1.5)$$

5 Now expression (1.2) can be rewritten into:

$$X = \frac{\ln(R_d(\lambda_2))}{\ln(R_d(\lambda_1))} = \frac{\langle l_1 \rangle \left(\psi \mu_{a,1}(\lambda_2) + (1 - \psi) \mu_{a,2}(\lambda_2) \right)}{\langle l_2 \rangle \left(\psi \mu_{a,1}(\lambda_1) + (1 - \psi) \mu_{a,2}(\lambda_1) \right)} \quad (1.6)$$

10 From expression (1.6) it is clear that, when the path lengths $\langle l_1 \rangle$ and $\langle l_2 \rangle$ are identical and the absorption properties of both absorbers are known, the concentration ratio ψ can be derived from the ratio of the natural logarithms of the reflectances. The strongest sensitivity is found when wavelengths λ_1 and λ_2 are chosen to be at wavelengths where the two absorbers have maximum differences. This ratio is often used in hyper or multi spectral imaging to quantify concentration ratios of chromophores.

15

The main problem in practical application of expression (1.6), lies in the assumption that $\langle l_1 \rangle = \langle l_2 \rangle$. In practice the path lengths, $\langle l_1 \rangle$ and $\langle l_2 \rangle$, are dependent on the optical properties.

20

There are three main reasons why the assumption $\langle l_1 \rangle = \langle l_2 \rangle$ is problematic in practical use.

25

A first reason is that the path length is dependent on both the absorption and the scattering properties. When aiming at a maximum sensitivity, the wavelengths λ_1 and λ_2 are chosen to be at wavelengths where the two absorbers have maximum differences. This results in very different path lengths and sampling volumes, even in homogeneous tissue.

30

A second reason is that the two tissue types with different concentrations of the two absorbers considered may be biologically very different. Hence it is likely that these tissues will also have very different scattering properties. This too causes large differences in path lengths and hence differences in sampling volumes.

35

A third reason is that the need for imaging implies inhomogeneous tissue, where different types of tissue with very different optical properties coexist. This results in very different path lengths and very different sampling volumes in the different tissue areas. Also it can generate artefacts in the boundary region between the two tissues. Detected photons show a preference for 'paths of lowest absorption': Light detected at the wavelength of maximum

absorption of a first absorber will have a stronger contribution from photons that have travelled preferentially through tissue containing a higher concentration of a second absorber, and vice versa. So, even if $\langle l_1 \rangle$ would be of the same length as $\langle l_2 \rangle$, the two paths would have sampled very different tissue volumes. This problem leads to an overestimation of the total amount of tissue components.

SUMMARY OF THE INVENTION

It would be desirable to provide a method and apparatus for optically imaging a spatial distribution of the ratio of concentrations of optical absorbers comprised in a medium. It would also be desirable to more accurately perform said optical imaging.

It would further be desirable to provide a method and device to optimize the spatial resolution in said imaging method.

15

To better address one or more of these concerns, in a first aspect of the invention a method of determining a property of an optically diffuse medium is provided. The medium comprises a first optical absorber having a first concentration and a second optical absorber having a second concentration. The method comprises the steps of:

20 (a) imaging a surface area of a volume of the medium at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber;

(b) determining a reflectance spectrum of the medium at the surface area at the multiple wavelengths;

25 (c) determining a derivative of the determined reflectance spectrum around the isosbestic wavelength; and

(c) estimating a concentration ratio of the first concentration and the second concentration from the derivative.

30 The invention is based on the approach that if there are problems due to differences in optical properties, then wavelengths should be used where the optical properties, in particular the optical path lengths are identical. Hereby, variations in sampling volume caused by differences in optical properties at the imaging wavelengths can be prevented.

A meaningful image can still be created when using such wavelengths, as explained below.

35 As no expression is available for the optical path length, the optical penetration depth, δ , is looked into, defined as:

$$(l) = 2\delta = \frac{z}{\sqrt{2\mu_a\mu'_s}} \tag{1.7}$$

wherein μ'_s is a reduced scattering coefficient. Expression (1.6) can be rewritten as:

$$5 \quad X = \frac{\ln(R_d(\lambda_2))}{\ln(R_d(\lambda_1))} = \frac{\mu_a(\lambda_2)}{\mu_a(\lambda_1)} \sqrt{\frac{\mu_a(\lambda_1)\mu'_s(\lambda_1)}{\mu_a(\lambda_2)\mu'_s(\lambda_2)}} = \sqrt{\frac{\mu_a(\lambda_2)\mu'_s(\lambda_1)}{\mu_a(\lambda_1)\mu'_s(\lambda_2)}} \tag{1.8}$$

A path length corrected absorption ratio can now be calculated with:

$$\left(\frac{\psi\mu_{a,d}(\lambda_2) + (1-\psi)\mu_{a,s}(\lambda_2)}{\psi\mu_{a,d}(\lambda_1) + (1-\psi)\mu_{a,s}(\lambda_1)} \right) = \frac{\mu'_s(\lambda_2)}{\mu'_s(\lambda_1)} X^2 \tag{1.9}$$

10

This clearly illustrates a part of the problem: to obtain ψ from X , the ratio of the scattering

coefficients $\frac{\mu'_s(\lambda_2)}{\mu'_s(\lambda_1)}$ needs to be known.

15

Now a wavelength, λ_0 , is looked for, for which the two tissues, first and second optical absorbers, have identical path lengths at the two wavelengths. This ensures that photons at both wavelengths do not see any differences between the two tissues and cannot do anything else than sampling the same volume.

20

The question, however, is that if the photons cannot 'see' differences between the tissues, how any differences can be imaged, i.e., the consequence of such a choice would be that at every concentration ratio the ratio X would be equal to 1.

25

Therefore, concentration ratios are obtained from the differences in shapes of the diffuse reflection spectra, by looking at the first or higher derivatives of the diffuse reflectance with respect to the wavelength measured closely around wavelengths that generate identical path lengths.

30

An imaging ratio DX is defined by subtracting two images taken $\Delta\lambda$ apart from a central wavelength λ_0 and dividing by the sum of the two images:

$$DX \equiv \frac{\ln(R_d(\lambda_0 + \Delta\lambda)) - \ln(R_d(\lambda_0 - \Delta\lambda))}{\ln(R_d(\lambda_0 + \Delta\lambda)) + \ln(R_d(\lambda_0 - \Delta\lambda))} \tag{1.10}$$

Because $\Delta\lambda$ is taken small with respect to λ_0 , $R_d(\lambda)$ around λ_0 can be linearized, so that:

$$DX = \frac{\ln\left(\left\langle R_d(\lambda_0 + \Delta\lambda) \right\rangle\right) - \ln\left(\left\langle R_d(\lambda_0 - \Delta\lambda) \right\rangle\right)}{2\ln\left(\left\langle R_d(\lambda_0) \right\rangle\right)} \tag{1.11}$$

This can be worked into

$$5 \quad DX = \frac{1}{2} \left(\frac{\ln\left(\left\langle R_d(\lambda_0 + \Delta\lambda) \right\rangle\right)}{\ln\left(\left\langle R_d(\lambda_0) \right\rangle\right)} - \frac{\ln\left(\left\langle R_d(\lambda_0 - \Delta\lambda) \right\rangle\right)}{\ln\left(\left\langle R_d(\lambda_0) \right\rangle\right)} \right) \tag{1.12}$$

Using a linear expansion of $\mu_a(\lambda)$ around λ_0 :

$$DX = \frac{1}{2} \left(\frac{\left\langle l_{\lambda_0 + \Delta\lambda} \right\rangle \mu_a(\lambda_0) + \Delta\lambda \left. \frac{\partial \mu_a}{\partial \lambda} \right|_{\lambda_0}}{\left\langle l_{\lambda_0} \right\rangle \mu_a(\lambda_0)} - \frac{\left\langle l_{\lambda_0 - \Delta\lambda} \right\rangle \mu_a(\lambda_0) - \Delta\lambda \left. \frac{\partial \mu_a}{\partial \lambda} \right|_{\lambda_0}}{\left\langle l_{\lambda_0} \right\rangle \mu_a(\lambda_0)} \right) \tag{1.13}$$

10

Rearranging:

$$DX = \frac{\left\langle l_{\lambda_0 + \Delta\lambda} \right\rangle - \left\langle l_{\lambda_0 - \Delta\lambda} \right\rangle}{2\left\langle l_{\lambda_0} \right\rangle} + \frac{\Delta\lambda \left. \frac{\partial \mu_a}{\partial \lambda} \right|_{\lambda_0}}{\mu_a(\lambda_0)} \left(\frac{\left\langle l_{\lambda_0 + \Delta\lambda} \right\rangle + \left\langle l_{\lambda_0 - \Delta\lambda} \right\rangle}{2\left\langle l_{\lambda_0} \right\rangle} \right) \tag{1.14}$$

15 leads to:

$$DX = \frac{\Delta\lambda \left. \frac{\partial \left\langle l_{\lambda_0} \right\rangle}{\partial \lambda} \right|_{\lambda_0}}{\left\langle l_{\lambda_0} \right\rangle} + \frac{\Delta\lambda \left. \frac{\partial \mu_a}{\partial \lambda} \right|_{\lambda_0}}{\mu_a(\lambda_0)} \tag{1.15}$$

Approximating $\left\langle l_{\lambda} \right\rangle$ according to expression (1.7) yields:

20

$$\frac{\Delta\lambda \left. \frac{\partial \left\langle l_{\lambda_0} \right\rangle}{\partial \lambda} \right|_{\lambda_0}}{\left\langle l_{\lambda_0} \right\rangle} = -\frac{\Delta\lambda}{2} \left(\frac{\left. \frac{\partial \mu_a}{\partial \lambda} \right|_{\lambda_0}}{\mu_a(\lambda_0)} + \frac{\left. \frac{\partial \mu_s}{\partial \lambda} \right|_{\lambda_0}}{\mu_s'(\lambda_0)} \right) \tag{1.16}$$

which leads to:

$$25 \quad DX = -\frac{\Delta\lambda}{2} \left(\frac{\left. \frac{\partial \mu_a}{\partial \lambda} \right|_{\lambda_0}}{\mu_a(\lambda_0)} + \frac{\left. \frac{\partial \mu_s}{\partial \lambda} \right|_{\lambda_0}}{\mu_s'(\lambda_0)} \right) + \Delta\lambda \frac{\left. \frac{\partial \mu_a}{\partial \lambda} \right|_{\lambda_0}}{\mu_a(\lambda_0)} \tag{1.17}$$

Thus:

$$DX = \frac{\Delta\lambda}{2} \left(\frac{\frac{\partial \mu_a}{\partial \lambda} \Big|_{\lambda_0}}{\mu_a(\lambda_0)} - \frac{\frac{\partial \mu'_a}{\partial \lambda} \Big|_{\lambda_0}}{\mu'_a(\lambda_0)} \right) \quad (1.18)$$

Differentiating expression (1.5) versus λ yields:

$$5 \quad \frac{\partial \mu_a}{\partial \lambda} \Big|_{\lambda_0} = (C_1 + C_2) \left(\psi \frac{\partial \mu_{a,1}}{\partial \lambda} \Big|_{\lambda_0} + (1 - \psi) \frac{\partial \mu_{a,2}}{\partial \lambda} \Big|_{\lambda_0} \right) \quad (1.19)$$

Before ψ can be solved from expression (1.19) some strategic choices are made. First, wavelengths for which the absorption spectra of the two absorbers have an isosbestic point are considered. From expression (1.5) it is derived:

10

$$\mu_a(\lambda_0) = (C_1 + C_2) \mu_{a,1}(\lambda_0) = (C_1 + C_2) \mu_{a,2}(\lambda_0) = (C_1 + C_2) \mu_{a,0} \quad (1.20)$$

Where λ_0 stands for the wavelength at which the two molar absorption coefficients $\mu_{a,1}(\lambda_0)$ and $\mu_{a,2}(\lambda_0)$ are identical. Now expression (1.16) turns into:

15

$$DX = \frac{\Delta\lambda}{2} \left(\frac{\left(\psi \frac{\partial \mu_{a,1}}{\partial \lambda} \Big|_{\lambda_0} + (1 - \psi) \frac{\partial \mu_{a,2}}{\partial \lambda} \Big|_{\lambda_0} \right)}{\mu_{a,0}} - \frac{\frac{\partial \mu'_a}{\partial \lambda} \Big|_{\lambda_0}}{\mu'_a(\lambda_0)} \right) \quad (1.21)$$

20

The first term in expression (1.21) has a linear dependence on ψ with constants that are related to the first derivative of the absorption spectra. Thus, the ratio image defined in expression (1.21) can be directly interpreted in terms of ψ if the absorption spectra of the two absorbers are known and sufficiently different in shape. Obviously, the ratio increases with increasing $\Delta\lambda$. Taking $\Delta\lambda$ too large will induce deviations from the premise that μ_a is constant. The order of magnitude of the first term depends on the steepness of the absorption curves. A simple estimate can be by considering a Gaussian absorption profile with bandwidth λ_w :

25

$$\mu_a(\lambda) = \alpha e^{-\frac{(\lambda - \lambda_c)^2}{2\lambda_w^2}} \quad (1.22)$$

where λ_c and λ_w stand for the central wavelength and the half bandwidth, respectively.

30

Now an estimate for the first term in expression (1.18) can be derived:

$$\frac{\frac{\Delta\lambda}{2} \frac{\partial \mu_a}{\partial \lambda}}{\mu_a(\lambda_0)} = \frac{\Delta\lambda}{2} \frac{\lambda - \lambda_c}{\lambda_w^2} \frac{\left(\frac{-(\lambda - \lambda_c)^2}{2\lambda_w^2} \right)}{\alpha e} = - \frac{\Delta\lambda \lambda - \lambda_c}{2 \lambda_w^2} \tag{1.23}$$

At the steepest point of the absorption curve, i.e. $\lambda - \lambda_c$ equals $\pm \lambda_w$. In case the isosbestic point λ_0 lies anywhere near the absorption band this turns into:

5

$$\frac{\Delta\lambda}{2} \frac{\partial \mu_a}{\partial \lambda} \approx \pm \frac{\Delta\lambda}{2\lambda_w} \tag{1.24}$$

The second term has a more complicated behaviour. It describes the relative change in scattering coefficient with wavelength. In the case of Mie scatter:

10

$$\mu_s(\lambda) = \mu_s(\lambda_0) \left(\frac{\lambda}{\lambda_0} \right)^b \tag{1.26}$$

The second term in expression (1.18) simply becomes:

15

$$\frac{\Delta\lambda}{2} \frac{\partial \mu_s'}{\partial \lambda} \Big|_{\lambda_0} = \frac{\Delta\lambda}{2} \frac{b}{\lambda_0} \mu_s(\lambda_0) \left(\frac{\lambda}{\lambda_0} \right)^{b-1} = \frac{\Delta\lambda}{2} \frac{b}{\lambda} \mu_s(\lambda_0) \left(\frac{\lambda}{\lambda_0} \right)^b \tag{1.27}$$

It is important to note that this is independent of the amount of scattering. It can be seen that the first term is much larger than the second term:

20

$$\frac{\Delta\lambda}{2\lambda_w} \gg \frac{\Delta\lambda}{\lambda} \frac{b}{2}, \quad \text{or} \quad \frac{\lambda_0}{\lambda_w} \gg b \tag{1.28}$$

While values for b for most tissues are in the range from 0.5 to 2.0 and typical values for λ_w and λ_0 of 40 nm and 1000 nm it is clear that the requirement of expressions (1.28) is fulfilled and that the first term in expression (1.18) is much bigger than the second term. An

25

estimation of the concentration ratio ψ can now with sufficiently high accuracy be calculated from:

$$\psi = \frac{\frac{\frac{\partial \mu_{a,2}}{\partial \lambda} \Big|_{\lambda_0}}{\mu_{a,2}}}{\frac{\frac{\partial \mu_{a,1}}{\partial \lambda} \Big|_{\lambda_0}}{\mu_{a,1}}} = \frac{\frac{\partial \mu_{a,2}}{\partial \lambda} \Big|_{\lambda_0}}{\frac{\partial \mu_{a,1}}{\partial \lambda} \Big|_{\lambda_0}} = \frac{f - g}{f - g} \tag{1.29}$$

where the constants f and g are the relative derivatives of the two absorbers defined by expressions (1.30):

$$f = \frac{\left. \frac{\partial \mu_{a,1}}{\partial \lambda} \right|_{\lambda_0}}{\mu_{a,0}} \quad \text{and} \quad g = \frac{\left. \frac{\partial \mu_{a,2}}{\partial \lambda} \right|_{\lambda_0}}{\mu_{a,0}}, \quad (1.30)$$

5

and where:

- λ indicates a wavelength;
- λ_0 indicates an isosbestic wavelength;
- $\Delta\lambda$ indicates a wavelength difference from the isosbestic wavelength;
- 10 - DX indicates an imaging ratio, defined by subtracting two images taken $\Delta\lambda$ apart from an isosbestic wavelength λ_0 and dividing by the sum of the two images;
- $\mu_{a,m}$ indicates a molar absorption coefficient of absorber m ;
- $\mu_{a,0}$ indicates a molar absorption coefficient at isosbestic wavelength.

15 Thus, the concentration ratio (ψ) may be determined using the partial derivative of a first molar absorption coefficient of the first absorber as a function of the wavelength, and the partial derivative of a second molar absorption coefficient of the second absorber as a function of the wavelength, both taken at a wavelength at which the first molar absorption coefficient and the second molar absorption coefficient are equal.

20 Instead of, or in addition to, the first derivative used here, similar considerations as the ones above may be applied to estimate the concentration ratio ψ , based on the use of second or higher derivatives. However, it is noted that the signal to noise ratio decreases substantially for the second or higher derivatives, which makes it at least less effective to use the second or higher derivatives.

25

In an embodiment, the method of the invention further comprises the steps of comparing the concentration ratio ψ to a predetermined concentration ratio range, and indicating a defect for the surface area when the concentration ratio is outside the concentration ratio range.

30 For various media, in particular media comprised of, or comprising, living tissue, a concentration ratio of a first absorber and a second absorber in healthy tissue differs, often substantially, from a concentration ratio of unhealthy tissue. As an example, in mammal breast tissue, normal tissue has a lower fat to water ratio than cancer tissue. Now, by defining a concentration ratio range applicable to normal tissue, cancer tissue may be detected from
35 its higher fat to water ratio, outside the concentration ratio range.

Accordingly, for one or more surface areas the concentration ratio may be determined, and when a defect is indicated for at least one surface area, an indication of a defect may be provided for the totality of surface areas. For example, an indication may be provided in an image of a plurality of surface areas, by colouring any surface area in which the concentration ratio is outside the concentration ratio range differently than other surface areas. From the image, a location and extension of cancer tissue may be recognized.

In an embodiment, the method of the invention further comprises the steps of repeating steps (a) to (d) for different isosbestic wavelengths of the first absorber and the second absorber in different wavelength regions to obtain estimated concentration ratios for each one of the isosbestic wavelengths.

Over a large range of wavelengths, there will often be several wavelength regions containing one or more isosbestic wavelengths where the optical penetration depth is independent of the ratio of the absorbers. The optical penetration depths in different wavelength regions can vary from a few centimetre to less than a millimetre. By performing the first and higher derivative imaging around isosbestic wavelengths in different wavelength regions, the optical penetration depth (sampling depth), and thereby the sampling volume of the imaging method can be changed.

In an embodiment, the method of the invention with repeating performing steps (a) to (d) for different isosbestic wavelengths further comprises comparing each one of the estimated concentration ratios to a predetermined concentration ratio range, and indicating a defect for the surface area when at least one of the concentration ratios is outside the concentration ratio range.

In an embodiment, the method of the invention with repeating performing steps (a) to (d) for different isosbestic wavelengths further comprises associating each isosbestic wavelength with a respective optical penetration depth, and indicating a defect and a depth thereof for the surface area if at least one of the concentration ratios is outside the concentration ratio range.

Thus, the method allows for providing not only indications of defects for surface areas, i.e. in two dimensions, but also for providing indications of the defects for sampling volumes, i.e. in three dimensions including a volume located beneath the imaged surface area.

In an embodiment, the method of the invention further comprises associating each isosbestic wavelength with a respective optical penetration depth, and, if small defects are to be located,

selecting an isosbestic wavelength among the different isosbestic wavelengths with a low optical penetration depth.

By changing isosbestic wavelengths, not only the optical penetration depth is changed, but also the spatial resolution. A relatively high optical penetration depth is associated with a
5 relatively low spatial resolution, and a relatively low optical penetration depth is associated with a relatively high spatial resolution. Thus, if particular sensitivity to small defects is required, a wavelength region with relatively low optical penetration depth is selected, to obtain a higher spatial resolution and hence a lower threshold for detection.

10

In a second aspect of the invention, an apparatus for determining a property of an optically diffuse medium is provided. The medium comprises a first optical absorber having a first concentration and a second optical absorber having a second concentration. The apparatus comprises:

15

a light source configured to illuminate a surface area of a volume of the medium;

a filtering device configured to receive reflected light from the surface area of the medium and to transmit filtered light to an optical imaging device, wherein the filtering device is configured to filter multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber;

20

a twodimensional, 2D, optical imaging device configured to receive the filtered light from the filtering device;

an image processing component configured to determine a reflectance spectrum of the medium at the surface area at the multiple wavelengths;

25

a calculating component configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength; and

an estimating component configured to estimate a concentration ratio of the first concentration and the second concentration from the derivative.

30

In a third aspect of the invention, an apparatus for determining a property of an optically diffuse medium is provided. The medium comprises a first optical absorber having a first concentration and a second optical absorber having a second concentration. The apparatus comprises:

a light source configured to illuminate a surface area of a volume of the medium;

35

a onedimensional, 1D, or twodimensional, 2D, hyperspectral optical imaging device configured to receive reflected light from the surface area of the medium;

an image processing component configured to determine the reflectance spectrum of the medium at the surface area at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber;

5 a calculating component configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength; and

an estimating component configured to estimate a concentration ratio of the first concentration and the second concentration from the derivative.

10 In an embodiment of the apparatus, wherein the optical imaging device is a 2D hyperspectral optical imaging device comprising an image sensor having a plurality of pixels, the image processing component is further configured to:

assign a cluster of pixels to the surface area of the medium;

determine the reflectance spectrum for each pixel of the image sensor; and

15 determine the reflectance spectrum of the medium at the surface area from the reflectance spectra of the pixels of the cluster.

20 In a fourth aspect of the invention, an apparatus for determining a property of an optically diffuse medium is provided. The medium comprises a first optical absorber having a first concentration and a second optical absorber having a second concentration. The apparatus comprises:

a light source configured to illuminate a surface area of a volume of the medium;

a spectrograph configured to receive reflected light from the surface area of the medium;

25 a light processing component configured to determine the reflectance spectrum of the medium at the surface area at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber;

a calculating component configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength; and

30 an estimating component configured to estimate a concentration ratio of the first concentration and the second concentration from the derivative.

In a fifth aspect of the invention, a computer program of computer program product is provided. The computer program comprises computer instructions which, when loaded in a processor, enable the processor to carry out the method of the invention.

In a sixth aspect of the invention, a non-volatile storage medium is provided. The storage medium stores computer instructions which, when loaded in a processor, enable the processor to carry out the method of the invention.

5 Briefly stated, the present invention provides a method, apparatus and software for accurate diffuse optical imaging. The apparatus comprises a light source emitting a plurality of wavelengths and an optical imaging device capable of producing images at a number of prescribed wavelength bands. Images taken at strategically chosen wavelength bands may be combined to produce an image representing a twodimensional spatial distribution of the
10 ratio of two absorbing substances.

It is noted that in diffuse reflection the light incident on the surface has entered the medium and has travelled through the medium, sometimes more than a cm, before being emitted again from the same surface.

15

In the case of specular reflection, a part of the incident light is directly reflected from the surface of the medium. It is desired to reduce such specular reflection, which is an additional benefit of the derivative algorithm outlined above.

20 In the case of specular reflection, expression 1.1 above becomes:

$$R_d = e^{-\mu_d(\ell)} + R_s \quad (2.1)$$

where R_s stands for the specular reflection. The specular reflection depends on the
25 differences in refractive indices of the medium and of air and is expressed by Fresnel's equations. For normal incidence the specular reflectance R_s is given by:

$$R_s = \left(\frac{n_{\text{medium}} - 1}{n_{\text{medium}} + 1} \right)^2 \quad (2.2)$$

30 The specular reflection can be substantial, e.g. for a refractive index of tissue of 1.35 the specular reflection is more than 2%. For oblique incidence this can be even higher. In general this can be more than the diffuse reflection. The specular reflection does not contain information about the composition of the tissue and is considered a disturbing factor in many situations. Derivative imaging offers a solution here.

35

The effect of specular reflection is shown by entering the specular reflection into expression (1.11):

$$DX = \frac{\ln\left(\left(R_s(\lambda_0 + \Delta\lambda) + R_d(\lambda_0 + \Delta\lambda)\right)\right) - \ln\left(\left(R_s(\lambda_0 + \Delta\lambda) + R_d(\lambda_0 - \Delta\lambda)\right)\right)}{2\ln\left(\left(R_s(\lambda_0 + \Delta\lambda) + R_d(\lambda_0)\right)\right)} \quad (2.3)$$

5

The analysis as described above leading to a fat-to-water ratio should be adapted: the contribution of R_c is large and cannot be neglected.

Hence, a slightly different approach is followed. The derivative is still used, but there is no normalization with respect to the reflection:

10

$$UDX = R_s(\lambda_0 + \Delta\lambda) + R_d(\lambda_0 + \Delta\lambda) - \left(R_s(\lambda_0 + \Delta\lambda) + R_d(\lambda_0 - \Delta\lambda)\right) \quad (2.4)$$

where UDX is an non-normalized imaging ratio.

15

For media such as human tissue, the refractive index varies only a few percent over a large wavelength region. For instance for human colon, the refractive index decreases from 1.37 at 500 nm to 1.32 at 1550 nm¹. In the Near Infrared the specular reflectance R_s (expression (2.2) changes only 10⁻⁶ per nanometre. As a consequence expression (2.4) changes into:

20

$$UDX = R_d(\lambda_0 + \Delta\lambda) - R_d(\lambda_0 - \Delta\lambda) \quad (2.5)$$

A linear expansion around λ_0 yields:

$$UDX = 2\Delta\lambda \frac{dR_d}{d\lambda} \quad (2.6)$$

Using expression (1.1), it can be derived:

$$UDX = -2\Delta\lambda R_d(\lambda_0) \frac{d\mu_a(\lambda)}{d\lambda} \quad (2.7)$$

30

Replacing $R_d(\lambda_0)$ by R_0 and introducing the concentration ratio ψ from expression (1.5) yields:

¹ Complex refractive index of normal and malignant human colorectal tissue in the visible and near-infrared. Giannios P, Koutsoumpos S, Toutouzas KG, Matiatou M, Zografos GC, Moutzouris K. J Biophotonics. 2016 Apr 19. doi: 10.1002/jbio.201600001

$$UDX = -2\Delta\lambda \frac{R_\sigma}{\mu_{\alpha,\sigma}} \left(\psi \frac{\partial \mu_{\alpha,1}}{\partial \lambda} \Big|_{\lambda_\sigma} + (1 - \psi) \frac{\partial \mu_{\alpha,2}}{\partial \lambda} \Big|_{\lambda_\sigma} \right) \quad (2.8)$$

The concentration ratio ψ can now be calculated from:

$$5 \quad \psi = UDX \frac{1}{2\Delta\lambda} \frac{\mu_{\alpha,\sigma}}{R_\sigma} \frac{\frac{\partial \mu_{\alpha,2}}{\partial \lambda} \Big|_{\lambda_\sigma}}{\frac{\partial \mu_{\alpha,1}}{\partial \lambda} \Big|_{\lambda_\sigma} - \frac{\partial \mu_{\alpha,2}}{\partial \lambda} \Big|_{\lambda_\sigma}} \quad (2.9)$$

In other words, by imaging UDX as defined in expression (2.4), a quantity is imaged that is proportional to the concentration ratio of the two chromophores that have an isosbestic point at λ_σ . A difference with the method above is that the proportionality constant is not known a priori, but the image is completely free of specular reflections. A similar outcome applies for
 10 second and higher derivatives.

These and other aspects of the invention will be more readily appreciated as the same becomes better understood by reference to the following detailed description and considered in connection with the accompanying drawings in which like reference symbols designate like
 15 parts.

BRIEF DESCRIPTION OF THE DRAWINGS

Figure 1 depicts a diagram showing graphs of an effective penetration depth [cm] at different
 20 wavelengths [nm] of a medium containing different ratios of fat and water.

Figure 2 depicts different diagrams showing graphs of effective penetration depth versus wavelength, diffuse reflectance versus wavelength, first derivative of diffuse reflectance versus wavelength, and second derivative of diffuse reflectance versus wavelength, all at
 25 three different wavelength ranges.

Figure 3 depicts a flow diagram illustrating steps of an embodiment of a method according to the invention.

30 Figure 4 schematically illustrates components of a first embodiment of an apparatus according to the invention.

Figure 5 schematically illustrates components of a second embodiment of an apparatus according to the invention.

Figure 6 schematically illustrates components of a third embodiment of an apparatus according to the invention.

5 Figure 7 schematically illustrates components of a fourth embodiment of an apparatus according to the invention.

Figures 8a, 8b, 8c and 8d illustrate a specular reflection reduction using a non-normalized imaging ratio.

10 Figures 9a, 9b, 9c, 9d, 9e and 9f illustrate test results in distinguishing invasive ductal carcinoma, IDC.

Figures 10a, 10b, 10c, 10d, 10e and 10f illustrate test results in distinguishing invasive ductal carcinoma, IDC, and ductal carcinoma in situ, DCIS.

15

Figures 11a, 11b, 11c, 12a, 12b, 12c and 12d illustrate a phantom study with an optical phantom.

DETAILED DESCRIPTION OF EMBODIMENTS

20

Figure 1 depicts a diagram showing graphs of an effective penetration depth [cm] at different wavelengths [nm] of a medium containing different ratios of fat (a first optical absorber having a first concentration) and water (a second optical absorber having a second concentration). The optical penetration depths have been calculated with diffusion theory.

25

In a first (low) graph, marked A, the effective penetration depth versus wavelength for a medium containing 10% by volume of fat and 90% by volume of water is shown. In a second (middle) graph, marked B, the effective penetration depth versus wavelength for a medium containing 50% by volume of fat and 50% by volume of water is shown. In a third (high) graph, marked C, the effective penetration depth versus wavelength for a medium containing 90% by volume of fat and 10% by volume of water is shown.

30

As indicated in Figure 1 by upwards pointing arrows, different isosbestic points (wavelengths) can be identified, at about 930 nm, about 1217 nm, about 1700 nm, and about 1736 nm, respectively. At these wavelengths, the effective (optical) penetration depth is the same for the different ratios of fat and water according to graphs A, B and C, and for other ratios of fat

35

and water. Also, at these wavelengths, inhomogeneous distributions of fat and water in the medium do not influence the optical penetration depth.

In case the optical properties of the two absorbers, i.e. the first optical absorber fat and the second optical absorber water, do not differ much, a concentration ratio cannot be determined at an isosbestic wavelength.

Hence, Figure 2 depicts different diagrams showing graphs of effective penetration depth versus wavelength, diffuse reflectance versus wavelength, first derivative of diffuse reflectance versus wavelength, and second derivative of diffuse reflectance versus wavelength, all at three different wavelength ranges around isosbestic wavelengths.

In a left column comprising four diagrams, from top to bottom an effective penetration depth, an associated diffuse reflectance, and associated first derivative of the diffuse reflectance, and an associated second derivative of the diffuse reflectance, respectively, are shown for three different concentration ratios of fat and water (10/90 (graph A), 50/50 (graph B), and 90/10 (graph C)), in a first wavelength range from 910 nm to 950 nm around an isosbestic wavelength of about 930 nm, as indicated by a vertical dashed line.

In a middle column comprising four diagrams, from top to bottom an effective penetration depth, an associated diffuse reflectance, and associated first derivative of the diffuse reflectance, and an associated second derivative of the diffuse reflectance, respectively, are shown for three different concentration ratios of fat and water (10/90 (graph A), 50/50 (graph B), and 90/10 (graph C)), in a second wavelength range from 1170 nm to 1240 nm around an isosbestic wavelength of about 1217 nm, as indicated by a vertical dashed line.

In a right column comprising four diagrams, from top to bottom an effective penetration depth, an associated diffuse reflectance, and associated first derivative of the diffuse reflectance, and an associated second derivative of the diffuse reflectance, respectively, are shown for three different concentration ratios of fat and water (10/90 (graph A), 50/50 (graph B), and 90/10 (graph C)), in a third wavelength range from 1680 nm to 1750 nm around isosbestic wavelengths of about 1700 and 1736 nm, as indicated by vertical dashed lines.

From the third and fourth rows, each comprising three diagrams showing the first derivative and the second derivative, respectively, it can be seen these first and second derivatives of the diffuse reflection spectra clearly are different at the isosbestic wavelengths for different concentration ratios of the first optical absorber fat and the second optical absorber water.

Based on this property of the diffuse reflection spectrum of a medium containing essentially two main optical absorbers, the following method is proposed, as illustrated in the flow diagram of Figure 3.

5 With the method, a property of an optically diffuse medium, in particular a concentration ratio of a first concentration of a first optical absorber and a second concentration of a second optical absorber, can be determined, i.e. estimated.

10 In a step 301, a surface area of the medium is imaged at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber. The surface area is selected to be in conformity with the spatial resolution to be expected for the medium. The imaging may be performed using an appropriate optical imaging device, as will be explained below by reference to Figures 4 to 7.

15 In Figure 3, in a subsequent step 302, a reflectance spectrum of the medium is determined at the surface area at the multiple wavelengths.

In a subsequent step 303, a derivative of the determined reflectance spectrum around the isosbestic wavelength is determined.

20 In a subsequent step 304, a concentration ratio of the first concentration and the second concentration is estimated from the derivative. In particular, a concentration ratio ψ may be determined using the following expression:

25
$$\psi = \frac{\frac{DX}{\Delta\lambda} \frac{\frac{\partial \mu_{a,2}}{\partial \lambda} |_{\lambda_0}}{\mu_{a,0}}}{\frac{\frac{\partial \mu_{a,1}}{\partial \lambda} |_{\lambda_0}}{\mu_{a,0}} - \frac{\frac{\partial \mu_{a,2}}{\partial \lambda} |_{\lambda_0}}{\mu_{a,0}}} = \frac{DX}{\Delta\lambda} \frac{f}{f-g}$$

where:

- f and g are defined by $f = \frac{\frac{\partial \mu_{a,1}}{\partial \lambda} |_{\lambda_0}}{\mu_{a,0}}$ and $g = \frac{\frac{\partial \mu_{a,2}}{\partial \lambda} |_{\lambda_0}}{\mu_{a,0}}$;
- λ indicates a wavelength;
- λ_0 indicates an isosbestic wavelength;
- 30 - $\Delta\lambda$ indicates a wavelength difference from the isosbestic wavelength;
- DX indicates an imaging ratio, defined by subtracting two images taken $\Delta\lambda$ apart from an isosbestic wavelength λ_0 and dividing by the sum of the two images;
- $\mu_{a,m}$ indicates a molar absorption coefficient of absorber m ;
- $\mu_{a,0}$ indicates a molar absorption coefficient at isosbestic wavelength.

Once the concentration ratio ψ has been determined, then according to step 305, it may be compared to a predetermined concentration ratio range. This predetermined concentration ratio range may have been established based on evidence that certain concentration ratios within the range are normal for the medium. Other concentration ratios outside the

5 concentration ratio range are abnormal for the medium, and therefore may be qualified as a defect. According to step 306, if the concentration ratio is outside the concentration ratio range, a defect is indicated for the surface area of the medium, in an appropriate way.

10 The steps 302 to 306 may be performed by components of one or more (data) processing devices and associated input and output devices.

Figure 4 schematically illustrates components of a first embodiment of an apparatus 400 according to the invention.

15 The apparatus 400 comprises a light source 402 configured to illuminate (as indicated by arrow 404) a surface area 406 of a volume of an optically diffuse medium. The medium comprises a first optical absorber having a first concentration and a second optical absorber having a second concentration. Light is reflected (as indicated by arrow 408) from the surface area 406 of the medium to a filtering device 410 configured to receive such reflected light from the surface area of the medium, and to transmit filtered light (as indicated by arrow 420)

20 to a twodimensional, 2D, optical imaging device 422 configured to receive the filtered light 420 from the filtering device 410.

The filtering device 410 is configured to filter multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber. The filtering device 410 comprises

25 a mechanical device 412 (also referred to as a filter wheel) comprising a filter support 414 with multiple different optical filters 416 each filtering a different wavelength from the reflected light 408. In the mechanical device 412, the filter support 414 which can be rotated in either one of directions indicated by double arrow 417 by an actuator 418 coupled to the filter support 414, whereby each one of the multiple optical filters 416 can be positioned in the

30 optical path of the reflected light 408 to provide light 420 filtered at different wavelengths.

The optical imaging device 422 may e.g. comprise a CCD camera, an InGaAs camera or a CMOS camera, and provides image data which are transferred to an image processing component 424 configured to determine a reflectance spectrum of the medium at the surface

35 area 406 at the multiple wavelengths. Data relating to the reflectance spectrum are transferred from the image processing component 424 to a calculating component 426 configured to determine a derivative of the determined reflectance spectrum around the

isobestic wavelength. Data relating to the derivative of the determined reflectance spectrum around the isobestic wavelength are transferred from the calculating component 426 to an estimating component 428 configured to estimate a concentration ratio ψ of the first concentration and the second concentration from the derivative. Data relating to an estimated concentration ratio ψ are transferred to a comparator 430 configured to compare the concentration ratio ψ to a predetermined concentration ratio range. The comparator 430 controls an output component 432, such as an image output component, configured to indicate a defect for the surface area 406 when the concentration ratio ψ is outside the concentration ratio range. For example, if a defect is to be indicated for the surface area 406, the output component 432 may display the surface area 406 having a color or any other indication to differentiate the surface area 406 from other surface areas for which a defect is not to be indicated.

Figure 5 schematically illustrates components of a second embodiment of an apparatus 500 according to the invention.

The apparatus 500 comprises a light source 502 configured to illuminate (as indicated by arrow 504) a surface area 506 of a volume of an optically diffuse medium. The medium comprises a first optical absorber having a first concentration and a second optical absorber having a second concentration. Light is reflected (as indicated by arrow 508) from the surface area 506 of the medium. The apparatus 500 further comprises a onedimensional, 1D, hyperspectral optical imaging device 510 configured to receive reflected light 508 from a scanned line 512 of the surface area 506 of the medium. For each pixel of the hyperspectral optical imaging device 510, an entire spectrum is acquired.

The medium is supported on a stage 520 coupled to an actuator 522 configured to drive the stage 520 to move the medium relative to the optical imaging device 510 in directions as indicated by double arrow 524. For example, the actuator 522 is coupled to the stage 520 through a rotatable spindle 526 to move the stage 520. By incrementally moving the stage 520, different lines of the surface area 506 are imaged.

The 1D optical imaging device 510 comprises, for example, a CCD, InGaAs or CMOS hyperspectral camera, and acquires image data in the form of a (x,y,λ) (x -coordinate, y -coordinate, wavelength) data cube, from which imaging wavelengths are chosen to select images. The optical imaging device 510 provides the selected image data which are transferred to an image processing component 530 configured to determine the reflectance spectrum of the medium at the surface area 506 at multiple wavelengths around an isobestic

wavelength of the first absorber and the second absorber. Data relating to the reflectance spectrum are transferred from the image processing component 530 to a calculating component 532 configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength. Data relating to the derivative of the determined

5 reflectance spectrum around the isosbestic wavelength are transferred from the calculating component 532 to an estimating component 534 configured to estimate a concentration ratio ψ of the first concentration and the second concentration from the derivative. Data relating to an estimated concentration ratio ψ are transferred to a comparator 536 configured to compare the concentration ratio ψ to a predetermined concentration ratio range. The

10 comparator 536 controls an output component 538, such as an image output component, configured to indicate a defect for the surface area 506 when the concentration ratio ψ is outside the concentration ratio range. For example, if a defect is to be indicated for the surface area 506, the output component 538 may display the surface area 506 having a color or any other indication to differentiate the surface area 506 from other surface areas for which

15 a defect is not to be indicated.

Figure 6 schematically illustrates components of a third embodiment of an apparatus 600 according to the invention.

20 The apparatus 600 comprises a light source 602 configured to illuminate (as indicated by arrow 604) a surface area 606 of a volume of an optically diffuse medium. The medium comprises a first optical absorber having a first concentration and a second optical absorber having a second concentration. Light is reflected (as indicated by arrow 608) from the surface area 606 of the medium. The apparatus 600 further comprises a twodimensional, 2D,

25 hyperspectral optical imaging device 610 configured to receive reflected light 608 from a scanned surface sub-area 612 of the surface area 606 of the medium. For each pixel of the hyperspectral optical imaging device 610, an entire spectrum is acquired.

The medium is supported on a stage 620 coupled to an actuator 622 configured to drive the

30 stage 620 to move the medium relative to the optical imaging device 610 in directions as indicated by double arrows 624, 625. By incrementally moving the stage 620, different surface sub-areas 612 of the surface area 606 are imaged.

In another embodiment lacking a stage, different surface sub-areas 612 of the surface area

35 606 may be scanned by use of movable scanning mirrors (not shown) directing the reflected light 608 to the 2D optical imaging device 610.

The 2D optical imaging device 610 comprises, for example, a CCD, InGaAs or CMOS hyperspectral camera, and acquires image data in the form of a (x,y,λ) (x-coordinate, y-coordinate, wavelength) data cube, from which imaging wavelengths are chosen to select images. The optical imaging device 610 provides the selected image data which are
5 transferred to an image processing component 630 configured to determine the reflectance spectrum of the medium at the surface area 606 at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber.

The 2D hyperspectral optical imaging device 610 comprises an image sensor having a
10 plurality of pixels. The image processing component 630 is configured to assign a cluster of pixels of the image sensor to the surface sub-area 612 of the medium, to determine the reflectance spectrum for each pixel of the image sensor, and to determine the reflectance spectrum of the medium at the surface area 606 from the reflectance spectra of the pixels of the cluster.

Data relating to the reflectance spectrum are transferred from the image processing
15 component 630 to a calculating component 632 configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength. Data relating to the derivative of the determined reflectance spectrum around the isosbestic wavelength are
20 transferred from the calculating component 632 to an estimating component 634 configured to estimate a concentration ratio ψ of the first concentration and the second concentration from the derivative. Data relating to an estimated concentration ratio ψ are transferred to a
comparator 636 configured to compare the concentration ratio ψ to a predetermined
25 concentration ratio range. The comparator 636 controls an output component 638, such as an image output component, configured to indicate a defect for the surface area 606 when the
concentration ratio ψ is outside the concentration ratio range. For example, if a defect is to be
indicated for the surface area 606, the output component 638 may display the surface area
606 having a color or any other indication to differentiate the surface area 606 from other
surface areas for which a defect is not to be indicated.

30 It is noted that, instead of a 2D hyperspectral optical imaging device 610, a spectrograph can be applied, as explained by reference to Figure 7.

Figure 7 schematically illustrates components of a fourth embodiment of an apparatus 700
35 according to the invention.

The apparatus 700 comprises a light source 702 configured to illuminate (as indicated by arrow 704) a surface area 706 of a volume of an optically diffuse medium. The medium comprises a first optical absorber having a first concentration and a second optical absorber having a second concentration. Light is reflected (as indicated by arrow 708) from the surface area 706 of the medium. The apparatus 700 further comprises a spectrograph 710 configured to receive reflected light 708 from a scanned surface sub-area 712 of the surface area 706 of the medium. For each pixel of the spectrograph 710, an entire spectrum is acquired.

The medium is supported on a stage 720 coupled to an actuator 722 configured to drive the stage 720 to move the medium relative to spectrograph 710 in directions as indicated by double arrows 724, 725. By incrementally moving the stage 720, different surface sub-areas 712 of the surface area 706 are imaged.

In another embodiment lacking a stage, different surface sub-areas 712 of the surface area 706 may be scanned by use of a scanning device comprising movable scanning mirrors (not shown) directing the reflected light 708 to the spectrograph 710.

The spectrograph 710 acquires image data in the form of a (x,y,λ) (x-coordinate, y-coordinate, wavelength) data cube, from which imaging wavelengths are chosen to select images. The spectrograph 710 provides the selected image data which are transferred to an image processing component 730 configured to determine the reflectance spectrum of the medium at the surface area 706 at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber.

Data relating to the reflectance spectrum are transferred from the image processing component 730 to a calculating component 732 configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength. Data relating to the derivative of the determined reflectance spectrum around the isosbestic wavelength are transferred from the calculating component 732 to an estimating component 734 configured to estimate a concentration ratio ψ of the first concentration and the second concentration from the derivative. Data relating to an estimated concentration ratio ψ are transferred to a comparator 736 configured to compare the concentration ratio ψ to a predetermined concentration ratio range. The comparator 736 controls an output component 738, such as an image output component, configured to indicate a defect for the surface area 706 when the concentration ratio ψ is outside the concentration ratio range. For example, if a defect is to be indicated for the surface area 706, the output component 738 may display the surface area

706 having a color or any other indication to differentiate the surface area 706 from other surface areas for which a defect is not to be indicated.

TEST RESULTS

5

Diffuse reflection

Figure 8a shows a number of spectra taken from different locations in a sample, indicating on a horizontal axis a wavelength in nm in steps of 100 nm from about 900 nm to about 1700 nm, and indicating on a vertical axis a diffuse reflectance (unscaled). Curves 801 and 802 show a spectrum from pixels of the sample with specular reflection, whereas the other curves show spectra from other areas. In a comparison of curve 803 with curve 801, and comparing curve 804 with curve 802, Figure 8a illustrates that the spectra affected by specular reflection do not change shape, but are only shifted, as a wavelength independent amount of reflection is added by the specular reflection to the diffuse reflection from the tissue.

Figures 8b, 8c and 8d illustrate that the first and second derivative can be helpful in decreasing the effects of specular reflection.

20 Figure 8b shows a spectral image of a breast tissue sample at 940 nm. This sample had a very irregular surface causing multiple specular reflections, recognised by the bright white areas 810 in the image.

25 Figure 8c shows a first derivative image around 1225 nm. A much smoother image with normal derivative values even in many of the pixels that were originally affected by specular reflection results. The dark areas as indicated by arrows 805 indicate pixels where one of the pixel values used to calculate the first derivative had reached a saturation level. In such a case the first derivative algorithm cannot compensate the specular reflection.

30 Figure 8d shows a second derivative image around 1225 nm. The image is noisier, but still a smooth image where only a few pixels remain affected by specular reflection, as indicated by arrow 805.

Distinguishing Invasive Ductal Carcinoma from normal tissue

35

Figures 9a, 9b, 9c, 9d, 9e and 9f illustrate test results in distinguishing invasive ductal carcinoma, IDC, confirming the excellent classification based on the first derivative. A

collected dataset consists of samples of eight different patients. All these samples contained both IDC and healthy tissue. Hyperspectral data cubes were obtained from all samples and first derivative images around 1225 nm were calculated. In all images, regions of interest, ROI, were selected that, based on the histopathology, only contained IDC or healthy tissue.

5 From these ROIs the first derivative values were calculated. A training set of 70% (randomly assigned) of these first derivative values was used to train a simple classification algorithm. An evaluation of the method was performed on the remaining 30% of the data.

10 Figure 9a shows a regular grey shade picture of a tissue slice of a patient sample that was assigned to the test dataset. The sample consisted of fatty tissue and a large centrally located invasive ductal carcinoma, IDC. The specimen was obtained fresh from surgery and sliced by the pathologist.

15 Figure 9b shows a hematoxylin-eosin, HE, stained section of the slice of Figure 9a. Line 901 encircles the area indicated by the pathologist as IDC.

Figure 9c shows the first derivative image of the slice taken around 1225 nm.

20 Figure 9d shows a tissue classification on the basis of the first derivative using a cut-off value of 0.044. The central dark grey area 902 indicates IDC, the surrounding light grey area is normal tissue and the further dark area is background, i.e. an area where there was no sample. Figure 9d clearly shows that the first derivative at 1225 nm can be used to distinguish IDC from normal tissue.

25 Figure 9e shows a histogram of first derivative values from the test set, and clearly shows that the first derivative at 1225 nm markedly differentiates IDC from healthy tissue according to the pathologic classification, as indicated by arrows indicating the values for the different tissues. A horizontal axis indicates a value from the first derivative around 1225 nm in steps of 0.02 running from 0 to 0.16, and a vertical axis indicates a number of spectra with that value in steps of 10 running from 0 to 80. The dotted line represents the cut-off value of 30 0.0438, derived from the training set.

35 Figure 9f shows a ROC curve indicating, on a horizontal axis, a false positive rate (1-Specificity) in steps of 0.2 running from 0 to 1, and on a vertical axis, a true positive rate (Sensitivity) in steps of 0.1 running from 0 to 1. Area under the curve, AUC, is 0.987 (confidence interval: 0.979-0.995).

Distinguishing invasive ductal carcinoma and ductal carcinoma in situ from normal tissue

Figures 10a, 10b, 10c, 10d, 10e and 10f illustrate test results in distinguishing invasive ductal carcinoma, IDC, and ductal carcinoma in situ, DCIS. The collected dataset consist of samples of ten different patients. All these samples contained healthy tissue, eight out of ten samples contained IDC and six out of ten samples contained DCIS. DCIS is a form of cancer that is in an earlier stage and is sometime present as small volumes of cancer cells. It is generally agreed that often positive resection margin are related to small pockets of DCIS that much more difficult to see by the surgeon than larger tumours.

In these samples, regions of interest, ROI, were selected that, based on the histopathology, only contained IDC, DCIS or healthy tissue. The spectra in these ROIs were used as a training (~70% of the spectra) and test set (~30% of the spectra) for the first derivative algorithm. Of each patient, the spectra in the ROIs were randomly assigned to either the training set or the test set.

Referring to Figure 10a, a regular grey shade picture of a patient sample is shown that was assigned to the test dataset. This sample, a tissue slice, exposes a large area of tumor tissue, and consists of fatty tissue, IDC and DCIS. The specimen was obtained fresh from surgery and coloured (according to standard pathology protocol) and sliced by the pathologist.

Figure 10b shows a hematoxylin-eosin, HE, stained section of the slice of Figure 10a. A line 1001 encircles the area indicated by the pathologist as IDC and DCIS.

Figure 10c shows the first derivative image of the slice taken at 1225 nm.

Figure 10d shows a tissue classification based on the first derivative using a cut-off value of 0.0486. The slice is indicated by a light grey area 1003. The central grey area 1002 indicates cancer, including both IDC and DCIS. Black areas appear in the sample where there were holes in the sample, or the sample thickness was too thin. As can be seen in Figures 10c and 10d, the large parts of IDC and DCIS were easily classified as tumor tissue. However, smaller DCIS branches were not detected. The width of these DCIS branches was less than 0.5 mm and since the path length of light is higher (2-2.5 mm), these small branches are missed. i.e. light acquired at this location has sampled a volume exceeding the DCIS branches and has picked up information from the surrounding fatty tissue making it look like fatty tissue.

Figure 10e shows a histogram of first derivative values from the test set, and clearly shows that the first derivative at 1225 nm markedly differentiates both IDC and DCIS from healthy tissue, as indicated by arrows indicating the values for the different tissues. A horizontal axis indicates a value from the first derivative around 1225 nm in steps of 0.02 running from 0 to 0.14, and a vertical axis indicates a number of spectra with that value in steps of 10 running from 0 to 80. The dashed line indicates the position of a cut-off point of 0.0486.

Figure 10f shows a ROC curve calculated with both IDC and DCIS spectra labelled as tumor and healthy spectra labelled as healthy tissue. Figure 10f indicates, on a horizontal axis, a false positive rate (1-Specificity) in steps of 0.2 running from 0 to 1, and on a vertical axis, a true positive rate (Sensitivity) in steps of 0.1 running from 0 to 1. Area under the curve, AUC, is 0.998 (confidence interval: 0.982-0.995).

Phantom study

15

Figures 11a, 11b, 11c, 12a, 12b, 12c and 12d illustrate test results in a phantom study using an optical phantom made out of pure coconut oil with embedded drops of Intralipid of different sizes. The coconut oil has a high fat content (close to 100%), Intralipid has a much lower fat content of 20%. This simulates a diffuse medium with inclusions of different sizes.

20

Figure 11a shows a diffuse reflection image of the optical phantom taken at 1210 nm. It clearly shows Intralipid drops of various dimensions.

Figure 11b shows the full spectra of locations 1101 marked with a dot in Figure 11a, together with a spectrum taken from pure Intralipid shown as a dashed line. A horizontal axis indicates a wavelength [nm] in steps of 100 nm running from 900 to 1700 nm. A vertical axis indicates a diffuse reflectance in steps of 0.1 running from 0 to 0.9. As has been recognized, spectra taken on larger Intralipid drops are more similar to the spectrum of pure Intralipid, while the spectra taken on smaller Intralipid drops are more similar to the spectra taken from pure coconut oil (displaying the largest dip at 1210 nm). Spectra taken on smaller intralipid drops indicate a mixture between Intralipid and coconut oil.

Figures 11a and 11b clearly show the effect of the sampling depth. The size of the larger Intralipid drops was taken to be larger than the sampling depth at 1210 nm. Hence the spectra taken at the centre of these spots closely resemble the spectra of pure Intralipid. The spectra taken from the centres of the smaller drops contain information from both the

35

Intralipid drop and their coconut oil environment and hence shows a mixture between the two spectra.

Figure 11c illustrates fat-to-water ratios calculated from the hyperspectral data in two different ways. In the top image, a diffusion model was fitted containing the known absorption spectra of water and fat. In the bottom image, the first derivative image around 1210 nm was calculated. Figure 11c clearly shows the effect of the wavelengths with large penetration depths on the fit of the diffusion model. The Intralipid droplets in the top image show a lower contrast for the smaller droplets.

Figures 12b and 12d show, for the optical phantom, fat-to-water ratios 1201, 1202 calculated with diffusion theory and the associated first derivatives 1203, 1204, respectively, calculated along the trajectories indicated in white lines in Figures 12a and 12c. The trajectories were chosen such that they cross several small Intralipid droplets. In general, the apparent fat content of the Intralipid droplets is much higher when estimated through the fit of the diffusion model. It can be clearly seen that diffusion theory overestimates the fat to water ratio in these small inclusions.

Figures 12a, 12b, 12c and 12d clearly illustrate the effect that the large sampling volume used when applying diffusion theory in a wide wavelength range dilutes the signal from the volume of the Intralipid droplet with a signal from the surrounding coconut oil.

Products, such as software or apparatus, for imaging relative concentrations of light absorbing components/absorbers in (samples of) optically diffuse materials, e.g. find applications in the following fields.

Healthcare:

- Diagnosis of disease by in vivo imaging of tissue components in suspect areas.
- Assessment of resection margins by ex vivo imaging the ratio between fat and water concentrations in resection margins in the resection specimen.
- Assessment of resection margins in vivo during surgery imaging the ratio between fat and water concentrations.
- Evaluation of stool samples for colon cancer risk stratification.

Agro/Food:

- Estimating freshness of fruit, meat or fish by quantitative imaging of essential components.

- Estimating nutritional value of food by quantitative imaging of nutritional components.
- Estimating a vase life of flowers.

Environmental:

- 5
- Quality determination and monitoring of surface water.

Forensic:

- Child abuse: Determination of the age of a bruise by accurate imaging of the spatial distribution of biochemical components occurring in haemoglobin breakdown.
- 10 • Crime scene evaluation: determination of the age of blood stains or other traces of other bodily fluids.
- Forensic pathology: Determination of the time of death by accurate imaging of the distribution of biochemical components occurring in livor mortis.

15 As explained in detail above, in a method and apparatus, a property of an optically diffuse medium comprising a first optical absorber having a first concentration and a second optical absorber having a second concentration is determined. A surface area of the medium is imaged at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber. A reflectance spectrum of the medium at the surface area at the multiple
20 wavelengths is determined. A derivative of the determined reflectance spectrum around the isosbestic wavelength is determined. From the derivative, a concentration ratio of the first concentration and the second concentration is estimated.

As required, detailed embodiments of the present invention are disclosed herein; however, it
25 is to be understood that the disclosed embodiments are merely exemplary of the invention, which can be embodied in various forms. Therefore, specific structural and functional details disclosed herein are not to be interpreted as limiting, but merely as a basis for the claims and as a representative basis for teaching one skilled in the art to variously employ the present invention in virtually any appropriately detailed structure. Further, the terms and phrases used
30 herein are not intended to be limiting, but rather, to provide an understandable description of the invention.

The terms "a"/"an", as used herein, are defined as one or more than one. The term plurality, as used herein, is defined as two or more than two. The term another, as used herein, is
35 defined as at least a second or more. The terms including and/or having, as used herein, are defined as comprising (i.e., open language, not excluding other elements or steps). Any

reference signs in the claims should not be construed as limiting the scope of the claims or the invention.

The mere fact that certain measures are recited in mutually different dependent claims does not indicate that a combination of these measures cannot be used to advantage.

The term coupled, as used herein, is defined as connected, although not necessarily directly, and not necessarily mechanically.

10 A single processor or other unit may fulfil the functions of several items recited in the claims.

The terms software, program, software application, and the like as used herein, are defined as a sequence of instructions designed for execution in a processor of a computer system.

15 A program, computer program, or software application may include a subroutine, a function, a procedure, an object method, an object implementation, an executable application, an applet, a servlet, a source code, an object code, a shared library/dynamic load library and/or other sequence of instructions designed for execution on a computer system.

20 A computer program may be stored and/or distributed on a suitable non-volatile medium, such as an optical storage medium or a solid-state medium supplied together with or as part of other hardware, but also be distributed in other forms, such as via the Internet or other wired or wireless telecommunication systems.

CLAIMS

1. A method of determining a property of an optically diffuse medium, the medium comprising a first optical absorber having a first concentration and a second optical absorber having a second concentration, the method comprising the steps of:

(a) imaging a surface area of the medium at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber;

(b) determining a reflectance spectrum of the medium at the surface area at the multiple wavelengths;

(c) determining a derivative of the determined reflectance spectrum around the isosbestic wavelength; and

(d) estimating a concentration ratio (ψ) of the first concentration and the second concentration from the derivative,

wherein the concentration ratio (ψ) is determined using the partial derivative of a first molar absorption coefficient of the first absorber as a function of the wavelength, and the partial derivative of a second molar absorption coefficient of the second absorber as a function of the wavelength, both taken at a wavelength at which the first molar absorption coefficient and the second molar absorption coefficient are equal.

2. The method according to claim 1, wherein the concentration ratio (ψ) is determined by:

$$\psi = \frac{\frac{\frac{\partial \mu_{a,1}}{\partial \lambda} \Big|_{\lambda_0}}{\Delta \lambda} \cdot \mu_{a,0}}{\frac{\partial \mu_{a,1}}{\partial \lambda} \Big|_{\lambda_0} \cdot \mu_{a,0} - \frac{\partial \mu_{a,2}}{\partial \lambda} \Big|_{\lambda_0} \cdot \mu_{a,0}}{\mu_{a,0}}} = \frac{\frac{\partial \mu_{a,1}}{\partial \lambda} \Big|_{\lambda_0}}{f-g}$$

where:

- f and g are defined by $f = \frac{\partial \mu_{a,1}}{\partial \lambda} \Big|_{\lambda_0}$ and $g = \frac{\partial \mu_{a,2}}{\partial \lambda} \Big|_{\lambda_0}$;

- λ indicates a wavelength;
- λ_0 indicates an isosbestic wavelength;
- $\Delta \lambda$ indicates a wavelength difference from the isosbestic wavelength;
- DX indicates an imaging ratio, defined by subtracting two images taken $\Delta \lambda$ apart from an isosbestic wavelength λ_0 and dividing by the sum of the two images;
- $\mu_{a,m}$ indicates a molar absorption coefficient of absorber m ;
- $\mu_{a,0}$ indicates a molar absorption coefficient at isosbestic wavelength.

3. The method according to any one of the preceding claims, wherein the derivative is any of a first and higher derivatives.

4. The method according to any one of claims 1 to 3, further comprising:
comparing the concentration ratio to a predetermined concentration ratio range; and
indicating a defect for the surface area if the concentration ratio is outside the
concentration ratio range.

5

5. The method according to any one of claims 1 to 3, further comprising:
repeating steps (a) to (d) for different isosbestic wavelengths of the first absorber and
the second absorber in different wavelength regions to obtain estimated concentration ratios
for each one of the isosbestic wavelengths.

10

6. The method according to claim 5, further comprising:
comparing each one of the estimated concentration ratios to a predetermined
concentration ratio range; and
indicating a defect for the surface area when at least one of the concentration ratios is
outside the concentration ratio range.

15

7. The method according to claim 5 or 6, further comprising:
associating each isosbestic wavelength with a respective optical penetration depth;
indicating a defect and a depth thereof for the surface area when at least one of the
concentration ratios is outside the concentration ratio range.

20

8. The method according to any one of the preceding claims, further comprising:
associating each isosbestic wavelength with a respective optical penetration depth; and
if small defects are to be located, selecting an isosbestic wavelength among different
isosbestic wavelengths with a low optical penetration depth.

25

9. An apparatus for determining a property of an optically diffuse medium, the medium
comprising a first optical absorber having a first concentration and a second optical absorber
having a second concentration, the apparatus comprising:

30

a light source configured to illuminate a surface area of a volume of the medium;

a filtering device configured to receive reflected light from the surface area of the
medium and to transmit filtered light to an optical imaging device, wherein the filtering device
is configured to filter multiple wavelengths around an isosbestic wavelength of the first
absorber and the second absorber;

35

a twodimensional, 2D, optical imaging device configured to receive the filtered light from
the filtering device;

an image processing component configured to determine a reflectance spectrum of the medium at the surface area at the multiple wavelengths;

a calculating component configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength; and

5 an estimating component configured to estimate a concentration ratio (ψ) of the first concentration and the second concentration from the derivative,

wherein the concentration ratio (ψ) is determined using the partial derivative of a first molar absorption coefficient of the first absorber as a function of the wavelength, and the partial derivative of a second molar absorption coefficient of the second absorber as a
10 function of the wavelength, both taken at a wavelength at which the first molar absorption coefficient and the second molar absorption coefficient are equal.

10. The apparatus according to claim 9, further comprising:

15 a comparator configured to compare the concentration ratio (ψ) to a predetermined concentration ratio range; and

an output component configured to indicate a defect for the surface area when the concentration ratio is outside the concentration ratio range.

11. The apparatus according to claim 9 or 10, wherein the filtering device comprises:

20 a filter support comprising a plurality of different optical filters; and

an actuator to position each one of the optical filters in an optical path from the surface area of the medium to the optical imaging device.

12. The apparatus according to any one of claims 9 to 11, wherein the optical imaging
25 device comprises a CCD camera, an InGaAs camera or a CMOS camera.

13. An apparatus for determining a property of an optically diffuse medium, the medium comprising a first optical absorber having a first concentration and a second optical absorber having a second concentration, the apparatus comprising:

30 a light source configured to illuminate a surface area of a volume of the medium;

a onedimensional, 1D, or twodimensional, 2D, hyperspectral optical imaging device configured to receive reflected light from the surface area of the medium;

35 an image processing component configured to determine the reflectance spectrum of the medium at the surface area at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber;

a calculating component configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength; and

an estimating component configured to estimate a concentration ratio (ψ) of the first concentration and the second concentration from the derivative,

wherein the concentration ratio (ψ) is determined using the partial derivative of a first molar absorption coefficient of the first absorber as a function of the wavelength, and the partial derivative of a second molar absorption coefficient of the second absorber as a function of the wavelength, both taken at a wavelength at which the first molar absorption coefficient and the second molar absorption coefficient are equal.

14. The apparatus according to claim 13, further comprising:

a stage configured to support the medium; and

an actuator configured to drive the stage to move the medium relative to the optical imaging device.

15. The apparatus according to claim 13 or 14, wherein the optical imaging device is a 2D

hyperspectral optical imaging device comprising an image sensor having a plurality of pixels, and wherein the image processing component is further configured to:

assign a cluster of pixels to the surface area of the medium;

determine the reflectance spectrum for each pixel of the image sensor; and

determine the reflectance spectrum of the medium at the surface area from the

reflectance spectra of the pixels of the cluster.

16. An apparatus for determining a property of an optically diffuse medium, the medium comprising a first optical absorber having a first concentration and a second optical absorber having a second concentration, the apparatus comprising:

a light source configured to illuminate a surface area of a volume of the medium;

a spectrograph configured to receive reflected light from the surface area of the

medium;

a light processing component configured to determine the reflectance spectrum of the medium at the surface area at multiple wavelengths around an isosbestic wavelength of the first absorber and the second absorber;

a calculating component configured to determine a derivative of the determined reflectance spectrum around the isosbestic wavelength; and

an estimating component configured to estimate a concentration ratio (ψ) of the first concentration and the second concentration from the derivative,

wherein the concentration ratio (ψ) is determined using the partial derivative of a first molar absorption coefficient of the first absorber as a function of the wavelength, and the partial derivative of a second molar absorption coefficient of the second absorber as a

function of the wavelength, both taken at a wavelength at which the first molar absorption coefficient and the second molar absorption coefficient are equal.

17. The apparatus according to claim 16, further comprising:
5 a stage configured to support the medium; and
an actuator configured to drive the stage to move the medium relative to the optical imaging device.
18. The apparatus according to claim 16 or 17, further comprising:
10 a scanning device configured to transmit reflected light from different surface areas of the medium to the spectrograph.
19. A computer program comprising computer instructions which, when loaded in a processor, enable the processor to carry out the method of any one of claims 1 to 8.
15
20. A non-volatile storage medium storing computer instructions which, when loaded in a processor, enable the processor to carry out the method of any one of claims 1 to 8.




FIG. 1





FIG. 2

3/12

**FIG. 3**




FIG. 4




FIG. 5




FIG. 6




FIG. 7




FIG. 8a




FIG. 8b










FIG. 8c



FIG. 8d

805








FIG. 11a




FIG. 11b




FIG. 11c




FIG. 12a-d

INTERNATIONAL SEARCH REPORT

International application No
PCT/NL2017/050086

A. CLASSIFICATION OF SUBJECT MATTER
 INV. G01N21/31 G01N21/47 G01N21/49 A61B5/00 G01J3/00
 G01N33/02 G01N33/12 G01N33/18
 ADD.
 According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED
 Minimum documentation searched (classification system followed by classification symbols)
 A61B G01N G01J

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)
 EPO-Internal, WPI Data, BIOSIS

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	US 2003/208111 A1 (MATTU MUTUA [US] ET AL) 6 November 2003 (2003-11-06)	1-8,16, 19,20
Y	paragraphs [0001], [0009] - [0010], [0017], [0018], [0025] - [0039]; claims 1,4,5,9-13; figures 1-5	9-15,17, 18
Y	US 2005/273011 A1 (HATTERY DAVID [US] ET AL) 8 December 2005 (2005-12-08) paragraphs [0143], [0237] - [0238]; figure 5	9-12
Y	US 2008/199080 A1 (SUBBIAH JEYAMKONDAN [US] ET AL) 21 August 2008 (2008-08-21) paragraphs [0004], [0017] - [0023], [0031], [0034], [0035], [0045]; claim 4; figure 1	13-15, 17,18

Further documents are listed in the continuation of Box C.

See patent family annex.

* Special categories of cited documents :

- "A" document defining the general state of the art which is not considered to be of particular relevance
- "E" earlier application or patent but published on or after the international filing date
- "L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)
- "O" document referring to an oral disclosure, use, exhibition or other means
- "P" document published prior to the international filing date but later than the priority date claimed

- "T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention
- "X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone
- "Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art
- "&" document member of the same patent family

Date of the actual completion of the international search 3 May 2017	Date of mailing of the international search report 12/05/2017
--	---

Name and mailing address of the ISA/ European Patent Office, P.B. 5818 Patentlaan 2 NL - 2280 HV Rijswijk Tel. (+31-70) 340-2040, Fax: (+31-70) 340-3016	Authorized officer Duijs, Eric
--	--

INTERNATIONAL SEARCH REPORT

Information on patent family members

International application No

PCT/NL2017/050086

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 2003208111 A1	06-11-2003	AU 7536601 A	24-12-2001
		TW 550064 B	01-09-2003
		US 2003208111 A1	06-11-2003
		WO 0195800 A2	20-12-2001

US 2005273011 A1	08-12-2005	US 2005273011 A1	08-12-2005
		US 2012013779 A1	19-01-2012
		US 2014092288 A1	03-04-2014

US 2008199080 A1	21-08-2008	NONE	
