




(86) **Date de dépôt PCT/PCT Filing Date:** 2010/08/18
(87) **Date publication PCT/PCT Publication Date:** 2011/03/24
(45) **Date de délivrance/Issue Date:** 2016/11/01
(85) **Entrée phase nationale/National Entry:** 2012/03/16
(86) **N° demande PCT/PCT Application No.:** US 2010/045883
(87) **N° publication PCT/PCT Publication No.:** 2011/034681
(30) **Priorité/Priority:** 2009/09/18 (US12/562,694)

(51) **Cl.Int./Int.Cl. A61N 1/378** (2006.01),
A61N 1/372 (2006.01)
(72) **Inventeur/Inventor:**
AGHASSIAN, DANIEL, US
(73) **Propriétaire/Owner:**
BOSTON SCIENTIFIC NEUROMODULATION
CORPORATION, US
(74) **Agent:** PARLEE MCLAWS LLP

(54) **Titre : CHARGEUR EXTERNE UTILISABLE AVEC UN DISPOSITIF MEDICAL IMPLANTABLE AYANT UN POINT DE REGLAGE
DE TEMPERATURE PROGRAMMABLE OU VARIANT DANS LE TEMPS**
(54) **Title: AN EXTERNAL CHARGER USABLE WITH AN IMPLANTABLE MEDICAL DEVICE HAVING A PROGRAMMABLE OR TIME-
VARYING TEMPERATURE SET POINT**



(57) **Abrégé/Abstract:**

An improved external charger for charging the battery within or providing power to an implantable medical device is disclosed. The improved external charger includes circuitry for detecting the temperature of the external charger and for controlling charging to prevent exceeding a maximum temperature. The external charger in some embodiments includes a user interface for allowing a patient to set the external charger's maximum temperature. The user interface can be used to select either constant maximum temperatures, or can allow the user to choose from a number of stored charging programs, which programs can control the maximum temperature to vary over time. Alternatively, a charging program in the external charger can vary the maximum temperature set point automatically. By controlling the maximum temperature of the external charger during charging in these manners, the time needed to charge can be minimized while still ensuring a temperature that is comfortable for that patient.

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property Organization
International Bureau(43) International Publication Date
24 March 2011 (24.03.2011)(10) International Publication Number
WO 2011/034681 A1

(51) International Patent Classification:

A61N 1/378 (2006.01) *A61N 1/372* (2006.01)

(21) International Application Number:

PCT/US2010/045883

(22) International Filing Date:

18 August 2010 (18.08.2010)

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

12/562,694 18 September 2009 (18.09.2009) US

(71) Applicant (for all designated States except US):
BOSTON SCIENTIFIC NEUROMODULATION CORPORATION [US/US]; 25155 Rye Canyon Loop,
Valencia, CA 91355 (US).

(72) Inventor; and

(75) Inventor/Applicant (for US only): **AGHASSIAN, Daniel**
[US/US]; 2503 Allanjay Place, Glendale, CA 91208 (US).(74) Agent: **LEWIS, Terril, G.**; Wong, Cabello, Lutsch,
Rutherford & Brucculeri, L.L.P., 20333 Tomball Park-
way, Suite 600, Houston, TX 77070 (US).(81) Designated States (unless otherwise indicated, for every
kind of national protection available): AE, AG, AL, AM,
AO, AT, AU, AZ, BA, BB, BG, BH, BR, BW, BY, BZ,CA, CH, CL, CN, CO, CR, CU, CZ, DE, DK, DM, DO,
DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT,
HN, HR, HU, ID, IL, IN, IS, JP, KE, KG, KM, KN, KP,
KR, KZ, LA, LC, LK, LR, LS, LT, LU, LY, MA, MD,
ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI,
NO, NZ, OM, PE, PG, PH, PL, PT, RO, RS, RU, SC, SD,
SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR,
TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.(84) Designated States (unless otherwise indicated, for every
kind of regional protection available): ARIPO (BW, GH,
GM, KE, LR, LS, MW, MZ, NA, SD, SL, SZ, TZ, UG,
ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, MD, RU, TJ,
TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK,
EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU,
LV, MC, MK, MT, NL, NO, PL, PT, RO, SE, SI, SK,
SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ,
GW, ML, MR, NE, SN, TD, TG).


Declarations under Rule 4.17:

- as to applicant's entitlement to apply for and be granted a patent (Rule 4.17(ii))
- as to the applicant's entitlement to claim the priority of the earlier application (Rule 4.17(iii))

Published:

- with international search report (Art. 21(3))
- with amended claims (Art. 19(1))

(54) Title: AN EXTERNAL CHARGER USABLE WITH AN IMPLANTABLE MEDICAL DEVICE HAVING PROGRAMMABLE OR TIME-VARYING TEMPERATURE SET POINT

**Figure 5A**

(57) Abstract: An improved external charger for charging the battery within or providing power to an implantable medical device is disclosed. The improved external charger includes circuitry for detecting the temperature of the external charger and for controlling charging to prevent exceeding a maximum temperature. The external charger in some embodiments includes a user interface for allowing a patient to set the external charger's maximum temperature. The user interface can be used to select either constant maximum temperatures, or can allow the user to choose from a number of stored charging programs, which programs can control the maximum temperature to vary over time. Alternatively, a charging program in the external charger can vary the maximum temperature set point automatically. By controlling the maximum temperature of the external charger during charging in these manners, the time needed to charge can be minimized while still ensuring a temperature that is comfortable for that patient.

**AN EXTERNAL CHARGER USABLE WITH AN IMPLANTABLE
MEDICAL DEVICE HAVING A PROGRAMMABLE OR TIME-VARYING
TEMPERATURE SET POINT**

TECHNICAL FIELD

[0002] The present invention relates generally to implantable medical devices, and more particularly, to an external charger for an implantable medical device having programmable temperature regulation.

BACKGROUND ART

[0003] Implantable stimulation devices are devices that generate and deliver electrical stimuli to body nerves and tissues for the therapy of various biological disorders, such as pacemakers to treat cardiac arrhythmia, defibrillators to treat cardiac fibrillation, cochlear stimulators to treat deafness, retinal stimulators to treat blindness, muscle stimulators to produce coordinated limb movement, spinal cord stimulators to treat chronic pain, cortical and deep brain stimulators to treat motor and psychological disorders, and other neural stimulators to treat urinary incontinence, sleep apnea, shoulder subluxation, etc. The present invention may find applicability in all such applications, although the description that follows will generally focus on the use of the invention within a Spinal Cord Stimulation (SCS) system, such as that disclosed in U.S. Patent 6,516,227 (“the ’227 patent”).

[0004] Spinal cord stimulation is a well-accepted clinical method for reducing pain in certain populations of patients. As shown in Figures 1A and 1B, a SCS system typically includes an Implantable Pulse Generator (IPG) 100, which includes a biocompatible case 30. The case 30 usually holds the circuitry and power source or battery necessary for the IPG to function. The IPG 100 is coupled to electrodes 106 via one or more electrode leads (two such leads 102 and 104 are shown), such that the electrodes 106 form an electrode array 110. The electrodes

106 are carried on a flexible body 108, which also houses the individual signal wires 112, 114, coupled to each electrode. The signal wires 112 and 114 are connected to the IPG 100 by way of an interface 115, which may be any suitable device that allows the leads 102 and 104 (or a lead extension, not shown) to be removably connected to the IPG 100. Interface 115 may comprise, for example, an electro-mechanical connector arrangement including lead connectors 38a and 38b configured to mate with corresponding connectors 119a and 119b on the leads 102 and 104. In the IPG 100 illustrated in Figure 1A, there are eight electrodes on lead 102, labeled E1-E8, and eight electrodes on lead 104, labeled E9-E16, although the number of leads and electrodes is application specific and therefore can vary. The electrode array 110 is typically implanted along the dura of the spinal cord, and the IPG 100 generates electrical pulses that are delivered through the electrodes 106 to the nerve fibers within the spinal column. The IPG 100 itself is then typically implanted somewhat distantly in the buttocks of the patient.

[0005] As shown in Figure 2, an IPG 100 typically includes an electronic substrate assembly 14 including a printed circuit board (PCB) 16, along with various electronic components 20, such as microprocessors, integrated circuits, and capacitors, mounted to the PCB 16. Ultimately, the electronic circuitry performs a therapeutic function, such as neurostimulation. A feedthrough assembly 24 routes the various electrode signals from the electronic substrate assembly 14 to the lead connectors 38a, 38b, which are in turn coupled to the leads 102 and 104 (see Figures 1A and 1B). The IPG 100 further comprises a header connector 36, which among other things houses the lead connectors 38a, 38b. The IPG 100 can further include a telemetry antenna or coil (not shown), which can be mounted within the header connector 36, for receipt and transmission of data to an external device such as a hand-held or clinician programmer (not shown). As noted earlier, the IPG 100 usually also includes a power source, typically a rechargeable battery 26.

[0006] Also shown in Figure 2 is an external charger 12 that is used to provide power to the IPG 100, which is explained in further detail below. The external charger 12 itself needs power to operate, and therefore may include its own battery 70, which may also be a battery that is rechargeable using a plug-in-the-

wall holster (“cradle”) or power cord connection much like a cellular telephone. Alternatively, the external charger 12 may lack a battery and instead draw its power directly from being plugged into a wall outlet (not shown). In any event, a primary function of the charger 12, as discussed further below, is to energize a charging coil 17. The external charger 12 can contain one or more circuit boards 72, 74, which contain the circuitry 76 needed to implement such functionality. In a preferred embodiment, and as shown in Figure 2, most of the circuitry 76 can be located on an orthogonal circuit board 74, which reduces interference and heating that might be produced by the charging coil 17, as is further explained in U.S. Patent Application Publication No. US 2008/0027500.

[0007] Further details concerning the structure and function of typical IPGs and IPG systems are disclosed in U.S. Patent No. 7,444,181.

[0008] If the battery 26 in the IPG 100 is rechargeable, it will be necessary to charge the battery 26 periodically using the external charger 12, i.e., a charger that is external to the patient in whom the IPG 100 is implanted. Because the IPG 100 may already be implanted in a patient, wireless recharging is greatly preferred to obviate the need to replace a power-depleted battery 26 via surgery.

[0009] To convey energy wirelessly between the external charger 12 and the IPG 100, and as shown in Figure 2, the charger 12 typically includes an energized alternating current (AC) coil 17 that supplies energy 29 to a similar charging coil 18 located in or on the IPG 100 via inductive coupling. In this regard, the coil 17 within the external charger 12 is wrapped in a plane 50, which lies substantially parallel to the plane 52 of the coil 18 within the IPG 100, as shown schematically in Figure 3. Such a means of inductive energy transfer can occur transcutaneously, i.e., through the patient’s tissue 25. The energy 29 received by the IPG 100’s coil 18 can then be rectified and stored in a rechargeable battery 26 within the IPG 100, which in turn powers the electronic circuitry that runs the IPG 100. Alternatively, the energy 29 received can be used to directly power the IPG 100’s electronic circuitry, which may lack a battery altogether.

[0010] Conventional external chargers 12 typically employ relatively simple user interfaces 94, which simplicity is warranted either because of the relative simplicity of the charging function, or because the external charger 12 may not be

visible to the patient while in use, which limits the utility of more complex visual user interfaces. For example, in an SCS application in which the IPG 100 is typically implanted in the buttocks, the external charger 12 is generally behind the patient while charging to align the external charger 12 with the IPG 100. Additionally, the external charger 12 may be covered by clothing while in use, again reducing the utility of a visual user interface. The user interface 94 of the conventional external charger 12 of Figure 2 therefore typically merely comprises an on/off switch that activates the charger 12, an LED to indicate the status of the on/off switch, and a speaker for emitting a “beep” at various times, such as when the charger is not properly aligned with the IPG 100 or when charging has completed.

[0011] Inductive charging between the two coils 17, 18 can produce significant heating in the external charger 12. Such external charger heating could, if unchecked, possibly discomfort or injure the patient. This possibility of injury is heightened because the external charger 12 is often held against the patient’s tissue 25 during charging. For example, in an SCS system, the external charger 12 is generally held in place against the buttocks of the patient by a “fanny pack.”

[0012] Accordingly, prior art external chargers have incorporated temperature monitoring and control circuitry to detect external charger temperatures, and to control charging accordingly. For example, and as shown in Figure 2, a prior art external charger 12 can include one or more temperature sensors 92, which for example can comprise thermistors or thermocouples affixed by heat conducting epoxy to the housing of the external charger 12. A hole 90 in the circuit board 72 can assist in connecting the temperature sensor 92 to the temperature sensing circuitry (not shown) resident on either of circuit boards 72 or 74. The temperature monitoring and control circuitry generally senses the temperature, $T(EC)$, of the external charger, and in particular sets a maximum temperature, $T_{max}(EC)$ for the external charger. The maximum temperature $T_{max}(EC)$ may be set to 41 °C (~106 °F) for example, which temperature is conservatively picked by the manufacturer of the external charger 12 as a temperature that should not discomfort or injure a normal healthy adult.

[0013] The temperature monitoring and control circuitry in the external charger 12 can operate as illustrated in Figure 4, which shows the temperature of the external charger, $T(EC)$, during a typical charging session. Initially, the charging circuitry in the external charger 12 is enabled, i.e., an AC current flows through coil 17 in the external charger 12 as previously discussed. As this occurs, $T(EC)$ increases. Eventually, $T(EC)$ equals $T_{max}(EC)$. At this point, the temperature sensing circuitry would inform the microcontroller in the external charger 12 to suspend charging, i.e., to cease current flow through coil 17. Once the current ceases, the $T(EC)$ will start to fall. At some point—for example after some time duration or when a minimum $T(EC)$ ($T_{min}(EC)$) is reached as illustrated—charging can be enabled until once again $T(EC)$ reaches $T_{max}(EC)$, etc. The result is that charging is duty cycled between enabled and disabled states.

[0014] Although the charging scheme illustrated in Figure 4 ensures that the external charger 12 never exceeds a predefined maximum safe temperature, $T_{max}(EC)$, the inventors consider such scheme non-optimal, because it fails to allow for differences between patients, and does not provide any way to control external charger 12 heating characteristics. For example, if a patient is not particularly heat sensitive, that patient may be able to tolerate a higher $T_{max}(EC)$, such as 42 °C for example. However, if $T_{max}(EC)$ is constrained to 41 °C by the manufacture of the external charger 12, charging will not take place as aggressively as that patient could tolerate: the current in the charging coil 17 would be limited, or charging would be suspended for greater amounts of time. In either case, the result is that charging will be performed too slowly for that patient. This is inconvenient, as patients would generally like charging to occur as quickly as possible. On the other hand, if a patient is unusually heat sensitive for some reason, perhaps because of a medical condition, that patient might be more comfortable with a lower $T_{max}(EC)$, say 40 °C for example. In this case, if $T_{max}(EC)$ were constrained to 41 °C by the manufacture, that patient would perceive charging as uncomfortably warm.

BRIEF DESCRIPTION OF DRAWINGS

[0015] Figures 1A and 1B illustrate an implantable pulse generator (IPG), and the manner in which an electrode array is coupled to the IPG in accordance with the prior art;

[0016] Figure 2 illustrates the IPG in relation to an external charger in accordance with the prior art;

[0017] Figure 3 illustrates the relationship of the charging coils in the external charger and in the IPG during charging in accordance with the prior art;

[0018] Figure 4 illustrates regulation of the external charger's temperature during IPG battery charging in accordance with the prior art;

[0019] Figures 5A and 5B illustrate two embodiments of an improved external charger that allow adjustment of temperature settings for the external charger;

[0020] Figure 6 illustrates temperature monitoring and control circuitry for one embodiment of the external chargers of Figures 5A and 5B;

[0021] Figure 6A illustrates temperature monitoring and control circuitry for another embodiment of the external chargers of Figures 5A and 5B;

[0022] Figure 7 illustrates regulation of the external charger's temperature during IPG battery charging in accordance with the embodiment illustrated in Figure 6A;

[0023] Figure 8 illustrates temperature monitoring and control circuitry for an external charger having a stored time-varying charging temperature program;

[0024] Figure 9 illustrates regulation of the external charger's temperature during IPG battery charging in accordance with the embodiment illustrated in Figure 8;

[0025] Figure 10 illustrates another example of regulation of the external charger's temperature during IPG battery charging in accordance with the embodiment illustrated in Figure 6A;

[0026] Figures 11A and 11B illustrate two embodiments of an improved external charger that allow the patient to select between multiple charging temperature programs;

[0027] Figure 12 illustrates a schematic of the temperature monitoring and control circuitry for the external chargers of Figures 11A and 11B;

[0028] Figure 13 illustrates regulation of the external charger's temperature during IPG battery charging in accordance with the embodiments illustrated in Figures 11A and 11B;

[0029] Figure 14 illustrates a schematic of the temperature monitoring and control circuitry for another embodiment of an external charger;

[0030] Figures 15-17 illustrates different examples of regulation of the external charger's temperature during IPG battery charging for the embodiment of Figure 14; and

[0031] Figure 18 illustrates programming the temperature setting of an external charger using the user interface of a different device such as an external controller or a computer.

DESCRIPTION OF EMBODIMENTS

[0032] An improved external charger for charging the battery within or providing power to an implantable medical device is disclosed. The improved external charger includes circuitry for detecting the temperature on the portion of the external charger that is applied to the patient and for controlling charging to prevent a maximum temperature from being exceeded. The external charger in some embodiments includes a user interface for allowing a patient to set the maximum temperature for the external charger. The user interface can be used to select either constant maximum temperatures, or can allow the user to choose from a number of charging programs, which programs can control the maximum temperature to vary over time. Alternatively, a charging program in the external charger can vary the maximum temperature set point automatically. By controlling the maximum temperature of the external charger during charging in these manners, the time needed to charge can be minimized while still ensuring a temperature that is comfortable for that patient.

[0033] The disclosed embodiments are not limited to use with an implantable pulse generator (IPG) system, but more generically are usable with any implantable medical device system in which an implanted medical device benefits

or requires wireless charging from an external source. For example, the disclosed embodiments may be used as part of a system involving a pacemaker, defibrillator, cochlear stimulator, retinal stimulator, a stimulator configured to produce coordinated limb movement, a cortical and deep brain stimulator, or in any other neural stimulator system configured to treat urinary incontinence, sleep apnea, shoulder subluxation, etc. The disclosed embodiments may also be used as part of a system in which the implanted medical device comprises a sensor or an active device not involving electrical stimulation (e.g., a drug pump).

[0034] Figures 5A and 5B illustrate two embodiments of an external charger 200 that allows adjustment of maximum temperature set point ($T_{\max}(\text{EC})$) for the external charger 200. In the embodiment of Figure 5A, a slide switch 210 user interface allows the patient to switch the external charger 200 between Fast, Normal, and Comfort settings (denoted as H, M, and L respectively in Figure 5A). Figure 5B provides a different user interface, including a display 215 and buttons for allowing a “charge Rate” to be selected. Although indicated to the user as an “Rate” setting, the reality underlying the setting is an adjustment to $T_{\max}(\text{EC})$ (and possibly also an adjustment to $T_{\min}(\text{EC})$). Of course, text accompanying the user interfaces could also reference “temperature” if that would be more intuitive to the patient, and could even reference the actual settings for $T_{\max}(\text{EC})$ (e.g., “42 °C,” “41 °C,” “40 °C.”). Although not illustrated in Figure 5B, the user interface might alternatively allow the patient to increment or decrement the charging intensity/temperature setting (e.g., in 0.5 °C increments) without having the patient choose an absolute setting. Although three settings are shown in Figures 5A, 5B, other numbers of settings can be provided and designated in any way desirable. Additionally, other user interfaces can be used beyond those depicted for selecting the maximum temperature, and no particular user interface is important to the implementation.

[0035] Figure 6 illustrates an embodiment of a temperature monitoring and control circuitry 250 useable in the external charger 200 of Figures 5A and 5B. By way of overview, temperature monitoring and control circuitry 250 operates to sense the temperature, $T(\text{EC})$, of the applied portion of the external charger 200, and to control the temperature between $T_{\max}(\text{EC})$ and $T_{\min}(\text{EC})$ by selectivity

enabling or disabling the charging coil 17 during charging. In this embodiment, two thermistors 602, 612 allow monitoring the temperature of more than one portion of the external charger 200, biased by resistors 604, 614. Although two thermistors 602, 614 are illustrated in Figure 6, the number of thermistors is illustrative only and any number of thermistors can be employed to monitor as many portions of the external charger 200 as desired.

[0036] The voltages produced by thermistors 602, 604 are converted to digital values by an analog to digital logic 620 of the microcontroller 251. The logic 620 can be incorporated into the microcontroller 251 as illustrated, or can be a separate logic external to the microcontroller as desired. The $T_{max}(EC)$ and $T_{min}(EC)$ values are then stored by the microcontroller 251, in any desired manner, such as in a memory (not shown) or in registers of the microcontroller 251. If the digitized voltage from thermistors 602, 612 exceeds $T_{max}(EC)$, then firmware in the microcontroller 251 can disable the charging coil 17 or reduce the charging rate. Similarly, if the digitized voltage from thermistors 602, 612 is less than $T_{min}(EC)$, then the firmware can enable the charging coil 17 or increase the charging rate. Direct digital synthesis logic 630 and amplifier logic 232 can be used to control the charging coil 17 by the microcontroller 251.

[0037] Figure 6A illustrates another embodiment of a temperature monitoring and control circuitry 250 useable in the external charger 200 of Figures 5A and 5B. By way of overview, temperature monitoring and control circuitry 250 operates to sense the temperature, $T(EC)$, of the applied portion of the external charger 200, and to control the temperature between $T_{max}(EC)$ and $T_{min}(EC)$ by selectivity enabling or disabling the charging coil 17 during charging. In the circuit of Figure 6A, $T_{max}(EC)$ and $T_{min}(EC)$ are offset by a predetermined amount, and so the patient's selection at the user interface of a particular charging rate/temperature acts to set both $T_{max}(EC)$ and $T_{min}(EC)$. However, this is not strictly necessary, and in other circuit implementations, $T_{max}(EC)$ and $T_{min}(EC)$ can each be separately selected by the patient, or only $T_{max}(EC)$ can be selected with $T_{min}(EC)$ being preset and not selectable. However, temperature monitoring and control circuits implementing such variations are not shown for simplicity, but, as

one skilled in the art will appreciate, would be easily implemented given the circuitry details that follow.

[0038] The patient uses the user interface (either slide switch 210, or display/buttons 215/220 for example), to input a charging rate/temperature. This indication is sent to the external charger 200's microcontroller 251, which in turn generates an associated control signal, such as L, M, or H (which, for ease of illustration, matches the low, medium, and high user interface selections in Figure 5A). These control signals are in turn sent to a variable resistive network 93, whose functionality will be described further below.

[0039] Temperature monitoring and control circuitry 250 further comprises two voltage dividers 258 and 259, a Tmax control circuit 260, and a Tmin control circuit 261. Voltage divider 258 includes thermistor 92 (see Figure 2), and two resistors R1 and R2. These elements, in conjunction with the external charger 200's battery voltage (Vdd), generate two control signals: Vbias(max), which is sent to Tmax control circuit 260, and Vbias(min), which is sent to Tmin control circuit 261. Voltage divider 259 includes the variable resistive network 93 and resistor R3, which elements generate a reference voltage, Vref, which is sent to both control circuits 260 and 261. As will be discussed further below, the relationship between Vbias(max) and Vref sets Tmax(EC), while the relationship between Vbias(min) and Vref sets Tmin(EC). These Tmax(EC) and Tmin(EC) temperatures are by default set by the various resistors to safe values that will not injure or cause discomfort to most patients.

[0040] Tmax control circuit 260 comprises a comparator 252 for receiving Vbias(max) and Vref, a transistor 252, and a pull up resistor R4. If Vbias(max) > Vref, comparator 252 output a logic 1, which turns on the transistor 256. This overcomes the effect of pull up resistor R4, and thus pulls control signal Tmax to a logic 0. By contrast, if Vbias(max) < Vref, then transistor 256 is off, and Tmax is pulled to a logic 1 by the pull up resistor R4. Generally speaking, Tmax indicates to the microcontroller 251 whether to enable charging (Tmax = 0) or disable charging (Tmax = 1), as will be discussed further below. Tmin control circuitry 261 is constructed and operates similarly to the Tmax control circuit 260, and thus such details are not reiterated.

[0041] At the onset of charging, and before the external charger 200 has had time to heat up, $V_{bias(max)}$ is set to be greater than V_{ref} . As just discussed, this sets a condition $T_{max} = 0$, which indicates to charge enable circuitry 230 associated with microcontroller 251 that charging can be enabled, i.e., that current can flow through coil 17. Charge enable circuit 230 is shown as part of the microcontroller 251, but it can be separate therefrom. As $T(EC)$ increases during charging, the resistance of thermistor 92 increases, causing the $V_{bias(max)}$ voltage to decrease. When the temperature $T(EC)$ reaches $T_{max}(EC)$, $V_{bias(max)}$ becomes less than V_{ref} , which sets $T_{max} = 1$. This indicates to the charge enable circuitry 230 that $T_{max}(EC)$ has been reached, and therefore that charging should be disabled. Rising edges of the T_{max} may be latched in the charge enable circuit 230 to ensure that the coil 17 remains disabled, and does not become enabled as soon as the external charger 200 cools to slightly below $T_{max}(EC)$, which would set the T_{max} line to a logic 0 again.

[0042] Once charging is disabled and the external charger 200 begins to cool, it will eventually be suitable to enable charging once again, and such is the function of T_{min} control circuit 261. In T_{min} control circuit 261, a comparator 253 compares V_{ref} and $V_{bias(min)}$. Initially, i.e., once cooling begins, $V_{ref} > V_{bias(min)}$, comparator 253 outputs a logic 1, and the T_{min} control signal is drawn to 0 by transistor 257, indicating that charging should not be enabled. As $T(EC)$ continues to fall, the resistance of the thermistor 92 decreases, and $V_{bias(min)}$ increases. Eventually, when $T(EC)$ reaches $T_{min}(EC)$, $V_{bias(min)} > V_{ref}$, causing comparator 253 to output a logic 0, which sets $T_{min} = 1$. This indicates to charge enable circuitry 230 that $T_{min}(EC)$ has been reached, and therefore that charging should once again be enabled. As with T_{max} , only rising edges of control signal T_{min} are latched at the charge enable circuit 230, avoiding disabling charging of the coil 17 as soon as the temperature $T(EC)$ rises slightly above $T_{min}(EC)$.

[0043] The default setting of $T_{max}(EC)$ in one embodiment is set to 41 °C while $T_{min}(EC)$ is set to 39 °C, although these values can be tailored to suit a particular implementation. As one skilled in the art will realize, setting such default values can be achieved through setting the varying resistances in the

voltage dividers 258 and 259, which is a matter of routine design. The default values in one embodiment can be set at a medium charging intensity/temperature, i.e., when $M = 1$.

[0044] Adjustment to $T_{\max}(\text{EC})$ (and $T_{\min}(\text{EC})$ in the illustrated example of temperature monitoring and control circuitry 250) occurs via control signals L, M, and H, which as noted above result from the patient's charging intensity/temperature selection at the user interface of the external charger 200. The goal of these control signals is to affect V_{ref} by affecting the resistance of the variable resistive network 93 in voltage divider 259. In this regard, both $T_{\max}(\text{EC})$ and $T_{\min}(\text{EC})$ are raised when V_{ref} is lowered, i.e., when the resistance of the variable resistive network 93 is raised. Conversely, $T_{\max}(\text{EC})$ and $T_{\min}(\text{EC})$ are lowered when V_{ref} is raised, i.e., when the resistance of the variable resistive network 93 is lowered.

[0045] An example variable resistive network 93 for achieving such functionality is shown in Figure 6A, which comprises three resistors of resistance R , $2R$, and $3R$ respectively gated by the H, M, and L control signals. When the patient selects the H setting, a transistor shorts out resistor R , and the variable resistive network 93's total resistance is $2R + 3R = 5R$, a relatively high resistance which results in a lower V_{ref} , and thus a higher $T_{\max}(\text{EC})/T_{\min}(\text{EC})$. When the patient selects the M setting, resistor $2R$ is shorted, providing a total resistance of $R + 3R = 4R$, which smaller resistance ultimately lowers $T_{\max}(\text{EC})/T_{\min}(\text{EC})$. When the patient selects the L setting, resistor $3R$ is shorted, providing an even smaller resistance of $R + 2R = 3R$, which lowers $T_{\max}(\text{EC})/T_{\min}(\text{EC})$ even further. However, it should be noted that the illustrated resistive network 93, and its control signals, are merely examples, and that other circuits and control signals could be used. Moreover, other techniques not involving resistive networks can be used to affect V_{ref} , and hence $T_{\max}(\text{EC})$ and $T_{\min}(\text{EC})$. For example, a patient's charging intensity/temperature selection at the user interface can be stored and used to set the output voltage of a bandgap reference voltage generator. Such an embodiment would dispense with the need for a voltage divider 259.

[0046] Once $T_{\max}(\text{EC})$ has been set by the user, it is preferably stored in a non-volatile temperature-parameter memory 275 in the external charger 200

(Figure 6A), which memory may be located in the microcontroller 251 or external to the microcontroller 251.

[0047] Figure 7 illustrates operation of the temperature monitoring and control circuitry 250 of Figure 6A for the patient-selected charging intensities/temperatures of high, medium, and low (H, M, and L). As illustrated in each of the graphs, the external temperature $T(EC)$ rises from its initial state to $T_{max}(EC)$ then oscillates between $T_{max}(EC)$ and $T_{min}(EC)$ for the remainder of the charging session. Note that the maximum temperature set points decrease with the patient selected setting ($T_{max1}(EC) > T_{max2}(EC) > T_{max3}(EC)$). (The $T_{min}(EC)$ set points also decrease in this embodiment, but as noted earlier that is not strictly necessary). Moreover, note that higher T_{max} values reduce the time necessary to complete charging ($t1 < t2 < t3$). This is understandable, because a higher permissible temperature will allow for a higher power output from the external charger 200 (e.g., a higher current in coil 17), which in turn reduces the time necessary to charge the battery 26 in the IPG 100 (see Figure 2). As a result, charging can be optimized for each particular patient: i.e., a $T_{max}(EC)$ can be chosen which is small enough to be comfortable, but larger enough so that the time necessary to charge is minimized.

[0048] In another embodiment illustrated in Figure 8, the external charger 200 varies $T_{max}(EC)$ (and again in this example, $T_{min}(EC)$) over time. This occurs not by a patient selection at the user interface but instead by automatically executing a temperature program ("temp prog 1") at the microcontroller 251 upon initiating a charging session.

[0049] The stored temperature program outputs control signals S1-S3 at particular times, as shown to the right on Figure 8. In the illustrated example, the temperature program asserts a control signal S1 for a time $t1$, followed by control signal S2 for a time $t2$, and followed by control signal S3 for a time $t3$. The S1, S2, and S3 control signals control the variable resistive network 93 similarly to the H, M, and L control signals of Figure 6A. Thus, and as shown in Figure 9, control signals S1-S3 set a high maximum temperature set point ($T_{max1}(EC)$) during $t1$, a medium $T_{max2}(EC)$ during time $t2$, and a low $T_{max3}(EC)$ during time $t3$.

(Again, the corresponding minimum temperatures $T_{min1}(EC)$ - $T_{min3}(EC)$ are also decreased, but this is not strictly necessary).

[0050] This variable-temperature program is logical from consideration of heat loading and comfort of the patient's tissue. Initially, the program chooses a high $T_{max1}(EC)$ on the assumption that the patient's tissue has not yet been subject to any heat, and therefore can probably tolerate relatively intense heat from the external charger 200 for at least a short while (t_1). During this short while, charging of the IPG's battery would be beneficially accelerated. As the tissue heats up, the patient may eventually no longer tolerate the relative high maximum temperature of $T_{max1}(EC)$. Therefore, the maximum temperature is decreased to $T_{max2}(EC)$ for a time t_2 . Although this lower temperature would tend to reduce discomfort, it would also lengthen the time necessary to charge the IPG's battery, and therefore t_2 may be greater than t_1 as shown (although this is not strictly necessary). For the same reasons, the maximum temperature can again be reduced to $T_{max3}(EC)$, although again this may lengthen the time necessary to charge (i.e., $t_3 > t_2 > t_1$).

[0051] The parameters of the temperature program (i.e., $T_{max}(EC)$ - $T_{max3}(EC)$, t_1 - t_3) could be specified and stored by the manufacture of the external charger 200. Alternatively, such parameters could be defined and stored by the patient after manufacture using the external charger's user interface. Although the temperature program could be implemented in microcode resident in the external charger's microcontroller 251, it could also easily be implemented using discrete circuitry components as one skilled in the art having benefit of this disclosure will appreciate.

[0052] Although Figure 9 illustrates a temperature program that varies the charging intensity/temperature in discrete intervals, other temperature programs executable at the microcontroller 251 can effect a smooth variance of $T_{max}(EC)$ (and $T_{min}(EC)$), as shown in Figure 10. Such smoothness can be achieved in many ways, again as one skilled in the art having benefit of this disclosure will appreciate.

[0053] Figures 11A and 11B illustrate examples of an improved external charger 200 in which the patient can select one of a plurality of predefined

external charger temperature programs, each of which varies $T_{\max}(\text{EC})$ in different ways. Similar to an earlier example (see Figures 5A and 5B), the user interface can employ a simple slide switch 210 or a display/buttons 215/220 to select between the temperature programs P1, P2, and P3. As illustrated in Figure 11B, the display 215 may provide a description of each program (“[desc n]”) to inform the patient about the basic features of the programs so that the user can make a meaningful selection. As before, the illustrated user interfaces are not mandatory, and other forms could be used.

[0054] Figure 12 illustrates the temperature monitoring and control circuitry 250 for the external chargers 200 of Figures 11A and 11B. Many of the elements depicted in Figure 12 were discussed in previous embodiments, and such details are not repeated here. However, of particular note to this embodiment, a plurality of temperature programs (P1–Pn) are stored for execution by the microcontroller 251. As with the embodiment of Figure 8, execution of a given program results in the issuance of a plurality of control signals S1–Sn. Consistent with earlier examples, the control signals are sent to voltage divider 259 to affect its resistance, which ultimately affects V_{ref} , and hence affects (at least) $T_{\max}(\text{EC})$. However, as before, a voltage divider 259 is not strictly needed; for example, the control signals S1–Sn could be sent to a bandgap reference voltage generator 259 instead. .

[0055] To illustrate operation of the circuit 250 of Figure 12, assume the patient can choose between three temperature programs as illustrated in Figure 13, each of which sets the $T_{\max}(\text{EC})/T_{\min}(\text{EC})$ values differently. In a first example program (P1), the external charger 200 sets the $T_{\max 1}(\text{EC})$ value to 41 °C and the $T_{\min 1}(\text{EC})$ value to 39 °C, and charges using that temperature range for the entire charging session. In a second example program (P2), the external charger 200 initially sets $T_{\max 2}(\text{EC})$ to 43 °C and $T_{\min 2}(\text{EC})$ to 41 °C, then continuously decreases those values over time. In a third example program (P3), the external charger 200 initially sets the $T_{\max 3}(\text{EC})/T_{\min 3}(\text{EC})$ values at 40 °C/38 °C for the first five minutes, then lowers them to 38 °C/36 °C for the next ten minutes, and finally to 37 °C/35 °C for the remainder of the charging session.

[0056] Although the above-described embodiments allow for adjustment of the $T_{\max}(\text{EC})$ and $T_{\min}(\text{EC})$ temperature settings together, other embodiments may allow separate control over each setting as previously noted, or may allow control only over $T_{\max}(\text{EC})$. Still other embodiments do not require the use of a $T_{\min}(\text{EC})$ at all, such as with the temperature monitoring and control circuitry 250 of Figure 14. In this embodiment, the $T_{\max}(\text{EC})$ and $T_{\min}(\text{EC})$ temperature values are effectively combined into a single $T_{\text{set}}(\text{EC})$ value. Similarly to the embodiment of Figure 6A, voltage dividers 258 and 259 in conjunction with T_{set} control circuitry 282 derive a control signal T_{set} , which signal is sent to the charge enable circuitry 230. Here, T_{set} is latched on both rising and falling edges at the charge enable circuitry 230, and thus will oscillate between a logic 1 when $T(\text{EC}) > T_{\text{set}}(\text{EC})$, which will disable charging, and a logic 0 when $T(\text{EC}) < T_{\text{set}}$, which will enable charging. To prevent the circuit 250 from switching too quickly between the enabled and disabled states, the charge enable circuitry can employ any of a number of digital filtering or digital integration techniques.

[0057] Figure 15 illustrates operation of the temperature monitoring and control circuitry 250 of Figure 14. When the external charger 200 exceeds $T_{\text{set}}(\text{EC})$, charging is suspended until the temperature falls below $T_{\text{set}}(\text{EC})$, at which point charging is resumed. Thus, the temperature $T(\text{EC})$ of the external charger 200 oscillates around the $T_{\text{set}}(\text{EC})$ value.

[0058] Use of a single temperature set point, $T_{\text{set}}(\text{EC})$, can also be used in conjunction with embodiments having temperature programs stored in the external charger, such as the single program example of Figures 8-10 and the multiple program example of Figures 11A-13. Figures 16 and 17 respectively illustrate operation of the temperature monitoring and control circuitry 250 of those program-based embodiments as modified per Figure 14 to provide only a single control signal, T_{set} , to the charge enable circuitry 230.

[0059] In one embodiment, the external charger 200 suspends charging after a charging session of four hours of charging, or whenever the IPG 100 indicates by back telemetry that the IPG 100 is fully charged. This maximum length for a charging session can alternatively be made adjustable by the patient, using the external charger's user interface for example. If the suspension occurs before the

[0001] IPG 100 is completely charged, the external charger can indicate via the user interface that the charging was incomplete, as a reminder to the patient to continue charging later.

[0002] In embodiments described to this point, the programming of the $T_{max}(EC)/T_{min}(EC)$ or $T_{set}(EC)$ temperatures into the improved external charger 200 occurs by use of a user interface on the external charger 200. However, this is not strictly necessary. For example, in Figure 18, the relevant user interface appears on an external controller 300, which is traditionally used by the patient to control the implant's therapy settings. In the example shown, the external controller 300's user interface comprises a display and buttons, and is thus similar to the user interface of Figure 5B described earlier. Once the charging rate/temperature setting is entered into the external controller 300, it can be wirelessly transmitted to the external charger 200 and stored in its memory 275 (Figure 6A). Such wireless transmission may comprise a short-range communication link such as Bluetooth for example, as described in further detail in U.S. Patent Application Publication No. 2010-0305663. Alternatively, the charging intensity/temperature can be programmed into the external charger 200 by using a wired connection to the patient's or clinician's computer 350 for example. In either case, the patient's ability to program the temperature after manufacture is preserved.

[0003] Although it is envisioned that the disclosed external chargers 200 would normally be used to charge a battery 26 within the IPG 100, the external charger 200 can also be used with IPGs or other implantable medical devices that lack a battery. This could occur for example in a system in which the IPG continually wirelessly receives energy from the external charger 200, which IPG in turn rectifies and uses this energy without storage.

[0004] Although in a preferred embodiment the external charger 200 employs a thermistor, other temperature sensors devices may be used, such as thermocouples, resistance temperature detectors (RTDs), semiconductor junction circuits, and/or circuits employing such devices.

[0063] The various examples of temperature monitoring and control circuitry 250 are used to enable or disable charging in conjunction with consideration of the set points $T_{\max}(\text{EC})/T_{\min}(\text{EC})$ or $T_{\text{set}}(\text{EC})$. However, such enabling or disabling of the external charger is not strictly required to control the external charger 200's temperature. For example, instead of completely disabling the external charger 200 when $T_{\max}(\text{EC})$ is exceeded, the charge enable circuitry 230 (Figure 6A) could instead reduce the current flow through the charging coil 17, instead of curtailing it altogether. Alternatively, the charge enable circuitry 230 could start to duty cycle current to the coil 17. In short, there are many different ways the power output of the external charger 200 could be controlled to keep the external charger's temperature in line with the temperature set points disclosed herein, and the disclosed techniques are not limited to the embodiments illustrated.

[0064] Furthermore, although the embodiments described above provide for a user interface to allow a patient to program the external charger 200, in other embodiments, the user interface can be omitted, allowing for programming the external charger 200 only by the factory. Alternatively, a clinician interface can be provided to allow a clinician to program the external charger 200, even if the external charger 200 has no user interface for patient programming of the external charger 200.

[0065] Use of the term "programming" should not be interpreted as requiring the ability to program software that controls the external charger 200, but should be understood to include any technique for controlling or modifying the functionality of the external charger 200 to modify its operating characteristics, such as maximum temperature settings, including hardware and circuitry techniques for accomplishing such control or modification.

[0066] In summary, an external charger 200 permits a patient to control the maximum temperature of the external charger 200 while charging. This controlled temperature regulation can provide faster charging with improved comfort and safety.

[0067] Although particular embodiments of the present invention have been shown and described, it should be understood that the above discussion is not intended to limit the present invention to these embodiments. It will be obvious to

those skilled in the art that various changes and modifications may be made without departing from the spirit and scope of the present invention. Thus, the present invention is intended to cover alternatives, modifications, and equivalents that may fall within the spirit and scope of the present invention as defined by the claims.

We Claim:

1. A system
an implantable medical device, comprising:
a first coil for receiving energy to provide power for operating the implantable medical device; and
an external charger, comprising:
a second coil for producing energy receivable by the implantable medical device to power or charge the implantable medical device; and
temperature monitoring and control circuitry for sensing a temperature of the external charger and controlling activation of the second coil in accordance with at least one temperature set point,
wherein the temperature set point is programmable to automatically vary over time during a charging session, and
wherein the temperature monitoring and control circuitry controls activation of the second coil in accordance with a maximum temperature set point and a minimum temperature set point.
2. The system of claim 1, wherein the temperature monitoring and control circuitry controls activation of the second coil by enabling and disabling operation of the second coil.
3. The system of claim 2, wherein the second coil is disabled when a maximum temperature set point is reached.
4. The system of any one of claims 1 to 3, wherein the temperature monitoring and control circuitry controls activation of the second coil by adjusting the charging current through the second coil.

5. The system of any one of claims 1 to 4, wherein the temperature monitoring and control circuitry disables activation of the second coil after a predetermined charging time.

6. The system of any one of claims 1 to 5 further comprising an external controller configured to allow a patient to control therapy settings in the implantable medical device and wherein the external controller comprises a user interface configured to allow a user to provide the temperature set point to the external charger.

7. An external charger for use with an implantable medical device, comprising:

a coil for producing energy receivable by the implantable medical device to power or charge the implantable medical device;

temperature monitoring and control circuitry for detecting a temperature of the external charger and controlling activation of the coil in accordance with a temperature set point; and

memory comprising a temperature program for indicating the temperature set point to the temperature monitoring and control circuitry during a charging session, wherein the temperature set point is configured to vary during a charging session.

8. The external charger of claim 7, wherein the temperature set point decreases during the charging session.

9. The external charger of claim 7 or 8, wherein the temperature set point is set to a first temperature during a first time period and is set to a second temperature during a second time period.

10. The external charger of claim 9, wherein the first temperature is greater than the second temperature, and wherein the first time period is less than the second time period.

11. The external charger of any one of claims 7 to 10, wherein the temperature program automatically executes at a beginning of the charging session.

12. The external charger of any one of claims 7 to 11, wherein the temperature monitoring and control circuitry disables activation of the coil after a predetermined charging time.

13. An external charger for use with an implantable medical device, comprising:

a coil, selectively activated for producing energy receivable by the implantable medical device to power or recharge the implantable medical device;

a temperature-sensitive element for detecting the temperature of the external charger;

control circuitry configured to disable or reduce activation of the coil when the external charger temperature during a charging session is above a programmable maximum temperature; and

wherein the maximum temperature is programmed to automatically vary over time during the charging session.

14. The external charger of claim 13, wherein the maximum temperature comprises a first programmable maximum temperature during a first portion of a charging session and a second programmable maximum temperature during a second portion of the charging session.

15. The external charger of claim 14, wherein the first programmable maximum temperature is higher than the second programmable maximum temperature.

16. The external charger of claim 13, 14 or 15, wherein the control circuitry disables activation of the coil after a predetermined charging time.

17. The external charger of any one of claims 13 to 16,
wherein the control circuitry is further configured to enable activation of the coil during the charging session when the external charger temperature is below a programmable minimum temperature, and

wherein the minimum temperature is programmed to automatically vary over time during the charging session.

18. The external charger of any one of claims 13 to 17, wherein the control circuitry is configured to enable stimulation of the coil during the charging session when the external charger temperature is below the programmable maximum temperature.

19. A system, comprising:
an implantable medical device, comprising:
a first coil for receiving energy to provide power for operating the implantable medical device; and
an external charger, comprising:
a second coil, selectively stimulated for producing energy receivable by the first coil; and
a control circuitry comprising a temperature-sensitive element for detecting the temperature of a portion of the external charger, the control circuitry configured to disable activation of the second coil when the external charger temperature is above a programmable maximum temperature,
wherein the programmable maximum temperature automatically varies over time during a charging session.
20. The system of claim 19, further comprising, a user interface for the external charger, configured to program the programmable maximum temperature.
21. The system of claim 20, wherein the user interface is configured *for direct entry of the programmable maximum temperature.*
22. The system of claim 20, wherein the user interface is configured to select the programmable maximum temperature from a plurality of pre-defined maximum temperatures.
23. The system of claim 20, wherein the user interface is configured to select a program of a plurality of temperature programs, the selected program establishing the programmable maximum temperature as:

a first maximum temperature during a first portion of a charging session,
and
a second maximum temperature during a second portion of the charging session, different from the first maximum temperature.

24. The system of claim 20,
wherein the user interface is configured to select a program from a plurality of temperature programs,
wherein a first program of the plurality of temperature programs establishes the programmable maximum temperature as a first maximum temperature for an entire charging session, and
wherein a second program of the plurality of temperature programs establishes the programmable maximum temperature as:
a first maximum temperature during a first portion of a charging session and
a second maximum temperature during a second portion of the charging session, different from the first maximum temperature.



25. The system of any one of claims 19 to 24, wherein the temperature detection circuitry is further configured to enable activation of the coil during the charging session when the external charger temperature is below a *programmable minimum temperature*, and
wherein the programmable minimum temperature depends on the programmable maximum temperature.

26. The system of any one of claims 19 to 25, wherein the temperature detection circuitry is configured to enable activation of the coil during a

charging session when the external charger temperature is below the programmable maximum temperature.

27. The system of any one of claims 19 to 26, wherein the control circuitry disables activation of the coil after a predetermined charging time.

28. The system of claim 19 further comprising an external controller configured to allow a patient to control therapy settings in the implantable medical device and wherein the external controller comprises a user interface configured to allow a user to provide the temperature set point to the external charger.



2/17




Figure 2
(prior art)




Figure 3
(prior art)

3/17




Figure 4
(prior art)




Figure 5A

Figure 5B

5/17





Figure 6






Figure 7




Figure 8





Figure 9

**Figure 10**




Figure 11B




Figure 11A




Figure 12






Figure 13



15/17

**Figure 15**




Figure 16

16/17





Figure 17

17/17

**Figure 18**

