



(86) **Date de dépôt PCT/PCT Filing Date:** 2009/01/29
(87) **Date publication PCT/PCT Publication Date:** 2009/08/06
(45) **Date de délivrance/Issue Date:** 2015/03/31
(85) **Entrée phase nationale/National Entry:** 2010/07/19
(86) **N° demande PCT/PCT Application No.:** US 2009/032412
(87) **N° publication PCT/PCT Publication No.:** 2009/097412
(30) **Priorités/Priorities:** 2008/01/30 (US61/024,778);
2008/01/30 (US61/024,737); 2008/03/20 (US61/038,281);
2008/10/13 (US61/104,870)

(51) **Cl.Int./Int.Cl. A61L 27/18** (2006.01),
A61L 27/50 (2006.01), **A61L 27/56** (2006.01)
(72) **Inventeurs/Inventors:**
CASE, KIRT L., US;
POPOOLA, OLUDELE O., US;
KISSLING, ROBBY, US;
MIMNAUGH, BRION R., US;
NEWSOME, ARCHIE W., US;
PANCHISON, CLARENCE M., US;
...
(73) **Propriétaire/Owner:**
ZIMMER, INC., US
(74) **Agent:** NORTON ROSE FULBRIGHT CANADA
LLP/S.E.N.C.R.L., S.R.L.

(54) **Titre : COMPOSANT ORTHOPEDIQUE POUR RIGIDITE BASSE**
(54) **Title: ORTHOPEDIC COMPONENT OF LOW STIFFNESS**

(57) **Abrége/Abstract:**

An orthopedic component having multiple layers that are selected to provide an overall modulus that is substantially lower than the modulus of known orthopedic components to more closely approximate the modulus of the bone into which the orthopedic component is implanted. In one exemplary embodiment, the orthopedic component is an acetabular shell. For example, the acetabular shell may include an outer layer configured for securement to the natural acetabulum of a patient and an inner layer configured to receive an acetabular liner. The head of a femoral prosthesis articulates against the acetabular liner to replicate the function of a natural hip joint. Alternatively, the inner layer of the acetabular shell may act as an integral acetabular liner against which the head of the femoral prosthesis articulates.



(72) **Inventeurs(suite)/Inventors(continued):** PLETCHER, DIRK L., US; SCHLEMMER, RANDY L., US; FANG, ZHIBIN, US;
VIVANCO, JUAN, US; RUFNER, ALICIA, US

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property Organization
International Bureau(43) International Publication Date
6 August 2009 (06.08.2009)

PCT

(10) International Publication Number
WO 2009/097412 A3

(51) International Patent Classification:

A61L 27/18 (2006.01) A61L 27/56 (2006.01)
A61L 27/50 (2006.01)

US]; 3557 Lakeshore Drive, Columbia City, IN 46725 (US).

(21) International Application Number:

PCT/US2009/032412

(74) Agent: SCHWARTZ, Michael, D.; Baker & Daniels LLP, 111 East Wayne Street, Suite 800, Fort Wayne, IN 46802 (US).

(22) International Filing Date:

29 January 2009 (29.01.2009)

(81) Designated States (*unless otherwise indicated, for every kind of national protection available*): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BR, BW, BY, BZ, CA, CH, CN, CO, CR, CU, CZ, DE, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IS, JP, KE, KG, KM, KN, KP, KR, KZ, LA, LC, LK, LR, LS, LT, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PG, PH, PL, PT, RO, RS, RU, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

61/024,737	30 January 2008 (30.01.2008)	US
61/024,778	30 January 2008 (30.01.2008)	US
61/038,281	20 March 2008 (20.03.2008)	US
61/104,870	13 October 2008 (13.10.2008)	US

(71) Applicant (*for all designated States except US*): ZIMMER, INC. [US/US]; P.O. Box 708, 345 East Main Street, Warsaw, IN 46580 (US).

(72) Inventors; and

(75) Inventors/Applicants (*for US only*): CASE, Kirt, L. [US/US]; 1832 E. 75 N., Warsaw, IN 46582 (US). POPOOLA, Oludele, O. [NG/US]; 50531 Prestonwood Court, Granger, IN 46530 (US). KISSLING, Robby [US/US]; 3328 N. Knollwood Drive, Warsaw, IN 46582 (US). MIMNAUGH, Brion, R. [US/US]; 8748 E. 500 N., North Webster, IN 46555 (US). NEWSOME, Archie, W. [US/US]; 7998 W. State Rd. 25, Mentone, IN 46539 (US). PANCHISON, Clarence, M. [US/US]; 2513 Westside Drive, Warsaw, IN 46580 (US). PLETCHER, Dirk, L. [US/US]; 505 Indiana Street, Walkerton, IN 46574 (US). SCHLEMMER, Randy, L. [US/US]; 3266 U.S. 6, Bremen, IN 46506 (US). FANG, Zhibin [US/US]; 2988 E. Lake Tahoe Trail, Warsaw, IN 46582 (US). VIVANCO, Juan [CL/US]; 1910 University Ave., Apt. A, Madison, WI 53726 (US). RUFNER, Alicia [US/(84) Designated States (*unless otherwise indicated, for every kind of regional protection available*): ARIPO (BW, GH, GM, KE, LS, MW, MZ, NA, SD, SL, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European (AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, SE, SI, SK, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, ML, MR, NE, SN, TD, TG).

Published:

- with international search report (Art. 21(3))
- before the expiration of the time limit for amending the claims and to be republished in the event of receipt of amendments (Rule 48.2(h))

(88) Date of publication of the international search report:

14 May 2010

(54) Title: OTHOPEDIC COMPONENT OF LOW STIFFNESS

(57) Abstract: An orthopedic component having multiple layers that are selected to provide an overall modulus that is substantially lower than the modulus of known orthopedic components to more closely approximate the modulus of the bone into which the orthopedic component is implanted. In one exemplary embodiment, the orthopedic component is an acetabular shell. For example, the acetabular shell may include an outer layer configured for securement to the natural acetabulum of a patient and an inner layer configured to receive an acetabular liner. The head of a femoral prosthesis articulates against the acetabular liner to replicate the function of a natural hip joint. Alternatively, the inner layer of the acetabular shell may act as an integral acetabular liner against which the head of the femoral prosthesis articulates.



WO 2009/097412 A3

BACKGROUND

1. Field of the Invention.

5 [0002] The present invention relates to orthopedics, and, particularly, to orthopedic implants.

2. Description of the Related Art.

10 [0003] During orthopedic surgery, such as a total hip arthroplasty, orthopedic prostheses are implanted into a patient's body. For example, a femoral stem may be implanted into the intramedullary canal of the patient's femur. After the stem is inserted into the intramedullary canal at a desired depth and orientation, a femoral head may be attached to the proximal end of the stem. The femoral head fits into a substantially hemispherically shaped socket of an

acetabular prosthesis and provides a surface for articulation between the femoral head and the acetabular prosthesis.

[0004] In certain embodiments, acetabular prostheses comprise two separate components, an acetabular backing component and an acetabular articulating component, which are often referred to as a shell and liner, respectively. The backing component is generally hemispherically shaped and is affixed and embedded into an acetabulum of a patient. Similarly, the liner has a hemispherical shape configured to mate with an internal cavity of the backing component. The liner further includes a socket to mate with a femoral head. The acetabular backing components are typically formed of solid metal, such as titanium or a titanium alloy, or from stainless steel. In contrast, the acetabular articulating components may be formed from an ultra-high molecular weight polyethylene polymer material ("UHMWPE"). These types of acetabular cups are often referred to in the art, and sometimes herein, as "metal backed" components.

[0005] Metal backed orthopedic components generally experience contact stresses during normal joint articulation. For example, in conventional metal backed acetabular components, previous studies have shown that polymer thickness and conformity between the femoral head and the polyethylene liner play a significant role in the level of contact stress experienced. For highly conforming acetabular component designs, the contact stress in the polymer liner is very sensitive to the polymer thickness when the polymer thickness is small. Specifically, it has been demonstrated that when the polymer thickness is less than 4-6 mm, the contact stress generally increases rapidly as the thickness decreases. Therefore, a minimum polymer thickness of 4-6 mm has been maintained for metal backed acetabular components in an attempt to lessen the contract stress. Further, having a 4 mm minimum thickness of the polyethylene liner in an acetabular component has been widely accepted as an industry design rule and has been traditionally relied upon in the manufacture of acetabular components.

[0006] Stress shielding is a phenomenon that can potentially occur with metal backed orthopedic components and can occur when a portion of the stress normally exerted on a patient's bone is instead borne by the orthopedic component, such as an acetabular cup. As result, the bone may begin to undergo atrophy and decalcification resulting in a weaker bone structure. In

severe stress shielding, the bone may be resorbed by the body, decreasing the amount of bone stock at the fracture site.

SUMMARY

[0007] The present invention, in one exemplary embodiment, provides an orthopedic component having multiple layers that are selected to provide an overall modulus that is substantially lower than the modulus of known orthopedic components to more closely approximate the modulus of the bone into which the orthopedic component is implanted. In one exemplary embodiment, the orthopedic component is an acetabular shell. For example, the acetabular shell may include an outer layer configured for securement to the natural acetabulum of a patient and an inner layer configured to receive an acetabular liner. The head of a femoral prosthesis articulates against the acetabular liner to replicate the function of a natural hip joint. Alternatively, the inner layer of the acetabular shell may act as an integral acetabular liner against which the head of the femoral prosthesis articulates.

[0008] In one exemplary embodiment, the outer layer of the acetabular shell is formed from a metal and the inner layer is formed from a biocompatible polymer. For example, the outer layer may be formed from a porous metal and the inner layer from a polyaryletherketone ("PAEK"), such as polyetheretherketone ("PEEK"), or from a UHMWPE, such as an antioxidant stabilized UHMWPE. To form an acetabular shell according to the present invention, the outer layer is formed as a cup-shaped body using traditional techniques. Once formed, the biocompatible polymer, such as PEEK or UHMWPE, is secured to the outer layer to form an inner layer, resulting in a substantially completed acetabular shell. In exemplary embodiments, the polymer is attached to the outer layer by injection or compression molding. In one exemplary embodiment, the orthopedic component is complete after attachment of the outer and inner layers, e.g., after injection or compression molding, and requires no further machining or modification prior to implantation. In another exemplary embodiment, the orthopedic component may be further machined after attachment of the outer and inner layers to provide the final details, dimensions, and/or features that are desired.

[0009] Attaching a biocompatible polymer to a porous material to form an orthopedic implant allows the polymer to be received within the pores of the material. When the polymer hardens, the interaction of the two materials provides a firm securement between the porous

layer and the polymer layer to form the orthopedic component. Additionally, when the biocompatible polymer is compression or injection molded, for example, to the porous material, a locking feature, such a groove configured to receive a corresponding snap ring, for example, may be integrally formed in and/or on the polymer. By integrally forming a locking feature into or on the orthopedic component, the machining steps necessary to finish the orthopedic component may be reduced and/or entirely eliminated.

[0010] Advantageously, by manufacturing an orthopedic component in accordance with embodiments of the present invention, the orthopedic component has a stiffness which is lower than the stiffness of other orthopedic components. As a result, the present invention lessens the effects of stress shielding on the natural bone stock, reducing the potential for osteolysis and bone resorption.

[0011] Additionally, the present invention provides a surgeon with the ability to drill bone screw receiving apertures into the orthopedic component at the most advantageous locations. Thus, when implanting an acetabular shell, for example, the surgeon may identify the areas of the acetabulum having the greatest bone stock and then drill apertures through the acetabular shell which are aligned to utilize this bone stock. As a result, each orthopedic component can be custom fit to the individual patient to facilitate the optimal retention of the orthopedic component by the natural bone stock. Additionally, allowing a surgeon to customize the location of the bone screw receiving apertures reduces inventory and machining costs by eliminating the need to manufacture and stock orthopedic components with various bone screw receiving aperture configurations. Further, the method of manufacture of the present invention may be used to create a smooth surface on the polymer of the orthopedic component against which another orthopedic component and/or natural bone stock may articulate. By creating a smooth surface on the polymer, the generation of wear debris may be lessened.

[0012] Further, by utilizing an antioxidant stabilized polymer, such as UHMWPE blended with vitamin E, the overall thickness of the polymer layer may be lessened. Thus, the polymer layer may have a thickness of less than 4 mm, for example, without demonstrating an unacceptable level of contact stress. This results in reduced material and manufacturing costs. Further, by utilizing an antioxidant stabilized polymer, long term oxidation of the polymer may be lessened, increasing the useful life of the orthopedic component.

[0013] In one form thereof, the present invention provides an orthopedic system including: a substantially hemispherical outer layer formed from a first porous material; and a substantially hemispherical inner layer formed from an antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene, wherein said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene at least partially permeates the pores of said first porous material so that at least a portion of said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene interdigitates with pores of said first porous material, said first porous material and said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene cooperating to define an acetabular shell having an effective stiffness between 0.1 GPa and 15 GPa, and wherein said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene is prepared by a process that includes combining ultrahigh molecular weight polyethylene with an antioxidant to form a blend; processing the blend to consolidate the blend, wherein the consolidated blend had a melting point; preheating the consolidated blend to a preheat temperature below the melting point of the consolidated blend; and irradiating the preheated consolidated blend while maintaining the consolidated blend at a temperature below the melting point of the consolidated blend.

[0014] In another form thereof, the present invention provides an acetabular component configured for use in a hip replacement surgery, the acetabular component including: a porous layer configured to contact and interface with bone tissue when the acetabular component is implanted; an inner layer formed from an antioxidant stabilized, crosslinked ultrahigh molecular weight polyethylene and having a thickness of less than six millimeters, the inner layer configured to receive a femoral component; and an interdigitation layer defined by the distance over which the antioxidant stabilized, crosslinked ultrahigh molecular weight polyethylene of the inner layer infiltrates pores of the porous layer; and wherein said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene is prepared by a process that includes combining ultrahigh molecular weight polyethylene with an antioxidant to form a blend; processing the blend to consolidate the blend, wherein the consolidated blend had a melting point; preheating the consolidated blend to a preheat temperature below the melting point of the consolidated blend; and irradiating the preheated consolidated blend while maintaining the consolidated blend at a temperature below the melting point of the consolidated blend.

[0015] In yet another form thereof, the present invention provides a method of manufacturing an orthopedic component for implantation into a bone, the orthopedic implant having a bone contacting layer, an interdigitation layer, and an inner layer, the method comprising the steps of: determining the elastic modulus of the bone; selecting a thickness of at least one of the bone contacting layer, an interdigitation layer, and an inner layer based on the elastic modulus of the bone; and molding the inner layer to the bone contacting layer to form at least one of the bone contacting layer, an interdigitation layer, and an inner layer to have the selected thickness; wherein said inner layer has an antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene prepared by a process that comprises combining ultrahigh molecular weight polyethylene with an antioxidant to form a blend; processing the blend to consolidate the blend, wherein the consolidated blend had a melting point; preheating the consolidated blend to a preheat temperature below the melting point of the consolidated blend; and irradiating the preheated consolidated blend while maintaining the consolidated blend at a temperature below the melting point of the consolidated blend.

15

BRIEF DESCRIPTION OF THE DRAWINGS

[0016] The above-mentioned and other features and advantages of this invention, and the manner of attaining them, will become more apparent and the invention itself will be better understood by reference to the following description of embodiments of the invention taken in conjunction with the accompanying drawings, wherein:

20

[0017] Fig. 1 is cross-sectional view of an exemplary orthopedic implant, depicted as an acetabular shell, made in accordance with an exemplary embodiment of the present invention;

- [0018] Fig. 2A is a fragmentary sectional view of a portion of the implant of Fig. 1;
- [0019] Fig. 2B is an enlarged schematic representation of a portion of the implant of Fig. 2A;
- [0020] Fig. 3 is an assembly view of the acetabular shell of Fig. 1 depicting alternative acetabular liners;
- [0021] Fig. 4 depicts a graphical illustration of the relationship between contact stress and the thickness of a polymer material;
- [0022] Fig. 5 depicts a graphical illustration of the stress/strain properties of different polymeric materials;
- [0023] Fig. 6 depicts a graphical illustration of an exemplary finite element model;
- [0024] Fig. 7 depicts a graphical illustration of the stress/strain properties of porous metal;
- [0025] Fig. 8 depicts a graphical illustration of the stress/strain properties of both conventional UHMWPE and antioxidant stabilized UHMWPE;
- [0026] Fig. 9 depicts a graphical illustration of an unconfined compression loading condition;
- [0027] Fig. 10 depicts a graphical illustration of the effect of porosity on a construct's stiffness;
- [0028] Fig. 11 depicts a graphical illustration of the effect of polymer thickness on a construct's stiffness;
- [0029] Fig. 12 depicts a graphical illustration of stress/strain properties for various polymer thicknesses;
- [0030] Fig. 13 depicts a graphical illustration of the stress/strain properties for different thicknesses of the interdigitation layer between the polymer and porous layers;
- [0031] Fig. 14 depicts a graphical illustration of stiffness properties for various materials;

[0032] Fig. 15 depicts a graphical illustration of the oxidative index of UHMWPE blend with 0.50 weight percent α -tocopherol acetate;

[0033] Fig. 16 depicts a graphical illustration of the trans-vinylene index of an irradiated monoblock component; and

5 [0034] Fig. 17 is a fragmentary, cross-sectional view of an acetabular cup construct depicting exemplary thickness measurements of a UHMWPE layer.

[0035] Corresponding reference characters indicate corresponding parts throughout the several views. The exemplifications set out herein illustrate preferred embodiments of the invention and such exemplifications are not to be construed as limiting the scope of the invention in any
10 manner.

DETAILED DESCRIPTION

[0036] As shown in Fig. 1, an orthopedic component is depicted as acetabular shell 10. While described and depicted herein with specific reference to an acetabular shell, the orthopedic component of the present invention may be any orthopedic component, such as a glenoid
15 component for use in a shoulder prosthesis system. Acetabular shell 10 has a substantially hemispherical shape and includes first, outer bone-contacting layer 12 and second, inner layer 14 against which a corresponding femoral orthopedic component and/or natural femoral head articulate. In one exemplary embodiment, second, inner layer 14 provides a surface against which an additional acetabular liner (Fig. 3) may be seated. Additionally, acetabular prosthesis may have
20 any suitable shape known in the art, such as hemispherical, and is generally configured to mate with a femoral head assembly, as indicated above, such as the femoral head assembly disclosed in U.S. Patent No. 7,306,629, entitled "FEMORAL HEAD ASSEMBLY WITH VARIABLE OFFSET". Similarly, prosthesis 10 may have a shape similar to that disclosed in U.S. Patent No. 5,879,398, entitled "ACETABULAR CUP".

25 [0037] In one exemplary embodiment, acetabular shell 10 has a stiffness which is less than the stiffness of other, known acetabular shells. Specifically, acetabular shell 10 has an effective stiffness which is substantially less than the effective stiffness of known acetabular

shells. As used herein, absent an indication to the contrary or use of a different term, "effective stiffness" refers to the overall stiffness or elastic modulus of the entire construct of an orthopedic component according to the experimental procedures set forth below in the Examples. As a result of the lower effective stiffness of the acetabular shell made in accordance with the present invention,

5 the effect of stress shielding on the natural acetabulum is lessened. As indicated above, stress shielding is a phenomenon in which the rigidity of implanted orthopedic components limits the transfer of forces to natural bone stock that the bone stock would normally receive, such as during the loading of a joint. As a result, osteolysis may occur at the joint, which may result in bone resorption and weakening of the bone stock surrounding the implanted orthopedic components.

10 Thus, due to the lower effective stiffness of acetabular shell 10, acetabular shell 10 allows for a greater amount of joint loading forces to be transferred to the natural bone stock.

[0038] In one exemplary embodiment, bone-contacting layer 12 is formed from a porous material, such as a porous metal. In one exemplary embodiment, bone-contacting layer 12 is formed using Trabecular Metal™ technology available from Zimmer, Inc., of Warsaw, Indiana. Trabecular

15 Metal™ is a trademark of Zimmer Technology, Inc. Such a material may be formed from a reticulated vitreous carbon foam substrate which is infiltrated and coated with a biocompatible metal, such as tantalum, etc., by a chemical vapor deposition ("CVD") process in the manner disclosed in detail in U.S. Patent No. 5,282,861, issued February 1, 1994, entitled "OPEN CELL TANTALUM STRUCTURES FOR CANCELLOUS BONE IMPLANTS AND CELL AND

20 TISSUE RECEPTORS", and in Levine, B.R., et al., "Experimental and Clinical Performance of Porous Tantalum in Orthopedic Surgery", Biomaterials 27 (2006) 4671-4681. In addition to tantalum, other metals such as niobium, or alloys of tantalum and niobium with one another or with other metals may also be used.

[0039] Generally, with reference to Figs. 2A and 2B, the porous tantalum structure of

25 bone-contacting layer 12 includes a large plurality of ligaments 16 defining open spaces, such as voids or channels 18, therebetween, with each ligament 16 generally including a carbon core covered by a thin film of metal such as tantalum, for example. The open spaces between ligaments 16 form a matrix of continuous channels having no dead ends, such that the growth of cancellous bone through the porous tantalum structure is uninhibited. The porous tantalum may

include up to 75%-85% or more void space therein. Thus, porous tantalum is a lightweight, strong porous structure which is substantially uniform and consistent in composition, and closely resembles the structure of natural cancellous bone, thereby providing a matrix into which cancellous bone may grow to anchor acetabular shell 10 in the surrounding bone of the pelvis.

[0040] The porous tantalum structure may be made in a variety of densities in order to selectively tailor the structure for particular applications. In particular, as discussed in the above-incorporated U.S. Patent No. 5,282,861, the porous tantalum may be fabricated to virtually any desired porosity and pore size, and can thus be matched with the surrounding natural bone in order to provide an improved matrix for bone in growth and mineralization.

[0041] While described herein as being formed from Trabecular Metal™ technology, first, bone-contacting layer 12 may be formed from any biocompatible metal, such as Ti-6Al-4V, or a porous material, such as a fiber metal. Additionally, in another exemplary embodiment, a nonporous metal, such as titanium, is used to form bone-contacting layer 12 and the inner surface thereof may be formed to incorporate a porous surface. The porous surface allows for inner layer 14 to be received within the pores of bone-contacting layer 12.

[0042] As indicated above, inner layer 14 is formed from a biocompatible polymer. In one exemplary embodiment, inner layer 14 is formed from a PAEK, such as PEEK. In another exemplary embodiment, inner layer 14 is formed from UHMWPE. In other exemplary embodiments, the biocompatible polymer may be a polyolefin, polyester, polyimide, polyamide, polyacrylate, and/or other suitable polymers. By utilizing a porous metal, such as a metal manufactured using Trabecular Metal™ technology, and a biocompatible polymer, such as PEEK or UHMWPE, in combination, the elastic moduli of both materials forming acetabular shell 10 may be selected to be between the elastic moduli of cortical bone and trabecular bone.

[0043] Specifically, the elastic modulus of porous tantalum made in accordance with Trabecular Metal™ technology is approximately 3 GPa (gigapascal), the elastic modulus of PEEK is approximately 3.6 GPa, and the elastic modulus of UHMWPE ranges from 0.5 GPa to 2.0 GPa, depending on the specific UHMWPE used. Thus, various values within the range of 0.5 GPa and 2.0 GPa have been used for the elastic modulus of UHMWPE in the calculations throughout the present application and the Examples contained herein. The specific value for the elastic modulus of UHMWPE that is used in any particular calculation is identified where

relevant. In addition, the elastic moduli of cortical bone and trabecular bone are 15 GPa and 0.1 GPa, respectively. In contrast, the elastic modulus of Cobalt-Chromium is 220 GPa and the elastic modulus of Ti-6Al-4V is 110 GPa. Thus, the substantially low elastic moduli of porous tantalum made in accordance with Trabecular Metal™ technology and PEEK and/or UHMWPE provide for the effective stiffness of acetabular shell 10 to be substantially lower than the effective stiffness of known acetabular shells. For example, the effective stiffness of acetabular shell 10 may be as low as 0.2, 0.3, 0.4, 0.5, 0.6, 0.7, 0.8, or 0.9 GPa or as high as 1.0, 1.1, 1.2, 1.3, 1.4, 1.5, 2.0, 2.5, 4.0, or 8.0 GPa. In one exemplary embodiment, the effective stiffness of acetabular shell 10 is between 0.3 GPa and 1.5 GPa.

10 [0044] When UHMWPE is used as the polymer for inner layer 14, in order to beneficially alter the material properties of UHMWPE and decrease its wear rate, the UHMWPE may be crosslinked. For example, UHMWPE may be subjected to electron beam or gamma irradiation, causing chain scissions of the individual polyethylene molecules as well as the breaking of C-H bonds to form free radicals on the polymer chains. The free radicals on adjacent polymer chains may then bond together to form crosslinked UHMWPE.

15 [0045] In another exemplary embodiment, inner layer 14 is formed from an antioxidant stabilized polymer, such as antioxidant stabilized UHMWPE. In one exemplary embodiment, UHMWPE powder is combined with an antioxidant to form an antioxidant stabilized UHMWPE. By forming inner layer 14 from an antioxidant stabilized polymer, some of the free radicals in the polymer are quenched, which reduces oxidation and, correspondingly, increases the useful life of the polymer. For example, the UHMWPE may include an antioxidant such as Vitamin C, lycopene, honey, and/or tocopherol, i.e., Vitamin E. Additionally, as any tocopherol may be used, such as d- α -tocopherol, d,l- α -tocopherol, or α -tocopherol acetate, unless otherwise specifically stated herein, the term
25 "tocopherol" in its generic form refers to all tocopherols. Exemplary methods for combining UHMWPE with antioxidants are disclosed in co-pending U.S. Patent Application US 2008/0319167A1, entitled "AN ANTIOXIDANT STABILIZED CROSSLINKED ULTRA-HIGH MOLECULAR WEIGHT POLYETHYLENE FOR MEDICAL DEVICE APPLICATIONS", filed April 10, 2008, assigned to the assignee of the present invention.

[0046] In combining UHMWPE and tocopherol to form an antioxidant stabilized UHMWPE, any mechanism and/or process achieving a substantially homogenous blend of the components may be utilized. In one exemplary embodiment, solvent blending is utilized. In solvent blending, tocopherol is mixed with a volatile solvent to lower the viscosity of the tocopherol and facilitate homogenous blending of the tocopherol with the UHMWPE. Once the tocopherol is mixed with the solvent, the tocopherol/solvent mixture may be combined with the UHMWPE, such as with a cone mixer. The solvent is then evaporated, leaving only the antioxidant stabilized UHMWPE. In another exemplary embodiment, tocopherol may be blended with UHMWPE by precision coating or atomization. For example, tocopherol may be precision coated onto the UHMWPE powder using a MP-1 MULTI-PROCESSOR[™] Fluid Bed connected to a laboratory module Precision Coater available from Niro Inc. of Columbia, Maryland. MULTI-PROCESSOR[™] is a trademark of Niro Inc.

[0047] In another exemplary embodiment, low intensity mixing may be used. Low intensity, i.e. low shear, mixing may be performed using a Diosna P100 Granulator, available from Diosna GmbH of Osnabrück, Germany, a subsidiary of Multimixing S.A. In another exemplary embodiment, high shear mixing may be used. High shear mixing of UHMWPE and tocopherol may be achieved using a RV02E or a R05T High Intensity Mixer, both commercially available from Eirich Machines of Gurnee, Illinois. Alternatively, high shear mixing may be achieved using a Collette ULTIMAPRO[™] 75 One Pot Processor available from Niro, Inc. of Columbia, Maryland. ULTIMAPRO[™] is a trademark of Niro, Inc. Based on the results of testing the above identified methods useful for combining UHMWPE and tocopherol, high shear mixing appears to provide favorable results, including an acceptable homogeneity and a low number of indications, i.e., areas of high tocopherol concentrations relative to the surrounding areas as determined by visual inspection under ultraviolet light or by chemical measurements, such as infrared spectroscopy or gas chromatography. Additionally, in other exemplary embodiments, fluidized bed, emulsion polymerization, electrostatic precipitation, wetting or coating of particles, and/or master batch blending may be used to combine the UHMWPE and tocopherol.

[0048] Irrespective of the method used to combine the UHMWPE and tocopherol to form the antioxidant stabilized UHMWPE, the components are combined in ratios necessary to achieve a tocopherol concentration of between 0.01 weight percent (wt. %) and 3 wt. %. In

exemplary embodiments, the tocopherol concentration may be as low as 0.01 wt. %, 0.05 wt. %, and 0.1 wt. %, or as high as 0.6 wt. %, 0.8 wt. %, and 1.0 wt. %, for example. In determining the appropriate amount of tocopherol, two competing concerns exist. Specifically, the amount selected must be high enough to quench free radicals in the UHMWPE, but must also be low enough to allow sufficient crosslinking so as to maintain acceptable wear properties of the antioxidant stabilized UHMWPE. In one exemplary embodiment, a range of tocopherol from 0.1 to 0.6 wt. % is used to successfully quench free radicals while still maintaining acceptable wear properties.

[0049] Once the antioxidant stabilized UHMWPE is substantially homogeneously blended and the amount of tocopherol is determined to be within an acceptable range, the antioxidant stabilized UHMWPE is secured to bone contacting layer 12 to form bone contacting layer 14. In one exemplary embodiment, the antioxidant stabilized UHMWPE is secured to bone contacting layer 12, as described in detail below. Once secured to bone contacting layer 12, the antioxidant stabilized UHMWPE may be exposed to crosslinking irradiation.

[0050] In preparing an antioxidant stabilized UHMWPE for exposure to crosslinking irradiation, the antioxidant stabilized UHMWPE may be preheated. In one exemplary embodiment, the antioxidant stabilized UHMWPE may be preheated to any temperature between room temperature, approximately 23°C, up to the melting point of the antioxidant stabilized UHMWPE, approximately 140°C. In another exemplary embodiment, the antioxidant stabilized UHMWPE is preheated to a temperature between 60°C and 130°C. In other exemplary embodiments, the antioxidant stabilized UHMWPE may be heated to a temperature as low as 60°C, 70°C, 80°C, 90°C, or 100°C or as high as 110°C, 120°C, 130°C, 135°C, 140°C. By preheating the antioxidant stabilized UHMWPE before irradiation, the material properties of the resulting irradiated antioxidant stabilized UHMWPE are affected. Thus, the material properties for an antioxidant stabilized UHMWPE irradiated at a relatively cold, e.g., approximately 40°C, temperature are substantially different than the material properties for an antioxidant stabilized UHMWPE irradiated at a relatively warm, e.g., approximately 120°C to approximately 140°C, temperature.

[0051] However, while the material properties of an antioxidant stabilized UHMWPE irradiated at a lower temperature may be superior, the wear properties, fatigue properties,

oxidation level, and free radical concentration are all negatively affected. In contrast, while irradiation of an antioxidant stabilized UHMWPE at a higher temperature may slightly diminish the material properties, it also results in a higher crosslinking efficiency due to higher chain mobility and adiabatic melting. Additionally, by irradiating at a higher temperature, a greater number of crosslinks are formed. Thus, there are less free radicals in the antioxidant stabilized UHMWPE and less tocopherol is consumed by reacting with the free radicals during irradiation and immediately thereafter. As a result, a greater amount of tocopherol remains in the blend that may react with free radicals during the antioxidant stabilized UHMWPE's lifecycle, i.e., after irradiation. This, in turn, increases the overall oxidative stability of the antioxidant stabilized UHMWPE.

[0052] Additionally, when the antioxidant stabilized UHMWPE forming inner layer 14 and the porous metal of bone contacting layer 12 are irradiated, bone contacting layer 12 may rapidly increase in temperature. Thus, the temperature increase of bone contacting layer 12 should be taken into account when determining the preheat temperature of the antioxidant stabilized UHMWPE forming inner layer 14 and the porous metal of bone contacting layer 12.

[0053] After the temperature of the antioxidant stabilized UHMWPE forming inner layer 14 reaches the desired preheat temperature, the antioxidant stabilized UHMWPE is subsequently irradiated to induce crosslinking of the UHMWPE. Thus, as used herein, "crosslinking irradiation" refers to exposing the antioxidant stabilized UHMWPE to ionizing irradiation to form free radicals which may later combine to form crosslinks. The irradiation may be performed in air at atmospheric pressure, in a vacuum chamber at a pressure substantially less than atmospheric pressure, or in an inert environment, i.e., in an argon environment, for example. The irradiation is, in one exemplary embodiment, electron beam irradiation. In another exemplary embodiment, the irradiation is gamma irradiation. In yet another exemplary embodiment, the crosslinking does not require irradiation, but instead utilize silane crosslinking. In one exemplary embodiment, crosslinking is induced by exposing the antioxidant stabilized UHMWPE to a total radiation dose between about 25 kGy and 1,000 kGy. In another exemplary embodiment, crosslinking is induced by exposing the antioxidant stabilized UHMWPE to a total radiation dose between about 50 kGy and 250 kGy in air. These doses are higher than doses commonly used to crosslink UHMWPE due to the presence of antioxidant in the antioxidant stabilized UHMWPE. Specifically, the antioxidant reacts with some of the polyethylene chains

that became free radicals during irradiation. As a result, a higher irradiation dose must be administered to the antioxidant stabilized UHMWPE to achieve the same level of crosslinking that would occur at a lower dose in conventional UHMWPE, i.e., UHMWPE absent an antioxidant.

[0054] Advantageously, by utilizing an antioxidant stabilized, crosslinked polymer to form inner layer 14 of acetabular component 10, the thickness of inner layer 14 may be substantially reduced. For example, the thickness of bearing layer 26, shown in Fig. 2B and described below, of inner layer 14 may be less than 4 mm and still maintain the same level of contact stress as known acetabular cups having a bearing layer thicker than 4 mm. As discussed above, it was previously believed by persons of ordinary skill in the art that bearing layer 26 of a metal backed orthopedic component must be equal to or greater than 4 mm. However, acetabular cup prosthesis 10 has a substantially lower elastic modulus than known acetabular cup designs. As a result, under the same loading conditions, porous metal backed acetabular components will deform more and provide more contact area so that the peak contact stress of acetabular cup prosthesis 10 is lowered, as shown in Fig. 4. Additionally, forming an acetabular component having a decreased thickness of bearing layer 26 allows for a greater range of motion of a mating femoral head component and also reduces the outer diameter of the acetabular component, which results in the preservation of a patient's natural bone stock. Additional information regarding the specific properties of an orthopedic component of the present invention having an antioxidant stabilized polymer inner layer 14 is set forth below in the corresponding Examples.

[0055] To manufacture acetabular shell 10, bone-contacting layer 12 is formed using traditional techniques, such as milling, molding, and/or machining. Once formed and/or machined to have the desired characteristics, inner layer 14 is attached to bone contacting layer 12. In one exemplary embodiment, inner layer 14 is attached to bone contacting layer 12 by injection or compression molding. By utilizing a porous metal, e.g., a material formed using Trabecular Metal™ technology or a metal having a porous coating to form bone-contacting layer 12, the attachment of the inner layer 14 allows the biocompatible polymer forming inner layer 14 to be received within the pores of bone-contacting layer 12. This interaction creates a rigid, mechanical bond between bone contacting layer 12 and inner layer 14 to fixedly attach the two

layers together, which allows for the transfer of the forces between bone contacting layer 12 and inner layer 14 that are encountered during the loading of the joint.

[0056] Referring to Figs. 2A and 2B, the polymeric material of inner layer 14 may be molded at least partially within the porous substrate of bone contacting layer 12 to a desired depth to thereby form a unified construct by which the polymeric material of inner layer 14 is connected to bone contacting layer 12 by interdigitation of the polymeric material of inner layer 14 at least partially within the pores or channels 18 of bone contacting layer 12. In this manner, referring to Fig. 2B, the implant construct generally includes three layers, including porous layer 22, which will contact and interface with bone tissue when acetabular cup 10 is implanted within a patient, interdigitation layer 24 in which the polymeric material of inner layer 14 is infiltrated within bone contacting layer 12, and bearing layer 26 comprising the polymeric material of inner layer 14 and defining bearing surface 20. Thus, interdigitation layer 24 defines an intermeshing region in which the material of bearing layer 26 intermeshes with the material of porous layer 22 and the interaction between the polymeric material of bearing layer 26 and ligaments 16 defining voids or channels 18 of porous layer 22 retains bearing layer 26 thereon.

[0057] By varying the thickness of each of these three layers, i.e., porous layer 22, interdigitation layer 24, and bearing layer 26, as well as the porosity of porous layer 22, orthopedic components having varying stiffness may be created. Advantageously, this allows for a surgeon or other medical professional to select an orthopedic component that has a stiffness that is substantially similar to the bone stiffness of an individual patient. Additionally, this allows for manufacturers of orthopedic components to design and manufacture orthopedic components that have a stiffness that is substantially similar to the stiffness of the bone into which the orthopedic component is designed to be implanted. For example, if the orthopedic component is an acetabular component, the thickness of porous layer 22, interdigitation layer 24, and bearing layer 26 may be selected to create an orthopedic component having an effective stiffness that is substantially similar to the elastic modulus of the bone of the pelvis. Similarly, if the orthopedic component is a glenoid component, the thickness of porous layer 22, interdigitation layer 24, and bearing layer 26 may be selected to create an orthopedic component having an effective stiffness that is substantially similar to the elastic modulus of the bone of the glenoid of the scapula. Additional detailed information regarding specific thicknesses of each of

the above-identified layers and the resulting effective stiffness achieved using the same are set forth below in the corresponding Examples. The porous layer may have a porosity of at least 55 percent.

[0058] Once the attachment, e.g., injection or compression molding, of inner layer 14 to bone contacting layer 12 is finished, acetabular shell 10 may be complete, i.e., ready for

5 implantation. If acetabular shell 10 is not finished at the conclusion of the attachment process, acetabular shell 10 may be further machined to create a final, implantable acetabular shell. In one exemplary embodiment, acetabular component 10, having bone contacting layer 12 and inner layer 14 forms the entire implantable acetabular prosthesis. Thus, in this embodiment, a corresponding femoral prosthesis would articulate against bearing surface 20. However, in another exemplary
10 embodiment, bearing surface 20 may define a contact and/or support surface for an acetabular liner. In this embodiment, in order to complete the acetabular component of a hip prosthesis system, an acetabular liner may be provided. The acetabular liner, such as liners 28, 30 described below, receive the head of a femoral prosthesis to replicate the natural hip joint. The liner may have a liner material having ultrahigh molecular weight polyethylene.

15 [0059] To facilitate attachment of an acetabular liner to acetabular shell 10, an integral locking feature may be incorporated into inner layer 14 during the process of attaching inner layer 14 to bone contacting layer 12. Referring to Fig. 3, in one exemplary embodiment, groove 32 may be formed in inner layer 14 on bearing surface 20. For example, groove 32 may extend around bearing surface 20 proximate to and/or at the equator of inner layer 14. Groove 32 may be
20 configured to mate with rib 34 formed on acetabular liner 28, which may be formed from polyethylene. Acetabular liner 28 may then be retained within acetabular shell 10 via the snap-fit connection of groove 32 and rib 34.

[0060] Additionally, in another exemplary embodiment, groove 32 may also be configured to mate with rib 36 formed on conversion ring 38. Conversion ring 38 includes an inner
25 tapered surface 40 which may be configured to mate with outer tapered surface 42 on acetabular liner 30. Acetabular liner 30 may be formed from metal or ceramic, for example. Thus, conversion ring 38 may be retained within acetabular shell 10 via the snap-fit connection of groove 32 and rib 36. Acetabular liner 30 may then be received by conversion ring 38 and retained within acetabular shell 10 by the locking engagement of inner tapered surface 40 of conversion ring 38 and outer
30 tapered surface 42 of acetabular liner 30. Additional details regarding the operation and use of conversion ring 28 are set forth in co-pending U.S. Patent

Application US 2007/0239283 A1, entitled "ACETABULAR CUP CONVERSION RING", filed April 11, 2006.

[0061] In another exemplary embodiment, the integral locking feature may be configured to receive a spring ring or any other locking component. In yet another exemplary embodiment, the integral locking feature may be a Morse taper. This Morse taper may be formed as a self- locking taper configured to mate with a corresponding acetabular liner without the use of conversion ring 38. Advantageously, the use of the integral locking feature allows for a surgeon to easily switch between different liners and may also lessen the number of components needed to form the complete acetabular cup assembly.

10 [0062] Additionally, when inner layer 14 is attached to bone contacting layer 12, inner layer 14 may be formed with smooth bearing surface 20 that may be smoother than the inner or bearing surface of known acetabular cups. For example, known acetabular cups have a bearing surface that has an arithmetical mean roughness (Ra) of no less than 13 microns. In contrast, smooth bearing surface 20 may have an Ra as low as 1, 2, 3, or 4 microns or as high as 7, 8, 9, or 15 microns. In one exemplary embodiment, smooth bearing surface 20 has an Ra between approximately 4 microns and 8 microns. The formation of smooth bearing surface 20 may minimize wear debris that could potentially be created between bearing surface 20 of inner layer 14 and outer surface 44 of liner 28 during joint articulation.

[0063] Further, if the attachment of inner layer 14 to bone contacting layer 12 is accomplished by injection molding, surface 18 will have the same finish, i.e., the same roughness, as the corresponding surface of the injection mold. Thus, the corresponding injection mold surface can be polished to a smooth finish and the need to further smooth or polish bearing surface 20 of acetabular shell 10 is substantially eliminated. Advantageously, this lowers manufacturing costs by shortening the manufacturing process and lessening the labor and tool costs of the same.

25 Additionally, the surface roughness of inner layer 14 may be substantially similar to the surface roughness of liner 28. As a result of the substantially similar surface roughness of inner layer 14 and liner 28, the creation of wear debris may be further minimized.

[0064] In preparation for the implantation of acetabular shell 10, a surgeon may examine the acetabulum to determine the location having the greatest amount of and/or strongest area of bone stock. Referring to Fig. 1, once this determination is made, the surgeon may drill bone

screw receiving apertures 46 within acetabular shell 10 that may align with the previously identified areas. Advantageously, by allowing a surgeon to customize the location of the bone screw receiving apertures in an orthopedic component, the costs of inventorying different orthopedic components with various bone screw receiving aperture configurations and machining the same are substantially eliminated. Then, once acetabular shell 10 is ready for implantation, acetabular shell 10 is inserted into the prepared, natural acetabulum and a bone screw, such as bone screw 48, is positioned within aperture 46 and threaded into the patient's natural bone stock. In one exemplary embodiment, once bone screw 48 is affixed to the patient's bone stock, one of liners 28, 30 may be connected to acetabular shell 10 in the manner described above. In another exemplary embodiment, bone screw 48 may be countersunk into aperture 46, such the substantial entirety of head 50 of bone screw 48 is positioned within outer layer 12. Alternatively, in another exemplary embodiment, the liner may be attached to acetabular shell 10 prior to implantation, and the two components implanted together and retained in position in the patient's acetabulum using bone cement, for example.

[0065] Advantageously, due to the attachment of bone contacting layer 12 and inner layer 14, any dust and/or debris generated during the drilling of bone screw receiving apertures 46 will be substantially contained within the polymeric debris of inner layer 14. This allows for a surgeon to easily remove any dust and/or debris generated during the creation of the bone screw receiving apertures from acetabular shell 10 by removing the metallic dust or debris and the larger polymer shavings that may have encapsulated additional metallic dust or debris during the drilling operation.

Examples

[0066] The following non-limiting Examples illustrate various features and characteristics of the present invention, which is not to be construed as limited thereto. The following abbreviations are used throughout the Examples unless otherwise indicated.

TABLE 1
Abbreviations

Abbreviation	Full Word
E_{eff}	effective stiffness
PEEK	polyetheretherketone
R	Radius

mm	millimeter
H	height
Δ	change
σ_{eff}	effective stress
ε_{eff}	effective strain
σ	stress
ε	strain
UHMWPE	ultrahigh molecular weight polyethylene
L	length
PM	porous material
GPa	Gigapascals
lbf	Pound feet
π	3.14159265
r	radial displacement
t	thickness
T_p	thickness of polymer layer
T_t	thickness of porous metal layer
T_i	thickness of interdigitation layer
ϕ	porosity
kGy	kilo Gray
min	minute
MeV	mega electron volt
m	meter
$^{\circ}$	degrees
C	Celsius
FTIR	Fourier Transform Infrared Spectroscopy
wt. %	weight percent
MPa	Megapascal
UTS	ultimate tensile strength
YS	yield strength
HXPE	highly crosslinked polyethylene
OI	Oxidation Index
T	Temperature
DSC	Differential Scanning Calorimetry
ml	milliliter
nm	nanometer
TVI	trans-vinylene index
VEI	d/l- α -tocopherol index
Mc	million cycles
	molecular weight between crosslinks
AVE	aged vitamin E percent
AVEI	aged vitamin E index
AV-OI	aged oxidation index
mg	milligram
cm	centimeter

IR	infrared
VE%	weight percent tocopherol
Vol	volume
wt.	weight
VE	tocopherol
g	gram
DMA	Dynamic Mechanical Analysis
kJ	kilojoule
Izod	Izod Impact Strength
Conc.	Concentration
dm	decimeter

Example 1

Effective Stiffness of PEEK Layer/Porous Layer Construct

[0067] An analytical model was used to study the effective stiffness of a layer of PEEK positioned adjacent to a layer of porous metal.

[0068] The effective stiffness of a layer of PEEK positioned adjacent to a layer of porous metal is dependent on both the material properties and geometries of the individual layers. Thus, to attempt to remove any geometric effects, an analytical model was created to include a rectangular PEEK layer positioned atop a corresponding rectangular porous metal layer. The model incorporated frictionless contact between the PEEK layer and porous metal layer and with none of the PEEK penetrating into the pores of the porous metal layer. Additionally, the porous metal layer is assumed to have properties substantially similar to a porous metal layer formed in accordance with Trabecular Metal™ technology, as described in detail above. The model was then subjected to uniaxial compression. The stress is equally carried by both materials and the compressive displacement of the materials is additive. Thus, the total displacement in the model is calculated as:

$$\Delta L_{total} = \Delta L_{PEEK} + \Delta L_{Porous} = \epsilon_{PEEK} L_{PEEK} + \epsilon_{Porous} L_{Porous} = \frac{\sigma}{E_{PEEK}} L_{PEEK} + \frac{\sigma}{E_{Porous}} L_{Porous}$$

[0069] Thus, the effective stiffness of the two material construct is:

$$E_{eff} = \frac{\sigma}{\epsilon_{total}} = \frac{\sigma L}{\Delta L_{total}} = \frac{L_{PEEK} + L_{Porous}}{L_{PEEK}/E_{PEEK} + L_{Porous}/E_{Porous}}$$

[0070] Since different stiffness values are available for a porous metal layer formed in accordance with Trabecular Metal™ technology and since different applications may require

different thicknesses of the PEEK layer and the porous metal layer, the effective stiffness was modeled under a range of values. Effective stiffnesses were modeled with the stiffness of the porous metal layer varying between 1-3 GPa, the thickness of the PEEK layer varying between 2.0 and 5.8 mm, and the thickness of the porous metal layer remaining constant at 4 mm. The resulting effective stiffnesses are shown in TABLE 2, in addition to some representative stiffnesses, i.e., Young's moduli, of various materials.

TABLE 2

Material	Stiffness (GPa)
CoCr	210
Ti	110
PEEK	4
TM	1-3
UHMWPE	1

[0071] Referring to TABLE 2, the results indicated that with a porous metal layer of 4 mm and a PEEK layer of 2 mm, the effective stiffness of the construct was 1.3 GPa when 1.0 GPa was used as the elastic modulus of the porous metal layer. Additionally, with a porous metal layer of 4 mm and a PEEK layer of 2.0 mm, the effective stiffness of the construct was 3.3 GPa when 3.0 GPa was used as the elastic modulus of the porous metal layer. Further, the results indicated that with a porous metal layer of 4 mm and a PEEK layer of 5.8 mm, the effective stiffness of the construct was 1.8 GPa when 1.0 GPa was used as the elastic modulus of the porous metal layer. Additionally, with a porous metal layer of 4 mm and a PEEK layer of 5.8 mm, the effective stiffness of the construct was 3.5 GPa when 3.0 GPa was used as the elastic modulus of the porous metal layer.

Example 2

Effects of Acetabular Cup Geometry on PEEK Layer/Porous Layer Stiffness

[0072] A Finite Element (FE) model was used to study the effective stiffness of an idealized acetabular hip cup model having a layer of PEEK interdigitated with a layer of porous metal.

[0073] A 3-D parametric FE model was developed using commercially available FE software ABAQUS 6.7 (ABAQUS, Inc., Providence RI, USA). The FE model included a femoral component, an acetabular cup, and a bone cement boundary. The femoral component was modeled to be rigid and spherical with a radius, R_{femoral} , of 16 mm. The acetabular cup was modeled to have a series of successive layers. The innermost layer is a UHMWPE layer which forms the acetabular liner of the acetabular construct and against which the femoral component articulates. The acetabular shell of the acetabular construct was modeled to have an innermost PEEK layer, an interdigitated PEEK/porous metal layer, and an outermost porous metal layer. The porous metal layer was modeled to have the properties of a material formed in accordance with Trabecular Metal™ technology, as described in detail above.

[0074] Additionally, the total polymer thickness, i.e., the combined thickness of the UHMWPE and PEEK layers, was modeled to be 7.8252 mm, with two separate distributions between the polymers tested. The first distribution was modeled with a UHMWPE layer thickness of 5.8252 mm and a PEEK layer thickness of 2.0 mm and the second distribution was modeled with a UHMWPE layer thickness of 2.0252 mm and a PEEK layer thickness of 5.8 mm. The PEEK/porous metal layer, i.e., the interdigitated layer, was modeled to have a constant thickness of 2.1248 mm and the porous metal layer was modeled to have a constant thickness of 2.077 mm. This entire construct was modeled to be embedded in a block of bone cement having an outer radius, R_{cement} , of 48.027 mm and a height, H_{cement} , of 48.027 mm.

[0075] Each of the material interfaces described above was also modeled to be in a completely bonded state, including the interface between the acetabular shell and the bone cement, and the coefficient of friction for articulation of the femoral component on the UHMWPE layer was modeled to be 0.02. The elastic modulus that was modeled for each of the layers is set forth in TABLE 3 below. All materials were modeled to have a Poisson's ratio of 0.3.

TABLE 3

<u>Layer</u>	<u>Elastic Modulus (GPa)</u>
UHMWPE	1.00
PEEK	4.00
PEEK/Porous metal	6.85
Porous metal	3.00
Bone Cement	3.00

[0076] Using the model set forth above, the femoral head was loaded by applying a 1700 lbf load to the femoral head at a 45 degree angle relative to the polar axis of the acetabular cup construct, with the outer boundary of the bone cement fully constrained, as shown in Fig. 6. The resulting displacement of the femoral component was calculated. The effective stiffness of the acetabular construct was calculated using the effective stress and the effective strain. Thus, the effective stress was calculated as the force applied to the femoral component, i.e., 1700 lbf, divided by the internal surface area of the cup, i.e., $2\pi R_{\text{femoral}}^2$. The effective strain was calculated as the radial displacement of the femoral component, i.e., Δr_{head} , divided by the total thickness of the acetabular cup construct, i.e., $t_{\text{cup}} = t_{\text{UHMWPE}} + t_{\text{PEEK}} + t_{\text{Porous / PEEK}} + t_{\text{Porous}}$.

Therefore, the effective stiffness is $E_{\text{eff}} = \frac{\sigma_{\text{eff}}}{\epsilon_{\text{eff}}} = \frac{F / 2\pi R_{\text{femoral}}^2}{\Delta r_{\text{head}} / t_{\text{cup}}}$

[0077] Utilizing the model described above, the effective stiffness of the acetabular construct having a UHMWPE layer thickness of 5.8252 mm and a PEEK layer thickness of 2.0 mm is $E_{\text{eff}} = 0.45$ GPa and the effective stiffness of the acetabular construct having a UHMWPE layer thickness of 2.0252 mm and a PEEK layer thickness of 5.8 mm is $E_{\text{eff}} = 0.54$ GPa.

Example 3

Effects of Layer Thickness on a Porous Metal/UHMWPE Orthopedic Component

[0078] A 2D stochastic microstructural FE model was developed to represent an idealized acetabular hip cup having a layer of UHMWPE interdigitated with a layer of porous metal. The internal geometry was based on a 2D random Voronoi structure generated by a

custom program in FORTRAN. This program also allows changing parametrically the thickness of the three layers, i.e., the UHMWPE layer, the interdigitated layer, and the porous metal layer, to create different designs. Tantalum microstructs, such as those formed in a material created using Trabecular Metal™ technology, are portions of ligaments 16, shown in Fig. 2a, having a loose end and defining the perimeter of the porous metal layer. These tantalum microstructs were modeled as a bilinear elastic material, as shown in Fig. 7, whereas UHMWPE was modeled as a multilinear elastic material, as shown in Fig. 8. Additionally, the various properties of the UHMWPE that were used in the model correspond to properties of an antioxidant stabilized, crosslinked UHMWPE, as set forth herein. The elements used in the FE model were PLANE183 which are eight-node quadratic elements with quadratic displacement behavior and are well suited to model irregular meshes in 2D. The FE models thus generated generally included more than 50,000 elements. Once the model was prepared, uniaxial compression was simulated using an unconfined compression loading condition, as shown in Fig. 9. Unconfined compression permits free dimensional change in directions transverse to the one in which load is applied. Fig. 9 depicts a change in dimension δ_1 in the horizontal dimension parallel to the x-axis. The object is free to deform in the vertical dimension parallel to the y-axis.

[0079] Using the above-identified assumptions, a three-parameter, two-level full factorial analysis was performed to investigate the effect of design parameters, i.e., the porosity (ϕ) of the porous metal, the thickness of the interdigitation layer (T_i), and the ratio of the UHMWPE layer thickness to the porous metal layer thickness (T_p/T_t) on the overall linear elastic stiffness, i.e., modulus, of the construct. The factors and their respective low and high levels are shown in TABLE 4 below. The overall construct thickness was kept constant with a total thickness of 10 mm. The porosity of the construct was assumed to be 65 and 85%, respectively.

TABLE 4

Factor	Low Level	High Level
Porosity, %	65	85
T_p/T_t	0.6	3.0
T_i [mm]	2	3

[0080] After determining the more significant effects from a Design of Experiments ("DOE") factorial analysis, a parametric study was performed to evaluate the mechanical behavior, such as the elastic modulus, of the modeled porous metal/UHMWPE construct. The DOE analysis determined which of the parameters identified in TABLE 4 have the greatest influence on predicted mechanical response (i.e., stiffness which defines the amount of deflection for a given load). Those parameters are % porosity, polyethylene thickness T_p and Trabecular Metal thickness T_t . Different and representative values of porosities and layer thicknesses were chosen based on current implant designs. The overall thickness of the entire construct and that of the interdigitation layer were kept constant at 10 mm and 2 mm, respectively. Additionally, values of 0.6 and 3.0 were used for the ratio of the thickness of the porous metal layer to the thickness of the polymer layer and values of 65% and 85% were used for the porosity of the porous metal layer. The matrix of thickness and porosity combinations investigated is shown below in TABLE 5. FE analysis was performed with ANSYS 10 finite analysis software (ANSYS, Inc., Canonsburg, PA, USA) to simulate uniaxial compression loading condition as set forth in Fig. 9.

TABLE 5

Analysis Factor Combinations																				
Porosity, %	65				70				75				78				85			
T _p (mm)	3	4	5	6	3	4	5	6	3	4	5	6	3	4	5	6	3	4	5	6
T _i (mm)	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2
T _t (mm)	5	4	3	2	5	4	3	2	5	4	3	2	5	4	3	2	5	4	3	2

[0081] The factorial analysis was performed for the response of the overall linear stiffness, i.e., elastic modulus, evaluated under compression loading conditions. The effects of overall response were evaluated considering a 95% confidence level. As shown in TABLE 6, the factorial analysis showed a significant interaction of the ratio of the porous metal layer thickness to the polymer layer thickness with the porosity of the metal. The thickness of the interdigitation layer was found to have little effect on the construct stiffness and was kept constant for the rest of the study. Specifically, it was determined that the percentage contribution of each significant factor and interaction to the model was 49.7% for porosity, 32.6% for the ratio of the porous metal layer thickness to the polymer layer thickness, 17.5% for the interaction of the ratio of the

porous metal layer thickness to the polymer layer thickness with the porosity of the metal, and 0.27% for the thickness of the interdigitation layer. The results are plotted in TABLE 6 and Figs. 10 and 11.

TABLE 6

Test	Ti (mm)	Tp/Tt Ratio	Porosity of Metal Layer (%)	Elastic Modulus (GPa)
1	2	0.6	65	2.413
2	3	0.6	65	2.46
3	2	3	65	1.325
4	3	3	65	1.343
5	2	0.6	85	1.134
6	3	0.6	85	1.235
7	2	3	85	0.982
8	3	3	85	1.046

[0082] Based on these results, it was shown that the higher the porosity level of the porous metal, the lower the overall linear elastic structural stiffness, or elastic modulus, of the construct. It was also found that in the linear region the increase of UHMWPE thickness produced a decrease in the overall linear stiffness, or elastic modulus. The effects of the thickness of the interdigitation layer and the polymer layer on the nonlinear overall mechanical behavior were also explored. It can be seen from Figs. 12 and 13 that in the linear region ($\epsilon < 0.02\%$), the increase of the UHMWPE layer thickness will decrease the elastic modulus (as indicated by the slope of the stress vs. strain curves) while the interdigitation layer thickness has little effect on the elastic modulus. In the nonlinear region ($\epsilon > 0.02\%$), the increase of the interdigitation layer increases the elastic modulus, while the increase in the polymer layer thickness has little change in the material.

[0083] Additionally, a clinically successful orthopedic component must have an appropriate stiffness to promote primary fixation, during the first months post-operatively, and secondary fixation with a constructive bone modeling/remodeling response. Bone resorption and subsequent implant loosening may be caused by high bone stresses as well as low bone stresses due to “stress shielding” from an overly stiff implant. Based on this, a preliminary design of a direct compression molded porous metal, such as a metal made in accordance with Trabecular Metal™ technology, and polymer, such as UHMWPE or PEEK, construct can be made to have a

stiffness that is substantially similar to the stiffness of the bone into which the construct is to be implanted. As shown by the results of the current study, the construct's bulk linear stiffness is principally affected by the polymer layer thickness and the porosity of the porous metal. Additionally, the mechanical properties of bone are nonhomogeneous and anisotropic, varying between anatomical sites. Therefore, an implant may be designed with a given porosity of the porous metal layer and thicknesses of the polymer, interdigitated, and porous metal layers that targets the structural stiffness of its host bone. In one exemplary embodiment, this can be achieved by varying the thickness of the polymer layer, as shown by the solid and dashed curves in Fig. 14 with respect to UHMWPE.

Example 4

Effects of Antioxidant Stabilizing on Layer Thickness

[0084] Using the previous FE model developed in Example 3 above, a porous metal and UHMWPE design using conventional UHMWPE, i.e., UHMWPE that has not been antioxidant stabilized, was compared with a porous metal and UHMWPE design using antioxidant stabilized UHMWPE, i.e., UHMWPE that incorporates an antioxidant, such as Vitamin E. The porosity of the porous metal was assumed to be 82%. Both conventional UHMWPE and antioxidant stabilized UHMWPE were modeled as multi-linear materials, as shown in Fig. 5, which indicates that the antioxidant stabilized UHMWPE has a lower elastic modulus but higher yield stress than the conventional UHMWPE. All models were simulated as under uniaxial compression and the geometry of all models was the same. Additionally, in the models, the thickness of both the conventional UHMWPE and the antioxidant stabilized UHMWPE layers was 3.5 mm and the thickness of the porous metal layer was 3.0 mm.

[0085] The stresses at the maximum stress point for both the conventional UHMWPE and the antioxidant stabilized UHMWPE were recorded during the loading process and the results are set forth in TABLE 7 below. Referring to TABLE 7, the results indicated that the overall response has less variation in the non-linear region and under compression loading. Regardless of the compression loading condition, it was shown that compressive stresses were up to 8% lower in the antioxidant stabilized UHMWPE than in the conventional UHMWPE. Additionally, shear stresses were up to 13% lower in the antioxidant stabilized UHMWPE than in the conventional UHMWPE. This indicates that under the same load condition, a porous

metal/UHMWPE construct utilizing an antioxidant stabilized UHMWPE may experience less overall stress than a similar construct utilizing conventional UHMWPE. Thus, in such a construct, the antioxidant stabilized UHMWPE may have equivalent or better mechanical performance than conventional UHMWPE.

TABLE 7

Von Mises Compression [MPa]			
loading time	strain	conventional	VE
0.00	0.0000	0.00	0.00
0.05	0.0025	5.20	4.47
0.10	0.0050	8.07	7.50
0.15	0.0075	10.74	9.33
0.20	0.0100	11.82	11.36
0.25	0.0125	12.93	12.63
0.30	0.0150	13.96	13.43
0.35	0.0175	14.91	14.16
0.40	0.0200	15.80	14.85
0.45	0.0225	16.56	15.51
0.50	0.0250	17.17	16.14
0.55	0.0275	17.75	16.74
0.60	0.0300	18.29	17.30
0.65	0.0325	18.74	17.78
0.70	0.0350	19.15	18.22
0.75	0.0375	19.51	18.62
0.80	0.0400	19.86	19.01
0.85	0.0425	20.22	19.39
0.90	0.0450	20.54	19.76
0.95	0.0475	20.82	20.08
1.00	0.0500	21.09	20.39

Example 5

Feasibility Study of α -Tocopherol Acetate

[0086] Throughout the various Examples 5-13 set forth below, irradiated UHMWPE blends (i.e., antioxidant stabilized, crosslinked UHMWPE) are used, which have been irradiated according to one of three different irradiation methods. As used in the Examples below, the term "UHMWPE blend" refers to an antioxidant stabilized UHMWPE or, if used to refer to an antioxidant stabilized UHMWPE after exposure to crosslinking irradiation, an antioxidant stabilized, crosslinked UHMWPE. As set forth above, differences in the irradiation conditions and techniques may affect the resulting material properties of the UHMWPE blend. Therefore, in order to properly analyze and compare the results set forth in the Examples and corresponding TABLES, each of the irradiated UHMWPE blends used in the Examples below are identified, where relevant, as having been irradiated according to the one of the methods set forth below in TABLE 8. Additionally, the electron beam source is calibrated by performing dosimetry at low irradiation doses and then parametrically determining the activation of the electron beam source

needed to achieve higher doses. As a result, at higher irradiation doses, differences may exist between the actual dose and the parametrically determined dose, which may cause differences in the material properties of the irradiated UHMWPE blends.

5

TABLE 8
Irradiation Methods

	Method A	Method B	Method C
Dose Rate (kGy-m/min)	30-75	16-25	75-240
Dose Level (kGy)	160-190	133-217	90-200
Electron Beam Energy (MeV)	10	12	10
Method of Dosimetry	Water Calorimeter	Aluminum Calorimeter	Radiochromic Film

[0087] The feasibility of blending α -tocopherol acetate with UHMWPE was investigated. α -tocopherol acetate was obtained from DSM Nutritional Products, Ltd. of Geleen, Netherlands and medical grade UHMWPE powder GURTM 1050 was obtained from Ticona, having North American headquarters located in Florence, Kentucky. Isopropanol was then added to the α -tocopherol acetate as a diluent and the α -tocopherol acetate was solvent blended with the UHMWPE powder. The blending continued until two different UHMWPE/ α -tocopherol acetate blends were obtained, one UHMWPE blend having 0.05 wt. % α -tocopherol acetate and the other UHMWPE blend having 0.5 wt. % α -tocopherol acetate. Each of the UHMWPE blends were then compression molded to form four one-inch-thick pucks. Two pucks of each UHMWPE blend, i.e., two pucks of the UHMWPE blend having 0.05 wt. % α -tocopherol acetate and two pucks of the UHMWPE blend having 0.5 wt. % α -tocopherol acetate, were preheated to 120°C in a Grieve convection oven, available from The Grieve Corporation of Round Lake, Illinois. The pucks were held at 120°C for 8 hours. After the expiration of 8 hours, the pucks were irradiated at 10 MeV, 50 kGy-m/min dose rate at 65 kGy and 100 kGy dose at Iotron Industries Canada Inc. located in Port Coquitlam, BC, Canada.

[0088] The remaining two pucks of each UHMWPE blend, i.e., two pucks of the UHMWPE blend having 0.05 wt. % α -tocopherol acetate and two pucks of the UHMWPE blend having 0.5 wt. % α -tocopherol acetate, were heated to 40°C overnight. The next morning, the

remaining two pucks of each UHMWPE blend were irradiated at 10 MeV, 50 kGy-m/min dose rate at 100 kGy dose at Iotron Industries Canada Inc. located in Port Coquitlam, BC, Canada.

[0089] After irradiation, all of the pucks were cut in half and a film was cut from the center of each puck. The films were then subjected to FTIR analysis using a Bruker Optics FTIR Spectrometer, available from Bruker Optics of Billerica, Massachusetts. Both halves of each puck were then machined into flat sheets approximately 1/8 inch thick. One half of the flat sheets were immediately subjected to FTIR. The other half of the flat sheets were then subjected to accelerated aging in accordance with the American Society for Testing and Materials (ASTM) Standard F-2003, Standard Practice for Accelerated Aging of Ultra-High Molecular Weight Polyethylene after Gamma Irradiation in Air. Tensile specimens formed from the flat sheets were subjected to accelerated aging and were then subjected to FTIR analysis. The OI and wt. % of α -tocopherol acetate were determined from the FTIR results, set forth in TABLE 9, below, and in Fig. 14. However, there were interference peaks in the FTIR results that prevented measurement of OI for the 0.5 wt. %, 65 kGy, unaged sample.

TABLE 9
FTIR Results

wt. % tocopherol	Dose, kGy	Condition	OI	wt. % tocopherol, meas.
0.50	100	Un-aged	<0 / <0	0.17 / 0.15
		Aged	0.0323	0.15
0.50	65	Un-aged	*interference*	0.14 / 0.15
		Aged	0.0083	0.14
0.05	100	Un-aged	0.0300 / 0.0948	0.01 / 0.00
		Aged	0.0647	<0
0.05	65	Un-aged	0.0376 / 0.0940	0.02 / 0.00
		Aged	0.0647	<0

[0090] The FTIR results revealed that the OI of the UHMWPE blend having 0.05 wt. % α -tocopherol acetate was generally higher than the OI of the UHMWPE blend having 0.50 wt. % α -tocopherol acetate. This is believed to be because these samples still contained α -tocopherol acetate after irradiation. As a result, the α -tocopherol acetate was still available in these samples to react with free radicals and reduce the oxidative degradation of the UHMWPE blend. Additionally, the FTIR results showed that virtually no α -tocopherol acetate was left after irradiation of the UHMWPE blend having 0.05 wt. % α -tocopherol acetate and that about one-

third of the α -tocopherol acetate was left after irradiation of the UHMWPE blend having 0.5 wt. % α -tocopherol acetate. Further, as shown in TABLE 10 below, tensile properties were similar for both the UHMWPE blends that were subjected to accelerated aging and the UHMWPE blends that were not subjected to accelerated aging. Finally, the FTIR results suggested that the UHMWPE blends containing α -tocopherol acetate have similar stabilization properties, i.e., a similar ability to prevent oxidative degeneration, as UHMWPE blends containing similar concentration of d, l- α -tocopherol.

TABLE 10
Mechanical Properties

wt. % α -tocopherol acetate	Dose, kGy (Temperature)	Condition	Elongation, %	Yield, MPa	UTS, MPa
0.50	100 (40°C)	Un-aged	356.7	23.1	61.7
		Aged	360.1	25.3	65.3
0.50	65 (120°C)	Un-aged	384.6	21.8	61.4
		Aged	378.4	24.2	65.1
0.05	100 (40°C)	Un-aged	342.9	21.5	56.7
		Aged	288.6	25.2	61.0
0.05	65 (120°C)	Un-aged	352.4	21.4	57.4
		Aged	287.7	25.2	59.6

Example 6

Chemical Properties of UHMWPE Blended with Tocopherol

[0091] The chemical properties of d/l- α -tocopherol mechanically blended with a UHMWPE powder which was slab molded into bars and electron beam irradiated were investigated. To perform this investigation, Design ExpertTM 6.0.10 software, obtained from Stat-Ease, Inc. Minneapolis, MN, was utilized to setup a modified fractional factorial Design of Experiment (DOE). The DOE evaluated five different variables: UHMWPE resin type, wt. % of d/l- α -tocopherol, preheat temperature, dose rate, and irradiation dose.

[0092] GURTM1050 and GURTM1020 medical grade UHMWPE powders were obtained from Ticona, having North American headquarters in Florence, Kentucky. d/l- α -tocopherol was obtained from DSM Nutritional Products, Ltd. of Geleen, Netherlands. The GURTM1050 and GURTM1020 were separately mechanically blended with the d/l- α -tocopherol by low intensity blending using a Diosna P100 Granulator, available from Diosna GmbH of Osnabrück,

Germany, a subsidiary of Multimixing S.A. Both the GUR 1050 and the GUR 1020 resins were mixed with the d/l- α -tocopherol in several batches to create UHMWPE blends of both resin types having 0.2 wt. %, 0.5 wt. %, and 1.0 wt. % d/l- α -tocopherol. Each batch of blended material was compression molded into a slab and cut into bars of various sizes. Each of the resulting bars was then preheated by heating to a preheat temperature in a Grieve convection oven, available from The Grieve Corporation of Round Lake, Illinois. The preheat temperature was selected from 40°C, 100°C, 110°C and 122.2°C, as set forth in TABLE 11 below.

[0093] After being preheated, the UHMWPE blend bars were electron beam irradiated according to Method C, set forth in TABLE 8 above, at a selected dose rate until a selected total irradiation dose was administered. The dose rate was selected from 75 kGy-m/min, 155 kGy-m/min, and 240 kGy-m/min and the total irradiation dose was selected from 90 kGy, 120 kGy, 150 kGy, and 200 kGy. The portion of each bar was then microtomed into 200 micron thick films. These films were then subjected to FTIR analysis on a Bruker Optics FTIR spectrometer, available from Bruker Optics of Billerica, Massachusetts. The FTIR results were analyzed to determine the VEI, wt. % d/l- α -tocopherol, the OI, and the TVI. The VEI and wt. % d/l- α -tocopherol were determined by calculating the ratio of the area under the d/l- α -tocopherol peak at 1275-1245 cm^{-1} on the resulting FTIR chart to the area under the polyethylene peak at 1392-1330 cm^{-1} and at 1985-1850 cm^{-1} . The OI was determined by calculating the ratio of the area under the carbonyl peak on the FTIR chart at 1765-1680 cm^{-1} to the area of the polyethylene peak at 1392-1330 cm^{-1} . The TVI was determined by calculating the ratio of the area on the FTIR chart under the vinyl peak at 980-947 cm^{-1} to the area under the polyethylene peak at 1392-1330 cm^{-1} .

[0094] After the initial VEI, wt. % d/l- α -tocopherol and TVI were determined from the FTIR analysis of the thin films, each of the thin films were accelerated aged according to ASTM Standard F-2003, Standard Practice for Accelerated Aging of Ultra-High Molecular Weight Polyethylene after Gamma Irradiation in Air. The accelerated aged films were again subjected to FTIR analysis on a Bruker Optics FTIR spectrometer, available from Bruker Optics of Billerica, Massachusetts. The resulting FTIR charts were analyzed to determine VEI, wt. % d/l- α -tocopherol, OI, and TVI according to the methods set forth above. Once subjected to FTIR analysis, the aged files were placed in boiling hexane and allowed to remain there for 24 hours to extract the d/l- α -tocopherol. After extraction of the d/l- α -tocopherol, the aged films were again

subjected to FTIR analysis on the Bruker Optics FTIR spectrometer. The resulting FTIR chart was then analyzed to determine the OI in accordance with the method set forth above. The additional FTIR analysis was performed to eliminate the d/l- α -tocopherol peak from interfering with the oxidation peaks. An analysis of the results set forth in TABLE 11 below indicate that selecting a warmer preheat temperature may result in a lower OI and may also result in some of the d/l- α -tocopherol remaining in the UHMWPE after irradiation.

TABLE 11

FTIR Results of Irradiated UHMWPE Blended with d/l- α -tocopherol

Run	Pre-heat (°C)	Dose (kGy)	VE level (°C)	Dose Rate (kGy-m/min)	Resin Type (GUR)
1	122	150	1	75	1020
2	40	200	0.2	155	1020
3	122	90	0.5	75	1020
4	122	200	0.2	155	1020
5	40	90	0.2	240	1050
6	122	90	0.2	75	1050
7	40	150	0.2	75	1050
8	122	150	0.2	75	1020
9	40	90	1	240	1050
10	40	200	0.5	155	1020
11	122	90	0.2	240	1020
12	40	90	1	75	1020
13	40	150	0.5	75	1020
14	122	150	0.2	240	1050
15	122	150	0.5	240	1020
16	40	150	0.2	240	1020
17	40	90	0.2	75	1020
18	122	200	0.5	155	1020
19	122	90	1	240	1020
20	40	150	1	240	1020
21	40	200	1	155	1020
22	122	200	1	155	1020
23	40	200	0.2	155	1050
24	122	200	0.2	155	1050
25	40	200	0.5	155	1050
26	122	200	0.5	155	1050
27	40	200	1	155	1050
28	122	200	1	155	1050
29	40	120	0.5	157.5	1050
30	122	120	0.5	157.5	1050
31	40	120	1	157.5	1050

32	122	120	1	157.5	1050
33	40	90	1	75	1050
34	122	90	0.5	75	1050
35	40	150	0.5	75	1050
36	122	150	1	75	1050
37	40	90	0.5	240	1050
38	122	90	1	240	1050
39	40	150	1	240	1050
40	122	150	0.5	240	1050

Run	VE% 1370nm IR peak	VE% 1900nm IR peak	VE Index 1370nm IR peak	VE Index 1900nm IR peak	VE% (aged) 1370 nm IR peak
1	0.803	0.682	0.046	0.171	0.493
2	0.040	0.048	0.004	0.015	0.022
3	0.359	0.321	0.021	0.082	0.248
4	0.045	0.054	0.004	0.016	0.037
5	0.047	0.055	0.004	0.016	0.063
6	0.061	0.071	0.005	0.020	0.073
7	0.011	0.025	0.003	0.009	0.017
8	0.031	0.042	0.004	0.013	0.033
9	0.194	0.165	0.012	0.044	0.272
10	0.731	0.626	0.042	0.157	0.545
11	0.075	0.078	0.006	0.022	0.081
12	0.882	0.738	0.050	0.185	0.417
13	0.286	0.222	0.017	0.058	0.274
14	0.058	0.072	0.005	0.021	0.056
15	0.162	0.151	0.011	0.040	0.279
16	0.051	0.053	0.005	0.016	0.050
17	0.078	0.076	0.006	0.022	0.044
18	0.721	0.634	0.041	0.159	0.524
19	0.769	0.688	0.044	0.173	0.430
20	0.781	0.597	0.044	0.150	0.531
21	0.769	0.591	0.044	0.149	0.560
22	0.765	0.607	0.044	0.153	0.575
23	0.028	0.034	0.003	0.011	0.016
24	0.051	0.053	0.005	0.016	0.041
25	0.288	0.249	0.018	0.064	0.281
26	0.320	0.282	0.019	0.073	0.309
27	0.284	0.222	0.017	0.058	0.281
28	0.308	0.241	0.019	0.062	0.295
29	0.613	0.550	0.035	0.139	0.489
30	0.753	0.700	0.043	0.176	0.445
31	0.283	0.240	0.017	0.062	0.279
32	0.306	0.288	0.019	0.074	0.259
33	0.779	0.706	0.044	0.177	0.429
34	0.328	0.314	0.020	0.080	0.209

35	0.143	0.125	0.010	0.034	0.247
36	0.803	0.758	0.046	0.190	0.442
37	0.332	0.291	0.020	0.075	0.262
38	0.741	0.731	0.042	0.183	0.390
39	0.790	0.658	0.045	0.165	0.524
40	0.327	0.301	0.020	0.077	0.282

Run	VE% (aged) 1900nm IR peak	VE Index (aged) 1370 nm IR peak	VE Index (aged) 1900nm IR peak	OI (Extraction- aged) FTIR	TVI FTIR	TVI (aged) FTIR
1	0.425	0.029	0.108	-0.010	0.061	0.062
2	0.033	0.003	0.011	0.078	0.073	0.072
3	0.226	0.015	0.059	-0.008	0.046	0.052
4	0.048	0.004	0.015	0.025	0.081	0.082
5	0.068	0.005	0.020	0.039	0.039	0.040
6	0.080	0.006	0.023	0.000	0.054	0.048
7	0.029	0.003	0.010	0.086	0.068	0.064
8	0.042	0.004	0.013	0.035	0.076	0.074
9	0.231	0.017	0.060	0.011	0.048	0.040
10	0.468	0.032	0.118	0.005	0.076	0.077
11	0.084	0.006	0.024	0.007	0.055	0.054
12	0.359	0.025	0.091	-0.001	0.034	0.035
13	0.211	0.017	0.055	0.082	0.075	0.074
14	0.070	0.005	0.020	0.001	0.072	0.072
15	0.244	0.017	0.063	0.025	0.081	0.084
16	0.052	0.005	0.016	0.075	0.063	0.062
17	0.048	0.004	0.015	0.098	0.045	0.045
18	0.458	0.030	0.116	-0.004	0.083	0.085
19	0.392	0.025	0.100	-0.010	0.041	0.047
20	0.409	0.031	0.104	0.087	0.080	0.078
21	0.429	0.032	0.109	0.061	0.081	0.087
22	0.457	0.033	0.116	-0.007	0.092	0.091
23	0.025	0.003	0.009	0.120	0.078	0.079
24	0.045	0.004	0.014	0.032	0.084	0.085
25	0.241	0.017	0.062	0.009	0.075	0.073
26	0.268	0.019	0.069	-0.002	0.085	0.083
27	0.220	0.017	0.057	0.024	0.080	0.079
28	0.229	0.018	0.059	0.042	0.094	0.096
29	0.429	0.028	0.109	0.040	0.053	0.058
30	0.421	0.026	0.107	0.000	0.063	0.064
31	0.236	0.017	0.061	0.063	0.061	0.061
32	0.244	0.016	0.063	0.004	0.065	0.066
33	0.397	0.025	0.101	0.036	0.040	0.041
34	0.205	0.013	0.053	-0.005	0.049	0.051
35	0.211	0.015	0.055	0.067	0.072	0.068
36	0.423	0.026	0.107	-0.012	0.055	0.058
37	0.234	0.016	0.061	0.029	0.041	0.048
38	0.391	0.023	0.099	-0.004	0.038	0.042
39	0.440	0.030	0.111	0.043	0.073	0.076
40	0.262	0.017	0.068	-0.004	0.068	0.068

Example 7Mechanical Properties of UHMWPE Blended with d/l- α -tocopherol

[0095] The mechanical properties of d/l- α -tocopherol mechanically blended with a UHMWPE powder which was slab molded into bars and electron beam irradiated were investigated. To perform this investigation, Design Expert 6.0.10 software, obtained from Stat-Ease, Inc. Minneapolis, MN, was utilized to setup a modified fractional factorial Design of Experiment (DOE). The DOE evaluated five different variables: UHMWPE resin type, weight percent of d/l- α -tocopherol, preheat temperature, dose rate, and irradiation dose.

[0096] GUR 1050 and GUR 1020 medical grade UHMWPE powders were obtained from Ticona, having North American headquarters in Florence, Kentucky. d/l- α -tocopherol was obtained from DSM Nutritional Products, Ltd. of Geleen, Netherlands. The GUR 1050 and GUR 1020 were separately mechanically blended with the d/l- α -tocopherol by low intensity blending using a Diosna P100 Granulator, available from Diosna GmbH of Osnabrück, Germany, a subsidiary of Multimixing S.A. Both the GUR 1050 and the GUR 1020 resins were mixed with the d/l- α -tocopherol in several batches to create UHMWPE blends of both resin types having 0.2 wt. %, 0.5 wt. %, and 1.0 wt. % d/l- α -tocopherol. Each batch of blended material was compression molded into a slab and cut into bars. Each of the resulting bars was then preheated by heating the bars in a Grieve convection oven, available from The Grieve Corporation of Round Lake, Illinois, to a preheat temperature. The preheat temperature was selected from 40°C, 100°C, 110°C and 122.2°C, as set forth in TABLE 12 below.

[0097] After being preheated, the UHMWPE blend bars were electron beam irradiated according to Method C, set forth in TABLE 8 above, at a selected dose rate until a selected total irradiation dose was administered. The dose rate was selected from 75 kGy-m/min, 155 kGy-m/min, and 240 kGy-m/min and the total irradiation dose was selected from 90 kGy, 120 kGy, 150 kGy, 200 kGy, and 250 kGy. Type V tensile specimens, as defined by the American Society for Testing and Materials (ASTM) Standard D638, Standard Test Method for Tensile Properties of Plastics, were machined from each of the UHMWPE blend bars. The Type V tensile specimens were then subjected to ultimate tensile elongation, UTS, and YS testing in accordance with ASTM Standard D638. Izod specimens were also machined from each of the UHMWPE blend bars and tested for izod impact strength according to ASTM Standard D256, Standard Test

Methods for Determining the Izod Pendulum Impact Resistance of Plastics. Dynamic mechanical analysis (DMA) specimens were also machined from each of the UHMWPE blend bars and tested using a Model DMA 2980 Dynamic Mechanical Analyzer from TA Instruments of New Castle, Delaware.

[0098] An analysis of the results indicates that the total irradiation dose had an influence on the izod impact strength, ultimate tensile elongation, and yield strength of the UHMWPE blends. Additionally, the preheat temperature had an influence on the ultimate tensile strength and yield strength. In contrast, the weight percent of d/l- α -tocopherol had an influence on ultimate tensile elongation and the dynamic mechanical analysis. Additional results from the testing are set forth below in TABLE 12, below.

TABLE 12

Mechanical Properties of UHMWPE Blended with d/l- α -tocopherol

Std	Preheat °C	Dose kGy	VE Conc.	Dose Rate kGy-m/min	Resin	Izod kJ/m ²	Elongation %	UTS M Pa	YS M Pa	DMA M Pa
1	40	90	0.2	75	1020	90.79	Not Tested			5.45
2	122.2	90	0.2	75	1050	74.8	348.8	50.38	21.72	6.12
3	40	150	0.2	75	1050	59.66	300.9	56.2	25.1	6.78
4	100	150	0.2	75	1020	66.05	314.8	52.42	25.24	5.76
5	40	90	1	75	1020	111.19	Not Tested			4.36
6	122.2	90	0.5	75	1020	91.55	Not Tested			5.1
7	40	150	0.5	75	1020	79.96	355.7	55.83	26.28	4.96
8	100	150	1	75	1020	81.25	Not Tested			4.78
9	40	90	0.2	240	1050	82.01	319.3	56.83	23.06	6.18
10	122.2	90	0.2	240	1020	84.5	Not Tested			5.43
11	40	150	0.2	240	1020	67.53	293.1	56.19	26.87	5.96
12	100	150	0.2	240	1050	67.75	307.2	52.95	25.21	6.31
13	40	90	1	240	1050	106.17	411.2	61.76	24.89	4.53
14	122.2	90	1	240	1020	94.66	Not Tested			4.83
15	40	150	1	240	1020	93.79	Not Tested			5.6
16	100	150	0.5	240	1020	73.08	342.2	51.54	23.39	5.22
17	40	120	0.5	157.5	1050	99.87	374.6	57.96	23.23	5.06
18	110	120	0.5	157.5	1050	90.67	363.6	50.97	22.15	5.42
19	40	120	1	157.5	1050	94.34	352.5	58.5	23.64	5.53
20	110	120	1	157.5	1050	85.01	344.9	48.98	21.95	5.72
21	40	90	1	75	1050	107.07	396.4	61.25	23.13	5.02
22	122.2	90	0.5	75	1050	93.44	375.8	51.47	21.92	5.7
23	40	150	0.5	75	1050	82.09	330.4	56.65	25.78	4.62
24	100	150	1	75	1050	88.28	Not Tested			5.4
25	40	90	0.5	240	1050	102.39	36.9	58.4	23.31	5.16
26	122.2	90	1	240	1050	96.6	381.9	50.3	21.29	5.44
27	40	150	1	240	1050	89.5	Not Tested			5.16
28	100	150	0.5	240	1050	78.51	332.9	50.18	22.08	5.6
29	40	200	0.2	155	1020	55.98	246.2	52.37	27.23	6.32
30	110	200	0.2	155	1020	52.98	268.4	48.28	24.82	6.05
31	40	200	0.5	155	1020	74.64	310.7	53.53	25.42	5.38
32	110	200	0.5	155	1020	65.47	309.1	49.13	24.21	5.31
33	40	200	1	155	1020	72.67	362.9	55.62	25.9	4.63
34	110	200	1	155	1020	66.62	349.7	50.45	24.24	4.93
35	40	200	0.2	155	1050	57.82	226.4	50.92	25.3	7.15
36	110	200	0.2	155	1050	59.04	259.4	46.4	23.7	6.57
37	40	200	0.5	155	1050	67.49	280.6	51.91	26.23	5.88
38	110	200	0.5	155	1050	64.6	304.8	51.23	25.14	5.7
39	40	200	1	155	1050	82.01	328.9	53.79	24.43	5.15
40	110	200	1	155	1050	69.42	329.7	49.54	23.12	5.27
41	100	150	0.2	240	1050	Not Tested	307.2	52.96	23.27	Not Tested
42	100	150	0.2	240	1020	Not Tested	288.6	49.28	23.35	Not Tested

Example 8Wear Properties of UHMWPE Mixed with d,l- α -tocopherol

[0099] The wear properties of UHMWPE mechanically blended with d/l- α -tocopherol and exposed to electron beam irradiation was investigated. To perform this investigation, Design Expert 6.0.10 software, obtained from Stat-Ease, Inc. Minneapolis, MN, was utilized to setup a

modified central composite Design of Experiment (DOE). The DOE evaluated five different variables: preheat temperature, dose rate, total dose administered, d,l- α -tocopherol concentration, and cooling period, i.e., the elapsed time from end of the preheat until initial exposure to irradiation.

[00100] GURTM1050 medical grade UHMWPE powder was obtained from Ticona, having

5 North American headquarters in Florence, Kentucky. d/l- α -tocopherol was obtained from DSM Nutritional Products, Ltd of Geleen, Netherlands. The GURTM1050 was mechanically mixed with the d/l- α -tocopherol using a High Intensity Mixer, available from Eirich Machines of Gurnee, Illinois. The GURTM1050 resin was mixed with the d/l- α -tocopherol in several batches to create UHMWPE blends having a selected wt. % of d/l- α -tocopherol. The wt. % of d/l- α -tocopherol was
10 selected from 0.14 wt. %, 0.19 wt. %, and 0.24 wt. % d/l- α -tocopherol. Each of the blends were then consolidated and formed into 2.5 inch diameter and 1 inch thick pucks. Each of the resulting pucks was then preheated by heating the pucks in a Grieve convection oven, available from The Grieve Corporation of Round Lake, Illinois, to a preheat temperature. The preheat temperature was selected from 85°C, 100°C, and 115°C, as set forth in TABLE 13 below.

15 [00101] After being preheated, the UHMWPE blend pucks were then removed from the convection oven for a cooling period. The cooling period was selected from 7 minutes, 14 minutes, and 21 minutes, as set forth in TABLE 13 below. The pucks were then electron beam irradiated according to Method A, set forth in TABLE 8 above, at a selected dose rate until a selected total irradiation dose was administered. The dose rate was selected from 30 kGy-m/min, 52.5 kGy-
20 m/min, and 75 kGy-m/min and the total irradiation dose was selected from 160 kGy, 175 kGy, and 190 kGy.

[00102] Pin-on-disc (POD) specimens in the form cylinders having a 9 mm diameter and 13 mm thickness were then machined from the UHMWPE blend pucks. A bidirectional pin-on-disc wear tester was then used to measure the wear rate of UHMWPE pins articulating against polished
25 cobalt-chrome discs lubricated by 100% bovine serum. These measurements were made in accordance with the teachings of Bragdon, C.R., et al., in A new pin-on-disk wear testing method for simulating wear of polyethylene on cobalt-chrome alloy in total hip arthroplasty, published in the Journal of Arthroplasty, Vol. 16, Issue 5, 2001, on pages 658-65. The bidirectional

motion for the pin-on-disc wear tester was generated by a computer controlled XY table, available from the Compumotor Division of Parker Hannifin of Cleveland, Ohio, which was programmed to move in a 10 mm by 5 mm rectangular pattern. Affixed atop the XY table was a basin containing six cobalt-chrome discs polished to an implant quality finish. The XY table and basin were mounted on a servo-hydraulic MTS machine, available from MTS of Eden Prairie, Minnesota. The MTS machine then loaded the UHMWPE blend pin specimens against the polished cobalt-chrome discs.

[00103] The MTS machine was programmed to produce a Paul-type curve in synchronization with the motion of the XY table. A Paul-type curve is explained in detail in Forces Transmitted By Joints in the Human Body by J.P. Paul and published by in the Proceedings Institution of Mechanical Engineers at Vol. 181, Part 37, pages 8- 15. The peak load of the Paul-type loading curve corresponded to a peak contact pressure of 6.5 MPa between each of the UHMWPE pin specimens and the cobalt-chrome discs. Tests were conducted at 2 Hz to a total of 1.128×10^6 cycles. Analysis of the results indicated that the wear properties are affected by both the concentration of d/l- α -tocopherol and the total irradiation dose. Specifically, the results indicated that increasing the d/l- α -tocopherol concentration increased the wear rate of the UHMWPE blends, while increasing the total irradiation dose decreased the wear rate of the UHMWPE blends. Additionally, the results indicated that both dose rate and the cooling period had substantially no impact on the wear rate of the UHMWPE.

TABLE 13

Wear Properties of UHMWPE Mixed with d/l- α -tocopherol

Run	Block	Preheat (°C)	Dose (kGy)	VE%	Dose Rate (kGy- m/min.)	Oven to Beam (minutes)	POD Wear (mg/Mc)
1	Block 1	85	190	0.11	30	7	0.96
2	Block 1	115	190	0.11	30	7	1.14
3	Block 1	115	190	0.11	30	21	0.76
4	Block 1	85	190	0.11	30	21	0.81
5	Block 1	115	160	0.11	30	7	1.86
6	Block 1	85	160	0.11	30	7	1.37
7	Block 1	115	160	0.11	30	21	1.53
8	Block 1	85	160	0.11	30	21	1.57
9	Block 1	85	160	0.22	75	7	2.94

10	Block 1	115	160	0.22	75	7	2.15
11	Block 1	85	160	0.22	75	21	2.41
12	Block 1	115	160	0.22	75	21	1.96
13	Block 1	115	190	0.22	75	7	2.57
14	Block 1	85	190	0.22	75	7	1.87
15	Block 1	115	190	0.22	75	21	1.87
16	Block 1	85	190	0.22	75	21	2.24
17	Block 1	100	175	0.165	52.5	14	0.89
18	Block 1	100	175	0.165	52.5	14	1.18
19	Block 1	100	175	0.165	52.5	14	1.24
20	Block 1	100	175	0.165	52.5	14	1.27

Example 9

Temperature Variations at the UHMWPE Blend/Substrate Interface

[00104] GUR 1050 medical grade UHMWPE powder was obtained from Ticona, having North American headquarters in Florence, Kentucky. d/l- α -tocopherol was obtained from DSM Nutritional Products, Ltd of Geleen, Netherlands. The GUR 1050 was mechanically blended with the d/l- α -tocopherol using a High Intensity Mixer, available from Eirich Machines of Gurnee, Illinois. The GUR 1050 resin was mixed with the d/l- α -tocopherol to create a UHMWPE blend having 0.2 wt. % d/l- α -tocopherol.

[00105] A portion of the UHMWPE blend was then compression molded into a block. Another portion of the UHMWPE blend was compression molded into a substrate to create a preform. The substrate was a 70 mm diameter porous metal substrate in the form of a near-net shape acetabular shell. The porous metal substrate was produced using Trabecular Metal™ technology generally available from Zimmer, Inc., of Warsaw, Indiana, and described in detail above. This process was repeated to create five different preforms. The preforms were then individually heated to a preheat temperature in a Grieve convection oven, available from The Grieve Corporation of Round Lake, Illinois. The preheat temperature was selected from 100°C, 120°C, and 125°C. Once heated to the selected preheat temperature, the preforms were irradiated using Method B, set forth in TABLE 8 above, until a total irradiation dose was received. The total irradiation dose was selected from 50 kGy, 75 kGy, and 150 kGy. Additionally, the UHMWPE block was heated to a preheat temperature of 100°C and irradiated using Method B until a total irradiation dose of 150 kGy was received by the UHMWPE block.

[00106] The temperature of the preforms was measured at the UHMWPE blend/substrate interface, at a point in the UHMWPE blend adjacent to the UHMWPE blend/substrate interface, and at a point in the center of the UHMWPE blend. Each of the temperature measures were taken using a Type J thermocouple. Additionally, the temperature at the center of the UHMWPE blend block was also measured using a Type J thermocouple. Based on the results, the presence of a porous substrate resulted in higher temperature readings in the UHMWPE blend. This is likely a result of substrate reaching a higher maximum temperature than the UHMWPE during irradiation.

Example 10

Effect of Substrate Orientation on UHMWPE Blend

[00107] GUR 1050 medical grade UHMWPE powder was obtained from Ticona, having North American headquarters in Florence, Kentucky. d/l- α -tocopherol was obtained from DSM Nutritional Products, Ltd of Geleen, Netherlands. The GUR 1050 was mechanically blended with the d/l- α -tocopherol using a High Intensity Mixer, available from Eirich Machines of Gurnee, Illinois. The GUR 1050 resin was mixed with the d/l- α -tocopherol to create a UHMWPE blend having 0.5 wt. % d/l- α -tocopherol.

[00108] A portion of the UHMWPE blend was compression molded into a substrate to create a preform. The substrate was a 70 mm diameter porous metal substrate in the form of a near-net shape acetabular shell. The porous metal substrate was produced using Trabecular Metal™ technology generally available from Zimmer, Inc., of Warsaw, Indiana, and described in detail above. This process was repeated to create three different preforms. The preforms were then heated in a convection oven to a preheat temperature of 110°C for a minimum of 12 hours. Two of the preforms were then irradiated using Method A, as set forth in TABLE 8 above, with the substrate of one of the preforms facing the irradiation source and the substrate of the other preform facing away from the irradiation source. With the preforms in these positions, they were exposed to a first, 100 kGy dose of irradiation. The preforms were then allowed to sit in ambient air for 20 minutes. After the expiration of 20 minutes, the preforms were exposed to a second, 100 kGy dose of irradiation, for a total irradiation dose of 200 kGy.

[00109] The remaining preform was irradiated using Method B, as set forth in TABLE 8 above, with the substrate of the preform facing the irradiation source. With the preform in this

position, the preform was exposed to a first, 100 kGy dose of irradiation. The preform was then placed in a convection oven which maintained a constant temperature of 110°C. After the expiration of four hours, the preform was removed from the convection oven and exposed to a second, 100 kGy dose of irradiation, for a total irradiation dose of 200 kGy.

[00110] Each of the preforms was then cut through the center and the substrate removed. The UHMWPE blend was then microtomed and subjected to FTIR analysis using a Bruker FTIR Spectrometer, available from Bruker Optics of Billerica, Massachusetts, to determine the TVI of the UHMWPE blend. This analysis was performed on the thickest part of the specimens. A sample of the UHMWPE blend was then subjected to DSC using a TA Instruments Q1000, available from TA Instruments of New Castle, Delaware, to determine the percent crystallinity of the UHMWPE blend. This analysis was repeated for samples of the UHMWPE blend taken from different locations.

[00111] In both of the monoblocks that were irradiated with the substrate facing the irradiation source, a band of discoloration, i.e., translucence, can be seen along the edge of the UHMWPE blend that interfaced with the substrate. As shown in Fig. 16, the FTIR analysis showed a substantial decline in the TVI of the UHMWPE blend at a point just past the interface between the UHMWPE blend and the substrate. Additionally, the percent crystallinity at a point in the center of the UHMWPE blend was approximately 59%. The percent crystallinity decreased as the UHMWPE blend approached the interface with the substrate, with the percent crystallinity reaching 48% in the translucent region near the UHMWPE blend/substrate interface, as shown in TABLE 14 below. In the preform that was irradiated with the substrate facing away from the irradiation source, the TVI of the UHMWPE blend was substantially more uniform throughout the UHMWPE blend and the percent crystallinity varied by only 2.2%. This may be a result of more uniform crosslinking occurring in the preform in which the substrate faced away from the irradiation source during irradiation.

TABLE 14
Percent Crystallinity of UHMWPE Blend

Specimen	% Crystallinity at the center of the UHMWPE Blend	% Crystallinity at the UHMWPE Blend/Substrate Interface
Substrate Toward Irradiation Source	59.29%	48.67%
Substrate Toward Irradiation Source	58.60%	47.96%
Substrate Away from Irradiation Source	59.88%	57.66%

Example 11
Effect of Irradiation Dose on UHMWPE Blend

[00112] Design Expert 6.0.10 software, obtained from Stat-Ease, Inc. Minneapolis, MN, was utilized to setup a central composite response surface Design of Experiment (DOE). The DOE evaluated four different variables: d,l- α -tocopherol concentration, preheat temperature, total irradiation dose administered, and irradiation dose per pass.

[00113] GUR 1050 medical grade UHMWPE powder was obtained from Ticona, having North American headquarters in Florence, Kentucky. d,l- α -tocopherol was obtained from DSM Nutritional Products, Ltd of Geleen, Netherlands. The GUR 1050 was mechanically mixed with the d,l- α -tocopherol using a High Intensity Mixer, available from Eirich Machines of Gurnee, Illinois. The GUR 1050 resin was mixed with the d,l- α -tocopherol in several batches to create UHMWPE blends having a selected wt. % of d,l- α -tocopherol. The wt. % of d,l- α -tocopherol was selected from 0.10 wt. %, 0.20 wt. %, 0.35 wt. %, 0.50 wt. %, and 0.60 wt. % d,l- α -tocopherol. Each of the blends was then compression molded into a substrate to create a preform. The substrate was a 70 mm outer diameter porous metal substrate in the form of a near-net shape acetabular shell. The porous metal substrate was produced using Trabecular Metal™ technology generally available from Zimmer, Inc., of Warsaw, Indiana, and described in detail above.

[00114] The resulting preforms were then placed inside a piece of expandable braided polyethylene terephthalate sleeving and vacuum sealed inside an aluminum-metallized plastic film pouch, such a pouch formed from a polyethylene terephthalate resin, such as Mylar®, which

has been coated with a metal, such as aluminum, to reduce gas diffusion rates through the film.

Mylar is a registered trademark of DuPont Teijin Films U.S. Limited Partnership of Wilmington, Delaware. The preforms remained in this condition until they were removed in preparation for exposing the preforms to irradiation. Prior to irradiation, each of the resulting preforms was

5 preheated by heating the preforms in a Grieve convection oven, available from The Grieve Corporation of Round Lake, Illinois, to a preheat temperature, which was held for a minimum of 12 hours. The preheat temperature was selected from 60°C, 70°C, 85°C, 100°C, and 110°C, as set forth in TABLE 15 below.

[00115] The preforms were then exposed to a selected total irradiation dose according to
10 Method B, as set forth above in TABLE 8. The total irradiation dose was selected from 133 kGy, 150 kGy, 175 kGy, 200 kGy, and 217 kGy. Additionally, the total irradiation dose was divided and administered to the preforms in either two equal passes or three equal passes, which are combined to achieve the total irradiation dose. Specifically, the preforms indicated to be "Block 1" in TABLE 15 below received the total irradiation dose in two equal passes, while the preforms indicated to be
15 "Block 2" in TABLE 15 received the total irradiation dose in three equal passes.

[00116] After irradiation, each of the UHMWPE blends was separated from the substrate and three Pin-on-Disc (POD) specimens in the shape of cylinders having a 9 mm diameter and 13 mm thickness were then machined from the UHMWPE blend pucks. A bidirectional pin-on-disc wear tester was then used to measure the wear rate of UHMWPE pins articulating against polished
20 cobalt-chrome discs lubricated by 100% bovine serum. These measurements were made in accordance with the teachings of Bragdon, C.R., et al., in A new pin-on-disk wear testing method for simulating wear of polyethylene on cobalt-chrome alloy in total hip arthroplasty, published in the Journal of Arthroplasty, Vol. 16, Issue 5, 2001, on pages 658-65. The bidirectional motion for the pin-on-disc wear tester was generated by a computer controlled XY table, available from the
25 Compumotor Division of Parker Hannifin of Cleveland, Ohio, which was programmed to move in a 10 mm by 5 mm rectangular pattern. Affixed atop the XY table was a basin containing six cobalt-chrome discs polished to an implant quality finish. The XY table and basin were mounted on a servo-hydraulic MTS machine, available from MTS of Eden Prairie,

Minnesota. The MTS machine then loaded the UHMWPE blend pin specimens against the polished cobalt-chrome discs.

5 [00117] The MTS machine was programmed to produce a Paul-type curve [2] in synchronization with the motion of the XY table. A Paul-type curve is explained in detail in Forces Transmitted By Joints in the Human Body by J.P. Paul and published in the Proceedings Institution of Mechanical Engineers at Vol. 181, Part 37, pages 8-15. The peak load of the Paul- type loading curve corresponded to a peak contact pressure of 6.5 MPa between each of the UHMWPE pin specimen and the cobalt-chrome discs. Tests were conducted at 2 Hz to a total of 1.128×10^6 cycles.

10 [00118] The remaining portions of the UHMWPE blends were cut in half to form microtome films that were subjected to FTIR analysis utilizing a Bruker Optics FTIR Spectrometer, available from Bruker Optics of Billerica, Massachusetts. The films were then accelerated aged according to ASTM Standard F2003, Standard Guide for Accelerated Aging of Ultra-High Molecular Weight Polyethylene. The OI of the post-aged films was then measured.

15 [00119] Once the measurements were taken, the post-aged films were placed in boiling hexane for 24 hours to extract any d/l- α -tocopherol remaining in the films. The percentage of d/l- α -tocopherol extracted from the UHMWPE blend films was then determined. The remaining UHMWPE blend from the monoblock was then machined into 1/16" flats and Type V tensile specimens, as defined by ASTM Standard D638, Standard Test Method for Tensile Properties of
20 Plastics, were machined from the flats.

[00120] An analysis of the results, set forth below in TABLE 15, indicated that wear increased with a lower total irradiation dose or with a higher concentration of d/l- α -tocopherol. Additionally, the d/l- α -tocopherol concentration had a significant impact on ultimate tensile elongation. The yield strength was affected the most by the preheat temperature,
25 whereas UTS was affected the most by the total irradiation dose and d/l- α -tocopherol concentration. The OI was decreased with higher preheat temperatures and higher concentration of d/l- α -tocopherol. Although the percentage of d/l- α -tocopherol decreased after irradiation and aging, a significant amount of d/l- α -tocopherol still remained in the UHMWPE blend after irradiation and aging.

TABLE 15

Effect of Irradiation Dose on UHMWPE Blend

Run	Block	Preheat (°C)	Dose (kGy)	VE%	POD Wear (mg/Mc)
1	Block 1	100.00	200.00	0.20	1.01
2	Block 1	100.00	150.00	0.50	3.84
3	Block 1	100.00	150.00	0.20	1.59
4	Block 1	100.00	200.00	0.50	1.78
5	Block 2	59.77	175.00	0.35	1.97
6	Block 1	70.00	150.00	0.20	1.76
7	Block 1	70.00	200.00	0.20	0.80
8	Block 1	70.00	150.00	0.50	3.91
9	Block 1	70.00	200.00	0.50	2.38
10	Block 2	110.23	175.00	0.35	2.05
11	Block 2	85.00	132.96	0.35	3.32
12	Block 2	85.00	175.00	0.60	2.34
13	Block 2	85.00	175.00	0.10	0.58
14	Block 2	85.00	175.00	0.35	2.28
15	Block 2	85.00	217.04	0.35	1.06
16	Block 1	85.00	175.00	0.35	1.94
17	Block 2	85.00	175.00	0.35	2.30

Run	Elongation %	YS (MPa)	UTS (MPa)	VE% (Aged)	OI (Aged)
1	248.90	21.86	41.18	0.04	0.04
2	306.40	22.85	47.62	0.27	0.02
3	268.10	22.50	46.06	0.07	0.03
4	293.00	22.03	43.03	0.25	0.02
5	261.80	24.45	49.27	0.12	0.08
6	248.10	23.08	45.95	0.06	0.06
7	223.00	23.14	43.93	0.05	0.07
8	310.00	24.04	51.23	0.25	0.03
9	272.30	23.86	48.34	0.24	0.02
10	273.20	23.76	46.95	0.17	0.04
11	288.90	23.92	49.37	0.17	0.04
12	289.60	24.37	49.24	0.29	0.04
13	213.20	23.21	45.01	-0.01	0.06
14	258.80	23.97	47.60	0.17	0.05
15	234.00	24.41	45.00	0.13	0.06
16	269.70	23.39	48.64	0.14	0.02
17	264.10	23.95	48.41	0.15	0.05

Example 12

Swell Ratio, Crosslink Density, and Molecular Weight Between Crosslinks

[00121] GUR 1050 medical grade UHMWPE powder was obtained from Ticona, having North American headquarters in Florence, Kentucky. d/l- α -tocopherol was obtained from DSM Nutritional Products, Ltd of Geleen, Netherlands. The GUR 1050 was mechanically blended with the d/l- α -tocopherol using a High Intensity Mixer, available from Eirich Machines of Gurnee, Illinois. The GUR 1050 resin was mixed with the d/l- α -tocopherol to create UHMWPE blends having 0.2, 0.5, or 1.0 weight percent d/l- α -tocopherol. The UHMWPE blends were then compression molded to form pucks that were then machined to form cubes having 5 mm sides. The UHMWPE cubes were then heated to a preheat temperature selected from 40°C, 100°C, and 110°C. Once heated to the selected preheat temperature, the UHMWPE blends were irradiated using Method C, set forth in TABLE 8 above, until a total irradiation dose was received. The total irradiation dose was selected from of 90 kGy, 120 kGy, 150 kGy, and 200 kGy.

[00122] The resulting UHMWPE blend cubes were then studied to investigate the polymer network parameters of the UHMWPE blend by measuring the materials' swell ratio (q_s) with a Swell Ratio Tester (SRT), Cambridge Polymer Group (Boston, MA), in accordance with ASTM F-2214-02. Knowing q_s , the Flory interaction parameter (χ_1), the molar volume of the solvent (ϕ_1), and the specific volume of the solvent ($\bar{v}$), the crosslink density (ν_x) and the molecular weight between crosslinks (M_c) of the material were calculated according the following equations:

$$\nu_x = -\frac{\ln(1 - q_s^{-1}) + q_s^{-1} + \chi_1 q_s^{-2}}{\phi_1 (q_s^{-1/3} - q_s^{-1}/2)}$$

$$M_c = \bar{v} \nu_x$$

[00123] Additionally, the swell ratio in stabilized o-xylene at 130°C was measured in the compression molded direction. The results of the testing are set forth in TABLE 16 below. For example, it was found that a UHMWPE blend having nominally 1.0 % weight percent of d/l- α -tocopherol when preheated to nominally 40°C and subsequently electron beam crosslinked with a total dose of nominally 200 kGy has a q_s less than about 4.3, a ν_x more than about 0.090 and a M_c less than about 11,142. It was also found that a UHMWPE blend having nominally 1.0 %

weight percent of d/l- α -tocopherol when preheated to nominally 110°C and subsequently electron beam crosslinked with a total dose of nominally 200 kGy has a q_s less than about 3.6, a v_x more than about 0.117 and a M_c less than about 8,577.

[00124] Also, it was found that a UHMWPE blend having nominally 0.5 wt.% weight percent of d/l- α -tocopherol when preheated to nominally 40°C and subsequently electron beam crosslinked with a total dose of nominally 200 kGy has a q_s less than about 3.8, a v_x more than about 0.119 and a M_c less than about 8,421. It was also found that a UHMWPE blend having nominally 0.5 % weight percent of d/l- α -tocopherol when preheated to nominally 110°C and subsequently electron beam crosslinked with a total dose of nominally 200 kGy has a q_s less than about 3.6, a v_x more than about 0.109 and a M_c less than about 9,166.

[00125] Further, it was found that a UHMWPE blend having nominally 0.2 wt. % of d/l- α -tocopherol when preheated to nominally 40°C and subsequently electron beam crosslinked with a total dose of nominally 200 kGy has a q_s less than about 2.8, a v_x more than about 0.187 and a M_c less than about 5,351. It was also found that the UHMWPE blend having nominally 0.2 wt. % of d/l- α -tocopherol when preheated to nominally 110°C and subsequently electron beam crosslinked with a total dose of nominally 200 kGy has a q_s less than about 3.0, a v_x more than about 0.164 and a M_c less than about 6,097.

[00126] Additionally, it was found that under some conditions the crosslinked UHMWPE blend exhibited a crosslink density of less than 0.200 moles/dm³. Under other conditions, the crosslinked UHMWPE blend having at least 0.1 wt. % antioxidant exhibited a crosslink density of less than 0.190 moles/dm³. Further, under certain conditions, the crosslinked UHMWPE blend having at least 0.1 wt. % antioxidant exhibited a crosslink density of more than 0.200 moles/dm³ and had a molecular weight between crosslinks of less than 11,200 daltons.

TABLE 16

Swell Ratio, Crosslink Density, and Molecular Weight Between Crosslinks

RUN	MATERIAL TYPE	SAMPLE DOSE (kGy)	PREHEAT TEMP (°C)	PERCENT VITAMIN E	DOSE RATE (kGy-m/min)
1	GUR 1020	90	40	0.2	75.00
2	GUR 1050	90	100	0.2	75.00
3	GUR 1050	150	40	0.2	75.00
4	GUR 1020	150	100	0.2	75.00
5	GUR 1020	90	40	1.0	75.00
6	GUR 1020	90	100	0.5	75.00
7	GUR 1020	150	40	0.5	75.00
8	GUR 1020	150	100	1.0	75.00
9	GUR 1050	90	40	0.2	240.00
10	GUR 1020	90	100	0.2	240.00
11	GUR 1020	150	40	0.2	240.00
12	GUR 1050	150	100	0.2	240.00
13	GUR 1050	90	40	1.0	240.00
14	GUR 1020	90	100	1.0	240.00
15	GUR 1020	150	40	1.0	240.00
16	GUR 1020	150	100	0.5	240.00
17	GUR 1050	120	40	0.5	157.50
18	GUR 1050	120	100	1.0	157.50
19	GUR 1050	120	40	1.0	157.50
20	GUR 1050	120	100	1.0	157.50
21	GUR 1050	90	40	1.0	75.00
22	GUR 1050	90	100	0.5	75.00
23	GUR 1050	150	40	0.5	75.00
24	GUR 1050	150	100	1.0	75.00
25	GUR 1050	90	40	0.5	240.00
26	GUR 1050	90	100	1.0	240.00
27	GUR 1050	150	40	1.0	240.00
28	GUR 1050	150	100	0.5	240.00
29	GUR 1050	2x100 = 200	40	0.2	240.00
30	GUR 1050	2x100 = 200	110	0.2	240.00
31	GUR 1050	2x100 = 200	40	1.0	240.00
32	GUR 1050	2x100 = 200	110	1.0	240.00
33	GUR 1050	2x100 = 200	40	0.5	240.00
34	GUR 1050	2x100 = 200	110	0.5	240.00

RUN	POD WEAR mg/1M CYCLES	SWELL RATIO V/V0 = q(s)	X	V _x = XLD moles/dm ³	Mc = MWbXL Daltons
1		5.09	0.44	0.068	14747
2		3.40	0.49	0.129	7764
3	2.65	3.15	0.50	0.147	6812
4	1.13	4.61	0.45	0.079	12652
5		6.15	0.42	0.051	19720
6		5.52	0.43	0.060	16706
7		4.22	0.46	0.091	11019
8		5.52	0.43	0.060	16706
9		3.75	0.48	0.110	9126
10		4.49	0.45	0.082	12143
11		3.84	0.47	0.105	9483
12	0.17	3.23	0.50	0.141	7115
13		7.13	0.41	0.040	24747
14		4.47	0.45	0.083	12059
15		5.69	0.43	0.057	17502
16		4.46	0.45	0.083	12017
17		4.50	0.45	0.082	12185
18		3.74	0.48	0.110	9087
19		5.32	0.43	0.063	15785
20		3.33	0.50	0.133	7496
21		5.78	0.43	0.056	17929
22		4.43	0.45	0.084	11891
23	3.92	3.84	0.47	0.105	9483
24		3.88	0.47	0.104	9642
25		5.59	0.43	0.059	17033
26		4.52	0.45	0.082	12270
27		4.37	0.46	0.086	11640
28	1.63	3.65	0.48	0.115	8733
29	0.02	2.76	0.53	0.187	5351
30	0.09	2.96	0.52	0.164	6097
31	1.46	4.25	0.46	0.090	11142
32	0.64	3.61	0.48	0.117	8577
33		3.57	0.48	0.119	8421
34		3.76	0.48	0.109	9166

Example 13

Free Radical Concentrations in UHMWPE Blended with d/l- α -tocopherol

[00127] The impact of mechanically blending d/l- α -tocopherol with UHMWPE powder on free radical concentration of electron beam irradiated UHMWPE blend molded pucks was investigated. To perform this investigation, Design Expert 6.0.10 software, obtained from Stat-Ease, Inc. Minneapolis, MN, was utilized to setup a modified central composite Design of

Experiment (DOE). The DOE evaluated five factors: preheat temperature, dose rate, irradiation dose, d/l- α -tocopherol concentration, and predetermined hold time, i.e., the time elapsed between removal of the UHMWPE blend from the oven until the initiation of electron beam irradiation.

[00128] GUR 1050 medical grade UHMWPE powder was obtained from Ticona, having North American headquarters in Florence, Kentucky. d/l- α -tocopherol was obtained from DSM Nutritional Products, Ltd. of Geleen, Netherlands. The GUR 1050 UHMWPE powder was mechanically blended with the d/l- α -tocopherol by high intensity blending using an Eirich Mixer, available from Eirich Machines, Inc. of Gurnee, Illinois. The GUR 1050 resin was mixed with the d/l- α -tocopherol in several batches to create UHMWPE blends having between 0.14 and 0.24 wt. % d/l- α -tocopherol, as set forth below in TABLE 17.

[00129] Each of the UHMWPE blends were then compression molded into 2.5 inch diameter and 1 inch thick pucks. Each of the resulting pucks was then preheated by heating in a Grieve convection oven, available from The Grieve Corporation of Round Lake, Illinois, to a preheat temperature. The preheat temperature was selected from between 85°C and 115°C, as set forth in TABLE 17 below. The pucks were then removed from the convection oven and held for a predetermined period of time ranging between 7 minutes and 21 minutes, as set forth in TABLE 17. After the expiration of the predetermined hold time, the pucks were electron beam irradiated utilizing Method A of TABLE 8. The pucks were irradiated at a dose rate selected from between 30 kGy-m/min and 75 kGy-m/min until a total dose selected from between 160 kGy and 190 kGy was administered, as set forth in TABLE 17. Cylindrical cores approximately 1 inch long were machined from the pucks. The cylindrical cores were then analyzed using a Bruker EMX/EPR (electron paramagnetic resonance) spectrometer, which has a detection limit of 0.01×10^{15} spins/gram and is available from Bruker Optics of Billerica, Massachusetts. The resulting analysis indicated that preheat temperature, percent d/l- α -tocopherol, and dose level were all significant factors in determining the resulting free radical concentration of the UHMWPE blend. Specifically, preheat temperature and d/l- α -tocopherol concentration had a negative correlation with the free radical concentration, while the total dose had a positive correlation with the free radical concentration.

TABLE 17

Free Radical Concentration of UHMWPE Blends After Various Processing

Run	Block	Preheat (°C)	Dose (kGy)	VE%	Dose Rate (kGy- m/min.)	Oven to Beam (minutes)	Free radicals (spins/gram x E10-16)
1	Block 1	85	190	0.11	30	7	2.87
2	Block 1	115	190	0.11	30	7	1.09
3	Block 1	115	190	0.11	30	21	2.01
4	Block 1	85	190	0.11	30	21	3.7
5	Block 1	115	160	0.11	30	7	1.09
6	Block 1	85	160	0.11	30	7	2.55
7	Block 1	115	160	0.11	30	21	1.5
8	Block 1	85	160	0.11	30	21	2.77
9	Block 1	85	160	0.22	75	7	2.37
10	Block 1	115	160	0.22	75	7	0.826
11	Block 1	85	160	0.22	75	21	2.46
12	Block 1	115	160	0.22	75	21	1.38
13	Block 1	115	190	0.22	75	7	0.786
14	Block 1	85	190	0.22	75	7	3.22
15	Block 1	115	190	0.22	75	21	1.28
16	Block 1	85	190	0.22	75	21	2.94
17	Block 1	100	175	0.165	52.5	14	2.46
18	Block 1	100	175	0.165	52.5	14	2.66
19	Block 1	100	175	0.165	52.5	14	2.98
20	Block 1	100	175	0.165	52.5	14	3.03

Example 14Effect of UHMWPE Thickness on the Fatigue Behavior of Metal Backed Acetabular Cups

[00130] The effect of reducing the thickness of an antioxidant stabilized UHMWPE layer in a metal backed acetabular cup design was investigated.

[00131] In order to prepare the test specimens, three Natural Cup™ metal acetabular shells made in accordance with Trabecular Metal™ technology, commercially available from Zimmer, Inc., of Warsaw, Indiana, having a shell size of 54 mm and a head size of 40 mm, were obtained. The metal acetabular shells were compression molded with an antioxidant stabilized UHMWPE to form monoblock acetabular cups. The antioxidant stabilized UHMWPE of the acetabular cups was then crosslinked by exposing the antioxidant stabilized UHMWPE to crosslinking irradiation. The inner or articular surface formed by the antioxidant stabilized UHMWPE layer

was then machined for receipt of a 40 mm femoral head. All of the test specimens were then cleaned to remove particles from the machining operation.

[00132] The thickness of the antioxidant stabilized UHMWPE layer of an acetabular cup is dependent on the specific shell size used and the desired femoral head size for a particular acetabular cup configuration. Additionally, the thickness of the antioxidant stabilized UHMWPE layer is dependent upon the position at which the thickness of the UHMWPE layer is measured. In order to facilitate a comparison of the test results, several different thickness measurements were taken for each of the acetabular cups in the test. Specifically, for the purposes of this Example, the "Minimum Thickness" is defined as the thickness of the antioxidant stabilized UHMWPE layer measured from a point on the metal shells that is just below the titanium ring, which is formed at the equator of the Natural CupTM metal acetabular shells, to a line tangent to the radius of the articular surface formed by the antioxidant stabilized UHMWPE layer, as shown in Fig. 17 as "MINIMUM". A second measurement, referred to as the "Wall Thickness", is defined, for purposes of this Example, as the thickness of the antioxidant stabilized UHMWPE layer measured from the inner edge of the titanium ring of the Natural CupTM metal acetabular shells to the outermost point of the articular surface formed by the antioxidant stabilized UHMWPE layer, as shown in Fig. 17 as "WALL". In addition, the thickness of the antioxidant stabilized UHMWPE layer at the pole of the cup and at a point that is 45° from the pole of the cup were also measured, as shown in Fig. 17 as "POLE" and "45", respectively. Based on the results of the measurements, the antioxidant stabilized UHMWPE layer had a minimum thickness of 3.66 mm, a Wall Thickness of 1.33 mm, a thickness at the pole of the cup of 4.49 mm, and a thickness at a point that is 45° from the pole of the cup of 4.05 mm.

[00133] Once prepared, each of the three acetabular cups was then potted in a block of poly methyl meth acrylate bone cement in a position such that each of the cups had an inclination angle of 60°. The 60° inclination angle was selected to capture a clinically relevant, but clinically steep, abduction angle. The cup specimens were then placed in a de- ionized water bath maintained at a temperature of $37 \pm 1^\circ\text{C}$ and subjected to uniaxial loading on a MTS 858 Mini BionixTM uniaxial fatigue test machine, commercially available from MTS Systems Corporation of Eden Prairie, Minnesota. The temperature of $37 \pm 1^\circ\text{C}$ is representative of the *in vivo*

environment, which is appropriate due to the temperature-dependency of the mechanical properties of the UHMWPE.

[00134] The uniaxial loading was applied as a cyclic compressive force using a Cobalt-Chromium-Molybdenum femoral head component, with a minimum/maximum stress ratio of $R = 0.1$. The tests were performed at a frequency of 3 Hz, with the first cup subjected to a force of 1000 lbs, the second cup subjected to a force of 1350 lbs, and the third cup subjected to a force of 1700 lbs, until reaching a stopping point of 5 million cycles or fracture, whichever occurs first. Once the tests were completed, each of the three acetabular cups were visually inspected for damage to the antioxidant stabilized UHMWPE layer and the porous metal layer.

10 [00135] In contrast to traditional crosslinked UHMWPE acetabular cups made with a thicker UHMWPE layer, such as a UHMWPE layer of at least 6 mm, the acetabular cups of the present investigation all survived the testing, i.e., the stopping point of 5 million cycles was reached, and no deformation of the antioxidant stabilized UHMWPE layer or the porous metal layer was observed. In contrast, traditional crosslinked UHMWPE acetabular cups made with a fiber
15 metal shell and having a UHMWPE layer that is not antioxidant stabilized and a thickness of approximately 6 mm have a fatigue strength of approximately 1400 lbs., as measured in a substantially similar manner as indicated above. Thus, the use of an antioxidant stabilized UHMWPE layer allows for the acetabular cups of the present invention to have a UHMWPE layer that has a thickness that is substantially less than the UHMWPE layer thickness of conventional
20 acetabular cups, while also providing an increase in the fatigue strength of the same.

[00136] While this invention has been described as having a preferred design, the present invention can be further modified. This application is therefore intended to cover any variations, uses, or adaptations of the invention using its general principles. The scope of the claims should not be limited by the preferred embodiments set forth herein, but should be given the broadest
25 interpretation consistent with the description as a whole.

CLAIMS,

1. An orthopedic system comprising:
 - a substantially hemispherical outer layer formed from a first porous material; and
 - a substantially hemispherical inner layer formed from an antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene, wherein said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene at least partially permeates the pores of said first porous material so that at least a portion of said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene interdigitates with pores of said first porous material, said first porous material and said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene cooperating to define an acetabular shell having an effective stiffness between 0.1 GPa and 15 GPa, and wherein said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene is prepared by a process that includes
 - combining ultrahigh molecular weight polyethylene with an antioxidant to form a blend;
 - processing the blend to consolidate the blend, wherein the consolidated blend had a melting point;
 - preheating the consolidated blend to a preheat temperature below the melting point of the consolidated blend; and
 - irradiating the preheated consolidated blend while maintaining the consolidated blend at a temperature below the melting point of the consolidated blend.
2. The orthopedic system of Claim 1, wherein said acetabular shell has an effective stiffness between 0.1 GPa and 10 GPa.
3. The orthopedic system of Claim 1 or 2, wherein said acetabular shell has an effective stiffness between 0.3 GPa and 1.5 GPa.
4. The orthopedic system of any one of Claims 1 to 3, wherein said first porous material has an elastic modulus between 0.1 GPa and 15 GPa and said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene has an elastic modulus between 0.1 GPa and 15 GPa.

5. The orthopedic system of any one of Claims 1 to 4, wherein said first porous material is a metal.
6. The orthopedic system of Claim 5, wherein said first porous material is a reticulated vitreous carbon foam substrate coated with a biocompatible metal by chemical vapor deposition.
7. The orthopedic system of any one of Claims 1 to 6, wherein said antioxidant is uniformly distributed throughout the antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene.
8. The orthopedic system of any one of Claims 1 to 7, further comprising a substantially hemispherical liner formed from a liner material, said liner configured for receipt within said acetabular shell.
9. The orthopedic system of Claim 8, wherein said liner material comprises ultrahigh molecular weight polyethylene.
10. The orthopedic system of Claim 1, wherein said substantially hemispherical inner layer further comprises an integral locking feature.
11. The orthopedic system of Claim 1, further comprising a substantially hemispherical liner configured for receipt within said acetabular shell, said substantially hemispherical liner having an integral locking feature, said integral locking feature of said substantially hemispherical liner configured for locking engagement with an integral locking feature of said substantially hemispherical inner layer, whereby translation of said substantially hemispherical liner relative to said acetabular shell is substantially prevented.
12. The orthopedic system of any one of Claims 1 to 11, wherein said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene is injection molded into said first porous material.

13. An acetabular component configured for use in a hip replacement surgery, the acetabular component comprising:
 - a porous layer configured to contact and interface with bone tissue when the acetabular component is implanted;
 - an inner layer formed from an antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene and having a thickness of less than six millimeters, said inner layer configured to receive a femoral component; and
 - an interdigitation layer defined by the distance over which said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene of said inner layer infiltrates pores of said porous layer; and wherein said antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene is prepared by a process that includes
 - combining ultrahigh molecular weight polyethylene with an antioxidant to form a blend;
 - processing the blend to consolidate the blend, wherein the consolidated blend had a melting point;
 - preheating the consolidated blend to a preheat temperature below the melting point of the consolidated blend; and
 - irradiating the preheated consolidated blend while maintaining the consolidated blend at a temperature below the melting point of the consolidated blend.
14. The acetabular component of Claim 13, wherein said inner layer has a thickness substantially equal to two millimeters.
15. The acetabular component of Claim 13 or 14, wherein said porous layer has a porosity of at least 55 percent.
16. The acetabular component of any one of Claims 13 to 15, wherein said porous layer is defined by a porous metal having an elastic modulus of less than 15 GPa.
17. The acetabular component of any one of Claims 13 to 16, having an effective stiffness between 0.1 GPa and 10 GPa.
18. The acetabular component of Claim 17, having an effective stiffness between 0.3 GPa and 1.5 GPa.
19. A method of manufacturing an orthopedic component for implantation into a bone, the orthopedic implant having a bone contacting layer, an interdigitation layer, and an inner layer, the method comprising the steps of:

determining the elastic modulus of the bone;

selecting a thickness of at least one of the bone contacting layer, an interdigitation layer, and an inner layer based on the elastic modulus of the bone; and

molding the inner layer to the bone contacting layer to form at least one of the bone contacting layer, an interdigitation layer, and an inner layer to have the selected thickness; wherein said inner layer has an antioxidant stabilized crosslinked ultrahigh molecular weight polyethylene prepared by a process that comprises:

combining ultrahigh molecular weight polyethylene with an antioxidant to form a blend;

processing the blend to consolidate the blend, wherein the consolidated blend had a melting point;

preheating the consolidated blend to a preheat temperature below the melting point of the consolidated blend; and

irradiating the preheated consolidated blend while maintaining the consolidated blend at a temperature below the melting point of the consolidated blend.

20. The method of Claim 19, further comprising, after the step of selecting a thickness, the step of selecting a thickness of another of at least one of the bone contacting layer, an interdigitation layer, and an inner layer based on the elastic modulus of the bone.

21. The method of Claim 19 or 20, wherein the bone contacting layer comprises a porous metal.

22. The method of any one of Claims 19 to 21, wherein the antioxidant is uniformly distributed throughout the consolidated blend.




FIG. 1




FIG. 2A




FIG. 2B




FIG. 3

Figure 4





Figure 5
Comparison Material Properties
Conventional UHMWPE vs. Vitamin E treated UHMEPE






Fig. 6




Fig. 7




Fig. 8




Fig. 9




Fig. 10




Fig. 11




Fig. 12




Fig. 13




Fig. 14




Fig. 15




Fig. 16




Fig. 17