



(86) Date de dépôt PCT/PCT Filing Date: 2013/09/27
(87) Date publication PCT/PCT Publication Date: 2014/04/03
(85) Entrée phase nationale/National Entry: 2015/03/18
(86) N° demande PCT/PCT Application No.: US 2013/062151
(87) N° publication PCT/PCT Publication No.: 2014/052741
(30) Priorité/Priority: 2012/09/28 (US61/707,070)

(51) Cl.Int./Int.Cl. *A61B 5/08* (2006.01)
(71) Demandeur/Applicant:
ARIZONA BOARD OF REGENTS ON BEHALF OF
ARIZONA STATE UNIVERSITY, US
(72) Inventeurs/Inventors:
FORZANI, ERICA, US;
TAO, NONGJIAN, US;
XIAN, XIAOJUN, US;
TSOW, FRANCIS, US
(74) Agent: SMART & BIGGAR

(54) Titre : **PIECE BUCCALE POUR LA DETECTION PRECISE DU NO EXHALE**
(54) Title: **MOUTHPIECE FOR ACCURATE DETECTION OF EXHALED NO**





FIG. 2

(57) **Abrégé/Abstract:**

A low back pressure mouthpiece for accurate detection of exhaled nitric oxide (NO) includes a conduit for receiving the exhaled breath from the subject. An oxidizing filter for sample conditioning, wherein the conduit and oxidizing filter operate to produce a back pressure of less than 4 cm H₂O; and a device for measuring the level of one or more components of the received exhaled breath.



(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property
Organization
International Bureau(43) International Publication Date
3 April 2014 (03.04.2014)(10) International Publication Number
WO 2014/052741 A1(51) International Patent Classification:
A61B 5/08 (2006.01)Laredo Ave., Gilbert, AZ 85296 (US). TSOW, Francis;
1941 E. Cairo Dr., Tempe, AZ 85282 (US).(21) International Application Number:
PCT/US2013/062151(74) Agent: LEONE, George; Citadel Patent Law, 9125
Bridgeport Way SW, Suite 105, Lakewood, Washington
98499 (US).(22) International Filing Date:
27 September 2013 (27.09.2013)


(25) Filing Language: English

(26) Publication Language: English

(30) Priority Data:
61/707,070 28 September 2012 (28.09.2012) US(71) Applicant: ARIZONA BOARD OF REGENTS, A
BODY CORPORATE OF THE STATE OF ARIZONA,
ACTING FOR AND ON BEHALF OF ARIZONA
STATE UNIVERSITY [US/US]; 1475 N. Scottsdale
Road, Suite 200, Scottsdale, AZ 85257-3538 (US).(81) Designated States (unless otherwise indicated, for every
kind of national protection available): AE, AG, AL, AM,
AO, AT, AU, AZ, BA, BB, BG, BH, BN, BR, BW, BY,
BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DK, DM,
DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT,
HN, HR, HU, ID, IL, IN, IS, JP, KE, KG, KN, KP, KR,
KZ, LA, LC, LK, LR, LS, LT, LU, LY, MA, MD, ME,
MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ,
OM, PA, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SA,
SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM,
TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM,
ZW.(72) Inventors: FORZANI, Erica; 1856 W. 6th Street, Mesa,
AZ 85201 (US). TAO, Nongjian; 13571 E. Estrella Aven-
ue, Scottsdale, AZ 85287 (US). XIAN, Xiaojun; 254 E.(84) Designated States (unless otherwise indicated, for every
kind of regional protection available): ARIPO (BW, GH,
GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, SZ, TZ,

[Continued on next page]

(54) Title: MOUTHPIECE FOR ACCURATE DETECTION OF EXHALED NO

(57) Abstract: A low back pressure mouthpiece for accurate
detection of exhaled nitric oxide (NO) includes a conduit for
receiving the exhaled breath from the subject. An oxidizing
filter for sample conditioning, wherein the conduit and oxid-
izing filter operate to produce a back pressure of less than 4
cm H₂O; and a device for measuring the level of one or more
components of the received exhaled breath.**FIG. 2**

WO 2014/052741 A1

UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, KM, ML, MR, NE, SN, TD, TG).

Published:
— *with international search report (Art. 21(3))*
— *before the expiration of the time limit for amending the claims and to be republished in the event of receipt of amendments (Rule 48.2(h))*

Declarations under Rule 4.17:

— *as to applicant's entitlement to apply for and be granted a patent (Rule 4.17(ii))*

MOUTHPIECE FOR ACCURATE DETECTION OF EXHALED NORELATED APPLICATION

5 This application claims priority from U.S. provisional application number 61/707,070 of Forzani et al., filed 09/28/2012, entitled "MOUTHPIECE FOR ACCURATE DETECTION OF EXHALED NO." U.S. application number 61/707,070, is hereby incorporated by reference.

TECHNICAL FIELD

10 The present invention relates to a mouthpiece for accurate detection of exhaled nitric oxide (NO), this invention relates to the measurement of components of exhaled breath from the lower respiratory tract.

BACKGROUND

15 Exhaled pulmonary nitric oxide (NO) may aid in monitoring pulmonary disease. Unfortunately, it has been recognized that, in measuring exhaled pulmonary nitric oxide (NO), there are obstacles that must be overcome. For example, nasal NO concentration can be higher than alveolar NO concentration, and, as a result, contamination with Nasal NO may occur.

20 One attempt to provide a solution to this problem was as disclosed by Silkoff et al. in a paper entitled "Marked flow-dependence of exhaled nitric oxide using a new technique to exclude nasal nitric oxide," (AMERICAN JOURNAL OF RESPIRATORY AND CRITICAL CARE MEDICINE, Volume: 155, Issue: 1, Pages: 260-267, Published: JAN 1997). There a technique was developed to measure pulmonary NO, without nasal NO, by having the subject maintain a positive
25 expiratory pressure (ensuring velum closure) in an attempt to prevent contamination by nasal NO.

 Unfortunately available techniques using exhalation against back pressure of 5 cm H₂O or larger can be difficult for people with limited lung expiratory force. Such subjects exhibit an inability to maintain constant exhalation flow for several seconds
30 (e.g., one commercially available device requires between 6 to 10 sec). Further, current commercial devices require pressure of 10-20 cm H₂O to perform the measurement, which makes it difficult to get the measurement done, especially in children.

In contrast to known methods, now presented is a new and novel low back pressure mouthpiece for measuring NO that overcomes difficulties in this area not adequately addressed until now.

5 BRIEF SUMMARY OF THE DISCLOSURE

This summary is provided to introduce a selection of concepts in a simplified form that are further described below in the Detailed Description. This summary is not intended to identify key features of the claimed subject matter, nor is it intended to be used as an aid in determining the scope of the claimed subject matter.

10 A low back pressure mouthpiece for accurate detection of exhaled nitric oxide (NO) comprising:

conduit means for receiving the exhaled breath from the subject;

an oxidizing filter means coupled to the conduit means for sample conditioning, wherein the oxidizing filter means has an outlet and wherein the
15 conduit means and oxidizing filter means operate to produce a back pressure of less than 4 cm H₂O; and

means for measuring the level of one or more components of exhaled breath received from the outlet.

20 BRIEF DESCRIPTION OF THE DRAWINGS

While the novel features of the invention are set forth with particularity in the appended claims, the invention, both as to organization and content, will be better understood and appreciated, along with other objects and features thereof, from the following detailed description taken in conjunction with the drawings, in which:

25 FIG. 1A shows a graphical illustration of exhaled NO measurements.

FIG. 1B shows a plot of NO concentration and airway opening vs. time as reported in a joint statement of the American Thoracic Society (ATS) and the European Respiratory Society (ERS).

FIG. 2 schematically shows an example of a configuration and picture of a
30 low back pressure mouthpiece.

FIG. 3 shows a schematic representation of a test configuration for measuring back pressure across a mouthpiece.

FIG. 4 graphically illustrates back pressure data of a mouthpiece.

FIG. 5 schematically shows an example of an experimental configuration of a correlation test on an NO device.

FIG. 6 schematically shows an example of electronics employed in an experimental configuration of a correlation test on an NO device.

5 FIG. 7 graphically shows an example of a correlation plot between an NO device and gold standard method.

FIG. 8 shows a typical plot of sensor response for one cycle of test with purging and sampling periods.

10 In the drawings, identical reference numbers identify similar elements or components. The sizes and relative positions of elements in the drawings are not necessarily drawn to scale. For example, the shapes of various elements and angles are not drawn to scale, and some of these elements are arbitrarily enlarged and positioned to improve drawing legibility. Further, the particular shapes of the elements as drawn, are not intended to convey any information regarding the actual
15 shape of the particular elements, and have been solely selected for ease of recognition in the drawings.

DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENTS

The following disclosure describes several embodiments for a low back pressure mouthpiece for NO measurement. Several features of methods and
20 systems in accordance with example embodiments are set forth and described in the Figures. It will be appreciated that methods and systems in accordance with other example embodiments can include additional procedures or features different than those shown in the Figures. Example embodiments are described herein with respect to analysis of pulmonary NO. However, it will be understood that these
25 examples are for the purpose of illustrating the principles, and that the invention is not so limited. Additionally, methods and systems in accordance with several example embodiments may not include all of the features shown in the Figures.

Unless the context requires otherwise, throughout the specification and claims which follow, the word "comprise" and variations thereof, such as,
30 "comprises" and "comprising" are to be construed in an open, inclusive sense that is as "including, but not limited to."

Reference throughout this specification to "one example" or "an example embodiment," "one embodiment," "an embodiment" or combinations and/or

variations of these terms means that a particular feature, structure or characteristic described in connection with the embodiment is included in at least one embodiment of the present disclosure. Thus, the appearances of the phrases "in one embodiment" or "in an embodiment" in various places throughout this specification
5 are not necessarily all referring to the same embodiment. Furthermore, the particular features, structures, or characteristics may be combined in any suitable manner in one or more embodiments.

Definitions

Generally, as used herein, the following terms have the following meanings
10 when used within the context of sample collection or analysis:

As used herein, "plurality" is understood to mean more than one. For example, a plurality refers to at least 3, 4, 5, 70, or more.

As used herein, "cellular telephone" (or "smart phone") has its generally accepted meaning and includes any portable device that can make and receive
15 telephone calls to and from a public telephone network, which includes other mobiles and fixed-line phones across the world. It also includes mobile devices that support a wide variety of other services such as text messaging, software applications, MMS, e-mail, Internet access, short-range wireless communications (for example, infrared and Bluetooth).

As used herein, "tablet computer" has its generally accepted meaning and includes any mobile computer including a complete mobile computer, larger than a mobile phone or personal digital assistant, integrated into a flat touch screen and primarily operated by touching the screen such as, for example, an Apple iPad®
20 tablet computer.

Example Embodiments

The inventors here have noted their own experience with NO, and other's experience, including ATS/ERS indicate the nasal contamination (if present can be washed out). Referring now to FIG. 1A, a graphical illustration of exhaled NO measurements published by Kharitonov, "Exhaled and Nasal Nitric Oxide
30 Measurements: Recommendations (Eur Respir J 1997, Vol. 10, pp. 1683-1693) illustrate the plateau of exhaled NO. According to this result, a back pressure of ~3 cm H₂O (=2.3 mmHg) is sufficient to produce a stable NO plateau at the end of a breath.

Referring briefly to FIG. 1B, there shown is a plot of NO concentration and airway opening vs. time as reported in a joint statement of the American Thoracic Society (ATS) and the European Respiratory Society (ERS). See “ATS/ERS Recommendations for Standardized Procedures for the Online and Offline
5 Measurement of Exhaled Lower Respiratory Nitric Oxide and Nasal Nitric Oxide, 2005,” (Am J Respir Crit Care Med, Vol.171, pp. 912-930, 2005). Note that in the exhaled NO pressure profile the NO plateau is essentially unaltered once the early peak has washed out. The inventor’s here exploited their experience and the noted data to arrive at a new configuration of a low back pressure mouthpiece (herein
10 referred to also as the “subject NO device” for measurement of exhaled NO).

Referring now to FIG. 2, an example of a low back pressure mouthpiece is schematically shown. A low back pressure mouthpiece apparatus 1 includes a breath inlet conduit 5, a coupler 10, an oxidizing filter housing 12, a first plurality of filters 14A, a second plurality of filters 14B and an outlet tube 20. Packed between
15 the first and second plurality of filters 14A, 14B are at least two types of filtering particles including a first type of filtering particles 15 and a second type of filtering particles 17. In one useful embodiment the filter housing 12, first and second plurality of filters 14A, 14B and filtering particles 15, 17 perform as an oxidizing filter.

In one useful embodiment, the inner diameter of the filter housing 12 is at
20 least 18 mm and the chemical particles are contained in the filter housing 12. In one useful embodiment the coupler 10 may be fabricated from an acrylic tube with inner diameter of at least 9.6 mm and outer diameter of at least 12.6 mm. The coupler 10 is used to guide the gas flow. In one embodiment, the first and second plurality of filters may advantageously comprise two felt pieces made from stiffened felt
25 comprising about 100% Eco-fi, a high quality polyester fiber, with diameter of 18 mm. The filters are used as stoppers to retain the chemical particles within the mouthpiece filter housing. In a preferred embodiment the elements operate in a low back pressure range of less than 4 cm H₂O and more preferably in a range from 1 to 3 cm H₂O.

30 In testing, as described in detail below, it has been shown that the mouthpiece has a capability of conditioning the breath at a flow of 50 ml/sec with an associated error of $\pm 10\%$ under sample collection condition with back pressure less than 4 cm H₂O. In one example embodiment, the breath inlet conduit 5 comprises a

commercial mouthpiece element for sample collection. The oxidizing filter elements operate to provide sample conditioning. In one example embodiment, the oxidizing filter housing 12 was made from acrylic tubing with chemical particles packed inside. Two kinds of chemical particles were used in the oxidizing filter: 1) desiccant
 5 including 300 mg CaSO_4 (indicating drierite, stock number 23001, from W.A. HAMMOND DRIERITE CO., LTD.) for reducing the humidity in the breath; and 2) oxidant containing 300 mg of a solid porous substrate impregnated with sodium permanganate (available under the trade name Purafil®) to provide optimum gas oxidation. The Purafil® media works under a wide range of humidity levels (e.g.
 10 from 10% to 95% RH).

Further examples of useful desiccants include activated alumina, aerogel, benzophenone, bentonite clay, calcium chloride, calcium sulfate, cobalt(ii) chloride, copper(ii) sulfate, lithium chloride, lithium bromide, magnesium sulfate, magnesium perchlorate, molecular sieve, potassium carbonate, silica gel, sodium, sodium
 15 chlorate, sodium chloride, sodium hydroxide, sodium sulfate, sucrose and the like. Further examples of useful oxidizing agents include oxygen (O_2), ozone (O_3), hydrogen peroxide (H_2O_2) and other inorganic peroxides, fluorine (F_2), chlorine (Cl_2), and other halogens, nitric acid (HNO_3) and nitrate compounds, sulfuric acid (H_2SO_4), peroxydisulfuric acid ($\text{H}_2\text{S}_2\text{O}_8$), peroxymonosulfuric acid (H_2SO_5), chlorite, chlorate,
 20 perchlorate, and other analogous halogen compounds, hypochlorite and other hypohalite compounds, including household bleach (NaClO), hexavalent chromium compounds such as chromic and dichromic acids and chromium trioxide, pyridinium chlorochromate (PCC), and chromate/dichromate compounds, permanganate compounds such as potassium permanganate, sodium perborate, nitrous oxide
 25 (N_2O), silver oxide (Ag_2O), osmium tetroxide (OsO_4) and the like.

FIG. 3 shows a schematic representation of a test configuration for measuring back pressure across a mouthpiece. The test configuration includes a mouthpiece 1, a source of clean air 30, a valve 32, a flow meter 34 and a pressure sensor 38.

30 In one exemplary process, the back pressure across the mouthpiece 1 was measured by following the procedure below.

- 1) Connecting the source of clean air 30, such as a gas cylinder, and the flow meter in series to the inlet 7 of the mouthpiece 1;
- 2) Connecting the pressure sensor 38 across the mouthpiece 1 by drilling two holes in two acrylic tubes of same diameter in the mouthpiece at both inlet and outlet. The two probes of the pressure sensor are connected across the mouthpiece by using hard tubing;
- 3) Turning on the valve on the clean air gas cylinder and adjust the flow rate to be 50 ml/sec.; and
- 4) Obtaining the pressure drop readings from the pressure sensor.

Referring now to FIG. 4, typical back pressure data of a mouthpiece made in accordance with the teachings herein is graphically illustrated. In tests conducted by the inventors three individual mouthpieces were prepared and tested to get the back pressure data. In this example, the back pressure of one embodiment of the mouthpiece in cm H₂O was plotted against the index of the mouthpiece and the mean and standard deviation was calculated. The results of the test were evaluated as acceptable if the back pressure of the mouthpiece is less than 4 cm H₂O at a flow rate of 50 ml/sec. As shown for these tests the mean was 1.027 cm H₂O with a standard deviation of 0.0006 cm H₂O.

Referring now to FIG. 5, an example of an experimental configuration of a correlation test on an NO device built in accordance with the principles disclosed herein is schematically shown. For the purposes of testing correlation an integrated NO device was constructed including a low back pressure mouthpiece 1, a valve 50, a zeroing filter 52, a pump 56, Nafion tubing 58, a sensor chip 62 within a sensor chamber 60, electronics on a printed circuit board (PCB) 500. Except for the new and novel low back pressure mouthpiece 1, the components may be standard components connected according to accepted engineering practice.

Referring now to FIG. 6, an example of electronics employed in an experimental configuration of a correlation test on an NO device is schematically shown. The electronics 500 include a microcontroller 502, a Bluetooth transceiver 504, a plurality of drivers 506, a feedback LED array 508, a power supply 512, and a switch 511. The microcontroller and drivers operate to execute a software application in obtaining and storing data and communicating to a user. A smart

phone 510 is employed in communication with the PCB 500 for data receiving, processing, and displaying. In one test an HTC HD2 Unlocked Phone with Windows Mobile 6.5 Professional was employed. The feedback LED array 508 comprised white, red, yellow and green LEDs in one example embodiment.

5 A software application written using standard computer science principles was installed in the smart phone to communicate with the Bluetooth of the NO device and display the test results. The application was written using Microsoft Visual Studio. In operation, the application asks the user to select the Bluetooth device from the list of devices that are visible to the phone. When the user selects
10 the sensor device, the connection is established. The device sends out raw data for the reference channel and the sensing channel. The absorbance value is calculated by taking the negative of the logarithm value of the ratio of intensity of sensing to reference channel. Then the difference between the slope of sampling and the slope of purging is calculated. This difference value is the quantity that is related to the
15 concentration of NO.

For correlation purposes chemiluminescence equipment used included a Nitric Oxide Analyzer (GE Analytical Instruments) Part number: NOA 280i. In one embodiment of the test setup the power supply comprised a commercially available battery charger, namely a TLP-2000 Tenergy Universal Smart Charger, from
20 Tenergy Corp of Fremont, California, Part No. 01211.

A correlation test was performed to compare the accuracy between nitric oxide levels detected with subject sensor and with chemiluminescence method (made by Sievers and sold by GE Analytical, Boulder, CO - gold standard method recognized by FDA) and an existing commercial device. The results of the test are
25 evaluated were considered acceptable if the correlation was larger than 90%.

Correlation of accuracy between the subject NO device and the gold standard method as well as the existing commercial device was completed by testing the NO level of real subjects. A new sensor chip and new mouthpiece is used for each test. In one test, nine different individuals were tested. Each subject may have been
30 tested multiple times on different days or even at different times during the same day. For some subjects, their NO concentration may have been subject to change in a very broad range, for example from 30 ppb to 200 ppb, depending on the

inflammation condition of their respiratory system. These subjects with more desirable NO concentration levels can be tested more.

Not all tests were completed on the same day and some over more than a month. One correlation test contains 65 data points. Each test consumed one mouthpiece and one sensor chip, so a total of 65 mouthpieces and 65 sensor chips were used for the correlation test. The device was scheduled for testing at ambient conditions, i.e. at room temperature, between 16°C and 30°C, and a relative humidity (RH) between 20% and 60% (non-condensing).

Testing Procedure:

- 10 The following steps were performed for each test:
 - 1) The batteries of the NO device were charged until they fully charged (tests were carried out without any external power supply).
 - 2) The as prepared sensor chip was inserted into the sensor chamber of the NO device.
 - 15 3) The switch was turned on and the device was warmed up for 20 minutes before the test.
 - 4) During the warm up, test the NO level of the subject with commercial device and NO analyzer (the Chemiluminescence equipment) by following the corresponding instructions. One test may be performed on each device respectively and the results are used for correlation comparison since they are well established NO testing methods.
 - 20 5) During the warm up, the mouthpiece was connected to the device via the mouthpiece adapter.
 - 6) During the warm up, the ambient air was continuously sucked into the device through the zeroing filter for purging and the white LED was always on to indicate the warm-up was ongoing.
 - 25 7) During the warm up, the software was run on the smart phone to communicate with the NO device. During the warm up, the smart phone may display "measuring" on the screen.

- 8) Once the warm-up was finished, the white LED was automatically off and yellow LED turned on.
- 9) The subject being tested then placed their mouth on the mouthpiece and blew. During the test, once a subject began supplying a sample within the correct flow rate window the green LED turned on. When the green LED was on the sampling time lasted for 6 seconds, during which time the subject must keep their breath at the correct flow rate for at least 6 seconds.
- 10) Note that if the subject was unable to make the flow rate in range, a red or yellow LED was turned on to give the subject feedback. If the red LED turned on, it indicated that the flow rate was too high and the subject must reduce their breath flow rate. Conversely, if the yellow LED turned on, it indicated that the flow rate was too low and the subject must increase their breath flow rate. A 10 second time window was set for the subject to adjust their breath flow rate. If the subject could not keep the flow rate in range for 6 seconds within this 10 second time window, the device would go back to baseline and purge the system for 60 seconds. At this time all the three LED indicators (red, green and yellow) would be turned off. Once the new baseline was built, the yellow LED would be turned on again and the subject can try to blow and do the test again.
- 11) Should the subject complete a test by holding the correct flow rate (maintaining the green LED lit) for 6 seconds continuously, the sampling period ended. All the LEDs (include white, red, green, and yellow) would automatically be turned on, which indicated that the NO device was sending data to the smart phone. The smart phone will continue to display "measuring" during this time. The subject can stop breath when all the LEDs are turn on.
- 12) When the data transmission was completed the smart phone screen would change to display concentration and temperature. This concentration value was calculated by using the given calibration curve in the smart phone application. In this case, since the given calibration curve was based on the artificial sample tests, which may be diverse a little bit

from the real breath test, the displayed concentration for the correlation comparison was not used. Instead, the real sensor response was calculated according the procedure mentioned in the data analysis section. The temperature value displayed was not a real breath temperature (it was a fixed value).

5

13) To complete the second test on the same sensor cartridge the device was not turned off and the sensor chip was left in the same position. The warm-up did not need to take place again as the device was left on. The mouthpiece was left in position as well. The smart phone application would need to be restarted by quitting the application from the smart phone then re-opening the application. When again connecting to the device, the smart phone screen will again display "measuring" once connected. Wait until the test was done.

10

14) Step 13 was repeated for a third test on the same sensor chip.

15

15) A predetermined "r" value correlation coefficient check function was integrated into the software program so that the smart phone application would automatically check the "r" value (correlation coefficient) of the data obtained to evaluate the quality of the test. If the "r" value was lower than a predetermined threshold, the application will display "test fail" on the screen, which means this test should not be considered as an acceptable measurement and a new test need to be carried out.

20

16) The device was turned off and a new mouthpiece was prepared while a new sensor chip was inserted into the chamber of the NO device. The test steps would be repeated from steps 3 to 14 with a new test subject to provide a different concentration level. If necessary the batteries of the device were charged. A fully charged device should be able to test three different sensor chips.

25

Data analysis

The raw data of each test was transmitted and automatically saved in the smart phone. In order to get a more accurate concentration, the concentration displayed on the phone was not used as the final result because it was based on the

30

calibration curve obtained from artificial sample tests. The procedure below was followed for the data analysis:

- 1) The raw data with txt format was copied from the smart phone and saved in a personal computer for data processing.
- 5 2) A txt file was opened with Origin (a common scientific data analysis software available from OriginLab of Northampton, MA), the unit of time was changed from HH:MM:SS to seconds, and the data plotted.
- 3) FIG. 8 shows a typical plot of the sensor response for one cycle of test with purging and sampling periods. Slopes from the signal as a function of
10 time are assessed for sampling and purging periods. Linear fitting was done for the purging period, which lasts about 60 seconds and the sampling period, which lasts for 6 seconds.
- 4) A sensor response was calculated as: $\text{Sensor response} = \text{Slope}_{\text{sampling}} - \text{Slope}_{\text{purging}}$. The value of the sensor response was proportional to the NO
15 concentration.
- 5) For each test the subject was been tested three times for a total of three readings. The mean of these three readings was calculated.
- 6) The mean of each test was plotted along with the NO concentrations from the NO analyzer (i.e. the chemiluminescence equipment, gold standard)
20 and linear fitting was applied. Then the linear fitting was used as an internal calibration curve.
- 7) Using the internal calibration curve obtained from step 6 the original NO sensor response (A.U./S) was converted to concentration (ppb).
- 8) A correlation plot for the sensor response (ppb) from the subject low back
25 pressure mouthpiece NO device was made comparing the corresponding response from the commercially available device. Another plot comparing the subject device with the gold standard (the chemiluminescence equipment) was also evaluated.
- 9) Linear fitting was done to these plots. Then the “r” values were obtained
30 from the linear fitting. An “r” value greater than 0.9, indicated that the

subject NO device shows correlation of better than 90% with the commercial device.

10) Based on the linear fitting parameters, the residuals and the standard error of estimate can be calculated according to the following equations.

5 Predicted Values: $Y_i = A + B \cdot X_i$

Where X_i was the concentration from the comparative method;
 Y_i was the predicted value according to the regression curve.

The residual was calculated by: $Residual_i = y_i - Y_i$

10 Where y_i was the corresponding concentration from the NO device.

The residuals were plotted against the corresponding concentration from the comparative method. And the standard error of the residuals was calculated for different range of NO concentrations: <50ppb, 50-100ppb, >100ppb.

15 Referring now to FIG. 7 an example of a correlation plot between a subject NO device and the chemiluminescence equipment method is shown. Since the chemiluminescence technique is generally regarded as the gold standard of NO detection, this technique was used to measure the real concentration of NO sample. The graphical representation includes an ordinate representing NO concentration as read from a mouthpiece under test in ppb compared to an abscissa representing NO
 20 concentration from the gold standard in ppb. The data points (X, Y) represent actual correlation test values from a correlation of an NO device comprising a low back pressure mouthpiece constructed in accordance with the principles disclosed herein with a "gold standard" (GS) measurement of NO concentration. The curve 100 is a linear fitting of the data showing a residual value R of 0.94062.

25 Briefly, 9 different subjects with the exhaled nitric oxide (eNO) concentration in the range of 10-210 ppb were tested by the gold standard method, another commercial device, and the presently disclosed low back pressure mouthpiece to evaluate the correlations. The low back pressure mouthpiece was tested at ambient conditions, i.e. at room temperature, between 16°C and 30°C, and a relative
 30 humidity (RH) between 20% and 60% (non-condensing).

The invention has been described herein in considerable detail in order to comply with the Patent Statutes and to provide those skilled in the art with the information needed to apply the novel principles of the present invention, and to construct and use such exemplary and specialized components as are required.

5 However, it is to be understood that the invention may be carried out by different equipment, and devices, and that various modifications, both as to the equipment details and operating procedures, may be accomplished without departing from the true spirit and scope of the present invention.

10 References

1. Kharitonov, S. et. al., "Exhaled and nasal nitric oxide measurements: recommendations," EUROPEAN RESPIRATORY JOURNAL, Volume: 10, Pages 1683 – 1693 Published: 1997.
- 15 2. Silkoff, PE et. al., "Marked flow-dependence of exhaled nitric oxide using a new technique to exclude nasal nitric oxide" AMERICAN JOURNAL OF RESPIRATORY AND CRITICAL CARE MEDICINE Volume: 155 Issue: 1 Pages: 260-267 Published: JAN 1997.
- 20 3. Kharitonov, Sergei A. et. al., "Nasal contribution to exhaled nitric oxide during exhalation against resistance or during breath holding," THORAX, Volume: 52 Pages: 540-544 Published 1997.
- 25 4. Hogman, M. et. al., "Nitric oxide from the human respiratory tract efficiently quantified by standardized single breath measurements," SCANDINAVIAN PHYSIOLOGICAL SOCIETY, Volume: 159, Pages: 345-346, Published: 1997.
- 30 5. Malmberg, L. P. et. al., "Exhaled nitric oxide rather than lung function distinguishes preschool children with probable asthma," THORAX, Volume: 58, Pages: 494-499, Published 1997.
- 35 6. American Thoracic Society and European Respiratory Society, "ATS/ERS Recommendations for Standardized Procedures for the Online and Offline Measurement of Exhaled Lower Respiratory Nitric Oxide and Nasal Nitric Oxide," AMERICAN JOURNAL OF RESPIRATORY AND CRITICAL CARE MEDICINE, Volume 171, Pages 912 – 930, Published, 2005.
- 40 7. Gustafsson, L. E. et. al., "Endogenous Nitric Oxide is Present in the Exhaled Air of Rabbits, Guinea Pigs and Humans," BIOCHEMICAL AND BIOPHYSICAL RESEARCH COMMUNICATIONS, Volume: 181, Number: 2, Pages: 852-857, Published December 1991.

8. Persson, Magnus et. al., "Single-breath nitric oxide measurements in asthmatic patients and smokers," LANCET, Volume: 343, Pages: 146-147, Published: 1994.
- 5 9. Matsumoto, Akihiro et. al., "Increased Nitric Oxide in the Exhaled Air of Patients with Decompensated Liver Cirrhosis," ANN. INTERN MED., Volume 123, Pages: 110-113, Published 1995.

Claims

What is claimed is:

1. A low back pressure mouthpiece for accurate detection of exhaled nitric oxide (NO) comprising:

5 conduit means for receiving the exhaled breath from the subject;

 an oxidizing filter means coupled to the conduit means for sample conditioning, wherein the oxidizing filter means has an outlet and wherein the conduit means and oxidizing filter means operate to produce a back pressure of less than 4 cm H₂O; and

10 means for measuring the level of one or more components of exhaled breath received from the outlet.

2. The apparatus of claim 1, wherein the oxidizing filter means comprises acrylic tubing containing chemical particles.

15

3. The apparatus of claim 1 wherein the chemical particles are selected from the group consisting of desiccant, an oxidant and a solid porous substrate impregnated with sodium permanganate.

20

4. The apparatus of claim 3 wherein the desiccant is selected from the group consisting of CaSO₄, activated alumina, aerogel, benzophenone, bentonite clay, calcium chloride, calcium sulfate, cobalt(ii) chloride, copper(ii) sulfate, lithium chloride, lithium bromide, magnesium sulfate, magnesium perchlorate, molecular sieve, potassium carbonate, silica gel, sodium, sodium chlorate, sodium chloride, sodium hydroxide, sodium sulfate, sucrose and combinations thereof.

25

5. The apparatus according to claims 3 or 4 wherein the oxidant is selected from the group consisting of Purafil media, oxygen (O₂), ozone (O₃), hydrogen peroxide (H₂O₂), inorganic peroxides, fluorine (F₂), chlorine (Cl₂), halogens, nitric acid (HNO₃) and nitrate compounds, sulfuric acid (H₂SO₄), peroxydisulfuric acid (H₂S₂O₈), peroxymonosulfuric acid (H₂SO₅), chlorite, chlorate, perchlorate, halogen compounds, hypochlorite, hypohalite compounds, household bleach (NaClO), hexavalent chromium compounds, chromic acid, dichromic acid, chromium trioxide,

30

pyridinium chlorochromate (PCC), chromate/dichromate compounds, permanganate compounds, potassium permanganate, sodium perborate, nitrous oxide (N_2O), silver oxide (Ag_2O), osmium tetroxide (OsO_4) and combinations thereof.

5 6. The apparatus according to claim 5, wherein the desiccant works under a range of relative humidity levels from 10% to 95% RH.

7. The apparatus of claim 2 wherein the inner diameter of the acrylic tubing is at least 18 mm and all the chemical particles are contained in the acrylic tubing.

10

8. The apparatus of claim 7 wherein the acrylic tubing has an inner diameter of at least 9.6 mm and outer diameter of at least 12.6 mm.

9. The apparatus of claim 1 wherein at least two felt pieces with diameter of 18 mm
15 are used as a stopper to keep the chemical particles in the mouthpiece.

10. The apparatus of claim 9 wherein the at least two felt pieces comprise polyester fiber.

20 11. The apparatus of claim 1 wherein the mouthpiece has a capability of conditioning the breath at a flow of 50 ml/sec with an associated error of $\pm 10\%$ under sample collection condition with back pressure less than 4 cm H_2O .

12. The apparatus of claim 9, further comprising means for providing the subject
25 with an instantaneous display of the pressure of the exhaled breath so that the subject can adjust the force of the exhalation to maintain a constant pressure.

13. The apparatus of claim 1, wherein said one or more components is selected from the group consisting of carbon dioxide, oxygen, nitric oxide, nitrogen, nitrogen
30 dioxide, hydrogen peroxide, proteins, surfactants, DNA, acetone, ammonia, sulfur compounds, acetylene, carbon monoxide, ethane and pentane.

14. The apparatus of claim 1, wherein said one or more components is nitric oxide.

15. A method for accurate detection of exhaled nitric oxide (NO) comprising:

5 receiving the exhaled breath from the subject through a conduit attached to a low back pressure mouthpiece, wherein the low back pressure mouthpiece has a back pressure value less than 4 cm H₂O;

using an oxidizing filter means for sample conditioning, wherein the low back pressure mouthpiece has a back pressure value less than 4 cm H₂O; and

10 measuring the level of one or more components of the received exhaled breath.

16. The method of claim 15, wherein the oxidizing filter means comprises acrylic tubing containing chemical particles.

15

17. The method of claim 16 wherein the chemical particles are selected from the group consisting of desiccant, an oxidant and a solid porous substrate impregnated with sodium permanganate.

20 18. The method of claim 17 wherein the desiccant is selected from the group consisting of CaSO₄, activated alumina, aerogel, benzophenone, bentonite clay, calcium chloride, calcium sulfate, cobalt(ii) chloride, copper(ii) sulfate, lithium chloride, lithium bromide, magnesium sulfate, magnesium perchlorate, molecular sieve, potassium carbonate, silica gel, sodium, sodium chlorate, sodium chloride, 25 sodium hydroxide, sodium sulfate, sucrose and combinations thereof.

19. The method according to claims 17 or 18 wherein the oxidant is selected from the group consisting of Purafil media, oxygen (O₂), ozone (O₃), hydrogen peroxide (H₂O₂), inorganic peroxides, fluorine (F₂), chlorine (Cl₂), halogens, nitric acid (HNO₃) 30 and nitrate compounds, sulfuric acid (H₂SO₄), peroxydisulfuric acid (H₂S₂O₈), peroxymonosulfuric acid (H₂SO₅), chlorite, chlorate, perchlorate, halogen compounds, hypochlorite, hypohalite compounds, household bleach (NaClO), hexavalent chromium compounds, chromic acid, dichromic acid, chromium trioxide,

pyridinium chlorochromate (PCC), chromate/dichromate compounds, permanganate compounds, potassium permanganate, sodium perborate, nitrous oxide (N₂O), silver oxide (Ag₂O), osmium tetroxide (OsO₄) and combinations thereof.

- 5 20. The method according to claim 19, wherein the desiccant works under a range of relative humidity levels from 10% to 95% RH.

21. The method of claim 15 wherein the mouthpiece has a capability of conditioning the breath at a flow of 50 ml/sec with an associated error of $\pm 10\%$ under sample
10 collection condition with back pressure less than 4 cm H₂O.


22. The method of claim 20, further comprising means for providing the subject with an instantaneous display of the pressure of the exhaled breath so that the subject can adjust the force of the exhalation to maintain a constant pressure.


15

23. The method of claim 15, wherein said one or more components is selected from the group consisting of carbon dioxide, oxygen, nitric oxide, nitrogen, nitrogen dioxide, hydrogen peroxide, proteins, surfactants, DNA, acetone, ammonia, sulfur compounds, acetylene, carbon monoxide, ethane and pentane.

20

24. The method of claim 15, wherein said one or more components is nitric oxide.

FIG. 1A

**FIG. 1B**




FIG. 2





FIG. 3


FIG. 4



Typical back pressure of the mouthpiece







Equation	$Y=A+B \cdot X$
Parameter	Value
A	5.05525E-4
B	1
R	0.94062

FIG. 7




FIG. 8




FIG. 2