



(51) International Patent Classification:

A61B 5/0452 (2006.01) A61B 18/14 (2006.01)
A61B 5/0472 (2006.01) A61B 6/00 (2006.01)
A61B 5/042 (2006.01) A61B 18/00 (2006.01)

(21) International Application Number:

PCT/EP2016/060046

(22) International Filing Date:

4 May 2016 (04.05.2016)

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

15305689.0 6 May 2015 (06.05.2015) EP

(71) Applicants: **CARDIOMED CONSULTING LTD** [NZ/NZ]; 31 A Bournemouth Terrace, Murrays Bay, 0630 Auckland (NZ). **CENTRE HOSPITALIER UNIVERSITAIRE DE BORDEAUX** [FR/FR]; 12, Rue Dubernat, 33404 Talence (FR). **UNIVERSITE DE BORDEAUX** [FR/FR]; 35, Place Pey Berland, 33000 Bordeaux (FR). **INSTITUT NATIONAL DE LA SANTE ET DE LA RECHERCHE MEDICALE** [FR/FR]; 101, Rue de Tolbiac, Paris 75013 (FR).

(72) Inventors: **HOOKS, Darren**; 77, Ola Karori Road Karori, 6011 Wellington (NZ). **MEILLET, Valentin**; 5 bis rue du 8 mai 1945, 33320 Eysines (FR). **DUBOIS, Rémi**; 408 Avenue de Verdun, 33700 Merignac (FR).

(74) Agents: **CAMUS, Olivier** et al.; 25 rue de maubeuge, 75009 Paris (FR).

(81) Designated States (unless otherwise indicated, for every kind of national protection available): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BN, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IR, IS, JP, KE, KG, KN, KP, KR, KZ, LA, LC, LK, LR, LS, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PA, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SA, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.


(84) Designated States (unless otherwise indicated, for every kind of regional protection available): ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, ST, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, KM, ML, MR, NE, SN, TD, TG).

Published:

- with international search report (Art. 21(3))
- before the expiration of the time limit for amending the claims and to be republished in the event of receipt of amendments (Rule 48.2(h))

(54) Title: SYSTEM FOR AUTOMATED RHYTHM-BASED CONTROL OF CARDIAC ABLATION CLOSE TO THE ATRIOVENTRICULAR NODE

FIG. 7



(57) Abstract: A monitoring system for ablation of cardiac tissue in the close proximity of the AV node of the heart, characterized in that it comprises : - At least one ablation catheter (12) powered by a power source (16); - At least one electrode (13) acquiring an in-tracardiac electrograms (IEGM), called IEGM electrode; - At least one set of electrodes (14) acquiring a surface electrocardiogram (ECG), called ECG electrodes; - At least one calculator (15, 17, 25) and a memory for measuring in real time signals acquired from the IEGM electrode and the ECG electrodes, treating both signals (IEGM, ECG) for automatically generating at least one stop command (Sc) when detecting at least one given arrhythmia event, the at least generated stop command (Sc) controlling the power source (16) of the ablation catheter (12) in order to automatically cut-out ablation catheter energy.

SYSTEM FOR AUTOMATED RHYTHM-BASED CONTROL OF CARDIAC ABLATION CLOSE TO THE ATRIOVENTRICULAR NODE

5 **Technical field**

The present invention relates to the field of catheter ablation of cardiac arrhythmia, and specifically enables safer ablation close to the atrioventricular node via automated cut-out of ablation in response to markers of risk for heart block.

10 **Background technology**

Cardiac arrhythmia occurs when the normal co-ordinated electrical rhythm of the heart is lost. A patient with cardiac arrhythmia may experience palpitations, fatigue, light-headedness, or, in extreme cases, loss of consciousness and death. Many cardiac arrhythmia are amenable to treatment with catheter based ablation. These procedures are typically performed with catheters inserted inside the heart via the femoral veins as represented in figure 1. Ablation energy, typically in the form of radiofrequency RF alternating current, or freezing (cryoablation), is delivered through an 'ablation catheter' 12 via a power source 16, and is used to permanently disrupt/extinguish (ablate) areas of cardiac tissue that are critical to the maintenance of the specific arrhythmia. Ablation catheters are designed to direct ablation energy to the tissue in a focal manner, via conducting tips usually between 4 and 8mm in length and ~2.5mm diameter. The ablation catheter is able to be flexed and rotated to reach all regions of the heart. The power source or generator unit 16 is capable of generating the ablation energy. In the case of RF ablation, current flows from the catheter tip to a return patch electrode on the body surface, causing localised tissue heating and cellular necrosis (death) around the catheter tip. In the case of cryoablation, the catheter tip is cooled initially to ~-30°C to assess effect, and then cooled further to ~-50°C to -70°C to cause permanent cellular necrosis and tissue disruption. Catheter ablation is an increasingly common procedure, with an estimated 472 procedures performed per million population in France in 2012, and frequently provides a cure to patients with arrhythmia.

In the normal heart, cardiac cells (myocytes) hold negative charge when at rest. During normal function, each heart beat is preceded by a change in cellular charge, so that cells become positively charged ('depolarised'). This process of depolarisation (or 'activation') travels through cardiac tissue as a self-sustaining

electrical impulse, travelling as a wavefront, at speeds of 30-100 cm/s, and triggers the heart muscle to contract.

Figure 2A shows a schematic of the heart cut-open to reveal the inside of the right atrium 37 and right ventricle 38 (cut surfaces shown in grey). The right atrium and right ventricle are separated by the tricuspid valve 34.

AV node

Figure 2A demonstrates the essential components of the electrical function of the heart. The electrical impulse originates in the sinoatrial node 35 (the pacemaker) located in the right atrium and propagates, as shown in figure 2B, through the atria to reach the atrioventricular node. The atrioventricular node is called the "AV node". In the normal case, the AV node is the sole electrical connection between the atria (upper chambers of the heart) and ventricles (lower chambers of the heart), and conducts the impulse to the ventricles. The AV node acts as a temporal filter, limiting the rate at which the ventricles can be electrically stimulated, which protects the ventricles against excessively fast atrial rhythms. As the sole connection between atria and ventricles in the normal heart, the AV node allows co-ordination of atrial and ventricular contraction.

Atrial side : Slow pathway, fast pathway

On the atrial side, the AV node has a diffuse structure into which the electrical impulse enters. This structure consists of limbs that display varying conduction speeds. These limbs are commonly known as the fast pathway 32 and slow pathway (33), as shown in figure 2A. However there may be more or less than two physiologically distinct limbs, and the limbs represent functional rather than anatomically defined inputs.

Ventricular side : His bundle, Purkinje fibre branches

On the ventricular side, the conducting limbs of the AV node coalesce to form a more discrete structure, known as the compact AV node 31, which funnels the propagating impulse to a cable-like structure known as the "His bundle" 30 which directs depolarisation to the ventricles, via the Purkinje fibre branches 41. Both the His bundle 30 and Purkinje fibres 41 support a high conduction velocity, enabling rapid propagation of the electrical impulse towards and throughout the ventricular tissue.

IEGM electrodes

The process of cardiac tissue depolarisation gives rise to electrical potential that can be recorded on electrodes placed inside the heart as so-called
5 'intracardiac electrograms (IEGMs)'. Such electrodes may be embedded into catheters placed into either the atria, coronary sinus, or ventricles. Electrograms recorded inside the heart are typically of 1-10 millivolts (mV) in amplitude.

ECG electrodes

10 Cardiac tissue depolarisation also gives rise to electrical potential on the body surface, which is recorded on the surface electrocardiogram called ECG and shown in figure 2C. Atrial tissue depolarisation causes a small voltage signal known as the p wave 44, whereas ventricular tissue depolarisation leads to generation of a larger electrical potential on the body surface, known as the QRS complex 45, which
15 is usually 1-5 mV in amplitude.

Measurement of the A-V interval

When ventricular depolarisation occurs via impulse propagation through the rapidly conducting fibres of the His bundle 30 and Purkinje branches 41, the
20 QRS complex is of narrow morphology, meaning typically < 120ms in width.

Conversely, when ventricular depolarisation arises from an abnormal focus within the ventricular tissue itself, the QRS complex becomes widened, specifically > 120ms. The interval between onset of the p wave and onset of the QRS complex is known as the PR interval 46, and indicates the time taken for the
25 electrical impulse to traverse between atria and ventricle.

Alternatively, the PR interval can be assessed in surrogate fashion by the interval between atrial and ventricular intracardiac electrogram recordings (described as the A-V interval).

30 Accessory pathway

In < 1% of humans, an abnormal connection termed 'accessory pathway'
50, shown in figure 3A, is present between the atria and ventricles. In that case, the electrical impulse is able to travel from atria to ventricles not only via the AV node, but also via the accessory pathway. The presence of an accessory pathway may
35 alter the morphology of the normal QRS complex 45, making it wider and with a

slurred upstroke. Figure 3B depicts such a QRS complex, referred to as a 'pre-excited' QRS. In addition, the normal PR interval 46 may be shorted to become < 120ms as represented in figure 3B. Individuals with a pre-excited QRS are said to have the Wolff-Parkinson-White syndrome, and in rare cases are prone to sudden cardiac death, due to loss of the temporal filtering properties of AV node conduction.

Arrhythmia / AVNRT & AVRT

In normal ('sinus') heart rhythm, activation of the heart proceeds in ordered fashion from the atria to the ventricles as shown in figure 2B and 3A. Arrhythmia occurs when there is disturbance of the normal sequence of spread of electrical impulse. Often, arrhythmia is caused by 're-entry', where impulse propagation re-enters previously depolarised (activated) tissue in a circular fashion, causing repeditive and rapid activation of tissue, which leads to sudden elevation of heart rate. For re-entry to occur, a circuit of cardiac tissue is required, typically around an anatomical or functional obstacle, through which the impulse wave is unable to propagate. This circuit can be formed by the two limbs of the AV node, in which case the arrhythmia is termed atrioventricular nodal re-entrant tachycardia called AVNRT and shown in figure 4A. Because these limbs have different functional (electrical) properties, they are apt to sustain re-entry in susceptible persons. Alternatively, in persons with the Wolff-Parkinson-White syndrome, the re-entrant circuit may be formed by the AV node and an accessory pathway, in which case the arrhythmia is termed atrioventricular reciprocating tachycardia called AVRT and shown in figure 4B.

Ablation for AVNRT / Slow pathway

Ablation for AVNRT is typically performed over the slow pathway of the AV node 33 to disrupt or modify the slow pathway to an extent that re-entrant impulse propagation is no longer possible.

Figure 5 depicts ablation over the slow pathway 33 using an ablation catheter 12, with an additional diagnostic catheter 13 placed in the position of the high right atrium.

Ablation for AVRT / Accessory pathway

Ablation for AVRT is typically performed over the accessory pathway 50, which may lie anywhere along the valve planes that separate the atria and

ventricles. In some cases, the accessory pathway 50 may be close, or adjacent, to the AV node, lying on the inter-atrial septum, as demonstrated in FIG. 3A. In these cases it is described as being located in mid- or antero- septal location. Ablation of a mid- or antero- septal accessory pathway may be undertaken for the purposes of cure of AVRT, or for the prevention of sudden cardiac death in individuals with the Wolff-Parkinson-White (WPW) syndrome.

Long-term cure rates of AVNRT with a single ablation procedure are ~95%. Ablation of mid- or antero- septal APs is successful in approximately 80% of cases with a single procedure.

Ablation procedure : Two catheters needed

Catheter ablation is generally performed following a diagnostic electrophysiology study, in which the mechanism of arrhythmia is determined. Such diagnostic study requires a minimum of two catheters, one of which may be the ablation catheter 12.

One catheter is placed in the right atrium or coronary sinus to record an atrial intra-cardiac electrogram A IEGM. Another catheter is placed alternately between the region of the His bundle to record a His electrogram, and the right ventricular apex to record a ventricular intra-cardiac electrogram V IEGM. Additional catheters may also be placed according to operator preference to allow the simultaneous recording of IEGMs in more locations.

During the diagnostic study, the heart rhythm is manipulated (stimulated) by pacing protocols in ways that define the arrhythmia mechanism. In general, programmed stimulation is required from both the atria and the ventricles in order to define the mechanism of arrhythmia. During ablation of AVNRT, AVRT, or a septal accessory pathway, a minimum of two catheters is generally required, one being the ablation catheter, and the other being the catheter recording an atrial intra-cardiac electrogram. The atrial intra-cardiac electrogram is required primarily because atrial electrical activity is not always well seen on the body surface ECG, due to atrial activity being of smaller amplitude compared to ventricular activity. At times, it is also necessary to perform atrial pacing during ablation.

Risks of the ablation procedure

Catheter ablation of arrhythmia is generally a well-tolerated procedure., however, carries a degree of risk. Specifically, ablation of either AVNRT, or a mid- or

antero- septal accessory pathway, risks damage to the compact AV node or His bundle due to the proximity of the ablation catheter to these structures as shown in figure 5.

AV block

5 Damage to the compact AV node or His bundle can cause heart block to occur, hereafter referred to as 'AV block'. AV block describes a condition where conduction of electrical impulses between the atria and ventricles is limited or completely extinguished. Heart block leads to a loss of co-ordination between the atria and the ventricles, and can lead to a lowering of ventricular heart rate, loss of
10 blood pressure, loss of consciousness, or even death.

Pacemaker

 Where damage to the compact AV node or His bundle causes permanent AV block, implantation of a permanent pacemaker is usually required to
15 restore atrial and ventricular synchrony. Permanent pacemakers carry significant morbidity due to the necessity of further surgery for generator changes every 5 to 12 years, and the possibilities of generator or lead failure, infection requiring removal, and detrimental long-term effects of pacing on ventricular function. It is therefore highly desirable to minimise the risk of AV block when performing ablation close to
20 the AV node. Currently, as many as 26% of cardiac ablation procedures are for AVNRT. With current ablation equipment and techniques, the complication of AV block and pacemaker insertion occurs in between 1:130-1:450 cases.

 Ablation for mid- or antero- septal AP comprises < 5% of all ablation procedures, however, the risk of AV block is higher than for AVNRT, and is
25 estimated at about 2-5%.

 Ablation of AVNRT or AVRT is generally performed during normal (sinus) heart rhythm, however may be performed during tachy-arrhythmia.

Cardiac rhythm

30 During ablation, heart rhythm may be expected to fluctuate between sinus rhythm as shown in figure 6A, junctional rhythm as shown in figure 6B, or ventricular ectopic rhythm as shown in figure 6C.

 In sinus rhythm, if AV node function is normal, each atrial depolarisation, as seen on the atrial intracardiac electrogram A IEGM, should be followed within
35

~250ms by a ventricular depolarisation, as seen either on a ventricular intracardiac electrogram V IEGM, or, on the surface ECG as a QRS complex, as shown in figure 6A.

Junctional rhythm is caused by heating to the slow pathway 33 and compact AV node 31 and/or His bundle 30, and describes a concentrically propagating wavefront emanating from the AV node or His bundle and causing near simultaneous atrial and ventricular activation, such that atrial depolarisation and ventricular depolarisation occur within ~130ms of each other.

In both sinus and junctional rhythms, activation of the ventricles is via the His bundle 30 and Purkinje fibres 41, and accordingly the QRS complex on the ECG is of narrow type as shown in figures 6A and 6B, excluding conditions of disease to the Purkinje fibres.

Ventricular ectopic rhythm describes activation commencing in the ventricular tissue, with retrograde AV nodal conduction to the atria, as illustrated in figure 6C, such that each atrial depolarisation should lag ventricular depolarisation by no more than ~ 250ms. Ventricular ectopic rhythm can occur spontaneously, or be induced by the presence of a catheter in the ventricle, in which case it is caused by the mechanical action of the catheter against the ventricular wall. In the case of ventricular ectopic rhythm, activation of the ventricles is not via the His and Purkinje fibres, and accordingly the QRS complex of the ECG is widened and of different morphology to normal as shown in figure 6C. The timing limits of 130ms and 250ms described here are a guide, and may differ based on exact location of intracardiac catheters or placement of ECG leads, and the heart rate and physiology of the individual patient. Nonetheless, absolute timings prove a reliable method of determining abnormal electrical impulse conduction in the AV node.

Indicators of AV block

AV block caused by ablation is often heralded by changes in the expected cardiac rhythms described above. There may be transient loss of atrio-ventricular association, such that each atrial depolarisation is not followed by a ventricular depolarisation, or vice versa, or the time limits for atrial or ventricular depolarisation outlined above are exceeded. In other cases, excessively rapid junctional rhythm (>130 beats per minute) may indicate impending AV block. In yet other cases, a prolongation of the normal PR interval 46 may indicate imminent AV block.

AV block may also be heralded by sudden displacement of the ablation catheter tip away from the desired location of ablation, towards the compact AV node 31 and/or His bundle 30. Such displacement can be detected radiographically by continuous low-intensity x-ray (fluoroscopy) of the catheter, or, via the monitoring of catheter tip location with 3-dimensional mapping systems 19 which typically use impedance and/or magnetic tracking methods to localise the catheter tip.

Manual termination of ablation

Currently, the risk of AV block is minimised by close attention of the operator (and ancillary staff) to cardiac rhythm and ablation catheter location during ablation, with manual termination of ablation energy, commonly by use of a foot pedal 20, where risk markers of heart block are identified. There is, therefore, a human visual feedback loop 22, as illustrated in figure 1, between cardiac rhythm and catheter location assessment, and manual control of the ablation generator.

However, visual monitoring of both cardiac rhythm and catheter location during ablation are complex tasks, each task competing with the needs of the other. In addition, human reaction times to changes of cardiac rhythm are variable, often exceed one second, and are frequently longer in trainees learning the art of cardiac ablation. A system to facilitate more rapid and consistent termination of ablation energy in response to indicators of AV block is therefore desirable.

Prior Art

The US patent No. 7,578,816 B2 (Boveja et al.) is directed toward control of ablation in similar cardiac rhythms. The Boveja application mainly discloses using automated monitoring of cardiac rhythm based exclusively on intracardiac signals, acquired from within both an atrial chamber and a ventricular chamber. This approach suffers several important limitations, which the present invention seeks to address.

Firstly, this solution (Boveja application) requires placement of a third catheter during ablation, namely a ventricular catheter. The use of an additional catheter adds considerable cost to the procedure.

Secondly, presence of a ventricular catheter during ablation significantly increases the burden of ventricular ectopic beats, termed here ventricular ectopic rhythm, as shown in figure 6C. This is related to the mechanical action of the

catheter on the ventricular tissue. Although ventricular ectopic rhythm can be reduced by careful catheter placement, it cannot usually be completely eliminated.

Evidence of such increase in ventricular ectopic rhythm is shown in FIG. 8, which documents the percentage of ventricular ectopic rhythm observed during 215 routine ablations performed in 20 patients, with a ventricular catheter placed by a clinician expert in the field of cardiac ablation, and during 196 ablations performed in 20 patients without a ventricular catheter placed. In the former group, only 36% of ablations were not complicated by ventricular ectopic rhythm, and the remaining 64% of ablations suffered from ventricular ectopic rhythm ranging from 0.1% to 20% of all heart beats. In the latter group, 95% of ablations were not complicated by ventricular ectopic rhythm, and only 5% of ablations suffered from ventricular ectopic rhythm, of <2% of all heart beats.

Ventricular ectopic rhythm during ablation is undesirable because it may limit the ability to discern the integrity of AV nodal conduction, and therefore continue ablation in a safe manner. This is because if a ventricular ectopic beat occurs in the period immediately following initiation of a sinus beat in the sinoatrial node 35, the state of AV nodal conduction is indeterminate, meaning that it cannot be determined if the electrical impulse was safely conducted by the AV node, or indeed, if AV block had occurred. The phenomenon of ventricular ectopic rhythm masking the integrity of AV nodal conduction is illustrated in FIG. 9A, and an example of signals acquired from a patient during ablation demonstrating this phenomenon is given in FIG. 9B.

Thirdly, reliance on ventricular intracardiac recordings for discerning cardiac rhythm limits a device's ability to distinguish various cardiac rhythms. For example, such a device is unable to:

(i) differentiate :

- junctional rhythm with impaired (slowed) AV nodal conduction, ie. atrial depolarisation lagging ventricular depolarisation by > 130ms, indicating impending heart block; as shown in figure 10A, from :
- ventricular ectopic rhythm with normal AV nodal conduction, ie. atrial depolarisation lagging ectopic ventricular depolarisation by < 250ms, as shown in figure 10B.

It can be seen from figures 10A and 10B that the intracardiac signals do not substantially differ in timing with each of these two rhythms.

(ii) differentiate :

- 5 ▪ sinus rhythm with normal AV nodal conduction, ie. atrial depolarisation followed by a normal ventricular depolarisation within 250ms, as illustrated in figure 11A, from :
- sinus rhythm with indeterminate AV nodal conduction due to occurrence of a ventricular ectopic, ie. atrial depolarisation followed by a ventricular ectopic depolarisation within 250ms, as illustrated in figure 11B.

10 It can be seen from figures 11A and 11B that the intracardiac signals do not substantially differ in timing with each of these two rhythms.

 (iii) distinguish pre-excited ventricular activation, as shown in figures 3A and 3B, from normal ventricular activation, as shown in figures 2B and 2C, independent of the speed/integrity of AV nodal conduction. Such distinction is
15 important during the ablation of an accessory pathway.

 Another important distinction is that the Boveja application is limited to cut-out of ablation via placement of a control switch within the power supply line of
20 the ablation generator, or between the ground patch and the ablation generator unit, or between the generator unit and the ablation catheter. These arrangements all suffer from the requirement to manually reset the generator unit after a cut-out condition has occurred. For example, in the Boveja application, figure 12 discloses "reset ablation and interface apparatus". Such manual reset is highly undesirable in
25 a clinical system designed for ablation, because cut-out conditions occur frequently, on average twelve times per patient, and most operators prefer a seamless way of recommencing ablation after checking the safety of catheter position, without the requirement to reset equipment.

 Yet another distinction is that the Boveja application does not address
30 the automated assessment of catheter tip location.

SUMMARY OF THE INVENTION

 The invention concerns a monitoring system for ablation of cardiac tissue in close proximity to the AV node of the heart. In another aspect the invention

concerns a monitoring system for improving the safety of ablation in close proximity to the AV.

The monitoring system of the invention comprises :

- At least one ablation catheter powered by a power source (16);
- 5 ▪ At least one electrode acquiring an intracardiac electrogram, called IEGM electrode ;
- At least one set of electrodes acquiring a surface electrocardiogram, called ECG electrodes ;
- 10 ▪ At least one calculator and a memory for measuring in real time signals acquired from the IEGM electrode and the ECG electrodes automatically in order to generate at least one stop command (Sc) when a cardiac risk indicator is generated by combining signals measured from the IEGM electrode and the ECG electrodes, the generated stop command (Sc) controlling the power source of the ablation catheter in order to
- 15 automatically cut-out ablation catheter energy.

In one embodiment of the invention, the monitoring system comprises at least one calculator and a memory for measuring in real time a displacement of the ablation catheter in order to automatically generate a stop command when detecting

20 a displacement of the ablation catheter exceeding a predefined limit.

In this application we describe such a system, which utilises computational real-time analysis of cardiac rhythm and ablation catheter tip location in order to automatically cut-out ablation, in a near-instantaneous fashion, when

25 markers of risk for AV block are detected.

The system of the invention is directed to improvement of the safety of ablation when applied close to the AV node. A system is described in figure 7 that allows real-time or near real-time monitoring of cardiac rhythm and assessment of

30 rhythm conditions that may herald development of AV block. The system also allows real-time or near real-time analysis of ablation catheter tip displacements that may risk AV block. The system allows automatic termination of ablation energy where cardiac rhythm or displacement conditions are identified that indicate risk of AV block, with faster reaction times and increased consistency compared to human

35 manual response.

An advantage of the invention compared to Prior Art is to utilise analysis of the body surface electrocardiogram ECG for determining ventricular depolarisation, in place of a ventricular intra-cardiac electrogram. The system employed incorporates both ECG QRS complex timing and morphology analysis.

In doing so, the system of the invention has the advantages of :

- one less catheter required during ablation because the ventricular catheter is not required,
- less ventricular ectopic rhythm due to absence of mechanical action of a ventricular catheter on the heart, and ;
- enhanced ability to discriminate cardiac rhythm. An outline of the enhanced cardiac rhythm discrimination can be made by reference to figures 10 and 11.

The surface ECG can be used to reliably distinguish the rhythms of figures 10A and 10B based on analysis of QRS complex morphology. In the case of junctional rhythm with impaired (slowed) AV nodal conduction, as shown in figure 10A, the QRS complex morphology is narrow, and ablation should be discontinued. However, in the case of ventricular ectopic rhythm with normal AV nodal conduction as shown in figure 10B, the QRS complex has a widened morphology, and it is safe to continue ablation. Similarly, the surface ECG can be used to reliably distinguish the rhythms of figures 11A and 11B based on analysis of QRS complex morphology. In the case of sinus rhythm with normal AV nodal conduction, as represented in figure 11A, the QRS complex morphology is narrow, and it is safe to continue ablation. However, in the case of sinus rhythm with indeterminate AV nodal conduction due to occurrence of a ventricular ectopic, see figure 11B, the QRS complex morphology is widened, and it may not be safe to continue.

Another advantage of the invention compared to Prior Art is to offer an ablation cut-out that mimics the manual release of a foot pedal connected to the ablation generator. The foot pedal is used to manually initiate and terminate delivery of ablation energy, and in addition, computer controlled cut-out of ablation delivery simulates release of the foot pedal. In this way, resumption of ablation following an automated cut-out condition can be made in seamless fashion, by merely releasing

and depressing the foot pedal again, without the need to perform a manual reset of the ablation generator or any other equipment.

Advantageously, the monitoring system may be achieved by the
5 integration of a calculator or a set of calculators into one or more of an electrophysiology (EP) recording system (17), an electrophysiology (EP) mapping system (19), a Pacing/Stimulation Unit (18), a junction box (15), an ablation generator (16) or into a stand-alone Rhythm Monitoring Device (25).

10 Advantageously, the monitoring system may be achieved by the integration of a memory or a set of memories into one or more of an electrophysiology (EP) recording system (17), an electrophysiology (EP) mapping system (19), a Pacing/Stimulation Unit (18), a junction box (15), an ablation generator (16) or into a stand-alone Rhythm Monitoring Device (25).

15 In one aspect of the invention, the calculator comprises a signal timing function in order to derive timing of atrial and ventricular depolarisation, the atrial depolarisation being measured by the IEGM electrode, and the ventricular depolarisation being measured by the ECG electrodes.

20 In one embodiment, the signal timing function incorporates a voting function allowing the determination of signal timing when more than one IEGM or more than one ECG signal is available to characterise depolarisation timing. One advantage is to avoid false detections of signal timing.

25 In one embodiment, the monitoring system comprises only one catheter which comprises at least one electrode allowing the acquisition of an intracardiac electrogram, as the IEGM electrode.

30 In one embodiment, the signal timing function compares the delay between atrial and ventricular depolarisations to a minimum predefined threshold of time and/or a maximum predefined threshold of time. The signal timing function is configured for generating a cardiac risk indicator when said measured delay exceeds at least one of the predefined thresholds.

35

In one embodiment, the calculator comprises a QRS complex signal morphology function in order to determine the nature of the ventricular depolarisation as being via the His bundle and Purkinje fibre branches , or, via a ventricular ectopic, or, via an accessory pathway.

5

In one embodiment, the signal morphology function compares an acquired QRS complex signal with one or more template QRS complexes stored in a memory, the comparison being processed using an auto-correlation function or a cross-correlation function.

10

In one embodiment, the signal morphology function is configured to compute the QRS complex width and/or to determinate deflection sequence of the complex QRS signal.

15

In one embodiment, the signal morphology function incorporates a voting function allowing the determination of signal morphology when more than one ECG signal is available to characterise depolarisation morphology. One advantage is to avoid false detections of signal morphology.

20

In one embodiment, the calculator is configured to characterise the real time or near real time cardiac rhythm given return values of the signal timing function and/or a return value of the signal morphology function. One advantage is to accurately define cardiac rhythm and detect rhythms which indicate a risk of AV block, allowing the automatic cut-out of ablation power.

25

In one embodiment, the cardiac risk indicator is a probability of AV block occurrence, said probability being characterised by at least one occurrence of :

30

- an atrio-ventricular interval prolongation ;
- a ventriculo-atrial interval prolongation ;
- an atrio-ventricular dissociation ;
- a ventriculo-atrial dissociation ;
- a fast junctional rhythm ;
- a PR interval prolongation ;
- a consecutive atrial depolarisations ;
- a consecutive ventricular depolarisations, or ;
- an indeterminate atrio-ventricular nodal conduction.

35

In one embodiment, the memory allows storage of data acquired by the electrodes, storage of QRS complex templates for the purpose of morphology analysis, and storage of predefined parameters such as the cardiac rhythms and ablation catheter tip displacements deemed to indicate risk of AV block.

In one embodiment, the calculator comprises a mapping function allowing the real-time or near real-time automatic tracking of the location of the tip of the ablation catheter, said location of the tip being determined in three-dimensions by an EP mapping system (19), and/or being determined by low dose x-ray (fluoroscopy). One advantage is to allow automatic cut-out of the ablation catheter power supply (16) in response to a sudden displacement of the catheter tip away from the desired or initial location of ablation.

In one embodiment, the calculator is configured to account for the type of ablation procedure, being one of AVNRT, AVRT, or accessory pathway (50) ablation. One advantage is to allow tolerance of different cardiac rhythms or the setting of different AV node conduction time limits depending on the type of ablation being performed.

In one embodiment, the calculator comprises a signal amplification function, a filtering function, an analog to digital conversion function, and computational analysis functions of signal timing and signal morphology. One advantage is elimination of noise from the IEGM and ECG signals, and conversion of signals to digital format thereby allowing digital processing by the signal timing and morphology functions.

In one embodiment, the monitoring system comprises a foot pedal (20) for manually switching off the power source (16) of the ablation catheter (12) via foot pedal release, and a switch unit (26) for automatically switching off the power source (16) in a fashion that mimics manual release of the foot pedal.

In one embodiment, the switch unit comprises :

- a first interface with the calculator allowing the transmission of the stop command to the power source and ;

- a second interface with the foot pedal allowing the transmission of the state of the foot pedal to the power source.
- A third interface with the foot pedal allowing the transmission of the state of the foot pedal to the calculator

5

In one embodiment, the monitoring system comprises a user interface allowing the selection of :

- a first set of parameters as inputs to the signal timing function, and/or ;
- 10 ▪ a second set of parameters as inputs to the signal morphology function, and/or ;
- a third set of parameters for processing the location of the ablation catheter tip, and/or ;
- a fourth set of parameters for defining the cardiac rhythms of risk which will trigger cut-out of ablation, and/or ;
- 15 ▪ a fifth set of parameters for defining the ablation catheter displacement limit which will trigger cut-out of ablation.

One advantage is to account for a specific patient's physiology, such that tolerance to various cardiac rhythms and/or atrio-ventricular conduction delays can be incorporated to avoid false positive cut-outs of ablation energy. Yet a further advantage is to allow for operators of various experience level, ranging from novice operators for whom cut-out parameters will be stringently set, to experienced operators for whom some cut-out parameters may be relaxed.

25

In one embodiment, the monitoring system comprises a screen displaying pulse markers in real time, the pulse markers corresponding to a onset or a peak of each electrophysiology recording acquired by the IEGM electrode and the ECM electrodes. One advantage is to allow an operator a visual check of monitoring system integrity.

30

In one embodiment, the monitoring system comprises a pacing unit that provides current pulses at programmed rates and patterns for stimulating the heart ;

In one embodiment, the monitoring system comprises a junction box that provides a common interface gathering the connections of ECG electrodes and IEGM electrode and delivers the acquired signals to the calculator.

5 Another object of the invention concerns a monitoring method achieving automatically the functions supported by the calculator of the system of the invention.

10 In one embodiment, the monitoring method increases safety of cardiac ablation when applied in close proximity to the AV node, by automated cut-out of ablation energy in response to cardiac rhythm related markers of risk for heart block, comprises the steps of:

- 15 ▪ Monitoring in real time or near real time, one or more atrial intracardiac signals and the body surface electrocardiogram signals of a patient, wherein said monitoring comprises the steps of signal amplification, filtering, analog to digital conversion, and computational analysis of signal timing and morphology.
- 20 ▪ Detecting in real time or near real time, an event that may herald imminent heart block, wherein said event is a condition tailored to the location of ablation and arrhythmia type being ablated, and said event may be one of atrio-ventricular dissociation, ventriculo-atrial dissociation, fast junctional rhythm, any junctional rhythm, atrio-ventricular interval prolongation, PR interval prolongation, consecutive atrial depolarisations, consecutive ventricular depolarisations, or indeterminate atrio-ventricular nodal conduction;
- 25 ▪ Automatically stopping energy delivery of said ablation based upon said markers of risk.

30 In one embodiment, computational analysis of signal timing comprises methods of determining onset or peak of electrogram depolarisation deflection, and electrocardiogram QRS complex.

35 In one embodiment, computational analysis of signal morphology comprises methods of determining the extent of electrocardiogram QRS complex match to a template QRS complex signal, and methods of determining electrocardiogram QRS complex width and sequence of deflections.

In one embodiment, computational analysis can be written using development tools and graphical programming applications or languages such as LAB WINDOWS®/CVI, LABVIEW®, MICROSOFT VISUAL BASIC®, MICROSOFT VISUAL C++®, or functionally equivalent software

5 In one embodiment, stopping energy delivery of said ablation is via a method that mimics the manual release of an ablation generator foot pedal and therefore does not require manual resetting of equipment, wherein such a method may be placement of a switch mechanism between a foot pedal and the ablation generator unit, or, integration of a switch mechanism into the circuitry of an ablation
10 generator.

In one embodiment, the results of computational analysis of signal timing are displayed in real-time or near real-time, in the form of pulse markers, on an electrophysiology recording system monitor, to allow a visual check of said computational analysis integrity.

15 In one embodiment, the method may be incorporated into or combined with or embedded with an ablation generator.

In one embodiment, the method may be incorporated into or combined with or embedded with an electrophysiology recording system.
20

In one embodiment, the method may be incorporated into or combined with or embedded with an electrophysiology mapping system.

In one embodiment, ablation is performed for atrio-ventricular nodal re-entrant tachycardia (AVNRT), or for mid-septal or antero-septal accessory pathway ablation.
25

In one embodiment, ablation is one of: radiofrequency (RF) ablation, cryoablation, high intensity focused ultrasound (HIFU) ablation, or microwave ablation.

30 In one embodiment, atrio-ventricular dissociation refers to an atrial depolarisation without a corresponding ventricular depolarisation within a given time limit.

In one embodiment, ventriculo-atrial dissociation refers to a ventricular depolarisation without a corresponding atrial depolarisation within a given time limit.

35

In one embodiment, a user interface is provided to the operator or ancillary staff in order to alter settings and parameters which trigger said automatic stopping of energy delivery, based on the type and location of ablation being performed, the nature of the patient's physiology, and operator personal preference.

5

BRIEF DESCRIPTION OF THE DRAWINGS

For the purpose of illustrating the invention, there are shown in accompanying drawings, forms which are presently preferred, it being understood that the invention is not intended to be limited to the precise arrangement and instrumentalities shown.

10

15

20

25

30

- FIG. 1 depicts the general set-up and equipment of the state of the art currently used in a standard cardiac ablation procedure. The "EP" acronym refers to "electrophysiology".
- FIG. 2A shows the cardiac anatomy with cut-away surface (shaded in grey) to show the right atrium and right ventricle, being separated by the tricuspid valve. The AV node consists of fast and slow pathway inputs, the compact AV node, and bundle of His. The "IVC" acronym refers to the "Inferior Vena Cava, and the "SVC" acronym refers to the "Superior Vena Cava".
- FIG. 2B depicts activation of the heart during normal sinus rhythm, with spread of electrical activation commencing in the sinoatrial node and spreading through the right atrium, and to the right ventricle via the AV node.
- FIG. 2C shows a typical body surface electrocardiogram recording, known as "ECG". Atrial depolarisation gives rise to a small deflection known as the 'p' wave 44, whereas ventricular depolarisation gives rise to a much larger amplitude deflection known as the QRS complex 45. The interval between onset of the p wave and QRS complex is known as the PR interval 46, and in the normal heart is at least 120ms in duration.
- FIG. 3A depicts activation sequence in the heart when a septal accessory pathway is present that allows antegrade conduction from the atria to the ventricles (so-called Wolff-Parkinson-White syndrome). The electrical impulse is propagated from the atria to the ventricles over both the AV node, and the accessory pathway.

- FIG. 3B shows a typical body surface electrocardiogram (ECG) recording from a patient with Wolff-Parkinson-White syndrome, in which the PR interval is shorter (< 120 ms) than normal, and the onset of the QRS complex has a slurred morphology.
- 5 ▪ FIG. 4A depicts electrical impulse propagation within the heart during atrio-ventricular nodal re-entrant tachycardia (AVNRT), with re-entrant (circular) activation of the AV node leading to near simultaneous atrial and ventricular activation.
- 10 ▪ FIG. 4B depicts electrical impulse propagation within the heart during atrio-ventricular reciprocating tachycardia (AVRT) utilising a septal accessory pathway, with re-entrant (circular) activation of the AV node and accessory pathway sequentially.
- 15 ▪ FIG. 5 shows placement of an ablation catheter over the slow input to the AV node for purposes of slow pathway ablation or modification in the treatment of AVNRT. A diagnostic (recording) catheter is placed in the high right atrium for purposes of recording an atrial intracardiac electrogram.
- 20 ▪ FIG. 6A depicts sinus rhythm during ablation, with electrical impulse propagation proceeding from the sinoatrial node, through the atrioventricular (AV) node, to the ventricles. In the case of normal AV node function, atrial intracardiac electrogram (A IEGM) should proceed the QRS complex of the surface electrocardiogram (ECG) by no more than approximately 250ms (nominal value).
- 25 ▪ FIG. 6B depicts junctional rhythm during ablation. Electrical activation of the heart originates from the compact AV node or His bundle and spreads outward to both atrial and ventricles simultaneously. In the case of normal AV node function, atrial intracardiac electrogram (A IEGM) should lag the QRS complex of the surface electrocardiogram (ECG) by no more than 130ms (nominal value).
- 30 ▪ FIG. 6C depicts ventricular ectopic beat (rhythm) during ablation. Electrical activation of the heart originates from the ventricular tissue and spreads retrogradely through the AV node to the atria. In the case of normal AV node function, atrial intracardiac electrogram (A IEGM) should lag the QRS complex of the surface electrocardiogram (ECG) by no more than 250ms (nominal value).
- 35

- FIG. 7 shows the current invention and its relationship to existing equipment for the purposes of cardiac ablation. The current invention includes a Rhythm Monitoring device 25, a user interface 27, and a switch unit 26.
- 5 ▪ FIG. 8 shows data on the percentage of ventricular ectopic beats (rhythm), as a function of whether an intracardiac catheter is placed within the right ventricle or not. In 215 ablation episodes from 20 patients in which a ventricular catheter was placed, 36% of ablations had no ventricular ectopic rhythm, whereas 64% had ventricular ectopic rhythm
10 varying between 0.1 to 20% of all beats. In 196 ablation episodes from 20 patients in which a ventricular catheter was not used, ventricular ectopic rhythm was much less, with 95% of ablations having no ventricular ectopic rhythm.
- 15 ▪ FIG. 9A depicts the phenomenon of indeterminate AV nodal function, due to a sinus beat being closely followed by a ventricular ectopic beat, such that the integrity of AV nodal function cannot be determined.
- 20 ▪ FIG. 9B gives an example of a sinus beat being closely followed by a ventricular ectopic beat, which leads to inability to assess AV nodal function (conduction) until the following beat.
- 25 ▪ FIG. 10A shows a junctional beat with slowed/damaged AV nodal conduction of the electrical impulse to the atria, giving rise to the same intracardiac electrograms as would a ventricular ectopic beat with normal AV nodal conduction (FIG. 10B). The QRS complex on the surface ECG is of narrow morphology and this allows differentiation of the two rhythms.
- 30 ▪ FIG. 10B shows a ventricular ectopic beat with normal AV nodal conduction of the electrical impulse to the atria, giving rise to the same intracardiac electrograms as would a junctional beat with slowed/damaged AV nodal conduction (FIG. 10A). The QRS complex on the surface ECG is of wide morphology and this allows differentiation of the two rhythms.
- 35 ▪ FIG. 11A shows a sinus beat with normal AV nodal conduction of the electrical impulse to the ventricles, giving rise to the same intracardiac electrograms as would a sinus beat followed by a ventricular ectopic beat (FIG. 11B) in which AV nodal conduction is indeterminate. The QRS

complex on the surface ECG is of narrow morphology and this allows differentiation of the two rhythms.

- 5 ▪ FIG. 11B shows a sinus beat followed by a ventricular ectopic beat (in which AV nodal conduction is indeterminate), giving rise to the same intracardiac electrograms as would a sinus beat with normal AV nodal conduction. The QRS complex on the surface ECG is of wide morphology and this allows differentiation of the two rhythms.
- 10 ▪ FIG. 12A shows an example of atrio-ventricular dissociation, in this case an atrial (A) depolarisation which is not followed by a ventricular (V) depolarisation within 250ms (A no V), causing cut-out of ablation energy. A IEGM = atrial intra-cardiac electrogram; ECG = electrocardiogram
- 15 ▪ FIG. 12B shows an example of atrio-ventricular dissociation, in this case a ventricular (V) depolarisation which is not followed by an atrial (A) depolarisation within 250ms (V no A), causing cut-out of ablation energy. A IEGM = atrial intra-cardiac electrogram; ECG = electrocardiogram.
- 20 ▪ FIG. 12C shows an example of Fast Junctional Rhythm (junctional rhythm exceeding 130bpm), causing cut-out of ablation energy. Cut-out occurs after the 2nd consecutively detected fast junctional beat. A IEGM = atrial intra-cardiac electrogram; ECG = electrocardiogram.
- 25 ▪ FIG. 12D shows an example of atrio-ventricular dissociation, as evidenced by consecutive atrial (A) depolarisations, without an intervening ventricular depolarisation (consecutive As), causing cut-out of ablation energy. A IEGM = atrial intra-cardiac electrogram; ECG = electrocardiogram.
- 30 ▪ FIG. 12E shows an example of PR interval prolongation exceeding 1.15 x baseline PR interval for 3 consecutive beats, causing cut-out of ablation energy. A IEGM = atrial intra-cardiac electrogram; ECG = electrocardiogram.
- 35 ▪ FIG. 13 shows a schematic of the Rhythm Monitoring Device's treatment of atrial intracardiac electrogram (IEGM) and ventricular electrocardiogram (ECG) signals, in order to discern cardiac rhythm.
- FIG. 14 shows real-time or near real-time display of pulsed signals output from the Rhythm Monitoring Device to the EP recording system, for the purposes of a visual check that the Rhythm Monitoring Device is accurately detecting atrial and ventricular depolarisations.

- FIG. 15 depicts an example of circuitry associated with the Switch Unit, and connections to the Foot Pedal, Ablation Generator, and Rhythm Monitoring Device.
- 5 ▪ FIG. 16 shows an example of an appropriate user interface on which parameters triggering cut-out of ablation can be set.
- FIG. 17 shows an example of the Rhythm Monitoring Device's computational algorithm employed during ablation of AVNRT.
- FIG. 18 shows an example of the Rhythm Monitoring Device's computational algorithm employed during ablation of AVRT
- 10 ▪ FIG. 19 shows an example of the Rhythm Monitoring Device's computational algorithm employed to allow continued ablation despite a single beat of dissociated atrio-ventricular activity.
- FIG. 20A depicts incorporation of the Rhythm Monitoring Device into an existing Ablation Generator
- 15 ▪ FIG. 20B depicts incorporation of the Rhythm Monitoring Device into an existing electrophysiology (EP) Recording System
- FIG. 21A shows the method for determining human operator reaction time to a condition requiring cessation of ablation energy. In this case, ablation was stopped because of an "A no V" condition. The time of manually applied cut-out of ablation (thin vertical dashed line) is derived from the end of ablation-induced artifact (noise) recorded on the distal ablation electrode (ABLd). The time of automated cut-out as determined by the Rhythm Monitoring Device (thick vertical solid line) is subtracted from the manual cut-out time to give 892ms of reaction time. A IEGM =
20 atrial intra-cardiac electrogram; ECG V1 = lead V1 of the electrocardiogram; RhyA = the automated rhythm analysis (refer to the key shown in figure).
- 25 ▪ FIG. 21B shows the method for determining human operator reaction time to a condition requiring cessation of ablation energy. In this case, ablation was stopped because of "Fast Junctional" rhythm. The time of manually applied cut-out of ablation (thin vertical dashed line) is derived from the end of ablation-induced artifact (noise) recorded on the distal ablation electrode (ABLd). The time of automated cut-out as determined by the Rhythm Monitoring Device (thick vertical solid line) is subtracted from the manual cut-out time to give 2210ms of reaction time. A IEGM =
30
- 35

atrial intra-cardiac electrogram; ECG V1 = lead V1 of the electrocardiogram; RhyA = the automated rhythm analysis (refer to the key shown in figure).

- 5 ▪ FIG. 22 summarises the results of testing of a prototype of the invention.
- FIG. 23A shows an example of under-detection of atrial depolarisation (dashed oval) due to sudden diminution of signal amplitude. A IEGM = atrial intra-cardiac electrogram; ECG V2 = lead V2 of the electrocardiogram.
- 10 ▪ FIG. 23B shows an example where accurate detection of atrial depolarisations (dashed ovals) is preserved despite sudden diminution of signal amplitude to a slightly lesser degree than FIG. 22A. A IEGM = atrial intra-cardiac electrogram; ECG V2 = lead V2 of the electrocardiogram.
- 15 ▪ FIG. 23C shows an example where accurate detection of atrial depolarisations is preserved despite the presence of ventricular far-field signal (dashed ovals) on the atrial recording. The far field signal is appropriately not detected as atrial depolarisations. A IEGM = atrial intra-cardiac electrogram; ECG V2 = lead V2 of the electrocardiogram.
- 20 ▪ FIG. 24 shows the distribution of reaction time improvements given by the invention during prototype testing. Reaction time improvement is the time difference between human (manual) cut-out of ablation, and automatic cut-out by the Rhythm Monitoring Device. Cut-outs were for both atrio-ventricular and ventriculo-atrial dissociation (grouped together as A-V dissociation; solid bars), and for Fast Junctional rhythm (grey bars).
- 25

DETAILED DESCRIPTION

30 The following description is of the best mode presently contemplated for carrying out the invention. This description is not to be taken in a limiting sense, but is made merely for the purpose of describing the general principles of the invention. The scope of the invention should be made with reference to the claims.

EQUIPMENT FOR ELECTROPHYSIOLOGY STUDY AND ABLATION

For the purposes of explaining the current applicant's system, a common setup of equipment for the purposes of electrophysiologic (EP) study and ablation is first reviewed in more detail (FIG. 1).

Catheters are placed inside of the heart 11 usually via the femoral veins.
5 A minimum of two catheters is generally required to diagnose arrhythmia mechanism.

Diagnostic Cathether

One or more diagnostic catheters 13 are placed to allow both recording
10 of intracardiac electrograms, and programmed stimulation of the heart. Diagnostic catheters are typically connected to junction box 15 which feeds signals to an electrophysiology (EP) recording system 17.

ECG (body surface)

15 Voltage signals are also recorded from the body surface in the form of an electrocardiogram (ECG) 14, typically of 12 leads in standard locations (only 3 leads shown in FIG. 1).

Ablation Catheter

20 An ablation catheter 12 is used to direct ablation energy to specific areas of the heart found to be critical to either arrhythmia triggering or maintenance.

Ablation generator

The ablation generator unit 16, which connects to the ablation catheter,
25 provides the energy for ablation. It is one example of the power source of the ablation catheter 12. This may be in the form of radiofrequency (RF) current, cryo (freezing), high intensity focused ultrasound (HIFU), microwave energy, or other forms; RF being the most commonly employed presently.

Examples of current commercially available RF generators include :

- 30
- EP-Shuttle and SmartAblate™ (Stockert),
 - Ampere™ (St. Jude Medical),
 - Maestro 3000™ (Boston Scientific).

The ablation generator can usually be controlled (that is to say, ablation energy commenced and terminated) by a switch on the unit, or, by a foot pedal 20.

EP recording system

The EP recording system 17 is used to condition (amplify and filter) the intracardiac and ECG signals, convert these to digital form, and display these to the operator and technical staff on a PC monitor(s) 21, for the purposes of performing the procedure.

Examples of current commercially available EP recording systems include :

- BARD® LABSYSTEM™ PRO EP (Boston Scientific),
- Sensis™ (Siemens),
- EP-WorkMate™ (St. Jude Medical),
- CardioLab™ (GE Healthcare).

EP mapping system

The EP mapping system 19 may also be used to condition and display intracardiac and ECG signals, but also performs the function of displaying three-dimensional catheter position relative to a reconstructed cardiac geometry, which augments standard x-ray based fluoroscopic methods of catheter navigation.

Examples of current commercially available EP mapping systems include :

- Carto™ (Biosense Webster),
- Rhythmia™ (Boston Scientific),
- EnSite™ and Velocity™ Cardiac Mapping Systems (St. Jude Medical).

Pacing/ stimulation Unit

The Pacing/Stimulation Unit 18 provides current pulses to stimulate the heart at programmed rates and patterns, to assist in the diagnosis of arrhythmia mechanism.

Examples of current commercially available Pacing/Stimulation Units include:

- StimCor™ (Boston Scientific).
- Micropace (Micropace EP Inc).

Variations of the above connection configurations can be implemented according to various embodiments of the invention. The figure 1 represents one

configuration of connectivity of items that may be implemented according to the invention.

For example, it is possible on some commercially available systems to plug all catheters directly into the EP mapping system bypassing the junction box, with the EP recording system, Pacing/Stimulation Unit, and ablation generator being slaved off the EP mapping system. Many other configurations are feasible, while still maintaining overall connectivity of the equipment items shown in FIG. 1. Moreover, it is possible to integrate items in FIG. 1 into combined systems, for example, the Pacing/Stimulation Unit may be integrated into an EP recording system.

PROCEEDING TO AN ABLATION CLOSE TO THE AV NODE

Using the above described set-up, ablation of cardiac tissue lying close to, or adjacent to, the AV node, is possible, but carries a risk of damage to the compact AV node 31 or His bundle 30, leading to permanent AV block, wherein there is failure of propagation of the electrical impulse from the atria to the ventricles, and co-ordination between atrial and ventricular contraction is lost.

Because permanent heart block generally requires implantation of a permanent pacemaker, this is considered a major complication, and may reduce the quality and quantity of the patient's life.

The invention allows prediction of occurrences of AV block by analysing recorded signals and ablation catheter tip location.

Atrio-ventricular and ventriculo-atrial dissociation

AV block may be heralded by changes seen in the cardiac signals displayed on the EP recording system monitor 21. Specifically, the occurrence of transient dissociation of atrial and ventricular depolarisation may indicate impending heart block, and can be detected by analysis of the intracardiac atrial electrogram IEGM together with the body surface ECG, wherein each atrial or ventricular signal is not associated with a counterpart ventricular or atrial signal within a defined time period as illustrated in figures 12A and 12B.

Fast junctional rhythm, consecutive atrial or ventricular depolarisations, PR interval prolongation

In addition, fast junctional rhythm, as illustrated in figure 12C, consecutive atrial or ventricular depolarisations, as illustrated in figure 12D, or PR

interval prolongation, as illustrated in figure 12D, may indicate impending heart block.

Visual checking

5 When ablating close to the AV node, the operator and technical assistant(s) closely watch the signals on the monitor 21, and if any of the aforementioned conditions arise, ablation is terminated manually by pressing on a switch on the ablation unit, or, alternatively by releasing the foot pedal 20. If a foot pedal is used, this may be controlled directly by the operator, or, by a technician.

10 Because the time of human reaction to said conditions indicating risk of heart block is typically 1-4 seconds, and because there is a competing need for the operator to visually monitor the x-ray projection of the catheter, watching for movement away from the desired location of ablation, this process of human visual feedback 22 suffers from significant limitations. Any delay in reaction to the markers of risk of

15 heart block may cause irreversible damage to tissue and render heart block more likely to occur.

AUTOMATED CUT-OFF OF ABLATION

20 The invention comprises a system for supplementing human visual feedback 22 with automated assessment of cardiac rhythm, and automatic cut-out of ablation energy when the conditions indicating risk of heart block occur. The system disclosed is able to react to risk conditions faster than the human reaction, and in a more consistent manner.

25 In the method and system of this invention, cardiac rhythm analysis is performed with use of one or more atrial intracardiac electrogram signal IEGM, and the use of one or more body surface ECG signal.

Junction box

30 In preferred embodiments of the invention, these signals are split from the Junction Box 15 shown in figure 7 into the Rhythm Monitoring Device 25, or alternatively are derived from connection between the Rhythm Monitoring Device 25 and the EP recording system 17 or EP mapping system 19. Other specific connectivities are possible, for example direct catheter connection to the Rhythm Monitoring Device 25 bypassing the Junction Box 15, and remain within the scope of

35 the invention. The Rhythm Monitoring Device conditions the signals by amplification,

filtering, and analog-to-digital conversion. Alternatively, partially or fully pre-conditioned signals may be obtained from the EP recording system 17 or EP mapping system 19 and input to the Rhythm Monitoring Device 25.

5 **Timing and Morphology analysis**

The Rhythm Monitoring Device performs real-time or near real-time computational analysis of signals in order to derive timing of atrial and ventricular depolarisation, in the former based on peak or onset of atrial electrogram deflection, and in the latter based on peak or onset of ECG QRS complex deflection as shown
10 in figure 13.

Analysis of ECG QRS complex morphology is also performed to determine the nature of ventricular depolarisation, specifically, whether ventricular depolarisation arises from a junctional or ectopic beat, and whether ventricular depolarisation is exclusively via the AV node, or pre-excited from an accessory
15 pathway.

Where there is conflict in the assessment of timing and morphology across leads of the ECG, a voting algorithm may be employed to determine a majority decision as shown in figure 13.

20 **Timing analysis**

In a first embodiment, timing analysis may be performed with a threshold of signal magnitude and/or derivative, with the threshold fixed to a percentage of the preceding n^{th} number of signal peaks.

In a second embodiment, an alternate method may employ signal phase analysis (for example utilising a Hilbert's transform) with signal timing defined by a
25 zero crossing of phase.

Morphology analysis

In one embodiment, morphology analysis may be performed using auto-correlation based matching of the real-time QRS complex with a template QRS, the
30 template QRS having been automatically stored from the same patient at procedure onset, prior to ablation.

In a second embodiment, additional morphology analysis methods may employ cross-correlation of QRS signal with generic (derived from other individuals)
35 template signals from normal and ectopic beats, as well as from other conditions

causing change to QRS morphology such as impaired conduction in the Purkinje fibre branches 41 (so-called "bundle branch block").

Yet further methods may employ determination of QRS complex width, and QRS signal positive/negative deflection sequence.

5

In a preferred embodiment of the invention, the Rhythm Monitoring Device 25 outputs pulsed (square wave) signals to the EP recording system 17, via connection to the Junction Box 15 (FIG. 7), to mark the timings of atrial and ventricular depolarisations, as determined by the Rhythm Monitoring Device. As shown in FIG. 14, these pulses are thus able to be displayed on the PC 21 in real or near-real time, in order to provide a visual check that the Rhythm Monitoring Device is functioning correctly.

Examples

An example of a suitable data acquisition (DAQ) card capable of fulfilling the requirements of the Rhythm Monitoring Device 25 is the sbRIO-9639 (National Instruments, NI). Similar DAQ cards from the same or other manufacturers could be employed. Processing of atrial IEGM and ECG signals can be performed in real-time on the microprocessor or the FPGA of the NI sbRIO-9639. Testing of the applicant's prototype has been performed using the aforementioned NI sbRIO-9639, with software written in LABVIEW®. Use of other programming languages or methods of data acquisition are considered within the scope of the invention.

SWITCH MECHANISM

In the present invention, automated cut-out of ablation mimics the manual release of an ablation generator foot pedal, when conditions indicating risk of AV block are detected.

A preferred embodiment of the invention is a switch mechanism 26 that is under control of the Rhythm Monitoring Device 25, and placed between the ablation generator 16 and foot pedal 20 as shown in figure 15. The switch mechanism may comprise a relay switch 80 that is opened by the Rhythm Monitoring Device to automatically cut-out ablation. In addition, the Rhythm Monitoring Device is able to monitor the state of the foot pedal, being either depressed or raised.

At onset of the procedure, the switch 80 is in a closed state due to application of a DC voltage from the Rhythm Monitoring Device.

Ablation is commenced in the usual fashion by the operator or technician depressing the foot pedal. If risk conditions are detected, the Rhythm Monitoring
5 Device is able to cut-out ablation by opening the switch 80, which has the effect of mimicking release of the foot pedal.

When the system of the invention does not comprise a foot pedal, risk conditions automatically detected by the Rhythm Monitoring Device will trigger cut-out of ablation by direct control of the ablation generator.

10 Audible feedback may be given to the operator explaining the cut-out, and the operator acknowledges said cut-out by releasing the foot pedal, which informs the Rhythm Monitoring Device to release the cut-out condition by setting the switch 80 back to the closed position. In this way, ablation is able to be recommenced simply by the operator re-depressing the pedal 20, without any need
15 to manually reset equipment. The operator holds responsibility for the decision to re-apply ablation, having taken into consideration the safety of continued ablation, and if necessary, repositioned the catheter to a safer location. The Rhythm Monitoring Device is able to give feed-back on the reaction time of the operator (releasing the foot pedal) to the cut-out condition.

20

USER INTERFACE

The Rhythm Monitoring Device is connected to a user interface 27, as represented in figure 7, which allows operator-led adjustment of the parameters causing automated cut-out.

25 An example of a suitable user interface is shown in figure 16.

For example, whether cut-out occurs for atrio-ventricular dissociation, as shown in figures 12A, 12B, and 12D, can be selected by the "Blocked Beats Cut-off" knob as shown in figure 16.

Whether cut-out occurs for PR interval prolongation as shown in figure
30 12E and at what degree of prolongation, 115% or 125%, can be selected using the "PR Prolongation" knob as shown in figure 16.

The heart rate causing cut-out, if any, for Fast Junctional Rhythm as shown in figure 12C can be selected using the "Junctional Rate Cut-off" knob as shown in figure 16.

Other more 'advanced settings' can also be adjusted, for example the maximum delay allowed between atrial and ventricular depolarisations, called the A-V interval, nominally set to 250ms.

5 The specific cut-out conditions set will depend on the type of ablation being performed, the particular patient's physiology, and individual operator preference. For example, a patient with pre-existing slow / diseased conduction of electrical impulse from atria to the ventricles, so-called first degree heart block, may need allowance of longer A-V intervals compared to a patient with normal baseline atrio-ventricular conduction. Similarly, an operator with many years of experience in
10 ablation may chose less conservative settings on the device compared to an operator with less experience.

DEVICE ALGORITHMS

There may be differences in the computational algorithms employed by
15 the Rhythm Monitoring Device 25 for different arrhythmia, to allow for safe ablation in each condition.

In the case of AVNRT, a marker of successful ablation is the development of slow junctional rhythm, as this indicates some degree of heating to the AV node, without risk of damage to the compact node or His bundle
20 components. However, fast junctional rhythm (exceeding ~130bpm) is to be avoided.

In contrast, in the case of AVRT or ablation of an accessory pathway close to the AV node, no junctional rhythm is desirable whilst the ventricle is pre-excited via the accessory pathway. This is because integrity of AV nodal conduction
25 is uncertain during junctional rhythm with pre-excitation. Algorithms that reflect these differences in strategy are employed on the Rhythm Monitoring Device.

The algorithm currently preferred for use in slow pathway ablation for AVNRT is shown in figure 17.

30 The algorithm currently used for accessory pathway ablation during pre-excitation of the ventricle is shown in figure 18.

The main differences between algorithms relates to the response to junctional rhythm and PR interval change.

In the case of accessory pathway ablation as shown in figure 18 :

- cut-out may occur for the presence of any junctional beats while the heart remains in pre-excited rhythm (because integrity of AV nodal conduction is unable to be assessed while in junctional rhythm), and;
- cut-out for PR prolongation is not enabled until pre-excitation is lost by successful ablation of the accessory pathway, to avoid unnecessary cut-out at the instant when conduction over the accessory pathway is extinguished.

An additional algorithm may also be employed that allows continued ablation despite the occurrence of a single dissociated beat, as for instance a single A without V, or a single V without A, corresponding to the cases illustrated in figures 12A and 12B, yet cuts-out ablation in response to multiple dissociated beats.

This algorithm is shown in figure 19, and is intended as an “advanced” mode of ablation, to be used for ablation in patients with a high number of non-conducted ectopic beats, or for operators of sufficient skill that cut-out for a single non-conducted beat is deemed unnecessary. The algorithm shown in figure 19 is based around requirements of two “strikes” for cut-out to occur :

- the first strike being for a non-conducted (dissociated) beat, and ;
- the second strike being for a second non-conducted beat, or, PR interval prolongation.

The first strike is able to be redeemed by the subsequent occurrence of two consecutive sinus beats of normal PR interval.

The algorithms displayed in figures 17, 18, and 19 are included for indicative purposes only and are not to be taken in a limiting sense with respect to the scope of the invention; many variations on these algorithms are possible. The algorithms given are particular to ablation in sinus rhythm. Further algorithms may also be employed for ablation during tachy-arrhythmia.

INTEGRATION TO EXISTING EQUIPMENT

In one embodiment of the invention, the Rhythm Monitoring Device 25 may be a stand-alone piece of equipment. In other embodiments, it could be variously incorporated into an ablation generator 16, as shown in figure 20A, an EP recording system 17, as shown in figure 20B, or an EP mapping system 19, (not shown).

In the case of incorporation into an ablation generator 16, the switch mechanism 26 would not need to be stand-alone, but rather could be incorporated into the circuitry of the ablation generator 16.

Equally, in the case of incorporation into other systems or devices, there may not be a requirement for a stand-alone user interface monitor 27. Rather, the functionality of the user interface 27 could be added to existing user interfaces associated with the EP recording system, the EP mapping system, or the ablation generator.

10 **ABLATION ENERGY TYPES**

The present invention can be equally used with any form of ablation energy, including but not limited to radiofrequency (RF) ablation, cryo-ablation, high intensity focused ultrasound (HIFU) ablation, and microwave ablation.

15 **PROTOTYPE TESTING**

A prototype has been developed using the preferred strategy outlined above, and tested on signals acquired from 19 patients (208 ablation episodes) during the course of AVNRT ablation via slow pathway ablation or modification. Intracardiac atrial electrograms recorded from either the high right atrium (HRA), or the coronary sinus (CS), were input to the Rhythm Monitoring Device 25, along with three leads from the standard ECG (leads I, V1, V4).

Activity in the atrium and ventricles was assessed using EGM and ECG peak detection, by way of voltage magnitude and 1st derivative adaptive thresholding as described above, incorporating a threshold decay function to allow resumption of detection after lead or catheter disconnection. Subsequent rhythm determination was by rule-based analysis, incorporating QRS complex morphology matching to distinguish junctional from ventricular ectopic beats. Post detection blanking intervals of 180ms and 200ms were applied to the atrial and ventricular signals respectively, to prevent double counting of activations. No blanking period was applied to the atrial signal following ventricular activation, to ensure that junctional rhythm could be adequately tracked. Cardiac rhythm was arbitrarily classified as 'sinus' where there was sequential atrial and ventricular depolarisation separated by ≥ 90 ms, 'ventricular' where there was sequential ventricular and atrial depolarisation separated by ≥ 90 ms, and 'junctional' where atrial and ventricular depolarisations were within 90ms of each other.

The algorithm shown in figure 17 was programmed in the graphical language LabView™ 12.0 (National Instruments; ni.com), and implemented on a National Instruments Single-board RIO microprocessor. The nominal settings that would trigger cut-out were :

- 5 ▪ A-V conduction time exceeding 250 ms, and V-A conduction time exceeding 250ms for a ventricular ectopic beat, or 130ms for a junctional beat,
- junctional rhythm > 130 bpm, sustained for 3 consecutive beats,
- PR interval prolongation > 1.15x pre-ablation baseline, sustained for
10 3 beats, and
- two consecutive atrial or ventricular depolarisations.

The timing of individual atrial and ventricular depolarisations (total of 30,754 EGM and ECG complexes), and the cardiac rhythm at each beat, was
15 assessed manually by an experienced clinician, to act as a gold-standard comparison to the algorithm. Database signals were fed to the device with real-time sampling to simulate a 'live' electrophysiologic study. The device signal interpretations and cut-out conditions were stored and later analysed.

Cut-out conditions of the device were compared to the clinical cut-outs,
20 which were with use of a foot pedal controlled RF generator (Stockert™, Biosense Webster). The foot pedal was controlled by a Physiologist instructed to terminate ablation in response to :

- a single event of failed or delayed atrio-ventricular conduction, or a single event of failed or delayed ventriculo-atrial conduction, or ;
- 25 ▪ junctional rhythm judged to be visually faster than ~130 bpm.

In rare cases, the Physiologist was instructed to continue ablation despite fast junctional rhythm, or the onset of AVNRT. The time of pedal release was able to be determined from the signal recorded at the distal bipole of the ablation catheter, which displayed an electrical artifact at the time of cut-out. This
30 allowed comparison of response speeds of human and device in all cases, with examples of A no V cut-out and Fast Junctional cut-out given in Figures 21A and 21B respectively.

The invention prototype performance is summarised in figure 22. The device assigned accurate times to atrial depolarisations in all but 10 / 15,317 cases,
35 and to ventricular depolarisations in all but 2 / 15,437 cases. The inaccurate timings

occurred in 3 patients. Of the 10 inaccurate EGM timings, 8 were due to under-detection of small amplitude EGM signals during periods of sudden diminution of signal amplitude (example given in figure 23A, with an example of successful detection in similar circumstances given in figure 23B). In two recordings, inaccurate assignment was due to sudden large amplitude fluctuation in EGM baseline, caused by patient movement. There were no cases of over-detection of ventricular far-field signal on the atrial EGM (example in figure 23C). The two cases of inaccurate timing assignment in ECG peak were due to the well described phenomenon of T wave over-sensing.

Of the 208 ablation episodes, 114 were terminated prematurely by the device, 53 for detection of fast junctional rhythm, 32 in response to failed atrio-ventricular conduction, and 29 in response to failed ventriculo-atrial conduction as represented in figure 22. There were three false positive terminations which occurred immediately following inaccurate assignment of atrial or ventricular depolarisation time during ablation (specificity of device termination = 96%, PPV 97%). There were no instances where the device missed a termination condition (sensitivity = 100%, NPV 100%).

For the 111 appropriate device cut-outs, response speed of the device was compared to the speed of manual RF foot-pedal release. The device performed a median of 1,238 ms faster (range 14 to 132,919) than the human response. In no cases did the device lag human response. The distribution of response time differences (manual cut-out time minus device cut-out time) is shown in figure 24, for episodes where this time difference was < 4s (n= 79; median = 971 ms). Cut-out time differences longer than 4s were assumed to be due to intentional continuation of ablation.

SUMMARY

We have described an invention and system for the automated detection of cardiac rhythms that indicate risk of AV block during ablation near to the AV node, with automatic cut-out of ablation in response to said rhythms. The present invention may be embodied in other specific forms without departing from the spirit or essential attributes thereof. It is therefore desired that the present embodiment be considered in all aspects as illustrative and not restrictive, reference being made to the appended claims rather than to the foregoing description to indicate the scope of the invention.

CLAIMS

- 5 1. A monitoring system for ablation of cardiac tissue in close proximity to the AV node of the heart, characterized in that it comprises :
- At least one ablation catheter (12) powered by a power source (16);
 - At least one electrode (13) acquiring an intracardiac electrogram (IEGM), called IEGM electrode ;
 - 10 ▪ At least one set of electrodes (14) acquiring a surface electrocardiogram (ECG), called ECG electrodes ;
 - At least one calculator (15, 16, 17, 18, 19, 25, 27) and a memory for measuring in real time signals acquired from the IEGM electrode and the ECG electrodes in order to automatically generate at least one stop
 - 15 command (Sc) when a cardiac risk indicator is triggered by combining signals measured from the IEGM electrode and the ECG electrodes, the generated stop command (Sc) controlling the power source (16) of the ablation catheter (12) in order to automatically cut-out ablation catheter energy.
 - 20
2. A monitoring system according to claim 1, characterized in that at least one calculator (15, 16, 17, 18, 19, 25, 27) and a memory is used for measuring in real time a displacement of the ablation catheter (12) in order to automatically generate a stop command (Sc) when detecting a
- 25 displacement of the ablation catheter (12) exceeding a predefined limit.
3. A monitoring system according to claim 1, characterized in that the calculator comprises a signal timing function in order to derive timing of atrial and ventricular depolarisation, the atrial depolarisation being
- 30 measured by the IEGM electrode (13), and the ventricular depolarisation being measured by the ECG electrodes (14).
4. A monitoring system according to claim 3, characterized in that the signal timing function compares the delay between atrial and ventricular depolarisations to a minimum predefined threshold of time and/or a
- 35 maximum predefined threshold of time, the signal timing function being


configured for generating a cardiac risk indicator when said measured delay exceeds at least one of the predefined thresholds.

- 5 5. A monitoring system according to claim 3, characterized in that the signal timing function comprises a first voting function allowing the determination of timing when more than one IEGM electrode or more than one ECG electrode is used.
- 10 6. A monitoring system according to claim 1, characterized in that the calculator comprises a signal morphology function in order to determine the nature of the ventricular depolarisation by analysing the QRS complex acquired by the ECG electrodes (14).
- 15 7. A monitoring system according to claim 6, characterized in that the signal morphology function compares an acquired QRS complex signal with a template QRS stored in a memory.
- 20 8. A monitoring system according to claim 6, characterized in that the signal morphology function is configured to compute the QRS complex width and/or to determinate deflection sequence of the QRS complex signal.
- 25 9. A monitoring system according to claim 6, characterized in that the signal morphology function comprises a second voting function allowing the determination of morphology when more than one ECG electrode is used.
- 30 10. A monitoring system according to claims 3 and 6, characterized in that the calculator is configured to determine in real-time or near real-time the cardiac rhythm using at least a return value of the signal timing function and/or a return value of the signal morphology function.
- 35 11. A monitoring system according to claim 2, characterized in that the calculator comprises a mapping function allowing automatic tracking of the location of the tip of the ablation catheter (12) in three-dimensional space, said location of the tip being determined by a electrophysiology (EP) mapping system (19) in real time or near real time.

- 5 12. A monitoring system according to claim 2, characterized in that the calculator comprises a mapping function allowing automatic tracking of the location of the tip of the ablation catheter (12), said tracking being determined by low-intensity x-ray (fluoroscopy) of the ablation catheter in real time or near real time.
- 10 13. A monitoring system according to claims 1 and 12, characterized in that the cardiac risk indicator is a probability of AV block occurrence, said probability being characterised by cardiac rhythm comprising at least one occurrence of :
- an atrio-ventricular interval prolongation ;
 - a ventriculo-atrial interval prolongation ;
 - an atrio-ventricular dissociation ;
 - a ventriculo-atrial dissociation ;
 - 15 • a fast junctional rhythm ;
 - a PR interval prolongation ;
 - consecutive atrial depolarisations ;
 - consecutive ventricular depolarisations ;
 - indeterminate atrio-ventricular nodal conduction.
- 20 14. A monitoring system according to claims 2, 11 or 12, characterized in that the cardiac risk indicator is a probability of AV block occurrence, said probability being characterised by displacement of the ablation catheter (12) in excess of a predefined limit away from the initial ablation position.
- 25 15. A monitoring system according to claim 1, characterized in that it comprises a foot pedal (20) for manually switching off the power source (16) of the ablation catheter (12), and a switch unit (26) for automatically switching off the power source (16), the switch unit (26) comprising :
- 30 ▪ a first interface with the calculator (15, 16, 17, 18, 19, 25, 27) allowing the transmission of the stop command (Sc) to the power source (16) and ;
 - a second interface with the foot pedal (20) allowing the transmission of the state of the foot pedal to the power source.
 - 35 ▪ A third interface with the foot pedal (20) allowing the transmission of the state of the foot pedal to the calculator

16. A monitoring system according to claims 1 to 3, 6, 13, 14, characterized in that it comprises a user interface allowing the selection of:
- a first set of parameters for processing the signal timing function and/or ;
 - a second set of parameters for processing the signal morphology function and/or ;
 - a third set of parameters for processing the mapping function and/or ;
 - a fourth set of parameters for defining the cardiac rhythms of risk which will trigger the stop command (Sc) and/or ;
 - a fifth set of parameters for defining the ablation catheter (12) displacement limit which will trigger the stop command (Sc).
17. A monitoring system according to claim 1, characterized in that it comprises a screen displaying pulse markers in real time, the pulse markers corresponding to a onset or a peak of each atrial and/or ventricular depolarisation as determined by recordings acquired by the IEGM electrode (13) and the ECG electrodes (14).
18. A monitoring system according to claim 1 to 17, characterized in that it utilizes at least one calculator and memory incorporated into at least one existing system used for the performance of cardiac ablation, said existing systems including inclusive of but not limited to :
- a pacing unit (18) that provides current pulses at programmed rates and patterns for stimulating the heart ;
 - a junction box (15) that provides a common interface gathering the connections of ECG electrodes and IEGM electrodes in order to deliver acquired signals to the calculator and existing systems ;
 - an electrophysiology (EP) recording system (17) that records and displays electrical signals from the heart for the purposes of diagnosis and treatment of arrhythmia ;
 - an electrophysiology (EP) mapping system (19) that records and displays the three-dimensional location of catheters ;

- an ablation generator (16) that provides power to the ablation catheter.



2 / 24




**Fig. 2A****Fig. 2B****Fig. 2C**

FIG. 3A




FIG. 3B




FIG. 4A




FIG. 4B




FIG. 5




FIG. 6A





FIG. 6B





FIG. 6C




FIG. 7





FIG. 8



9/24

FIG. 9A





FIG. 9B



10/24

FIG. 10A




FIG. 10B




FIG. 11A





FIG. 11B



12 / 24









FIG. 13



pulse outputs from the Rhythm Monitoring Device to the EP recording system, to allow a visual check of atrial and ventricular depolarisation timing, as determined by the Rhythm Monitoring Device.





Fig. 14

FIG. 15




**Fig. 16**

FIG. 17




FIG. 19




FIG. 20A




FIG. 20B




Fig. 21A





Fig. 21B



22/24

FIG. 22

	Device Performance
No. of patients	19
No. of ablations	208
No. of activation events	30,754
No. of inaccurate timings	
<i>Atrial IEGM</i>	10 /15,317
<i>ECG QRS complex</i>	2 /15,437
Total # of cut-outs	114
<i>A no V</i>	32
<i>V no A</i>	29
<i>Fast Junctional</i>	53
<i>PR prolongation</i>	-
<i>Consecutive As or Vs</i>	-
Specificity	0.96
Sensitivity	1
PPV	0.97
NPV	1
Median pedal reaction time (s)	1.24

FIG 23A




FIG 23B




FIG 23C





FIG. 24



INTERNATIONAL SEARCH REPORT

International application No
PCT/EP2016/060046

A. CLASSIFICATION OF SUBJECT MATTER

INV. A61B5/0452 A61B5/0472 A61B5/042 A61B18/14 A61B6/00
ADD. A61B18/00

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

A61B

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)

EPO-Internal, WPI Data

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	US 7 578 816 B2 (BOVEJA BIRINDER R [US] ET AL) 25 August 2009 (2009-08-25) cited in the application abstract; figures 1-3 column 4, line 16 - column 5, line 53 -----	1-18
A	US 7 588 567 B2 (BOVEJA BIRINDER R [US] ET AL) 15 September 2009 (2009-09-15) abstract; figures 1, 2, 4B column 3, line 48 - column 4, line 2 column 4, line 30 - column 5, line 34 column 6, lines 26-32 -----	1-18
A	US 2004/059237 A1 (NARAYAN SANJIV MATHUR [US] ET AL) 25 March 2004 (2004-03-25) abstract; figures 1A-B, 4-5 paragraphs [0075] - [0087] ----- -/-	1-18



Further documents are listed in the continuation of Box C.



See patent family annex.

* Special categories of cited documents :

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier application or patent but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art

"&" document member of the same patent family

Date of the actual completion of the international search

19 August 2016

Date of mailing of the international search report

30/08/2016

Name and mailing address of the ISA/

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+31-70) 340-2040,
Fax: (+31-70) 340-3016

Authorized officer

Lahorte, Philippe

INTERNATIONAL SEARCH REPORT

International application No

PCT/EP2016/060046

C(Continuation). DOCUMENTS CONSIDERED TO BE RELEVANT		
Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	US 2015/099992 A1 (DENO DON CURTIS [US]) 9 April 2015 (2015-04-09) the whole document -----	1-18

INTERNATIONAL SEARCH REPORT

Information on patent family members

International application No

PCT/EP2016/060046

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 7578816	B2	25-08-2009	NONE
US 7588567	B2	15-09-2009	NONE
US 2004059237	A1	25-03-2004	AU 2003299035 A1 08-04-2004
			US 2004059237 A1 25-03-2004
			US 2007021679 A1 25-01-2007
			WO 2004026123 A2 01-04-2004
US 2015099992	A1	09-04-2015	CN 105555187 A 04-05-2016
			EP 3021744 A1 25-05-2016
			US 2015099992 A1 09-04-2015
			WO 2015054048 A1 16-04-2015