



US012133660B2

(12) **United States Patent**
Nott et al.

(10) **Patent No.:** **US 12,133,660 B2**
(45) **Date of Patent:** **Nov. 5, 2024**

- (54) **CONTROLLING A TEMPERATURE OF AN ULTRASONIC ELECTROMECHANICAL BLADE ACCORDING TO FREQUENCY**
- (71) Applicant: **Cilag GmbH International**, Zug (CH)
- (72) Inventors: **Cameron R. Nott**, Fairfield, OH (US); **Fergus P. Quigley**, Mason, OH (US); **Frederick E. Shelton, IV**, Hillsboro, OH (US); **Kevin L. Houser**, Springboro, OH (US); **David C. Yates**, Morrow, OH (US); **Patrick J. Scoggins**, Loveland, OH (US); **Craig N. Faller**, Batavia, OH (US); **Madeleine C. Jayme**, Cincinnati, OH (US)
- (73) Assignee: **Cilag GmbH International**, Zug (CH)
- (*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 899 days.

- (58) **Field of Classification Search**
CPC A61B 17/320092; A61B 17/320068; A61B 2017/00084; A61B 2017/320094; A61B 2018/00601; A61B 2018/00666
See application file for complete search history.

(56) **References Cited**
U.S. PATENT DOCUMENTS

1,853,416 A 4/1932 Hall
2,222,125 A 11/1940 Stehlik
(Continued)

FOREIGN PATENT DOCUMENTS

AU 2015201140 A1 3/2015
CA 2795323 A1 5/2014
(Continued)

OTHER PUBLICATIONS

US 10,504,709 B2, 12/2019, Karancsi et al. (withdrawn)
(Continued)

Primary Examiner — Jaymi E Della

(57) **ABSTRACT**

An ultrasonic electromechanical system for an ultrasonic surgical instrument may include an ultrasonic blade, a clamp arm disposed opposite the ultrasonic blade, an ultrasonic transducer configured to oscillate the ultrasonic blade in response to a drive signal, and a control circuit coupled to the ultrasonic transducer. The control circuit can be configured to determine a temperature of the ultrasonic blade, increase an amount of power of the drive signal when the temperature of the ultrasonic blade is less than a first predetermined value, and decrease the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than a second predetermined value. The second predetermined value may be greater than the first predetermined value. An ultrasonic generator connectable to the ultrasonic electromechanical system may include the control circuit.


- (21) Appl. No.: **17/128,552**
- (22) Filed: **Dec. 21, 2020**
- (65) **Prior Publication Data**
US 2021/0153889 A1 May 27, 2021

Related U.S. Application Data

- (63) Continuation of application No. 16/115,239, filed on Aug. 28, 2018.
(Continued)

- (51) **Int. Cl.**
A61B 17/32 (2006.01)
A61B 5/06 (2006.01)
(Continued)
- (52) **U.S. Cl.**
CPC **A61B 17/320092** (2013.01); **A61B 5/068** (2013.01); **A61B 17/072** (2013.01);
(Continued)

17 Claims, 81 Drawing Sheets



Related U.S. Application Data

(60) Provisional application No. 62/721,995, filed on Aug. 23, 2018, provisional application No. 62/721,998, filed on Aug. 23, 2018, provisional application No. 62/721,999, filed on Aug. 23, 2018, provisional application No. 62/721,994, filed on Aug. 23, 2018, provisional application No. 62/721,996, filed on Aug. 23, 2018, provisional application No. 62/692,747, filed on Jun. 30, 2018, provisional application No. 62/692,748, filed on Jun. 30, 2018, provisional application No. 62/692,768, filed on Jun. 30, 2018, provisional application No. 62/640,417, filed on Mar. 8, 2018, provisional application No. 62/640,415, filed on Mar. 8, 2018, provisional application No. 62/650,898, filed on Mar. 30, 2018, provisional application No. 62/650,887, filed on Mar. 30, 2018, provisional application No. 62/650,882, filed on Mar. 30, 2018, provisional application No. 62/650,877, filed on Mar. 30, 2018, provisional application No. 62/611,341, filed on Dec. 28, 2017, provisional application No. 62/611,340, filed on Dec. 28, 2017, provisional application No. 62/611,339, filed on Dec. 28, 2017.

(51) **Int. Cl.**

A61B 17/072 (2006.01)
A61B 17/295 (2006.01)
A61B 18/12 (2006.01)
A61B 18/14 (2006.01)
A61B 34/37 (2016.01)
A61B 90/00 (2016.01)
A61B 90/30 (2016.01)
A61B 17/00 (2006.01)
A61B 18/00 (2006.01)
A61B 34/00 (2016.01)
A61B 34/30 (2016.01)

(52) **U.S. Cl.**

CPC *A61B 17/295* (2013.01); *A61B 17/320068* (2013.01); *A61B 18/12* (2013.01); *A61B 18/1445* (2013.01); *A61B 34/37* (2016.02); *A61B 90/30* (2016.02); *A61B 90/361* (2016.02); *A61B 2017/00017* (2013.01); *A61B 2017/00022* (2013.01); *A61B 2017/00026* (2013.01); *A61B 2017/0003* (2013.01); *A61B 2017/00057* (2013.01); *A61B 2017/00075* (2013.01); *A61B 2017/00084* (2013.01); *A61B 2017/00106* (2013.01); *A61B 2017/00115* (2013.01); *A61B 2017/00123* (2013.01); *A61B 2017/00181* (2013.01); *A61B 2017/00199* (2013.01); *A61B 2017/00221* (2013.01); *A61B 2017/00398* (2013.01); *A61B 2017/00464* (2013.01); *A61B 2017/00734* (2013.01); *A61B 2017/07285* (2013.01); *A61B 2017/32007* (2017.08); *A61B 2017/320074* (2017.08); *A61B 2017/320093* (2017.08); *A61B 2017/320094* (2017.08); *A61B 2017/320095* (2017.08); *A61B 2017/320097* (2017.08); *A61B 2018/00589* (2013.01); *A61B 2018/00595* (2013.01); *A61B 2018/00601* (2013.01); *A61B 2018/00619* (2013.01); *A61B 2018/0063* (2013.01); *A61B 2018/00666* (2013.01); *A61B 2018/00803* (2013.01); *A61B 2018/00875* (2013.01); *A61B 2018/00904* (2013.01); *A61B 2018/00994* (2013.01); *A61B 18/1206* (2013.01); *A61B 2018/1253*

(2013.01); *A61B 2018/126* (2013.01); *A61B 2018/1273* (2013.01); *A61B 18/1442* (2013.01); *A61B 34/25* (2016.02); *A61B 34/30* (2016.02); *A61B 2090/064* (2016.02); *A61B 2090/065* (2016.02); *A61B 2090/066* (2016.02); *A61B 2090/0809* (2016.02); *A61B 2090/0811* (2016.02); *A61B 90/37* (2016.02); *A61B 2090/373* (2016.02); *A61B 2217/005* (2013.01); *A61B 2217/007* (2013.01); *A61B 2218/008* (2013.01); *A61B 2505/05* (2013.01)

(56)

References Cited

U.S. PATENT DOCUMENTS

3,082,426	A	3/1963	Miles
3,503,396	A	3/1970	Pierie et al.
3,584,628	A	6/1971	Green
3,626,457	A	12/1971	Duerr et al.
3,633,584	A	1/1972	Farrell
3,759,017	A	9/1973	Young
3,863,118	A	1/1975	Lander et al.
3,898,545	A	8/1975	Coppa et al.
3,912,121	A	10/1975	Steffen
3,915,271	A	10/1975	Harper
3,932,812	A	1/1976	Milligan
4,041,362	A	8/1977	Ichiyanagi
4,052,649	A	10/1977	Greenwell et al.
4,087,730	A	5/1978	Goles
4,157,859	A	6/1979	Terry
4,171,700	A	10/1979	Farin
4,202,722	A	5/1980	Paquin
4,412,539	A	11/1983	Jarvik
4,448,193	A	5/1984	Ivanov
4,523,695	A	6/1985	Braun et al.
4,608,160	A	8/1986	Zoch
4,614,366	A	9/1986	North et al.
4,633,874	A	1/1987	Chow et al.
4,701,193	A	10/1987	Robertson et al.
4,735,603	A	4/1988	Goodson et al.
4,788,977	A	12/1988	Farin et al.
4,849,752	A	7/1989	Bryant
D303,787	S	10/1989	Messenger et al.
4,892,244	A	1/1990	Fox et al.
4,976,173	A	12/1990	Yang
5,010,341	A	4/1991	Huntley et al.
5,026,387	A	6/1991	Thomas
5,035,692	A	7/1991	Lyon et al.
5,042,460	A	8/1991	Sakurai et al.
5,047,043	A	9/1991	Kubota et al.
5,084,057	A	1/1992	Green et al.
5,100,402	A	3/1992	Fan
D327,061	S	6/1992	Soren et al.
5,129,570	A	7/1992	Schulze et al.
5,151,102	A	9/1992	Kamiyama et al.
5,156,315	A	10/1992	Green et al.
5,158,585	A	10/1992	Saho et al.
5,171,247	A	12/1992	Hughett et al.
5,189,277	A	2/1993	Boisvert et al.
5,197,962	A	3/1993	Sansom et al.
5,204,669	A	4/1993	Dorfe et al.
5,217,003	A	6/1993	Wilk
5,242,474	A	9/1993	Herbst et al.
5,253,793	A	10/1993	Green et al.
5,271,543	A	12/1993	Grant et al.
RE34,519	E	1/1994	Fox et al.
5,275,323	A	1/1994	Schulze et al.
5,318,516	A	6/1994	Cosmescu
5,318,563	A	6/1994	Malis et al.
5,322,055	A	6/1994	Davison et al.
5,342,349	A	8/1994	Kaufman
5,364,003	A	11/1994	Williamson, IV
5,383,880	A	1/1995	Hooven
5,385,544	A	1/1995	Edwards et al.
5,396,900	A	3/1995	Slater et al.
5,397,046	A	3/1995	Savage et al.
5,403,312	A	4/1995	Yates et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

5,403,327	A	4/1995	Thornton et al.	6,030,437	A	2/2000	Gourrier et al.
5,413,267	A	5/1995	Solyntjes et al.	6,036,637	A	3/2000	Kudo
5,415,335	A	5/1995	Knodell, Jr.	6,039,734	A	3/2000	Goble
5,417,699	A	5/1995	Klein et al.	6,039,735	A	3/2000	Greep
5,439,468	A	8/1995	Schulze et al.	6,059,799	A	5/2000	Aranyi et al.
5,445,304	A	8/1995	Plyley et al.	6,066,137	A	5/2000	Greep
5,462,545	A	10/1995	Wang et al.	6,079,606	A	6/2000	Milliman et al.
5,465,895	A	11/1995	Knodel et al.	6,090,107	A	7/2000	Borgmeier et al.
5,467,911	A	11/1995	Tsuruta et al.	6,099,537	A	8/2000	Sugai et al.
5,474,566	A	12/1995	Alesi et al.	6,102,907	A	8/2000	Smethers et al.
5,485,947	A	1/1996	Olson et al.	6,109,500	A	8/2000	Alli et al.
5,496,315	A	3/1996	Weaver et al.	6,113,598	A	9/2000	Baker
5,496,317	A	3/1996	Goble et al.	6,126,592	A	10/2000	Proch et al.
5,503,320	A	4/1996	Webster et al.	6,126,658	A	10/2000	Baker
5,529,235	A	6/1996	Boiarski et al.	6,131,789	A	10/2000	Schulze et al.
5,531,743	A	7/1996	Nettekoven et al.	6,155,473	A	12/2000	Tompkins et al.
5,545,148	A	8/1996	Wurster	6,214,000	B1	4/2001	Fleenor et al.
5,552,685	A	9/1996	Young et al.	6,258,105	B1	7/2001	Hart et al.
5,560,372	A	10/1996	Cory	6,269,411	B1	7/2001	Reasoner
5,584,425	A	12/1996	Savage et al.	6,273,887	B1	8/2001	Yamauchi et al.
5,610,379	A	3/1997	Muz et al.	6,283,960	B1	9/2001	Ashley
5,610,811	A	3/1997	Honda	6,301,495	B1	10/2001	Gueziec et al.
5,613,966	A	3/1997	Makower et al.	6,302,881	B1	10/2001	Farin
5,624,452	A	4/1997	Yates	6,308,089	B1	10/2001	von der Ruhr et al.
D379,346	S	5/1997	Mieki	6,325,808	B1	12/2001	Bernard et al.
5,626,587	A	5/1997	Bishop et al.	6,325,811	B1	12/2001	Messerly
5,643,291	A	7/1997	Pier et al.	6,331,181	B1	12/2001	Tierney et al.
5,654,750	A	8/1997	Weil et al.	6,341,164	B1	1/2002	Dilkie et al.
5,673,841	A	10/1997	Schulze et al.	6,391,102	B1	5/2002	Bodden et al.
5,673,842	A	10/1997	Bittner et al.	6,423,057	B1	7/2002	He et al.
5,675,227	A	10/1997	Roos et al.	6,434,416	B1	8/2002	Mizoguchi et al.
5,693,042	A	12/1997	Boiarski et al.	6,443,973	B1	9/2002	Whitman
5,693,052	A	12/1997	Weaver	6,451,015	B1	9/2002	Rittman, III et al.
5,695,502	A	12/1997	Pier et al.	6,454,781	B1	9/2002	Witt et al.
5,697,926	A	12/1997	Weaver	6,457,625	B1	10/2002	Tormala et al.
5,706,998	A	1/1998	Plyley et al.	6,461,352	B2	10/2002	Morgan et al.
5,718,359	A	2/1998	Palmer et al.	6,466,817	B1	10/2002	Kaula et al.
5,724,468	A	3/1998	Leone et al.	6,480,796	B2	11/2002	Wiener
5,725,536	A	3/1998	Oberlin et al.	6,524,307	B1	2/2003	Palmerton et al.
5,725,542	A	3/1998	Yoon	6,530,933	B1	3/2003	Yeung et al.
5,735,445	A	4/1998	Vidal et al.	6,551,243	B2	4/2003	Bocionek et al.
5,735,848	A	4/1998	Yates et al.	6,569,109	B2	5/2003	Sakurai et al.
5,746,209	A	5/1998	Yost et al.	6,582,424	B2	6/2003	Fleenor et al.
5,749,362	A	5/1998	Funda et al.	6,584,358	B2	6/2003	Carter et al.
5,749,893	A	5/1998	Vidal et al.	6,585,791	B1	7/2003	Garito et al.
5,752,644	A	5/1998	Bolanos et al.	6,611,793	B1	8/2003	Burnside et al.
5,762,255	A	6/1998	Chrisman et al.	6,618,626	B2	9/2003	West, Jr. et al.
5,762,458	A	6/1998	Wang et al.	6,633,234	B2	10/2003	Wiener et al.
5,766,186	A	6/1998	Faraz et al.	6,648,223	B2	11/2003	Boukhny et al.
5,769,791	A	6/1998	Benaron et al.	6,678,552	B2	1/2004	Pearlman
5,775,331	A	7/1998	Raymond et al.	6,679,899	B2	1/2004	Wiener et al.
5,797,537	A	8/1998	Oberlin et al.	6,685,704	B2	2/2004	Greep
5,800,350	A	9/1998	Coppleson et al.	6,699,187	B2	3/2004	Webb et al.
D399,561	S	10/1998	Ellingson	6,731,514	B2	5/2004	Evans
5,817,093	A	10/1998	Williamson, IV et al.	6,742,895	B2	6/2004	Robin
5,820,009	A	10/1998	Melling et al.	6,752,816	B2	6/2004	Culp et al.
5,833,690	A	11/1998	Yates et al.	6,760,616	B2	7/2004	Hoey et al.
5,836,849	A	11/1998	Mathiak et al.	6,770,072	B1	8/2004	Truckai et al.
5,836,869	A	11/1998	Kudo et al.	6,773,444	B2	8/2004	Messerly
5,836,909	A	11/1998	Cosmescu	6,775,575	B2	8/2004	Bommannan et al.
5,843,080	A	12/1998	Fleenor et al.	6,778,846	B1	8/2004	Martinez et al.
5,846,237	A	12/1998	Nettekoven	6,781,683	B2	8/2004	Kacyra et al.
5,849,022	A	12/1998	Sakashita et al.	6,783,524	B2	8/2004	Anderson et al.
5,873,873	A	2/1999	Smith et al.	6,783,525	B2	8/2004	Greep et al.
5,878,938	A	3/1999	Bittner et al.	6,793,652	B1	9/2004	Whitman et al.
5,893,849	A	4/1999	Weaver	6,793,663	B2	9/2004	Kneifel et al.
5,906,625	A	5/1999	Bito et al.	6,824,539	B2	11/2004	Novak
5,942,333	A	8/1999	Arnett et al.	6,846,308	B2	1/2005	Whitman et al.
5,947,996	A	9/1999	Logeman	6,849,074	B2	2/2005	Chen et al.
5,968,032	A	10/1999	Sleister	6,852,219	B2	2/2005	Hammond
5,980,510	A	11/1999	Tsonton et al.	6,863,650	B1	3/2005	Irion
5,987,346	A	11/1999	Benaron et al.	6,869,430	B2	3/2005	Balbierz et al.
5,997,528	A	12/1999	Bisch et al.	6,869,435	B2	3/2005	Blake, III
6,004,269	A	12/1999	Crowley et al.	6,911,033	B2	6/2005	de Guillebon et al.
6,010,054	A	1/2000	Johnson et al.	6,913,471	B2	7/2005	Smith
				6,937,892	B2	8/2005	Leyde et al.
				6,945,981	B2	9/2005	Donofrio et al.
				6,951,559	B1	10/2005	Greep
				6,962,587	B2	11/2005	Johnson et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

6,978,921 B2	12/2005	Shelton, IV et al.	7,621,192 B2	11/2009	Conti et al.
6,988,649 B2	1/2006	Shelton, IV et al.	7,621,898 B2	11/2009	Lalomia et al.
7,000,818 B2	2/2006	Shelton, IV et al.	7,631,793 B2	12/2009	Rethy et al.
7,009,511 B2	3/2006	Mazar et al.	7,637,410 B2	12/2009	Marczyk
7,030,146 B2	4/2006	Baynes et al.	7,637,907 B2	12/2009	Blaha
7,032,798 B2	4/2006	Whitman et al.	7,641,092 B2	1/2010	Kruszynski et al.
7,041,941 B2	5/2006	Faries, Jr. et al.	7,644,848 B2	1/2010	Swayze et al.
7,044,352 B2	5/2006	Shelton, IV et al.	7,667,592 B2	2/2010	Ohyama et al.
7,044,911 B2	5/2006	Drinan et al.	7,667,839 B2	2/2010	Bates
7,044,949 B2	5/2006	Orszulak et al.	7,670,334 B2	3/2010	Hueil et al.
7,048,775 B2	5/2006	Jornitz et al.	7,694,865 B2	4/2010	Scirica
7,053,752 B2	5/2006	Wang et al.	7,699,860 B2	4/2010	Huitema et al.
7,055,730 B2	6/2006	Ehrenfels et al.	7,720,306 B2	5/2010	Gardiner et al.
7,073,765 B2	7/2006	Newkirk	7,721,934 B2	5/2010	Shelton, IV et al.
7,077,853 B2	7/2006	Kramer et al.	7,721,936 B2	5/2010	Shalton, IV et al.
7,077,856 B2	7/2006	Whitman	7,736,357 B2	6/2010	Lee, Jr. et al.
7,081,096 B2	7/2006	Brister et al.	7,742,176 B2	6/2010	Braunecker et al.
7,097,640 B2	8/2006	Wang et al.	7,743,960 B2	6/2010	Whitman et al.
7,103,688 B2	9/2006	Strong	7,753,245 B2	7/2010	Boudreaux et al.
7,104,949 B2	9/2006	Anderson et al.	7,757,028 B2	7/2010	Druke et al.
7,118,564 B2	10/2006	Ritchie et al.	7,766,207 B2	8/2010	Mather et al.
7,121,460 B1	10/2006	Parsons et al.	7,766,905 B2	8/2010	Paterson et al.
7,137,980 B2	11/2006	Buysse et al.	7,770,773 B2	8/2010	Whitman et al.
7,140,528 B2	11/2006	Shelton, IV	7,771,429 B2	8/2010	Ballard et al.
7,143,923 B2	12/2006	Shelton, IV et al.	7,776,037 B2	8/2010	Odom
7,143,925 B2	12/2006	Shelton, IV et al.	7,782,789 B2	8/2010	Stultz et al.
7,147,139 B2	12/2006	Schwemberger et al.	7,784,663 B2	8/2010	Shelton, IV
7,155,316 B2	12/2006	Sutherland et al.	7,803,151 B2	9/2010	Whitman
7,164,940 B2	1/2007	Hareyama et al.	7,810,692 B2	10/2010	Hall et al.
7,169,145 B2	1/2007	Isaacson et al.	7,818,041 B2	10/2010	Kim et al.
7,177,533 B2	2/2007	McFarlin et al.	7,819,298 B2	10/2010	Hall et al.
7,182,775 B2	2/2007	de Guillebon et al.	7,832,612 B2	11/2010	Baxter, III et al.
7,207,472 B2	4/2007	Wukusick et al.	7,833,219 B2	11/2010	Tashiro et al.
7,208,005 B2	4/2007	Frecker et al.	7,836,085 B2	11/2010	Petakov et al.
7,217,269 B2	5/2007	El-Galley et al.	7,837,079 B2	11/2010	Holsten et al.
7,230,529 B2	6/2007	Ketcherside, Jr. et al.	7,837,680 B2	11/2010	Isaacson et al.
7,232,447 B2	6/2007	Gellman et al.	7,841,980 B2	11/2010	Minosawa et al.
7,236,817 B2	6/2007	Papas et al.	7,845,537 B2	12/2010	Shelton, IV et al.
7,246,734 B2	7/2007	Shelton, IV	7,857,185 B2	12/2010	Swayze et al.
7,252,664 B2	8/2007	Nasab et al.	D631,252 S	1/2011	Leslie
7,278,563 B1	10/2007	Green	7,862,560 B2	1/2011	Marion
7,294,106 B2	11/2007	Birkenbach et al.	7,862,579 B2	1/2011	Ortiz et al.
7,294,116 B1	11/2007	Ellman et al.	7,865,236 B2	1/2011	Cory et al.
7,296,724 B2	11/2007	Green et al.	7,884,735 B2	2/2011	Newkirk
7,317,955 B2	1/2008	McGreevy	7,887,530 B2	2/2011	Zemlok et al.
7,328,828 B2	2/2008	Ortiz et al.	7,892,337 B2	2/2011	Palmerton et al.
7,334,717 B2	2/2008	Rethy et al.	7,907,166 B2	3/2011	Lamprecht et al.
7,343,565 B2	3/2008	Ying et al.	7,913,891 B2	3/2011	Doll et al.
7,344,532 B2	3/2008	Goble et al.	7,918,230 B2	4/2011	Whitman et al.
7,353,068 B2	4/2008	Tanaka et al.	7,918,377 B2	4/2011	Measamer et al.
7,362,228 B2	4/2008	Nycz et al.	7,920,706 B2	4/2011	Asokan et al.
7,371,227 B2	5/2008	Zeiner	7,927,014 B2	4/2011	Dehler
7,380,695 B2	6/2008	Doll et al.	7,932,826 B2	4/2011	Fritchie et al.
7,383,088 B2	6/2008	Spinelli et al.	7,942,300 B2	5/2011	Rethy et al.
7,391,173 B2	6/2008	Schena	7,945,065 B2	5/2011	Menzl et al.
7,407,074 B2	8/2008	Ortiz et al.	7,945,342 B2	5/2011	Tsai et al.
7,408,439 B2	8/2008	Wang et al.	7,951,148 B2	5/2011	McClurken
7,422,136 B1	9/2008	Marczyk	7,954,682 B2	6/2011	Giordano et al.
7,422,139 B2	9/2008	Shelton, IV et al.	7,955,322 B2	6/2011	Devengenzo et al.
7,423,972 B2	9/2008	Shaham et al.	7,956,620 B2	6/2011	Gilbert
D579,876 S	11/2008	Novotney et al.	7,963,433 B2	6/2011	Whitman et al.
7,457,804 B2	11/2008	Uber, III et al.	7,966,269 B2	6/2011	Bauer et al.
D583,328 S	12/2008	Chiang	7,967,180 B2	6/2011	Scirica
7,464,847 B2	12/2008	Viola et al.	7,976,553 B2	7/2011	Shelton, IV et al.
7,464,849 B2	12/2008	Shelton, IV et al.	7,979,157 B2	7/2011	Anvari
7,496,418 B2	2/2009	Kim et al.	7,980,443 B2	7/2011	Scheib et al.
D589,447 S	3/2009	Sasada et al.	7,982,776 B2	7/2011	Dunki-Jacobs et al.
7,515,961 B2	4/2009	Germanson et al.	7,988,028 B2	8/2011	Farascioni et al.
7,518,502 B2	4/2009	Austin et al.	7,993,140 B2	8/2011	Sakezles
7,554,343 B2	6/2009	Bromfield	7,995,045 B2	8/2011	Dunki-Jacobs
7,563,259 B2	7/2009	Takahashi	8,005,947 B2	8/2011	Morris et al.
7,568,604 B2	8/2009	Ehrenfels et al.	8,007,494 B1	8/2011	Taylor et al.
7,575,144 B2	8/2009	Ortiz et al.	8,007,513 B2	8/2011	Nalagatla et al.
7,597,731 B2	10/2009	Palmerton et al.	8,010,180 B2	8/2011	Quaid et al.
7,617,137 B2	11/2009	Kreiner et al.	8,012,170 B2	9/2011	Whitman et al.
			8,015,976 B2	9/2011	Shah
			8,016,855 B2	9/2011	Whitman et al.
			8,019,094 B2	9/2011	Hsieh et al.
			8,025,199 B2	9/2011	Whitman et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

8,027,710	B1	9/2011	Dannan	8,403,946	B2	3/2013	Whitfield et al.
8,035,685	B2	10/2011	Jensen	8,406,859	B2	3/2013	Zuzak et al.
8,038,686	B2	10/2011	Huitema et al.	8,411,034	B2	4/2013	Boillot et al.
8,038,693	B2	10/2011	Allen	8,413,871	B2	4/2013	Racenet et al.
8,043,560	B2	10/2011	Okumoto et al.	8,422,035	B2	4/2013	Hinderling et al.
8,054,184	B2	11/2011	Cline et al.	8,423,182	B2	4/2013	Robinson et al.
8,054,752	B2	11/2011	Druke et al.	8,428,722	B2	4/2013	Verhoef et al.
8,062,306	B2	11/2011	Nobis et al.	8,429,153	B2	4/2013	Birdwell et al.
8,062,330	B2	11/2011	Prommersberger et al.	8,439,910	B2	5/2013	Greep et al.
8,066,721	B2	11/2011	Kortenbach et al.	8,444,663	B2	5/2013	Houser et al.
8,074,861	B2	12/2011	Ehrenfels et al.	8,452,615	B2	5/2013	Abri
8,075,571	B2	12/2011	Vitali et al.	8,454,506	B2	6/2013	Rothman et al.
8,096,459	B2	1/2012	Ortiz et al.	8,461,744	B2	6/2013	Wiener et al.
8,118,206	B2	2/2012	Zand et al.	8,468,030	B2	6/2013	Stroup et al.
8,120,301	B2	2/2012	Goldberg et al.	8,469,973	B2	6/2013	Meade et al.
8,123,764	B2	2/2012	Meade et al.	8,472,630	B2	6/2013	Konrad et al.
D655,678	S	3/2012	Kobayashi et al.	D687,146	S	7/2013	Juzkiw et al.
8,128,625	B2	3/2012	Odom	8,476,227	B2	7/2013	Kaplan et al.
8,131,565	B2	3/2012	Dicks et al.	8,478,418	B2	7/2013	Fahey
8,136,712	B2	3/2012	Zingman	8,489,235	B2	7/2013	Moll et al.
D657,368	S	4/2012	Magee et al.	8,499,992	B2	8/2013	Whitman et al.
8,147,486	B2	4/2012	Honour et al.	8,500,728	B2	8/2013	Newton et al.
8,155,479	B2	4/2012	Hoffman et al.	8,500,756	B2	8/2013	Papa et al.
8,157,145	B2	4/2012	Shelton, IV et al.	8,503,759	B2	8/2013	Greer et al.
8,157,150	B2	4/2012	Viola et al.	8,505,801	B2	8/2013	Ehrenfels et al.
8,157,151	B2	4/2012	Ingmanson et al.	8,506,478	B2	8/2013	Mizuyoshi
8,160,098	B1	4/2012	Yan et al.	8,512,325	B2	8/2013	Mathonnet
8,160,690	B2	4/2012	Willfley et al.	8,512,365	B2	8/2013	Wiener et al.
8,161,977	B2	4/2012	Shelton, IV et al.	8,515,520	B2	8/2013	Brunnett et al.
8,170,396	B2	5/2012	Kuspa et al.	8,517,239	B2	8/2013	Scheib et al.
8,172,836	B2	5/2012	Ward	8,521,331	B2	8/2013	Itkowitz
8,181,839	B2	5/2012	Beetel	8,523,043	B2	9/2013	Ullrich et al.
8,185,409	B2	5/2012	Putnam et al.	8,540,709	B2	9/2013	Allen
8,206,345	B2	6/2012	Abboud et al.	8,546,996	B2	10/2013	Messerly et al.
8,208,707	B2	6/2012	Mendonca et al.	8,554,697	B2	10/2013	Claus et al.
8,210,411	B2	7/2012	Yates et al.	8,560,047	B2	10/2013	Haider et al.
8,214,007	B2	7/2012	Baker et al.	8,561,870	B2	10/2013	Baxter, III et al.
8,216,849	B2	7/2012	Petty	8,562,598	B2	10/2013	Falkenstein et al.
8,220,688	B2	7/2012	Laurent et al.	8,566,115	B2	10/2013	Moore
8,225,643	B2	7/2012	Abboud et al.	8,567,393	B2	10/2013	Hickle et al.
8,225,979	B2	7/2012	Farascioni et al.	8,571,598	B2	10/2013	Valavi
8,226,629	B1*	7/2012	Keilman A61M 37/0092 604/20	8,573,459	B2	11/2013	Smith et al.
8,229,549	B2	7/2012	Whitman et al.	8,573,465	B2	11/2013	Shelton, IV
8,231,042	B2	7/2012	Hessler et al.	8,574,229	B2	11/2013	Eder et al.
8,239,066	B2	8/2012	Jennings et al.	8,585,631	B2	11/2013	Dacquay
8,241,322	B2	8/2012	Whitman et al.	8,585,694	B2	11/2013	Amoah et al.
8,255,045	B2	8/2012	Gharib et al.	8,590,762	B2	11/2013	Hess et al.
D667,838	S	9/2012	Magee et al.	8,591,536	B2	11/2013	Robertson
8,257,387	B2	9/2012	Cunningham	8,595,607	B2	11/2013	Nekoomaram et al.
8,260,016	B2	9/2012	Maeda et al.	8,596,513	B2	12/2013	Olson et al.
8,262,560	B2	9/2012	Whitman	8,596,515	B2	12/2013	Okoniewski
8,292,639	B2	10/2012	Achammer et al.	8,604,709	B2	12/2013	Jalbout et al.
8,292,888	B2	10/2012	Whitman	8,608,044	B2	12/2013	Hueil et al.
8,295,902	B2	10/2012	Salahieh et al.	8,608,045	B2	12/2013	Smith et al.
8,308,040	B2	11/2012	Huang et al.	8,616,431	B2	12/2013	Timm et al.
8,321,581	B2	11/2012	Katis et al.	8,617,155	B2	12/2013	Johnson et al.
8,322,590	B2	12/2012	Patel et al.	8,620,055	B2	12/2013	Barratt et al.
8,328,065	B2	12/2012	Shah	8,620,473	B2	12/2013	Diolaiti et al.
8,335,590	B2	12/2012	Costa et al.	8,622,275	B2	1/2014	Baxter, III et al.
D675,164	S	1/2013	Kobayashi et al.	8,623,027	B2	1/2014	Price et al.
8,343,065	B2	1/2013	Bartol et al.	8,627,483	B2	1/2014	Rachlin et al.
8,346,392	B2	1/2013	Walsler et al.	8,627,993	B2	1/2014	Smith et al.
8,360,299	B2	1/2013	Zemlok et al.	8,627,995	B2	1/2014	Smith et al.
8,364,222	B2	1/2013	Cook et al.	8,628,518	B2	1/2014	Blumenkranz et al.
D676,392	S	2/2013	Gassauer	8,628,545	B2	1/2014	Cabrera et al.
8,365,975	B1	2/2013	Manoux et al.	8,631,987	B2	1/2014	Shelton, IV et al.
D678,196	S	3/2013	Miyauchi et al.	8,632,525	B2	1/2014	Kerr et al.
D678,304	S	3/2013	Yakoub et al.	8,636,190	B2	1/2014	Zemlok et al.
8,388,652	B2	3/2013	Viola	8,636,736	B2	1/2014	Yates et al.
8,393,514	B2	3/2013	Shelton, IV et al.	8,641,621	B2	2/2014	Razzaque et al.
8,397,972	B2	3/2013	Kostrzewski	8,652,086	B2	2/2014	Gerg et al.
8,398,541	B2	3/2013	DiMaio et al.	8,652,121	B2	2/2014	Quick et al.
8,403,944	B2	3/2013	Pain et al.	8,652,128	B2	2/2014	Ward
8,403,945	B2	3/2013	Whitfield et al.	8,657,176	B2	2/2014	Shelton, IV et al.
				8,657,177	B2	2/2014	Scirica et al.
				8,663,220	B2	3/2014	Wiener et al.
				8,666,544	B2	3/2014	Moll et al.
				8,679,114	B2	3/2014	Chapman et al.
				8,682,049	B2	3/2014	Zhao et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

8,682,489	B2	3/2014	Itkowitz et al.	8,991,678	B2	3/2015	Wellman et al.
8,685,056	B2	4/2014	Evans et al.	8,992,565	B2	3/2015	Brisson et al.
8,688,188	B2	4/2014	Heller et al.	8,998,797	B2	4/2015	Omori
8,690,864	B2	4/2014	Hoarau	9,002,518	B2	4/2015	Manzo et al.
8,701,962	B2	4/2014	Kostrzewski	9,010,608	B2	4/2015	Casasanta, Jr. et al.
D704,839	S	5/2014	Juzkiw et al.	9,010,611	B2	4/2015	Ross et al.
8,719,061	B2	5/2014	Birchall	9,011,366	B2	4/2015	Dean et al.
8,720,766	B2	5/2014	Hess et al.	9,011,427	B2	4/2015	Price et al.
8,733,613	B2	5/2014	Huitema et al.	9,016,539	B2	4/2015	Kostrzewski et al.
8,740,840	B2	6/2014	Foley et al.	9,017,326	B2	4/2015	DiNardo et al.
8,740,866	B2	6/2014	Reasoner et al.	9,020,240	B2	4/2015	Pettersson et al.
8,747,238	B2	6/2014	Shelton, IV et al.	D729,267	S	5/2015	Yoo et al.
8,752,749	B2	6/2014	Moore et al.	9,023,032	B2	5/2015	Robinson
8,757,465	B2	6/2014	Woodard, Jr. et al.	9,023,071	B2	5/2015	Miller et al.
8,761,717	B1	6/2014	Buchheit	9,027,431	B2	5/2015	Tang et al.
8,763,879	B2	7/2014	Shelton, IV et al.	9,028,494	B2	5/2015	Shelton, IV et al.
8,768,251	B2	7/2014	Claus et al.	9,035,568	B2	5/2015	Ganton et al.
8,771,270	B2	7/2014	Burbank	9,038,882	B2	5/2015	Racenet et al.
8,775,196	B2	7/2014	Simpson et al.	9,043,027	B2	5/2015	Durant et al.
8,779,648	B2	7/2014	Giordano et al.	9,044,227	B2	6/2015	Shelton, IV et al.
8,790,253	B2	7/2014	Sunagawa et al.	9,044,244	B2	6/2015	Ludwin et al.
8,794,497	B2	8/2014	Zingman	9,044,261	B2	6/2015	Houser
8,795,001	B1	8/2014	Lam et al.	9,050,063	B2	6/2015	Roe et al.
8,799,008	B2	8/2014	Johnson et al.	9,050,083	B2	6/2015	Yates et al.
8,799,009	B2	8/2014	Mellin et al.	9,050,120	B2	6/2015	Swarup et al.
8,800,838	B2	8/2014	Shelton, IV	9,052,809	B2	6/2015	Vesto
8,801,703	B2	8/2014	Gregg et al.	9,055,035	B2	6/2015	Porsch et al.
8,814,996	B2	8/2014	Giurgiutiu et al.	9,055,870	B2	6/2015	Meador et al.
8,818,556	B2	8/2014	Sanchez et al.	9,060,770	B2	6/2015	Shelton, IV et al.
8,819,581	B2	8/2014	Nakamura et al.	9,060,775	B2	6/2015	Wiener et al.
8,820,603	B2	9/2014	Shelton, IV et al.	9,066,650	B2	6/2015	Sekiguchi
8,820,608	B2	9/2014	Miyamoto	9,072,523	B2	7/2015	Houser et al.
8,827,134	B2	9/2014	Viola et al.	9,072,535	B2	7/2015	Shelton, IV et al.
8,840,003	B2	9/2014	Morgan et al.	9,072,536	B2	7/2015	Shelton, IV et al.
D716,333	S	10/2014	Chotin et al.	9,078,653	B2	7/2015	Leimbach et al.
8,851,354	B2	10/2014	Swensgard et al.	9,078,727	B2	7/2015	Miller
8,852,174	B2	10/2014	Burbank	9,084,606	B2	7/2015	Greep
8,875,973	B2	11/2014	Whitman	9,089,360	B2	7/2015	Messerly et al.
8,876,857	B2	11/2014	Burbank	9,095,362	B2	8/2015	Dachs, II et al.
8,882,662	B2	11/2014	Charles	9,095,367	B2	8/2015	Olson et al.
8,886,790	B2	11/2014	Harrang et al.	9,099,863	B2	8/2015	Smith et al.
8,893,949	B2	11/2014	Shelton, IV et al.	9,101,358	B2	8/2015	Kerr et al.
8,899,479	B2	12/2014	Cappuzzo et al.	9,101,359	B2	8/2015	Smith et al.
8,905,977	B2	12/2014	Shelton et al.	9,101,374	B1	8/2015	Hoch et al.
8,912,746	B2	12/2014	Reid et al.	9,106,270	B2	8/2015	Puterbaugh et al.
8,914,098	B2	12/2014	Brennan et al.	9,107,573	B2	8/2015	Birnkrant
8,917,513	B1	12/2014	Hazzard	9,107,662	B2	8/2015	Kostrzewski
8,918,207	B2	12/2014	Prisco	9,107,684	B2	8/2015	Ma
8,920,186	B2	12/2014	Shishikura	9,107,688	B2	8/2015	Kimball et al.
8,920,414	B2	12/2014	Stone et al.	9,107,689	B2	8/2015	Robertson et al.
8,920,433	B2	12/2014	Barrier et al.	9,107,694	B2	8/2015	Hendriks et al.
8,930,203	B2	1/2015	Kiaie et al.	9,111,548	B2	8/2015	Nandy et al.
8,930,214	B2	1/2015	Woolford	9,113,880	B2	8/2015	Zemlok et al.
8,931,679	B2	1/2015	Kostrzewski	9,114,494	B1	8/2015	Mah
8,936,614	B2	1/2015	Allen, IV	9,116,597	B1	8/2015	Gulasky
8,945,095	B2	2/2015	Blumenkranz et al.	9,119,617	B2	9/2015	Souls et al.
8,945,163	B2	2/2015	Voegele et al.	9,119,655	B2	9/2015	Bowling et al.
8,955,732	B2	2/2015	Zemlok et al.	9,119,657	B2	9/2015	Shelton, IV et al.
8,956,581	B2	2/2015	Rosenbaum et al.	9,123,155	B2	9/2015	Cunningham et al.
8,960,519	B2	2/2015	Whitman et al.	9,125,644	B2	9/2015	Lane et al.
8,960,520	B2	2/2015	McCuen	9,129,054	B2	9/2015	Nawana et al.
8,962,062	B2	2/2015	Podhajsky et al.	9,137,254	B2	9/2015	Bilbrey et al.
8,967,443	B2	3/2015	McCuen	9,138,129	B2	9/2015	Diolaiti
8,967,455	B2	3/2015	Zhou	9,138,225	B2	9/2015	Huang et al.
8,968,276	B2	3/2015	Zemlok et al.	9,149,322	B2	10/2015	Knowlton
8,968,296	B2	3/2015	McPherson	9,155,503	B2	10/2015	Cadwell
8,968,309	B2	3/2015	Roy et al.	9,160,853	B1	10/2015	Daddi et al.
8,968,312	B2	3/2015	Marczyk et al.	9,161,803	B2	10/2015	Yates et al.
8,968,337	B2	3/2015	Whitfield et al.	9,168,054	B2	10/2015	Turner et al.
8,968,358	B2	3/2015	Reschke	9,168,091	B2	10/2015	Janssen et al.
8,974,429	B2	3/2015	Gordon et al.	9,168,104	B2	10/2015	Dein
8,979,890	B2	3/2015	Boudreaux	9,179,912	B2	11/2015	Yates et al.
8,986,288	B2	3/2015	Konishi	9,183,723	B2	11/2015	Sherman et al.
8,986,302	B2	3/2015	Aldridge et al.	9,186,143	B2	11/2015	Timm et al.
8,989,903	B2	3/2015	Weir et al.	9,192,375	B2	11/2015	Skinlo et al.
				9,192,447	B2	11/2015	Choi et al.
				9,192,707	B2	11/2015	Gerber et al.
				9,198,711	B2	12/2015	Joseph
				9,202,078	B2	12/2015	Abuelsead et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

9,204,830	B2	12/2015	Zand et al.	9,393,017	B2	7/2016	Flanagan et al.
9,204,879	B2	12/2015	Shelton, IV	9,393,037	B2	7/2016	Olson et al.
9,204,995	B2	12/2015	Scheller et al.	9,398,905	B2	7/2016	Martin
9,211,120	B2	12/2015	Scheib et al.	9,398,911	B2	7/2016	Auld
9,216,062	B2	12/2015	Duque et al.	9,402,629	B2	8/2016	Ehrenfels et al.
9,218,053	B2	12/2015	Komuro et al.	9,414,776	B2	8/2016	Sillay et al.
9,220,502	B2	12/2015	Zemlok et al.	9,414,940	B2	8/2016	Stein et al.
9,226,689	B2	1/2016	Jacobsen et al.	9,419,018	B2	8/2016	Sasagawa et al.
9,226,751	B2	1/2016	Shelton, IV et al.	9,421,014	B2	8/2016	Ingmanson et al.
9,226,766	B2	1/2016	Aldridge et al.	9,433,470	B2	9/2016	Choi
9,226,767	B2	1/2016	Stulen et al.	9,439,622	B2	9/2016	Case et al.
9,226,791	B2	1/2016	McCarthy et al.	9,439,668	B2	9/2016	Timm et al.
9,232,883	B2	1/2016	Ozawa et al.	9,439,736	B2	9/2016	Olson
9,237,891	B2	1/2016	Shelton, IV	9,445,764	B2	9/2016	Gross et al.
9,237,921	B2	1/2016	Messerly et al.	9,445,813	B2	9/2016	Shelton, IV et al.
9,241,728	B2	1/2016	Price et al.	9,450,701	B2	9/2016	Do et al.
9,241,730	B2	1/2016	Babaev	9,451,949	B2	9/2016	Gorek et al.
9,241,731	B2	1/2016	Boudreaux et al.	9,451,958	B2	9/2016	Shelton, IV et al.
9,247,996	B1	2/2016	Merana et al.	9,463,022	B2	10/2016	Swayze et al.
9,250,172	B2	2/2016	Harris et al.	9,463,646	B2	10/2016	Payne et al.
9,255,907	B2	2/2016	Heanue et al.	9,468,438	B2	10/2016	Baber et al.
9,265,429	B2	2/2016	St. Pierre et al.	9,474,565	B2	10/2016	Shikhman et al.
9,265,585	B2	2/2016	Wingardner et al.	D772,252	S	11/2016	Myers et al.
9,272,406	B2	3/2016	Aronhalt et al.	9,480,492	B2	11/2016	Aranyi et al.
9,277,956	B2	3/2016	Zhang	9,485,475	B2	11/2016	Speier et al.
9,277,961	B2	3/2016	Panescu et al.	9,486,271	B2	11/2016	Dunning
9,277,969	B2	3/2016	Brannan et al.	9,492,146	B2	11/2016	Kostrzewski et al.
9,280,884	B1	3/2016	Schultz et al.	9,492,237	B2	11/2016	Kang et al.
9,282,962	B2	3/2016	Schmid et al.	9,493,807	B2	11/2016	Little et al.
9,282,974	B2	3/2016	Shelton, IV	9,498,182	B2	11/2016	Case et al.
9,283,045	B2	3/2016	Rhee et al.	9,498,215	B2	11/2016	Duque et al.
9,283,054	B2	3/2016	Morgan et al.	9,498,231	B2	11/2016	Haider et al.
9,289,211	B2	3/2016	Williams et al.	9,516,239	B2	12/2016	Blanquart et al.
9,289,212	B2	3/2016	Shelton, IV et al.	9,519,753	B1	12/2016	Gerdeman et al.
9,295,514	B2	3/2016	Shelton, IV et al.	9,522,003	B2	12/2016	Weir et al.
9,301,691	B2	4/2016	Hufnagel et al.	9,526,407	B2	12/2016	Hoeg et al.
9,301,753	B2	4/2016	Aldridge et al.	9,526,499	B2	12/2016	Kostrzewski et al.
9,301,759	B2	4/2016	Spivey et al.	9,526,587	B2	12/2016	Zhao et al.
9,301,810	B2	4/2016	Amiri et al.	9,532,827	B2	1/2017	Morgan et al.
9,302,213	B2	4/2016	Manahan et al.	9,532,845	B1	1/2017	Dossett et al.
9,307,894	B2	4/2016	von Grunberg et al.	9,539,007	B2	1/2017	Dhakad et al.
9,307,914	B2	4/2016	Fahey	9,539,020	B2	1/2017	Conlon et al.
9,307,986	B2	4/2016	Hall et al.	9,542,481	B2	1/2017	Halter et al.
9,314,246	B2	4/2016	Shelton, IV et al.	9,546,662	B2	1/2017	Shener-Irmakoglu et al.
9,314,308	B2	4/2016	Parihar et al.	9,549,781	B2	1/2017	He et al.
9,320,563	B2	4/2016	Brustad et al.	9,554,692	B2	1/2017	Levy
9,325,732	B1	4/2016	Stickle et al.	9,554,794	B2	1/2017	Baber et al.
9,326,767	B2	5/2016	Koch et al.	9,554,854	B2	1/2017	Yates et al.
9,326,770	B2	5/2016	Shelton, IV et al.	9,561,038	B2	2/2017	Shelton, IV et al.
9,331,422	B2	5/2016	Nazzaro et al.	9,561,045	B2	2/2017	Hinman et al.
9,332,987	B2	5/2016	Leimbach et al.	9,561,082	B2	2/2017	Yen et al.
9,333,042	B2	5/2016	Diolaiti et al.	9,561,982	B2	2/2017	Enicks et al.
9,336,385	B1	5/2016	Spencer et al.	9,566,708	B2	2/2017	Kurnianto
9,341,704	B2	5/2016	Picard et al.	9,572,592	B2	2/2017	Price et al.
9,345,481	B2	5/2016	Hall et al.	9,579,099	B2	2/2017	Penna et al.
9,345,490	B2	5/2016	Ippisch	9,579,503	B2	2/2017	McKinney et al.
9,345,546	B2	5/2016	Toth et al.	9,585,657	B2	3/2017	Shelton, IV et al.
9,345,900	B2	5/2016	Wu et al.	9,592,095	B2	3/2017	Panescu et al.
9,351,726	B2	5/2016	Leimbach et al.	9,597,081	B2	3/2017	Swayze et al.
9,351,727	B2	5/2016	Leimbach et al.	9,600,031	B2	3/2017	Kaneko et al.
9,358,003	B2	6/2016	Hall et al.	9,600,138	B2	3/2017	Thomas et al.
9,358,685	B2	6/2016	Meier et al.	9,603,024	B2	3/2017	Wang et al.
9,360,449	B2	6/2016	Duric	9,603,277	B2	3/2017	Morgan et al.
9,364,230	B2	6/2016	Shelton, IV et al.	D783,675	S	4/2017	Yagisawa et al.
9,364,231	B2	6/2016	Wenchell	D784,270	S	4/2017	Bhattacharya
9,364,249	B2	6/2016	Kimball et al.	9,610,114	B2	4/2017	Baxter, III et al.
9,364,294	B2	6/2016	Razzaque et al.	9,610,412	B2	4/2017	Zemlok et al.
9,370,400	B2	6/2016	Parihar	9,615,877	B2	4/2017	Tyrrell et al.
9,375,282	B2	6/2016	Nau, Jr. et al.	9,622,684	B2	4/2017	Wybo
9,375,539	B2	6/2016	Stearns et al.	9,622,808	B2	4/2017	Beller et al.
9,381,003	B2	7/2016	Todor et al.	9,628,501	B2	4/2017	Datta Ray et al.
9,381,058	B2	7/2016	Houser et al.	9,629,560	B2	4/2017	Joseph
9,386,984	B2	7/2016	Aronhalt et al.	9,629,623	B2	4/2017	Lytle, IV et al.
9,386,988	B2	7/2016	Baxter, III et al.	9,629,628	B2	4/2017	Aranyi
9,387,295	B1	7/2016	Mastri et al.	9,629,629	B2	4/2017	Leimbach et al.
				9,630,318	B2	4/2017	Ibarz Gabardos et al.
				9,636,112	B2	5/2017	Penna et al.
				9,636,188	B2	5/2017	Gattani et al.
				9,636,239	B2	5/2017	Durand et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

9,636,825	B2	5/2017	Penn et al.	9,795,436	B2	10/2017	Yates et al.
9,641,596	B2	5/2017	Unagami et al.	9,797,486	B2	10/2017	Zergiebel et al.
9,641,815	B2	5/2017	Richardson et al.	9,801,531	B2	10/2017	Morita et al.
9,642,620	B2	5/2017	Baxter, III et al.	9,801,626	B2	10/2017	Parihar et al.
9,643,022	B2	5/2017	Mashiach et al.	9,801,627	B2	10/2017	Harris et al.
9,649,110	B2	5/2017	Parihar et al.	9,801,679	B2	10/2017	Trees et al.
9,649,111	B2	5/2017	Shelton, IV et al.	9,802,033	B2	10/2017	Hibner et al.
9,649,126	B2	5/2017	Robertson et al.	9,804,618	B2	10/2017	Leimbach et al.
9,649,169	B2	5/2017	Cinquin et al.	9,805,472	B2	10/2017	Chou et al.
9,652,655	B2	5/2017	Satish et al.	9,808,244	B2	11/2017	Leimbach et al.
9,655,616	B2	5/2017	Aranyi	9,808,245	B2	11/2017	Richard et al.
9,656,092	B2	5/2017	Golden	9,808,246	B2	11/2017	Shelton, IV et al.
9,662,116	B2	5/2017	Smith et al.	9,808,248	B2	11/2017	Hoffman
9,662,177	B2	5/2017	Weir et al.	9,808,249	B2	11/2017	Shelton, IV
9,668,729	B2	6/2017	Williams et al.	9,814,457	B2	11/2017	Martin et al.
9,668,732	B2	6/2017	Patel et al.	9,814,460	B2	11/2017	Kimsey et al.
9,668,765	B2	6/2017	Grace et al.	9,814,462	B2	11/2017	Woodard, Jr. et al.
9,671,860	B2	6/2017	Ogawa et al.	9,814,463	B2	11/2017	Williams et al.
9,675,264	B2	6/2017	Acquista et al.	9,820,699	B2	11/2017	Bingley et al.
9,675,354	B2	6/2017	Weir et al.	9,820,738	B2	11/2017	Lytle, IV et al.
9,681,870	B2	6/2017	Baxter, III et al.	9,820,741	B2	11/2017	Kostrzewski
9,686,306	B2	6/2017	Chizeck et al.	9,820,768	B2	11/2017	Gee et al.
9,687,230	B2	6/2017	Leimbach et al.	9,826,976	B2	11/2017	Parihar et al.
9,690,362	B2	6/2017	Leimbach et al.	9,826,977	B2	11/2017	Leimbach et al.
9,700,292	B2	7/2017	Nawana et al.	9,827,054	B2	11/2017	Richmond et al.
9,700,309	B2	7/2017	Jaworek et al.	9,827,059	B2	11/2017	Robinson et al.
9,700,312	B2	7/2017	Kostrzewski et al.	9,830,424	B2	11/2017	Dixon et al.
9,700,320	B2	7/2017	Dinardo et al.	9,833,241	B2	12/2017	Huitema et al.
9,706,993	B2	7/2017	Hessler et al.	9,833,254	B1	12/2017	Barral et al.
9,710,214	B2	7/2017	Lin et al.	9,839,419	B2	12/2017	Deck et al.
9,710,644	B2	7/2017	Reybok et al.	9,839,424	B2	12/2017	Zergiebel et al.
9,713,424	B2	7/2017	Spaide	9,839,428	B2	12/2017	Baxter, III et al.
9,713,503	B2	7/2017	Goldschmidt	9,839,467	B2	12/2017	Harper et al.
9,717,141	B1	7/2017	Tegg	9,839,470	B2	12/2017	Gilbert et al.
9,717,498	B2	8/2017	Aranyi et al.	9,839,487	B2	12/2017	Dachs, II
9,717,525	B2	8/2017	Ahluwalia et al.	9,844,321	B1	12/2017	Ekvall et al.
9,717,548	B2	8/2017	Couture	9,844,368	B2	12/2017	Boudreaux et al.
9,724,094	B2	8/2017	Baber et al.	9,844,369	B2	12/2017	Huitema et al.
9,724,100	B2	8/2017	Scheib et al.	9,844,374	B2	12/2017	Lytle, IV et al.
9,724,118	B2	8/2017	Schulte et al.	9,844,375	B2	12/2017	Overmyer et al.
9,733,663	B2	8/2017	Leimbach et al.	9,844,376	B2	12/2017	Baxter, III et al.
9,737,301	B2	8/2017	Baber et al.	9,844,379	B2	12/2017	Shelton, IV et al.
9,737,310	B2	8/2017	Whitfield et al.	9,848,058	B2	12/2017	Johnson et al.
9,737,335	B2	8/2017	Butler et al.	9,848,877	B2	12/2017	Shelton, IV et al.
9,737,355	B2	8/2017	Yates et al.	9,861,354	B2	1/2018	Saliman et al.
9,740,826	B2	8/2017	Raghavan et al.	9,861,363	B2	1/2018	Chen et al.
9,743,016	B2	8/2017	Nestares et al.	9,861,428	B2	1/2018	Trees et al.
9,743,929	B2	8/2017	Leimbach et al.	9,864,839	B2	1/2018	Baym et al.
9,743,946	B2	8/2017	Faller et al.	9,867,612	B2	1/2018	Parihar et al.
9,743,947	B2	8/2017	Price et al.	9,867,651	B2	1/2018	Wham
9,750,499	B2	9/2017	Leimbach et al.	9,867,670	B2	1/2018	Brannan et al.
9,750,500	B2	9/2017	Malkowski	9,867,914	B2	1/2018	Bonano et al.
9,750,522	B2	9/2017	Scheib et al.	9,872,609	B2	1/2018	Levy
9,750,523	B2	9/2017	Tsubuku	9,872,683	B2	1/2018	Hopkins et al.
9,750,563	B2	9/2017	Shikhman et al.	9,877,718	B2	1/2018	Weir et al.
9,753,135	B2	9/2017	Bosch	9,877,721	B2	1/2018	Schellin et al.
9,753,568	B2	9/2017	McMillen	9,883,860	B2	2/2018	Leimbach
9,757,126	B2	9/2017	Cappola	9,888,864	B2	2/2018	Rondoni et al.
9,757,128	B2	9/2017	Baber et al.	9,888,914	B2	2/2018	Martin et al.
9,757,142	B2	9/2017	Shimizu	9,888,919	B2	2/2018	Leimbach et al.
9,757,152	B2	9/2017	Ogilvie et al.	9,888,921	B2	2/2018	Williams et al.
9,763,741	B2	9/2017	Alvarez et al.	9,888,975	B2	2/2018	Auld
9,764,164	B2	9/2017	Wiener et al.	9,895,148	B2	2/2018	Shelton, IV et al.
9,770,541	B2	9/2017	Carr et al.	9,900,787	B2	2/2018	Ou
9,775,611	B2	10/2017	Kostrzewski	9,901,342	B2	2/2018	Shelton, IV et al.
9,777,913	B2	10/2017	Talbert et al.	9,901,406	B2	2/2018	State et al.
9,782,164	B2	10/2017	Mumaw et al.	9,905,000	B2	2/2018	Chou et al.
9,782,169	B2	10/2017	Kimsey et al.	9,907,196	B2	2/2018	Susini et al.
9,782,212	B2	10/2017	Wham et al.	9,907,550	B2	3/2018	Sniffin et al.
9,782,214	B2	10/2017	Houser et al.	9,913,642	B2	3/2018	Leimbach et al.
9,788,835	B2	10/2017	Morgan et al.	9,913,645	B2	3/2018	Zerkle et al.
9,788,836	B2	10/2017	Overmyer et al.	9,918,326	B2	3/2018	Gilson et al.
9,788,851	B2	10/2017	Dannaher et al.	9,918,730	B2	3/2018	Trees et al.
9,788,902	B2	10/2017	Inoue et al.	9,918,778	B2	3/2018	Walberg et al.
9,788,907	B1	10/2017	Alvi et al.	9,918,788	B2	3/2018	Paul et al.
				9,922,304	B2	3/2018	DeBusk et al.
				9,924,941	B2	3/2018	Burbank
				9,924,944	B2	3/2018	Shelton, IV et al.
				9,924,961	B2	3/2018	Shelton, IV et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

9,931,040 B2	4/2018	Homyk et al.	10,080,618 B2	9/2018	Marshall et al.
9,931,118 B2	4/2018	Shelton, IV et al.	10,084,833 B2	9/2018	McDonnell et al.
9,931,124 B2	4/2018	Gokharu	D831,209 S	10/2018	Huitema et al.
9,936,863 B2	4/2018	Tesar	10,085,748 B2	10/2018	Morgan et al.
9,936,942 B2	4/2018	Chin et al.	10,085,749 B2	10/2018	Cappola et al.
9,936,955 B2	4/2018	Miller et al.	10,092,355 B1	10/2018	Hannaford et al.
9,936,961 B2	4/2018	Chien et al.	10,095,942 B2	10/2018	Mentese et al.
9,937,012 B2	4/2018	Hares et al.	10,097,578 B2	10/2018	Baldonado et al.
9,937,014 B2	4/2018	Bowling et al.	10,098,527 B2	10/2018	Weisenburgh, II et al.
9,937,626 B2	4/2018	Rockrohr	10,098,635 B2	10/2018	Burbank
9,938,972 B2	4/2018	Walley	10,098,642 B2	10/2018	Baxter, III et al.
9,943,230 B2	4/2018	Kaku et al.	10,098,705 B2	10/2018	Brisson et al.
9,943,309 B2	4/2018	Shelton, IV et al.	10,102,926 B1	10/2018	Leonardi
9,943,312 B2	4/2018	Posada et al.	10,105,140 B2	10/2018	Malinouskas et al.
9,943,377 B2	4/2018	Yates et al.	10,105,142 B2	10/2018	Baxter, III et al.
9,943,379 B2	4/2018	Gregg, II et al.	10,105,470 B2	10/2018	Reasoner et al.
9,943,918 B2	4/2018	Grogan et al.	10,111,658 B2	10/2018	Chowaniec et al.
9,949,785 B2	4/2018	Price et al.	10,111,665 B2	10/2018	Aranyi et al.
9,962,157 B2	5/2018	Sapre	10,111,679 B2	10/2018	Baber et al.
9,968,355 B2	5/2018	Shelton, IV et al.	10,111,703 B2	10/2018	Cosman, Jr. et al.
9,974,595 B2	5/2018	Anderson et al.	D834,541 S	11/2018	You et al.
9,980,140 B1	5/2018	Spencer et al.	10,117,649 B2	11/2018	Baxter et al.
9,980,769 B2	5/2018	Trees et al.	10,117,651 B2	11/2018	Whitman et al.
9,980,778 B2	5/2018	Ohline et al.	10,117,702 B2	11/2018	Danziger et al.
9,987,000 B2	6/2018	Shelton, IV et al.	10,118,119 B2	11/2018	Sappok et al.
9,987,068 B2	6/2018	Anderson et al.	10,130,359 B2	11/2018	Hess et al.
9,987,072 B2	6/2018	McPherson	10,130,360 B2	11/2018	Olson et al.
9,990,856 B2	6/2018	Kuchenbecker et al.	10,130,361 B2	11/2018	Yates et al.
9,993,248 B2	6/2018	Shelton, IV et al.	10,130,367 B2	11/2018	Cappola et al.
9,993,258 B2	6/2018	Shelton, IV et al.	10,133,248 B2	11/2018	Fitzsimmons et al.
9,993,305 B2	6/2018	Andersson	10,135,242 B2	11/2018	Baber et al.
10,004,491 B2	6/2018	Martin et al.	10,136,887 B2	11/2018	Shelton, IV et al.
10,004,497 B2	6/2018	Overmyer et al.	10,136,891 B2	11/2018	Shelton, IV et al.
10,004,500 B2	6/2018	Shelton, IV et al.	10,136,949 B2	11/2018	Felder et al.
10,004,501 B2	6/2018	Shelton, IV et al.	10,136,954 B2	11/2018	Johnson et al.
10,004,527 B2	6/2018	Gee et al.	10,137,245 B2	11/2018	Melker et al.
10,004,557 B2	6/2018	Gross	10,143,526 B2	12/2018	Walker et al.
D822,206 S	7/2018	Shelton, IV et al.	10,143,948 B2	12/2018	Bonifas et al.
10,010,322 B2	7/2018	Shelton, IV et al.	10,147,148 B2	12/2018	Wu et al.
10,010,324 B2	7/2018	Huitema et al.	10,149,680 B2	12/2018	Parihar et al.
10,013,049 B2	7/2018	Leimbach et al.	10,152,789 B2	12/2018	Carnes et al.
10,016,199 B2	7/2018	Baber et al.	10,154,841 B2	12/2018	Weaner et al.
10,021,318 B2	7/2018	Hugosson et al.	10,159,044 B2	12/2018	Hrabak
10,022,090 B2	7/2018	Whitman	10,159,481 B2	12/2018	Whitman et al.
10,022,120 B2	7/2018	Martin et al.	10,159,483 B2	12/2018	Beckman et al.
10,022,391 B2	7/2018	Ruderman Chen et al.	10,164,466 B2	12/2018	Calderoni
10,022,568 B2	7/2018	Messerly et al.	10,166,025 B2	1/2019	Leimbach et al.
10,028,402 B1	7/2018	Walker	10,166,061 B2	1/2019	Berry et al.
10,028,744 B2	7/2018	Shelton, IV et al.	10,169,862 B2	1/2019	Andre et al.
10,028,761 B2	7/2018	Leimbach et al.	10,172,618 B2	1/2019	Shelton, IV et al.
10,028,788 B2	7/2018	Kang	10,172,687 B2	1/2019	Garbus et al.
10,034,704 B2	7/2018	Asher et al.	10,175,096 B2	1/2019	Dickerson
10,037,641 B2	7/2018	Hyde et al.	10,175,127 B2	1/2019	Collins et al.
10,037,715 B2	7/2018	Toly et al.	10,178,992 B2	1/2019	Wise et al.
D826,405 S	8/2018	Shelton, IV et al.	10,179,413 B2	1/2019	Rockrohr
10,039,546 B2	8/2018	Williams et al.	10,180,463 B2	1/2019	Beckman et al.
10,039,564 B2	8/2018	Hibner et al.	10,182,814 B2	1/2019	Okoniewski
10,039,565 B2	8/2018	Vezzu	10,182,816 B2	1/2019	Shelton, IV et al.
10,039,589 B2	8/2018	Virshek et al.	10,182,818 B2	1/2019	Hensel et al.
10,041,822 B2	8/2018	Zemlok	10,188,385 B2	1/2019	Kerr et al.
10,044,791 B2	8/2018	Kamen et al.	10,189,157 B2	1/2019	Schlegel et al.
10,045,704 B2	8/2018	Fagin et al.	10,190,888 B2	1/2019	Hryb et al.
10,045,776 B2	8/2018	Shelton, IV et al.	10,194,891 B2	2/2019	Jeong et al.
10,045,779 B2	8/2018	Savage et al.	10,194,907 B2	2/2019	Marczyk et al.
10,045,781 B2	8/2018	Cropper et al.	10,194,913 B2	2/2019	Nalagatla et al.
10,045,782 B2	8/2018	Murthy Aravalli	10,194,972 B2	2/2019	Yates et al.
10,045,813 B2	8/2018	Mueller	10,197,803 B2	2/2019	Badiali et al.
10,048,379 B2	8/2018	Markendorf et al.	10,198,965 B2	2/2019	Hart
10,052,044 B2	8/2018	Shelton, IV et al.	10,201,311 B2	2/2019	Chou et al.
10,052,102 B2	8/2018	Baxter, III et al.	10,201,349 B2	2/2019	Leimbach et al.
10,052,104 B2	8/2018	Shelton, IV et al.	10,201,364 B2	2/2019	Leimbach et al.
10,054,441 B2	8/2018	Schorr et al.	10,201,365 B2	2/2019	Boudreaux et al.
10,058,393 B2	8/2018	Bonutti et al.	10,205,708 B1	2/2019	Fletcher et al.
10,069,633 B2	9/2018	Gulati et al.	10,206,605 B2	2/2019	Shelton, IV et al.
10,076,326 B2	9/2018	Yates et al.	10,206,752 B2	2/2019	Hares et al.
			10,213,201 B2	2/2019	Shelton, IV et al.
			10,213,203 B2	2/2019	Swayze et al.
			10,213,266 B2	2/2019	Zemlok et al.
			10,213,268 B2	2/2019	Dachs, II

(56)

References Cited

U.S. PATENT DOCUMENTS

10,219,491 B2	3/2019	Stiles, Jr. et al.	10,342,602 B2	7/2019	Strobl et al.
10,220,522 B2	3/2019	Rockrohr	10,342,623 B2	7/2019	Huelman et al.
10,222,750 B2	3/2019	Bang et al.	10,343,102 B2	7/2019	Reasoner et al.
10,226,249 B2	3/2019	Jaworek et al.	10,349,824 B2	7/2019	Claude et al.
10,226,250 B2	3/2019	Beckman et al.	10,349,939 B2	7/2019	Shelton, IV et al.
10,226,254 B2	3/2019	Cabrera et al.	10,349,941 B2	7/2019	Marczyk et al.
10,226,302 B2	3/2019	Lacal et al.	10,350,016 B2	7/2019	Burbank et al.
10,231,634 B2	3/2019	Zand et al.	10,357,184 B2	7/2019	Crawford et al.
10,231,733 B2	3/2019	Ehrenfels et al.	10,357,246 B2	7/2019	Shelton, IV et al.
10,231,775 B2	3/2019	Shelton, IV et al.	10,357,247 B2	7/2019	Shelton, IV et al.
10,238,413 B2	3/2019	Hibner et al.	10,362,179 B2	7/2019	Harris
10,245,027 B2	4/2019	Shelton, IV et al.	10,363,032 B2	7/2019	Scheib et al.
10,245,028 B2	4/2019	Shelton, IV et al.	10,363,037 B2	7/2019	Aronhalt et al.
10,245,029 B2	4/2019	Hunter et al.	10,368,861 B2	8/2019	Baxter, III et al.
10,245,030 B2	4/2019	Hunter et al.	10,368,865 B2	8/2019	Harris et al.
10,245,033 B2	4/2019	Overmyer et al.	10,368,867 B2	8/2019	Harris et al.
10,245,037 B2	4/2019	Conklin et al.	10,368,876 B2	8/2019	Bhatnagar et al.
10,245,038 B2	4/2019	Hopkins et al.	10,368,894 B2	8/2019	Madan et al.
10,245,040 B2	4/2019	Milliman	10,368,903 B2	8/2019	Morales et al.
10,251,661 B2	4/2019	Collings et al.	10,376,263 B2	8/2019	Morgan et al.
10,251,725 B2	4/2019	Valentine et al.	10,376,305 B2	8/2019	Yates et al.
10,258,331 B2	4/2019	Shelton, IV et al.	10,376,337 B2	8/2019	Kilroy et al.
10,258,359 B2	4/2019	Kapadia	10,376,338 B2	8/2019	Taylor et al.
10,258,362 B2	4/2019	Conlon	10,378,893 B2	8/2019	Mankovskii
10,258,363 B2	4/2019	Worrell et al.	10,383,518 B2	8/2019	Abu-Tarif et al.
10,258,415 B2	4/2019	Harrah et al.	10,383,699 B2	8/2019	Kilroy et al.
10,258,418 B2	4/2019	Shelton, IV et al.	10,384,021 B2	8/2019	Koeth et al.
10,258,425 B2	4/2019	Mustufa et al.	10,386,990 B2	8/2019	Shikhman et al.
10,263,171 B2	4/2019	Wiener et al.	10,390,718 B2	8/2019	Chen et al.
10,265,004 B2	4/2019	Yamaguchi et al.	10,390,794 B2	8/2019	Kuroiwa et al.
10,265,035 B2	4/2019	Fehre et al.	10,390,825 B2	8/2019	Shelton, IV et al.
10,265,066 B2	4/2019	Measamer et al.	10,390,831 B2	8/2019	Holsten et al.
10,265,068 B2	4/2019	Harris et al.	10,390,895 B2	8/2019	Henderson et al.
10,265,072 B2	4/2019	Shelton, IV et al.	10,398,348 B2	9/2019	Osadchy et al.
10,265,090 B2	4/2019	Ingmanson et al.	10,398,434 B2	9/2019	Shelton, IV et al.
10,265,130 B2	4/2019	Hess et al.	10,398,517 B2	9/2019	Eckert et al.
10,271,840 B2	4/2019	Sapre	10,398,521 B2	9/2019	Itkowitz et al.
10,271,844 B2	4/2019	Valentine et al.	10,404,521 B2	9/2019	McChord et al.
10,271,850 B2	4/2019	Williams	10,404,801 B2	9/2019	Martch
10,271,851 B2	4/2019	Shelton, IV et al.	10,405,857 B2	9/2019	Shelton, IV et al.
D847,989 S	5/2019	Shelton, IV et al.	10,405,863 B2	9/2019	Wise et al.
10,278,698 B2	5/2019	Racenet	10,413,291 B2	9/2019	Worthington et al.
10,278,778 B2	5/2019	State et al.	10,413,293 B2	9/2019	Shelton, IV et al.
10,283,220 B2	5/2019	Azizian et al.	10,413,297 B2	9/2019	Harris et al.
10,285,694 B2	5/2019	Viola et al.	10,417,446 B2	9/2019	Takeyama
10,285,698 B2	5/2019	Cappola et al.	10,420,552 B2	9/2019	Shelton, IV et al.
10,285,700 B2	5/2019	Scheib	10,420,558 B2	9/2019	Nalagatla et al.
10,285,705 B2	5/2019	Shelton, IV et al.	10,420,559 B2	9/2019	Marczyk et al.
10,292,704 B2	5/2019	Harris et al.	10,420,620 B2	9/2019	Rockrohr
10,292,707 B2	5/2019	Shelton, IV et al.	10,420,865 B2	9/2019	Reasoner et al.
10,292,758 B2	5/2019	Boudreaux et al.	10,422,727 B2	9/2019	Pliskin
10,292,769 B1	5/2019	Yu	10,426,466 B2	10/2019	Contini et al.
10,292,771 B2	5/2019	Wood et al.	10,426,467 B2	10/2019	Miller et al.
10,293,129 B2	5/2019	Fox et al.	10,426,468 B2	10/2019	Contini et al.
10,299,792 B2	5/2019	Huitema et al.	10,426,471 B2	10/2019	Shelton, IV et al.
10,299,870 B2	5/2019	Connolly et al.	10,426,481 B2	10/2019	Aronhalt et al.
10,305,926 B2	5/2019	Mihan et al.	10,433,837 B2	10/2019	Worthington et al.
D850,617 S	6/2019	Shelton, IV et al.	10,433,844 B2	10/2019	Shelton, IV et al.
10,307,159 B2	6/2019	Harris et al.	10,433,849 B2	10/2019	Shelton, IV et al.
10,307,170 B2	6/2019	Parfett et al.	10,433,918 B2	10/2019	Shelton, IV et al.
10,307,199 B2	6/2019	Farritor et al.	10,441,279 B2	10/2019	Shelton, IV et al.
10,311,036 B1	6/2019	Hussam et al.	10,441,345 B2	10/2019	Aldridge et al.
10,313,137 B2	6/2019	Aarnio et al.	10,448,948 B2	10/2019	Shelton, IV et al.
10,314,577 B2	6/2019	Laurent et al.	10,448,950 B2	10/2019	Shelton, IV et al.
10,314,582 B2	6/2019	Shelton, IV et al.	10,456,137 B2	10/2019	Vendely et al.
10,321,907 B2	6/2019	Shelton, IV et al.	10,456,140 B2	10/2019	Shelton, IV et al.
10,321,964 B2	6/2019	Grover et al.	10,456,193 B2	10/2019	Yates et al.
10,327,764 B2	6/2019	Harris et al.	10,463,365 B2	11/2019	Williams
10,327,779 B2	6/2019	Richard et al.	10,463,367 B2	11/2019	Kostrzewski et al.
10,335,147 B2	7/2019	Rector et al.	10,463,371 B2	11/2019	Kostrzewski
10,335,149 B2	7/2019	Baxter, III et al.	10,463,436 B2	11/2019	Jackson et al.
10,335,180 B2	7/2019	Johnson et al.	10,470,762 B2	11/2019	Leimbach et al.
10,335,227 B2	7/2019	Heard	10,470,764 B2	11/2019	Baxter, III et al.
10,339,496 B2	7/2019	Matson et al.	10,470,768 B2	11/2019	Harris et al.
10,342,543 B2	7/2019	Shelton, IV et al.	10,470,791 B2	11/2019	Houser
			10,471,254 B2	11/2019	Sano et al.
			10,478,181 B2	11/2019	Shelton, IV et al.
			10,478,185 B2	11/2019	Nicholas
			10,478,189 B2	11/2019	Bear et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

10,478,190	B2	11/2019	Miller et al.	10,610,223	B2	4/2020	Wellman et al.
10,478,544	B2	11/2019	Friederichs et al.	10,610,224	B2	4/2020	Shelton, IV et al.
10,485,450	B2	11/2019	Gupta et al.	10,610,286	B2	4/2020	Wiener et al.
10,485,542	B2	11/2019	Shelton, IV et al.	10,610,313	B2	4/2020	Bailey et al.
10,485,543	B2	11/2019	Shelton, IV et al.	10,617,412	B2	4/2020	Shelton, IV et al.
10,492,783	B2	12/2019	Shelton, IV et al.	10,617,414	B2	4/2020	Shelton, IV et al.
10,492,784	B2	12/2019	Beardsley et al.	10,617,482	B2	4/2020	Houser et al.
10,492,785	B2	12/2019	Overmyer et al.	10,617,484	B2	4/2020	Kilroy et al.
10,496,788	B2	12/2019	Amarasingham et al.	10,624,635	B2	4/2020	Harris et al.
10,498,269	B2	12/2019	Zemlok et al.	10,624,667	B2	4/2020	Faller et al.
10,499,847	B2	12/2019	Latimer et al.	10,624,691	B2	4/2020	Wiener et al.
10,499,891	B2	12/2019	Chaplin et al.	10,631,423	B2	4/2020	Collins et al.
10,499,914	B2	12/2019	Huang et al.	10,631,858	B2	4/2020	Burbank
10,499,915	B2	12/2019	Aranyi	10,631,912	B2	4/2020	McFarlin et al.
10,499,994	B2	12/2019	Luks et al.	10,631,916	B2	4/2020	Horner et al.
10,507,068	B2	12/2019	Kopp et al.	10,631,917	B2	4/2020	Ineson
10,512,413	B2	12/2019	Schepis et al.	10,631,939	B2	4/2020	Dachs, II et al.
10,512,461	B2	12/2019	Gupta et al.	10,639,027	B2	5/2020	Shelton, IV et al.
10,512,499	B2	12/2019	McHenry et al.	10,639,034	B2	5/2020	Harris et al.
10,512,514	B2	12/2019	Nowlin et al.	10,639,035	B2	5/2020	Shelton, IV et al.
10,517,588	B2	12/2019	Gupta et al.	10,639,036	B2	5/2020	Yates et al.
10,517,595	B2	12/2019	Hunter et al.	10,639,037	B2	5/2020	Shelton, IV et al.
10,517,596	B2	12/2019	Hunter et al.	10,639,039	B2	5/2020	Vendely et al.
10,517,686	B2	12/2019	Vokrot et al.	10,639,098	B2	5/2020	Cosman et al.
10,524,789	B2	1/2020	Swayze et al.	10,639,111	B2	5/2020	Kopp
10,531,579	B2	1/2020	Hsiao et al.	10,639,185	B2	5/2020	Agrawal et al.
10,531,874	B2	1/2020	Morgan et al.	10,653,413	B2	5/2020	Worthington et al.
10,531,929	B2	1/2020	Widenhouse et al.	10,653,476	B2	5/2020	Ross
10,532,330	B2	1/2020	Diallo et al.	10,653,489	B2	5/2020	Kopp
10,536,617	B2	1/2020	Liang et al.	10,656,720	B1	5/2020	Holz
10,537,324	B2	1/2020	Shelton, IV et al.	10,660,705	B2	5/2020	Piron et al.
10,537,325	B2	1/2020	Bakos et al.	10,667,809	B2	6/2020	Bakos et al.
10,537,351	B2	1/2020	Shelton, IV et al.	10,667,810	B2	6/2020	Shelton, IV et al.
10,542,978	B2	1/2020	Chowaniec et al.	10,667,811	B2	6/2020	Harris et al.
10,542,979	B2	1/2020	Shelton, IV et al.	10,667,877	B2	6/2020	Kapadia
10,542,982	B2	1/2020	Beckman et al.	10,674,897	B2	6/2020	Levy
10,542,991	B2	1/2020	Shelton, IV et al.	10,675,021	B2	6/2020	Harris et al.
D876,466	S	2/2020	Kobayashi et al.	10,675,023	B2	6/2020	Cappola
10,548,504	B2	2/2020	Shelton, IV et al.	10,675,024	B2	6/2020	Shelton, IV et al.
10,548,612	B2	2/2020	Martinez et al.	10,675,025	B2	6/2020	Swayze et al.
10,548,673	B2	2/2020	Harris et al.	10,675,026	B2	6/2020	Harris et al.
10,552,574	B2	2/2020	Sweeney	10,675,035	B2	6/2020	Zingman
10,555,675	B2	2/2020	Satish et al.	10,675,100	B2	6/2020	Frushour
10,555,748	B2	2/2020	Yates et al.	10,675,104	B2	6/2020	Kapadia
10,555,750	B2	2/2020	Conlon et al.	10,677,764	B2	6/2020	Ross et al.
10,555,769	B2	2/2020	Worrell et al.	10,679,758	B2	6/2020	Fox et al.
10,561,422	B2	2/2020	Schellin et al.	10,682,136	B2	6/2020	Harris et al.
10,561,471	B2	2/2020	Nichogi	10,682,138	B2	6/2020	Shelton, IV et al.
10,561,753	B2	2/2020	Thompson et al.	10,686,805	B2	6/2020	Reybok, Jr. et al.
10,568,625	B2	2/2020	Harris et al.	10,687,806	B2	6/2020	Shelton, IV et al.
10,568,626	B2	2/2020	Shelton, IV et al.	10,687,809	B2	6/2020	Shelton, IV et al.
10,568,632	B2	2/2020	Miller et al.	10,687,810	B2	6/2020	Shelton, IV et al.
10,568,704	B2	2/2020	Savall et al.	10,687,884	B2	6/2020	Wiener et al.
10,575,868	B2	3/2020	Hall et al.	10,687,905	B2	6/2020	Kostrzewski
10,582,928	B2	3/2020	Hunter et al.	10,695,055	B2	6/2020	Shelton, IV et al.
10,582,931	B2	3/2020	Mujawar	10,695,081	B2	6/2020	Shelton, IV et al.
10,582,964	B2	3/2020	Weinberg et al.	10,695,134	B2	6/2020	Barral et al.
10,586,074	B2	3/2020	Rose et al.	10,702,270	B2	7/2020	Shelton, IV et al.
10,588,623	B2	3/2020	Schmid et al.	10,702,271	B2	7/2020	Aranyi et al.
10,588,625	B2	3/2020	Weaner et al.	10,709,446	B2	7/2020	Harris et al.
10,588,629	B2	3/2020	Malinouskas et al.	10,716,489	B2	7/2020	Kalvoy et al.
10,588,630	B2	3/2020	Shelton, IV et al.	10,716,615	B2	7/2020	Shelton, IV et al.
10,588,631	B2	3/2020	Shelton, IV et al.	10,716,639	B2	7/2020	Kapadia et al.
10,588,632	B2	3/2020	Shelton, IV et al.	10,717,194	B2	7/2020	Griffiths et al.
10,588,711	B2	3/2020	DiCarlo et al.	10,722,222	B2	7/2020	Aranyi
10,592,067	B2	3/2020	Merdan et al.	10,722,233	B2	7/2020	Wellman
10,595,844	B2	3/2020	Nawana et al.	10,722,292	B2	7/2020	Arya et al.
10,595,882	B2	3/2020	Parfett et al.	D893,717	S	8/2020	Messlerly et al.
10,595,887	B2	3/2020	Shelton, IV et al.	10,729,458	B2	8/2020	Stoddard et al.
10,595,930	B2	3/2020	Scheib et al.	10,729,509	B2	8/2020	Shelton, IV et al.
10,595,952	B2	3/2020	Forrest et al.	10,733,267	B2	8/2020	Pedersen
10,602,007	B2	3/2020	Takano	10,736,219	B2	8/2020	Seow et al.
10,602,848	B2	3/2020	Magana	10,736,616	B2	8/2020	Scheib et al.
10,603,036	B2	3/2020	Hunter et al.	10,736,628	B2	8/2020	Yates et al.
10,603,128	B2	3/2020	Zergiebel et al.	10,736,629	B2	8/2020	Shelton, IV et al.
				10,736,636	B2	8/2020	Baxter, III et al.
				10,736,705	B2	8/2020	Scheib et al.
				10,743,872	B2	8/2020	Leimbach et al.
				10,748,115	B2	8/2020	Laster et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

10,751,052 B2	8/2020	Stokes et al.	10,898,186 B2	1/2021	Bakos et al.
10,751,136 B2	8/2020	Farritor et al.	10,898,189 B2	1/2021	McDonald, II
10,751,768 B2	8/2020	Hersey et al.	10,898,256 B2	1/2021	Yates et al.
10,755,813 B2	8/2020	Shelton, IV et al.	10,898,280 B2	1/2021	Kopp
D896,379 S	9/2020	Shelton, IV et al.	10,898,622 B2	1/2021	Shelton, IV et al.
10,758,229 B2	9/2020	Shelton, IV et al.	10,902,944 B1	1/2021	Casey et al.
10,758,230 B2	9/2020	Shelton, IV et al.	10,903,685 B2	1/2021	Yates et al.
10,758,294 B2	9/2020	Jones	10,905,415 B2	2/2021	DiNardo et al.
10,758,310 B2	9/2020	Shelton, IV et al.	10,905,418 B2	2/2021	Shelton, IV et al.
10,765,376 B2	9/2020	Brown, III et al.	10,905,420 B2	2/2021	Jasemian et al.
10,765,424 B2	9/2020	Baxter, III et al.	10,912,559 B2	2/2021	Harris et al.
10,765,427 B2	9/2020	Shelton, IV et al.	10,912,567 B2	2/2021	Shelton, IV et al.
10,765,470 B2	9/2020	Yates et al.	10,912,580 B2	2/2021	Green et al.
10,772,630 B2	9/2020	Wixey	10,912,619 B2	2/2021	Jarc et al.
10,772,651 B2	9/2020	Shelton, IV et al.	10,916,415 B2	2/2021	Karancsi et al.
10,772,673 B2	9/2020	Allen, IV et al.	10,918,385 B2	2/2021	Overmyer et al.
10,772,688 B2	9/2020	Peine et al.	10,930,400 B2	2/2021	Robbins et al.
10,779,818 B2	9/2020	Zemlok et al.	D914,878 S	3/2021	Shelton, IV et al.
10,779,821 B2	9/2020	Harris et al.	10,932,705 B2	3/2021	Muhsin et al.
10,779,823 B2	9/2020	Shelton, IV et al.	10,932,772 B2	3/2021	Shelton, IV et al.
10,779,897 B2	9/2020	Rockrohr	10,932,784 B2	3/2021	Mozdzierz et al.
10,779,900 B2	9/2020	Pedros et al.	10,932,804 B2	3/2021	Scheib et al.
10,783,634 B2	9/2020	Nye et al.	10,932,806 B2	3/2021	Shelton, IV et al.
10,786,298 B2	9/2020	Johnson	10,932,872 B2	3/2021	Shelton, IV et al.
10,786,317 B2	9/2020	Zhou et al.	10,939,313 B2	3/2021	Eom et al.
10,786,327 B2	9/2020	Anderson et al.	10,943,454 B2	3/2021	Shelton, IV et al.
10,792,038 B2	10/2020	Becerra et al.	10,944,728 B2	3/2021	Wiener et al.
10,792,118 B2	10/2020	Prpa et al.	10,945,727 B2	3/2021	Shelton, IV et al.
10,792,422 B2	10/2020	Douglas et al.	10,950,982 B2	3/2021	Regnier et al.
10,799,304 B2	10/2020	Kapadia et al.	10,952,708 B2	3/2021	Scheib et al.
10,803,977 B2	10/2020	Sanmugalingham	10,954,935 B2	3/2021	O'Shea et al.
10,806,445 B2	10/2020	Penna et al.	10,959,727 B2	3/2021	Hunter et al.
10,806,453 B2	10/2020	Chen et al.	10,959,729 B2	3/2021	Ehrenfels et al.
10,806,454 B2	10/2020	Kopp	10,959,744 B2	3/2021	Shelton, IV et al.
10,806,499 B2	10/2020	Castaneda et al.	10,959,788 B2	3/2021	Grover et al.
10,806,506 B2	10/2020	Gaspedres et al.	10,960,150 B2	3/2021	Zergiebel et al.
10,806,532 B2	10/2020	Grubbs et al.	10,966,791 B2	4/2021	Harris et al.
10,813,638 B2	10/2020	Shelton, IV et al.	10,966,798 B2	4/2021	Tesar et al.
10,813,703 B2	10/2020	Swayze et al.	10,973,516 B2	4/2021	Shelton, IV et al.
10,818,383 B2	10/2020	Sharifi Sedeh et al.	10,973,517 B2	4/2021	Wixey
10,828,028 B2	11/2020	Harris et al.	10,973,520 B2	4/2021	Shelton, IV et al.
10,828,030 B2	11/2020	Weir et al.	10,973,682 B2	4/2021	Vezzu et al.
10,835,245 B2	11/2020	Swayze et al.	10,980,536 B2	4/2021	Weaner et al.
10,835,246 B2	11/2020	Shelton, IV et al.	10,980,537 B2	4/2021	Shelton, IV et al.
10,835,247 B2	11/2020	Shelton, IV et al.	10,980,560 B2	4/2021	Shelton, IV et al.
10,842,473 B2	11/2020	Scheib et al.	10,980,610 B2	4/2021	Rosenberg et al.
10,842,490 B2	11/2020	DiNardo et al.	10,987,102 B2	4/2021	Gonzalez et al.
10,842,492 B2	11/2020	Shelton, IV et al.	10,987,178 B2	4/2021	Shelton, IV et al.
10,842,522 B2	11/2020	Messerly et al.	10,992,698 B2	4/2021	Patel et al.
10,842,523 B2	11/2020	Shelton, IV et al.	10,993,715 B2	5/2021	Shelton, IV et al.
10,842,575 B2	11/2020	Panescu et al.	10,998,098 B2	5/2021	Greene et al.
10,842,897 B2	11/2020	Schwartz et al.	11,000,276 B2	5/2021	Shelton, IV et al.
D904,612 S	12/2020	Wynn et al.	11,000,278 B2	5/2021	Shelton, IV et al.
10,849,697 B2	12/2020	Yates et al.	11,007,004 B2	5/2021	Shelton, IV et al.
10,849,700 B2	12/2020	Kopp et al.	11,007,022 B2	5/2021	Shelton, IV et al.
10,856,768 B2	12/2020	Osadchy et al.	11,013,563 B2	5/2021	Shelton, IV et al.
10,856,867 B2	12/2020	Shelton, IV et al.	11,026,687 B2	6/2021	Shelton, IV et al.
10,856,868 B2	12/2020	Shelton, IV et al.	11,026,712 B2	6/2021	Shelton, IV et al.
10,856,870 B2	12/2020	Harris et al.	11,026,713 B2	6/2021	Stokes et al.
10,863,984 B2	12/2020	Shelton, IV et al.	11,026,751 B2	6/2021	Shelton, IV et al.
10,864,037 B2	12/2020	Mun et al.	11,039,834 B2	6/2021	Harris et al.
10,864,050 B2	12/2020	Tabandeh et al.	11,045,191 B2	6/2021	Shelton, IV et al.
10,872,684 B2	12/2020	McNutt et al.	11,045,192 B2	6/2021	Harris et al.
10,881,399 B2	1/2021	Shelton, IV et al.	11,045,197 B2	6/2021	Shelton, IV et al.
10,881,401 B2	1/2021	Baber et al.	11,045,591 B2	6/2021	Shelton, IV et al.
10,881,446 B2	1/2021	Strobl	11,051,817 B2	7/2021	Shelton, IV et al.
10,881,464 B2	1/2021	Odermatt et al.	11,051,836 B2	7/2021	Shelton, IV et al.
10,888,321 B2	1/2021	Shelton, IV et al.	11,051,873 B2	7/2021	Wiener et al.
10,888,322 B2	1/2021	Morgan et al.	11,051,876 B2	7/2021	Shelton, IV et al.
10,892,899 B2	1/2021	Shelton, IV et al.	11,056,244 B2	7/2021	Shelton, IV et al.
10,892,995 B2	1/2021	Shelton, IV et al.	11,058,423 B2	7/2021	Shelton, IV et al.
10,893,863 B2	1/2021	Shelton, IV et al.	11,058,498 B2	7/2021	Shelton, IV et al.
10,893,864 B2	1/2021	Harris et al.	11,058,501 B2	7/2021	Tokarchuk et al.
10,893,884 B2	1/2021	Stoddard et al.	11,064,997 B2	7/2021	Shelton, IV et al.
10,898,183 B2	1/2021	Shelton, IV et al.	11,069,012 B2	7/2021	Shelton, IV et al.
			11,071,560 B2	7/2021	Deck et al.
			11,071,595 B2	7/2021	Johnson et al.
			11,076,921 B2	8/2021	Shelton, IV et al.
			11,083,458 B2	8/2021	Harris et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

11,090,047 B2	8/2021	Shelton, IV et al.	11,298,148 B2	4/2022	Jayme et al.
11,090,048 B2	8/2021	Fanelli et al.	11,304,699 B2	4/2022	Shelton, IV et al.
11,090,075 B2	8/2021	Hunter et al.	11,304,720 B2	4/2022	Kimball et al.
11,096,688 B2	8/2021	Shelton, IV et al.	11,304,745 B2	4/2022	Shelton, IV et al.
11,096,693 B2	8/2021	Shelton, IV et al.	11,304,763 B2	4/2022	Shelton, IV et al.
11,100,631 B2	8/2021	Yates et al.	11,308,075 B2	4/2022	Shelton, IV et al.
11,103,268 B2	8/2021	Shelton, IV et al.	11,311,306 B2	4/2022	Shelton, IV et al.
11,109,866 B2	9/2021	Shelton, IV et al.	11,311,342 B2	4/2022	Parihar et al.
11,109,878 B2	9/2021	Shelton, IV et al.	D950,728 S	5/2022	Bakos et al.
11,114,195 B2	9/2021	Shelton, IV et al.	D952,144 S	5/2022	Boudreaux
11,116,485 B2	9/2021	Scheib et al.	11,317,915 B2	5/2022	Boudreaux et al.
11,123,070 B2	9/2021	Shelton, IV et al.	11,317,919 B2	5/2022	Shelton, IV et al.
11,129,611 B2	9/2021	Shelton, IV et al.	11,317,937 B2	5/2022	Nott et al.
11,129,634 B2	9/2021	Scheib et al.	11,324,557 B2	5/2022	Shelton, IV et al.
11,129,636 B2	9/2021	Shelton, IV et al.	11,331,100 B2	5/2022	Boudreaux et al.
11,129,669 B2	9/2021	Stulen et al.	11,331,101 B2	5/2022	Harris et al.
11,129,670 B2	9/2021	Shelton, IV et al.	11,337,746 B2	5/2022	Boudreaux
11,132,462 B2	9/2021	Shelton, IV et al.	11,344,326 B2	5/2022	Faller et al.
11,134,942 B2	10/2021	Harris et al.	11,350,959 B2	6/2022	Messerly et al.
11,141,160 B2	10/2021	Shelton, IV et al.	11,350,978 B2	6/2022	Henderson et al.
11,141,213 B2	10/2021	Yates et al.	2001/0056237 A1	12/2001	Cane et al.
11,147,607 B2	10/2021	Yates et al.	2002/0049551 A1	4/2002	Friedman et al.
11,160,551 B2	11/2021	Shelton, IV et al.	2002/0052616 A1	5/2002	Wiener et al.
11,160,605 B2	11/2021	Shelton, IV et al.	2002/0072746 A1	6/2002	Lingenfelder et al.
11,166,716 B2	11/2021	Shelton, IV et al.	2002/0138642 A1	9/2002	Miyazawa et al.
11,166,772 B2	11/2021	Shelton, IV et al.	2002/0169584 A1	11/2002	Fu et al.
11,179,150 B2	11/2021	Yates et al.	2003/0009111 A1	1/2003	Cory et al.
11,179,151 B2	11/2021	Shelton, IV et al.	2003/0018329 A1	1/2003	Hooven
11,179,155 B2	11/2021	Shelton, IV et al.	2003/0046109 A1	3/2003	Uchikubo
11,179,175 B2	11/2021	Houser et al.	2003/0069573 A1	4/2003	Kadhiresan et al.
11,179,204 B2	11/2021	Shelton, IV et al.	2003/0093503 A1	5/2003	Yamaki et al.
11,179,208 B2	11/2021	Yates et al.	2003/0114851 A1	6/2003	Truckai et al.
11,185,325 B2	11/2021	Shelton, IV et al.	2003/0130711 A1	7/2003	Pearson et al.
11,185,330 B2	11/2021	Huitema et al.	2003/0210812 A1	11/2003	Khamene et al.
11,191,539 B2	12/2021	Overmyer et al.	2003/0223877 A1	12/2003	Anstine et al.
11,191,540 B2	12/2021	Aronhalt et al.	2004/0015053 A1	1/2004	Bieger et al.
11,197,668 B2	12/2021	Shelton, IV et al.	2004/0078236 A1	4/2004	Stoodley et al.
11,202,570 B2	12/2021	Shelton, IV et al.	2004/0108825 A1	6/2004	Lee et al.
11,207,065 B2	12/2021	Harris et al.	2004/0199180 A1	10/2004	Knodel et al.
11,207,067 B2	12/2021	Shelton, IV et al.	2004/0199659 A1	10/2004	Ishikawa et al.
11,207,090 B2	12/2021	Shelton, IV et al.	2004/0206365 A1	10/2004	Knowlton
11,213,293 B2	1/2022	Worthington et al.	2004/0243147 A1	12/2004	Lipow
11,213,294 B2	1/2022	Shelton, IV et al.	2004/0243148 A1	12/2004	Wasielewski
11,213,359 B2	1/2022	Shelton, IV et al.	2004/0243435 A1	12/2004	Williams
11,218,822 B2	1/2022	Morgan et al.	2005/0020909 A1	1/2005	Moctezuma de la Barrera et al.
11,219,453 B2	1/2022	Shelton, IV et al.	2005/0020918 A1	1/2005	Wilk et al.
11,224,426 B2	1/2022	Shelton, IV et al.	2005/0021027 A1	1/2005	Shields et al.
11,229,436 B2	1/2022	Shelton, IV et al.	2005/0023324 A1	2/2005	Doll et al.
11,229,471 B2	1/2022	Shelton, IV et al.	2005/0063575 A1	3/2005	Ma et al.
11,234,756 B2	2/2022	Shelton, IV et al.	2005/0065438 A1	3/2005	Miller
11,241,230 B2	2/2022	Shelton, IV et al.	2005/0100867 A1	5/2005	Hilscher et al.
11,253,256 B2	2/2022	Harris et al.	2005/0131390 A1	6/2005	Heinrich et al.
11,253,315 B2	2/2022	Yates et al.	2005/0143759 A1	6/2005	Kelly
11,257,589 B2	2/2022	Shelton, IV et al.	2005/0149001 A1	7/2005	Uchikubo et al.
11,259,806 B2	3/2022	Shelton, IV et al.	2005/0149356 A1	7/2005	Cyr et al.
11,259,807 B2	3/2022	Shelton, IV et al.	2005/0165390 A1	7/2005	Mauti et al.
11,259,830 B2	3/2022	Nott et al.	2005/0182655 A1	8/2005	Merzlak et al.
11,266,409 B2	3/2022	Huitema et al.	2005/0192633 A1	9/2005	Montpetit
11,266,468 B2	3/2022	Shelton, IV et al.	2005/0203380 A1	9/2005	Sauer et al.
11,272,931 B2	3/2022	Boudreaux et al.	2005/0203384 A1	9/2005	Sati et al.
11,273,001 B2	3/2022	Shelton, IV et al.	2005/0203504 A1	9/2005	Wham et al.
11,278,280 B2	3/2022	Shelton, IV et al.	2005/0222631 A1	10/2005	Dalal et al.
11,278,281 B2	3/2022	Shelton, IV et al.	2005/0228425 A1*	10/2005	Boukhny A61B 17/320068 606/167
11,284,890 B2	3/2022	Nalagatla et al.	2005/0236474 A1	10/2005	Onuma et al.
11,284,936 B2	3/2022	Shelton, IV et al.	2005/0251233 A1	11/2005	Kanzius
11,291,440 B2	4/2022	Harris et al.	2005/0277913 A1	12/2005	McCary
11,291,441 B2	4/2022	Giordano et al.	2006/0020272 A1	1/2006	Gildenberg
11,291,444 B2	4/2022	Boudreaux et al.	2006/0025816 A1	2/2006	Shelton
11,291,445 B2	4/2022	Shelton, IV et al.	2006/0059018 A1	3/2006	Shiobara et al.
11,291,465 B2	4/2022	Parihar et al.	2006/0069388 A1	3/2006	Truckai et al.
11,291,495 B2	4/2022	Yates et al.	2006/0079872 A1	4/2006	Eggleston
11,291,510 B2	4/2022	Shelton, IV et al.	2006/0079874 A1	4/2006	Faller et al.
11,298,128 B2	4/2022	Messerly et al.	2006/0116908 A1	6/2006	Dew et al.
11,298,129 B2	4/2022	Bakos et al.	2006/0136622 A1	6/2006	Rouvelin et al.
11,298,130 B2	4/2022	Bakos et al.	2006/0184160 A1	8/2006	Ozaki et al.
			2006/0241399 A1	10/2006	Fabian
			2007/0010838 A1	1/2007	Shelton et al.
			2007/0016235 A1	1/2007	Tanaka et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

2007/0016979	A1	1/2007	Damaj et al.	2009/0326336	A1	12/2009	Lemke et al.
2007/0027459	A1	2/2007	Horvath et al.	2010/0036374	A1	2/2010	Ward
2007/0038080	A1	2/2007	Salisbury et al.	2010/0036405	A1	2/2010	Giordano et al.
2007/0049947	A1	3/2007	Menn et al.	2010/0038403	A1	2/2010	D'Arcangelo
2007/0066970	A1	3/2007	Ineson	2010/0057106	A1	3/2010	Sorrentino et al.
2007/0078678	A1	4/2007	DiSilvestro et al.	2010/0065604	A1	3/2010	Weng
2007/0084896	A1	4/2007	Doll et al.	2010/0069939	A1	3/2010	Konishi
2007/0156019	A1	7/2007	Larkin et al.	2010/0069942	A1	3/2010	Shelton, IV
2007/0167702	A1	7/2007	Hasser et al.	2010/0070417	A1	3/2010	Flynn et al.
2007/0168461	A1	7/2007	Moore	2010/0120266	A1	5/2010	Rimborg
2007/0173803	A1	7/2007	Wham et al.	2010/0132334	A1	6/2010	Duclos et al.
2007/0175955	A1	8/2007	Shelton et al.	2010/0137845	A1	6/2010	Ramstein et al.
2007/0179482	A1	8/2007	Anderson	2010/0137886	A1	6/2010	Zergiebel et al.
2007/0179508	A1	8/2007	Arndt	2010/0168561	A1	7/2010	Anderson
2007/0191713	A1	8/2007	Eichmann et al.	2010/0179831	A1	7/2010	Brown et al.
2007/0203744	A1	8/2007	Scholl	2010/0191100	A1	7/2010	Anderson et al.
2007/0225556	A1	9/2007	Ortiz et al.	2010/0198200	A1	8/2010	Horvath
2007/0225690	A1	9/2007	Sekiguchi et al.	2010/0198248	A1	8/2010	Vakharia
2007/0244478	A1	10/2007	Bahney	2010/0204717	A1	8/2010	Knodel
2007/0249990	A1	10/2007	Cosmescu	2010/0217991	A1	8/2010	Choi
2007/0270660	A1	11/2007	Caylor et al.	2010/0234996	A1	9/2010	Schreiber et al.
2007/0282195	A1	12/2007	Masini et al.	2010/0235689	A1	9/2010	Tian et al.
2007/0282321	A1	12/2007	Shah et al.	2010/0250571	A1	9/2010	Pierce et al.
2007/0282333	A1	12/2007	Fortson et al.	2010/0258327	A1	10/2010	Esenwein et al.
2007/0293218	A1	12/2007	Meylan et al.	2010/0292535	A1	11/2010	Paskar
2008/0013460	A1	1/2008	Allen et al.	2010/0292684	A1	11/2010	Cybulski et al.
2008/0015664	A1	1/2008	Podhajsky	2010/0301095	A1	12/2010	Shelton, IV et al.
2008/0015912	A1	1/2008	Rosenthal et al.	2011/0022032	A1	1/2011	Zemlok et al.
2008/0033404	A1	2/2008	Romoda et al.	2011/0036890	A1	2/2011	Ma
2008/0040151	A1	2/2008	Moore	2011/0071530	A1	3/2011	Carson
2008/0059658	A1	3/2008	Williams	2011/0077512	A1	3/2011	Boswell
2008/0077158	A1	3/2008	Haider et al.	2011/0087238	A1	4/2011	Wang et al.
2008/0083414	A1	4/2008	Messerges	2011/0092972	A1*	4/2011	Allen A61B 17/320092 606/45
2008/0091071	A1	4/2008	Kumar et al.	2011/0105895	A1	5/2011	Kornblau et al.
2008/0114212	A1	5/2008	Messerges	2011/0112569	A1	5/2011	Friedman et al.
2008/0114350	A1	5/2008	Park et al.	2011/0118708	A1	5/2011	Burbank et al.
2008/0129465	A1	6/2008	Rao	2011/0119075	A1	5/2011	Dhoble
2008/0140090	A1	6/2008	Aranyi et al.	2011/0125149	A1	5/2011	El-Galley et al.
2008/0177258	A1	7/2008	Govari et al.	2011/0152712	A1	6/2011	Cao et al.
2008/0177362	A1	7/2008	Phillips et al.	2011/0163147	A1	7/2011	Laurent et al.
2008/0200940	A1	8/2008	Eichmann et al.	2011/0166883	A1	7/2011	Palmer et al.
2008/0234708	A1	9/2008	Houser et al.	2011/0196398	A1	8/2011	Robertson et al.
2008/0255413	A1	10/2008	Zemlok et al.	2011/0237883	A1	9/2011	Chun
2008/0262654	A1	10/2008	Omori et al.	2011/0251612	A1	10/2011	Faller et al.
2008/0272172	A1	11/2008	Zemlok et al.	2011/0264000	A1	10/2011	Paul et al.
2008/0281301	A1	11/2008	DeBoer et al.	2011/0264078	A1	10/2011	Lipow et al.
2008/0281678	A1	11/2008	Keuls et al.	2011/0273465	A1	11/2011	Konishi et al.
2008/0296346	A1	12/2008	Shelton, IV et al.	2011/0278343	A1	11/2011	Knodel et al.
2008/0306759	A1	12/2008	Ilkin et al.	2011/0290024	A1	12/2011	Lefler
2008/0312953	A1	12/2008	Claus	2011/0295270	A1	12/2011	Giordano et al.
2009/0017910	A1	1/2009	Rofougaran et al.	2011/0306840	A1	12/2011	Allen et al.
2009/0030437	A1	1/2009	Houser et al.	2012/0021684	A1	1/2012	Schultz et al.
2009/0036750	A1	2/2009	Weinstein et al.	2012/0022519	A1	1/2012	Huang et al.
2009/0036794	A1	2/2009	Stubhaug et al.	2012/0029354	A1	2/2012	Mark et al.
2009/0036914	A1*	2/2009	Houser A61B 17/29 606/169	2012/0046662	A1	2/2012	Gilbert
2009/0043253	A1	2/2009	Podaima	2012/0059684	A1	3/2012	Hampapur et al.
2009/0046146	A1	2/2009	Hoyt	2012/0078247	A1	3/2012	Worrell et al.
2009/0048589	A1	2/2009	Takashino et al.	2012/0080336	A1	4/2012	Shelton, IV et al.
2009/0048611	A1	2/2009	Funda et al.	2012/0083786	A1	4/2012	Artale et al.
2009/0076409	A1	3/2009	Wu et al.	2012/0100517	A1	4/2012	Bowditch et al.
2009/0090763	A1	4/2009	Zemlok et al.	2012/0101488	A1	4/2012	Aldridge et al.
2009/0099866	A1	4/2009	Newman	2012/0116265	A1	5/2012	Houser et al.
2009/0114699	A1	5/2009	Viola	2012/0116381	A1	5/2012	Houser et al.
2009/0128084	A1	5/2009	Johnson et al.	2012/0116394	A1	5/2012	Timm et al.
2009/0182577	A1	7/2009	Squilla et al.	2012/0130217	A1	5/2012	Kauphusman et al.
2009/0206131	A1	8/2009	Weisenburgh, II et al.	2012/0145714	A1	6/2012	Farascioni et al.
2009/0217932	A1	9/2009	Voegele	2012/0172696	A1	7/2012	Kallback et al.
2009/0234352	A1	9/2009	Behnke et al.	2012/0190981	A1	7/2012	Harris et al.
2009/0259149	A1	10/2009	Tahara et al.	2012/0191091	A1	7/2012	Allen
2009/0259221	A1	10/2009	Tahara et al.	2012/0191162	A1	7/2012	Villa
2009/0299214	A1	12/2009	Wu et al.	2012/0197619	A1	8/2012	Namer Yelin et al.
2009/0306581	A1	12/2009	Claus	2012/0203785	A1	8/2012	Awada
2009/0307681	A1	12/2009	Armado et al.	2012/0211542	A1	8/2012	Racenet
2009/0326321	A1	12/2009	Jacobsen et al.	2012/0226150	A1	9/2012	Balicki et al.
				2012/0245958	A1	9/2012	Lawrence et al.
				2012/0253329	A1	10/2012	Zemlok et al.
				2012/0253847	A1	10/2012	Dell'Anno et al.
				2012/0265555	A1	10/2012	Cappuzzo et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

2012/0292367	A1	11/2012	Morgan et al.	2014/0263541	A1	9/2014	Leimbach et al.
2012/0319859	A1	12/2012	Taub et al.	2014/0263552	A1	9/2014	Hall et al.
2013/0001121	A1	1/2013	Metzger	2014/0275760	A1	9/2014	Lee et al.
2013/0006241	A1	1/2013	Takashino	2014/0276749	A1	9/2014	Johnson
2013/0008677	A1	1/2013	Huifu	2014/0287393	A1	9/2014	Kumar et al.
2013/0024213	A1	1/2013	Poon	2014/0303660	A1	10/2014	Boyden et al.
2013/0046182	A1	2/2013	Hegg et al.	2014/0303990	A1	10/2014	Schoenefeld et al.
2013/0046279	A1	2/2013	Niklewski et al.	2014/0337052	A1	11/2014	Pellini et al.
2013/0046295	A1	2/2013	Kerr et al.	2014/0364691	A1	12/2014	Krivopisk et al.
2013/0066647	A1	3/2013	Andrie et al.	2015/0006201	A1	1/2015	Pait et al.
2013/0090526	A1	4/2013	Suzuki et al.	2015/0025549	A1	1/2015	Kilroy et al.
2013/0093829	A1	4/2013	Rosenblatt et al.	2015/0032150	A1	1/2015	Ishida et al.
2013/0096597	A1	4/2013	Anand et al.	2015/0051452	A1	2/2015	Ciaccio
2013/0116218	A1	5/2013	Kaplan et al.	2015/0051598	A1	2/2015	Orszulak et al.
2013/0144284	A1	6/2013	Behnke, II et al.	2015/0051617	A1	2/2015	Takemura et al.
2013/0165776	A1	6/2013	Blomqvist	2015/0053737	A1	2/2015	Leimbach et al.
2013/0178853	A1	7/2013	Hyink et al.	2015/0053743	A1	2/2015	Yates et al.
2013/0191647	A1	7/2013	Ferrara, Jr. et al.	2015/0057675	A1	2/2015	Akeel et al.
2013/0193188	A1	8/2013	Shelton, IV et al.	2015/0066000	A1	3/2015	An et al.
2013/0197531	A1	8/2013	Boukhny et al.	2015/0070187	A1	3/2015	Wiesner et al.
2013/0206813	A1	8/2013	Nalagatla	2015/0073400	A1	3/2015	Sverdluk et al.
2013/0214025	A1	8/2013	Zemlok et al.	2015/0077528	A1	3/2015	Awdeh
2013/0253480	A1	9/2013	Kimball et al.	2015/0108198	A1	4/2015	Estrella
2013/0256373	A1	10/2013	Schmid et al.	2015/0133945	A1	5/2015	Dushyant et al.
2013/0267874	A1	10/2013	Marcotte et al.	2015/0136833	A1	5/2015	Shelton, IV et al.
2013/0268283	A1	10/2013	Vann et al.	2015/0140982	A1	5/2015	Postrel
2013/0277410	A1	10/2013	Fernandez et al.	2015/0145682	A1	5/2015	Harris
2013/0296908	A1*	11/2013	Schulte A61B 17/320068 606/169	2015/0148830	A1	5/2015	Stulen et al.
2013/0317837	A1	11/2013	Ballantyne et al.	2015/0157354	A1	6/2015	Bales, Jr. et al.
2013/0321425	A1	12/2013	Greene et al.	2015/0173673	A1	6/2015	Toth et al.
2013/0325809	A1	12/2013	Kim et al.	2015/0173756	A1	6/2015	Baxter, III et al.
2013/0331873	A1	12/2013	Ross et al.	2015/0182220	A1	7/2015	Yates et al.
2013/0331874	A1*	12/2013	Ross G01N 29/44 702/19	2015/0182220	A1	7/2015	Shelton, IV et al.
2013/0331875	A1	12/2013	Ross et al.	2015/0196295	A1	7/2015	Lee
2014/0001231	A1	1/2014	Shelton, IV et al.	2015/0199109	A1	7/2015	Kim et al.
2014/0001234	A1	1/2014	Shelton, IV et al.	2015/0202014	A1	7/2015	Kim et al.
2014/0005640	A1	1/2014	Shelton, IV et al.	2015/0208934	A1	7/2015	Sztrubel et al.
2014/0006132	A1	1/2014	Barker	2015/0223725	A1	8/2015	Engel et al.
2014/0009894	A1	1/2014	Yu	2015/0223868	A1	8/2015	Brandt et al.
2014/0013565	A1	1/2014	MacDonald et al.	2015/0237502	A1	8/2015	Schmidt et al.
2014/0018788	A1	1/2014	Engelman et al.	2015/0272557	A1	10/2015	Overmyer et al.
2014/0029411	A1	1/2014	Nayak et al.	2015/0272571	A1	10/2015	Leimbach et al.
2014/0033926	A1	2/2014	Fassel et al.	2015/0272580	A1	10/2015	Leimbach et al.
2014/0035762	A1	2/2014	Shelton, IV et al.	2015/0272582	A1	10/2015	Leimbach et al.
2014/0066700	A1	3/2014	Wilson et al.	2015/0272694	A1	10/2015	Charles
2014/0073893	A1	3/2014	Bencini	2015/0282821	A1	10/2015	Look et al.
2014/0074076	A1	3/2014	Gertner	2015/0297200	A1	10/2015	Fitzsimmons et al.
2014/0081255	A1	3/2014	Johnson et al.	2015/0297222	A1	10/2015	Huitema et al.
2014/0081659	A1	3/2014	Nawana et al.	2015/0297228	A1	10/2015	Huitema et al.
2014/0084949	A1	3/2014	Smith et al.	2015/0297311	A1	10/2015	Tesar
2014/0087999	A1	3/2014	Kaplan et al.	2015/0302157	A1	10/2015	Collar et al.
2014/0092089	A1	4/2014	Kasuya et al.	2015/0310174	A1	10/2015	Coudert et al.
2014/0107697	A1	4/2014	Patani et al.	2015/0313538	A1	11/2015	Bechtel et al.
2014/0108035	A1	4/2014	Akbay et al.	2015/0317899	A1	11/2015	Dumbauld et al.
2014/0108983	A1	4/2014	William et al.	2015/0320423	A1	11/2015	Aranyi
2014/0121669	A1	5/2014	Claus	2015/0324114	A1	11/2015	Hurley et al.
2014/0148729	A1	5/2014	Schmitz et al.	2015/0328474	A1	11/2015	Flyash et al.
2014/0148803	A1	5/2014	Taylor	2015/0332003	A1	11/2015	Stamm et al.
2014/0163359	A1	6/2014	Sholev et al.	2015/0332196	A1	11/2015	Stiller et al.
2014/0166724	A1	6/2014	Schellin et al.	2015/0335344	A1	11/2015	Aljuri et al.
2014/0171778	A1	6/2014	Tsusaka et al.	2015/0374259	A1	12/2015	Garbey et al.
2014/0187856	A1	7/2014	Holoien et al.	2016/0000437	A1	1/2016	Giordano et al.
2014/0188440	A1	7/2014	Donhowe et al.	2016/0001411	A1	1/2016	Alberti
2014/0194864	A1	7/2014	Martin et al.	2016/0015471	A1	1/2016	Piron et al.
2014/0195052	A1	7/2014	Tsusaka et al.	2016/0022374	A1	1/2016	Haider et al.
2014/0204190	A1	7/2014	Rosenblatt, III et al.	2016/0034648	A1	2/2016	Mohlenbrock et al.
2014/0226572	A1	8/2014	Thota et al.	2016/0038253	A1	2/2016	Piron et al.
2014/0243799	A1	8/2014	Parihar	2016/0048780	A1	2/2016	Sethumadhavan et al.
2014/0243809	A1	8/2014	Gelfand et al.	2016/0058439	A1	3/2016	Shelton, IV et al.
2014/0243811	A1	8/2014	Reschke et al.	2016/0066913	A1	3/2016	Swayze et al.
2014/0246475	A1	9/2014	Hall et al.	2016/0106516	A1	4/2016	Mesallum
2014/0249557	A1	9/2014	Koch et al.	2016/0106934	A1	4/2016	Hiraga et al.
2014/0252064	A1	9/2014	Mozdzierz et al.	2016/0121143	A1	5/2016	Mumaw et al.
				2016/0157717	A1	6/2016	Gaster
				2016/0158468	A1	6/2016	Tang et al.
				2016/0174998	A1	6/2016	Lal et al.
				2016/0175025	A1	6/2016	Strobl
				2016/0180045	A1	6/2016	Syed
				2016/0184054	A1	6/2016	Lowe
				2016/0192960	A1	7/2016	Bueno et al.

(56)	References Cited						
	U.S. PATENT DOCUMENTS						
2016/0206202	A1	7/2016	Frangioni	2017/0303984	A1	10/2017	Malackowski
2016/0206362	A1	7/2016	Mehta et al.	2017/0304007	A1	10/2017	Piron et al.
2016/0224760	A1	8/2016	Petak et al.	2017/0304020	A1	10/2017	Ng et al.
2016/0225551	A1	8/2016	Shedletsky	2017/0311777	A1	11/2017	Hirayama et al.
2016/0228061	A1	8/2016	Kallback et al.	2017/0312456	A1	11/2017	Phillips
2016/0228204	A1	8/2016	Quaid et al.	2017/0325876	A1	11/2017	Nakadate et al.
2016/0235303	A1	8/2016	Fleming et al.	2017/0325878	A1	11/2017	Messerly et al.
2016/0242836	A1	8/2016	Eggers et al.	2017/0337043	A1	11/2017	Brincat et al.
2016/0249910	A1	9/2016	Shelton, IV et al.	2017/0360358	A1	12/2017	Amiot et al.
2016/0249920	A1	9/2016	Gupta et al.	2017/0360499	A1	12/2017	Greep et al.
2016/0278841	A1	9/2016	Panescu et al.	2017/0367583	A1	12/2017	Black et al.
2016/0287312	A1	10/2016	Tegg et al.	2017/0367754	A1	12/2017	Narisawa
2016/0287316	A1	10/2016	Worrell et al.	2017/0367771	A1	12/2017	Tako et al.
2016/0287912	A1	10/2016	Warnking	2017/0367772	A1	12/2017	Gunn et al.
2016/0292456	A1	10/2016	Dubey et al.	2017/0370710	A1	12/2017	Chen et al.
2016/0296246	A1	10/2016	Schaller	2018/0008359	A1	1/2018	Randle
2016/0302210	A1	10/2016	Thornton et al.	2018/0011983	A1	1/2018	Zuhars et al.
2016/0310055	A1	10/2016	Zand et al.	2018/0021058	A1	1/2018	Meglan
2016/0310204	A1	10/2016	McHenry et al.	2018/0042659	A1	2/2018	Rupp et al.
2016/0314716	A1	10/2016	Grubbs	2018/0050196	A1	2/2018	Pawsey et al.
2016/0314717	A1	10/2016	Grubbs	2018/0052971	A1	2/2018	Hanina et al.
2016/0321400	A1	11/2016	Durrant et al.	2018/0065248	A1	3/2018	Barral et al.
2016/0323283	A1	11/2016	Kang et al.	2018/0078170	A1	3/2018	Panescu et al.
2016/0331460	A1	11/2016	Cheatham, III et al.	2018/0098816	A1	4/2018	Govari et al.
2016/0342753	A1	11/2016	Feazell	2018/0108438	A1	4/2018	Ryan et al.
2016/0342916	A1	11/2016	Arceneaux et al.	2018/0116662	A1	5/2018	Shelton, IV et al.
2016/0345857	A1	12/2016	Jensrud et al.	2018/0116735	A1	5/2018	Tierney et al.
2016/0350490	A1	12/2016	Martinez et al.	2018/0122506	A1	5/2018	Grantcharov et al.
2016/0354162	A1	12/2016	Yen et al.	2018/0132895	A1	5/2018	Silver
2016/0361070	A1	12/2016	Ardel et al.	2018/0144243	A1	5/2018	Hsieh et al.
2016/0367305	A1	12/2016	Hareland	2018/0153574	A1	6/2018	Faller et al.
2016/0367401	A1	12/2016	Claus	2018/0153632	A1	6/2018	Tokarchuk et al.
2016/0374723	A1	12/2016	Frankhouser et al.	2018/0154297	A1	6/2018	Maletich et al.
2016/0374762	A1	12/2016	Case et al.	2018/0161062	A1	6/2018	Kaga et al.
2016/0379504	A1	12/2016	Bailey et al.	2018/0161716	A1	6/2018	Li et al.
2017/0027603	A1	2/2017	Pandey	2018/0165780	A1	6/2018	Romeo
2017/0042604	A1	2/2017	McFarland et al.	2018/0168575	A1	6/2018	Simms et al.
2017/0068792	A1	3/2017	Reiner	2018/0168577	A1	6/2018	Aronhalt et al.
2017/0079530	A1	3/2017	DiMaio et al.	2018/0168579	A1	6/2018	Aronhalt et al.
2017/0079730	A1	3/2017	Azizian et al.	2018/0168592	A1	6/2018	Overmyer et al.
2017/0086829	A1	3/2017	Vendely et al.	2018/0168598	A1	6/2018	Shelton, IV et al.
2017/0086930	A1	3/2017	Thompson et al.	2018/0168608	A1	6/2018	Shelton, IV et al.
2017/0105754	A1	4/2017	Boudreaux et al.	2018/0168609	A1	6/2018	Fanelli et al.
2017/0116873	A1	4/2017	Lendvay et al.	2018/0168610	A1	6/2018	Shelton, IV et al.
2017/0127499	A1	5/2017	Unoson et al.	2018/0168615	A1	6/2018	Shelton, IV et al.
2017/0132374	A1	5/2017	Lee et al.	2018/0168618	A1	6/2018	Scott et al.
2017/0132385	A1	5/2017	Hunter et al.	2018/0168619	A1	6/2018	Scott et al.
2017/0132785	A1	5/2017	Wshah et al.	2018/0168623	A1	6/2018	Simms et al.
2017/0143284	A1	5/2017	Sehnert et al.	2018/0168625	A1	6/2018	Posada et al.
2017/0164997	A1	6/2017	Johnson et al.	2018/0168633	A1	6/2018	Shelton, IV et al.
2017/0165008	A1	6/2017	Finley	2018/0168647	A1	6/2018	Shelton, IV et al.
2017/0165012	A1	6/2017	Chaplin et al.	2018/0168648	A1	6/2018	Shelton, IV et al.
2017/0172550	A1	6/2017	Mukherjee et al.	2018/0168650	A1	6/2018	Shelton, IV et al.
2017/0172565	A1	6/2017	Heneveld	2018/0177383	A1	6/2018	Noonan et al.
2017/0172614	A1	6/2017	Scheib et al.	2018/0182475	A1	6/2018	Cossler et al.
2017/0173262	A1	6/2017	Veltz	2018/0193579	A1	7/2018	Hanrahan et al.
2017/0177807	A1	6/2017	Fabian	2018/0206884	A1	7/2018	Beaupre
2017/0196583	A1	7/2017	Sugiyama	2018/0206905	A1	7/2018	Batchelor et al.
2017/0202595	A1	7/2017	Shelton, IV	2018/0211726	A1	7/2018	Courtemanche et al.
2017/0202608	A1	7/2017	Shelton, IV et al.	2018/0214025	A1	8/2018	Homyk et al.
2017/0215944	A1	8/2017	Keffeler	2018/0221005	A1	8/2018	Hamel et al.
2017/0224332	A1	8/2017	Hunter et al.	2018/0221598	A1	8/2018	Silver
2017/0224428	A1	8/2017	Kopp	2018/0228557	A1	8/2018	Darisse et al.
2017/0231628	A1	8/2017	Shelton, IV et al.	2018/0233222	A1	8/2018	Daley et al.
2017/0245809	A1	8/2017	Ma et al.	2018/0235719	A1	8/2018	Jarc
2017/0249431	A1	8/2017	Shelton, IV et al.	2018/0235722	A1	8/2018	Baghdadi et al.
2017/0249432	A1	8/2017	Grantcharov	2018/0242967	A1	8/2018	Meade
2017/0262604	A1	9/2017	Francois	2018/0247128	A1	8/2018	Alvi et al.
2017/0265864	A1	9/2017	Hessler et al.	2018/0247711	A1	8/2018	Terry
2017/0265943	A1	9/2017	Sela et al.	2018/0250086	A1	9/2018	Grubbs
2017/0273715	A1	9/2017	Piron et al.	2018/0263699	A1	9/2018	Murphy et al.
2017/0281186	A1	10/2017	Shelton, IV et al.	2018/0263710	A1	9/2018	Sakaguchi et al.
2017/0296173	A1	10/2017	Shelton, IV et al.	2018/0268320	A1	9/2018	Shekhar
2017/0296185	A1	10/2017	Swensgard et al.	2018/0271520	A1	9/2018	Shelton, IV et al.
2017/0296213	A1	10/2017	Swensgard et al.	2018/0271603	A1	9/2018	Nir et al.
				2018/0289427	A1	10/2018	Griffiths et al.
				2018/0294060	A1	10/2018	Kassab
				2018/0296286	A1	10/2018	Peine et al.
				2018/0303552	A1	10/2018	Ryan et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

2018/0304471	A1	10/2018	Tokuchi	2019/0201026	A1	7/2019	Shelton, IV et al.
2018/0310986	A1	11/2018	Batchelor et al.	2019/0201027	A1	7/2019	Shelton, IV et al.
2018/0315492	A1	11/2018	Bishop et al.	2019/0201028	A1	7/2019	Shelton, IV et al.
2018/0333188	A1	11/2018	Nott et al.	2019/0201029	A1	7/2019	Shelton, IV et al.
2018/0333207	A1	11/2018	Moctezuma De la Barrera	2019/0201030	A1	7/2019	Shelton, IV et al.
2018/0333209	A1	11/2018	Frushour et al.	2019/0201033	A1	7/2019	Yates et al.
2018/0353186	A1	12/2018	Mozdzierz et al.	2019/0201034	A1	7/2019	Shelton, IV et al.
2018/0357383	A1	12/2018	Allen et al.	2019/0201036	A1	7/2019	Nott et al.
2018/0360456	A1	12/2018	Shelton, IV et al.	2019/0201038	A1	7/2019	Yates et al.
2018/0366213	A1	12/2018	Fidone et al.	2019/0201039	A1	7/2019	Widenhouse et al.
2018/0368930	A1	12/2018	Esterberg et al.	2019/0201040	A1	7/2019	Messerly et al.
2019/0000569	A1	1/2019	Crawford et al.	2019/0201042	A1	7/2019	Nott et al.
2019/0001079	A1	1/2019	Zergiebel et al.	2019/0201043	A1	7/2019	Shelton, IV et al.
2019/0005641	A1	1/2019	Yamamoto	2019/0201044	A1	7/2019	Shelton, IV et al.
2019/0006047	A1	1/2019	Gorek et al.	2019/0201045	A1	7/2019	Yates et al.
2019/0025040	A1	1/2019	Andreason et al.	2019/0201046	A1	7/2019	Shelton, IV et al.
2019/0036688	A1	1/2019	Wasily et al.	2019/0201047	A1	7/2019	Yates et al.
2019/0038335	A1	2/2019	Mohr et al.	2019/0201073	A1	7/2019	Nott et al.
2019/0038364	A1	2/2019	Enoki	2019/0201075	A1	7/2019	Shelton, IV et al.
2019/0046198	A1	2/2019	Stokes et al.	2019/0201080	A1	7/2019	Messerly et al.
2019/0053801	A1	2/2019	Wixey et al.	2019/0201083	A1	7/2019	Shelton, IV et al.
2019/0053866	A1	2/2019	Seow et al.	2019/0201085	A1	7/2019	Shelton, IV et al.
2019/0059986	A1	2/2019	Shelton, IV et al.	2019/0201087	A1	7/2019	Shelton, IV et al.
2019/0069949	A1	3/2019	Vrba et al.	2019/0201090	A1	7/2019	Shelton, IV et al.
2019/0069964	A1	3/2019	Hagn	2019/0201091	A1	7/2019	Yates et al.
2019/0069966	A1	3/2019	Petersen et al.	2019/0201102	A1	7/2019	Shelton, IV et al.
2019/0070550	A1	3/2019	Lalomia et al.	2019/0201104	A1	7/2019	Shelton, IV et al.
2019/0070731	A1	3/2019	Bowling et al.	2019/0201112	A1	7/2019	Wiener et al.
2019/0083190	A1	3/2019	Graves et al.	2019/0201113	A1	7/2019	Shelton, IV et al.
2019/0087544	A1	3/2019	Peterson	2019/0201115	A1	7/2019	Shelton, IV et al.
2019/0099221	A1	4/2019	Schmidt et al.	2019/0201118	A1	7/2019	Shelton, IV et al.
2019/0104919	A1	4/2019	Shelton, IV et al.	2019/0201120	A1	7/2019	Shelton, IV et al.
2019/0110828	A1	4/2019	Despatie	2019/0201123	A1	7/2019	Shelton, IV et al.
2019/0110855	A1	4/2019	Barral et al.	2019/0201124	A1	7/2019	Shelton, IV et al.
2019/0115108	A1	4/2019	Hegedus et al.	2019/0201126	A1	7/2019	Shelton, IV et al.
2019/0125320	A1	5/2019	Shelton, IV et al.	2019/0201127	A1	7/2019	Shelton, IV et al.
2019/0125335	A1	5/2019	Shelton, IV et al.	2019/0201128	A1	7/2019	Yates et al.
2019/0125336	A1	5/2019	Deck et al.	2019/0201130	A1	7/2019	Shelton, IV et al.
2019/0125338	A1	5/2019	Shelton, IV et al.	2019/0201136	A1	7/2019	Shelton, IV et al.
2019/0125358	A1	5/2019	Shelton, IV et al.	2019/0201137	A1	7/2019	Shelton, IV et al.
2019/0125361	A1	5/2019	Shelton, IV et al.	2019/0201139	A1	7/2019	Shelton, IV et al.
2019/0125377	A1	5/2019	Shelton, IV	2019/0201140	A1	7/2019	Yates et al.
2019/0125378	A1	5/2019	Shelton, IV et al.	2019/0201142	A1	7/2019	Shelton, IV et al.
2019/0125388	A1	5/2019	Shelton, IV et al.	2019/0201146	A1	7/2019	Shelton, IV et al.
2019/0125430	A1	5/2019	Shelton, IV et al.	2019/0201158	A1	7/2019	Shelton, IV et al.
2019/0125431	A1	5/2019	Shelton, IV et al.	2019/0201594	A1	7/2019	Shelton, IV et al.
2019/0125432	A1	5/2019	Shelton, IV et al.	2019/0204201	A1	7/2019	Shelton, IV et al.
2019/0125454	A1	5/2019	Stokes et al.	2019/0205001	A1	7/2019	Messerly et al.
2019/0125455	A1	5/2019	Shelton, IV et al.	2019/0205567	A1	7/2019	Shelton, IV et al.
2019/0125458	A1	5/2019	Shelton, IV et al.	2019/0206003	A1	7/2019	Harris et al.
2019/0125459	A1	5/2019	Shelton, IV et al.	2019/0206551	A1	7/2019	Yates et al.
2019/0125476	A1	5/2019	Shelton, IV et al.	2019/0206555	A1	7/2019	Morgan et al.
2019/0133703	A1	5/2019	Seow et al.	2019/0206561	A1	7/2019	Shelton, IV et al.
2019/0142535	A1	5/2019	Seow et al.	2019/0206562	A1	7/2019	Shelton, IV et al.
2019/0145942	A1	5/2019	Dutriez et al.	2019/0206564	A1	7/2019	Shelton, IV et al.
2019/0150975	A1	5/2019	Kawasaki et al.	2019/0206565	A1	7/2019	Shelton, IV
2019/0159778	A1	5/2019	Shelton, IV et al.	2019/0206569	A1	7/2019	Shelton, IV et al.
2019/0163875	A1	5/2019	Allen et al.	2019/0208641	A1	7/2019	Yates et al.
2019/0167296	A1	6/2019	Tsubuku et al.	2019/0224434	A1	7/2019	Silver et al.
2019/0192157	A1	6/2019	Scott et al.	2019/0254759	A1	8/2019	Azizian
2019/0200844	A1	7/2019	Shelton, IV et al.	2019/0261984	A1	8/2019	Nelson et al.
2019/0200905	A1	7/2019	Shelton, IV et al.	2019/0269476	A1	9/2019	Bowling et al.
2019/0200906	A1	7/2019	Shelton, IV et al.	2019/0272917	A1	9/2019	Couture et al.
2019/0200977	A1	7/2019	Shelton, IV et al.	2019/0274662	A1	9/2019	Rockman et al.
2019/0200980	A1	7/2019	Shelton, IV et al.	2019/0274705	A1	9/2019	Sawhney et al.
2019/0200981	A1	7/2019	Harris et al.	2019/0274706	A1	9/2019	Nott et al.
2019/0200984	A1	7/2019	Shelton, IV et al.	2019/0274707	A1	9/2019	Sawhney et al.
2019/0200985	A1	7/2019	Shelton, IV et al.	2019/0274709	A1	9/2019	Scoggins
2019/0200986	A1	7/2019	Shelton, IV et al.	2019/0274710	A1	9/2019	Black
2019/0200987	A1	7/2019	Shelton, IV et al.	2019/0274711	A1	9/2019	Scoggins et al.
2019/0200988	A1	7/2019	Shelton, IV	2019/0274713	A1	9/2019	Scoggins et al.
2019/0200997	A1	7/2019	Shelton, IV et al.	2019/0274714	A1	9/2019	Cuti et al.
2019/0201021	A1	7/2019	Shelton, IV et al.	2019/0274718	A1	9/2019	Denzinger et al.
2019/0201024	A1	7/2019	Shelton, IV et al.	2019/0274719	A1	9/2019	Stulen
2019/0201025	A1	7/2019	Shelton, IV et al.	2019/0274720	A1	9/2019	Gee et al.
				2019/0274749	A1	9/2019	Brady et al.
				2019/0274752	A1	9/2019	Denzinger et al.
				2019/0278262	A1	9/2019	Taylor et al.
				2019/0282311	A1	9/2019	Nowlin et al.

(56)

References Cited

FOREIGN PATENT DOCUMENTS

EP	3056923	A1	8/2016
EP	3095399	A2	11/2016
EP	3120781	A2	1/2017
EP	3135225	A2	3/2017
EP	3141181	A1	3/2017
FR	2838234	A1	10/2003
GB	2037167	A1	7/1980
GB	2509523	A	7/2014
JP	S5373315	A	6/1978
JP	2001029353	A	2/2001
JP	2007123394	A	5/2007
JP	2010057642	A	3/2010
JP	2017513561	A	6/2017
KR	20140104587	A	8/2014
KR	101587721	B1	1/2016
WO	WO-9734533	A1	9/1997
WO	WO-0024322	A1	5/2000
WO	WO-0108578	A1	2/2001
WO	WO-0112089	A1	2/2001
WO	WO-0120892	A2	3/2001
WO	WO-03079909	A2	10/2003
WO	WO-2007137304	A2	11/2007
WO	WO-2008053485	A1	5/2008
WO	WO-2008056618	A2	5/2008
WO	WO-2008069816	A1	6/2008
WO	WO-2008147555	A2	12/2008
WO	WO-2011112931	A1	9/2011
WO	WO-2013143573	A1	10/2013
WO	WO-2014031800	A1	2/2014
WO	WO-2014071184	A1	5/2014
WO	WO-2014134196	A1	9/2014
WO	WO-2015129395	A1	9/2015
WO	WO-2016100719	A1	6/2016
WO	WO-2016118752	A1	7/2016
WO	WO-2016206015	A1	12/2016
WO	WO-2017011382	A1	1/2017
WO	WO-2017011646	A1	1/2017
WO	WO-2017058617	A1	4/2017
WO	WO-2017058695	A1	4/2017
WO	WO-2017151996	A1	9/2017
WO	WO-2017189317	A1	11/2017
WO	WO-2017205308	A1	11/2017
WO	WO-2017210499	A1	12/2017
WO	WO-2017210501	A1	12/2017
WO	WO-2018116247	A1	6/2018
WO	WO-2018152141	A1	8/2018
WO	WO-2018176414	A1	10/2018

OTHER PUBLICATIONS

Engel et al. "A safe robot system for craniofacial surgery", 2013 IEEE International Conference on Robotics and Automation (ICRA); May 6-10, 2013; Karlsruhe, Germany, vol. 2, Jan. 1, 2001, pp. 2020-2024.

Miller, et al., "Impact of Powered and Tissue-Specific Endoscopic Stapling Technology on Clinical and Economic Outcomes of Video-Assisted Thoracic Surgery Lobectomy Procedures: A Retrospective, Observational Study," Article, Apr. 2018, pp. 707-723, vol. 35 (Issue 5), *Advances in Therapy*.

Choi et al., "A haptic augmented reality surgeon console for a laparoscopic surgery robot system, 2013, IEEE, p. 355-357 (Year: 2013).

"Surgical instrumentation: the true cost of instrument trays and a potential strategy for optimization"; Mhlaba et al.; Sep. 23, 2015 (Year: 2015).

Shi et al., "An intuitive control console for robotic surgery system, 2014, IEEE, p. 404-407 (Year: 2014).

Slocinski et al., "Distance measure for impedance spectra for quantified evaluations," *Lecture Notes on Impedance Spectroscopy*, vol. 3, Taylor and Francis Group (Jul. 2012)—Book Not Attached.

Zoccali, Bruno, "A Method for Approximating Component Temperatures at Altitude Conditions Based on CFD Analysis at Sea

Level Conditions," (white paper), www.tdmginc.com, Dec. 6, 2018 (9 pages).

Flores et al., "Large-scale Offloading in the Internet of Things," 2017 IEEE International Conference on Pervasive Computing and Communications Workshops (PERCOM Workshops), IEEE, pp. 479-484, Mar. 13, 2017.

Kalantarian et al., "Computation Offloading for Real-Time Health-Monitoring Devices," 2016 38th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EBMC), IEEE, pp. 4971-4974, Aug. 16, 2016.

Yuyi Mao et al., "A Survey on Mobile Edge Computing: The Communication Perspective," *IEEE Communications Surveys & Tutorials*, pp. 2322-2358, Jun. 13, 2017.

Khazaei et al., "Health Informatics for Neonatal Intensive Care Units: An Analytical Modeling Perspective, *IEEE Journal of Translational Engineering in Health and Medicine*," vol. 3, pp. 1-9, Oct. 21, 2015.

Benkmann et al., "Concept of iterative optimization of minimally invasive surgery," 2017 22nd International Conference on Methods and Models in Automation and Robotics (MMAR), IEEE pp. 443-446, Aug. 28, 2017.

Trautman, Peter, "Breaking the Human-Robot Deadlock: Surpassing Shared Control Performance Limits with Sparse Human-Robot Interaction," *Robotics: Science and Systems XIII*, pp. 1-10, Jul. 12, 2017.

Miksch et al., "Utilizing temporal data abstraction for data validation and therapy planning for artificially ventilated newborn infants," *Artificial Intelligence in Medicine*, vol. 8, No. 6, pp. 543-576 (1996).

Horn et al., "Effective data validation of high-frequency data: Time-point-time-interval-, and trend-based methods," *Computers in Biology and Medic*, New York, NY, vol. 27, No. 5, pp. 389-409 (1997).

Stacey et al., "Temporal abstraction in intelligent clinical data analysis: A survey," *Artificial Intelligence in Medicine*, vol. 39, No. 1, pp. 1-24 (2006).

Yang et al., "A dynamic strategy for packet scheduling and bandwidth allocation based on channel quality in IEEE 802.16e OFDMA system," *Journal of Network and Computer Applications*, vol. 39, pp. 52-60, May 2, 2013.

Hsiao-Wei Tang, "ARCM", Video, Sep. 2012, YouTube, 5 screenshots, Retrieved from internet: <<https://www.youtube.com/watch?v=UldQaxb3fRw&feature=youtu.be>>.

Giannios, et al., "Visible to near-infrared refractive properties of freshly-excised human-liver tissues: marking hepatic malignancies," Article, Jun. 14, 2016, pp. 1-10, *Scientific Reports* 6, Article No. 27910, Nature.

Vander Heiden, et al., "Understanding the Warburg effect: the metabolic requirements of cell proliferation," Article, May 22, 2009, pp. 1-12, vol. 324, Issue 5930, *Science*.

Hirayama et al., "Quantitative Metabolome Profiling of Colon and Stomach Cancer Microenvironment by Capillary Electrophoresis Time-of-Flight Mass Spectrometry," Article, Jun. 2009, pp. 4918-4925, vol. 69, Issue 11, *Cancer Research*.

Cengiz, et al., "A Tale of Two Compartments: Interstitial Versus Blood Glucose Monitoring," Article, Jun. 2009, pp. S11-S16, vol. 11, Supplement 1, *Diabetes Technology & Therapeutics*.

Shen, et al., "An iridium nanoparticles dispersed carbon based thick film electrochemical biosensor and its application for a single use, disposable glucose biosensor," Article, Feb. 3, 2007, pp. 106-113, vol. 125, Issue 1, *Sensors and Actuators B: Chemical*, Science Direct.

IEEE Std No. 177, "Standard Definitions and Methods of Measurement for Piezoelectric Vibrators," published May 1966, The Institute of Electrical and Electronics Engineers, Inc., New York, N.Y. CRC Press, "The Measurement, Instrumentation and Sensors Handbook," 1999, Section VII, Chapter 41, Peter O'Shea, "Phase Measurement," pp. 1303-1321, ISBN 0-8493-2145-X.

"ATM-MPLS Network Interworking Version 2.0, af-aic-0178.001" ATM Standard, The ATM Forum Technical Committee, published Aug. 2003.

IEEE Std 802.3-2012 (Revision of IEEE Std 802.3-2008), published Dec. 28, 2012.

(56)

References Cited

OTHER PUBLICATIONS

Bonaci et al., "To Make a Robot Secure: An Experimental Analysis of Cyber Security Threats Against Teleoperated Surgical Robots," May 13, 2015. Retrieved from the Internet: URL:<https://arxiv.org/pdf/1504.04339v2.pdf> [retrieved on Aug. 25, 2019].

Homa Alemzadeh et al., "Targeted Attacks on Teleoperated Surgical Robots: Dynamic Model-Based Detection and Mitigation," 2016 46th Annual IEEE/IFIP International Conference on Dependable Systems and Networks (DSN), IEEE, Jun. 28, 2016, pp. 395-406.

Harold I. Brandon and V. Leroy Young, Mar. 1997, Surgical Services Management vol. 3 No. 3. retrieved from the internet <<https://www.surgimedics.com/Research%20Articles/Electrosurgical%20Plume/Characterization%20And%20Removal%20Of%20Electrosurgical%20Smoke.pdf>> (Year: 1997).

Marshall Brain, How Microcontrollers Work, 2006, retrieved from the internet <<https://web.archive.org/web/20060221235221/http://electronics.howstuffworks.com/microcontroller.htm/printable>> (Year: 2006).

Staub et al., "Contour-based Surgical Instrument Tracking Supported by Kinematic Prediction," Proceedings of the 2010 3rd IEEE RAS & EMBS International Conference on Biomedical Robotics and Biomechanics, Sep. 1, 2010, pp. 746-752.

Phumzile Malindi, "5. QoS in Telemedicine," "Telemedicine," Jun. 20, 2011, IntechOpen, pp. 119-138.

Allan et al., "3-D Pose Estimation of Articulated Instruments in Robotic Minimally Invasive Surgery," IEEE Transactions on Medical Imaging, vol. 37, No. 5, May 1, 2018, pp. 1204-1213.

Kassahun et al., "Surgical Robotics Beyond Enhanced Dexterity Instrumentation: A Survey of the Machine Learning Techniques and their Role in Intelligent and Autonomous Surgical Actions." International Journal of Computer Assisted Radiology and Surgery, vol. 11, No. 4, Oct. 8, 2015, pp. 553-568.

Weede et al., "An Intelligent and Autonomous Endoscopic Guidance System for Minimally Invasive Surgery," 2013 IEEE International Conference on Robotics and Automation (ICRA), May 6-10, 2013. Karlsruhe, Germany, May 1, 2011, pp. 5762-5768.

Altenberg et al., "Genes of Glycolysis are Ubiquitously Overexpressed in 24 Cancer Classes," Genomics, vol. 84, pp. 1014-1020 (2004).

Jiang, "'Sound of Silence': a secure indoor wireless ultrasonic communication system," Article, 2014, pp. 46-50, Snapshots of Doctoral Research at University College Cork, School of Engineering—Electrical & Electronic Engineering, UCC, Cork, Ireland.

Li, et al., "Short-range ultrasonic communications in air using quadrature modulation," Journal, Oct. 30, 2009, pp. 2060-2072, IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control, vol. 56, No. 10, IEEE.

Salamon, "AI Detects Polyps Better Than Colonoscopists" Online Article, Jun. 3, 2018, Medscape Medical News, Digestive Disease Week (DDW) 2018: Presentation 133.

Misawa, et al. "Artificial Intelligence-Assisted Polyp Detection for Colonoscopy: Initial Experience," Article, Jun. 2018, pp. 2027-2029, vol. 154, Issue 8, American Gastroenterology Association.

Dottorato, "Analysis and Design of the Rectangular Microstrip Patch Antennas for TM_{0n0} operating mode," Article, Oct. 8, 2010, pp. 1-9, Microwave Journal.

Xie et al., Development of stereo vision and master-slave controller for a compact surgical robot system, 2015, IEEE, p. 403-407 (Year: 2015).

Sun et al., Innovative effector design for simulation training in robotic surgery, 2010, IEEE, p. 1755-1759 (Year: 2010).

Nabil Simaan et al., "Intelligent Surgical Robots with Situational Awareness: From Good to Great Surgeons", DOI: 10.1115/1.2015-Sep-6 external link, Sep. 2015 (Sep. 2015), p. 3-6, Retrieved from the Internet: URL:<http://memagazineselect.asmedigitalcollection.asme.org/data/journals/meena/936888/me-2015-sep6.pdf> XP055530863.

Anonymous, "Internet of Things Powers Connected Surgical Device Infrastructure Case Study", Dec. 31, 2016 (Dec. 31, 2016), Retrieved from the Internet: URL:https://www.cognizant.com/services-resources/150110_IoT_connected_surgical_devices.pdf.

Roy D Cullum, "Handbook of Engineering Design", ISBN: 9780408005586, Jan. 1, 1988 (Jan. 1, 1988), XP055578597, ISBN: 9780408005586, 10-20, Chapter 6, p. 138, right-hand col. paragraph 3.

Draijer, Matthijs et al., "Review of laser speckle contrast techniques for visualizing tissue perfusion," Lasers in Medical Science, Springer-Verlag, I.O. vol. 24, No. 4, Dec. 3, 2008, pp. 639-651.

Takahashi et al., "Automatic smoke evacuation in laparoscopic surgery: a simplified method for objective evaluation," Surgical Endoscopy, vol. 27, No. 8, pp. 2980-2987, Feb. 23, 2013.

Anonymous: "Titanium Key Chain Tool 1.1, Ultralight Multipurpose Key Chain Tool, Forward Cutting Can Opener—Vargo Titanium," vargooutdoors.com, Jul. 5, 2014 (Jul. 5, 2014), retrieved from the internet: <https://vargooutdoors.com/titanium-key-chain-tool-1-1.html>.

Anonymous: "Screwdriver—Wikipedia", en.wikipedia.org, Jun. 23, 2019, XP055725151, Retrieved from the Internet: URL:<https://en.wikipedia.org/w/index.php?title=Screwdriver&oldid=903111203> [retrieved on Mar. 20, 2021].

Nordlinger, Christopher, "The Internet of Things and the Operating Room of the Future," May 4, 2015, <https://medium.com/@chrisnordlinger/the-internet-of-things-and-the-operating-room-of-the-future-8999a143d7b1>, retrieved from the internet on Apr. 27, 2021, 9 pages.

Screen captures from YouTube video clip entitled "Four ways to use the Lego Brick Separator Tool," 2 pages, uploaded on May 29, 2014 by user "Sarah Lewis". Retrieved from internet: <https://www.youtube.com/watch?v=ucKiRD6U1LU> (Year: 2014).

Sorrells, P., "Application Note AN680. Passive RFID Basics," retrieved from <http://ww1.microchip.com/downloads/en/AppNotes/00680b.pdf> on Feb. 26, 2020, Dec. 31, 1998, pp. 1-7.

* cited by examiner





FIG. 1






FIG. 3




FIG. 5




FIG. 4




FIG. 6




FIG. 7




FIG. 8




FIG. 9




FIG. 10




FIG. 11




FIG. 12




FIG. 13




FIG. 14




FIG. 15




FIG. 16




FIG. 17




FIG. 18




FIG. 19




FIG. 20




FIG. 21




FIG. 22




FIG. 23




FIG. 24




FIG. 25




FIG. 27A




FIG. 27B




FIG. 27C




FIG. 28A




FIG. 28B




FIG. 29




FIG. 30




FIG. 32




FIG. 33




FIG. 34

CONTROL CIRCUIT 3200 FIG. 30




FIG. 35




FIG. 36




FIG. 37




FIG. 38




FIG. 39




FIG. 40




FIG. 41




FIG. 42




FIG. 43A




FIG. 43B




FIG. 44




FIG. 45




FIG. 46A




FIG. 46B




FIG. 47A




FIG. 47B




FIG. 48




FIG. 49




FIG. 50




FIG. 51




FIG. 52




FIG. 53




FIG. 54




FIG. 55




FIG. 56




FIG. 57




FIG. 58




FIG. 59




FIG. 60




FIG. 61




FIG. 62




FIG. 63




FIG. 64B




FIG. 64A




FIG. 65




FIG. 66




FIG. 67




FIG. 68




FIG. 69




FIG. 71




FIG. 72




FIG. 73





FIG. 74






FIG. 76




FIG. 77




FIG. 78




FIG. 79




FIG. 80




FIG. 81




FIG. 82




FIG. 83




FIG. 84




FIG. 85

**CONTROLLING A TEMPERATURE OF AN
ULTRASONIC ELECTROMECHANICAL
BLADE ACCORDING TO FREQUENCY**

CROSS-REFERENCE TO RELATED
APPLICATIONS

The present application claims priority under 35 U.S.C. § 120 to U.S. patent application Ser. No. 16/115,239, titled DETERMINING THE STATE OF AN ULTRASONIC ELECTROMECHANICAL SYSTEM ACCORDING TO FREQUENCY SHIFT, filed on Aug. 28, 2018, now U.S. Patent Application Publication No. 2019/0201042, the disclosure of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 claims priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application No. 62/721,995, titled CONTROLLING AN ULTRASONIC SURGICAL INSTRUMENT ACCORDING TO TISSUE LOCATION, filed on Aug. 23, 2018, the disclosure of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 claims priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application No. 62/721,998, titled SITUATIONAL AWARENESS OF ELECTROSURGICAL SYSTEMS, filed on Aug. 23, 2018, the disclosure of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 claims priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application No. 62/721,999, titled INTERRUPTION OF ENERGY DUE TO INADVERTENT CAPACITIVE COUPLING, filed on Aug. 23, 2018, the disclosure of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 claims priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application No. 62/721,994, titled BIPOLAR COMBINATION DEVICE THAT AUTOMATICALLY ADJUSTS PRESSURE BASED ON ENERGY MODALITY, filed on Aug. 23, 2018, the disclosure of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 claims priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application No. 62/721,996, titled RADIO FREQUENCY ENERGY DEVICE FOR DELIVERING COMBINED ELECTRICAL SIGNALS, filed on Aug. 23, 2018, the disclosure of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 also claims priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application No. 62/692,747, titled SMART ACTIVATION OF AN ENERGY DEVICE BY ANOTHER DEVICE, filed on Jun. 30, 2018, to U.S. Provisional Patent Application No. 62/692,748, titled SMART ENERGY ARCHITECTURE, filed on Jun. 30, 2018, and to U.S. Provisional Patent Application No. 62/692,768, titled SMART ENERGY DEVICES, filed on Jun. 30, 2018, the disclosure of each of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 also claims the benefit of priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application Ser. No. 62/640,417, titled TEMPERATURE CONTROL IN ULTRASONIC DEVICE AND CONTROL SYSTEM THEREFOR, filed Mar. 8, 2018, and to U.S. Provisional Patent Application Ser. No. 62/640,415, titled ESTIMATING STATE OF ULTRASONIC END EFFECTOR AND CONTROL SYSTEM

THEREFOR, filed Mar. 8, 2018, the disclosure of each of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 also claims the benefit of priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application No. 62/650,898 filed on Mar. 30, 2018, titled CAPACITIVE COUPLED RETURN PATH PAD WITH SEPARABLE ARRAY ELEMENTS, to U.S. Provisional Patent Application Ser. No. 62/650,887, titled SURGICAL SYSTEMS WITH OPTIMIZED SENSING CAPABILITIES, filed Mar. 30, 2018, to U.S. Provisional Patent Application Ser. No. 62/650,882, titled SMOKE EVACUATION MODULE FOR INTERACTIVE SURGICAL PLATFORM, filed Mar. 30, 2018, and to U.S. Provisional Patent Application Ser. No. 62/650,877, titled SURGICAL SMOKE EVACUATION SENSING AND CONTROLS, filed Mar. 30, 2018, the disclosure of each of which is herein incorporated by reference in its entirety.

U.S. patent application Ser. No. 16/115,239 also claims the benefit of priority under 35 U.S.C. § 119(e) to U.S. Provisional Patent Application Ser. No. 62/611,341, titled INTERACTIVE SURGICAL PLATFORM, filed Dec. 28, 2017, to U.S. Provisional Patent Application Ser. No. 62/611,340, titled CLOUD-BASED MEDICAL ANALYTICS, filed Dec. 28, 2017, and to U.S. Provisional Patent Application Ser. No. 62/611,339, titled ROBOT ASSISTED SURGICAL PLATFORM, filed Dec. 28, 2017, the disclosure of each of which is herein incorporated by reference in its entirety.

BACKGROUND

In a surgical environment, smart energy devices may be needed in a smart energy architecture environment.

SUMMARY

In one general aspect, an ultrasonic electromechanical system for an ultrasonic surgical instrument. The ultrasonic electromechanical system comprises: an ultrasonic blade, an ultrasonic transducer acoustically coupled to the ultrasonic blade, and a control circuit coupled to the ultrasonic transducer. The ultrasonic transducer is configured to ultrasonically oscillate the ultrasonic blade in response to a drive signal. The control circuit is configured to: determine a first resonant frequency of the ultrasonic electromechanical system, determine a second resonant frequency of the ultrasonic electromechanical system as the ultrasonic blade oscillates against a tissue, and determine a state of the ultrasonic surgical instrument according to a comparison between the first resonant frequency and the second resonant frequency.

In another general aspect, an ultrasonic generator connectable to an ultrasonic electromechanical system for an ultrasonic surgical instrument. The ultrasonic electromechanical system comprises an ultrasonic blade and an ultrasonic transducer acoustically coupled to the ultrasonic blade. The ultrasonic generator comprises a control circuit coupleable to the ultrasonic transducer. The control circuit is configured to: apply a drive signal to the ultrasonic transducer to cause the ultrasonic blade to oscillate, determine a first resonant frequency of the ultrasonic electromechanical system, determine a second resonant frequency of the ultrasonic electromechanical system as the ultrasonic blade oscillates against a tissue, and determine a state of the ultrasonic surgical instrument according to a comparison between the first resonant frequency and the second resonant frequency.

In another general aspect, a method of controlling an ultrasonic electromechanical system for an ultrasonic surgi-

3

cal instrument. The ultrasonic electromechanical system comprises an ultrasonic blade and an ultrasonic transducer acoustically coupled to the ultrasonic blade. The method comprises: determining, by a control circuit coupled to the ultrasonic electromechanical system, a natural resonant frequency of the ultrasonic electromechanical system; monitoring, by the control circuit, a resonant frequency of the ultrasonic electromechanical system as the ultrasonic blade oscillates; and determine, by the control circuit, whether a state change of the ultrasonic electromechanical system has occurred according to whether the resonant frequency of the ultrasonic electromechanical system has deviated from the natural resonant frequency.

In various embodiments, an ultrasonic electromechanical system for an ultrasonic surgical instrument is disclosed. The ultrasonic electromechanical system comprises an ultrasonic blade, a clamp arm disposed opposite the ultrasonic blade, an ultrasonic transducer acoustically coupled to the ultrasonic blade, and a control circuit coupled to the ultrasonic transducer. The ultrasonic transducer configured to ultrasonically oscillate the ultrasonic blade in response to a drive signal. The control circuit is configured to determine a temperature of the ultrasonic blade. The control circuit is configured to increase an amount of power of the drive signal when the temperature of the ultrasonic blade is less than a first predetermined value. The first predetermined value is a value less than a temperature configured to cut a tissue. The control circuit is configured to decrease the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than a second predetermined value. The second predetermined value is greater than the first predetermined value.

In various embodiments, an ultrasonic generator connectable to an ultrasonic electromechanical system for an ultrasonic surgical instrument is disclosed. The ultrasonic electromechanical system comprises an ultrasonic blade, a clamp arm disposed opposite the ultrasonic blade, and an ultrasonic transducer acoustically coupled to the ultrasonic blade. The ultrasonic generator comprises a control circuit coupleable to the ultrasonic transducer. The control circuit is configured to apply a drive signal to the ultrasonic transducer, determine a temperature of the ultrasonic blade, and increase an amount of power of the drive signal when the temperature of the ultrasonic blade is less than a first predetermined value. The first predetermined value is a value less than a temperature configured to cut a tissue. The control circuit is configured to decrease the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than a second predetermined value. The second predetermined value is greater than the first predetermined value.

In various embodiments, a method of controlling an ultrasonic electromechanical system for an ultrasonic surgical instrument is disclosed. The ultrasonic electromechanical system comprises an ultrasonic blade, a clamp arm disposed opposite the ultrasonic blade, a control circuit, and an ultrasonic transducer acoustically coupled to the ultrasonic blade. The ultrasonic transducer configured to ultrasonically oscillate the ultrasonic blade in response to a drive signal. The method comprises determining, by the control circuit, a temperature of the ultrasonic blade.

The method comprises increasing, by the control circuit, an amount of power of the drive signal when the temperature of the ultrasonic blade is less than a first predetermined value. The first predetermined value is a value less than a temperature configured to cut a tissue. The method comprises decreasing, by the control circuit, the amount of

4

power of the drive signal when the temperature of the ultrasonic blade is greater than a second predetermined value. The second predetermined value is greater than the first predetermined value.

FIGURES

The features of various aspects are set forth with particularity in the appended claims. The various aspects, however, both as to organization and methods of operation, together with further objects and advantages thereof, may best be understood by reference to the following description, taken in conjunction with the accompanying drawings as follows.

FIG. 1 is a block diagram of a computer-implemented interactive surgical system, in accordance with at least one aspect of the present disclosure.

FIG. 2 is a surgical system being used to perform a surgical procedure in an operating room, in accordance with at least one aspect of the present disclosure.

FIG. 3 is a surgical hub paired with a visualization system, a robotic system, and an intelligent instrument, in accordance with at least one aspect of the present disclosure.

FIG. 4 is a partial perspective view of a surgical hub enclosure, and of a combo generator module slidably receivable in a drawer of the surgical hub enclosure, in accordance with at least one aspect of the present disclosure.

FIG. 5 is a perspective view of a combo generator module with bipolar, ultrasonic, and monopolar contacts and a smoke evacuation component, in accordance with at least one aspect of the present disclosure.

FIG. 6 illustrates individual power bus attachments for a plurality of lateral docking ports of a lateral modular housing configured to receive a plurality of modules, in accordance with at least one aspect of the present disclosure.

FIG. 7 illustrates a vertical modular housing configured to receive a plurality of modules, in accordance with at least one aspect of the present disclosure.

FIG. 8 illustrates a surgical data network comprising a modular communication hub configured to connect modular devices located in one or more operating theaters of a healthcare facility, or any room in a healthcare facility specially equipped for surgical operations, to the cloud, in accordance with at least one aspect of the present disclosure.

FIG. 9 illustrates a computer-implemented interactive surgical system, in accordance with at least one aspect of the present disclosure.

FIG. 10 illustrates a surgical hub comprising a plurality of modules coupled to the modular control tower, in accordance with at least one aspect of the present disclosure.

FIG. 11 illustrates one aspect of a Universal Serial Bus (USB) network hub device, in accordance with at least one aspect of the present disclosure.

FIG. 12 illustrates a logic diagram of a control system of a surgical instrument or tool, in accordance with at least one aspect of the present disclosure.

FIG. 13 illustrates a control circuit configured to control aspects of the surgical instrument or tool, in accordance with at least one aspect of the present disclosure.

FIG. 14 illustrates a combinational logic circuit configured to control aspects of the surgical instrument or tool, in accordance with at least one aspect of the present disclosure.

FIG. 15 illustrates a sequential logic circuit configured to control aspects of the surgical instrument or tool, in accordance with at least one aspect of the present disclosure.

FIG. 16 illustrates a surgical instrument or tool comprising a plurality of motors which can be activated to perform various functions, in accordance with at least one aspect of the present disclosure.

FIG. 17 is a schematic diagram of a robotic surgical instrument configured to operate a surgical tool described herein, in accordance with at least one aspect of the present disclosure.

FIG. 18 illustrates a block diagram of a surgical instrument programmed to control the distal translation of a displacement member, in accordance with at least one aspect of the present disclosure.

FIG. 19 is a schematic diagram of a surgical instrument configured to control various functions, in accordance with at least one aspect of the present disclosure.

FIG. 20 is a system configured to execute adaptive ultrasonic blade control algorithms in a surgical data network comprising a modular communication hub, in accordance with at least one aspect of the present disclosure.

FIG. 21 illustrates an example of a generator, in accordance with at least one aspect of the present disclosure.

FIG. 22 is a surgical system comprising a generator and various surgical instruments usable therewith, in accordance with at least one aspect of the present disclosure.

FIG. 23 is an end effector, in accordance with at least one aspect of the present disclosure.

FIG. 24 is a diagram of the surgical system of FIG. 22, in accordance with at least one aspect of the present disclosure.

FIG. 25 is a model illustrating motional branch current, in accordance with at least one aspect of the present disclosure.

FIG. 26 is a structural view of a generator architecture, in accordance with at least one aspect of the present disclosure.

FIGS. 27A-27C are functional views of a generator architecture, in accordance with at least one aspect of the present disclosure.

FIGS. 28A-28B are structural and functional aspects of a generator, in accordance with at least one aspect of the present disclosure.

FIG. 29 is a schematic diagram of one aspect of an ultrasonic drive circuit.

FIG. 30 is a schematic diagram of a control circuit, in accordance with at least one aspect of the present disclosure.

FIG. 31 shows a simplified block circuit diagram illustrating another electrical circuit contained within a modular ultrasonic surgical instrument, in accordance with at least one aspect of the present disclosure.

FIG. 32 illustrates a generator circuit partitioned into multiple stages, in accordance with at least one aspect of the present disclosure.

FIG. 33 illustrates a generator circuit partitioned into multiple stages where a first stage circuit is common to the second stage circuit, in accordance with at least one aspect of the present disclosure.

FIG. 34 is a schematic diagram of one aspect of a drive circuit configured for driving a high-frequency current (RF), in accordance with at least one aspect of the present disclosure.

FIG. 35 illustrates one aspect of a fundamental architecture for a digital synthesis circuit such as a direct digital synthesis (DDS) circuit configured to generate a plurality of wave shapes for the electrical signal waveform for use in a surgical instrument, in accordance with at least one aspect of the present disclosure.

FIG. 36 illustrates one aspect of direct digital synthesis (DDS) circuit configured to generate a plurality of wave

shapes for the electrical signal waveform for use in surgical instrument, in accordance with at least one aspect of the present disclosure.

FIG. 37 illustrates one cycle of a discrete time digital electrical signal waveform, in accordance with at least one aspect of the present disclosure of an analog waveform (shown superimposed over a discrete time digital electrical signal waveform for comparison purposes), in accordance with at least one aspect of the present disclosure.

FIG. 38 is a diagram of a control system configured to provide progressive closure of a closure member as it advances distally to close the clamp arm to apply a closure force load at a desired rate according to one aspect of this disclosure.

FIG. 39 illustrates a proportional-integral-derivative (PID) controller feedback control system according to one aspect of this disclosure.

FIG. 40 is a system diagram of a segmented circuit comprising a plurality of independently operated circuit segments, in accordance with at least one aspect of the present disclosure.

FIG. 41 is a circuit diagram of various components of a surgical instrument with motor control functions, in accordance with at least one aspect of the present disclosure.

FIG. 42 is an alternative system for controlling the frequency of an ultrasonic electromechanical system and detecting the impedance thereof, in accordance with at least one aspect of the present disclosure.

FIG. 43A is a graphical representation of impedance phase angle as a function of resonant frequency of the same ultrasonic device with a cold (blue) and hot (red) ultrasonic blade; and

FIG. 43B is a graphical representation of impedance magnitude as a function of resonant frequency of the same ultrasonic device with a cold (blue) and hot (red) ultrasonic blade.

FIG. 44 is a diagram of a Kalman filter to improve temperature estimator and state space model based on impedance across an ultrasonic transducer measured at a variety of frequencies, in accordance with at least one aspect of the present disclosure.

FIG. 45 are three probability distributions employed by a state estimator of the Kalman filter shown in FIG. 44 to maximize estimates, in accordance with at least one aspect of the present disclosure.

FIG. 46A is a graphical representation of temperature versus time of an ultrasonic device with no temperature control reaching a maximum temperature of 490° C.

FIG. 46B is a graphical representation of temperature versus time of an ultrasonic device with temperature control reaching a maximum temperature of 320° C., in accordance with at least one aspect of the present disclosure.

FIGS. 47A-47B are graphical representations of feedback control to adjust ultrasonic power applied to an ultrasonic transducer when a sudden drop in temperature of an ultrasonic blade is detected, where

FIG. 47A is a graphical representation of ultrasonic power as a function of time; and

FIG. 47B is a plot of ultrasonic blade temperature as a function of time, in accordance with at least one aspect of the present disclosure.

FIG. 48 is a logic flow diagram of a process depicting a control program or a logic configuration to control the temperature of an ultrasonic blade, in accordance with at least one aspect of the present disclosure.

FIG. 49 is a graphical representation of ultrasonic blade temperature as a function of time during a vessel firing, in accordance with at least one aspect of the present disclosure.

FIG. 50 is a logic flow diagram of a process depicting a control program or a logic configuration to control the temperature of an ultrasonic blade between two temperature set points, in accordance with at least one aspect of the present disclosure.

FIG. 51 is a logic flow diagram of a process depicting a control program or a logic configuration to determine the initial temperature of an ultrasonic blade, in accordance with at least one aspect of the present disclosure.

FIG. 52 is a logic flow diagram of a process depicting a control program or a logic configuration to determine when an ultrasonic blade is approaching instability and then adjusting the power to the ultrasonic transducer to prevent instability of the ultrasonic transducer, in accordance with at least one aspect of the present disclosure.

FIG. 53 is a logic flow diagram of a process depicting a control program or a logic configuration to provide ultrasonic sealing with temperature control, in accordance with at least one aspect of the present disclosure.

FIG. 54 are graphical representations of ultrasonic transducer current and ultrasonic blade temperature as a function of time, in accordance with at least one aspect of the present disclosure.

FIG. 55 is a bottom view of an ultrasonic end effector showing a clamp arm and ultrasonic blade delineating tissue positioning within the ultrasonic end effector, in accordance with at least one aspect of the present disclosure.

FIG. 56 is a graphical representation depicting change in ultrasonic transducer impedance as a function tissue location within the ultrasonic end effector over a range of predetermined ultrasonic generator power level increases, in accordance with at least one aspect of the present disclosure.

FIG. 57 is a graphical representation depicting change in ultrasonic transducer impedance as a function of time relative to the location of tissue within the ultrasonic end effector, in accordance with at least one aspect of the present disclosure.

FIG. 58 is a logic flow diagram of a process depicting a control program or a logic configuration to identify operation in a non-therapeutic range of power applied to the ultrasonic transducer to determine tissue positioning, in accordance with at least one aspect of the present disclosure.

FIG. 59 illustrates one aspect of an end effector of an ultrasonic surgical instrument comprising infrared (IR) sensors located on the jaw member, in accordance with at least one aspect of the present disclosure.

FIG. 60 illustrates one aspect of a flexible circuit on which the IR sensors shown in FIG. 59 may be mounted or formed integrally with, in accordance to one aspect of the present disclosure.

FIG. 61 is a sectional view of an ultrasonic end effector comprising a clamp arm and an ultrasonic blade, in accordance with at least one aspect of the present disclosure.

FIG. 62 illustrates IR refractivity detection sensor circuits mounted on a flexible circuit substrate shown in plan view, in accordance with at least one aspect of the present disclosure.

FIG. 63 is a logic flow diagram of a process depicting a control program or a logic configuration to measure IR reflectance to determine tissue composition to tune the amplitude of the ultrasonic transducer, in accordance with at least one aspect of the present disclosure.

FIG. 64A is a graphical representation of the rate of closure of the clamp arm versus time to identify the collagen

transformation point according to various aspects of the present disclosure where time is shown along the horizontal axis and change in clamp arm position is shown along the vertical axis, in accordance with at least one aspect of the present disclosure.

FIG. 64B is a magnified portion of the graphical representation shown in FIG. 64A.

FIG. 65 is a logic flow diagram of a process depicting a control program or a logic configuration to detect the collagen transformation point to control the rate of closure of the of the clamp arm or the amplitude of the ultrasonic transducer, in accordance with at least one aspect of the present disclosure.

FIG. 66 is a graphical representation of the identification of the collagen transformation temperature point to identify the collagen/elastin ratio according to various aspects of the present disclosure, where tissue temperature is shown along the horizontal axis and ultrasonic transducer impedance is shown along the vertical axis, in accordance with at least one aspect of the present disclosure.

FIG. 67 is a logic flow diagram of a process depicting a control program or a logic configuration to identify the collagen transformation temperature to identify the collagen/elastin ratio, in accordance with at least one aspect of the present disclosure.

FIG. 68 is a graphical representation of the distribution of compression load across an ultrasonic blade, in accordance with at least one aspect of the present disclosure.

FIG. 69 is a graphical representation of pressure applied to tissue versus time, in accordance with at least one aspect of the present disclosure.

FIG. 70 illustrates an end effector including a single-jaw electrode array for detecting tissue location, in accordance with at least one aspect of the present disclosure.

FIG. 71 is an activation matrix for the single-jaw electrode array of FIG. 70, in accordance with at least one aspect of the present disclosure.

FIG. 72 illustrates an end effector including a dual-jaw electrode array for detecting tissue location, in accordance with at least one aspect of the present disclosure.

FIG. 73 is an activation matrix for the dual-jaw electrode array of FIG. 72, in accordance with at least one aspect of the present disclosure.

FIG. 74 illustrates opposing sets of electrodes overlaid a tissue grasped by an end effector corresponding to the activation matrix on FIG. 73, in accordance with at least one aspect of the present disclosure.

FIG. 75 illustrates an end effector including dual-jaw segmented electrode array, according to at least one aspect of the present disclosure.

FIG. 76 illustrates a tissue overlaid a jaw including a segmented electrode array, according to at least one aspect of the present disclosure.

FIG. 77 is a schematic diagram of a segmented electrode array circuit including band-pass filters, in accordance with at least one aspect of the present disclosure.

FIG. 78 is a graphical representation of the frequency response corresponding to the tissue grasped in FIG. 76, in accordance with at least one aspect of the present disclosure.

FIG. 79 is a graphical representation of the frequency of the ultrasonic transducer system as a function of drive frequency and ultrasonic blade temperature drift, in accordance with at least one aspect of the present disclosure.

FIG. 80 is a graphical representation of temperature of the ultrasonic transducer as a function of time, in accordance with at least one aspect of the present disclosure.

FIG. 81 is a graphical representation of the modal shift of resonant frequency based on the temperature of the ultrasonic blade moving the resonant frequency as a function of the temperature of the ultrasonic blade, in accordance with at least one aspect of the present disclosure.

FIG. 82 is a spectra of an ultrasonic surgical instrument with a variety of different states and conditions of the end effector where phase and magnitude of the impedance of an ultrasonic transducer are plotted as a function of frequency, in accordance with at least one aspect of the present disclosure.

FIG. 83 is a methodology for classification of data based on a set of training data S, where ultrasonic transducer impedance magnitude and phase are plotted as a function of frequency, in accordance with at least one aspect of the present disclosure.

FIG. 84 is a logic flow diagram depicting a control program or a logic configuration to determine jaw conditions based on the complex impedance characteristic pattern (fingerprint), in accordance with at least one aspect of the present disclosure.

FIG. 85 is a timeline depicting situational awareness of a surgical hub, in accordance with at least one aspect of the present disclosure.

DESCRIPTION

Applicant of the present application owns the following U.S. patent applications, filed on Aug. 28, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

- U.S. patent application Ser. No. 16/115,214, titled ESTIMATING STATE OF ULTRASONIC END EFFECTOR AND CONTROL SYSTEM THEREFOR, now U.S. Patent Application Publication No. 2019/0201073;
- U.S. patent application Ser. No. 16/115,205, titled TEMPERATURE CONTROL OF ULTRASONIC END EFFECTOR AND CONTROL SYSTEM THEREFOR, now U.S. Patent Application Publication No. 2019/0201036;
- U.S. patent application Ser. No. 16/115,233, titled RADIO FREQUENCY ENERGY DEVICE FOR DELIVERING COMBINED ELECTRICAL SIGNALS, now U.S. Patent Application Publication No. 2019/0201091;
- U.S. patent application Ser. No. 16/115,208, titled CONTROLLING AN ULTRASONIC SURGICAL INSTRUMENT ACCORDING TO TISSUE LOCATION, now U.S. Patent Application Publication No. 2019/0201037;
- U.S. patent application Ser. No. 16/115,220, titled CONTROLLING ACTIVATION OF AN ULTRASONIC SURGICAL INSTRUMENT ACCORDING TO THE PRESENCE OF TISSUE, now U.S. Patent Application Publication No. 2019/0201040;
- U.S. patent application Ser. No. 16/115,232, titled DETERMINING TISSUE COMPOSITION VIA AN ULTRASONIC SYSTEM, now U.S. Patent Application Publication No. 2019/0201038;
- U.S. patent application Ser. No. 16/115,247, titled DETERMINING THE STATE OF AN ULTRASONIC END EFFECTOR, now U.S. Patent Application Publication No. 2019/0274716;

U.S. patent application Ser. No. 16/115,211, titled SITUATIONAL AWARENESS OF ELECTROSURGICAL SYSTEMS, now U.S. Patent Application Publication No. 2019/0201039;

U.S. patent application Ser. No. 16/115,226, titled MECHANISMS FOR CONTROLLING DIFFERENT ELECTROMECHANICAL SYSTEMS OF AN ELECTROSURGICAL INSTRUMENT, now U.S. Patent Application Publication No. 2019/0201075;

U.S. patent application Ser. No. 16/115,240, titled DETECTION OF END EFFECTOR IMMERSION IN LIQUID, now U.S. Patent Application Publication No. 2019/0201043;

U.S. patent application Ser. No. 16/115,249, titled INTERRUPTION OF ENERGY DUE TO INADVERTENT CAPACITIVE COUPLING, now U.S. Patent Application Publication No. 2019/0201077;

U.S. patent application Ser. No. 16/115,256, titled INCREASING RADIO FREQUENCY TO CREATE PAD-LESS MONOPOLAR LOOP, now U.S. Patent Application Publication No. 2019/0201092;

U.S. patent application Ser. No. 16/115,223, titled BIPO-LAR COMBINATION DEVICE THAT AUTOMATICALLY ADJUSTS PRESSURE BASED ON ENERGY MODALITY, now U.S. Patent Application Publication No. 2019/0201074; and

U.S. patent application Ser. No. 16/115,238, titled ACTIVATION OF ENERGY DEVICES, now U.S. Patent Application Publication No. 2019/0201041.

Applicant of the present application owns the following U.S. patent applications, filed on Aug. 23, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

- U.S. Provisional Patent Application No. 62/721,995, titled CONTROLLING AN ULTRASONIC SURGICAL INSTRUMENT ACCORDING TO TISSUE LOCATION;
- U.S. Provisional Patent Application No. 62/721,998, titled SITUATIONAL AWARENESS OF ELECTROSURGICAL SYSTEMS;
- U.S. Provisional Patent Application No. 62/721,999, titled INTERRUPTION OF ENERGY DUE TO INADVERTENT CAPACITIVE COUPLING;
- U.S. Provisional Patent Application No. 62/721,994, titled BIPO-LAR COMBINATION DEVICE THAT AUTOMATICALLY ADJUSTS PRESSURE BASED ON ENERGY MODALITY; and
- U.S. Provisional Patent Application No. 62/721,996, titled RADIO FREQUENCY ENERGY DEVICE FOR DELIVERING COMBINED ELECTRICAL SIGNALS.

Applicant of the present application owns the following U.S. patent applications, filed on Jun. 30, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

- U.S. Provisional Patent Application No. 62/692,747, titled SMART ACTIVATION OF AN ENERGY DEVICE BY ANOTHER DEVICE;
- U.S. Provisional Patent Application No. 62/692,748, titled SMART ENERGY ARCHITECTURE; and
- U.S. Provisional Patent Application No. 62/692,768, titled SMART ENERGY DEVICES.

Applicant of the present application owns the following U.S. patent applications, filed on Jun. 29, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

11

U.S. patent application Ser. No. 16/024,090, titled CAPACITIVE COUPLED RETURN PATH PAD WITH SEPARABLE ARRAY ELEMENTS, now U.S. Patent Application Publication No. 2019/0201090;

U.S. patent application Ser. No. 16/024,057, titled CONTROLLING A SURGICAL INSTRUMENT ACCORDING TO SENSED CLOSURE PARAMETERS, now U.S. Pat. No. 10,695,081;

U.S. patent application Ser. No. 16/024,067, titled SYSTEMS FOR ADJUSTING END EFFECTOR PARAMETERS BASED ON PERIOPERATIVE INFORMATION, now U.S. Pat. No. 10,595,887;

U.S. patent application Ser. No. 16/024,075, titled SAFETY SYSTEMS FOR SMART POWERED SURGICAL STAPLING, now U.S. Patent Application Publication No. 2019/0201146;

U.S. patent application Ser. No. 16/024,083, titled SAFETY SYSTEMS FOR SMART POWERED SURGICAL STAPLING, now U.S. Patent Application Publication No. 2019/0200984;

U.S. patent application Ser. No. 16/024,094, titled SURGICAL SYSTEMS FOR DETECTING END EFFECTOR TISSUE DISTRIBUTION IRREGULARITIES, now U.S. Patent Application Publication No. 2019/0201020;

U.S. patent application Ser. No. 16/024,138, titled SYSTEMS FOR DETECTING PROXIMITY OF SURGICAL END EFFECTOR TO CANCEROUS TISSUE, now U.S. Patent Application Publication No. 2019/0200985;

U.S. patent application Ser. No. 16/024,150, titled SURGICAL INSTRUMENT CARTRIDGE SENSOR ASSEMBLIES, now U.S. Patent Application Publication No. 2019/0200986;

U.S. patent application Ser. No. 16/024,160, titled VARIABLE OUTPUT CARTRIDGE SENSOR ASSEMBLY, now U.S. Patent Application Publication No. 2019/0200987;

U.S. patent application Ser. No. 16/024,124, titled SURGICAL INSTRUMENT HAVING A FLEXIBLE ELECTRODE, now U.S. Patent Application Publication No. 2019/0201079;

U.S. patent application Ser. No. 16/024,132, titled SURGICAL INSTRUMENT HAVING A FLEXIBLE CIRCUIT, now U.S. Patent Application Publication No. 2019/0201021;

U.S. patent application Ser. No. 16/024,141, titled SURGICAL INSTRUMENT WITH A TISSUE MARKING ASSEMBLY, now U.S. Patent Application Publication No. 2019/0201159;

U.S. patent application Ser. No. 16/024,162, titled SURGICAL SYSTEMS WITH PRIORITIZED DATA TRANSMISSION CAPABILITIES, now U.S. Patent Application Publication No. 2019/0200988;

U.S. patent application Ser. No. 16/024,066, titled SURGICAL EVACUATION SENSING AND MOTOR CONTROL, now U.S. Patent Application Publication No. 2019/0201082;

U.S. patent application Ser. No. 16/024,096, titled SURGICAL EVACUATION SENSOR ARRANGEMENTS, now U.S. Patent Application Publication No. 2019/0201083;

U.S. patent application Ser. No. 16/024,116, titled SURGICAL EVACUATION FLOW PATHS, now U.S. Patent Application Publication No. 2019/0201084;

12

U.S. patent application Ser. No. 16/024,149, titled SURGICAL EVACUATION SENSING AND GENERATOR CONTROL, now U.S. Patent Application Publication No. 2019/0201085;

U.S. patent application Ser. No. 16/024,180, titled SURGICAL EVACUATION SENSING AND DISPLAY, now U.S. Patent Application Publication No. 2019/0201086;

U.S. patent application Ser. No. 16/024,245, titled COMMUNICATION OF SMOKE EVACUATION SYSTEM PARAMETERS TO HUB OR CLOUD IN SMOKE EVACUATION MODULE FOR INTERACTIVE SURGICAL PLATFORM, now U.S. Pat. No. 10,755,813;

U.S. patent application Ser. No. 16/024,258, titled SMOKE EVACUATION SYSTEM INCLUDING A SEGMENTED CONTROL CIRCUIT FOR INTERACTIVE SURGICAL PLATFORM, now U.S. Patent Application Publication No. 2019/0201087;

U.S. patent application Ser. No. 16/024,265, titled SURGICAL EVACUATION SYSTEM WITH A COMMUNICATION CIRCUIT FOR COMMUNICATION BETWEEN A FILTER AND A SMOKE EVACUATION DEVICE, now U.S. Patent Application Publication No. 2019/0201088; and

U.S. patent application Ser. No. 16/024,273, titled DUAL IN-SERIES LARGE AND SMALL DROPLET FILTERS, now U.S. Patent Application Publication No. 2019/0201597.

Applicant of the present application owns the following U.S. Provisional patent applications, filed on Jun. 28, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

U.S. Provisional Patent Application Ser. No. 62/691,228, titled A METHOD OF USING REINFORCED FLEX CIRCUITS WITH MULTIPLE SENSORS WITH ELECTROSURGICAL DEVICES;

U.S. Provisional Patent Application Ser. No. 62/691,227, titled CONTROLLING A SURGICAL INSTRUMENT ACCORDING TO SENSED CLOSURE PARAMETERS;

U.S. Provisional Patent Application Ser. No. 62/691,230, titled SURGICAL INSTRUMENT HAVING A FLEXIBLE ELECTRODE;

U.S. Provisional Patent Application Ser. No. 62/691,219, titled SURGICAL EVACUATION SENSING AND MOTOR CONTROL;

U.S. Provisional Patent Application Ser. No. 62/691,257, titled COMMUNICATION OF SMOKE EVACUATION SYSTEM PARAMETERS TO HUB OR CLOUD IN SMOKE EVACUATION MODULE FOR INTERACTIVE SURGICAL PLATFORM;

U.S. Provisional Patent Application Ser. No. 62/691,262, titled SURGICAL EVACUATION SYSTEM WITH A COMMUNICATION CIRCUIT FOR COMMUNICATION BETWEEN A FILTER AND A SMOKE EVACUATION DEVICE; and

U.S. Provisional Patent Application Ser. No. 62/691,251, titled DUAL IN-SERIES LARGE AND SMALL DROPLET FILTERS.

Applicant of the present application owns the following U.S. Provisional patent application, filed on Apr. 19, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

U.S. Provisional Patent Application Ser. No. 62/659,900, titled METHOD OF HUB COMMUNICATION.

Applicant of the present application owns the following U.S. Provisional patent applications, filed on Mar. 30, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

- U.S. Provisional Patent Application No. 62/650,898 filed on Mar. 30, 2018, titled CAPACITIVE COUPLED RETURN PATH PAD WITH SEPARABLE ARRAY ELEMENTS; 5
- U.S. Provisional Patent Application Ser. No. 62/650,887, titled SURGICAL SYSTEMS WITH OPTIMIZED SENSING CAPABILITIES; 10
- U.S. Provisional Patent Application Ser. No. 62/650,882, titled SMOKE EVACUATION MODULE FOR INTERACTIVE SURGICAL PLATFORM; and 15
- U.S. Provisional Patent Application Ser. No. 62/650,877, titled SURGICAL SMOKE EVACUATION SENSING AND CONTROLS

Applicant of the present application owns the following U.S. patent applications, filed on Mar. 29, 2018, the disclosure of each of which is herein incorporated by reference in its entirety: 20

- U.S. patent application Ser. No. 15/940,641, titled INTERACTIVE SURGICAL SYSTEMS WITH ENCRYPTED COMMUNICATION CAPABILITIES, now U.S. Patent Application Publication No. 2019/0207911; 25
- U.S. patent application Ser. No. 15/940,648, titled INTERACTIVE SURGICAL SYSTEMS WITH CONDITION HANDLING OF DEVICES AND DATA CAPABILITIES, now U.S. Patent Application Publication No. 2019/0206004; 30
- U.S. patent application Ser. No. 15/940,656, titled SURGICAL HUB COORDINATION OF CONTROL AND COMMUNICATION OF OPERATING ROOM DEVICES, now U.S. Patent Application Publication No. 2019/0201141; 35
- U.S. patent application Ser. No. 15/940,666, titled SPATIAL AWARENESS OF SURGICAL HUBS IN OPERATING ROOMS, now U.S. Patent Application Publication No. 2019/0206551; 40
- U.S. patent application Ser. No. 15/940,670, titled COOPERATIVE UTILIZATION OF DATA DERIVED FROM SECONDARY SOURCES BY INTELLIGENT SURGICAL HUBS, now U.S. Patent Application Publication No. 2019/0201116; 45
- U.S. patent application Ser. No. 15/940,677, titled SURGICAL HUB CONTROL ARRANGEMENTS, now U.S. Patent Application Publication No. 2019/0201143; 50
- U.S. patent application Ser. No. 15/940,632, titled DATA STRIPPING METHOD TO INTERROGATE PATIENT RECORDS AND CREATE ANONYMIZED RECORD, now U.S. Patent Application Publication No. 2019/0205566; 55
- U.S. patent application Ser. No. 15/940,640, titled COMMUNICATION HUB AND STORAGE DEVICE FOR STORING PARAMETERS AND STATUS OF A SURGICAL DEVICE TO BE SHARED WITH CLOUD BASED ANALYTICS SYSTEMS, now U.S. Patent Application Publication No. 2019/02005863; 60
- U.S. patent application Ser. No. 15/940,645, titled SELF DESCRIBING DATA PACKETS GENERATED AT AN ISSUING INSTRUMENT, now U.S. Patent Application Publication No. 2019/0207773; 65
- U.S. patent application Ser. No. 15/940,649, titled DATA PAIRING TO INTERCONNECT A DEVICE MEA-

- SURED PARAMETER WITH AN OUTCOME, now U.S. Patent Application Publication No. 2019/0205567;
- U.S. patent application Ser. No. 15/940,654, titled SURGICAL HUB SITUATIONAL AWARENESS, now U.S. Patent Application Publication No. 2019/0201140;
- U.S. patent application Ser. No. 15/940,663, titled SURGICAL SYSTEM DISTRIBUTED PROCESSING, now U.S. Patent Application Publication No. 2019/0201033;
- U.S. patent application Ser. No. 15/940,668, titled AGGREGATION AND REPORTING OF SURGICAL HUB DATA, now U.S. Patent Application Publication No. 2019/0201115;
- U.S. patent application Ser. No. 15/940,671, titled SURGICAL HUB SPATIAL AWARENESS TO DETERMINE DEVICES IN OPERATING THEATER, now U.S. Patent Application Publication No. 2019/0201104;
- U.S. patent application Ser. No. 15/940,686, titled DISPLAY OF ALIGNMENT OF STAPLE CARTRIDGE TO PRIOR LINEAR STAPLE LINE, now U.S. Patent Application Publication No. 2019/0201105;
- U.S. patent application Ser. No. 15/940,700, titled STERILE FIELD INTERACTIVE CONTROL DISPLAYS, now U.S. Patent Application Publication No. 2019/0205001;
- U.S. patent application Ser. No. 15/940,629, titled COMPUTER IMPLEMENTED INTERACTIVE SURGICAL SYSTEMS, now U.S. Patent Application Publication No. 2019/0201112;
- U.S. patent application Ser. No. 15/940,704, titled USE OF LASER LIGHT AND RED-GREEN-BLUE COLORATION TO DETERMINE PROPERTIES OF BACK SCATTERED LIGHT, now U.S. Patent Application Publication No. 2019/0206050;
- U.S. patent application Ser. No. 15/940,722, titled CHARACTERIZATION OF TISSUE IRREGULARITIES THROUGH THE USE OF MONO-CHROMATIC LIGHT REFRACTIVITY, now U.S. Patent Application Publication No. 2019/0200905; and
- U.S. patent application Ser. No. 15/940,742, titled DUAL CMOS ARRAY IMAGING, now U.S. Patent Application Publication No. 2019/0200906.
- U.S. patent application Ser. No. 15/940,636, titled ADAPTIVE CONTROL PROGRAM UPDATES FOR SURGICAL DEVICES, now U.S. Patent Application Publication No. 2019/0206003;
- U.S. patent application Ser. No. 15/940,653, titled ADAPTIVE CONTROL PROGRAM UPDATES FOR SURGICAL HUBS, now U.S. Patent Application Publication No. 2019/0201114;
- U.S. patent application Ser. No. 15/940,660, titled CLOUD-BASED MEDICAL ANALYTICS FOR CUSTOMIZATION AND RECOMMENDATIONS TO A USER, now U.S. Patent Application Publication No. 2019/0206555;
- U.S. patent application Ser. No. 15/940,679, titled CLOUD-BASED MEDICAL ANALYTICS FOR LINKING OF LOCAL USAGE TRENDS WITH THE RESOURCE ACQUISITION BEHAVIORS OF LARGER DATA SET, now U.S. Patent Application Publication No. 2019/0201144;
- U.S. patent application Ser. No. 15/940,694, titled CLOUD-BASED MEDICAL ANALYTICS FOR MEDICAL FACILITY SEGMENTED INDIVIDUALIZATION OF INSTRUMENT FUNCTION, now U.S. Patent Application Publication No. 2019/0201119;

U.S. patent application Ser. No. 15/940,634, titled CLOUD-BASED MEDICAL ANALYTICS FOR SECURITY AND AUTHENTICATION TRENDS AND REACTIVE MEASURES, now U.S. Patent Application Publication No. 2019/0201138; 5

U.S. patent application Ser. No. 15/940,706, titled DATA HANDLING AND PRIORITIZATION IN A CLOUD ANALYTICS NETWORK, now U.S. Patent Application Publication No. 2019/0206561; and

U.S. patent application Ser. No. 15/940,675, titled CLOUD INTERFACE FOR COUPLED SURGICAL DEVICES, now U.S. Pat. No. 10,849,697. 10

U.S. patent application Ser. No. 15/940,627, titled DRIVE ARRANGEMENTS FOR ROBOT-ASSISTED SURGICAL PLATFORMS, now U.S. Patent Application Publication No. 2019/0201111; 15

U.S. patent application Ser. No. 15/940,637, titled COMMUNICATION ARRANGEMENTS FOR ROBOT-ASSISTED SURGICAL PLATFORMS, now U.S. Patent Application Publication No. 2019/0201139; 20

U.S. patent application Ser. No. 15/940,642, titled CONTROLS FOR ROBOT-ASSISTED SURGICAL PLATFORMS, now U.S. Patent Application Publication No. 2019/0201113;

U.S. patent application Ser. No. 15/940,676, titled AUTOMATIC TOOL ADJUSTMENTS FOR ROBOT-ASSISTED SURGICAL PLATFORMS, now U.S. Patent Application Publication No. 2019/0201142; 25

U.S. patent application Ser. No. 15/940,680, titled CONTROLLERS FOR ROBOT-ASSISTED SURGICAL PLATFORMS, now U.S. Patent Application Publication No. 2019/0201135; 30

U.S. patent application Ser. No. 15/940,683, titled COOPERATIVE SURGICAL ACTIONS FOR ROBOT-ASSISTED SURGICAL PLATFORMS, now U.S. Patent Application Publication No. 2019/0201145; 35

U.S. patent application Ser. No. 15/940,690, titled DISPLAY ARRANGEMENTS FOR ROBOT-ASSISTED SURGICAL PLATFORMS, now U.S. Patent Application Publication No. 2019/0201118; and 40

U.S. patent application Ser. No. 15/940,711, titled SENSING ARRANGEMENTS FOR ROBOT-ASSISTED SURGICAL PLATFORMS, now U.S. Patent Application Publication No. 2019/0201120.

Applicant of the present application owns the following 45 U.S. Provisional patent applications, filed on Mar. 28, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

U.S. Provisional Patent Application Ser. No. 62/649,302, titled INTERACTIVE SURGICAL SYSTEMS WITH ENCRYPTED COMMUNICATION CAPABILITIES; 50

U.S. Provisional Patent Application Ser. No. 62/649,294, titled DATA STRIPPING METHOD TO INTERROGATE PATIENT RECORDS AND CREATE ANONYMIZED RECORD; 55

U.S. Provisional Patent Application Ser. No. 62/649,300, titled SURGICAL HUB SITUATIONAL AWARENESS;

U.S. Provisional Patent Application Ser. No. 62/649,309, titled SURGICAL HUB SPATIAL AWARENESS TO DETERMINE DEVICES IN OPERATING THEATER; 60

U.S. Provisional Patent Application Ser. No. 62/649,310, titled COMPUTER IMPLEMENTED INTERACTIVE SURGICAL SYSTEMS;

U.S. Provisional Patent Application Ser. No. 62/649,291, titled USE OF LASER LIGHT AND RED-GREEN-

BLUE COLORATION TO DETERMINE PROPERTIES OF BACK SCATTERED LIGHT;

U.S. Provisional Patent Application Ser. No. 62/649,296, titled ADAPTIVE CONTROL PROGRAM UPDATES FOR SURGICAL DEVICES;

U.S. Provisional Patent Application Ser. No. 62/649,333, titled CLOUD-BASED MEDICAL ANALYTICS FOR CUSTOMIZATION AND RECOMMENDATIONS TO A USER;

U.S. Provisional Patent Application Ser. No. 62/649,327, titled CLOUD-BASED MEDICAL ANALYTICS FOR SECURITY AND AUTHENTICATION TRENDS AND REACTIVE MEASURES;

U.S. Provisional Patent Application Ser. No. 62/649,315, titled DATA HANDLING AND PRIORITIZATION IN A CLOUD ANALYTICS NETWORK;

U.S. Provisional Patent Application Ser. No. 62/649,313, titled CLOUD INTERFACE FOR COUPLED SURGICAL DEVICES;

U.S. Provisional Patent Application Ser. No. 62/649,320, titled DRIVE ARRANGEMENTS FOR ROBOT-ASSISTED SURGICAL PLATFORMS;

U.S. Provisional Patent Application Ser. No. 62/649,307, titled AUTOMATIC TOOL ADJUSTMENTS FOR ROBOT-ASSISTED SURGICAL PLATFORMS; and

U.S. Provisional Patent Application Ser. No. 62/649,323, titled SENSING ARRANGEMENTS FOR ROBOT-ASSISTED SURGICAL PLATFORMS.

Applicant of the present application owns the following 30 U.S. Provisional patent applications, filed on Mar. 8, 2018, the disclosure of each of which is herein incorporated by reference in its entirety:

U.S. Provisional Patent Application Ser. No. 62/640,417, titled TEMPERATURE CONTROL IN ULTRASONIC DEVICE AND CONTROL SYSTEM THEREFOR; and

U.S. Provisional Patent Application Ser. No. 62/640,415, titled ESTIMATING STATE OF ULTRASONIC END EFFECTOR AND CONTROL SYSTEM THEREFOR. 40

Applicant of the present application owns the following U.S. Provisional patent applications, filed on Dec. 28, 2017, the disclosure of each of which is herein incorporated by reference in its entirety:

U.S. Provisional Patent Application Ser. No. 62/611,341, titled INTERACTIVE SURGICAL PLATFORM; 45

U.S. Provisional Patent Application Ser. No. 62/611,340, titled CLOUD-BASED MEDICAL ANALYTICS; and

U.S. Provisional Patent Application Ser. No. 62/611,339, titled ROBOT ASSISTED SURGICAL PLATFORM.

Before explaining various aspects of surgical devices and generators in detail, it should be noted that the illustrative 50 examples are not limited in application or use to the details of construction and arrangement of parts illustrated in the accompanying drawings and description. The illustrative examples may be implemented or incorporated in other 55 aspects, variations and modifications, and may be practiced or carried out in various ways. Further, unless otherwise indicated, the terms and expressions employed herein have been chosen for the purpose of describing the illustrative examples for the convenience of the reader and are not for the purpose of limitation thereof. Also, it will be appreciated that one or more of the following-described aspects, expressions of aspects, and/or examples, can be combined with any one or more of the other following-described aspects, 60 expressions of aspects and/or examples.

Various aspects are directed to improved ultrasonic surgical devices, electrosurgical devices and generators for use 65

therewith. Aspects of the ultrasonic surgical devices can be configured for transecting and/or coagulating tissue during surgical procedures, for example. Aspects of the electrosurgical devices can be configured for transecting, coagulating, scaling, welding and/or desiccating tissue during surgical procedures, for example.

Referring to FIG. 1, a computer-implemented interactive surgical system 100 includes one or more surgical systems 102 and a cloud-based system (e.g., the cloud 104 that may include a remote server 113 coupled to a storage device 105). Each surgical system 102 includes at least one surgical hub 106 in communication with the cloud 104 that may include a remote server 113. In one example, as illustrated in FIG. 1, the surgical system 102 includes a visualization system 108, a robotic system 110, and a handheld intelligent surgical instrument 112, which are configured to communicate with one another and/or the hub 106. In some aspects, a surgical system 102 may include an M number of hubs 106, an N number of visualization systems 108, an O number of robotic systems 110, and a P number of handheld intelligent surgical instruments 112, where M, N, O, and P are integers greater than or equal to one.

FIG. 3 depicts an example of a surgical system 102 being used to perform a surgical procedure on a patient who is lying down on an operating table 114 in a surgical operating room 116. A robotic system 110 is used in the surgical procedure as a part of the surgical system 102. The robotic system 110 includes a surgeon's console 118, a patient side cart 120 (surgical robot), and a surgical robotic hub 122. The patient side cart 120 can manipulate at least one removably coupled surgical tool 117 through a minimally invasive incision in the body of the patient while the surgeon views the surgical site through the surgeon's console 118. An image of the surgical site can be obtained by a medical imaging device 124, which can be manipulated by the patient side cart 120 to orient the imaging device 124. The robotic hub 122 can be used to process the images of the surgical site for subsequent display to the surgeon through the surgeon's console 118.

Other types of robotic systems can be readily adapted for use with the surgical system 102. Various examples of robotic systems and surgical tools that are suitable for use with the present disclosure are described in U.S. Provisional Patent Application Ser. No. 62/611,339, titled ROBOT ASSISTED SURGICAL PLATFORM, filed Dec. 28, 2017, the disclosure of which is herein incorporated by reference in its entirety.

Various examples of cloud-based analytics that are performed by the cloud 104, and are suitable for use with the present disclosure, are described in U.S. Provisional Patent Application Ser. No. 62/611,340, titled CLOUD-BASED MEDICAL ANALYTICS, filed Dec. 28, 2017, the disclosure of which is herein incorporated by reference in its entirety.

In various aspects, the imaging device 124 includes at least one image sensor and one or more optical components. Suitable image sensors include, but are not limited to, Charge-Coupled Device (CCD) sensors and Complementary Metal-Oxide Semiconductor (CMOS) sensors.

The optical components of the imaging device 124 may include one or more illumination sources and/or one or more lenses. The one or more illumination sources may be directed to illuminate portions of the surgical field. The one or more image sensors may receive light reflected or refracted from the surgical field, including light reflected or refracted from tissue and/or surgical instruments.

The one or more illumination sources may be configured to radiate electromagnetic energy in the visible spectrum as well as the invisible spectrum. The visible spectrum, sometimes referred to as the optical spectrum or luminous spectrum, is that portion of the electromagnetic spectrum that is visible to (i.e., can be detected by) the human eye and may be referred to as visible light or simply light. A typical human eye will respond to wavelengths in air that are from about 380 nm to about 750 nm.

The invisible spectrum (i.e., the non-luminous spectrum) is that portion of the electromagnetic spectrum that lies below and above the visible spectrum (i.e., wavelengths below about 380 nm and above about 750 nm). The invisible spectrum is not detectable by the human eye. Wavelengths greater than about 750 nm are longer than the red visible spectrum, and they become invisible infrared (IR), microwave, and radio electromagnetic radiation. Wavelengths less than about 380 nm are shorter than the violet spectrum, and they become invisible ultraviolet, x-ray, and gamma ray electromagnetic radiation.

In various aspects, the imaging device 124 is configured for use in a minimally invasive procedure. Examples of imaging devices suitable for use with the present disclosure include, but not limited to, an arthroscope, angioscope, bronchoscope, choledochoscope, colonoscope, cytoscope, duodenoscope, enteroscope, esophagogastro-duodenoscope (gastroscope), endoscope, laryngoscope, nasopharyngoscopy, sigmoidoscope, thoracoscope, and ureteroscopy.

In one aspect, the imaging device employs multi-spectrum monitoring to discriminate topography and underlying structures. A multi-spectral image is one that captures image data within specific wavelength ranges across the electromagnetic spectrum. The wavelengths may be separated by filters or by the use of instruments that are sensitive to particular wavelengths, including light from frequencies beyond the visible light range, e.g., IR and ultraviolet. Spectral imaging can allow extraction of additional information the human eye fails to capture with its receptors for red, green, and blue. The use of multi-spectral imaging is described in greater detail under the heading "Advanced Imaging Acquisition Module" in U.S. Provisional Patent Application Ser. No. 62/611,341, titled INTERACTIVE SURGICAL PLATFORM, filed Dec. 28, 2017, the disclosure of which is herein incorporated by reference in its entirety. Multi-spectrum monitoring can be a useful tool in relocating a surgical field after a surgical task is completed to perform one or more of the previously described tests on the treated tissue.

It is axiomatic that strict sterilization of the operating room and surgical equipment is required during any surgery. The strict hygiene and sterilization conditions required in a "surgical theater," i.e., an operating or treatment room, necessitate the highest possible sterility of all medical devices and equipment. Part of that sterilization process is the need to sterilize anything that comes in contact with the patient or penetrates the sterile field, including the imaging device 124 and its attachments and components. It will be appreciated that the sterile field may be considered a specified area, such as within a tray or on a sterile towel, that is considered free of microorganisms, or the sterile field may be considered an area, immediately around a patient, who has been prepared for a surgical procedure. The sterile field may include the scrubbed team members, who are properly attired, and all furniture and fixtures in the area.

In various aspects, the visualization system 108 includes one or more imaging sensors, one or more image-processing

units, one or more storage arrays, and one or more displays that are strategically arranged with respect to the sterile field, as illustrated in FIG. 2. In one aspect, the visualization system 108 includes an interface for HL7, PACS, and EMR. Various components of the visualization system 108 are described under the heading “Advanced Imaging Acquisition Module” in U.S. Provisional Patent Application Ser. No. 62/611,341, titled INTERACTIVE SURGICAL PLATFORM, filed Dec. 28, 2017, the disclosure of which is herein incorporated by reference in its entirety.

As illustrated in FIG. 2, a primary display 119 is positioned in the sterile field to be visible to an operator at the operating table 114. In addition, a visualization tower 111 is positioned outside the sterile field. The visualization tower 111 includes a first non-sterile display 107 and a second non-sterile display 109, which face away from each other. The visualization system 108, guided by the hub 106, is configured to utilize the displays 107, 109, and 119 to coordinate information flow to operators inside and outside the sterile field. For example, the hub 106 may cause the visualization system 108 to display a snapshot of a surgical site, as recorded by an imaging device 124, on a non-sterile display 107 or 109, while maintaining a live feed of the surgical site on the primary display 119. The snapshot on the non-sterile display 107 or 109 can permit a non-sterile operator to perform a diagnostic step relevant to the surgical procedure, for example.

In one aspect, the hub 106 is also configured to route a diagnostic input or feedback entered by a non-sterile operator at the visualization tower 111 to the primary display 119 within the sterile field, where it can be viewed by a sterile operator at the operating table. In one example, the input can be in the form of a modification to the snapshot displayed on the non-sterile display 107 or 109, which can be routed to the primary display 119 by the hub 106.

Referring to FIG. 2, a surgical instrument 112 is being used in the surgical procedure as part of the surgical system 102. The hub 106 is also configured to coordinate information flow to a display of the surgical instrument 112. For example, in U.S. Provisional Patent Application Ser. No. 62/611,341, titled INTERACTIVE SURGICAL PLATFORM, filed Dec. 28, 2017, the disclosure of which is herein incorporated by reference in its entirety. A diagnostic input or feedback entered by a non-sterile operator at the visualization tower 111 can be routed by the hub 106 to the surgical instrument display 115 within the sterile field, where it can be viewed by the operator of the surgical instrument 112. Example surgical instruments that are suitable for use with the surgical system 102 are described under the heading SURGICAL INSTRUMENT HARDWARE and in U.S. Provisional Patent Application Ser. No. 62/611,341, titled INTERACTIVE SURGICAL PLATFORM, filed Dec. 28, 2017, the disclosure of which is herein incorporated by reference in its entirety, for example.

Referring now to FIG. 3, a hub 106 is depicted in communication with a visualization system 108, a robotic system 110, and a handheld intelligent surgical instrument 112. The hub 106 includes a hub display 135, an imaging module 138, a generator module 140, a communication module 130, a processor module 132, and a storage array 134. In certain aspects, as illustrated in FIG. 3, the hub 106 further includes a smoke evacuation module 126 and/or a suction/irrigation module 128.

During a surgical procedure, energy application to tissue, for sealing and/or cutting, is generally associated with smoke evacuation, suction of excess fluid, and/or irrigation of the tissue. Fluid, power, and/or data lines from different

sources are often entangled during the surgical procedure. Valuable time can be lost addressing this issue during a surgical procedure. Detangling the lines may necessitate disconnecting the lines from their respective modules, which may require resetting the modules. The hub modular enclosure 136 offers a unified environment for managing the power, data, and fluid lines, which reduces the frequency of entanglement between such lines.

Aspects of the present disclosure present a surgical hub for use in a surgical procedure that involves energy application to tissue at a surgical site. The surgical hub includes a hub enclosure and a combo generator module slidably receivable in a docking station of the hub enclosure. The docking station includes data and power contacts. The combo generator module includes two or more of an ultrasonic energy generator component, a bipolar RF energy generator component, and a monopolar RF energy generator component that are housed in a single unit. In one aspect, the combo generator module also includes a smoke evacuation component, at least one energy delivery cable for connecting the combo generator module to a surgical instrument, at least one smoke evacuation component configured to evacuate smoke, fluid, and/or particulates generated by the application of therapeutic energy to the tissue, and a fluid line extending from the remote surgical site to the smoke evacuation component.

In one aspect, the fluid line is a first fluid line and a second fluid line extends from the remote surgical site to a suction and irrigation module slidably received in the hub enclosure. In one aspect, the hub enclosure comprises a fluid interface.

Certain surgical procedures may require the application of more than one energy type to the tissue. One energy type may be more beneficial for cutting the tissue, while another different energy type may be more beneficial for sealing the tissue. For example, a bipolar generator can be used to seal the tissue while an ultrasonic generator can be used to cut the sealed tissue. Aspects of the present disclosure present a solution where a hub modular enclosure 136 is configured to accommodate different generators, and facilitate an interactive communication therebetween. One of the advantages of the hub modular enclosure 136 is enabling the quick removal and/or replacement of various modules.

Aspects of the present disclosure present a modular surgical enclosure for use in a surgical procedure that involves energy application to tissue. The modular surgical enclosure includes a first energy-generator module, configured to generate a first energy for application to the tissue, and a first docking station comprising a first docking port that includes first data and power contacts, wherein the first energy-generator module is slidably movable into an electrical engagement with the power and data contacts and wherein the first energy-generator module is slidably movable out of the electrical engagement with the first power and data contacts,

Further to the above, the modular surgical enclosure also includes a second energy-generator module configured to generate a second energy, different than the first energy, for application to the tissue, and a second docking station comprising a second docking port that includes second data and power contacts, wherein the second energy-generator module is slidably movable into an electrical engagement with the power and data contacts, and wherein the second energy-generator module is slidably movable out of the electrical engagement with the second power and data contacts.

In addition, the modular surgical enclosure also includes a communication bus between the first docking port and the

second docking port, configured to facilitate communication between the first energy-generator module and the second energy-generator module.

Referring to FIGS. 3-7, aspects of the present disclosure are presented for a hub modular enclosure **136** that allows the modular integration of a generator module **140**, a smoke evacuation module **126**, and a suction/irrigation module **128**. The hub modular enclosure **136** further facilitates interactive communication between the modules **140**, **126**, **128**. As illustrated in FIG. 5, the generator module **140** can be a generator module with integrated monopolar, bipolar, and ultrasonic components supported in a single housing unit **139** slidably insertable into the hub modular enclosure **136**. As illustrated in FIG. 5, the generator module **140** can be configured to connect to a monopolar device **146**, a bipolar device **147**, and an ultrasonic device **148**. Alternatively, the generator module **140** may comprise a series of monopolar, bipolar, and/or ultrasonic generator modules that interact through the hub modular enclosure **136**. The hub modular enclosure **136** can be configured to facilitate the insertion of multiple generators and interactive communication between the generators docked into the hub modular enclosure **136** so that the generators would act as a single generator.

In one aspect, the hub modular enclosure **136** comprises a modular power and communication backplane **149** with external and wireless communication headers to enable the removable attachment of the modules **140**, **126**, **128** and interactive communication therebetween.

In one aspect, the hub modular enclosure **136** includes docking stations, or drawers, **151**, herein also referred to as drawers, which are configured to slidably receive the modules **140**, **126**, **128**. FIG. 4 illustrates a partial perspective view of a surgical hub enclosure **136**, and a combo generator module **145** slidably receivable in a docking station **151** of the surgical hub enclosure **136**. A docking port **152** with power and data contacts on a rear side of the combo generator module **145** is configured to engage a corresponding docking port **150** with power and data contacts of a corresponding docking station **151** of the hub modular enclosure **136** as the combo generator module **145** is slid into position within the corresponding docking station **151** of the hub module enclosure **136**. In one aspect, the combo generator module **145** includes a bipolar, ultrasonic, and monopolar module and a smoke evacuation module integrated together into a single housing unit **139**, as illustrated in FIG. 5.

In various aspects, the smoke evacuation module **126** includes a fluid line **154** that conveys captured/collected smoke and/or fluid away from a surgical site and to, for example, the smoke evacuation module **126**. Vacuum suction originating from the smoke evacuation module **126** can draw the smoke into an opening of a utility conduit at the surgical site. The utility conduit, coupled to the fluid line, can be in the form of a flexible tube terminating at the smoke evacuation module **126**. The utility conduit and the fluid line define a fluid path extending toward the smoke evacuation module **126** that is received in the hub enclosure **136**.

In various aspects, the suction/irrigation module **128** is coupled to a surgical tool comprising an aspiration fluid line and a suction fluid line. In one example, the aspiration and suction fluid lines are in the form of flexible tubes extending from the surgical site toward the suction/irrigation module **128**. One or more drive systems can be configured to cause irrigation and aspiration of fluids to and from the surgical site.

In one aspect, the surgical tool includes a shaft having an end effector at a distal end thereof and at least one energy

treatment associated with the end effector, an aspiration tube, and an irrigation tube. The aspiration tube can have an inlet port at a distal end thereof and the aspiration tube extends through the shaft. Similarly, an irrigation tube can extend through the shaft and can have an inlet port in proximity to the energy deliver implement. The energy deliver implement is configured to deliver ultrasonic and/or RF energy to the surgical site and is coupled to the generator module **140** by a cable extending initially through the shaft.

The irrigation tube can be in fluid communication with a fluid source, and the aspiration tube can be in fluid communication with a vacuum source. The fluid source and/or the vacuum source can be housed in the suction/irrigation module **128**. In one example, the fluid source and/or the vacuum source can be housed in the hub enclosure **136** separately from the suction/irrigation module **128**. In such example, a fluid interface can be configured to connect the suction/irrigation module **128** to the fluid source and/or the vacuum source.

In one aspect, the modules **140**, **126**, **128** and/or their corresponding docking stations on the hub modular enclosure **136** may include alignment features that are configured to align the docking ports of the modules into engagement with their counterparts in the docking stations of the hub modular enclosure **136**. For example, as illustrated in FIG. 4, the combo generator module **145** includes side brackets **155** that are configured to slidably engage with corresponding brackets **156** of the corresponding docking station **151** of the hub modular enclosure **136**. The brackets cooperate to guide the docking port contacts of the combo generator module **145** into an electrical engagement with the docking port contacts of the hub modular enclosure **136**.

In some aspects, the drawers **151** of the hub modular enclosure **136** are the same, or substantially the same size, and the modules are adjusted in size to be received in the drawers **151**. For example, the side brackets **155** and/or **156** can be larger or smaller depending on the size of the module. In other aspects, the drawers **151** are different in size and are each designed to accommodate a particular module.

Furthermore, the contacts of a particular module can be keyed for engagement with the contacts of a particular drawer to avoid inserting a module into a drawer with mismatching contacts.

As illustrated in FIG. 4, the docking port **150** of one drawer **151** can be coupled to the docking port **150** of another drawer **151** through a communications link **157** to facilitate an interactive communication between the modules housed in the hub modular enclosure **136**. The docking ports **150** of the hub modular enclosure **136** may alternatively, or additionally, facilitate a wireless interactive communication between the modules housed in the hub modular enclosure **136**. Any suitable wireless communication can be employed, such as for example Air Titan-Bluetooth.

FIG. 6 illustrates individual power bus attachments for a plurality of lateral docking ports of a lateral modular housing **160** configured to receive a plurality of modules of a surgical hub **206**. The lateral modular housing **160** is configured to laterally receive and interconnect the modules **161**. The modules **161** are slidably inserted into docking stations **162** of lateral modular housing **160**, which includes a backplane for interconnecting the modules **161**. As illustrated in FIG. 6, the modules **161** are arranged laterally in the lateral modular housing **160**. Alternatively, the modules **161** may be arranged vertically in a lateral modular housing.

FIG. 7 illustrates a vertical modular housing **164** configured to receive a plurality of modules **165** of the surgical hub **106**. The modules **165** are slidably inserted into docking

stations, or drawers, **167** of vertical modular housing **164**, which includes a backplane for interconnecting the modules **165**. Although the drawers **167** of the vertical modular housing **164** are arranged vertically, in certain instances, a vertical modular housing **164** may include drawers that are arranged laterally. Furthermore, the modules **165** may interact with one another through the docking ports of the vertical modular housing **164**. In the example of FIG. 7, a display **177** is provided for displaying data relevant to the operation of the modules **165**. In addition, the vertical modular housing **164** includes a master module **178** housing a plurality of sub-modules that are slidably received in the master module **178**.

In various aspects, the imaging module **138** comprises an integrated video processor and a modular light source and is adapted for use with various imaging devices. In one aspect, the imaging device is comprised of a modular housing that can be assembled with a light source module and a camera module. The housing can be a disposable housing. In at least one example, the disposable housing is removably coupled to a reusable controller, a light source module, and a camera module. The light source module and/or the camera module can be selectively chosen depending on the type of surgical procedure. In one aspect, the camera module comprises a CCD sensor. In another aspect, the camera module comprises a CMOS sensor. In another aspect, the camera module is configured for scanned beam imaging. Likewise, the light source module can be configured to deliver a white light or a different light, depending on the surgical procedure.

During a surgical procedure, removing a surgical device from the surgical field and replacing it with another surgical device that includes a different camera or a different light source can be inefficient. Temporarily losing sight of the surgical field may lead to undesirable consequences. The module imaging device of the present disclosure is configured to permit the replacement of a light source module or a camera module midstream during a surgical procedure, without having to remove the imaging device from the surgical field.

In one aspect, the imaging device comprises a tubular housing that includes a plurality of channels. A first channel is configured to slidably receive the camera module, which can be configured for a snap-fit engagement with the first channel. A second channel is configured to slidably receive the light source module, which can be configured for a snap-fit engagement with the second channel. In another example, the camera module and/or the light source module can be rotated into a final position within their respective channels. A threaded engagement can be employed in lieu of the snap-fit engagement.

In various examples, multiple imaging devices are placed at different positions in the surgical field to provide multiple views. The imaging module **138** can be configured to switch between the imaging devices to provide an optimal view. In various aspects, the imaging module **138** can be configured to integrate the images from the different imaging device.

Various image processors and imaging devices suitable for use with the present disclosure are described in U.S. Pat. No. 7,995,045, titled COMBINED SBI AND CONVENTIONAL IMAGE PROCESSOR, which issued on Aug. 9, 2011, which is herein incorporated by reference in its entirety. In addition, U.S. Pat. No. 7,982,776, titled SBI MOTION ARTIFACT REMOVAL APPARATUS AND METHOD, which issued on Jul. 19, 2011, which is herein incorporated by reference in its entirety, describes various systems for removing motion artifacts from image data. Such systems can be integrated with the imaging module

138. Furthermore, U.S. Patent Application Publication No. 2011/0306840, titled CONTROLLABLE MAGNETIC SOURCE TO FIXTURE INTRACORPOREAL APPARATUS, which published on Dec. 15, 2011, and U.S. Pat. No. 10,098,527, titled SYSTEM FOR PERFORMING A MINIMALLY INVASIVE SURGICAL PROCEDURE, which issued on Oct. 16, 2018, each of which is herein incorporated by reference in its entirety.

FIG. 8 illustrates a surgical data network **201** comprising a modular communication hub **203** configured to connect modular devices located in one or more operating theaters of a healthcare facility, or any room in a healthcare facility specially equipped for surgical operations, to a cloud-based system (e.g., the cloud **204** that may include a remote server **213** coupled to a storage device **205**). In one aspect, the modular communication hub **203** comprises a network hub **207** and/or a network switch **209** in communication with a network router. The modular communication hub **203** also can be coupled to a local computer system **210** to provide local computer processing and data manipulation. The surgical data network **201** may be configured as passive, intelligent, or switching. A passive surgical data network serves as a conduit for the data, enabling it to go from one device (or segment) to another and to the cloud computing resources. An intelligent surgical data network includes additional features to enable the traffic passing through the surgical data network to be monitored and to configure each port in the network hub **207** or network switch **209**. An intelligent surgical data network may be referred to as a manageable hub or switch. A switching hub reads the destination address of each packet and then forwards the packet to the correct port.

Modular devices **1a-1n** located in the operating theater may be coupled to the modular communication hub **203**. The network hub **207** and/or the network switch **209** may be coupled to a network router **211** to connect the devices **1a-1n** to the cloud **204** or the local computer system **210**. Data associated with the devices **1a-1n** may be transferred to cloud-based computers via the router for remote data processing and manipulation. Data associated with the devices **1a-1n** may also be transferred to the local computer system **210** for local data processing and manipulation. Modular devices **2a-2m** located in the same operating theater also may be coupled to a network switch **209**. The network switch **209** may be coupled to the network hub **207** and/or the network router **211** to connect to the devices **2a-2m** to the cloud **204**. Data associated with the devices **2a-2n** may be transferred to the cloud **204** via the network router **211** for data processing and manipulation. Data associated with the devices **2a-2m** may also be transferred to the local computer system **210** for local data processing and manipulation.

It will be appreciated that the surgical data network **201** may be expanded by interconnecting multiple network hubs **207** and/or multiple network switches **209** with multiple network routers **211**. The modular communication hub **203** may be contained in a modular control tower configured to receive multiple devices **1a-1n/2a-2m**. The local computer system **210** also may be contained in a modular control tower. The modular communication hub **203** is connected to a display **212** to display images obtained by some of the devices **1a-1n/2a-2m**, for example during surgical procedures. In various aspects, the devices **1a-1n/2a-2m** may include, for example, various modules such as an imaging module **138** coupled to an endoscope, a generator module **140** coupled to an energy-based surgical device, a smoke evacuation module **126**, a suction/irrigation module **128**, a communication module **130**, a processor module **132**, a

storage array **134**, a surgical device coupled to a display, and/or a non-contact sensor module, among other modular devices that may be connected to the modular communication hub **203** of the surgical data network **201**.

In one aspect, the surgical data network **201** may comprise a combination of network hub(s), network switch(es), and network router(s) connecting the devices **1a-1n/2a-2m** to the cloud. Any one of or all of the devices **1a-1n/2a-2m** coupled to the network hub or network switch may collect data in real time and transfer the data to cloud computers for data processing and manipulation. It will be appreciated that cloud computing relies on sharing computing resources rather than having local servers or personal devices to handle software applications. The word “cloud” may be used as a metaphor for “the Internet,” although the term is not limited as such. Accordingly, the term “cloud computing” may be used herein to refer to “a type of Internet-based computing,” where different services—such as servers, storage, and applications—are delivered to the modular communication hub **203** and/or computer system **210** located in the surgical theater (e.g., a fixed, mobile, temporary, or field operating room or space) and to devices connected to the modular communication hub **203** and/or computer system **210** through the Internet. The cloud infrastructure may be maintained by a cloud service provider. In this context, the cloud service provider may be the entity that coordinates the usage and control of the devices **1a-1n/2a-2m** located in one or more operating theaters. The cloud computing services can perform a large number of calculations based on the data gathered by smart surgical instruments, robots, and other computerized devices located in the operating theater. The hub hardware enables multiple devices or connections to be connected to a computer that communicates with the cloud computing resources and storage.

Applying cloud computer data processing techniques on the data collected by the devices **1a-1n/2a-2m**, the surgical data network provides improved surgical outcomes, reduced costs, and improved patient satisfaction. At least some of the devices **1a-1n/2a-2m** may be employed to view tissue states to assess leaks or perfusion of sealed tissue after a tissue sealing and cutting procedure. At least some of the devices **1a-1n/2a-2m** may be employed to identify pathology, such as the effects of diseases, using the cloud-based computing to examine data including images of samples of body tissue for diagnostic purposes. This includes localization and margin confirmation of tissue and phenotypes. At least some of the devices **1a-1n/2a-2m** may be employed to identify anatomical structures of the body using a variety of sensors integrated with imaging devices and techniques such as overlaying images captured by multiple imaging devices. The data gathered by the devices **1a-1n/2a-2m**, including image data, may be transferred to the cloud **204** or the local computer system **210** or both for data processing and manipulation including image processing and manipulation. The data may be analyzed to improve surgical procedure outcomes by determining if further treatment, such as the application of endoscopic intervention, emerging technologies, a targeted radiation, targeted intervention, and precise robotics to tissue-specific sites and conditions, may be pursued. Such data analysis may further employ outcome analytics processing, and using standardized approaches may provide beneficial feedback to either confirm surgical treatments and the behavior of the surgeon or suggest modifications to surgical treatments and the behavior of the surgeon.

In one implementation, the operating theater devices **1a-1n** may be connected to the modular communication hub

203 over a wired channel or a wireless channel depending on the configuration of the devices **1a-1n** to a network hub. The network hub **207** may be implemented, in one aspect, as a local network broadcast device that works on the physical layer of the Open System Interconnection (OSI) model. The network hub provides connectivity to the devices **1a-1n** located in the same operating theater network. The network hub **207** collects data in the form of packets and sends them to the router in half duplex mode. The network hub **207** does not store any media access control/Internet Protocol (MAC/IP) to transfer the device data. Only one of the devices **1a-1n** can send data at a time through the network hub **207**. The network hub **207** has no routing tables or intelligence regarding where to send information and broadcasts all network data across each connection and to a remote server **213** (FIG. 9) over the cloud **204**. The network hub **207** can detect basic network errors such as collisions, but having all information broadcast to multiple ports can be a security risk and cause bottlenecks.

In another implementation, the operating theater devices **2a-2m** may be connected to a network switch **209** over a wired channel or a wireless channel. The network switch **209** works in the data link layer of the OSI model. The network switch **209** is a multicast device for connecting the devices **2a-2m** located in the same operating theater to the network. The network switch **209** sends data in the form of frames to the network router **211** and works in full duplex mode. Multiple devices **2a-2m** can send data at the same time through the network switch **209**. The network switch **209** stores and uses MAC addresses of the devices **2a-2m** to transfer data.

The network hub **207** and/or the network switch **209** are coupled to the network router **211** for connection to the cloud **204**. The network router **211** works in the network layer of the OSI model. The network router **211** creates a route for transmitting data packets received from the network hub **207** and/or network switch **211** to cloud-based computer resources for further processing and manipulation of the data collected by any one of or all the devices **1a-1n/2a-2m**. The network router **211** may be employed to connect two or more different networks located in different locations, such as, for example, different operating theaters of the same healthcare facility or different networks located in different operating theaters of different healthcare facilities. The network router **211** sends data in the form of packets to the cloud **204** and works in full duplex mode. Multiple devices can send data at the same time. The network router **211** uses IP addresses to transfer data.

In one example, the network hub **207** may be implemented as a USB hub, which allows multiple USB devices to be connected to a host computer. The USB hub may expand a single USB port into several tiers so that there are more ports available to connect devices to the host system computer. The network hub **207** may include wired or wireless capabilities to receive information over a wired channel or a wireless channel. In one aspect, a wireless USB short-range, high-bandwidth wireless radio communication protocol may be employed for communication between the devices **1a-1n** and devices **2a-2m** located in the operating theater.

In other examples, the operating theater devices **1a-1n/2a-2m** may communicate to the modular communication hub **203** via Bluetooth wireless technology standard for exchanging data over short distances (using short-wavelength UHF radio waves in the ISM band from 2.4 to 2.485 GHz) from fixed and mobile devices and building personal area networks (PANs). In other aspects, the operating theater

devices **1a-1n/2a-2m** may communicate to the modular communication hub **203** via a number of wireless or wired communication standards or protocols, including but not limited to Wi-Fi (IEEE 802.11 family), WiMAX (IEEE 802.16 family), IEEE 802.20, long-term evolution (LTE), and Ev-DO, HSPA+, HSDPA+, HSUPA+, EDGE, GSM, GPRS, CDMA, TDMA, DECT, and Ethernet derivatives thereof, as well as any other wireless and wired protocols that are designated as 3G, 4G, 5G, and beyond. The computing module may include a plurality of communication modules. For instance, a first communication module may be dedicated to shorter-range wireless communications such as Wi-Fi and Bluetooth, and a second communication module may be dedicated to longer-range wireless communications such as GPS, EDGE, GPRS, CDMA, WiMAX, LTE, Ev-DO, and others.

The modular communication hub **203** may serve as a central connection for one or all of the operating theater devices **1a-1n/2a-2m** and handles a data type known as frames. Frames carry the data generated by the devices **1a-1n/2a-2m**. When a frame is received by the modular communication hub **203**, it is amplified and transmitted to the network router **211**, which transfers the data to the cloud computing resources by using a number of wireless or wired communication standards or protocols, as described herein.

The modular communication hub **203** can be used as a standalone device or be connected to compatible network hubs and network switches to form a larger network. The modular communication hub **203** is generally easy to install, configure, and maintain, making it a good option for networking the operating theater devices **1a-1n/2a-2m**.

FIG. 9 illustrates a computer-implemented interactive surgical system **200**. The computer-implemented interactive surgical system **200** is similar in many respects to the computer-implemented interactive surgical system **100**. For example, the computer-implemented interactive surgical system **200** includes one or more surgical systems **202**, which are similar in many respects to the surgical systems **102**. Each surgical system **202** includes at least one surgical hub **206** in communication with a cloud **204** that may include a remote server **213**. In one aspect, the computer-implemented interactive surgical system **200** comprises a modular control tower **236** connected to multiple operating theater devices such as, for example, intelligent surgical instruments, robots, and other computerized devices located in the operating theater. As shown in FIG. 10, the modular control tower **236** comprises a modular communication hub **203** coupled to a computer system **210**. As illustrated in the example of FIG. 9, the modular control tower **236** is coupled to an imaging module **238** that is coupled to an endoscope **239**, a generator module **240** that is coupled to an energy device **241**, a smoke evacuator module **226**, a suction/irrigation module **228**, a communication module **230**, a processor module **232**, a storage array **234**, a smart device/instrument **235** optionally coupled to a display **237**, and a non-contact sensor module **242**. The operating theater devices are coupled to cloud computing resources and data storage via the modular control tower **236**. A robot hub **222** also may be connected to the modular control tower **236** and to the cloud computing resources. The devices/instruments **235**, visualization systems **208**, among others, may be coupled to the modular control tower **236** via wired or wireless communication standards or protocols, as described herein. The modular control tower **236** may be coupled to a hub display **215** (e.g., monitor, screen) to display and overlay images received from the imaging module, device/instrument display, and/or other visualization systems **208**.

The hub display also may display data received from devices connected to the modular control tower in conjunction with images and overlaid images.

FIG. 10 illustrates a surgical hub **206** comprising a plurality of modules coupled to the modular control tower **236**. The modular control tower **236** comprises a modular communication hub **203**, e.g., a network connectivity device, and a computer system **210** to provide local processing, visualization, and imaging, for example. As shown in FIG. 10, the modular communication hub **203** may be connected in a tiered configuration to expand the number of modules (e.g., devices) that may be connected to the modular communication hub **203** and transfer data associated with the modules to the computer system **210**, cloud computing resources, or both. As shown in FIG. 10, each of the network hubs/switches in the modular communication hub **203** includes three downstream ports and one upstream port. The upstream network hub/switch is connected to a processor to provide a communication connection to the cloud computing resources and a local display **217**. Communication to the cloud **204** may be made either through a wired or a wireless communication channel.

The surgical hub **206** employs a non-contact sensor module **242** to measure the dimensions of the operating theater and generate a map of the surgical theater using either ultrasonic or laser-type non-contact measurement devices. An ultrasound-based non-contact sensor module scans the operating theater by transmitting a burst of ultrasound and receiving the echo when it bounces off the perimeter walls of an operating theater as described under the heading "Surgical Hub Spatial Awareness Within an Operating Room" in U.S. Provisional Patent Application Ser. No. 62/611,341, titled INTERACTIVE SURGICAL PLATFORM, filed Dec. 28, 2017, which is herein incorporated by reference in its entirety, in which the sensor module is configured to determine the size of the operating theater and to adjust Bluetooth-pairing distance limits. A laser-based non-contact sensor module scans the operating theater by transmitting laser light pulses, receiving laser light pulses that bounce off the perimeter walls of the operating theater, and comparing the phase of the transmitted pulse to the received pulse to determine the size of the operating theater and to adjust Bluetooth pairing distance limits, for example.

The computer system **210** comprises a processor **244** and a network interface **245**. The processor **244** is coupled to a communication module **247**, storage **248**, memory **249**, non-volatile memory **250**, and input/output interface **251** via a system bus. The system bus can be any of several types of bus structure(s) including the memory bus or memory controller, a peripheral bus or external bus, and/or a local bus using any variety of available bus architectures including, but not limited to, 9-bit bus, Industrial Standard Architecture (ISA), Micro-Charmel Architecture (MSA), Extended ISA (EISA), Intelligent Drive Electronics (IDE), VESA Local Bus (VLB), Peripheral Component Interconnect (PCI), USB, Advanced Graphics Port (AGP), Personal Computer Memory Card International Association bus (PCMCIA), Small Computer Systems Interface (SCSI), or any other proprietary bus.

The processor **244** may be any single-core or multicore processor such as those known under the trade name ARM Cortex by Texas Instruments. In one aspect, the processor may be an LM4F230H5QR ARM Cortex-M4F Processor Core, available from Texas Instruments, for example, comprising an on-chip memory of 256 KB single-cycle flash memory, or other non-volatile memory, up to 40 MHz, a prefetch buffer to improve performance above 40 MHz, a 32

KB single-cycle serial random access memory (SRAM), an internal read-only memory (ROM) loaded with StellarisWare® software, a 2 KB electrically erasable programmable read-only memory (EEPROM), and/or one or more pulse width modulation (PWM) modules, one or more quadrature encoder inputs (QEI) analogs, one or more 12-bit analog-to-digital converters (ADCs) with 12 analog input channels, details of which are available for the product datasheet.

In one aspect, the processor **244** may comprise a safety controller comprising two controller-based families such as TMS570 and RM4x, known under the trade name Hercules ARM Cortex R4, also by Texas Instruments. The safety controller may be configured specifically for IEC 61508 and ISO 26262 safety critical applications, among others, to provide advanced integrated safety features while delivering scalable performance, connectivity, and memory options.

The system memory includes volatile memory and non-volatile memory. The basic input/output system (BIOS), containing the basic routines to transfer information between elements within the computer system, such as during start-up, is stored in non-volatile memory. For example, the non-volatile memory can include ROM, programmable ROM (PROM), electrically programmable ROM (EPROM), EEPROM, or flash memory. Volatile memory includes random-access memory (RAM), which acts as external cache memory. Moreover, RAM is available in many forms such as SRAM, dynamic RAM (DRAM), synchronous DRAM (SDRAM), double data rate SDRAM (DDR SDRAM), enhanced SDRAM (ESDRAM), Synchlink DRAM (SL-DRAM), and direct Rambus RAM (DRRAM).

The computer system **210** also includes removable/non-removable, volatile/non-volatile computer storage media, such as for example disk storage. The disk storage includes, but is not limited to, devices like a magnetic disk drive, floppy disk drive, tape drive, Jaz drive, Zip drive, LS-60 drive, flash memory card, or memory stick. In addition, the disk storage can include storage media separately or in combination with other storage media including, but not limited to, an optical disc drive such as a compact disc ROM device (CD-ROM), compact disc recordable drive (CD-R Drive), compact disc rewritable drive (CD-RW Drive), or a digital versatile disc ROM drive (DVD-ROM). To facilitate the connection of the disk storage devices to the system bus, a removable or non-removable interface may be employed.

It is to be appreciated that the computer system **210** includes software that acts as an intermediary between users and the basic computer resources described in a suitable operating environment. Such software includes an operating system. The operating system, which can be stored on the disk storage, acts to control and allocate resources of the computer system. System applications take advantage of the management of resources by the operating system through program modules and program data stored either in the system memory or on the disk storage. It is to be appreciated that various components described herein can be implemented with various operating systems or combinations of operating systems.

A user enters commands or information into the computer system **210** through input device(s) coupled to the I/O interface **251**. The input devices include, but are not limited to, a pointing device such as a mouse, trackball, stylus, touch pad, keyboard, microphone, joystick, game pad, satellite dish, scanner, TV tuner card, digital camera, digital video camera, web camera, and the like. These and other input devices connect to the processor through the system bus via interface port(s). The interface port(s) include, for example,

a serial port, a parallel port, a game port, and a USB. The output device(s) use some of the same types of ports as input device(s). Thus, for example, a USB port may be used to provide input to the computer system and to output information from the computer system to an output device. An output adapter is provided to illustrate that there are some output devices like monitors, displays, speakers, and printers, among other output devices that require special adapters. The output adapters include, by way of illustration and not limitation, video and sound cards that provide a means of connection between the output device and the system bus. It should be noted that other devices and/or systems of devices, such as remote computer(s), provide both input and output capabilities.

The computer system **210** can operate in a networked environment using logical connections to one or more remote computers, such as cloud computer(s), or local computers. The remote cloud computer(s) can be a personal computer, server, router, network PC, workstation, micro-processor-based appliance, peer device, or other common network node, and the like, and typically includes many or all of the elements described relative to the computer system. For purposes of brevity, only a memory storage device is illustrated with the remote computer(s). The remote computer(s) is logically connected to the computer system through a network interface and then physically connected via a communication connection. The network interface encompasses communication networks such as local area networks (LANs) and wide area networks (WANs). LAN technologies include Fiber Distributed Data Interface (FDDI), Copper Distributed Data Interface (CDDI), Ethernet/IEEE 802.3, Token Ring/IEEE 802.5 and the like. WAN technologies include, but are not limited to, point-to-point links, circuit-switching networks like Integrated Services Digital Networks (ISDN) and variations thereon, packet-switching networks, and Digital Subscriber Lines (DSL).

In various aspects, the computer system **210** of FIG. **10**, the imaging module **238** and/or visualization system **208**, and/or the processor module **232** of FIGS. **9-10**, may comprise an image processor, image-processing engine, media processor, or any specialized digital signal processor (DSP) used for the processing of digital images. The image processor may employ parallel computing with single instruction, multiple data (SIMD) or multiple instruction, multiple data (MIMD) technologies to increase speed and efficiency. The digital image-processing engine can perform a range of tasks. The image processor may be a system on a chip with multicore processor architecture.

The communication connection(s) refers to the hardware/software employed to connect the network interface to the bus. While the communication connection is shown for illustrative clarity inside the computer system, it can also be external to the computer system **210**. The hardware/software necessary for connection to the network interface includes, for illustrative purposes only, internal and external technologies such as modems, including regular telephone-grade modems, cable modems, and DSL modems, ISDN adapters, and Ethernet cards.

FIG. **11** illustrates a functional block diagram of one aspect of a USB network hub **300** device, in accordance with at least one aspect of the present disclosure. In the illustrated aspect, the USB network hub device **300** employs a TUSB2036 integrated circuit hub by Texas Instruments. The USB network hub **300** is a CMOS device that provides an upstream USB transceiver port **302** and up to three downstream USB transceiver ports **304**, **306**, **308** in compliance

with the USB 2.0 specification. The upstream USB transceiver port **302** is a differential root data port comprising a differential data minus (DM0) input paired with a differential data plus (DP0) input. The three downstream USB transceiver ports **304**, **306**, **308** are differential data ports where each port includes differential data plus (DP1-DP3) outputs paired with differential data minus (DM1-DM3) outputs.

The USB network hub **300** device is implemented with a digital state machine instead of a microcontroller, and no firmware programming is required. Fully compliant USB transceivers are integrated into the circuit for the upstream USB transceiver port **302** and all downstream USB transceiver ports **304**, **306**, **308**. The downstream USB transceiver ports **304**, **306**, **308** support both full-speed and low-speed devices by automatically setting the slew rate according to the speed of the device attached to the ports. The USB network hub **300** device may be configured either in bus-powered or self-powered mode and includes a hub power logic **312** to manage power.

The USB network hub **300** device includes a serial interface engine **310** (SIE). The SIE **310** is the front end of the USB network hub **300** hardware and handles most of the protocol described in chapter 8 of the USB specification. The SIE **310** typically comprehends signaling up to the transaction level. The functions that it handles could include: packet recognition, transaction sequencing, SOP, EOP, RESET, and RESUME signal detection/generation, clock/data separation, non-return-to-zero invert (NRZI) data encoding/decoding and bit-stuffing, CRC generation and checking (token and data), packet ID (PID) generation and checking/decoding, and/or serial-parallel/parallel-serial conversion. The **310** receives a clock input **314** and is coupled to a suspend/resume logic and frame timer **316** circuit and a hub repeater circuit **318** to control communication between the upstream USB transceiver port **302** and the downstream USB transceiver ports **304**, **306**, **308** through port logic circuits **320**, **322**, **324**. The SIE **310** is coupled to a command decoder **326** via interface logic to control commands from a serial EEPROM via a serial EEPROM interface **330**.

In various aspects, the USB network hub **300** can connect **127** functions configured in up to six logical layers (tiers) to a single computer. Further, the USB network hub **300** can connect to all peripherals using a standardized four-wire cable that provides both communication and power distribution. The power configurations are bus-powered and self-powered modes. The USB network hub **300** may be configured to support four modes of power management: a bus-powered hub, with either individual-port power management or ganged-port power management, and the self-powered hub, with either individual-port power management or ganged-port power management. In one aspect, using a USB cable, the USB network hub **300**, the upstream USB transceiver port **302** is plugged into a USB host controller, and the downstream USB transceiver ports **304**, **306**, **308** are exposed for connecting USB compatible devices, and so forth.

Surgical Instrument Hardware

FIG. 12 illustrates a logic diagram of a control system **470** of a surgical instrument or tool in accordance with one or more aspects of the present disclosure. The system **470** comprises a control circuit. The control circuit includes a microcontroller **461** comprising a processor **462** and a memory **468**. One or more of sensors **472**, **474**, **476**, for example, provide real-time feedback to the processor **462**. A motor **482**, driven by a motor driver **492**, operably couples a longitudinally movable displacement member to drive a

clamp arm closure member. A tracking system **480** is configured to determine the position of the longitudinally movable displacement member. The position information is provided to the processor **462**, which can be programmed or configured to determine the position of the longitudinally movable drive member as well as the position of the closure member. Additional motors may be provided at the tool driver interface to control closure tube travel, shaft rotation, articulation, or clamp arm closure, or a combination of the above. A display **473** displays a variety of operating conditions of the instruments and may include touch screen functionality for data input. Information displayed on the display **473** may be overlaid with images acquired via endoscopic imaging modules.

In one aspect, the microcontroller **461** may be any single-core or multicore processor such as those known under the trade name ARM Cortex by Texas Instruments. In one aspect, the main microcontroller **461** may be an LM4F230H5QR ARM Cortex-M4F Processor Core, available from Texas Instruments, for example, comprising an on-chip memory of 256 KB single-cycle flash memory, or other non-volatile memory, up to 40 MHz, a prefetch buffer to improve performance above 40 MHz, a 32 KB single-cycle SRAM, and internal ROM loaded with StellarisWare® software, a 2 KB EEPROM, one or more PWM modules, one or more QEI analogs, and/or one or more 12-bit ADCs with 12 analog input channels, details of which are available for the product datasheet.

In one aspect, the microcontroller **461** may comprise a safety controller comprising two controller-based families such as TMS570 and RM4x, known under the trade name Hercules ARM Cortex R4, also by Texas Instruments. The safety controller may be configured specifically for IEC 61508 and ISO 26262 safety critical applications, among others, to provide advanced integrated safety features while delivering scalable performance, connectivity, and memory options.

The microcontroller **461** may be programmed to perform various functions such as precise control over the speed and position of the knife, articulation systems, clamp arm, or a combination of the above. In one aspect, the microcontroller **461** includes a processor **462** and a memory **468**. The electric motor **482** may be a brushed direct current (DC) motor with a gearbox and mechanical links to an articulation or knife system. In one aspect, a motor driver **492** may be an A3941 available from Allegro Microsystems, Inc. Other motor drivers may be readily substituted for use in the tracking system **480** comprising an absolute positioning system. A detailed description of an absolute positioning system is described in U.S. Patent Application Publication No. 2017/0296213, titled SYSTEMS AND METHODS FOR CONTROLLING A SURGICAL STAPLING AND CUTTING INSTRUMENT, which published on Oct. 19, 2017, which is herein incorporated by reference in its entirety.

The microcontroller **461** may be programmed to provide precise control over the speed and position of displacement members and articulation systems. The microcontroller **461** may be configured to compute a response in the software of the microcontroller **461**. The computed response is compared to a measured response of the actual system to obtain an "observed" response, which is used for actual feedback decisions. The observed response is a favorable, tuned value that balances the smooth, continuous nature of the simulated response with the measured response, which can detect outside influences on the system.

In one aspect, the motor **482** may be controlled by the motor driver **492** and can be employed by the firing system of the surgical instrument or tool. In various forms, the motor **482** may be a brushed DC driving motor having a maximum rotational speed of approximately 25,000 RPM. In other arrangements, the motor **482** may include a brushless motor, a cordless motor, a synchronous motor, a stepper motor, or any other suitable electric motor. The motor driver **492** may comprise an H-bridge driver comprising field-effect transistors (FETs), for example. The motor **482** can be powered by a power assembly releasably mounted to the handle assembly or tool housing for supplying control power to the surgical instrument or tool. The power assembly may comprise a battery which may include a number of battery cells connected in series that can be used as the power source to power the surgical instrument or tool. In certain circumstances, the battery cells of the power assembly may be replaceable and/or rechargeable battery cells. In at least one example, the battery cells can be lithium-ion batteries which can be coupleable to and separable from the power assembly.

The motor driver **492** may be an A3941 available from Allegro Microsystems, Inc. The A3941 **492** is a full-bridge controller for use with external N-channel power metal-oxide semiconductor field-effect transistors (MOSFETs) specifically designed for inductive loads, such as brush DC motors. The driver **492** comprises a unique charge pump regulator that provides full (>10 V) gate drive for battery voltages down to 7 V and allows the A3941 to operate with a reduced gate drive, down to 5.5 V. A bootstrap capacitor may be employed to provide the above battery supply voltage required for N-channel MOSFETs. An internal charge pump for the high-side drive allows DC (100% duty cycle) operation. The full bridge can be driven in fast or slow decay modes using diode or synchronous rectification. In the slow decay mode, current recirculation can be through the high-side or the low-side FETs. The power FETs are protected from shoot-through by resistor-adjustable dead time. Integrated diagnostics provide indications of undervoltage, overtemperature, and power bridge faults and can be configured to protect the power MOSFETs under most short circuit conditions. Other motor drivers may be readily substituted for use in the tracking system **480** comprising an absolute positioning system.

The tracking system **480** comprises a controlled motor drive circuit arrangement comprising a position sensor **472** according to one aspect of this disclosure. The position sensor **472** for an absolute positioning system provides a unique position signal corresponding to the location of a displacement member. In one aspect, the displacement member represents a longitudinally movable drive member comprising a rack of drive teeth for meshing engagement with a corresponding drive gear of a gear reducer assembly. In other aspects, the displacement member represents the firing member, which could be adapted and configured to include a rack of drive teeth. In yet another aspect, the displacement member represents a longitudinal displacement member to open and close a clamp arm, which can be adapted and configured to include a rack of drive teeth. In other aspects, the displacement member represents a clamp arm closure member configured to close and to open a clamp arm of a stapler, ultrasonic, or electrosurgical device, or combinations of the above. Accordingly, as used herein, the term displacement member is used generically to refer to any movable member of the surgical instrument or tool such as the drive member, the clamp arm, or any element that can be displaced. Accordingly, the absolute positioning system can,

in effect, track the displacement of the clamp arm by tracking the linear displacement of the longitudinally movable drive member.

In other aspects, the absolute positioning system can be configured to track the position of a clamp arm in the process of closing or opening. In various other aspects, the displacement member may be coupled to any position sensor **472** suitable for measuring linear displacement. Thus, the longitudinally movable drive member, or clamp arm, or combinations thereof, may be coupled to any suitable linear displacement sensor. Linear displacement sensors may include contact or non-contact displacement sensors. Linear displacement sensors may comprise linear variable differential transformers (LVDT), differential variable reluctance transducers (DVRT), a slide potentiometer, a magnetic sensing system comprising a movable magnet and a series of linearly arranged Hall effect sensors, a magnetic sensing system comprising a fixed magnet and a series of movable, linearly arranged Hall effect sensors, an optical sensing system comprising a movable light source and a series of linearly arranged photo diodes or photo detectors, an optical sensing system comprising a fixed light source and a series of movable linearly, arranged photo diodes or photo detectors, or any combination thereof.

The electric motor **482** can include a rotatable shaft that operably interfaces with a gear assembly that is mounted in meshing engagement with a set, or rack, of drive teeth on the displacement member. A sensor element may be operably coupled to a gear assembly such that a single revolution of the position sensor **472** element corresponds to some linear longitudinal translation of the displacement member. An arrangement of gearing and sensors can be connected to the linear actuator, via a rack and pinion arrangement, or a rotary actuator, via a spur gear or other connection. A power source supplies power to the absolute positioning system and an output indicator may display the output of the absolute positioning system. The displacement member represents the longitudinally movable drive member comprising a rack of drive teeth formed thereon for meshing engagement with a corresponding drive gear of the gear reducer assembly. The displacement member represents the longitudinally movable firing member to open and close a clamp arm.

A single revolution of the sensor element associated with the position sensor **472** is equivalent to a longitudinal linear displacement d_1 of the of the displacement member, where d_1 is the longitudinal linear distance that the displacement member moves from point "a" to point "b" after a single revolution of the sensor element coupled to the displacement member. The sensor arrangement may be connected via a gear reduction that results in the position sensor **472** completing one or more revolutions for the full stroke of the displacement member. The position sensor **472** may complete multiple revolutions for the full stroke of the displacement member.

A series of switches, where n is an integer greater than one, may be employed alone or in combination with a gear reduction to provide a unique position signal for more than one revolution of the position sensor **472**. The state of the switches are fed back to the microcontroller **461** that applies logic to determine a unique position signal corresponding to the longitudinal linear displacement $d_1+d_2+ \dots d_n$ of the displacement member. The output of the position sensor **472** is provided to the microcontroller **461**. The position sensor **472** of the sensor arrangement may comprise a magnetic sensor, an analog rotary sensor like a potentiometer, or an array of analog Hall-effect elements, which output a unique combination of position signals or values.

The position sensor **472** may comprise any number of magnetic sensing elements, such as, for example, magnetic sensors classified according to whether they measure the total magnetic field or the vector components of the magnetic field. The techniques used to produce both types of magnetic sensors encompass many aspects of physics and electronics. The technologies used for magnetic field sensing include search coil, fluxgate, optically pumped, nuclear precession, SQUID, Hall-effect, anisotropic magnetoresistance, giant magnetoresistance, magnetic tunnel junctions, giant magnetoimpedance, magnetostrictive/piezoelectric composites, magnetodiode, magnetotransistor, fiber-optic, magneto-optic, and microelectromechanical systems-based magnetic sensors, among others.

In one aspect, the position sensor **472** for the tracking system **480** comprising an absolute positioning system comprises a magnetic rotary absolute positioning system. The position sensor **472** may be implemented as an AS5055EFT single-chip magnetic rotary position sensor available from Austria Microsystems, AG. The position sensor **472** is interfaced with the microcontroller **461** to provide an absolute positioning system. The position sensor **472** is a low-voltage and low-power component and includes four Hall-effect elements in an area of the position sensor **472** that is located above a magnet. A high-resolution ADC and a smart power management controller are also provided on the chip. A coordinate rotation digital computer (CORDIC) processor, also known as the digit-by-digit method and Volder's algorithm, is provided to implement a simple and efficient algorithm to calculate hyperbolic and trigonometric functions that require only addition, subtraction, bitshift, and table lookup operations. The angle position, alarm bits, and magnetic field information are transmitted over a standard serial communication interface, such as a serial peripheral interface (SPI) interface, to the microcontroller **461**. The position sensor **472** provides 12 or 14 bits of resolution. The position sensor **472** may be an AS5055 chip provided in a small QFN 16-pin 4x4x0.85 mm package.

The tracking system **480** comprising an absolute positioning system may comprise and/or be programmed to implement a feedback controller, such as a PID, state feedback, and adaptive controller. A power source converts the signal from the feedback controller into a physical input to the system: in this case the voltage. Other examples include a PWM of the voltage, current, and force. Other sensor(s) may be provided to measure physical parameters of the physical system in addition to the position measured by the position sensor **472**. In some aspects, the other sensor(s) can include sensor arrangements such as those described in U.S. Pat. No. 9,345,481, titled STAPLE CARTRIDGE TISSUE THICKNESS SENSOR SYSTEM, which issued on May 24, 2016, which is herein incorporated by reference in its entirety; U.S. Patent Application Publication No. 2014/0263552, titled STAPLE CARTRIDGE TISSUE THICKNESS SENSOR SYSTEM, which published on Sep. 18, 2014, which is herein incorporated by reference in its entirety; and U.S. patent application Ser. No. 15/628,175, titled TECHNIQUES FOR ADAPTIVE CONTROL OF MOTOR VELOCITY OF A SURGICAL STAPLING AND CUTTING INSTRUMENT, filed Jun. 20, 2017, now U.S. Patent Application Publication No. 2018/0360452, which is herein incorporated by reference in its entirety. In a digital signal processing system, an absolute positioning system is coupled to a digital data acquisition system where the output of the absolute positioning system will have a finite resolution and sampling frequency. The absolute positioning

system may comprise a compare-and-combine circuit to combine a computed response with a measured response using algorithms, such as a weighted average and a theoretical control loop, that drive the computed response towards the measured response. The computed response of the physical system takes into account properties like mass, inertia, viscous friction, inductance resistance, etc., to predict what the states and outputs of the physical system will be by knowing the input.

The absolute positioning system provides an absolute position of the displacement member upon power-up of the instrument, without retracting or advancing the displacement member to a reset (zero or home) position as may be required with conventional rotary encoders that merely count the number of steps forwards or backwards that the motor **482** has taken to infer the position of a device actuator, drive bar, knife, or the like.

A sensor **474**, such as, for example, a strain gauge or a micro-strain gauge, is configured to measure one or more parameters of the end effector, such as, for example, the amplitude of the strain exerted on the anvil during a clamping operation, which can be indicative of the closure forces applied to the anvil. The measured strain is converted to a digital signal and provided to the processor **462**. Alternatively, or in addition to the sensor **474**, a sensor **476**, such as, for example, a load sensor, can measure the closure force applied by the closure drive system to the anvil in a stapler or a clamp arm in an ultrasonic or electrosurgical instrument. The sensor **476**, such as, for example, a load sensor, can measure the firing force applied to a closure member coupled to a clamp arm of the surgical instrument or tool or the force applied by a clamp arm to tissue located in the jaws of an ultrasonic or electrosurgical instrument. Alternatively, a current sensor **478** can be employed to measure the current drawn by the motor **482**. The displacement member also may be configured to engage a clamp arm to open or close the clamp arm. The force sensor may be configured to measure the clamping force on tissue. The force required to advance the displacement member can correspond to the current drawn by the motor **482**, for example. The measured force is converted to a digital signal and provided to the processor **462**.

In one form, the strain gauge sensor **474** can be used to measure the force applied to the tissue by the end effector. A strain gauge can be coupled to the end effector to measure the force on the tissue being treated by the end effector. A system for measuring forces applied to the tissue grasped by the end effector comprises a strain gauge sensor **474**, such as, for example, a micro-strain gauge, that is configured to measure one or more parameters of the end effector, for example. In one aspect, the strain gauge sensor **474** can measure the amplitude or magnitude of the strain exerted on a jaw member of an end effector during a clamping operation, which can be indicative of the tissue compression. The measured strain is converted to a digital signal and provided to a processor **462** of the microcontroller **461**. A load sensor **476** can measure the force used to operate the knife element, for example, to cut the tissue captured between the anvil and the staple cartridge. A load sensor **476** can measure the force used to operate the clamp arm element, for example, to capture tissue between the clamp arm and an ultrasonic blade or to capture tissue between the clamp arm and a jaw of an electrosurgical instrument. A magnetic field sensor can be employed to measure the thickness of the captured tissue. The measurement of the magnetic field sensor also may be converted to a digital signal and provided to the processor **462**.

The measurements of the tissue compression, the tissue thickness, and/or the force required to close the end effector on the tissue, as respectively measured by the sensors **474**, **476**, can be used by the microcontroller **461** to characterize the selected position of the firing member and/or the corresponding value of the speed of the firing member. In one instance, a memory **468** may store a technique, an equation, and/or a lookup table which can be employed by the microcontroller **461** in the assessment.

The control system **470** of the surgical instrument or tool also may comprise wired or wireless communication circuits to communicate with the modular communication hub as shown in FIGS. **8-11**.

FIG. **13** illustrates a control circuit **500** configured to control aspects of the surgical instrument or tool according to one aspect of this disclosure. The control circuit **500** can be configured to implement various processes described herein. The control circuit **500** may comprise a microcontroller comprising one or more processors **502** (e.g., microprocessor, microcontroller) coupled to at least one memory circuit **504**. The memory circuit **504** stores machine-executable instructions that, when executed by the processor **502**, cause the processor **502** to execute machine instructions to implement various processes described herein. The processor **502** may be any one of a number of single-core or multicore processors known in the art. The memory circuit **504** may comprise volatile and non-volatile storage media. The processor **502** may include an instruction processing unit **506** and an arithmetic unit **508**. The instruction processing unit may be configured to receive instructions from the memory circuit **504** of this disclosure.

FIG. **14** illustrates a combinational logic circuit **510** configured to control aspects of the surgical instrument or tool according to one aspect of this disclosure. The combinational logic circuit **510** can be configured to implement various processes described herein. The combinational logic circuit **510** may comprise a finite state machine comprising a combinational logic **512** configured to receive data associated with the surgical instrument or tool at an input **514**, process the data by the combinational logic **512**, and provide an output **516**.

FIG. **15** illustrates a sequential logic circuit **520** configured to control aspects of the surgical instrument or tool according to one aspect of this disclosure. The sequential logic circuit **520** or the combinational logic **522** can be configured to implement various processes described herein. The sequential logic circuit **520** may comprise a finite state machine. The sequential logic circuit **520** may comprise a combinational logic **522**, at least one memory circuit **524**, and a clock **529**, for example. The at least one memory circuit **524** can store a current state of the finite state machine. In certain instances, the sequential logic circuit **520** may be synchronous or asynchronous. The combinational logic **522** is configured to receive data associated with the surgical instrument or tool from an input **526**, process the data by the combinational logic **522**, and provide an output **528**. In other aspects, the circuit may comprise a combination of a processor (e.g., processor **502**, FIG. **13**) and a finite state machine to implement various processes herein. In other aspects, the finite state machine may comprise a combination of a combinational logic circuit (e.g., combinational logic circuit **510**, FIG. **14**) and the sequential logic circuit **520**.

FIG. **16** illustrates a surgical instrument or tool comprising a plurality of motors which can be activated to perform various functions. In certain instances, a first motor can be activated to perform a first function, a second motor can be

activated to perform a second function, a third motor can be activated to perform a third function, a fourth motor can be activated to perform a fourth function, and so on. In certain instances, the plurality of motors of robotic surgical instrument **600** can be individually activated to cause firing, closure, and/or articulation motions in the end effector. The firing, closure, and/or articulation motions can be transmitted to the end effector through a shaft assembly, for example.

In certain instances, the surgical instrument system or tool may include a firing motor **602**. The firing motor **602** may be operably coupled to a firing motor drive assembly **604** which can be configured to transmit firing motions, generated by the motor **602** to the end effector, in particular to displace the clamp arm closure member. The closure member may be retracted by reversing the direction of the motor **602**, which also causes the clamp arm to open.

In certain instances, the surgical instrument or tool may include a closure motor **603**. The closure motor **603** may be operably coupled to a closure motor drive assembly **605** which can be configured to transmit closure motions, generated by the motor **603** to the end effector, in particular to displace a closure tube to close the anvil and compress tissue between the anvil and the staple cartridge. The closure motor **603** may be operably coupled to a closure motor drive assembly **605** which can be configured to transmit closure motions, generated by the motor **603** to the end effector, in particular to displace a closure tube to close the clamp arm and compress tissue between the clamp arm and either an ultrasonic blade or jaw member of an electrosurgical device. The closure motions may cause the end effector to transition from an open configuration to an approximated configuration to capture tissue, for example. The end effector may be transitioned to an open position by reversing the direction of the motor **603**.

In certain instances, the surgical instrument or tool may include one or more articulation motors **606a**, **606b**, for example. The motors **606a**, **606b** may be operably coupled to respective articulation motor drive assemblies **608a**, **608b**, which can be configured to transmit articulation motions generated by the motors **606a**, **606b** to the end effector. In certain instances, the articulation motions may cause the end effector to articulate relative to the shaft, for example.

As described above, the surgical instrument or tool may include a plurality of motors which may be configured to perform various independent functions. In certain instances, the plurality of motors of the surgical instrument or tool can be individually or separately activated to perform one or more functions while the other motors remain inactive. For example, the articulation motors **606a**, **606b** can be activated to cause the end effector to be articulated while the firing motor **602** remains inactive. Alternatively, the firing motor **602** can be activated to fire the plurality of staples, and/or to advance the cutting edge, while the articulation motor **606** remains inactive. Furthermore, the closure motor **603** may be activated simultaneously with the firing motor **602** to cause the closure tube or closure member to advance distally as described in more detail hereinbelow.

In certain instances, the surgical instrument or tool may include a common control module **610** which can be employed with a plurality of motors of the surgical instrument or tool. In certain instances, the common control module **610** may accommodate one of the plurality of motors at a time. For example, the common control module **610** can be couplable to and separable from the plurality of motors of the robotic surgical instrument individually. In certain instances, a plurality of the motors of the surgical

instrument or tool may share one or more common control modules such as the common control module 610. In certain instances, a plurality of motors of the surgical instrument or tool can be individually and selectively engaged with the common control module 610. In certain instances, the common control module 610 can be selectively switched from interfacing with one of a plurality of motors of the surgical instrument or tool to interfacing with another one of the plurality of motors of the surgical instrument or tool.

In at least one example, the common control module 610 can be selectively switched between operable engagement with the articulation motors 606a, 606b and operable engagement with either the firing motor 602 or the closure motor 603. In at least one example, as illustrated in FIG. 16, a switch 614 can be moved or transitioned between a plurality of positions and/or states. In a first position 616, the switch 614 may electrically couple the common control module 610 to the firing motor 602; in a second position 617, the switch 614 may electrically couple the common control module 610 to the closure motor 603; in a third position 618a, the switch 614 may electrically couple the common control module 610 to the first articulation motor 606a; and in a fourth position 618b, the switch 614 may electrically couple the common control module 610 to the second articulation motor 606b, for example. In certain instances, separate common control modules 610 can be electrically coupled to the firing motor 602, the closure motor 603, and the articulations motor 606a, 606b at the same time. In certain instances, the switch 614 may be a mechanical switch, an electromechanical switch, a solid-state switch, or any suitable switching mechanism.

Each of the motors 602, 603, 606a, 606b may comprise a torque sensor to measure the output torque on the shaft of the motor. The force on an end effector may be sensed in any conventional manner, such as by force sensors on the outer sides of the jaws or by a torque sensor for the motor actuating the jaws.

In various instances, as illustrated in FIG. 16, the common control module 610 may comprise a motor driver 626 which may comprise one or more H-Bridge FETs. The motor driver 626 may modulate the power transmitted from a power source 628 to a motor coupled to the common control module 610 based on input from a microcontroller 620 (the “controller”), for example. In certain instances, the microcontroller 620 can be employed to determine the current drawn by the motor, for example, while the motor is coupled to the common control module 610, as described above.

In certain instances, the microcontroller 620 may include a microprocessor 622 (the “processor”) and one or more non-transitory computer-readable mediums or memory units 624 (the “memory”). In certain instances, the memory 624 may store various program instructions, which when executed may cause the processor 622 to perform a plurality of functions and/or calculations described herein. In certain instances, one or more of the memory units 624 may be coupled to the processor 622, for example. In various aspects, the microcontroller 620 may communicate over a wired or wireless channel, or combinations thereof.

In certain instances, the power source 628 can be employed to supply power to the microcontroller 620, for example. In certain instances, the power source 628 may comprise a battery (or “battery pack” or “power pack”), such as a lithium-ion battery, for example. In certain instances, the battery pack may be configured to be releasably mounted to a handle for supplying power to the surgical instrument 600. A number of battery cells connected in series may be

used as the power source 628. In certain instances, the power source 628 may be replaceable and/or rechargeable, for example.

In various instances, the processor 622 may control the motor driver 626 to control the position, direction of rotation, and/or velocity of a motor that is coupled to the common control module 610. In certain instances, the processor 622 can signal the motor driver 626 to stop and/or disable a motor that is coupled to the common control module 610. It should be understood that the term “processor” as used herein includes any suitable microprocessor, microcontroller, or other basic computing device that incorporates the functions of a computer’s central processing unit (CPU) on an integrated circuit or, at most, a few integrated circuits. The processor 622 is a multipurpose, programmable device that accepts digital data as input, processes it according to instructions stored in its memory, and provides results as output. It is an example of sequential digital logic, as it has internal memory. Processors operate on numbers and symbols represented in the binary numeral system.

In one instance, the processor 622 may be any single-core or multicore processor such as those known under the trade name ARM Cortex by Texas Instruments. In certain instances, the microcontroller 620 may be an LM4F230H5QR, available from Texas Instruments, for example. In at least one example, the Texas Instruments LM4F230H5QR is an ARM Cortex-M4F Processor Core comprising an on-chip memory of 256 KB single-cycle flash memory, or other non-volatile memory, up to 40 MHz, a prefetch buffer to improve performance above 40 MHz, a 32 KB single-cycle SRAM, an internal ROM loaded with StellarisWare® software, a 2 KB EEPROM, one or more PWM modules, one or more QEI analogs, one or more 12-bit ADCs with 12 analog input channels, among other features that are readily available for the product datasheet. Other microcontrollers may be readily substituted for use with the module 4410. Accordingly, the present disclosure should not be limited in this context.

In certain instances, the memory 624 may include program instructions for controlling each of the motors of the surgical instrument 600 that are couplable to the common control module 610. For example, the memory 624 may include program instructions for controlling the firing motor 602, the closure motor 603, and the articulation motors 606a, 606b. Such program instructions may cause the processor 622 to control the firing, closure, and articulation functions in accordance with inputs from algorithms or control programs of the surgical instrument or tool.

In certain instances, one or more mechanisms and/or sensors such as, for example, sensors 630 can be employed to alert the processor 622 to the program instructions that should be used in a particular setting. For example, the sensors 630 may alert the processor 622 to use the program instructions associated with firing, closing, and articulating the end effector. In certain instances, the sensors 630 may comprise position sensors which can be employed to sense the position of the switch 614, for example. Accordingly, the processor 622 may use the program instructions associated with firing the closure member coupled to the clamp arm of the end effector upon detecting, through the sensors 630 for example, that the switch 614 is in the first position 616; the processor 622 may use the program instructions associated with closing the anvil upon detecting, through the sensors 630 for example, that the switch 614 is in the second position 617; and the processor 622 may use the program instructions associated with articulating the end effector upon detecting,

through the sensors 630 for example, that the switch 614 is in the third or fourth position 618a, 618b.

FIG. 17 is a schematic diagram of a robotic surgical instrument 700 configured to operate a surgical tool described herein according to one aspect of this disclosure. The robotic surgical instrument 700 may be programmed or configured to control distal/proximal translation of a displacement member, distal/proximal displacement of a closure tube, shaft rotation, and articulation, either with single or multiple articulation drive links. In one aspect, the surgical instrument 700 may be programmed or configured to individually control a firing member, a closure member, a shaft member, or one or more articulation members, or combinations thereof. The surgical instrument 700 comprises a control circuit 710 configured to control motor-driven firing members, closure members, shaft members, or one or more articulation members, or combinations thereof.

In one aspect, the robotic surgical instrument 700 comprises a control circuit 710 configured to control a clamp arm 716 and a closure member 714 portion of an end effector 702, an ultrasonic blade 718 coupled to an ultrasonic transducer 719 excited by an ultrasonic generator 721, a shaft 740, and one or more articulation members 742a, 742b via a plurality of motors 704a-704e. A position sensor 734 may be configured to provide position feedback of the closure member 714 to the control circuit 710. Other sensors 738 may be configured to provide feedback to the control circuit 710. A timer/counter 731 provides timing and counting information to the control circuit 710. An energy source 712 may be provided to operate the motors 704a-704e, and a current sensor 736 provides motor current feedback to the control circuit 710. The motors 704a-704e can be operated individually by the control circuit 710 in an open-loop or closed-loop feedback control.

In one aspect, the control circuit 710 may comprise one or more microcontrollers, microprocessors, or other suitable processors for executing instructions that cause the processor or processors to perform one or more tasks. In one aspect, a timer/counter 731 provides an output signal, such as the elapsed time or a digital count, to the control circuit 710 to correlate the position of the closure member 714 as determined by the position sensor 734 with the output of the timer/counter 731 such that the control circuit 710 can determine the position of the closure member 714 at a specific time (t) relative to a starting position or the time (t) when the closure member 714 is at a specific position relative to a starting position. The timer/counter 731 may be configured to measure elapsed time, count external events, or time external events.

In one aspect, the control circuit 710 may be programmed to control functions of the end effector 702 based on one or more tissue conditions. The control circuit 710 may be programmed to sense tissue conditions, such as thickness, either directly or indirectly, as described herein. The control circuit 710 may be programmed to select a firing control program or closure control program based on tissue conditions. A firing control program may describe the distal motion of the displacement member. Different firing control programs may be selected to better treat different tissue conditions. For example, when thicker tissue is present, the control circuit 710 may be programmed to translate the displacement member at a lower velocity and/or with lower power. When thinner tissue is present, the control circuit 710 may be programmed to translate the displacement member at a higher velocity and/or with higher power. A closure control program may control the closure force applied to the

tissue by the clamp arm 716. Other control programs control the rotation of the shaft 740 and the articulation members 742a, 742b.

In one aspect, the control circuit 710 may generate motor set point signals. The motor set point signals may be provided to various motor controllers 708a-708e. The motor controllers 708a-708e may comprise one or more circuits configured to provide motor drive signals to the motors 704a-704e to drive the motors 704a-704e as described herein. In some examples, the motors 704a-704e may be brushed DC electric motors. For example, the velocity of the motors 704a-704e may be proportional to the respective motor drive signals. In some examples, the motors 704a-704e may be brushless DC electric motors, and the respective motor drive signals may comprise a PWM signal provided to one or more stator windings of the motors 704a-704e. Also, in some examples, the motor controllers 708a-708e may be omitted and the control circuit 710 may generate the motor drive signals directly.

In one aspect, the control circuit 710 may initially operate each of the motors 704a-704e in an open-loop configuration for a first open-loop portion of a stroke of the displacement member. Based on the response of the robotic surgical instrument 700 during the open-loop portion of the stroke, the control circuit 710 may select a firing control program in a closed-loop configuration. The response of the instrument may include a translation distance of the displacement member during the open-loop portion, a time elapsed during the open-loop portion, the energy provided to one of the motors 704a-704e during the open-loop portion, a sum of pulse widths of a motor drive signal, etc. After the open-loop portion, the control circuit 710 may implement the selected firing control program for a second portion of the displacement member stroke. For example, during a closed-loop portion of the stroke, the control circuit 710 may modulate one of the motors 704a-704e based on translation data describing a position of the displacement member in a closed-loop manner to translate the displacement member at a constant velocity.

In one aspect, the motors 704a-704e may receive power from an energy source 712. The energy source 712 may be a DC power supply driven by a main alternating current power source, a battery, a super capacitor, or any other suitable energy source. The motors 704a-704e may be mechanically coupled to individual movable mechanical elements such as the closure member 714, clamp arm 716, shaft 740, articulation 742a, and articulation 742b via respective transmissions 706a-706e. The transmissions 706a-706e may include one or more gears or other linkage components to couple the motors 704a-704e to movable mechanical elements. A position sensor 734 may sense a position of the closure member 714. The position sensor 734 may be or include any type of sensor that is capable of generating position data that indicate a position of the closure member 714. In some examples, the position sensor 734 may include an encoder configured to provide a series of pulses to the control circuit 710 as the closure member 714 translates distally and proximally. The control circuit 710 may track the pulses to determine the position of the closure member 714. Other suitable position sensors may be used, including, for example, a proximity sensor. Other types of position sensors may provide other signals indicating motion of the closure member 714. Also, in some examples, the position sensor 734 may be omitted. Where any of the motors 704a-704e is a stepper motor, the control circuit 710 may track the position of the closure member 714 by aggregating the number and direction of steps that the

motor 704 has been instructed to execute. The position sensor 734 may be located in the end effector 702 or at any other portion of the instrument. The outputs of each of the motors 704a-704e include a torque sensor 744a-744e to sense force and have an encoder to sense rotation of the drive shaft.

In one aspect, the control circuit 710 is configured to drive a firing member such as the closure member 714 portion of the end effector 702. The control circuit 710 provides a motor set point to a motor control 708a, which provides a drive signal to the motor 704a. The output shaft of the motor 704a is coupled to a torque sensor 744a. The torque sensor 744a is coupled to a transmission 706a which is coupled to the closure member 714. The transmission 706a comprises movable mechanical elements such as rotating elements and a firing member to control the movement of the closure member 714 distally and proximally along a longitudinal axis of the end effector 702. In one aspect, the motor 704a may be coupled to the knife gear assembly, which includes a knife gear reduction set that includes a first knife drive gear and a second knife drive gear. A torque sensor 744a provides a firing force feedback signal to the control circuit 710. The firing force signal represents the force required to fire or displace the closure member 714. A position sensor 734 may be configured to provide the position of the closure member 714 along the firing stroke or the position of the firing member as a feedback signal to the control circuit 710. The end effector 702 may include additional sensors 738 configured to provide feedback signals to the control circuit 710. When ready to use, the control circuit 710 may provide a firing signal to the motor control 708a. In response to the firing signal, the motor 704a may drive the firing member distally along the longitudinal axis of the end effector 702 from a proximal stroke start position to a stroke end position distal to the stroke start position. As the closure member 714 translates distally, the clamp arm 716 closes towards the ultrasonic blade 718.

In one aspect, the control circuit 710 is configured to drive a closure member such as the clamp arm 716 portion of the end effector 702. The control circuit 710 provides a motor set point to a motor control 708b, which provides a drive signal to the motor 704b. The output shaft of the motor 704b is coupled to a torque sensor 744b. The torque sensor 744b is coupled to a transmission 706b which is coupled to the clamp arm 716. The transmission 706b comprises movable mechanical elements such as rotating elements and a closure member to control the movement of the clamp arm 716 from the open and closed positions. In one aspect, the motor 704b is coupled to a closure gear assembly, which includes a closure reduction gear set that is supported in meshing engagement with the closure spur gear. The torque sensor 744b provides a closure force feedback signal to the control circuit 710. The closure force feedback signal represents the closure force applied to the clamp arm 716. The position sensor 734 may be configured to provide the position of the closure member as a feedback signal to the control circuit 710. Additional sensors 738 in the end effector 702 may provide the closure force feedback signal to the control circuit 710. The pivotable clamp arm 716 is positioned opposite the ultrasonic blade 718. When ready to use, the control circuit 710 may provide a closure signal to the motor control 708b. In response to the closure signal, the motor 704b advances a closure member to grasp tissue between the clamp arm 716 and the ultrasonic blade 718.

In one aspect, the control circuit 710 is configured to rotate a shaft member such as the shaft 740 to rotate the end effector 702. The control circuit 710 provides a motor set

point to a motor control 708c, which provides a drive signal to the motor 704c. The output shaft of the motor 704c is coupled to a torque sensor 744c. The torque sensor 744c is coupled to a transmission 706c which is coupled to the shaft 740. The transmission 706c comprises movable mechanical elements such as rotating elements to control the rotation of the shaft 740 clockwise or counterclockwise up to and over 360°. In one aspect, the motor 704c is coupled to the rotational transmission assembly, which includes a tube gear segment that is formed on (or attached to) the proximal end of the proximal closure tube for operable engagement by a rotational gear assembly that is operably supported on the tool mounting plate. The torque sensor 744c provides a rotation force feedback signal to the control circuit 710. The rotation force feedback signal represents the rotation force applied to the shaft 740. The position sensor 734 may be configured to provide the position of the closure member as a feedback signal to the control circuit 710. Additional sensors 738 such as a shaft encoder may provide the rotational position of the shaft 740 to the control circuit 710.

In one aspect, the control circuit 710 is configured to articulate the end effector 702. The control circuit 710 provides a motor set point to a motor control 708d, which provides a drive signal to the motor 704d. The output shaft of the motor 704d is coupled to a torque sensor 744d. The torque sensor 744d is coupled to a transmission 706d which is coupled to an articulation member 742a. The transmission 706d comprises movable mechanical elements such as articulation elements to control the articulation of the end effector 702 ±65°. In one aspect, the motor 704d is coupled to an articulation nut, which is rotatably journaled on the proximal end portion of the distal spine portion and is rotatably driven thereon by an articulation gear assembly. The torque sensor 744d provides an articulation force feedback signal to the control circuit 710. The articulation force feedback signal represents the articulation force applied to the end effector 702. Sensors 738, such as an articulation encoder, may provide the articulation position of the end effector 702 to the control circuit 710.

In another aspect, the articulation function of the robotic surgical system 700 may comprise two articulation members, or links, 742a, 742b. These articulation members 742a, 742b are driven by separate disks on the robot interface (the rack) which are driven by the two motors 708d, 708e. When the separate firing motor 704a is provided, each of articulation links 742a, 742b can be antagonistically driven with respect to the other link in order to provide a resistive holding motion and a load to the head when it is not moving and to provide an articulation motion as the head is articulated. The articulation members 742a, 742b attach to the head at a fixed radius as the head is rotated. Accordingly, the mechanical advantage of the push-and-pull link changes as the head is rotated. This change in the mechanical advantage may be more pronounced with other articulation link drive systems.

In one aspect, the one or more motors 704a-704e may comprise a brushed DC motor with a gearbox and mechanical links to a firing member, closure member, or articulation member. Another example includes electric motors 704a-704e that operate the movable mechanical elements such as the displacement member, articulation links, closure tube, and shaft. An outside influence is an unmeasured, unpredictable influence of things like tissue, surrounding bodies, and friction on the physical system. Such outside influence can be referred to as drag, which acts in opposition to one of electric motors 704a-704e. The outside influence, such as

drag, may cause the operation of the physical system to deviate from a desired operation of the physical system.

In one aspect, the position sensor **734** may be implemented as an absolute positioning system. In one aspect, the position sensor **734** may comprise a magnetic rotary absolute positioning system implemented as an AS5055EQFT single-chip magnetic rotary position sensor available from Austria Microsystems, AG. The position sensor **734** may interface with the control circuit **710** to provide an absolute positioning system. The position may include multiple Hall-effect elements located above a magnet and coupled to a CORDIC processor, also known as the digit-by-digit method and Volder's algorithm, that is provided to implement a simple and efficient algorithm to calculate hyperbolic and trigonometric functions that require only addition, subtraction, bitshift, and table lookup operations.

In one aspect, the control circuit **710** may be in communication with one or more sensors **738**. The sensors **738** may be positioned on the end effector **702** and adapted to operate with the robotic surgical instrument **700** to measure the various derived parameters such as the gap distance versus time, tissue compression versus time, and anvil strain versus time. The sensors **738** may comprise a magnetic sensor, a magnetic field sensor, a strain gauge, a load cell, a pressure sensor, a force sensor, a torque sensor, an inductive sensor such as an eddy current sensor, a resistive sensor, a capacitive sensor, an optical sensor, and/or any other suitable sensor for measuring one or more parameters of the end effector **702**. The sensors **738** may include one or more sensors. The sensors **738** may be located on the clamp arm **716** to determine tissue location using segmented electrodes. The torque sensors **744a-744e** may be configured to sense force such as firing force, closure force, and/or articulation force, among others. Accordingly, the control circuit **710** can sense (1) the closure load experienced by the distal closure tube and its position, (2) the firing member at the rack and its position, (3) what portion of the ultrasonic blade **718** has tissue on it, and (4) the load and position on both articulation rods.

In one aspect, the one or more sensors **738** may comprise a strain gauge, such as a micro-strain gauge, configured to measure the magnitude of the strain in the clamp arm **716** during a clamped condition. The strain gauge provides an electrical signal whose amplitude varies with the magnitude of the strain. The sensors **738** may comprise a pressure sensor configured to detect a pressure generated by the presence of compressed tissue between the clamp arm **716** and the ultrasonic blade **718**. The sensors **738** may be configured to detect impedance of a tissue section located between the clamp arm **716** and the ultrasonic blade **718** that is indicative of the thickness and/or fullness of tissue located therebetween.

In one aspect, the sensors **738** may be implemented as one or more limit switches, electromechanical devices, solid-state switches, Hall-effect devices, magneto-resistive (MR) devices, giant magneto-resistive (GMR) devices, magnetometers, among others. In other implementations, the sensors **738** may be implemented as solid-state switches that operate under the influence of light, such as optical sensors, IR sensors, ultraviolet sensors, among others. Still, the switches may be solid-state devices such as transistors (e.g., FET, junction FET, MOSFET, bipolar, and the like). In other implementations, the sensors **738** may include electrical conductorless switches, ultrasonic switches, accelerometers, and inertial sensors, among others.

In one aspect, the sensors **738** may be configured to measure forces exerted on the clamp arm **716** by the closure

drive system. For example, one or more sensors **738** can be at an interaction point between the closure tube and the clamp arm **716** to detect the closure forces applied by the closure tube to the clamp arm **716**. The forces exerted on the clamp arm **716** can be representative of the tissue compression experienced by the tissue section captured between the clamp arm **716** and the ultrasonic blade **718**. The one or more sensors **738** can be positioned at various interaction points along the closure drive system to detect the closure forces applied to the clamp arm **716** by the closure drive system. The one or more sensors **738** may be sampled in real time during a clamping operation by the processor of the control circuit **710**. The control circuit **710** receives real-time sample measurements to provide and analyze time-based information and assess, in real time, closure forces applied to the clamp arm **716**.

In one aspect, a current sensor **736** can be employed to measure the current drawn by each of the motors **704a-704e**. The force required to advance any of the movable mechanical elements such as the closure member **714** corresponds to the current drawn by one of the motors **704a-704e**. The force is converted to a digital signal and provided to the control circuit **710**. The control circuit **710** can be configured to simulate the response of the actual system of the instrument in the software of the controller. A displacement member can be actuated to move the closure member **714** in the end effector **702** at or near a target velocity. The robotic surgical instrument **700** can include a feedback controller, which can be one of any feedback controllers, including, but not limited to a PID, a state feedback, a linear-quadratic (LQR), and/or an adaptive controller, for example. The robotic surgical instrument **700** can include a power source to convert the signal from the feedback controller into a physical input such as case voltage, PWM voltage, frequency modulated voltage, current, torque, and/or force, for example. Additional details are disclosed in U.S. patent application Ser. No. 15/636,829, titled CLOSED LOOP VELOCITY CONTROL TECHNIQUES FOR ROBOTIC SURGICAL INSTRUMENT, filed Jun. 29, 2017, now U.S. Patent Application Publication No. 2019/0000446, which is herein incorporated by reference in its entirety.

FIG. **18** illustrates a schematic diagram of a surgical instrument **750** configured to control the distal translation of a displacement member according to one aspect of this disclosure. In one aspect, the surgical instrument **750** is programmed to control the distal translation of a displacement member such as the closure member **764**. The surgical instrument **750** comprises an end effector **752** that may comprise a clamp arm **766**, a closure member **764**, and an ultrasonic blade **768** coupled to an ultrasonic transducer **769** driven by an ultrasonic generator **771**.

The position, movement, displacement, and/or translation of a linear displacement member, such as the closure member **764**, can be measured by an absolute positioning system, sensor arrangement, and position sensor **784**. Because the closure member **764** is coupled to a longitudinally movable drive member, the position of the closure member **764** can be determined by measuring the position of the longitudinally movable drive member employing the position sensor **784**. Accordingly, in the following description, the position, displacement, and/or translation of the closure member **764** can be achieved by the position sensor **784** as described herein. A control circuit **760** may be programmed to control the translation of the displacement member, such as the closure member **764**. The control circuit **760**, in some examples, may comprise one or more microcontrollers, microprocessors, or other suitable processors for executing

instructions that cause the processor or processors to control the displacement member, e.g., the closure member **764**, in the manner described. In one aspect, a timer/counter **781** provides an output signal, such as the elapsed time or a digital count, to the control circuit **760** to correlate the position of the closure member **764** as determined by the position sensor **784** with the output of the timer/counter **781** such that the control circuit **760** can determine the position of the closure member **764** at a specific time (t) relative to a starting position. The timer/counter **781** may be configured to measure elapsed time, count external events, or time external events.

The control circuit **760** may generate a motor set point signal **772**. The motor set point signal **772** may be provided to a motor controller **758**. The motor controller **758** may comprise one or more circuits configured to provide a motor drive signal **774** to the motor **754** to drive the motor **754** as described herein. In some examples, the motor **754** may be a brushed DC electric motor. For example, the velocity of the motor **754** may be proportional to the motor drive signal **774**. In some examples, the motor **754** may be a brushless DC electric motor and the motor drive signal **774** may comprise a PWM signal provided to one or more stator windings of the motor **754**. Also, in some examples, the motor controller **758** may be omitted, and the control circuit **760** may generate the motor drive signal **774** directly.

The motor **754** may receive power from an energy source **762**. The energy source **762** may be or include a battery, a super capacitor, or any other suitable energy source. The motor **754** may be mechanically coupled to the closure member **764** via a transmission **756**. The transmission **756** may include one or more gears or other linkage components to couple the motor **754** to the closure member **764**. A position sensor **784** may sense a position of the closure member **764**. The position sensor **784** may be or include any type of sensor that is capable of generating position data that indicate a position of the closure member **764**. In some examples, the position sensor **784** may include an encoder configured to provide a series of pulses to the control circuit **760** as the closure member **764** translates distally and proximally. The control circuit **760** may track the pulses to determine the position of the closure member **764**. Other suitable position sensors may be used, including, for example, a proximity sensor. Other types of position sensors may provide other signals indicating motion of the closure member **764**. Also, in some examples, the position sensor **784** may be omitted. Where the motor **754** is a stepper motor, the control circuit **760** may track the position of the closure member **764** by aggregating the number and direction of steps that the motor **754** has been instructed to execute. The position sensor **784** may be located in the end effector **752** or at any other portion of the instrument.

The control circuit **760** may be in communication with one or more sensors **788**. The sensors **788** may be positioned on the end effector **752** and adapted to operate with the surgical instrument **750** to measure the various derived parameters such as gap distance versus time, tissue compression versus time, and anvil strain versus time. The sensors **788** may comprise a magnetic sensor, a magnetic field sensor, a strain gauge, a pressure sensor, a force sensor, an inductive sensor such as an eddy current sensor, a resistive sensor, a capacitive sensor, an optical sensor, and/or any other suitable sensor for measuring one or more parameters of the end effector **752**. The sensors **788** may include one or more sensors.

The one or more sensors **788** may comprise a strain gauge, such as a micro-strain gauge, configured to measure

the magnitude of the strain in the clamp arm **766** during a clamped condition. The strain gauge provides an electrical signal whose amplitude varies with the magnitude of the strain. The sensors **788** may comprise a pressure sensor configured to detect a pressure generated by the presence of compressed tissue between the clamp arm **766** and the ultrasonic blade **768**. The sensors **788** may be configured to detect impedance of a tissue section located between the clamp arm **766** and the ultrasonic blade **768** that is indicative of the thickness and/or fullness of tissue located therebetween.

The sensors **788** may be is configured to measure forces exerted on the clamp arm **766** by a closure drive system. For example, one or more sensors **788** can be at an interaction point between a closure tube and the clamp arm **766** to detect the closure forces applied by a closure tube to the clamp arm **766**. The forces exerted on the clamp arm **766** can be representative of the tissue compression experienced by the tissue section captured between the clamp arm **766** and the ultrasonic blade **768**. The one or more sensors **788** can be positioned at various interaction points along the closure drive system to detect the closure forces applied to the clamp arm **766** by the closure drive system. The one or more sensors **788** may be sampled in real time during a clamping operation by a processor of the control circuit **760**. The control circuit **760** receives real-time sample measurements to provide and analyze time-based information and assess, in real time, closure forces applied to the clamp arm **766**.

A current sensor **786** can be employed to measure the current drawn by the motor **754**. The force required to advance the closure member **764** corresponds to the current drawn by the motor **754**. The force is converted to a digital signal and provided to the control circuit **760**.

The control circuit **760** can be configured to simulate the response of the actual system of the instrument in the software of the controller. A displacement member can be actuated to move a closure member **764** in the end effector **752** at or near a target velocity. The surgical instrument **750** can include a feedback controller, which can be one of any feedback controllers, including, but not limited to a PID, a state feedback, LQR, and/or an adaptive controller, for example. The surgical instrument **750** can include a power source to convert the signal from the feedback controller into a physical input such as case voltage, PWM voltage, frequency modulated voltage, current, torque, and/or force, for example.

The actual drive system of the surgical instrument **750** is configured to drive the displacement member, cutting member, or closure member **764**, by a brushed DC motor with gearbox and mechanical links to an articulation and/or knife system. Another example is the electric motor **754** that operates the displacement member and the articulation driver, for example, of an interchangeable shaft assembly. An outside influence is an unmeasured, unpredictable influence of things like tissue, surrounding bodies and friction on the physical system. Such outside influence can be referred to as drag which acts in opposition to the electric motor **754**. The outside influence, such as drag, may cause the operation of the physical system to deviate from a desired operation of the physical system.

Various example aspects are directed to a surgical instrument **750** comprising an end effector **752** with motor-driven surgical sealing and cutting implements. For example, a motor **754** may drive a displacement member distally and proximally along a longitudinal axis of the end effector **752**. The end effector **752** may comprise a pivotable clamp arm **766** and, when configured for use, an ultrasonic blade **768**

positioned opposite the clamp arm **766**. A clinician may grasp tissue between the clamp arm **766** and the ultrasonic blade **768**, as described herein. When ready to use the instrument **750**, the clinician may provide a firing signal, for example by depressing a trigger of the instrument **750**. In response to the firing signal, the motor **754** may drive the displacement member distally along the longitudinal axis of the end effector **752** from a proximal stroke begin position to a stroke end position distal of the stroke begin position. As the displacement member translates distally, the closure member **764** with a cutting element positioned at a distal end, may cut the tissue between the ultrasonic blade **768** and the clamp arm **766**.

In various examples, the surgical instrument **750** may comprise a control circuit **760** programmed to control the distal translation of the displacement member, such as the closure member **764**, for example, based on one or more tissue conditions. The control circuit **760** may be programmed to sense tissue conditions, such as thickness, either directly or indirectly, as described herein. The control circuit **760** may be programmed to select a control program based on tissue conditions. A control program may describe the distal motion of the displacement member. Different control programs may be selected to better treat different tissue conditions. For example, when thicker tissue is present, the control circuit **760** may be programmed to translate the displacement member at a lower velocity and/or with lower power. When thinner tissue is present, the control circuit **760** may be programmed to translate the displacement member at a higher velocity and/or with higher power.

In some examples, the control circuit **760** may initially operate the motor **754** in an open loop configuration for a first open loop portion of a stroke of the displacement member. Based on a response of the instrument **750** during the open loop portion of the stroke, the control circuit **760** may select a firing control program. The response of the instrument may include, a translation distance of the displacement member during the open loop portion, a time elapsed during the open loop portion, energy provided to the motor **754** during the open loop portion, a sum of pulse widths of a motor drive signal, etc. After the open loop portion, the control circuit **760** may implement the selected firing control program for a second portion of the displacement member stroke. For example, during the closed loop portion of the stroke, the control circuit **760** may modulate the motor **754** based on translation data describing a position of the displacement member in a closed loop manner to translate the displacement member at a constant velocity. Additional details are disclosed in U.S. patent application Ser. No. 15/720,852, titled SYSTEM AND METHODS FOR CONTROLLING A DISPLAY OF A SURGICAL INSTRUMENT, filed Sep. 29, 2017, which issued on Aug. 18, 2020 as U.S. Pat. No. 10,743,872, which is herein incorporated by reference in its entirety.

FIG. **19** is a schematic diagram of a surgical instrument **790** configured to control various functions according to one aspect of this disclosure. In one aspect, the surgical instrument **790** is programmed to control distal translation of a displacement member such as the closure member **764**. The surgical instrument **790** comprises an end effector **792** that may comprise a clamp arm **766**, a closure member **764**, and an ultrasonic blade **768** which may be interchanged with or work in conjunction with one or more RF electrodes **796** (shown in dashed line). The ultrasonic blade **768** is coupled to an ultrasonic transducer **769** driven by an ultrasonic generator **771**.

In one aspect, sensors **788** may be implemented as a limit switch, electromechanical device, solid-state switches, Hall-effect devices, MR devices, GMR devices, magnetometers, among others. In other implementations, the sensors **638** may be solid-state switches that operate under the influence of light, such as optical sensors, IR sensors, ultraviolet sensors, among others. Still, the switches may be solid-state devices such as transistors (e.g., FET, junction FET, MOS-FET, bipolar, and the like). In other implementations, the sensors **788** may include electrical conductorless switches, ultrasonic switches, accelerometers, and inertial sensors, among others.

In one aspect, the position sensor **784** may be implemented as an absolute positioning system comprising a magnetic rotary absolute positioning system implemented as an AS5055EQFT single-chip magnetic rotary position sensor available from Austria Microsystems, AG. The position sensor **784** may interface with the control circuit **760** to provide an absolute positioning system. The position may include multiple Hall-effect elements located above a magnet and coupled to a CORDIC processor, also known as the digit-by-digit method and Volder's algorithm, that is provided to implement a simple and efficient algorithm to calculate hyperbolic and trigonometric functions that require only addition, subtraction, bitshift, and table lookup operations.

In some examples, the position sensor **784** may be omitted. Where the motor **754** is a stepper motor, the control circuit **760** may track the position of the closure member **764** by aggregating the number and direction of steps that the motor has been instructed to execute. The position sensor **784** may be located in the end effector **792** or at any other portion of the instrument.

The control circuit **760** may be in communication with one or more sensors **788**. The sensors **788** may be positioned on the end effector **792** and adapted to operate with the surgical instrument **790** to measure the various derived parameters such as gap distance versus time, tissue compression versus time, and anvil strain versus time. The sensors **788** may comprise a magnetic sensor, a magnetic field sensor, a strain gauge, a pressure sensor, a force sensor, an inductive sensor such as an eddy current sensor, a resistive sensor, a capacitive sensor, an optical sensor, and/or any other suitable sensor for measuring one or more parameters of the end effector **792**. The sensors **788** may include one or more sensors.

An RF energy source **794** is coupled to the end effector **792** and is applied to the RF electrode **796** when the RF electrode **796** is provided in the end effector **792** in place of the ultrasonic blade **768** or to work in conjunction with the ultrasonic blade **768**. For example, the ultrasonic blade is made of electrically conductive metal and may be employed as the return path for electrosurgical RF current. The control circuit **760** controls the delivery of the RF energy to the RF electrode **796**.

Additional details are disclosed in U.S. patent application Ser. No. 15/636,096, titled SURGICAL SYSTEM COUPLABLE WITH STAPLE CARTRIDGE AND RADIO FREQUENCY CARTRIDGE, AND METHOD OF USING SAME, filed Jun. 28, 2017, now U.S. Patent Application Publication No. 2019/0000478, which is herein incorporated by reference in its entirety.

Generator Hardware
Adaptive Ultrasonic Blade Control Algorithms

In various aspects smart ultrasonic energy devices may comprise adaptive algorithms to control the operation of the ultrasonic blade. In one aspect, the ultrasonic blade adaptive

control algorithms are configured to identify tissue type and adjust device parameters. In one aspect, the ultrasonic blade control algorithms are configured to parameterize tissue type. An algorithm to detect the collagen/elastic ratio of tissue to tune the amplitude of the distal tip of the ultrasonic blade is described in the following section of the present disclosure. Various aspects of smart ultrasonic energy devices are described herein in connection with FIGS. 1-85, for example. Accordingly, the following description of adaptive ultrasonic blade control algorithms should be read in conjunction with FIGS. 1-85 and the description associated therewith.

Tissue Type Identification and Device Parameter Adjustments

In certain surgical procedures it would be desirable to employ adaptive ultrasonic blade control algorithms. In one aspect, adaptive ultrasonic blade control algorithms may be employed to adjust the parameters of the ultrasonic device based on the type of tissue in contact with the ultrasonic blade. In one aspect, the parameters of the ultrasonic device may be adjusted based on the location of the tissue within the jaws of the ultrasonic end effector, for example, the location of the tissue between the clamp arm and the ultrasonic blade. The impedance of the ultrasonic transducer may be employed to differentiate what percentage of the tissue is located in the distal or proximal end of the end effector. The reactions of the ultrasonic device may be based on the tissue type or compressibility of the tissue. In another aspect, the parameters of the ultrasonic device may be adjusted based on the identified tissue type or parameterization. For example, the mechanical displacement amplitude of the distal tip of the ultrasonic blade may be tuned based on the ration of collagen to elastin tissue detected during the tissue identification procedure. The ratio of collagen to elastin tissue may be detected used a variety of techniques including infrared (IR) surface reflectance and emissivity. The force applied to the tissue by the clamp arm and/or the stroke of the clamp arm to produce gap and compression. Electrical continuity across a jaw equipped with electrodes may be employed to determine what percentage of the jaw is covered with tissue.

FIG. 20 is a system 800 configured to execute adaptive ultrasonic blade control algorithms in a surgical data network comprising a modular communication hub, in accordance with at least one aspect of the present disclosure. In one aspect, the generator module 240 is configured to execute the adaptive ultrasonic blade control algorithm(s) 802 as described herein with reference to FIGS. 43A-54. In another aspect, the device/instrument 235 is configured to execute the adaptive ultrasonic blade control algorithm(s) 804 as described herein with reference to FIGS. 43A-54. In another aspect, both the device/instrument 235 and the device/instrument 235 are configured to execute the adaptive ultrasonic blade control algorithms 802, 804 as described herein with reference to FIGS. 43A-54.

The generator module 240 may comprise a patient isolated stage in communication with a non-isolated stage via a power transformer. A secondary winding of the power transformer is contained in the isolated stage and may comprise a tapped configuration (e.g., a center-tapped or a non-center-tapped configuration) to define drive signal outputs for delivering drive signals to different surgical instruments, such as, for example, an ultrasonic surgical instrument, an RF electrosurgical instrument, and a multifunction surgical instrument which includes ultrasonic and RF energy modes that can be delivered alone or simultaneously. In particular, the drive signal outputs may output an ultrasonic

drive signal (e.g., a 420V root-mean-square (RMS) drive signal) to an ultrasonic surgical instrument 241, and the drive signal outputs may output an RF electrosurgical drive signal (e.g., a 100V RMS drive signal) to an RF electrosurgical instrument 241. Aspects of the generator module 240 are described herein with reference to FIGS. 21-28B.

The generator module 240 or the device/instrument 235 or both are coupled to the modular control tower 236 connected to multiple operating theater devices such as, for example, intelligent surgical instruments, robots, and other computerized devices located in the operating theater, as described with reference to FIGS. 8-11, for example.

FIG. 21 illustrates an example of a generator 900, which is one form of a generator configured to couple to an ultrasonic instrument and further configured to execute adaptive ultrasonic blade control algorithms in a surgical data network comprising a modular communication hub as shown in FIG. 20. The generator 900 is configured to deliver multiple energy modalities to a surgical instrument. The generator 900 provides RF and ultrasonic signals for delivering energy to a surgical instrument either independently or simultaneously. The RF and ultrasonic signals may be provided alone or in combination and may be provided simultaneously. As noted above, at least one generator output can deliver multiple energy modalities (e.g., ultrasonic, bipolar or monopolar RF, irreversible and/or reversible electroporation, and/or microwave energy, among others) through a single port, and these signals can be delivered separately or simultaneously to the end effector to treat tissue. The generator 900 comprises a processor 902 coupled to a waveform generator 904. The processor 902 and waveform generator 904 are configured to generate a variety of signal waveforms based on information stored in a memory coupled to the processor 902, not shown for clarity of disclosure. The digital information associated with a waveform is provided to the waveform generator 904 which includes one or more DAC circuits to convert the digital input into an analog output. The analog output is fed to an amplifier 1106 for signal conditioning and amplification. The conditioned and amplified output of the amplifier 906 is coupled to a power transformer 908. The signals are coupled across the power transformer 908 to the secondary side, which is in the patient isolation side. A first signal of a first energy modality is provided to the surgical instrument between the terminals labeled ENERGY₁ and RETURN. A second signal of a second energy modality is coupled across a capacitor 910 and is provided to the surgical instrument between the terminals labeled ENERGY₂ and RETURN. It will be appreciated that more than two energy modalities may be output and thus the subscript "n" may be used to designate that up to n ENERGY_n terminals may be provided, where n is a positive integer greater than 1. It also will be appreciated that up to "n" return paths RETURN_n may be provided without departing from the scope of the present disclosure.

A first voltage sensing circuit 912 is coupled across the terminals labeled ENERGY₁ and the RETURN path to measure the output voltage therebetween. A second voltage sensing circuit 924 is coupled across the terminals labeled ENERGY₂ and the RETURN path to measure the output voltage therebetween. A current sensing circuit 914 is disposed in series with the RETURN leg of the secondary side of the power transformer 908 as shown to measure the output current for either energy modality. If different return paths are provided for each energy modality, then a separate current sensing circuit should be provided in each return leg. The outputs of the first and second voltage sensing circuits

912, 924 are provided to respective isolation transformers 916, 922 and the output of the current sensing circuit 914 is provided to another isolation transformer 918. The outputs of the isolation transformers 916, 928, 922 in the on the primary side of the power transformer 908 (non-patient isolated side) are provided to a one or more ADC circuit 926. The digitized output of the ADC circuit 926 is provided to the processor 902 for further processing and computation. The output voltages and output current feedback information can be employed to adjust the output voltage and current provided to the surgical instrument and to compute output impedance, among other parameters. Input/output communications between the processor 902 and patient isolated circuits is provided through an interface circuit 920. Sensors also may be in electrical communication with the processor 902 by way of the interface circuit 920.

In one aspect, the impedance may be determined by the processor 902 by dividing the output of either the first voltage sensing circuit 912 coupled across the terminals labeled ENERGY₁/RETURN or the second voltage sensing circuit 924 coupled across the terminals labeled ENERGY₂/RETURN by the output of the current sensing circuit 914 disposed in series with the RETURN leg of the secondary side of the power transformer 908. The outputs of the first and second voltage sensing circuits 912, 924 are provided to separate isolations transformers 916, 922 and the output of the current sensing circuit 914 is provided to another isolation transformer 916. The digitized voltage and current sensing measurements from the ADC circuit 926 are provided the processor 902 for computing impedance. As an example, the first energy modality ENERGY₁ may be ultrasonic energy and the second energy modality ENERGY₂ may be RF energy. Nevertheless, in addition to ultrasonic and bipolar or monopolar RF energy modalities, other energy modalities include irreversible and/or reversible electroporation and/or microwave energy, among others. Also, although the example illustrated in FIG. 21 shows a single return path RETURN may be provided for two or more energy modalities, in other aspects, multiple return paths RETURN_n may be provided for each energy modality ENERGY_n. Thus, as described herein, the ultrasonic transducer impedance may be measured by dividing the output of the first voltage sensing circuit 912 by the current sensing circuit 914 and the tissue impedance may be measured by dividing the output of the second voltage sensing circuit 924 by the current sensing circuit 914.

As shown in FIG. 21, the generator 900 comprising at least one output port can include a power transformer 908 with a single output and with multiple taps to provide power in the form of one or more energy modalities, such as ultrasonic, bipolar or monopolar RF, irreversible and/or reversible electroporation, and/or microwave energy, among others, for example, to the end effector depending on the type of treatment of tissue being performed. For example, the generator 900 can deliver energy with higher voltage and lower current to drive an ultrasonic transducer, with lower voltage and higher current to drive RF electrodes for sealing tissue, or with a coagulation waveform for spot coagulation using either monopolar or bipolar RF electrosurgical electrodes. The output waveform from the generator 900 can be steered, switched, or filtered to provide the frequency to the end effector of the surgical instrument. The connection of an ultrasonic transducer to the generator 900 output would be preferably located between the output labeled ENERGY₁ and RETURN as shown in FIG. 21. In one example, a connection of RF bipolar electrodes to the generator 900 output would be preferably located between the output

labeled ENERGY₂ and RETURN. In the case of monopolar output, the preferred connections would be active electrode (e.g., pencil or other probe) to the ENERGY₂ output and a suitable return pad connected to the RETURN output.

Additional details are disclosed in U.S. Patent Application Publication No. 2017/0086914, titled TECHNIQUES FOR OPERATING GENERATOR FOR DIGITALLY GENERATING ELECTRICAL SIGNAL WAVEFORMS AND SURGICAL INSTRUMENTS, which published on Mar. 30, 2017, now U.S. Pat. No. 10,624,691, which is herein incorporated by reference in its entirety.

As used throughout this description, the term “wireless” and its derivatives may be used to describe circuits, devices, systems, methods, techniques, communications channels, etc., that may communicate data through the use of modulated electromagnetic radiation through a non-solid medium. The term does not imply that the associated devices do not contain any wires, although in some aspects they might not. The communication module may implement any of a number of wireless or wired communication standards or protocols, including but not limited to Wi-Fi (IEEE 802.11 family), WiMAX (IEEE 802.16 family), IEEE 802.20, long term evolution (LTE), Ev-DO, HSPA+, HSDPA+, HSUPA+, EDGE, GSM, GPRS, CDMA, TDMA, DECT, Bluetooth, Ethernet derivatives thereof, as well as any other wireless and wired protocols that are designated as 3G, 4G, 5G, and beyond. The computing module may include a plurality of communication modules. For instance, a first communication module may be dedicated to shorter range wireless communications such as Wi-Fi and Bluetooth and a second communication module may be dedicated to longer range wireless communications such as GPS, EDGE, GPRS, CDMA, WiMAX, LTE, Ev-DO, and others.

As used herein a processor or processing unit is an electronic circuit which performs operations on some external data source, usually memory or some other data stream. The term is used herein to refer to the central processor (central processing unit) in a system or computer systems (especially systems on a chip (SoCs)) that combine a number of specialized “processors.”

As used herein, a system on a chip or system on chip (SoC or SOC) is an integrated circuit (also known as an “IC” or “chip”) that integrates all components of a computer or other electronic systems. It may contain digital, analog, mixed-signal, and often radio-frequency functions-all on a single substrate. A SoC integrates a microcontroller (or microprocessor) with advanced peripherals like graphics processing unit (GPU), Wi-Fi module, or coprocessor. A SoC may or may not contain built-in memory.

As used herein, a microcontroller or controller is a system that integrates a microprocessor with peripheral circuits and memory. A microcontroller (or MCU for microcontroller unit) may be implemented as a small computer on a single integrated circuit. It may be similar to a SoC; an SoC may include a microcontroller as one of its components. A microcontroller may contain one or more core processing units (CPUs) along with memory and programmable input/output peripherals. Program memory in the form of Ferroelectric RAM, NOR flash or OTP ROM is also often included on chip, as well as a small amount of RAM. Microcontrollers may be employed for embedded applications, in contrast to the microprocessors used in personal computers or other general purpose applications consisting of various discrete chips.

As used herein, the term controller or microcontroller may be a stand-alone IC or chip device that interfaces with a peripheral device. This may be a link between two parts of

a computer or a controller on an external device that manages the operation of (and connection with) that device.

Any of the processors or microcontrollers described herein, may be implemented by any single core or multicore processor such as those known under the trade name ARM Cortex by Texas Instruments. In one aspect, the processor may be an LM4F230H5QR ARM Cortex-M4F Processor Core, available from Texas Instruments, for example, comprising on-chip memory of 256 KB single-cycle flash memory, or other non-volatile memory, up to 40 MHz, a prefetch buffer to improve performance above 40 MHz, a 32 KB single-cycle serial random access memory (SRAM), internal read-only memory (ROM) loaded with StellarisWare® software, 2 KB electrically erasable programmable read-only memory (EEPROM), one or more pulse width modulation (PWM) modules, one or more quadrature encoder inputs (QEI) analog, one or more 12-bit Analog-to-Digital Converters (ADC) with 12 analog input channels, details of which are available for the product datasheet.

In one aspect, the processor may comprise a safety controller comprising two controller-based families such as TMS570 and RM4x known under the trade name Hercules ARM Cortex R4, also by Texas Instruments. The safety controller may be configured specifically for IEC 61508 and ISO 26262 safety critical applications, among others, to provide advanced integrated safety features while delivering scalable performance, connectivity, and memory options.

Modular devices include the modules (as described in connection with FIGS. 3 and 9, for example) that are receivable within a surgical hub and the surgical devices or instruments that can be connected to the various modules in order to connect or pair with the corresponding surgical hub. The modular devices include, for example, intelligent surgical instruments, medical imaging devices, suction/irrigation devices, smoke evacuators, energy generators, ventilators, insufflators, and displays. The modular devices described herein can be controlled by control algorithms. The control algorithms can be executed on the modular device itself, on the surgical hub to which the particular modular device is paired, or on both the modular device and the surgical hub (e.g., via a distributed computing architecture). In some exemplifications, the modular devices' control algorithms control the devices based on data sensed by the modular device itself (i.e., by sensors in, on, or connected to the modular device). This data can be related to the patient being operated on (e.g., tissue properties or insufflation pressure) or the modular device itself (e.g., the rate at which a knife is being advanced, motor current, or energy levels). For example, a control algorithm for a surgical stapling and cutting instrument can control the rate at which the instrument's motor drives its knife through tissue according to resistance encountered by the knife as it advances.

FIG. 22 illustrates one form of a surgical system 1000 comprising a generator 1100 and various surgical instruments 1104, 1106, 1108 usable therewith, where the surgical instrument 1104 is an ultrasonic surgical instrument, the surgical instrument 1106 is an RF electrosurgical instrument, and the multifunction surgical instrument 1108 is a combination ultrasonic/RF electrosurgical instrument. The generator 1100 is configurable for use with a variety of surgical instruments. According to various forms, the generator 1100 may be configurable for use with different surgical instruments of different types including, for example, ultrasonic surgical instruments 1104, RF electrosurgical instruments 1106, and multifunction surgical instruments 1108 that integrate RF and ultrasonic energies delivered simultaneously

from the generator 1100. Although in the form of FIG. 22 the generator 1100 is shown separate from the surgical instruments 1104, 1106, 1108 in one form, the generator 1100 may be formed integrally with any of the surgical instruments 1104, 1106, 1108 to form a unitary surgical system. The generator 1100 comprises an input device 1110 located on a front panel of the generator 1100 console. The input device 1110 may comprise any suitable device that generates signals suitable for programming the operation of the generator 1100. The generator 1100 may be configured for wired or wireless communication.

The generator 1100 is configured to drive multiple surgical instruments 1104, 1106, 1108. The first surgical instrument is an ultrasonic surgical instrument 1104 and comprises a handpiece 1105 (HP), an ultrasonic transducer 1120, a shaft 1126, and an end effector 1122. The end effector 1122 comprises an ultrasonic blade 1128 acoustically coupled to the ultrasonic transducer 1120 and a clamp arm 1140. The handpiece 1105 comprises a trigger 1143 to operate the clamp arm 1140 and a combination of the toggle buttons 1134a, 1134b, 1134c to energize and drive the ultrasonic blade 1128 or other function. The toggle buttons 1134a, 1134b, 1134c can be configured to energize the ultrasonic transducer 1120 with the generator 1100.

The generator 1100 also is configured to drive a second surgical instrument 1106. The second surgical instrument 1106 is an RF electrosurgical instrument and comprises a handpiece 1107 (HP), a shaft 1127, and an end effector 1124. The end effector 1124 comprises electrodes in clamp arms 1142a, 1142b and return through an electrical conductor portion of the shaft 1127. The electrodes are coupled to and energized by a bipolar energy source within the generator 1100. The handpiece 1107 comprises a trigger 1145 to operate the clamp arms 1142a, 1142b and an energy button 1135 to actuate an energy switch to energize the electrodes in the end effector 1124.

The generator 1100 also is configured to drive a multifunction surgical instrument 1108. The multifunction surgical instrument 1108 comprises a handpiece 1109 (HP), a shaft 1129, and an end effector 1125. The end effector 1125 comprises an ultrasonic blade 1149 and a clamp arm 1146. The ultrasonic blade 1149 is acoustically coupled to the ultrasonic transducer 1120. The handpiece 1109 comprises a trigger 1147 to operate the clamp arm 1146 and a combination of the toggle buttons 1137a, 1137b, 1137c to energize and drive the ultrasonic blade 1149 or other function. The toggle buttons 1137a, 1137b, 1137c can be configured to energize the ultrasonic transducer 1120 with the generator 1100 and energize the ultrasonic blade 1149 with a bipolar energy source also contained within the generator 1100.

The generator 1100 is configurable for use with a variety of surgical instruments. According to various forms, the generator 1100 may be configurable for use with different surgical instruments of different types including, for example, the ultrasonic surgical instrument 1104, the RF electrosurgical instrument 1106, and the multifunction surgical instrument 1108 that integrates RF and ultrasonic energies delivered simultaneously from the generator 1100. Although in the form of FIG. 22 the generator 1100 is shown separate from the surgical instruments 1104, 1106, 1108, in another form the generator 1100 may be formed integrally with any one of the surgical instruments 1104, 1106, 1108 to form a unitary surgical system. As discussed above, the generator 1100 comprises an input device 1110 located on a front panel of the generator 1100 console. The input device 1110 may comprise any suitable device that generates signals suitable for programming the operation of the generator

1100. The generator **1100** also may comprise one or more output devices **1112**. Further aspects of generators for digitally generating electrical signal waveforms and surgical instruments are described in US patent publication US-2017-0086914-A1, which is herein incorporated by reference in its entirety.

FIG. **23** is an end effector **1122** of the example ultrasonic device **1104**, in accordance with at least one aspect of the present disclosure. The end effector **1122** may comprise a blade **1128** that may be coupled to the ultrasonic transducer **1120** via a wave guide. When driven by the ultrasonic transducer **1120**, the blade **1128** may vibrate and, when brought into contact with tissue, may cut and/or coagulate the tissue, as described herein. According to various aspects, and as illustrated in FIG. **23**, the end effector **1122** may also comprise a clamp arm **1140** that may be configured for cooperative action with the blade **1128** of the end effector **1122**. With the blade **1128**, the clamp arm **1140** may comprise a set of jaws. The clamp arm **1140** may be pivotally connected at a distal end of a shaft **1126** of the instrument portion **1104**. The clamp arm **1140** may include a clamp arm tissue pad **1163**, which may be formed from TEFLON® or other suitable low-friction material. The pad **1163** may be mounted for cooperation with the blade **1128**, with pivotal movement of the clamp arm **1140** positioning the clamp pad **1163** in substantially parallel relationship to, and in contact with, the blade **1128**. By this construction, a tissue bite to be clamped may be grasped between the tissue pad **1163** and the blade **1128**. The tissue pad **1163** may be provided with a sawtooth-like configuration including a plurality of axially spaced, proximally extending gripping teeth **1161** to enhance the gripping of tissue in cooperation with the blade **1128**. The clamp arm **1140** may transition from the open position shown in FIG. **23** to a closed position (with the clamp arm **1140** in contact with or proximity to the blade **1128**) in any suitable manner. For example, the handpiece **1105** may comprise a jaw closure trigger. When actuated by a clinician, the jaw closure trigger may pivot the clamp arm **1140** in any suitable manner.

The generator **1100** may be activated to provide the drive signal to the ultrasonic transducer **1120** in any suitable manner. For example, the generator **1100** may comprise a foot switch **1430** (FIG. **24**) coupled to the generator **1100** via a footswitch cable **1432**. A clinician may activate the ultrasonic transducer **1120**, and thereby the ultrasonic transducer **1120** and blade **1128**, by depressing the foot switch **1430**. In addition, or instead of the foot switch **1430**, some aspects of the device **1104** may utilize one or more switches positioned on the handpiece **1105** that, when activated, may cause the generator **1100** to activate the ultrasonic transducer **1120**. In one aspect, for example, the one or more switches may comprise a pair of toggle buttons **1134a**, **1134b**, **1134c** (FIG. **22**), for example, to determine an operating mode of the device **1104**. When the toggle button **1134a** is depressed, for example, the ultrasonic generator **1100** may provide a maximum drive signal to the ultrasonic transducer **1120**, causing it to produce maximum ultrasonic energy output. Depressing toggle button **1134b** may cause the ultrasonic generator **1100** to provide a user-selectable drive signal to the ultrasonic transducer **1120**, causing it to produce less than the maximum ultrasonic energy output. The device **1104** additionally or alternatively may comprise a second switch to, for example, indicate a position of a jaw closure trigger for operating the jaws via the clamp arm **1140** of the end effector **1122**. Also, in some aspects, the ultrasonic generator **1100** may be activated based on the position of the jaw closure

trigger, (e.g., as the clinician depresses the jaw closure trigger to close the jaws via the clamp arm **1140**, ultrasonic energy may be applied).

Additionally or alternatively, the one or more switches may comprise a toggle button **1134c** that, when depressed, causes the generator **1100** to provide a pulsed output (FIG. **22**). The pulses may be provided at any suitable frequency and grouping, for example. In certain aspects, the power level of the pulses may be the power levels associated with toggle buttons **1134a**, **1134b** (maximum, less than maximum), for example.

It will be appreciated that a device **1104** may comprise any combination of the toggle buttons **1134a**, **1134b**, **1134c** (FIG. **22**). For example, the device **1104** could be configured to have only two toggle buttons: a toggle button **1134a** for producing maximum ultrasonic energy output and a toggle button **1134c** for producing a pulsed output at either the maximum or less than maximum power level per. In this way, the drive signal output configuration of the generator **1100** could be five continuous signals, or any discrete number of individual pulsed signals (1, 2, 3, 4, or 5). In certain aspects, the specific drive signal configuration may be controlled based upon, for example, EEPROM settings in the generator **1100** and/or user power level selection(s).

In certain aspects, a two-position switch may be provided as an alternative to a toggle button **1134c** (FIG. **22**). For example, a device **1104** may include a toggle button **1134a** for producing a continuous output at a maximum power level and a two-position toggle button **1134b**. In a first detented position, toggle button **1134b** may produce a continuous output at a less than maximum power level, and in a second detented position the toggle button **1134b** may produce a pulsed output (e.g., at either a maximum or less than maximum power level, depending upon the EEPROM settings).

In some aspects, the RF electro-surgical end effector **1124**, **1125** (FIG. **22**) may also comprise a pair of electrodes. The electrodes may be in communication with the generator **1100**, for example, via a cable. The electrodes may be used, for example, to measure an impedance of a tissue bite present between the clamp arm **1142a**, **1146** and the blade **1142b**, **1149**. The generator **1100** may provide a signal (e.g., a non-therapeutic signal) to the electrodes. The impedance of the tissue bite may be found, for example, by monitoring the current, voltage, etc. of the signal.

In various aspects, the generator **1100** may comprise several separate functional elements, such as modules and/or blocks, as shown in FIG. **24**, a diagram of the surgical system **1000** of FIG. **22**. Different functional elements or modules may be configured for driving the different kinds of surgical devices **1104**, **1106**, **1108**. For example an ultrasonic generator module may drive an ultrasonic device, such as the ultrasonic device **1104**. An electro-surgery/RF generator module may drive the electro-surgical device **1106**. The modules may generate respective drive signals for driving the surgical devices **1104**, **1106**, **1108**. In various aspects, the ultrasonic generator module and/or the electro-surgery/RF generator module each may be formed integrally with the generator **1100**. Alternatively, one or more of the modules may be provided as a separate circuit module electrically coupled to the generator **1100**. (The modules are shown in phantom to illustrate this option.) Also, in some aspects, the electro-surgery/RF generator module may be formed integrally with the ultrasonic generator module, or vice versa.

In accordance with the described aspects, the ultrasonic generator module may produce a drive signal or signals of particular voltages, currents, and frequencies (e.g. 55,500

cycles per second, or Hz). The drive signal or signals may be provided to the ultrasonic device **1104**, and specifically to the transducer **1120**, which may operate, for example, as described above. In one aspect, the generator **1100** may be configured to produce a drive signal of a particular voltage, current, and/or frequency output signal that can be stepped with high resolution, accuracy, and repeatability.

In accordance with the described aspects, the electrosurgery/RF generator module may generate a drive signal or signals with output power sufficient to perform bipolar electrosurgery using radio frequency (RF) energy. In bipolar electrosurgery applications, the drive signal may be provided, for example, to the electrodes of the electrosurgical device **1106**, for example, as described above. Accordingly, the generator **1100** may be configured for therapeutic purposes by applying electrical energy to the tissue sufficient for treating the tissue (e.g., coagulation, cauterization, tissue welding, etc.).

The generator **1100** may comprise an input device **2150** (FIG. 27B) located, for example, on a front panel of the generator **1100** console. The input device **2150** may comprise any suitable device that generates signals suitable for programming the operation of the generator **1100**. In operation, the user can program or otherwise control operation of the generator **1100** using the input device **2150**. The input device **2150** may comprise any suitable device that generates signals that can be used by the generator (e.g., by one or more processors contained in the generator) to control the operation of the generator **1100** (e.g., operation of the ultrasonic generator module and/or electrosurgery/RF generator module). In various aspects, the input device **2150** includes one or more of: buttons, switches, thumbwheels, keyboard, keypad, touch screen monitor, pointing device, remote connection to a general purpose or dedicated computer. In other aspects, the input device **2150** may comprise a suitable user interface, such as one or more user interface screens displayed on a touch screen monitor, for example. Accordingly, by way of the input device **2150**, the user can set or program various operating parameters of the generator, such as, for example, current (I), voltage (V), frequency (f), and/or period (T) of a drive signal or signals generated by the ultrasonic generator module and/or electrosurgery/RF generator module.

The generator **1100** may also comprise an output device **2140** (FIG. 27B) located, for example, on a front panel of the generator **1100** console. The output device **2140** includes one or more devices for providing a sensory feedback to a user. Such devices may comprise, for example, visual feedback devices (e.g., an LCD display screen, LED indicators), audio feedback devices (e.g., a speaker, a buzzer) or tactile feedback devices (e.g., haptic actuators).

Although certain modules and/or blocks of the generator **1100** may be described by way of example, it can be appreciated that a greater or lesser number of modules and/or blocks may be used and still fall within the scope of the aspects. Further, although various aspects may be described in terms of modules and/or blocks to facilitate description, such modules and/or blocks may be implemented by one or more hardware components, e.g., processors, Digital Signal Processors (DSPs), Programmable Logic Devices (PLDs), Application Specific Integrated Circuits (ASICs), circuits, registers and/or software components, e.g., programs, subroutines, logic and/or combinations of hardware and software components.

In one aspect, the ultrasonic generator drive module and electrosurgery/RF drive module **1110** (FIG. 22) may comprise one or more embedded applications implemented as

firmware, software, hardware, or any combination thereof. The modules may comprise various executable modules such as software, programs, data, drivers, application program interfaces (APIs), and so forth. The firmware may be stored in nonvolatile memory (NVM), such as in bit-masked read-only memory (ROM) or flash memory. In various implementations, storing the firmware in ROM may preserve flash memory. The NVM may comprise other types of memory including, for example, programmable ROM (PROM), erasable programmable ROM (EPROM), electrically erasable programmable ROM (EEPROM), or battery backed random-access memory (RAM) such as dynamic RAM (DRAM), Double-Data-Rate DRAM (DDRAM), and/or synchronous DRAM (SDRAM).

In one aspect, the modules comprise a hardware component implemented as a processor for executing program instructions for monitoring various measurable characteristics of the devices **1104**, **1106**, **1108** and generating a corresponding output drive signal or signals for operating the devices **1104**, **1106**, **1108**. In aspects in which the generator **1100** is used in conjunction with the device **1104**, the drive signal may drive the ultrasonic transducer **1120** in cutting and/or coagulation operating modes. Electrical characteristics of the device **1104** and/or tissue may be measured and used to control operational aspects of the generator **1100** and/or provided as feedback to the user. In aspects in which the generator **1100** is used in conjunction with the device **1106**, the drive signal may supply electrical energy (e.g., RF energy) to the end effector **1124** in cutting, coagulation and/or desiccation modes. Electrical characteristics of the device **1106** and/or tissue may be measured and used to control operational aspects of the generator **1100** and/or provided as feedback to the user. In various aspects, as previously discussed, the hardware components may be implemented as DSP, PLD, ASIC, circuits, and/or registers. In one aspect, the processor may be configured to store and execute computer software program instructions to generate the step function output signals for driving various components of the devices **1104**, **1106**, **1108**, such as the ultrasonic transducer **1120** and the end effectors **1122**, **1124**, **1125**.

An electromechanical ultrasonic system includes an ultrasonic transducer, a waveguide, and an ultrasonic blade. The electromechanical ultrasonic system has an initial resonant frequency defined by the physical properties of the ultrasonic transducer, the waveguide, and the ultrasonic blade. The ultrasonic transducer is excited by an alternating voltage $V_g(t)$ and current $I_g(t)$ signal equal to the resonant frequency of the electromechanical ultrasonic system. When the electromechanical ultrasonic system is at resonance, the phase difference between the voltage $V_g(t)$ and current $I_g(t)$ signals is zero. Stated another way, at resonance the inductive impedance is equal to the capacitive impedance. As the ultrasonic blade heats up, the compliance of the ultrasonic blade (modeled as an equivalent capacitance) causes the resonant frequency of the electromechanical ultrasonic system to shift. Thus, the inductive impedance is no longer equal to the capacitive impedance causing a mismatch between the drive frequency and the resonant frequency of the electromechanical ultrasonic system. The system is now operating "off-resonance." The mismatch between the drive frequency and the resonant frequency is manifested as a phase difference between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer. The generator electronics can easily monitor the phase difference between the voltage $V_g(t)$ and current $I_g(t)$ signals and can continuously adjust the drive frequency until the phase difference is once again zero. At this point, the new drive frequency is

equal to the new resonant frequency of the electromechanical ultrasonic system. The change in phase and/or frequency can be used as an indirect measurement of the ultrasonic blade temperature.

As shown in FIG. 25, the electromechanical properties of the ultrasonic transducer may be modeled as an equivalent circuit comprising a first branch having a static capacitance and a second “motional” branch having a serially connected inductance, resistance and capacitance that define the electromechanical properties of a resonator. Known ultrasonic generators may include a tuning inductor for tuning out the static capacitance at a resonant frequency so that substantially all of generator’s drive signal current flows into the motional branch. Accordingly, by using a tuning inductor, the generator’s drive signal current represents the motional branch current, and the generator is thus able to control its drive signal to maintain the ultrasonic transducer’s resonant frequency. The tuning inductor may also transform the phase impedance plot of the ultrasonic transducer to improve the generator’s frequency lock capabilities. However, the tuning inductor must be matched with the specific static capacitance of an ultrasonic transducer at the operational resonance frequency. In other words, a different ultrasonic transducer having a different static capacitance requires a different tuning inductor.

FIG. 25 illustrates an equivalent circuit 1500 of an ultrasonic transducer, such as the ultrasonic transducer 1120, according to one aspect. The circuit 1500 comprises a first “motional” branch having a serially connected inductance L_s , resistance R_s and capacitance C_s that define the electromechanical properties of the resonator, and a second capacitive branch having a static capacitance C_0 . Drive current $I_g(t)$ may be received from a generator at a drive voltage $V_g(t)$, with motional current $I_m(t)$ flowing through the first branch and current $I_g(t) - I_m(t)$ flowing through the capacitive branch. Control of the electromechanical properties of the ultrasonic transducer may be achieved by suitably controlling $I_g(t)$ and $V_g(t)$. As explained above, known generator architectures may include a tuning inductor L_t (shown in phantom in FIG. 25) in a parallel resonance circuit for tuning out the static capacitance C_0 at a resonant frequency so that substantially all of the generator’s current output $I_g(t)$ flows through the motional branch. In this way, control of the motional branch current $I_m(t)$ is achieved by controlling the generator current output $I_g(t)$. The tuning inductor L_t is specific to the static capacitance C_0 of an ultrasonic transducer, however, and a different ultrasonic transducer having a different static capacitance requires a different tuning inductor L_t . Moreover, because the tuning inductor L_t is matched to the nominal value of the static capacitance C_0 at a single resonant frequency, accurate control of the motional branch current $I_m(t)$ is assured only at that frequency. As frequency shifts down with transducer temperature, accurate control of the motional branch current is compromised.

Various aspects of the generator 1100 may not rely on a tuning inductor L_t to monitor the motional branch current $I_m(t)$. Instead, the generator 1100 may use the measured value of the static capacitance C_0 in between applications of power for a specific ultrasonic surgical device 1104 (along with drive signal voltage and current feedback data) to determine values of the motional branch current $I_m(t)$ on a dynamic and ongoing basis (e.g., in real-time). Such aspects of the generator 1100 are therefore able to provide virtual tuning to simulate a system that is tuned or resonant with any value of static capacitance C_0 at any frequency, and not just at a single resonant frequency dictated by a nominal value of the static capacitance C_0 .

FIG. 26 is a simplified block diagram of one aspect of the generator 1100 for providing inductorless tuning as described above, among other benefits. FIGS. 27A-27C illustrate an architecture of the generator 1100 of FIG. 26 according to one aspect. With reference to FIG. 26, the generator 1100 may comprise a patient isolated stage 1520 in communication with a non-isolated stage 1540 via a power transformer 1560. A secondary winding 1580 of the power transformer 1560 is contained in the isolated stage 1520 and may comprise a tapped configuration (e.g., a center-tapped or non-center tapped configuration) to define drive signal outputs 1600a, 1600b, 1600c for outputting drive signals to different surgical devices, such as, for example, an ultrasonic surgical device 1104 and an electro-surgical device 1106. In particular, drive signal outputs 1600a, 1600b, 1600c may output a drive signal (e.g., a 420V RMS drive signal) to an ultrasonic surgical device 1104, and drive signal outputs 1600a, 1600b, 1600c may output a drive signal (e.g., a 100V RMS drive signal) to an electrosurgical device 1106, with output 1600b corresponding to the center tap of the power transformer 1560. The non-isolated stage 1540 may comprise a power amplifier 1620 having an output connected to a primary winding 1640 of the power transformer 1560. In certain aspects the power amplifier 1620 may comprise a push-pull amplifier, for example. The non-isolated stage 1540 may further comprise a programmable logic device 1660 for supplying a digital output to a digital-to-analog converter (DAC) 1680, which in turn supplies a corresponding analog signal to an input of the power amplifier 1620. In certain aspects the programmable logic device 1660 may comprise a field-programmable gate array (FPGA), for example. The programmable logic device 1660, by virtue of controlling the power amplifier’s 1620 input via the DAC 1680, may therefore control any of a number of parameters (e.g., frequency, waveform shape, waveform amplitude) of drive signals appearing at the drive signal outputs 1600a, 1600b, 1600c. In certain aspects and as discussed below, the programmable logic device 1660, in conjunction with a processor (e.g., processor 1740 discussed below), may implement a number of digital signal processing (DSP)-based and/or other control algorithms to control parameters of the drive signals output by the generator 1100.

Power may be supplied to a power rail of the power amplifier 1620 by a switch-mode regulator 1700. In certain aspects the switch-mode regulator 1700 may comprise an adjustable buck regulator, for example. As discussed above, the non-isolated stage 1540 may further comprise a processor 1740, which in one aspect may comprise a DSP processor such as an ADSP-21469 SHARC DSP, available from Analog Devices, Norwood, Mass., for example. In certain aspects the processor 1740 may control operation of the switch-mode power converter 1700 responsive to voltage feedback data received from the power amplifier 1620 by the processor 1740 via an analog-to-digital converter (ADC) 1760. In one aspect, for example, the processor 1740 may receive as input, via the ADC 1760, the waveform envelope of a signal (e.g., an RF signal) being amplified by the power amplifier 1620. The processor 1740 may then control the switch-mode regulator 1700 (e.g., via a pulse-width modulated (PWM) output) such that the rail voltage supplied to the power amplifier 1620 tracks the waveform envelope of the amplified signal. By dynamically modulating the rail voltage of the power amplifier 1620 based on the waveform envelope, the efficiency of the power amplifier 1620 may be significantly improved relative to a fixed rail voltage amplifier scheme. The processor 1740 may be configured for wired or wireless communication.

In certain aspects and as discussed in further detail in connection with FIGS. 28A-28B, the programmable logic device **1660**, in conjunction with the processor **1740**, may implement a direct digital synthesizer (DDS) control scheme to control the waveform shape, frequency and/or amplitude of drive signals output by the generator **1100**. In one aspect, for example, the programmable logic device **1660** may implement a DDS control algorithm **2680** (FIG. 28A) by recalling waveform samples stored in a dynamically-updated look-up table (LUT), such as a RAM LUT which may be embedded in an FPGA. This control algorithm is particularly useful for ultrasonic applications in which an ultrasonic transducer, such as the ultrasonic transducer **1120**, may be driven by a clean sinusoidal current at its resonant frequency. Because other frequencies may excite parasitic resonances, minimizing or reducing the total distortion of the motional branch current may correspondingly minimize or reduce undesirable resonance effects. Because the waveform shape of a drive signal output by the generator **1100** is impacted by various sources of distortion present in the output drive circuit (e.g., the power transformer **1560**, the power amplifier **1620**), voltage and current feedback data based on the drive signal may be input into an algorithm, such as an error control algorithm implemented by the processor **1740**, which compensates for distortion by suitably pre-distorting or modifying the waveform samples stored in the LUT on a dynamic, ongoing basis (e.g., in real-time). In one aspect, the amount or degree of pre-distortion applied to the LUT samples may be based on the error between a computed motional branch current and a desired current waveform shape, with the error being determined on a sample-by-sample basis. In this way, the pre-distorted LUT samples, when processed through the drive circuit, may result in a motional branch drive signal having the desired waveform shape (e.g., sinusoidal) for optimally driving the ultrasonic transducer. In such aspects, the LUT waveform samples will therefore not represent the desired waveform shape of the drive signal, but rather the waveform shape that is required to ultimately produce the desired waveform shape of the motional branch drive signal when distortion effects are taken into account.

The non-isolated stage **1540** may further comprise an ADC **1780** and an ADC **1800** coupled to the output of the power transformer **1560** via respective isolation transformers **1820**, **1840** for respectively sampling the voltage and current of drive signals output by the generator **1100**. In certain aspects, the ADCs **1780**, **1800** may be configured to sample at high speeds (e.g., 80 Msps) to enable oversampling of the drive signals. In one aspect, for example, the sampling speed of the ADCs **1780**, **1800** may enable approximately 200x (depending on drive frequency) oversampling of the drive signals. In certain aspects, the sampling operations of the ADCs **1780**, **1800** may be performed by a single ADC receiving input voltage and current signals via a two-way multiplexer. The use of high-speed sampling in aspects of the generator **1100** may enable, among other things, calculation of the complex current flowing through the motional branch (which may be used in certain aspects to implement DDS-based waveform shape control described above), accurate digital filtering of the sampled signals, and calculation of real power consumption with a high degree of precision. Voltage and current feedback data output by the ADCs **1780**, **1800** may be received and processed (e.g., FIFO buffering, multiplexing) by the programmable logic device **1660** and stored in data memory for subsequent retrieval by, for example, the processor **1740**. As noted above, voltage and current feedback data may be used as

input to an algorithm for pre-distorting or modifying LUT waveform samples on a dynamic and ongoing basis. In certain aspects, this may require each stored voltage and current feedback data pair to be indexed based on, or otherwise associated with, a corresponding LUT sample that was output by the programmable logic device **1660** when the voltage and current feedback data pair was acquired. Synchronization of the LUT samples and the voltage and current feedback data in this manner contributes to the correct timing and stability of the pre-distortion algorithm.

In certain aspects, the voltage and current feedback data may be used to control the frequency and/or amplitude (e.g., current amplitude) of the drive signals. In one aspect, for example, voltage and current feedback data may be used to determine impedance phase, e.g., the phase difference between the voltage and current drive signals. The frequency of the drive signal may then be controlled to minimize or reduce the difference between the determined impedance phase and an impedance phase setpoint (e.g., 0°), thereby minimizing or reducing the effects of harmonic distortion and correspondingly enhancing impedance phase measurement accuracy. The determination of phase impedance and a frequency control signal may be implemented in the processor **1740**, for example, with the frequency control signal being supplied as input to a DDS control algorithm implemented by the programmable logic device **1660**.

The impedance phase may be determined through Fourier analysis. In one aspect, the phase difference between the generator voltage $V_g(t)$ and generator current $I_g(t)$ driving signals may be determined using the Fast Fourier Transform (FFT) or the Discrete Fourier Transform (DFT) as follows:

$$V_g(t) = A_1 \cos(2\pi f_0 t + \varphi_1)$$

$$I_g(t) = A_2 \cos(2\pi f_0 t + \varphi_2)$$

$$V_g(f) = \frac{A_1}{2} (\delta(f - f_0) + \delta(f + f_0)) \exp\left(j2\pi f \frac{\varphi_1}{2\pi f_0}\right)$$

$$I_g(f) = \frac{A_2}{2} (\delta(f - f_0) + \delta(f + f_0)) \exp\left(j2\pi f \frac{\varphi_2}{2\pi f_0}\right)$$

Evaluating the Fourier Transform at the frequency of the sinusoid yields:

$$V_g(f_0) = \frac{A_1}{2} \delta(0) \exp(j\varphi_1) \arg V(f_0) = \varphi_1$$

$$I_g(f_0) = \frac{A_2}{2} \delta(0) \exp(j\varphi_2) \arg I(f_0) = \varphi_2$$

Other approaches include weighted least-squares estimation, Kalman filtering, and space-vector-based techniques. Virtually all of the processing in an FFT or DFT technique may be performed in the digital domain with the aid of the 2-channel high speed ADC **1780**, **1800**, for example. In one technique, the digital signal samples of the voltage and current signals are Fourier transformed with an FFT or a DFT. The phase angle φ at any point in time can be calculated by:

$$\varphi = 2\pi f t + \varphi_0$$

where φ is the phase angle, f is the frequency, t is time, and φ_0 is the phase at $t=0$.

Another technique for determining the phase difference between the voltage $V_g(t)$ and current $I_g(t)$ signals is the

zero-crossing method and produces highly accurate results. For voltage $V_g(t)$ and current $I_g(t)$ signals having the same frequency, each negative to positive zero-crossing of voltage signal $V_g(t)$ triggers the start of a pulse, while each negative to positive zero-crossing of current signal $I_g(t)$ triggers the end of the pulse. The result is a pulse train with a pulse width proportional to the phase angle between the voltage signal and the current signal. In one aspect, the pulse train may be passed through an averaging filter to yield a measure of the phase difference. Furthermore, if the positive to negative zero crossings also are used in a similar manner, and the results averaged, any effects of DC and harmonic components can be reduced. In one implementation, the analog voltage $V_g(t)$ and current $I_g(t)$ signals are converted to digital signals that are high if the analog signal is positive and low if the analog signal is negative. High accuracy phase estimates require sharp transitions between high and low. In one aspect, a Schmitt trigger along with an RC stabilization network may be employed to convert the analog signals into digital signals. In other aspects, an edge triggered RS flip-flop and ancillary circuitry may be employed. In yet another aspect, the zero-crossing technique may employ an eXclusive OR (XOR) gate.

Other techniques for determining the phase difference between the voltage and current signals include Lissajous figures and monitoring the image; methods such as the three-voltmeter method, the crossed-coil method, vector voltmeter and vector impedance methods; and using phase standard instruments, phase-locked loops, and other techniques as described in Phase Measurement, Peter O'Shea, 2000 CRC Press LLC, <<http://www.engnetbase.com>>, which is incorporated herein by reference.

In another aspect, for example, the current feedback data may be monitored in order to maintain the current amplitude of the drive signal at a current amplitude setpoint. The current amplitude setpoint may be specified directly or determined indirectly based on specified voltage amplitude and power setpoints. In certain aspects, control of the current amplitude may be implemented by control algorithm, such as, for example, a proportional-integral-derivative (PID) control algorithm, in the processor 1740. Variables controlled by the control algorithm to suitably control the current amplitude of the drive signal may include, for example, the scaling of the LUT waveform samples stored in the programmable logic device 1660 and/or the full-scale output voltage of the DAC 1680 (which supplies the input to the power amplifier 1620) via a DAC 1860.

The non-isolated stage 1540 may further comprise a processor 1900 for providing, among other things, user interface (UI) functionality. In one aspect, the processor 1900 may comprise an Atmel AT91 SAM9263 processor having an ARM 926EJ-S core, available from Atmel Corporation, San Jose, Calif., for example. Examples of UI functionality supported by the processor 1900 may include audible and visual user feedback, communication with peripheral devices (e.g., via a Universal Serial Bus (USB) interface), communication with a foot switch 1430, communication with an input device 2150 (e.g., a touch screen display) and communication with an output device 2140 (e.g., a speaker). The processor 1900 may communicate with the processor 1740 and the programmable logic device (e.g., via a serial peripheral interface (SPI) bus). Although the processor 1900 may primarily support UI functionality, it may also coordinate with the processor 1740 to implement hazard mitigation in certain aspects. For example, the processor 1900 may be programmed to monitor various aspects of user input and/or other inputs (e.g., touch screen inputs

2150, foot switch 1430 inputs, temperature sensor inputs 2160) and may disable the drive output of the generator 1100 when an erroneous condition is detected.

In certain aspects, both the processor 1740 (FIG. 26, 27A) and the processor 1900 (FIG. 26, 27B) may determine and monitor the operating state of the generator 1100. For processor 1740, the operating state of the generator 1100 may dictate, for example, which control and/or diagnostic processes are implemented by the processor 1740. For processor 1900, the operating state of the generator 1100 may dictate, for example, which elements of a user interface (e.g., display screens, sounds) are presented to a user. The processors 1740, 1900 may independently maintain the current operating state of the generator 1100 and recognize and evaluate possible transitions out of the current operating state. The processor 1740 may function as the master in this relationship and determine when transitions between operating states are to occur. The processor 1900 may be aware of valid transitions between operating states and may confirm if a particular transition is appropriate. For example, when the processor 1740 instructs the processor 1900 to transition to a specific state, the processor 1900 may verify that the requested transition is valid. In the event that a requested transition between states is determined to be invalid by the processor 1900, the processor 1900 may cause the generator 1100 to enter a failure mode.

The non-isolated stage 1540 may further comprise a controller 1960 (FIG. 26, 27B) for monitoring input devices 2150 (e.g., a capacitive touch sensor used for turning the generator 1100 on and off, a capacitive touch screen). In certain aspects, the controller 1960 may comprise at least one processor and/or other controller device in communication with the processor 1900. In one aspect, for example, the controller 1960 may comprise a processor (e.g., a Mega168 8-bit controller available from Atmel) configured to monitor user input provided via one or more capacitive touch sensors. In one aspect, the controller 1960 may comprise a touch screen controller (e.g., a QT5480 touch screen controller available from Atmel) to control and manage the acquisition of touch data from a capacitive touch screen.

In certain aspects, when the generator 1100 is in a "power off" state, the controller 1960 may continue to receive operating power (e.g., via a line from a power supply of the generator 1100, such as the power supply 2110 (FIG. 26) discussed below). In this way, the controller 1960 may continue to monitor an input device 2150 (e.g., a capacitive touch sensor located on a front panel of the generator 1100) for turning the generator 1100 on and off. When the generator 1100 is in the "power off" state, the controller 1960 may wake the power supply (e.g., enable operation of one or more DC/DC voltage converters 2130 (FIG. 26) of the power supply 2110) if activation of the "on/off" input device 2150 by a user is detected. The controller 1960 may therefore initiate a sequence for transitioning the generator 1100 to a "power on" state. Conversely, the controller 1960 may initiate a sequence for transitioning the generator 1100 to the "power off" state if activation of the "on/off" input device 2150 is detected when the generator 1100 is in the "power on" state. In certain aspects, for example, the controller 1960 may report activation of the "on/off" input device 2150 to the processor 1900, which in turn implements the necessary process sequence for transitioning the generator 1100 to the "power off" state. In such aspects, the controller 1960 may have no independent ability for causing the removal of power from the generator 1100 after its "power on" state has been established.

In certain aspects, the controller **1960** may cause the generator **1100** to provide audible or other sensory feedback for alerting the user that a “power on” or “power off” sequence has been initiated. Such an alert may be provided at the beginning of a “power on” or “power off” sequence and prior to the commencement of other processes associated with the sequence.

In certain aspects, the isolated stage **1520** may comprise an instrument interface circuit **1980** to, for example, provide a communication interface between a control circuit of a surgical device (e.g., a control circuit comprising handpiece switches) and components of the non-isolated stage **1540**, such as, for example, the programmable logic device **1660**, the processor **1740** and/or the processor **1900**. The instrument interface circuit **1980** may exchange information with components of the non-isolated stage **1540** via a communication link that maintains a suitable degree of electrical isolation between the stages **1520**, **1540**, such as, for example, an infrared (IR)-based communication link. Power may be supplied to the instrument interface circuit **1980** using, for example, a low-dropout voltage regulator powered by an isolation transformer driven from the non-isolated stage **1540**.

In one aspect, the instrument interface circuit **1980** may comprise a programmable logic device **2000** (e.g., an FPGA) in communication with a signal conditioning circuit **2020** (FIG. 26 and FIG. 27C). The signal conditioning circuit **2020** may be configured to receive a periodic signal from the programmable logic device **2000** (e.g., a 2 kHz square wave) to generate a bipolar interrogation signal having an identical frequency. The interrogation signal may be generated, for example, using a bipolar current source fed by a differential amplifier. The interrogation signal may be communicated to a surgical device control circuit (e.g., by using a conductive pair in a cable that connects the generator **1100** to the surgical device) and monitored to determine a state or configuration of the control circuit. For example, the control circuit may comprise a number of switches, resistors and/or diodes to modify one or more characteristics (e.g., amplitude, rectification) of the interrogation signal such that a state or configuration of the control circuit is uniquely discernible based on the one or more characteristics. In one aspect, for example, the signal conditioning circuit **2020** may comprise an ADC for generating samples of a voltage signal appearing across inputs of the control circuit resulting from passage of interrogation signal therethrough. The programmable logic device **2000** (or a component of the non-isolated stage **1540**) may then determine the state or configuration of the control circuit based on the ADC samples.

In one aspect, the instrument interface circuit **1980** may comprise a first data circuit interface **2040** to enable information exchange between the programmable logic device **2000** (or other element of the instrument interface circuit **1980**) and a first data circuit disposed in or otherwise associated with a surgical device. In certain aspects, for example, a first data circuit **2060** may be disposed in a cable integrally attached to a surgical device handpiece, or in an adaptor for interfacing a specific surgical device type or model with the generator **1100**. In certain aspects, the first data circuit may comprise a non-volatile storage device, such as an electrically erasable programmable read-only memory (EEPROM) device. In certain aspects and referring again to FIG. 26, the first data circuit interface **2040** may be implemented separately from the programmable logic device **2000** and comprise suitable circuitry (e.g., discrete logic devices, a processor) to enable communication

between the programmable logic device **2000** and the first data circuit. In other aspects, the first data circuit interface **2040** may be integral with the programmable logic device **2000**.

In certain aspects, the first data circuit **2060** may store information pertaining to the particular surgical device with which it is associated. Such information may include, for example, a model number, a serial number, a number of operations in which the surgical device has been used, and/or any other type of information. This information may be read by the instrument interface circuit **1980** (e.g., by the programmable logic device **2000**), transferred to a component of the non-isolated stage **1540** (e.g., to programmable logic device **1660**, processor **1740** and/or processor **1900**) for presentation to a user via an output device **2140** and/or for controlling a function or operation of the generator **1100**. Additionally, any type of information may be communicated to first data circuit **2060** for storage therein via the first data circuit interface **2040** (e.g., using the programmable logic device **2000**). Such information may comprise, for example, an updated number of operations in which the surgical device has been used and/or dates and/or times of its usage.

As discussed previously, a surgical instrument may be detachable from a handpiece (e.g., instrument **1106** may be detachable from handpiece **1107**) to promote instrument interchangeability and/or disposability. In such cases, known generators may be limited in their ability to recognize particular instrument configurations being used and to optimize control and diagnostic processes accordingly. The addition of readable data circuits to surgical device instruments to address this issue is problematic from a compatibility standpoint, however. For example, it may be impractical to design a surgical device to maintain backward compatibility with generators that lack the requisite data reading functionality due to, for example, differing signal schemes, design complexity and cost. Other aspects of instruments address these concerns by using data circuits that may be implemented in existing surgical instruments economically and with minimal design changes to preserve compatibility of the surgical devices with current generator platforms.

Additionally, aspects of the generator **1100** may enable communication with instrument-based data circuits. For example, the generator **1100** may be configured to communicate with a second data circuit (e.g., a data circuit) contained in an instrument (e.g., instrument **1104**, **1106** or **1108**) of a surgical device. The instrument interface circuit **1980** may comprise a second data circuit interface **2100** to enable this communication. In one aspect, the second data circuit interface **2100** may comprise a tri-state digital interface, although other interfaces may also be used. In certain aspects, the second data circuit may generally be any circuit for transmitting and/or receiving data. In one aspect, for example, the second data circuit may store information pertaining to the particular surgical instrument with which it is associated. Such information may include, for example, a model number, a serial number, a number of operations in which the surgical instrument has been used, and/or any other type of information. Additionally or alternatively, any type of information may be communicated to the second data circuit for storage therein via the second data circuit interface **2100** (e.g., using the programmable logic device **2000**). Such information may comprise, for example, an updated number of operations in which the instrument has been used and/or dates and/or times of its usage. In certain aspects, the second data circuit may transmit data acquired by one or more sensors (e.g., an instrument-based tempera-

ture sensor). In certain aspects, the second data circuit may receive data from the generator **1100** and provide an indication to a user (e.g., an LED indication or other visible indication) based on the received data.

In certain aspects, the second data circuit and the second data circuit interface **2100** may be configured such that communication between the programmable logic device **2000** and the second data circuit can be effected without the need to provide additional conductors for this purpose (e.g., dedicated conductors of a cable connecting a handpiece to the generator **1100**). In one aspect, for example, information may be communicated to and from the second data circuit using a one-wire bus communication scheme implemented on existing cabling, such as one of the conductors used to transmit interrogation signals from the signal conditioning circuit **2020** to a control circuit in a handpiece. In this way, design changes or modifications to the surgical device that might otherwise be necessary are minimized or reduced. Moreover, because different types of communications can be implemented over a common physical channel (either with or without frequency-band separation), the presence of a second data circuit may be “invisible” to generators that do not have the requisite data reading functionality, thus enabling backward compatibility of the surgical device instrument.

In certain aspects, the isolated stage **1520** may comprise at least one blocking capacitor **2960-1** (FIG. **27C**) connected to the drive signal output **1600b** to prevent passage of DC current to a patient. A single blocking capacitor may be required to comply with medical regulations or standards, for example. While failure in single-capacitor designs is relatively uncommon, such failure may nonetheless have negative consequences. In one aspect, a second blocking capacitor **2960-2** may be provided in series with the blocking capacitor **2960-1**, with current leakage from a point between the blocking capacitors **2960-1**, **2960-2** being monitored by, for example, an ADC **2980** for sampling a voltage induced by leakage current. The samples may be received by the programmable logic device **2000**, for example. Based on changes in the leakage current (as indicated by the voltage samples in the aspect of FIG. **26**), the generator **1100** may determine when at least one of the blocking capacitors **2960-1**, **2960-2** has failed. Accordingly, the aspect of FIG. **26** may provide a benefit over single-capacitor designs having a single point of failure.

In certain aspects, the non-isolated stage **1540** may comprise a power supply **2110** for outputting DC power at a suitable voltage and current. The power supply may comprise, for example, a 400 W power supply for outputting a 48 VDC system voltage. As discussed above, the power supply **2110** may further comprise one or more DC/DC voltage converters **2130** for receiving the output of the power supply to generate DC outputs at the voltages and currents required by the various components of the generator **1100**. As discussed above in connection with the controller **1960**, one or more of the DC/DC voltage converters **2130** may receive an input from the controller **1960** when activation of the “on/off” input device **2150** by a user is detected by the controller **1960** to enable operation of, or wake, the DC/DC voltage converters **2130**.

FIGS. **28A-28B** illustrate certain functional and structural aspects of one aspect of the generator **1100**. Feedback indicating current and voltage output from the secondary winding **1580** of the power transformer **1560** is received by the ADCs **1780**, **1800**, respectively. As shown, the ADCs **1780**, **1800** may be implemented as a 2-channel ADC and may sample the feedback signals at a high speed (e.g., 80

MSPs) to enable oversampling (e.g., approximately 200× oversampling) of the drive signals. The current and voltage feedback signals may be suitably conditioned in the analog domain (e.g., amplified, filtered) prior to processing by the ADCs **1780**, **1800**. Current and voltage feedback samples from the ADCs **1780**, **1800** may be individually buffered and subsequently multiplexed or interleaved into a single data stream within block **2120** of the programmable logic device **1660**. In the aspect of FIGS. **28A-28B**, the programmable logic device **1660** comprises an FPGA.

The multiplexed current and voltage feedback samples may be received by a parallel data acquisition port (PDAP) implemented within block **2144** of the processor **1740**. The PDAP may comprise a packing unit for implementing any of a number of methodologies for correlating the multiplexed feedback samples with a memory address. In one aspect, for example, feedback samples corresponding to a particular LUT sample output by the programmable logic device **1660** may be stored at one or more memory addresses that are correlated or indexed with the LUT address of the LUT sample. In another aspect, feedback samples corresponding to a particular LUT sample output by the programmable logic device **1660** may be stored, along with the LUT address of the LUT sample, at a common memory location. In any event, the feedback samples may be stored such that the address of the LUT sample from which a particular set of feedback samples originated may be subsequently ascertained. As discussed above, synchronization of the LUT sample addresses and the feedback samples in this way contributes to the correct timing and stability of the pre-distortion algorithm. A direct memory access (DMA) controller implemented at block **2166** of the processor **1740** may store the feedback samples (and any LUT sample address data, where applicable) at a designated memory location **2180** of the processor **1740** (e.g., internal RAM).

Block **2200** of the processor **1740** may implement a pre-distortion algorithm for pre-distorting or modifying the LUT samples stored in the programmable logic device **1660** on a dynamic, ongoing basis. As discussed above, pre-distortion of the LUT samples may compensate for various sources of distortion present in the output drive circuit of the generator **1100**. The pre-distorted LUT samples, when processed through the drive circuit, will therefore result in a drive signal having the desired waveform shape (e.g., sinusoidal) for optimally driving the ultrasonic transducer.

At block **2220** of the pre-distortion algorithm, the current through the motional branch of the ultrasonic transducer is determined. The motional branch current may be determined using Kirchhoff’s Current Law based on, for example, the current and voltage feedback samples stored at memory location **2180** (which, when suitably scaled, may be representative of I_g and V_g in the model of FIG. **25** discussed above), a value of the ultrasonic transducer static capacitance C_o (measured or known a priori) and a known value of the drive frequency. A motional branch current sample for each set of stored current and voltage feedback samples associated with a LUT sample may be determined.

At block **2240** of the pre-distortion algorithm, each motional branch current sample determined at block **2220** is compared to a sample of a desired current waveform shape to determine a difference, or sample amplitude error, between the compared samples. For this determination, the sample of the desired current waveform shape may be supplied, for example, from a waveform shape LUT **2260** containing amplitude samples for one cycle of a desired current waveform shape. The particular LUT sample of the desired current waveform shape from the LUT **2260** used for

the comparison may be dictated by the LUT sample address associated with the motional branch current sample used in the comparison. Accordingly, the input of the motional branch current to block **2240** may be synchronized with the input of its associated LUT sample address to block **2240**. The LUT samples stored in the programmable logic device **1660** and the LUT samples stored in the waveform shape LUT **2260** may therefore be equal in number. In certain aspects, the desired current waveform shape represented by the LUT samples stored in the waveform shape LUT **2260** may be a fundamental sine wave. Other waveform shapes may be desirable. For example, it is contemplated that a fundamental sine wave for driving main longitudinal motion of an ultrasonic transducer superimposed with one or more other drive signals at other frequencies, such as a third order harmonic for driving at least two mechanical resonances for beneficial vibrations of transverse or other modes, could be used.

Each value of the sample amplitude error determined at block **2240** may be transmitted to the LUT of the programmable logic device **1660** (shown at block **2280** in FIG. **28A**) along with an indication of its associated LUT address. Based on the value of the sample amplitude error and its associated address (and, optionally, values of sample amplitude error for the same LUT address previously received), the LUT **2280** (or other control block of the programmable logic device **1660**) may pre-distort or modify the value of the LUT sample stored at the LUT address such that the sample amplitude error is reduced or minimized. It will be appreciated that such pre-distortion or modification of each LUT sample in an iterative manner across the entire range of LUT addresses will cause the waveform shape of the generator's output current to match or conform to the desired current waveform shape represented by the samples of the waveform shape LUT **2260**.

Current and voltage amplitude measurements, power measurements and impedance measurements may be determined at block **2300** of the processor **1740** based on the current and voltage feedback samples stored at memory location **2180**. Prior to the determination of these quantities, the feedback samples may be suitably scaled and, in certain aspects, processed through a suitable filter **2320** to remove noise resulting from, for example, the data acquisition process and induced harmonic components. The filtered voltage and current samples may therefore substantially represent the fundamental frequency of the generator's drive output signal. In certain aspects, the filter **2320** may be a finite impulse response (FIR) filter applied in the frequency domain. Such aspects may use the Fast Fourier Transform (FFT) of the output drive signal current and voltage signals. In certain aspects, the resulting frequency spectrum may be used to provide additional generator functionality. In one aspect, for example, the ratio of the second and/or third order harmonic component relative to the fundamental frequency component may be used as a diagnostic indicator.

At block **2340** (FIG. **28B**), a root mean square (RMS) calculation may be applied to a sample size of the current feedback samples representing an integral number of cycles of the drive signal to generate a measurement I_{rms} representing the drive signal output current.

At block **2360**, a root mean square (RMS) calculation may be applied to a sample size of the voltage feedback samples representing an integral number of cycles of the drive signal to determine a measurement V_{rms} representing the drive signal output voltage.

At block **2380**, the current and voltage feedback samples may be multiplied point by point, and a mean calculation is

applied to samples representing an integral number of cycles of the drive signal to determine a measurement P_r of the generator's real output power.

At block **2400**, measurement P_a of the generator's apparent output power may be determined as the product $V_{rms} \cdot I_{rms}$.

At block **2420**, measurement Z_m of the load impedance magnitude may be determined as the quotient V_{rms}/I_{rms} .

In certain aspects, the quantities I_{rms} , V_{rms} , P_r , P_a and Z_m determined at blocks **2340**, **2360**, **2380**, **2400** and **2420** may be used by the generator **1100** to implement any of a number of control and/or diagnostic processes. In certain aspects, any of these quantities may be communicated to a user via, for example, an output device **2140** integral with the generator **1100** or an output device **2140** connected to the generator **1100** through a suitable communication interface (e.g., a USB interface). Various diagnostic processes may include, without limitation, handpiece integrity, instrument integrity, instrument attachment integrity, instrument overload, approaching instrument overload, frequency lock failure, over-voltage condition, over-current condition, over-power condition, voltage sense failure, current sense failure, audio indication failure, visual indication failure, short circuit condition, power delivery failure, or blocking capacitor failure, for example.

Block **2440** of the processor **1740** may implement a phase control algorithm for determining and controlling the impedance phase of an electrical load (e.g., the ultrasonic transducer) driven by the generator **1100**. As discussed above, by controlling the frequency of the drive signal to minimize or reduce the difference between the determined impedance phase and an impedance phase setpoint (e.g., 0°), the effects of harmonic distortion may be minimized or reduced, and the accuracy of the phase measurement increased.

The phase control algorithm receives as input the current and voltage feedback samples stored in the memory location **2180**. Prior to their use in the phase control algorithm, the feedback samples may be suitably scaled and, in certain aspects, processed through a suitable filter **2460** (which may be identical to filter **2320**) to remove noise resulting from the data acquisition process and induced harmonic components, for example. The filtered voltage and current samples may therefore substantially represent the fundamental frequency of the generator's drive output signal.

At block **2480** of the phase control algorithm, the current through the motional branch of the ultrasonic transducer is determined. This determination may be identical to that described above in connection with block **2220** of the pre-distortion algorithm. The output of block **2480** may thus be, for each set of stored current and voltage feedback samples associated with a LUT sample, a motional branch current sample.

At block **2500** of the phase control algorithm, impedance phase is determined based on the synchronized input of motional branch current samples determined at block **2480** and corresponding voltage feedback samples. In certain aspects, the impedance phase is determined as the average of the impedance phase measured at the rising edge of the waveforms and the impedance phase measured at the falling edge of the waveforms.

At block **2520** of the of the phase control algorithm, the value of the impedance phase determined at block **2220** is compared to phase setpoint **2540** to determine a difference, or phase error, between the compared values.

At block **2560** (FIG. **28A**) of the phase control algorithm, based on a value of phase error determined at block **2520** and the impedance magnitude determined at block **2420**, a

frequency output for controlling the frequency of the drive signal is determined. The value of the frequency output may be continuously adjusted by the block 2560 and transferred to a DDS control block 2680 (discussed below) in order to maintain the impedance phase determined at block 2500 at the phase setpoint (e.g., zero phase error). In certain aspects, the impedance phase may be regulated to a 0° phase setpoint. In this way, any harmonic distortion will be centered about the crest of the voltage waveform, enhancing the accuracy of phase impedance determination.

Block 2580 of the processor 1740 may implement an algorithm for modulating the current amplitude of the drive signal in order to control the drive signal current, voltage and power in accordance with user specified setpoints, or in accordance with requirements specified by other processes or algorithms implemented by the generator 1100. Control of these quantities may be realized, for example, by scaling the LUT samples in the LUT 2280 and/or by adjusting the full-scale output voltage of the DAC 1680 (which supplies the input to the power amplifier 1620) via a DAC 1860. Block 2600 (which may be implemented as a PID controller in certain aspects) may receive, as input, current feedback samples (which may be suitably scaled and filtered) from the memory location 2180. The current feedback samples may be compared to a “current demand” I_d value dictated by the controlled variable (e.g., current, voltage or power) to determine if the drive signal is supplying the necessary current. In aspects in which drive signal current is the control variable, the current demand I_d may be specified directly by a current setpoint 2620A (I_{sp}). For example, an RMS value of the current feedback data (determined as in block 2340) may be compared to user-specified RMS current setpoint I_{sp} to determine the appropriate controller action. If, for example, the current feedback data indicates an RMS value less than the current setpoint I_{sp} , LUT scaling and/or the full-scale output voltage of the DAC 1680 may be adjusted by the block 2600 such that the drive signal current is increased. Conversely, block 2600 may adjust LUT scaling and/or the full-scale output voltage of the DAC 1680 to decrease the drive signal current when the current feedback data indicates an RMS value greater than the current setpoint I_{sp} .

In aspects in which the drive signal voltage is the control variable, the current demand I_d may be specified indirectly, for example, based on the current required to maintain a desired voltage setpoint 2620B (V_{sp}) given the load impedance magnitude Z_m measured at block 2420 (e.g. $I_d = V_{sp} / Z_m$). Similarly, in aspects in which drive signal power is the control variable, the current demand I_d may be specified indirectly, for example, based on the current required to maintain a desired power setpoint 2620C (P_{sp}) given the voltage V_{rms} measured at blocks 2360 (e.g. $I_d = P_{sp} / V_{rms}$).

Block 2680 (FIG. 28A) may implement a DDS control algorithm for controlling the drive signal by recalling LUT samples stored in the LUT 2280. In certain aspects, the DDS control algorithm may be a numerically-controlled oscillator (NCO) algorithm for generating samples of a waveform at a fixed clock rate using a point (memory location)-skipping technique. The NCO algorithm may implement a phase accumulator, or frequency-to-phase converter, that functions as an address pointer for recalling LUT samples from the LUT 2280. In one aspect, the phase accumulator may be a D step size, modulo N phase accumulator, where D is a positive integer representing a frequency control value, and N is the number of LUT samples in the LUT 2280. A frequency control value of D=1, for example, may cause the phase accumulator to sequentially point to every address of

the LUT 2280, resulting in a waveform output replicating the waveform stored in the LUT 2280. When $D > 1$, the phase accumulator may skip addresses in the LUT 2280, resulting in a waveform output having a higher frequency. Accordingly, the frequency of the waveform generated by the DDS control algorithm may therefore be controlled by suitably varying the frequency control value. In certain aspects, the frequency control value may be determined based on the output of the phase control algorithm implemented at block 2440. The output of block 2680 may supply the input of DAC 1680, which in turn supplies a corresponding analog signal to an input of the power amplifier 1620.

Block 2700 of the processor 1740 may implement a switch-mode converter control algorithm for dynamically modulating the rail voltage of the power amplifier 1620 based on the waveform envelope of the signal being amplified, thereby improving the efficiency of the power amplifier 1620. In certain aspects, characteristics of the waveform envelope may be determined by monitoring one or more signals contained in the power amplifier 1620. In one aspect, for example, characteristics of the waveform envelope may be determined by monitoring the minima of a drain voltage (e.g., a MOSFET drain voltage) that is modulated in accordance with the envelope of the amplified signal. A minima voltage signal may be generated, for example, by a voltage minima detector coupled to the drain voltage. The minima voltage signal may be sampled by ADC 1760, with the output minima voltage samples being received at block 2720 of the switch-mode converter control algorithm. Based on the values of the minima voltage samples, block 2740 may control a PWM signal output by a PWM generator 2760, which, in turn, controls the rail voltage supplied to the power amplifier 1620 by the switch-mode regulator 1700. In certain aspects, as long as the values of the minima voltage samples are less than a minima target 2780 input into block 2720, the rail voltage may be modulated in accordance with the waveform envelope as characterized by the minima voltage samples. When the minima voltage samples indicate low envelope power levels, for example, block 2740 may cause a low rail voltage to be supplied to the power amplifier 1620, with the full rail voltage being supplied only when the minima voltage samples indicate maximum envelope power levels. When the minima voltage samples fall below the minima target 2780, block 2740 may cause the rail voltage to be maintained at a minimum value suitable for ensuring proper operation of the power amplifier 1620.

FIG. 29 is a schematic diagram of one aspect of an electrical circuit 2900, suitable for driving an ultrasonic transducer, such as ultrasonic transducer 1120, in accordance with at least one aspect of the present disclosure. The electrical circuit 2900 comprises an analog multiplexer 2980. The analog multiplexer 2980 multiplexes various signals from the upstream channels SCL-A, SDA-A such as ultrasonic, battery, and power control circuit. A current sensor 2982 is coupled in series with the return or ground leg of the power supply circuit to measure the current supplied by the power supply. A field effect transistor (FET) temperature sensor 2984 provides the ambient temperature. A pulse width modulation (PWM) watchdog timer 2988 automatically generates a system reset if the main program neglects to periodically service it. It is provided to automatically reset the electrical circuit 2900 when it hangs or freezes because of a software or hardware fault. It will be appreciated that the electrical circuit 2900 may be configured as an RF driver circuit for driving the ultrasonic transducer or for driving RF electrodes such as the electrical circuit 3600 shown in FIG. 34, for example. Accordingly,

with reference now back to FIG. 29, the electrical circuit 2900 can be used to drive both ultrasonic transducers and RF electrodes interchangeably. If driven simultaneously, filter circuits may be provided in the corresponding first stage circuits 3404 (FIG. 32) to select either the ultrasonic waveform or the RF waveform. Such filtering techniques are described in commonly owned U.S. Pat. Pub. No. US-2017-0086910-A1, titled TECHNIQUES FOR CIRCUIT TOPOLOGIES FOR COMBINED GENERATOR, which is herein incorporated by reference in its entirety.

A drive circuit 2986 provides left and right ultrasonic energy outputs. A digital signal that represents the signal waveform is provided to the SCL-A, SDA-A inputs of the analog multiplexer 2980 from a control circuit, such as the control circuit 3200 (FIG. 30). A digital-to-analog converter 2990 (DAC) converts the digital input to an analog output to drive a PWM circuit 2992 coupled to an oscillator 2994. The PWM circuit 2992 provides a first signal to a first gate drive circuit 2996a coupled to a first transistor output stage 2998a to drive a first Ultrasonic (LEFT) energy output. The PWM circuit 2992 also provides a second signal to a second gate drive circuit 2996b coupled to a second transistor output stage 2998b to drive a second Ultrasonic (RIGHT) energy output. A voltage sensor 2999 is coupled between the Ultrasonic LEFT/RIGHT output terminals to measure the output voltage. The drive circuit 2986, the first and second drive circuits 2996a, 2996b, and the first and second transistor output stages 2998a, 2998b define a first stage amplifier circuit. In operation, the control circuit 3200 (FIG. 30) generates a digital waveform 4300 (FIG. 37) employing circuits such as direct digital synthesis (DDS) circuits 4100, 4200 (FIGS. 35 and 36). The DAC 2990 receives the digital waveform 4300 and converts it into an analog waveform, which is received and amplified by the first stage amplifier circuit.

FIG. 30 is a schematic diagram of a control circuit 3200, such as control circuit 3212, in accordance with at least one aspect of the present disclosure. The control circuit 3200 is located within a housing of the battery assembly. The battery assembly is the energy source for a variety of local power supplies 3215. The control circuit comprises a main processor 3214 coupled via an interface master 3218 to various downstream circuits by way of outputs SCL-A and SDA-A, SCL-B and SDA-B, SCL-C and SDA-C, for example. In one aspect, the interface master 3218 is a general purpose serial interface such as an I²C serial interface. The main processor 3214 also is configured to drive switches 3224 through general purposes input/output (GPIO) 3220, a display 3226 (e.g., and LCD display), and various indicators 3228 through GPIO 3222. A watchdog processor 3216 is provided to control the main processor 3214. A switch 3230 is provided in series with a battery 3211 to activate the control circuit 3212 upon insertion of the battery assembly into a handle assembly of a surgical instrument.

In one aspect, the main processor 3214 is coupled to the electrical circuit 2900 (FIG. 29) by way of output terminals SCL-A, SDA-A. The main processor 3214 comprises a memory for storing tables of digitized drive signals or waveforms that are transmitted to the electrical circuit 2900 for driving the ultrasonic transducer 1120, for example. In other aspects, the main processor 3214 may generate a digital waveform and transmit it to the electrical circuit 2900 or may store the digital waveform for later transmission to the electrical circuit 2900. The main processor 3214 also may provide RF drive by way of output terminals SCL-B, SDA-B and various sensors (e.g., Hall-effect sensors, magneto-rheological fluid (MRF) sensors, etc.) by way of output

terminals SCL-C, SDA-C. In one aspect, the main processor 3214 is configured to sense the presence of ultrasonic drive circuitry and/or RF drive circuitry to enable appropriate software and user interface functionality.

In one aspect, the main processor 3214 may be an LM 4F230H5QR, available from Texas Instruments, for example. In at least one example, the Texas Instruments LM4F230H5QR is an ARM Cortex-M4F Processor Core comprising on-chip memory of 256 KB single-cycle flash memory, or other non-volatile memory, up to 40 MHz, a prefetch buffer to improve performance above 40 MHz, a 32 KB single-cycle serial random access memory (SRAM), internal read-only memory (ROM) loaded with StellarisWare® software, 2 KB electrically erasable programmable read-only memory (EEPROM), one or more pulse width modulation (PWM) modules, one or more quadrature encoder inputs (QED analog, one or more 12-bit Analog-to-Digital Converters (ADC) with 12 analog input channels, among other features that are readily available from the product datasheet. Other processors may be readily substituted and, accordingly, the present disclosure should not be limited in this context.

FIG. 31 shows a simplified block circuit diagram illustrating another electrical circuit 3300 contained within a modular ultrasonic surgical instrument 3334, in accordance with at least one aspect of the present disclosure. The electrical circuit 3300 includes a processor 3302, a clock 3330, a memory 3326, a power supply 3304 (e.g., a battery), a switch 3306, such as a metal-oxide semiconductor field effect transistor (MOSFET) power switch, a drive circuit 3308 (PLL), a transformer 3310, a signal smoothing circuit 3312 (also referred to as a matching circuit and can be, for example, a tank circuit), a sensing circuit 3314, a transducer 1120, and a shaft assembly (e.g. shaft assembly 1126, 1129) comprising an ultrasonic transmission waveguide that terminates at an ultrasonic blade (e.g. ultrasonic blade 1128, 1149) which may be referred to herein simply as the waveguide.

One feature of the present disclosure that severs dependency on high voltage (120 VAC) input power (a characteristic of general ultrasonic cutting devices) is the utilization of low-voltage switching throughout the wave-forming process and the amplification of the driving signal only directly before the transformer stage. For this reason, in one aspect of the present disclosure, power is derived from only a battery, or a group of batteries, small enough to fit either within a handle assembly. State-of-the-art battery technology provides powerful batteries of a few centimeters in height and width and a few millimeters in depth. By combining the features of the present disclosure to provide a self-contained and self-powered ultrasonic device, a reduction in manufacturing cost may be achieved.

The output of the power supply 3304 is fed to and powers the processor 3302. The processor 3302 receives and outputs signals and, as will be described below, functions according to custom logic or in accordance with computer programs that are executed by the processor 3302. As discussed above, the electrical circuit 3300 can also include a memory 3326, preferably, random access memory (RAM), that stores computer-readable instructions and data.

The output of the power supply 3304 also is directed to the switch 3306 having a duty cycle controlled by the processor 3302. By controlling the on-time for the switch 3306, the processor 3302 is able to dictate the total amount of power that is ultimately delivered to the transducer 1120. In one aspect, the switch 3306 is a MOSFET, although other switches and switching configurations are adaptable as well.

The output of the switch **3306** is fed to a drive circuit **3308** that contains, for example, a phase detecting phase-locked loop (PLL) and/or a low-pass filter and/or a voltage-controlled oscillator. The output of the switch **3306** is sampled by the processor **3302** to determine the voltage and current of the output signal (V_{IN} and I_{IN} , respectively). These values are used in a feedback architecture to adjust the pulse width modulation of the switch **3306**. For instance, the duty cycle of the switch **3306** can vary from about 20% to about 80%, depending on the desired and actual output from the switch **3306**.

The drive circuit **3308**, which receives the signal from the switch **3306**, includes an oscillatory circuit that turns the output of the switch **3306** into an electrical signal having an ultrasonic frequency, e.g., 55 kHz (VCO). As explained above, a smoothed-out version of this ultrasonic waveform is ultimately fed to the ultrasonic transducer **1120** to produce a resonant sine wave along an ultrasonic transmission waveguide.

At the output of the drive circuit **3308** is a transformer **3310** that is able to step up the low voltage signal(s) to a higher voltage. It is noted that upstream switching, prior to the transformer **3310**, is performed at low (e.g., battery driven) voltages, something that, to date, has not been possible for ultrasonic cutting and cautery devices. This is at least partially due to the fact that the device advantageously uses low on-resistance MOSFET switching devices. Low on-resistance MOSFET switches are advantageous, as they produce lower switching losses and less heat than a traditional MOSFET device and allow higher current to pass through. Therefore, the switching stage (pre-transformer) can be characterized as low voltage/high current. To ensure the lower on-resistance of the amplifier MOSFET(s), the MOSFET(s) are run, for example, at 10 V. In such a case, a separate 10 VDC power supply can be used to feed the MOSFET gate, which ensures that the MOSFET is fully on and a reasonably low on resistance is achieved. In one aspect of the present disclosure, the transformer **3310** steps up the battery voltage to 120 V root-mean-square (RMS). Transformers are known in the art and are, therefore, not explained here in detail.

In the circuit configurations described, circuit component degradation can negatively impact the circuit performance of the circuit. One factor that directly affects component performance is heat. Known circuits generally monitor switching temperatures (e.g., MOSFET temperatures). However, because of the technological advancements in MOSFET designs, and the corresponding reduction in size, MOSFET temperatures are no longer a valid indicator of circuit loads and heat. For this reason, in accordance with at least one aspect of the present disclosure, the sensing circuit **3314** senses the temperature of the transformer **3310**. This temperature sensing is advantageous as the transformer **3310** is run at or very close to its maximum temperature during use of the device. Additional temperature will cause the core material, e.g., the ferrite, to break down and permanent damage can occur. The present disclosure can respond to a maximum temperature of the transformer **3310** by, for example, reducing the driving power in the transformer **3310**, signaling the user, turning the power off, pulsing the power, or other appropriate responses.

In one aspect of the present disclosure, the processor **3302** is communicatively coupled to the end effector (e.g. **1122**, **1125**), which is used to place material in physical contact with the ultrasonic blade (e.g. **1128**, **1149**). Sensors are provided that measure, at the end effector, a clamping force value (existing within a known range) and, based upon the

received clamping force value, the processor **3302** varies the motional voltage V_M . Because high force values combined with a set motional rate can result in high blade temperatures, a temperature sensor **3332** can be communicatively coupled to the processor **3302**, where the processor **3302** is operable to receive and interpret a signal indicating a current temperature of the blade from the temperature sensor **3336** and to determine a target frequency of blade movement based upon the received temperature. In another aspect, force sensors such as strain gages or pressure sensors may be coupled to the trigger (e.g. **1143**, **1147**) to measure the force applied to the trigger by the user. In another aspect, force sensors such as strain gages or pressure sensors may be coupled to a switch button such that displacement intensity corresponds to the force applied by the user to the switch button.

In accordance with at least one aspect of the present disclosure, the PLL portion of the drive circuit **3308**, which is coupled to the processor **3302**, is able to determine a frequency of waveguide movement and communicate that frequency to the processor **3302**. The processor **3302** stores this frequency value in the memory **3326** when the device is turned off. By reading the clock **3330**, the processor **3302** is able to determine an elapsed time after the device is shut off and retrieve the last frequency of waveguide movement if the elapsed time is less than a predetermined value. The device can then start up at the last frequency, which, presumably, is the optimum frequency for the current load. Modular Battery Powered Handheld Surgical Instrument with Multistage Generator Circuits

In another aspect, the present disclosure provides a modular battery powered handheld surgical instrument with multistage generator circuits. Disclosed is a surgical instrument that includes a battery assembly, a handle assembly, and a shaft assembly where the battery assembly and the shaft assembly are configured to mechanically and electrically connect to the handle assembly. The battery assembly includes a control circuit configured to generate a digital waveform. The handle assembly includes a first stage circuit configured to receive the digital waveform, convert the digital waveform into an analog waveform, and amplify the analog waveform. The shaft assembly includes a second stage circuit coupled to the first stage circuit to receive, amplify, and apply the analog waveform to a load.

In one aspect, the present disclosure provides a surgical instrument, comprising: a battery assembly, comprising a control circuit comprising a battery, a memory coupled to the battery, and a processor coupled to the memory and the battery, wherein the processor is configured to generate a digital waveform; a handle assembly comprising a first stage circuit coupled to the processor, the first stage circuit comprising a digital-to-analog (DAC) converter and a first stage amplifier circuit, wherein the DAC is configured to receive the digital waveform and convert the digital waveform into an analog waveform, wherein the first stage amplifier circuit is configured to receive and amplify the analog waveform; and a shaft assembly comprising a second stage circuit coupled to the first stage amplifier circuit to receive the analog waveform, amplify the analog waveform, and apply the analog waveform to a load; wherein the battery assembly and the shaft assembly are configured to mechanically and electrically connect to the handle assembly.

The load may comprise any one of an ultrasonic transducer, an electrode, or a sensor, or any combinations thereof. The first stage circuit may comprise a first stage ultrasonic drive circuit and a first stage high-frequency current drive circuit. The control circuit may be configured to drive the

first stage ultrasonic drive circuit and the first stage high-frequency current drive circuit independently or simultaneously. The first stage ultrasonic drive circuit may be configured to couple to a second stage ultrasonic drive circuit. The second stage ultrasonic drive circuit may be configured to couple to an ultrasonic transducer. The first stage high-frequency current drive circuit may be configured to couple to a second stage high-frequency drive circuit. The second stage high-frequency drive circuit may be configured to couple to an electrode.

The first stage circuit may comprise a first stage sensor drive circuit. The first stage sensor drive circuit may be configured to a second stage sensor drive circuit. The second stage sensor drive circuit may be configured to couple to a sensor.

In another aspect, the present disclosure provides a surgical instrument, comprising: a battery assembly, comprising a control circuit comprising a battery, a memory coupled to the battery, and a processor coupled to the memory and the battery, wherein the processor is configured to generate a digital waveform; a handle assembly comprising a common first stage circuit coupled to the processor, the common first stage circuit comprising a digital-to-analog (DAC) converter and a common first stage amplifier circuit, wherein the DAC is configured to receive the digital waveform and convert the digital waveform into an analog waveform, wherein the common first stage amplifier circuit is configured to receive and amplify the analog waveform; and a shaft assembly comprising a second stage circuit coupled to the common first stage amplifier circuit to receive the analog waveform, amplify the analog waveform, and apply the analog waveform to a load; wherein the battery assembly and the shaft assembly are configured to mechanically and electrically connect to the handle assembly.

The load may comprise any one of an ultrasonic transducer, an electrode, or a sensor, or any combinations thereof. The common first stage circuit may be configured to drive ultrasonic, high-frequency current, or sensor circuits. The common first stage drive circuit may be configured to couple to a second stage ultrasonic drive circuit, a second stage high-frequency drive circuit, or a second stage sensor drive circuit. The second stage ultrasonic drive circuit may be configured to couple to an ultrasonic transducer, the second stage high-frequency drive circuit is configured to couple to an electrode, and the second stage sensor drive circuit is configured to couple to a sensor.

In another aspect, the present disclosure provides a surgical instrument, comprising a control circuit comprising a memory coupled to a processor, wherein the processor is configured to generate a digital waveform; a handle assembly comprising a common first stage circuit coupled to the processor, the common first stage circuit configured to receive the digital waveform, convert the digital waveform into an analog waveform, and amplify the analog waveform; and a shaft assembly comprising a second stage circuit coupled to the common first stage circuit to receive and amplify the analog waveform; wherein the shaft assembly is configured to mechanically and electrically connect to the handle assembly.

The common first stage circuit may be configured to drive ultrasonic, high-frequency current, or sensor circuits. The common first stage drive circuit may be configured to couple to a second stage ultrasonic drive circuit, a second stage high-frequency drive circuit, or a second stage sensor drive circuit. The second stage ultrasonic drive circuit may be configured to couple to an ultrasonic transducer, the second stage high-frequency drive circuit is configured to couple to

an electrode, and the second stage sensor drive circuit is configured to couple to a sensor.

FIG. 32 illustrates a generator circuit 3400 partitioned into a first stage circuit 3404 and a second stage circuit 3406, in accordance with at least one aspect of the present disclosure. In one aspect, the surgical instruments of surgical system 1000 described herein may comprise a generator circuit 3400 partitioned into multiple stages. For example, surgical instruments of surgical system 1000 may comprise the generator circuit 3400 partitioned into at least two circuits: the first stage circuit 3404 and the second stage circuit 3406 of amplification enabling operation of RF energy only, ultrasonic energy only, and/or a combination of RF energy and ultrasonic energy. A combination modular shaft assembly 3414 may be powered by the common first stage circuit 3404 located within a handle assembly 3412 and the modular second stage circuit 3406 integral to the modular shaft assembly 3414. As previously discussed throughout this description in connection with the surgical instruments of surgical system 1000, a battery assembly 3410 and the shaft assembly 3414 are configured to mechanically and electrically connect to the handle assembly 3412. The end effector assembly is configured to mechanically and electrically connect the shaft assembly 3414.

Turning now to FIG. 32, the generator circuit 3400 is partitioned into multiple stages located in multiple modular assemblies of a surgical instrument, such as the surgical instruments of surgical system 1000 described herein. In one aspect, a control stage circuit 3402 may be located in the battery assembly 3410 of the surgical instrument. The control stage circuit 3402 is a control circuit 3200 as described in connection with FIG. 30. The control circuit 3200 comprises a processor 3214, which includes internal memory 3217 (FIG. 32) (e.g., volatile and non-volatile memory), and is electrically coupled to a battery 3211. The battery 3211 supplies power to the first stage circuit 3404, the second stage circuit 3406, and a third stage circuit 3408, respectively. As previously discussed, the control circuit 3200 generates a digital waveform 4300 (FIG. 37) using circuits and techniques described in connection with FIGS. 35 and 36. Returning to FIG. 32, the digital waveform 4300 may be configured to drive an ultrasonic transducer, high-frequency (e.g., RF) electrodes, or a combination thereof either independently or simultaneously. If driven simultaneously, filter circuits may be provided in the corresponding first stage circuits 3404 to select either the ultrasonic waveform or the RF waveform. Such filtering techniques are described in commonly owned U.S. Patent Application Publication No. 2017/0086910, titled TECHNIQUES FOR CIRCUIT TOPOLOGIES FOR COMBINED GENERATOR, now U.S. Pat. No. 10,610,286, which is herein incorporated by reference in its entirety.

The first stage circuits 3404 (e.g., the first stage ultrasonic drive circuit 3420, the first stage RF drive circuit 3422, and the first stage sensor drive circuit 3424) are located in a handle assembly 3412 of the surgical instrument. The control circuit 3200 provides the ultrasonic drive signal to the first stage ultrasonic drive circuit 3420 via outputs SCL-A, SDA-A of the control circuit 3200. The first stage ultrasonic drive circuit 3420 is described in detail in connection with FIG. 29. The control circuit 3200 provides the RF drive signal to the first stage RF drive circuit 3422 via outputs SCL-B, SDA-B of the control circuit 3200. The first stage RF drive circuit 3422 is described in detail in connection with FIG. 34. The control circuit 3200 provides the sensor drive signal to the first stage sensor drive circuit 3424 via

outputs SCL-C, SDA-C of the control circuit 3200. Generally, each of the first stage circuits 3404 includes a digital-to-analog (DAC) converter and a first stage amplifier section to drive the second stage circuits 3406. The outputs of the first stage circuits 3404 are provided to the inputs of the second stage circuits 3406.

The control circuit 3200 is configured to detect which modules are plugged into the control circuit 3200. For example, the control circuit 3200 is configured to detect whether the first stage ultrasonic drive circuit 3420, the first stage RF drive circuit 3422, or the first stage sensor drive circuit 3424 located in the handle assembly 3412 is connected to the battery assembly 3410. Likewise, each of the first stage circuits 3404 can detect which second stage circuits 3406 are connected thereto and that information is provided back to the control circuit 3200 to determine the type of signal waveform to generate. Similarly, each of the second stage circuits 3406 can detect which third stage circuits 3408 or components are connected thereto and that information is provided back to the control circuit 3200 to determine the type of signal waveform to generate.

In one aspect, the second stage circuits 3406 (e.g., the ultrasonic drive second stage circuit 3430, the RF drive second stage circuit 3432, and the sensor drive second stage circuit 3434) are located in the shaft assembly 3414 of the surgical instrument. The first stage ultrasonic drive circuit 3420 provides a signal to the second stage ultrasonic drive circuit 3430 via outputs US-Left/US-Right. The second stage ultrasonic drive circuit 3430 can include, for example, a transformer, filter, amplifier, and/or signal conditioning circuits. The first stage high-frequency (RF) current drive circuit 3422 provides a signal to the second stage RF drive circuit 3432 via outputs RF-Left/RF-Right. In addition to a transformer and blocking capacitors, the second stage RF drive circuit 3432 also may include filter, amplifier, and signal conditioning circuits. The first stage sensor drive circuit 3424 provides a signal to the second stage sensor drive circuit 3434 via outputs Sensor-1/Sensor-2. The second stage sensor drive circuit 3434 may include filter, amplifier, and signal conditioning circuits depending on the type of sensor. The outputs of the second stage circuits 3406 are provided to the inputs of the third stage circuits 3408.

In one aspect, the third stage circuits 3408 (e.g., the ultrasonic transducer 1120, the RF electrodes 3074a, 3074b, and the sensors 3440) may be located in various assemblies 3416 of the surgical instruments. In one aspect, the second stage ultrasonic drive circuit 3430 provides a drive signal to the ultrasonic transducer 1120 piezoelectric stack. In one aspect, the ultrasonic transducer 1120 is located in the ultrasonic transducer assembly of the surgical instrument. In other aspects, however, the ultrasonic transducer 1120 may be located in the handle assembly 3412, the shaft assembly 3414, or the end effector. In one aspect, the second stage RF drive circuit 3432 provides a drive signal to the RF electrodes 3074a, 3074b, which are generally located in the end effector portion of the surgical instrument. In one aspect, the second stage sensor drive circuit 3434 provides a drive signal to various sensors 3440 located throughout the surgical instrument.

FIG. 33 illustrates a generator circuit 3500 partitioned into multiple stages where a first stage circuit 3504 is common to the second stage circuit 3506, in accordance with at least one aspect of the present disclosure. In one aspect, the surgical instruments of surgical system 1000 described herein may comprise generator circuit 3500 partitioned into multiple stages. For example, the surgical instruments of surgical system 1000 may comprise the generator circuit

3500 partitioned into at least two circuits: the first stage circuit 3504 and the second stage circuit 3506 of amplification enabling operation of high-frequency (RF) energy only, ultrasonic energy only, and/or a combination of RF energy and ultrasonic energy. A combination modular shaft assembly 3514 may be powered by a common first stage circuit 3504 located within the handle assembly 3512 and a modular second stage circuit 3506 integral to the modular shaft assembly 3514. As previously discussed throughout this description in connection with the surgical instruments of surgical system 1000, a battery assembly 3510 and the shaft assembly 3514 are configured to mechanically and electrically connect to the handle assembly 3512. The end effector assembly is configured to mechanically and electrically connect the shaft assembly 3514.

As shown in the example of FIG. 33, the battery assembly 3510 portion of the surgical instrument comprises a first control circuit 3502, which includes the control circuit 3200 previously described. The handle assembly 3512, which connects to the battery assembly 3510, comprises a common first stage drive circuit 3420. As previously discussed, the first stage drive circuit 3420 is configured to drive ultrasonic, high-frequency (RF) current, and sensor loads. The output of the common first stage drive circuit 3420 can drive any one of the second stage circuits 3506 such as the second stage ultrasonic drive circuit 3430, the second stage high-frequency (RF) current drive circuit 3432, and/or the second stage sensor drive circuit 3434. The common first stage drive circuit 3420 detects which second stage circuit 3506 is located in the shaft assembly 3514 when the shaft assembly 3514 is connected to the handle assembly 3512. Upon the shaft assembly 3514 being connected to the handle assembly 3512, the common first stage drive circuit 3420 determines which one of the second stage circuits 3506 (e.g., the second stage ultrasonic drive circuit 3430, the second stage RF drive circuit 3432, and/or the second stage sensor drive circuit 3434) is located in the shaft assembly 3514. The information is provided to the control circuit 3200 located in the handle assembly 3512 in order to supply a suitable digital waveform 4300 (FIG. 37) to the second stage circuit 3506 to drive the appropriate load, e.g., ultrasonic, RF, or sensor. It will be appreciated that identification circuits may be included in various assemblies 3516 in third stage circuit 3508 such as the ultrasonic transducer 1120, the electrodes 3074a, 3074b, or the sensors 3440. Thus, when a third stage circuit 3508 is connected to a second stage circuit 3506, the second stage circuit 3506 knows the type of load that is required based on the identification information.

FIG. 34 is a schematic diagram of one aspect of an electrical circuit 3600 configured to drive a high-frequency current (RF), in accordance with at least one aspect of the present disclosure. The electrical circuit 3600 comprises an analog multiplexer 3680. The analog multiplexer 3680 multiplexes various signals from the upstream channels SCL-A, SDA-A such as RF, battery, and power control circuit. A current sensor 3682 is coupled in series with the return or ground leg of the power supply circuit to measure the current supplied by the power supply. A field effect transistor (FET) temperature sensor 3684 provides the ambient temperature. A pulse width modulation (PWM) watchdog timer 3688 automatically generates a system reset if the main program neglects to periodically service it. It is provided to automatically reset the electrical circuit 3600 when it hangs or freezes because of a software or hardware fault. It will be appreciated that the electrical circuit 3600 may be configured for driving RF electrodes or for driving the ultrasonic transducer 1120 as described in connection with FIG. 29, for

example. Accordingly, with reference now back to FIG. 34, the electrical circuit 3600 can be used to drive both ultrasonic and RF electrodes interchangeably.

A drive circuit 3686 provides Left and Right RF energy outputs. A digital signal that represents the signal waveform is provided to the SCL-A, SDA-A inputs of the analog multiplexer 3680 from a control circuit, such as the control circuit 3200 (FIG. 30). A digital-to-analog converter 3690 (DAC) converts the digital input to an analog output to drive a PWM circuit 3692 coupled to an oscillator 3694. The PWM circuit 3692 provides a first signal to a first gate drive circuit 3696a coupled to a first transistor output stage 3698a to drive a first RF+ (Left) energy output. The PWM circuit 3692 also provides a second signal to a second gate drive circuit 3696b coupled to a second transistor output stage 3698b to drive a second RF- (Right) energy output. A voltage sensor 3699 is coupled between the RF Left/RF output terminals to measure the output voltage. The drive circuit 3686, the first and second drive circuits 3696a, 3696b, and the first and second transistor output stages 3698a, 3698b define a first stage amplifier circuit. In operation, the control circuit 3200 (FIG. 30) generates a digital waveform 4300 (FIG. 37) employing circuits such as direct digital synthesis (DDS) circuits 4100, 4200 (FIGS. 35 and 36). The DAC 3690 receives the digital waveform 4300 and converts it into an analog waveform, which is received and amplified by the first stage amplifier circuit.

In one aspect, the ultrasonic or high-frequency current generators of the surgical system 1000 may be configured to generate the electrical signal waveform digitally such that the desired using a predetermined number of phase points stored in a lookup table to digitize the wave shape. The phase points may be stored in a table defined in a memory, a field programmable gate array (FPGA), or any suitable non-volatile memory. FIG. 35 illustrates one aspect of a fundamental architecture for a digital synthesis circuit such as a direct digital synthesis (DDS) circuit 4100 configured to generate a plurality of wave shapes for the electrical signal waveform. The generator software and digital controls may command the FPGA to scan the addresses in the lookup table 4104 which in turn provides varying digital input values to a DAC circuit 4108 that feeds a power amplifier. The addresses may be scanned according to a frequency of interest. Using such a lookup table 4104 enables generating various types of wave shapes that can be fed into tissue or into a transducer, an RF electrode, multiple transducers simultaneously, multiple RF electrodes simultaneously, or a combination of RF and ultrasonic instruments. Furthermore, multiple lookup tables 4104 that represent multiple wave shapes can be created, stored, and applied to tissue from a generator.

The waveform signal may be configured to control at least one of an output current, an output voltage, or an output power of an ultrasonic transducer and/or an RF electrode, or multiples thereof (e.g. two or more ultrasonic transducers and/or two or more RF electrodes). Further, where the surgical instrument comprises an ultrasonic components, the waveform signal may be configured to drive at least two vibration modes of an ultrasonic transducer of the at least one surgical instrument. Accordingly, a generator may be configured to provide a waveform signal to at least one surgical instrument wherein the waveform signal corresponds to at least one wave shape of a plurality of wave shapes in a table. Further, the waveform signal provided to the two surgical instruments may comprise two or more wave shapes. The table may comprise information associated with a plurality of wave shapes and the table may be

stored within the generator. In one aspect or example, the table may be a direct digital synthesis table, which may be stored in an FPGA of the generator. The table may be addressed by anyway that is convenient for categorizing wave shapes. According to one aspect, the table, which may be a direct digital synthesis table, is addressed according to a frequency of the waveform signal. Additionally, the information associated with the plurality of wave shapes may be stored as digital information in the table.

The analog electrical signal waveform may be configured to control at least one of an output current, an output voltage, or an output power of an ultrasonic transducer and/or an RF electrode, or multiples thereof (e.g., two or more ultrasonic transducers and/or two or more RF electrodes). Further, where the surgical instrument comprises ultrasonic components, the analog electrical signal waveform may be configured to drive at least two vibration modes of an ultrasonic transducer of the at least one surgical instrument. Accordingly, the generator circuit may be configured to provide an analog electrical signal waveform to at least one surgical instrument wherein the analog electrical signal waveform corresponds to at least one wave shape of a plurality of wave shapes stored in a lookup table 4104. Further, the analog electrical signal waveform provided to the two surgical instruments may comprise two or more wave shapes. The lookup table 4104 may comprise information associated with a plurality of wave shapes and the lookup table 4104 may be stored either within the generator circuit or the surgical instrument. In one aspect or example, the lookup table 4104 may be a direct digital synthesis table, which may be stored in an FPGA of the generator circuit or the surgical instrument. The lookup table 4104 may be addressed by anyway that is convenient for categorizing wave shapes. According to one aspect, the lookup table 4104, which may be a direct digital synthesis table, is addressed according to a frequency of the desired analog electrical signal waveform. Additionally, the information associated with the plurality of wave shapes may be stored as digital information in the lookup table 4104.

With the widespread use of digital techniques in instrumentation and communications systems, a digitally-controlled method of generating multiple frequencies from a reference frequency source has evolved and is referred to as direct digital synthesis. The basic architecture is shown in FIG. 35. In this simplified block diagram, a DDS circuit is coupled to a processor, controller, or a logic device of the generator circuit and to a memory circuit located in the generator circuit of the surgical system 1000. The DDS circuit 4100 comprises an address counter 4102, lookup table 4104, a register 4106, a DAC circuit 4108, and a filter 4112. A stable clock f , is received by the address counter 4102 and the register 4106 drives a programmable-read-only-memory (PROM) which stores one or more integral number of cycles of a sinewave (or other arbitrary waveform) in a lookup table 4104. As the address counter 4102 steps through memory locations, values stored in the lookup table 4104 are written to the register 4106, which is coupled to the DAC circuit 4108. The corresponding digital amplitude of the signal at the memory location of the lookup table 4104 drives the DAC circuit 4108, which in turn generates an analog output signal 4110. The spectral purity of the analog output signal 4110 is determined primarily by the DAC circuit 4108. The phase noise is basically that of the reference clock f . The first analog output signal 4110 output from the DAC circuit 4108 is filtered by the filter 4112 and a second analog output signal 4114 output by the filter 4112

is provided to an amplifier having an output coupled to the output of the generator circuit. The second analog output signal has a frequency f_{out} .

Because the DDS circuit 4100 is a sampled data system, issues involved in sampling must be considered: quantization noise, aliasing, filtering, etc. For instance, the higher order harmonics of the DAC circuit 4108 output frequencies fold back into the Nyquist bandwidth, making them unfilterable, whereas, the higher order harmonics of the output of phase-locked-loop (PLL) based synthesizers can be filtered. The lookup table 4104 contains signal data for an integral number of cycles. The final output frequency f_{out} can be changed changing the reference clock frequency f , or by reprogramming the PROM.

The DDS circuit 4100 may comprise multiple lookup tables 4104 where the lookup table 4104 stores a waveform represented by a predetermined number of samples, wherein the samples define a predetermined shape of the waveform. Thus multiple waveforms having a unique shape can be stored in multiple lookup tables 4104 to provide different tissue treatments based on instrument settings or tissue feedback. Examples of waveforms include high crest factor RF electrical signal waveforms for surface tissue coagulation, low crest factor RF electrical signal waveform for deeper tissue penetration, and electrical signal waveforms that promote efficient touch-up coagulation. In one aspect, the DDS circuit 4100 can create multiple wave shape lookup tables 4104 and during a tissue treatment procedure (e.g., “on-the-fly” or in virtual real time based on user or sensor inputs) switch between different wave shapes stored in separate lookup tables 4104 based on the tissue effect desired and/or tissue feedback.

Accordingly, switching between wave shapes can be based on tissue impedance and other factors, for example. In other aspects, the lookup tables 4104 can store electrical signal waveforms shaped to maximize the power delivered into the tissue per cycle (i.e., trapezoidal or square wave). In other aspects, the lookup tables 4104 can store wave shapes synchronized in such way that they make maximizing power delivery by the multifunction surgical instrument of surgical system 1000 while delivering RF and ultrasonic drive signals. In yet other aspects, the lookup tables 4104 can store electrical signal waveforms to drive ultrasonic and RF therapeutic, and/or sub-therapeutic, energy simultaneously while maintaining ultrasonic frequency lock. Custom wave shapes specific to different instruments and their tissue effects can be stored in the non-volatile memory of the generator circuit or in the non-volatile memory (e.g., EEPROM) of the surgical system 1000 and be fetched upon connecting the multifunction surgical instrument to the generator circuit. An example of an exponentially damped sinusoid, as used in many high crest factor “coagulation” waveforms is shown in FIG. 37.

Amore flexible and efficient implementation of the DDS circuit 4100 employs a digital circuit called a Numerically Controlled Oscillator (NCO). A block diagram of a more flexible and efficient digital synthesis circuit such as a DDS circuit 4200 is shown in FIG. 36. In this simplified block diagram, a DDS circuit 4200 is coupled to a processor, controller, or a logic device of the generator and to a memory circuit located either in the generator or in any of the surgical instruments of surgical system 1000. The DDS circuit 4200 comprises a load register 4202, a parallel delta phase register 4204, an adder circuit 4216, a phase register 4208, a lookup table 4210 (phase-to-amplitude converter), a DAC circuit 4212, and a filter 4214. The adder circuit 4216 and the phase register 4208 form part of a phase accumulator

4206. A clock frequency f , is applied to the phase register 4208 and a DAC circuit 4212. The load register 4202 receives a tuning word that specifies output frequency as a fraction of the reference clock frequency signal f_c . The output of the load register 4202 is provided to the parallel delta phase register 4204 with a tuning word M .

The DDS circuit 4200 includes a sample clock that generates the clock frequency f , the phase accumulator 4206, and the lookup table 4210 (e.g., phase to amplitude converter). The content of the phase accumulator 4206 is updated once per clock cycle f . When time the phase accumulator 4206 is updated, the digital number, M , stored in the parallel delta phase register 4204 is added to the number in the phase register 4208 by the adder circuit 4216. Assuming that the number in the parallel delta phase register 4204 is 00 . . . 01 and that the initial contents of the phase accumulator 4206 is 00 . . . 00. The phase accumulator 4206 is updated by 00 . . . 01 per clock cycle. If the phase accumulator 4206 is 32-bits wide, 232 clock cycles (over 4 billion) are required before the phase accumulator 4206 returns to 00 . . . 00, and the cycle repeats.

A truncated output 4218 of the phase accumulator 4206 is provided to a phase-to amplitude converter lookup table 4210 and the output of the lookup table 4210 is coupled to a DAC circuit 4212. The truncated output 4218 of the phase accumulator 4206 serves as the address to a sine (or cosine) lookup table. An address in the lookup table corresponds to a phase point on the sinewave from 0° to 360°. The lookup table 4210 contains the corresponding digital amplitude information for one complete cycle of a sinewave. The lookup table 4210 therefore maps the phase information from the phase accumulator 4206 into a digital amplitude word, which in turn drives the DAC circuit 4212. The output of the DAC circuit is a first analog signal 4220 and is filtered by a filter 4214. The output of the filter 4214 is a second analog signal 4222, which is provided to a power amplifier coupled to the output of the generator circuit.

In one aspect, the electrical signal waveform may be digitized into 1024 (210) phase points, although the wave shape may be digitized is any suitable number of $2n$ phase points ranging from 256 (28) to 281,474,976,710,656 (248), where n is a positive integer, as shown in TABLE 1. The electrical signal waveform may be expressed as $A_n(\theta_n)$, where a normalized amplitude A_n , at a point n is represented by a phase angle θ_n , is referred to as a phase point at point n . The number of discrete phase points n determines the tuning resolution of the DDS circuit 4200 (as well as the DDS circuit 4100 shown in FIG. 35).

TABLE 1 specifies the electrical signal waveform digitized into a number of phase points.

TABLE 1

N	Number of Phase Points 2 ⁿ
8	256
10	1, 024
12	4, 096
14	16, 384
16	65, 536
18	262, 144
20	1, 048, 576
22	4, 194, 304
24	16, 777, 216

TABLE 1-continued

N	Number of Phase Points 2 ⁿ
26	67, 108, 864
28	268, 435, 456
...	...
32	4, 294, 967, 296
...	...
48	281, 474, 976, 710, 656
...	...

The generator circuit algorithms and digital control circuits scan the addresses in the lookup table 4210, which in turn provides varying digital input values to the DAC circuit 4212 that feeds the filter 4214 and the power amplifier. The addresses may be scanned according to a frequency of interest. Using the lookup table enables generating various types of shapes that can be converted into an analog output signal by the DAC circuit 4212, filtered by the filter 4214, amplified by the power amplifier coupled to the output of the generator circuit, and fed to the tissue in the form of RF energy or fed to an ultrasonic transducer and applied to the tissue in the form of ultrasonic vibrations which deliver energy to the tissue in the form of heat. The output of the amplifier can be applied to an RF electrode, multiple RF electrodes simultaneously, an ultrasonic transducer, multiple ultrasonic transducers simultaneously, or a combination of RF and ultrasonic transducers, for example. Furthermore, multiple wave shape tables can be created, stored, and applied to tissue from a generator circuit.

With reference back to FIG. 35, for n=32, and M=1, the phase accumulator 4206 steps through 232 possible outputs before it overflows and restarts. The corresponding output wave frequency is equal to the input clock frequency divided by 232. If M=2, then the phase register 1708 “rolls over” twice as fast, and the output frequency is doubled. This can be generalized as follows.

For a phase accumulator 4206 configured to accumulate n-bits (n generally ranges from 24 to 32 in most DDS systems, but as previously discussed n may be selected from a wide range of options), there are 2ⁿ possible phase points. The digital word in the delta phase register, M, represents the amount the phase accumulator is incremented per clock cycle. If f_c is the clock frequency, then the frequency of the output sinewave is equal to:

$$f_0 = \frac{M \cdot f_c}{2^n}$$

The above equation is known as the DDS “tuning equation.” Note that the frequency resolution of the system is equal to

$$\frac{f_0}{2^n}$$

For n=32, the resolution is greater than one part in four billion. In one aspect of the DDS circuit 4200, not all of the bits out of the phase accumulator 4206 are passed on to the lookup table 4210, but are truncated, leaving only the first 13 to 15 most significant bits (MSBs), for example. This reduces the size of the lookup table 4210 and does not affect the frequency resolution. The phase truncation only adds a small but acceptable amount of phase noise to the final output.

The electrical signal waveform may be characterized by a current, voltage, or power at a predetermined frequency. Further, where any one of the surgical instruments of surgical system 1000 comprises ultrasonic components, the electrical signal waveform may be configured to drive at least two vibration modes of an ultrasonic transducer of the at least one surgical instrument. Accordingly, the generator circuit may be configured to provide an electrical signal waveform to at least one surgical instrument wherein the electrical signal waveform is characterized by a predetermined wave shape stored in the lookup table 4210 (or lookup table 4104, FIG. 35). Further, the electrical signal waveform may be a combination of two or more wave shapes. The lookup table 4210 may comprise information associated with a plurality of wave shapes. In one aspect or example, the lookup table 4210 may be generated by the DDS circuit 4200 and may be referred to as a direct digital synthesis table. DDS works by first storing a large repetitive waveform in onboard memory. A cycle of a waveform (sine, triangle, square, arbitrary) can be represented by a predetermined number of phase points as shown in TABLE 1 and stored into memory. Once the waveform is stored into memory, it can be generated at very precise frequencies. The direct digital synthesis table may be stored in a non-volatile memory of the generator circuit and/or may be implemented with a FPGA circuit in the generator circuit. The lookup table 4210 may be addressed by any suitable technique that is convenient for categorizing wave shapes. According to one aspect, the lookup table 4210 is addressed according to a frequency of the electrical signal waveform. Additionally, the information associated with the plurality of wave shapes may be stored as digital information in a memory or as part of the lookup table 4210.

In one aspect, the generator circuit may be configured to provide electrical signal waveforms to at least two surgical instruments simultaneously. The generator circuit also may be configured to provide the electrical signal waveform, which may be characterized two or more wave shapes, via an output channel of the generator circuit to the two surgical instruments simultaneously. For example, in one aspect the electrical signal waveform comprises a first electrical signal to drive an ultrasonic transducer (e.g., ultrasonic drive signal), a second RF drive signal, and/or a combination thereof. In addition, an electrical signal waveform may comprise a plurality of ultrasonic drive signals, a plurality of RF drive signals, and/or a combination of a plurality of ultrasonic and RF drive signals.

In addition, a method of operating the generator circuit according to the present disclosure comprises generating an electrical signal waveform and providing the generated electrical signal waveform to any one of the surgical instruments of surgical system 1000, where generating the electrical signal waveform comprises receiving information associated with the electrical signal waveform from a memory. The generated electrical signal waveform comprises at least one wave shape. Furthermore, providing the generated electrical signal waveform to the at least one surgical instrument comprises providing the electrical signal waveform to at least two surgical instruments simultaneously.

The generator circuit as described herein may allow for the generation of various types of direct digital synthesis tables. Examples of wave shapes for RF/Electrosurgery signals suitable for treating a variety of tissue generated by the generator circuit include RF signals with a high crest factor (which may be used for surface coagulation in RF mode), a low crest factor RF signals (which may be used for

deeper tissue penetration), and waveforms that promote efficient touch-up coagulation. The generator circuit also may generate multiple wave shapes employing a direct digital synthesis lookup table **4210** and, on the fly, can switch between particular wave shapes based on the desired tissue effect. Switching may be based on tissue impedance and/or other factors.

In addition to traditional sine/cosine wave shapes, the generator circuit may be configured to generate wave shape(s) that maximize the power into tissue per cycle (i.e., trapezoidal or square wave). The generator circuit may provide wave shape(s) that are synchronized to maximize the power delivered to the load when driving RF and ultrasonic signals simultaneously and to maintain ultrasonic frequency lock, provided that the generator circuit includes a circuit topology that enables simultaneously driving RF and ultrasonic signals. Further, custom wave shapes specific to instruments and their tissue effects can be stored in a non-volatile memory (NVM) or an instrument EEPROM and can be fetched upon connecting any one of the surgical instruments of surgical system **1000** to the generator circuit.

The DDS circuit **4200** may comprise multiple lookup tables **4104** where the lookup table **4210** stores a waveform represented by a predetermined number of phase points (also may be referred to as samples), wherein the phase points define a predetermined shape of the waveform. Thus multiple waveforms having a unique shape can be stored in multiple lookup tables **4210** to provide different tissue treatments based on instrument settings or tissue feedback. Examples of waveforms include high crest factor RF electrical signal waveforms for surface tissue coagulation, low crest factor RF electrical signal waveform for deeper tissue penetration, and electrical signal waveforms that promote efficient touch-up coagulation. In one aspect, the DDS circuit **4200** can create multiple wave shape lookup tables **4210** and during a tissue treatment procedure (e.g., “on-the-fly” or in virtual real time based on user or sensor inputs) switch between different wave shapes stored in different lookup tables **4210** based on the tissue effect desired and/or tissue feedback.

Accordingly, switching between wave shapes can be based on tissue impedance and other factors, for example. In other aspects, the lookup tables **4210** can store electrical signal waveforms shaped to maximize the power delivered into the tissue per cycle (i.e., trapezoidal or square wave). In other aspects, the lookup tables **4210** can store wave shapes synchronized in such way that they make maximizing power delivery by any one of the surgical instruments of surgical system **1000** when delivering RF and ultrasonic drive signals. In yet other aspects, the lookup tables **4210** can store electrical signal waveforms to drive ultrasonic and RF therapeutic, and/or sub-therapeutic, energy simultaneously while maintaining ultrasonic frequency lock. Generally, the output wave shape may be in the form of a sine wave, cosine wave, pulse wave, square wave, and the like. Nevertheless, the more complex and custom wave shapes specific to different instruments and their tissue effects can be stored in the non-volatile memory of the generator circuit or in the non-volatile memory (e.g., EEPROM) of the surgical instrument and be fetched upon connecting the surgical instrument to the generator circuit. One example of a custom wave shape is an exponentially damped sinusoid as used in many high crest factor “coagulation” waveforms, as shown in FIG. **37**.

FIG. **37** illustrates one cycle of a discrete time digital electrical signal waveform **4300**, in accordance with at least one aspect of the present disclosure of an analog waveform

4304 (shown superimposed over the discrete time digital electrical signal waveform **4300** for comparison purposes). The horizontal axis represents Time (t) and the vertical axis represents digital phase points. The digital electrical signal waveform **4300** is a digital discrete time version of the desired analog waveform **4304**, for example. The digital electrical signal waveform **4300** is generated by storing an amplitude phase point **4302** that represents the amplitude per clock cycle T_{clk} over one cycle or period T_o . The digital electrical signal waveform **4300** is generated over one period T_o by any suitable digital processing circuit. The amplitude phase points are digital words stored in a memory circuit. In the example illustrated in FIGS. **35** and **36**, the digital word is a six-bit word that is capable of storing the amplitude phase points with a resolution of 26 or 64 bits. It will be appreciated that the examples shown in FIGS. **35** and **36** is for illustrative purposes and in actual implementations the resolution can be much higher. The digital amplitude phase points **4302** over one cycle T_o are stored in the memory as a string of string words in a lookup table **4104**, **4210** as described in connection with FIGS. **35** and **36**, for example. To generate the analog version of the analog waveform **4304**, the amplitude phase points **4302** are read sequentially from the memory from 0 to T_o per clock cycle T_{clk} and are converted by a DAC circuit **4108**, **4212**, also described in connection with FIGS. **35** and **36**. Additional cycles can be generated by repeatedly reading the amplitude phase points **4302** of the digital electrical signal waveform **4300** the from 0 to T_o for as many cycles or periods as may be desired. The smooth analog version of the analog waveform **4304** is achieved by filtering the output of the DAC circuit **4108**, **4212** by a filter **4112**, **4214** (FIGS. **35** and **36**). The filtered analog output signal **4114**, **4222** (FIGS. **35** and **36**) is applied to the input of a power amplifier.

FIG. **38** is a diagram of a control system **12950** configured to provide progressive closure of a closure member (e.g., closure tube) when the displacement member advances distally and couples into a clamp arm (e.g., anvil) to lower the closure force load on the closure member at a desired rate and decrease the firing force load on the firing member according to one aspect of this disclosure. In one aspect, the control system **12950** may be implemented as a nested PID feedback controller. A PID controller is a control loop feedback mechanism (controller) to continuously calculate an error value as the difference between a desired set point and a measured process variable and applies a correction based on proportional, integral, and derivative terms (sometimes denoted P, I, and D respectively). The nested PID controller feedback control system **12950** includes a primary controller **12952**, in a primary (outer) feedback loop **12954** and a secondary controller **12955** in a secondary (inner) feedback loop **12956**. The primary controller **12952** may be a PID controller **12972** as shown in FIG. **39**, and the secondary controller **12955** also may be a PID controller **12972** as shown in FIG. **39**. The primary controller **12952** controls a primary process **12958** and the secondary controller **12955** controls a secondary process **12960**. The output **12966** of the primary process **12958** is subtracted from a primary set point SP_1 by a first summer **12962**. The first summer **12962** produces a single sum output signal which is applied to the primary controller **12952**. The output of the primary controller **12952** is the secondary set point SP_2 . The output **12968** of the secondary process **12960** is subtracted from the secondary set point SP_2 by a second summer **12964**.

In the context of controlling the displacement of a closure tube, the control system **12950** may be configured such that

the primary set point SP_1 is a desired closure force value and the primary controller **12952** is configured to receive the closure force from a torque sensor coupled to the output of a closure motor and determine a set point SP_2 motor velocity for the closure motor. In other aspects, the closure force may be measured with strain gauges, load cells, or other suitable force sensors. The closure motor velocity set point SP_2 is compared to the actual velocity of the closure tube, which is determined by the secondary controller **12955**. The actual velocity of the closure tube may be measured by comparing measuring the displacement of the closure tube with the position sensor and measuring elapsed time with a timer/counter. Other techniques, such as linear or rotary encoders may be employed to measure displacement of the closure tube. The output **12968** of the secondary process **12960** is the actual velocity of the closure tube. This closure tube velocity output **12968** is provided to the primary process **12958** which determines the force acting on the closure tube and is fed back to the adder **12962**, which subtracts the measured closure force from the primary set point SP_1 . The primary set point SP_1 may be an upper threshold or a lower threshold. Based on the output of the adder **12962**, the primary controller **12952** controls the velocity and direction of the closure motor. The secondary controller **12955** controls the velocity of the closure motor based on the actual velocity of closure tube measured by the secondary process **12960** and the secondary set point SP_2 , which is based on a comparison of the actual firing force and the firing force upper and lower thresholds.

FIG. **39** illustrates a PID feedback control system **12970** according to one aspect of this disclosure. The primary controller **12952** or the secondary controller **12955**, or both, may be implemented as a PID controller **12972**. In one aspect, the PID controller **12972** may comprise a proportional element **12974** (P), an integral element **12976** (I), and a derivative element **12978** (D). The outputs of the P, I, D elements **12974**, **12976**, **12978** are summed by a summer **12986**, which provides the control variable $\mu(t)$ to the process **12980**. The output of the process **12980** is the process variable $y(t)$. A summer **12984** calculates the difference between a desired set point $r(t)$ and a measured process variable $y(t)$. The PID controller **12972** continuously calculates an error value $e(t)$ (e.g., difference between closure force threshold and measured closure force) as the difference between a desired set point $r(t)$ (e.g., closure force threshold) and a measured process variable $y(t)$ (e.g., velocity and direction of closure tube) and applies a correction based on the proportional, integral, and derivative terms calculated by the proportional element **12974** (P), integral element **12976** (I), and derivative element **12978** (D), respectively. The PID controller **12972** attempts to minimize the error $e(t)$ over time by adjustment of the control variable $\mu(t)$ (e.g., velocity and direction of the closure tube).

In accordance with the PID algorithm, the “P” element **12974** accounts for present values of the error. For example, if the error is large and positive, the control output will also be large and positive. In accordance with the present disclosure, the error term $e(t)$ is the different between the desired closure force and the measured closure force of the closure tube. The “I” element **12976** accounts for past values of the error. For example, if the current output is not sufficiently strong, the integral of the error will accumulate over time, and the controller will respond by applying a stronger action. The “D” element **12978** accounts for possible future trends of the error, based on its current rate of change. For example, continuing the P example above, when the large positive control output succeeds in bringing the

error closer to zero, it also puts the process on a path to large negative error in the near future. In this case, the derivative turns negative and the D module reduces the strength of the action to prevent this overshoot.

It will be appreciated that other variables and set points may be monitored and controlled in accordance with the feedback control systems **12950**, **12970**. For example, the adaptive closure member velocity control algorithm described herein may measure at least two of the following parameters: firing member stroke location, firing member load, displacement of cutting element, velocity of cutting element, closure tube stroke location, closure tube load, among others.

Ultrasonic surgical devices, such as ultrasonic scalpels, are finding increasingly widespread applications in surgical procedures by virtue of their unique performance characteristics. Depending upon specific device configurations and operational parameters, ultrasonic surgical devices can provide substantially simultaneous transection of tissue and homeostasis by coagulation, desirably minimizing patient trauma. An ultrasonic surgical device may comprise a handpiece containing an ultrasonic transducer, and an instrument coupled to the ultrasonic transducer having a distally-mounted end effector (e.g., a blade tip) to cut and seal tissue. In some cases, the instrument may be permanently affixed to the handpiece. In other cases, the instrument may be detachable from the handpiece, as in the case of a disposable instrument or an interchangeable instrument. The end effector transmits ultrasonic energy to tissue brought into contact with the end effector to realize cutting and sealing action. Ultrasonic surgical devices of this nature can be configured for open surgical use, laparoscopic, or endoscopic surgical procedures including robotic-assisted procedures.

Ultrasonic energy cuts and coagulates tissue using temperatures lower than those used in electrosurgical procedures and can be transmitted to the end effector by an ultrasonic generator in communication with the handpiece. Vibrating at high frequencies (e.g., 55,500 cycles per second), the ultrasonic blade denatures protein in the tissue to form a sticky coagulum. Pressure exerted on tissue by the blade surface collapses blood vessels and allows the coagulum to form a hemostatic seal. A surgeon can control the cutting speed and coagulation by the force applied to the tissue by the end effector, the time over which the force is applied, and the selected excursion level of the end effector.

The ultrasonic transducer may be modeled as an equivalent circuit comprising a first branch having a static capacitance and a second “motional” branch having a serially connected inductance, resistance and capacitance that define the electromechanical properties of a resonator. Known ultrasonic generators may include a tuning inductor for tuning out the static capacitance at a resonant frequency so that substantially all of a generator’s drive signal current flows into the motional branch. Accordingly, by using a tuning inductor, the generator’s drive signal current represents the motional branch current, and the generator is thus able to control its drive signal to maintain the ultrasonic transducer’s resonant frequency. The tuning inductor may also transform the phase impedance plot of the ultrasonic transducer to improve the generator’s frequency lock capabilities. However, the tuning inductor must be matched with the specific static capacitance of an ultrasonic transducer at the operational resonant frequency. In other words, a different ultrasonic transducer having a different static capacitance requires a different tuning inductor.

Additionally, in some ultrasonic generator architectures, the generator’s drive signal exhibits asymmetrical harmonic

distortion that complicates impedance magnitude and phase measurements. For example, the accuracy of impedance phase measurements may be reduced due to harmonic distortion in the current and voltage signals.

Moreover, electromagnetic interference in noisy environments decreases the ability of the generator to maintain lock on the ultrasonic transducer's resonant frequency, increasing the likelihood of invalid control algorithm inputs.

Electrosurgical devices for applying electrical energy to tissue in order to treat and/or destroy the tissue are also finding increasingly widespread applications in surgical procedures. An electrosurgical device may comprise a handpiece and an instrument having a distally-mounted end effector (e.g., one or more electrodes). The end effector can be positioned against the tissue such that electrical current is introduced into the tissue. Electrosurgical devices can be configured for bipolar or monopolar operation. During bipolar operation, current is introduced into and returned from the tissue by active and return electrodes, respectively, of the end effector. During monopolar operation, current is introduced into the tissue by an active electrode of the end effector and returned through a return electrode (e.g., a grounding pad) separately located on a patient's body. Heat generated by the current flowing through the tissue may form hemostatic seals within the tissue and/or between tissues and thus may be particularly useful for sealing blood vessels, for example. The end effector of an electrosurgical device may also comprise a cutting member that is movable relative to the tissue and the electrodes to transect the tissue.

Electrical energy applied by an electrosurgical device can be transmitted to the instrument by a generator in communication with the handpiece. The electrical energy may be in the form of radio frequency (RF) energy. RF energy is a form of electrical energy that may be in the frequency range of 300 kHz to 1 MHz, as described in EN60601-2-2:2009+A11:2011, Definition 201.3.218—HIGH FREQUENCY. For example, the frequencies in monopolar RF applications are typically restricted to less than 5 MHz. However, in bipolar RF applications, the frequency can be almost any value. Frequencies above 200 kHz are typically used for monopolar applications in order to avoid the unwanted stimulation of nerves and muscles which would result from the use of low frequency current. Lower frequencies may be used for bipolar techniques if a risk analysis shows the possibility of neuromuscular stimulation has been mitigated to an acceptable level. Normally, frequencies above 5 MHz are not used in order to minimize the problems associated with high frequency leakage currents. It is generally recognized that 10 mA is the lower threshold of thermal effects on tissue.

During its operation, an electrosurgical device can transmit low frequency RF energy through tissue, which causes ionic agitation, or friction, in effect resistive heating, thereby increasing the temperature of the tissue. Because a sharp boundary may be created between the affected tissue and the surrounding tissue, surgeons can operate with a high level of precision and control, without sacrificing un-targeted adjacent tissue. The low operating temperatures of RF energy may be useful for removing, shrinking, or sculpting soft tissue while simultaneously sealing blood vessels. RF energy may work particularly well on connective tissue, which is primarily comprised of collagen and shrinks when contacted by heat.

Due to their unique drive signal, sensing and feedback needs, ultrasonic and electrosurgical devices have generally required different generators. Additionally, in cases where the instrument is disposable or interchangeable with a hand-

piece, ultrasonic and electrosurgical generators are limited in their ability to recognize the particular instrument configuration being used and to optimize control and diagnostic processes accordingly. Moreover, capacitive coupling between the non-isolated and patient-isolated circuits of the generator, especially in cases where higher voltages and frequencies are used, may result in exposure of a patient to unacceptable levels of leakage current.

Furthermore, due to their unique drive signal, sensing and feedback needs, ultrasonic and electrosurgical devices have generally required different user interfaces for the different generators. In such conventional ultrasonic and electrosurgical devices, one user interface is configured for use with an ultrasonic instrument whereas a different user interface may be configured for use with an electrosurgical instrument. Such user interfaces include hand and/or foot activated user interfaces such as hand activated switches and/or foot activated switches. As various aspects of combined generators for use with both ultrasonic and electrosurgical instruments are contemplated in the subsequent disclosure, additional user interfaces that are configured to operate with both ultrasonic and/or electrosurgical instrument generators also are contemplated.

Additional user interfaces for providing feedback, whether to the user or other machine, are contemplated within the subsequent disclosure to provide feedback indicating an operating mode or status of either an ultrasonic and/or electrosurgical instrument. Providing user and/or machine feedback for operating a combination ultrasonic and/or electrosurgical instrument will require providing sensory feedback to a user and electrical/mechanical/electromechanical feedback to a machine. Feedback devices that incorporate visual feedback devices (e.g., an LCD display screen, LED indicators), audio feedback devices (e.g., a speaker, a buzzer) or tactile feedback devices (e.g., haptic actuators) for use in combined ultrasonic and/or electrosurgical instruments are contemplated in the subsequent disclosure.

Other electrical surgical instruments include, without limitation, irreversible and/or reversible electroporation, and/or microwave technologies, among others. Accordingly, the techniques disclosed herein are applicable to ultrasonic, bipolar or monopolar RF (electrosurgical), irreversible and/or reversible electroporation, and/or microwave based surgical instruments, among others.

Various aspects are directed to improved ultrasonic surgical devices, electrosurgical devices and generators for use therewith. Aspects of the ultrasonic surgical devices can be configured for transecting and/or coagulating tissue during surgical procedures, for example. Aspects of the electrosurgical devices can be configured for transecting, coagulating, scaling, welding and/or desiccating tissue during surgical procedures, for example.

Aspects of the generator utilize high-speed analog-to-digital sampling (e.g., approximately 200x oversampling, depending on frequency) of the generator drive signal current and voltage, along with digital signal processing, to provide a number of advantages and benefits over known generator architectures. In one aspect, for example, based on current and voltage feedback data, a value of the ultrasonic transducer static capacitance, and a value of the drive signal frequency, the generator may determine the motional branch current of an ultrasonic transducer. This provides the benefit of a virtually tuned system, and simulates the presence of a system that is tuned or resonant with any value of the static capacitance (e.g., C_0 in FIG. 25) at any frequency. Accordingly, control of the motional branch current may be realized

by tuning out the effects of the static capacitance without the need for a tuning inductor. Additionally, the elimination of the tuning inductor may not degrade the generator's frequency lock capabilities, as frequency lock can be realized by suitably processing the current and voltage feedback data.

High-speed analog-to-digital sampling of the generator drive signal current and voltage, along with digital signal processing, may also enable precise digital filtering of the samples. For example, aspects of the generator may utilize a low-pass digital filter (e.g., a finite impulse response (FIR) filter) that rolls off between a fundamental drive signal frequency and a second-order harmonic to reduce the asymmetrical harmonic distortion and EMI-induced noise in current and voltage feedback samples. The filtered current and voltage feedback samples represent substantially the fundamental drive signal frequency, thus enabling a more accurate impedance phase measurement with respect to the fundamental drive signal frequency and an improvement in the generator's ability to maintain resonant frequency lock. The accuracy of the impedance phase measurement may be further enhanced by averaging falling edge and rising edge phase measurements, and by regulating the measured impedance phase to 0°.

Various aspects of the generator may also utilize the high-speed analog-to-digital sampling of the generator drive signal current and voltage, along with digital signal processing, to determine real power consumption and other quantities with a high degree of precision. This may allow the generator to implement a number of useful algorithms, such as, for example, controlling the amount of power delivered to tissue as the impedance of the tissue changes and controlling the power delivery to maintain a constant rate of tissue impedance increase. Some of these algorithms are used to determine the phase difference between the generator drive signal current and voltage signals. At resonance, the phase difference between the current and voltage signals is zero. The phase changes as the ultrasonic system goes off-resonance. Various algorithms may be employed to detect the phase difference and adjust the drive frequency until the ultrasonic system returns to resonance, i.e., the phase difference between the current and voltage signals goes to zero. The phase information also may be used to infer the conditions of the ultrasonic blade. As discussed with particularity below, the phase changes as a function of the temperature of the ultrasonic blade. Therefore, the phase information may be employed to control the temperature of the ultrasonic blade. This may be done, for example, by reducing the power delivered to the ultrasonic blade when the ultrasonic blade runs too hot and increasing the power delivered to the ultrasonic blade when the ultrasonic blade runs too cold.

Various aspects of the generator may have a wide frequency range and increased output power necessary to drive both ultrasonic surgical devices and electrosurgical devices. The lower voltage, higher current demand of electrosurgical devices may be met by a dedicated tap on a wideband power transformer, thereby eliminating the need for a separate power amplifier and output transformer. Moreover, sensing and feedback circuits of the generator may support a large dynamic range that addresses the needs of both ultrasonic and electrosurgical applications with minimal distortion.

Various aspects may provide a simple, economical means for the generator to read from, and optionally write to, a data circuit (e.g., a single-wire bus device, such as a one-wire protocol EEPROM known under the trade name "1-Wire") disposed in an instrument attached to the handpiece using

existing multi-conductor generator/handpiece cables. In this way, the generator is able to retrieve and process instrument-specific data from an instrument attached to the handpiece. This may enable the generator to provide better control and improved diagnostics and error detection. Additionally, the ability of the generator to write data to the instrument makes possible new functionality in terms of, for example, tracking instrument usage and capturing operational data. Moreover, the use of frequency band permits the backward compatibility of instruments containing a bus device with existing generators.

Disclosed aspects of the generator provide active cancellation of leakage current caused by unintended capacitive coupling between non-isolated and patient-isolated circuits of the generator. In addition to reducing patient risk, the reduction of leakage current may also lessen electromagnetic emissions.

FIG. 40 is a system diagram 7400 of a segmented circuit 7401 comprising a plurality of independently operated circuit segments 7402, 7414, 7416, 7420, 7424, 7428, 7434, 7440, in accordance with at least one aspect of the present disclosure. A circuit segment of the plurality of circuit segments of the segmented circuit 7401 comprises one or more circuits and one or more sets of machine executable instructions stored in one or more memory devices. The one or more circuits of a circuit segment are coupled to for electrical communication through one or more wired or wireless connection media. The plurality of circuit segments are configured to transition between three modes comprising a sleep mode, a standby mode and an operational mode.

In one aspect shown, the plurality of circuit segments 7402, 7414, 7416, 7420, 7424, 7428, 7434, 7440 start first in the standby mode, transition second to the sleep mode, and transition third to the operational mode. However, in other aspects, the plurality of circuit segments may transition from any one of the three modes to any other one of the three modes. For example, the plurality of circuit segments may transition directly from the standby mode to the operational mode. Individual circuit segments may be placed in a particular state by the voltage control circuit 7408 based on the execution by a processor of machine executable instructions. The states comprise a deenergized state, a low energy state, and an energized state. The deenergized state corresponds to the sleep mode, the low energy state corresponds to the standby mode, and the energized state corresponds to the operational mode. Transition to the low energy state may be achieved by, for example, the use of a potentiometer.

In one aspect, the plurality of circuit segments 7402, 7414, 7416, 7420, 7424, 7428, 7434, 7440 may transition from the sleep mode or the standby mode to the operational mode in accordance with an energization sequence. The plurality of circuit segments also may transition from the operational mode to the standby mode or the sleep mode in accordance with a deenergization sequence. The energization sequence and the deenergization sequence may be different. In some aspects, the energization sequence comprises energizing only a subset of circuit segments of the plurality of circuit segments. In some aspects, the deenergization sequence comprises deenergizing only a subset of circuit segments of the plurality of circuit segments.

Referring back to the system diagram 7400 in FIG. 40, the segmented circuit 7401 comprise a plurality of circuit segments comprising a transition circuit segment 7402, a processor circuit segment 7414, a handle circuit segment 7416, a communication circuit segment 7420, a display circuit segment 7424, a motor control circuit segment 7428, an energy treatment circuit segment 7434, and a shaft circuit

segment **7440**. The transition circuit segment comprises a wake up circuit **7404**, a boost current circuit **7406**, a voltage control circuit **7408**, a safety controller **7410** and a POST controller **7412**. The transition circuit segment **7402** is configured to implement a deenergization and an energiza-

tion sequence, a safety detection protocol, and a POST. In some aspects, the wake up circuit **7404** comprises an accelerometer button sensor **7405**. In aspects, the transition circuit segment **7402** is configured to be in an energized state while other circuit segments of the plurality of circuit segments of the segmented circuit **7401** are configured to be in a low energy state, a deenergized state or an energized state. The accelerometer button sensor **7405** may monitor movement or acceleration of the surgical instrument **6480** described herein. For example, the movement may be a change in orientation or rotation of the surgical instrument. The surgical instrument may be moved in any direction relative to a three dimensional Euclidean space by for example, a user of the surgical instrument. When the accelerometer button sensor **7405** senses movement or acceleration, the accelerometer button sensor **7405** sends a signal to the voltage control circuit **7408** to cause the voltage control circuit **7408** to apply voltage to the processor circuit segment **7414** to transition the processor and a volatile memory to an energized state. In aspects, the processor and the volatile memory are in an energized state before the voltage control circuit **7409** applies voltage to the processor and the volatile memory. In the operational mode, the processor may initiate an energization sequence or a deenergization sequence. In various aspects, the accelerometer button sensor **7405** may also send a signal to the processor to cause the processor to initiate an energization sequence or a deenergization sequence. In some aspects, the processor initiates an energization sequence when the majority of individual circuit segments are in a low energy state or a deenergized state. In other aspects, the processor initiates a deenergization sequence when the majority of individual circuit segments are in an energized state.

Additionally or alternatively, the accelerometer button sensor **7405** may sense external movement within a predetermined vicinity of the surgical instrument. For example, the accelerometer button sensor **7405** may sense a user of the surgical instrument **6480** described herein moving a hand of the user within the predetermined vicinity. When the accelerometer button sensor **7405** senses this external movement, the accelerometer button sensor **7405** may send a signal to the voltage control circuit **7408** and a signal to the processor, as previously described. After receiving the sent signal, the processor may initiate an energization sequence or a deenergization sequence to transition one or more circuit segments between the three modes. In aspects, the signal sent to the voltage control circuit **7408** is sent to verify that the processor is in operational mode. In some aspects, the accelerometer button sensor **7405** may sense when the surgical instrument has been dropped and send a signal to the processor based on the sensed drop. For example, the signal can indicate an error in the operation of an individual circuit segment. One or more sensors may sense damage or malfunctioning of the affected individual circuit segments. Based on the sensed damage or malfunctioning, the POST controller **7412** may perform a POST of the corresponding individual circuit segments.

An energization sequence or a deenergization sequence may be defined based on the accelerometer button sensor **7405**. For example, the accelerometer button sensor **7405** may sense a particular motion or a sequence of motions that indicates the selection of a particular circuit segment of the

plurality of circuit segments. Based on the sensed motion or series of sensed motions, the accelerometer button sensor **7405** may transmit a signal comprising an indication of one or more circuit segments of the plurality of circuit segments to the processor when the processor is in an energized state. Based on the signal, the processor determines an energization sequence comprising the selected one or more circuit segments. Additionally or alternatively, a user of the surgical instruments **6480** described herein may select a number and order of circuit segments to define an energization sequence or a deenergization sequence based on interaction with a graphical user interface (GUI) of the surgical instrument.

In various aspects, the accelerometer button sensor **7405** may send a signal to the voltage control circuit **7408** and a signal to the processor only when the accelerometer button sensor **7405** detects movement of the surgical instrument **6480** described herein or external movement within a predetermined vicinity above a predetermined threshold. For example, a signal may only be sent if movement is sensed for 5 or more seconds or if the surgical instrument is moved 5 or more inches. In other aspects, the accelerometer button sensor **7405** may send a signal to the voltage control circuit **7408** and a signal to the processor only when the accelerometer button sensor **7405** detects oscillating movement of the surgical instrument. A predetermined threshold reduces inadvertent transition of circuit segments of the surgical instrument. As previously described, the transition may comprise a transition to operational mode according to an energization sequence, a transition to low energy mode according to a deenergization sequence, or a transition to sleep mode according to a deenergization sequence. In some aspects, the surgical instrument comprises an actuator that may be actuated by a user of the surgical instrument. The actuation is sensed by the accelerometer button sensor **7405**. The actuator may be a slider, a toggle switch, or a momentary contact switch. Based on the sensed actuation, the accelerometer button sensor **7405** may send a signal to the voltage control circuit **7408** and a signal to the processor.

The boost current circuit **7406** is coupled to a battery. The boost current circuit **7406** is a current amplifier, such as a relay or transistor, and is configured to amplify the magnitude of a current of an individual circuit segment. The initial magnitude of the current corresponds to the source voltage provided by the battery to the segmented circuit **7401**. Suitable relays include solenoids. Suitable transistors include field-effect transistors (FET), MOSFET, and bipolar junction transistors (BJT). The boost current circuit **7406** may amplify the magnitude of the current corresponding to an individual circuit segment or circuit which requires more current draw during operation of the surgical instruments **6480** described herein. For example, an increase in current to the motor control circuit segment **7428** may be provided when a motor of the surgical instrument requires more input power. The increase in current provided to an individual circuit segment may cause a corresponding decrease in current of another circuit segment or circuit segments. Additionally or alternatively, the increase in current may correspond to voltage provided by an additional voltage source operating in conjunction with the battery.

The voltage control circuit **7408** is coupled to the battery. The voltage control circuit **7408** is configured to provide voltage to or remove voltage from the plurality of circuit segments. The voltage control circuit **7408** is also configured to increase or reduce voltage provided to the plurality of circuit segments of the segmented circuit **7401**. In various aspects, the voltage control circuit **7408** comprises a combinational logic circuit such as a multiplexer (MUX) to

select inputs, a plurality of electronic switches, and a plurality of voltage converters. An electronic switch of the plurality of electronic switches may be configured to switch between an open and closed configuration to disconnect or connect an individual circuit segment to or from the battery. The plurality of electronic switches may be solid state devices such as transistors or other types of switches such as wireless switches, ultrasonic switches, accelerometers, inertial sensors, among others. The combinational logic circuit is configured to select an individual electronic switch for switching to an open configuration to enable application of voltage to the corresponding circuit segment. The combination logic circuit also is configured to select an individual electronic switch for switching to a closed configuration to enable removal of voltage from the corresponding circuit segment. By selecting a plurality of individual electronic switches, the combination logic circuit may implement a deenergization sequence or an energization sequence. The plurality of voltage converters may provide a stepped-up voltage or a stepped-down voltage to the plurality of circuit segments. The voltage control circuit **7408** may also comprise a microprocessor and memory device.

The safety controller **7410** is configured to perform safety checks for the circuit segments. In some aspects, the safety controller **7410** performs the safety checks when one or more individual circuit segments are in the operational mode. The safety checks may be performed to determine whether there are any errors or defects in the functioning or operation of the circuit segments. The safety controller **7410** may monitor one or more parameters of the plurality of circuit segments. The safety controller **7410** may verify the identity and operation of the plurality of circuit segments by comparing the one or more parameters with predefined parameters. For example, if an RF energy modality is selected, the safety controller **7410** may verify that an articulation parameter of the shaft matches a predefined articulation parameter to verify the operation of the RF energy modality of the surgical instrument **6480** described herein. In some aspects, the safety controller **7410** may monitor, by the sensors, a predetermined relationship between one or more properties of the surgical instrument to detect a fault. A fault may arise when the one or more properties are inconsistent with the predetermined relationship. When the safety controller **7410** determines that a fault exists, an error exists, or that some operation of the plurality of circuit segments was not verified, the safety controller **7410** prevents or disables operation of the particular circuit segment where the fault, error or verification failure originated.

The POST controller **7412** performs a POST to verify proper operation of the plurality of circuit segments. In some aspects, the POST is performed for an individual circuit segment of the plurality of circuit segments prior to the voltage control circuit **7408** applying a voltage to the individual circuit segment to transition the individual circuit segment from standby mode or sleep mode to operational mode. If the individual circuit segment does not pass the POST, the particular circuit segment does not transition from standby mode or sleep mode to operational mode. POST of the handle circuit segment **7416** may comprise, for example, testing whether the handle control sensors **7418** sense an actuation of a handle control of the surgical instrument **6480** described herein. In some aspects, the POST controller **7412** may transmit a signal to the accelerometer button sensor **7405** to verify the operation of the individual circuit segment as part of the POST. For example, after receiving the signal, the accelerometer button sensor **7405** may prompt a user of

the surgical instrument to move the surgical instrument to a plurality of varying locations to confirm operation of the surgical instrument. The accelerometer button sensor **7405** may also monitor an output of a circuit segment or a circuit of a circuit segment as part of the POST. For example, the accelerometer button sensor **7405** can sense an incremental motor pulse generated by the motor **7432** to verify operation. A motor controller of the motor control circuit **7430** may be used to control the motor **7432** to generate the incremental motor pulse.

In various aspects, the surgical instrument **6480** described herein may comprise additional accelerometer button sensors. The POST controller **7412** may also execute a control program stored in the memory device of the voltage control circuit **7408**. The control program may cause the POST controller **7412** to transmit a signal requesting a matching encrypted parameter from a plurality of circuit segments. Failure to receive a matching encrypted parameter from an individual circuit segment indicates to the POST controller **7412** that the corresponding circuit segment is damaged or malfunctioning. In some aspects, if the POST controller **7412** determines based on the POST that the processor is damaged or malfunctioning, the POST controller **7412** may send a signal to one or more secondary processors to cause one or more secondary processors to perform critical functions that the processor is unable to perform. In some aspects, if the POST controller **7412** determines based on the POST that one or more circuit segments do not operate properly, the POST controller **7412** may initiate a reduced performance mode of those circuit segments operating properly while locking out those circuit segments that fail POST or do not operate properly. A locked out circuit segment may function similarly to a circuit segment in standby mode or sleep mode.

The processor circuit segment **7414** comprises the processor and the volatile memory. The processor is configured to initiate an energization or a deenergization sequence. To initiate the energization sequence, the processor transmits an energizing signal to the voltage control circuit **7408** to cause the voltage control circuit **7408** to apply voltage to the plurality or a subset of the plurality of circuit segments in accordance with the energization sequence. To initiate the deenergization sequence, the processor transmits a deenergizing signal to the voltage control circuit **7408** to cause the voltage control circuit **7408** to remove voltage from the plurality or a subset of the plurality of circuit segments in accordance with the deenergization sequence.

The handle circuit segment **7416** comprises handle control sensors **7418**. The handle control sensors **7418** may sense an actuation of one or more handle controls of the surgical instrument **6480** described herein. In various aspects, the one or more handle controls comprise a clamp control, a release button, an articulation switch, an energy activation button, and/or any other suitable handle control. The user may activate the energy activation button to select between an RF energy mode, an ultrasonic energy mode or a combination RF and ultrasonic energy mode. The handle control sensors **7418** may also facilitate attaching a modular handle to the surgical instrument. For example, the handle control sensors **7418** may sense proper attachment of the modular handle to the surgical instrument and indicate the sensed attachment to a user of the surgical instrument. The LCD display **7426** may provide a graphical indication of the sensed attachment. In some aspects, the handle control sensors **7418** senses actuation of the one or more handle

controls. Based on the sensed actuation, the processor may initiate either an energization sequence or a deenergization sequence.

The communication circuit segment **7420** comprises a communication circuit **7422**. The communication circuit **7422** comprises a communication interface to facilitate signal communication between the individual circuit segments of the plurality of circuit segments. In some aspects, the communication circuit **7422** provides a path for the modular components of the surgical instrument **6480** described herein to communicate electrically. For example, a modular shaft and a modular transducer, when attached together to the handle of the surgical instrument, can upload control programs to the handle through the communication circuit **7422**.

The display circuit segment **7424** comprises a LCD display **7426**. The LCD display **7426** may comprise a liquid crystal display screen, LED indicators, etc. In some aspects, the LCD display **7426** is an organic light-emitting diode (OLED) screen. A display may be placed on, embedded in, or located remotely from the surgical instrument **6480** described herein. For example, the display can be placed on the handle of the surgical instrument. The display is configured to provide sensory feedback to a user. In various aspects, the LCD display **7426** further comprises a backlight. In some aspects, the surgical instrument may also comprise audio feedback devices such as a speaker or a buzzer and tactile feedback devices such as a haptic actuator.

The motor control circuit segment **7428** comprises a motor control circuit **7430** coupled to a motor **7432**. The motor **7432** is coupled to the processor by a driver and a transistor, such as a FET. In various aspects, the motor control circuit **7430** comprises a motor current sensor in signal communication with the processor to provide a signal indicative of a measurement of the current draw of the motor to the processor. The processor transmits the signal to the display. The display receives the signal and displays the measurement of the current draw of the motor **7432**. The processor may use the signal, for example, to monitor that the current draw of the motor **7432** exists within an acceptable range, to compare the current draw to one or more parameters of the plurality of circuit segments, and to determine one or more parameters of a patient treatment site. In various aspects, the motor control circuit **7430** comprises a motor controller to control the operation of the motor. For example, the motor control circuit **7430** controls various motor parameters, such as by adjusting the velocity, torque and acceleration of the motor **7432**. The adjusting is done based on the current through the motor **7432** measured by the motor current sensor.

In various aspects, the motor control circuit **7430** comprises a force sensor to measure the force and torque generated by the motor **7432**. The motor **7432** is configured to actuate a mechanism of the surgical instruments **6480** described herein. For example, the motor **7432** is configured to control actuation of the shaft of the surgical instrument to realize clamping, rotation and articulation functionality. For example, the motor **7432** may actuate the shaft to realize a clamping motion with jaws of the surgical instrument. The motor controller may determine whether the material clamped by the jaws is tissue or metal. The motor controller may also determine the extent to which the jaws clamp the material. For example, the motor controller may determine how open or closed the jaws are based on the derivative of sensed motor current or motor voltage. In some aspects, the motor **7432** is configured to actuate the transducer to cause the transducer to apply torque to the handle or to control

articulation of the surgical instrument. The motor current sensor may interact with the motor controller to set a motor current limit. When the current meets the predefined threshold limit, the motor controller initiates a corresponding change in a motor control operation. For example, exceeding the motor current limit causes the motor controller to reduce the current draw of the motor.

The energy treatment circuit segment **7434** comprises a RF amplifier and safety circuit **7436** and an ultrasonic signal generator circuit **7438** to implement the energy modular functionality of the surgical instrument **6480** described herein. In various aspects, the RF amplifier and safety circuit **7436** is configured to control the RF modality of the surgical instrument by generating an RF signal. The ultrasonic signal generator circuit **7438** is configured to control the ultrasonic energy modality by generating an ultrasonic signal. The RF amplifier and safety circuit **7436** and an ultrasonic signal generator circuit **7438** may operate in conjunction to control the combination RF and ultrasonic energy modality.

The shaft circuit segment **7440** comprises a shaft module controller **7442**, a modular control actuator **7444**, one or more end effector sensors **7446**, and a non volatile memory **7448**. The shaft module controller **7442** is configured to control a plurality of shaft modules comprising the control programs to be executed by the processor. The plurality of shaft modules implements a shaft modality, such as ultrasonic, combination ultrasonic and RF, RF I-blade, and RF-opposable jaw. The shaft module controller **7442** can select shaft modality by selecting the corresponding shaft module for the processor to execute. The modular control actuator **7444** is configured to actuate the shaft according to the selected shaft modality. After actuation is initiated, the shaft articulates the end effector according to the one or more parameters, routines or programs specific to the selected shaft modality and the selected end effector modality. The one or more end effector sensors **7446** located at the end effector may include force sensors, temperature sensors, current sensors or motion sensors. The one or more end effector sensors **7446** transmit data about one or more operations of the end effector, based on the energy modality implemented by the end effector. In various aspects, the energy modalities include an ultrasonic energy modality, a RF energy modality, or a combination of the ultrasonic energy modality and the RF energy modality. The non volatile memory **7448** stores the shaft control programs. A control program comprises one or more parameters, routines or programs specific to the shaft. In various aspects, the non volatile memory **7448** may be a ROM, EPROM, EEPROM or flash memory. The non volatile memory **7448** stores the shaft modules corresponding to the selected shaft of the surgical instrument **6480** described herein in. The shaft modules may be changed or upgraded in the non volatile memory **7448** by the shaft module controller **7442**, depending on the surgical instrument shaft to be used in operation.

FIG. **41** is a schematic diagram of a circuit **7925** of various components of a surgical instrument with motor control functions, in accordance with at least one aspect of the present disclosure. In various aspects, the surgical instrument **6480** described herein may include a drive mechanism **7930** which is configured to drive shafts and/or gear components in order to perform the various operations associated with the surgical instrument **6480**. In one aspect, the drive mechanism **7930** includes a rotation drivetrain **7932** configured to rotate an end effector, for example, about a longitudinal axis relative to handle housing. The drive mechanism **7930** further includes a closure drivetrain **7934** configured to close a jaw member to grasp tissue with the

end effector. In addition, the drive mechanism **7930** includes a firing drive train **7936** configured to open and close a clamp arm portion of the end effector to grasp tissue with the end effector.

The drive mechanism **7930** includes a selector gearbox assembly **7938** that can be located in the handle assembly of the surgical instrument. Proximal to the selector gearbox assembly **7938** is a function selection module which includes a first motor **7942** that functions to selectively move gear elements within the selector gearbox assembly **7938** to selectively position one of the drivetrains **7932**, **7934**, **7936** into engagement with an input drive component of an optional second motor **7944** and motor drive circuit **7946** (shown in dashed line to indicate that the second motor **7944** and motor drive circuit **7946** are optional components).

Still referring to FIG. **41**, the motors **7942**, **7944** are coupled to motor control circuits **7946**, **7948**, respectively, which are configured to control the operation of the motors **7942**, **7944** including the flow of electrical energy from a power source **7950** to the motors **7942**, **7944**. The power source **7950** may be a DC battery (e.g., rechargeable lead-based, nickel-based, lithium-ion based, battery etc.) or any other power source suitable for providing electrical energy to the surgical instrument.

The surgical instrument further includes a microcontroller **7952** (“controller”). In certain instances, the controller **7952** may include a microprocessor **7954** (“processor”) and one or more computer readable mediums or memory units **7956** (“memory”). In certain instances, the memory **7956** may store various program instructions, which when executed may cause the processor **7954** to perform a plurality of functions and/or calculations described herein. The power source **7950** can be configured to supply power to the controller **7952**, for example.

The processor **7954** may be in communication with the motor control circuit **7946**. In addition, the memory **7956** may store program instructions, which when executed by the processor **7954** in response to a user input **7958** or feedback elements **7960**, may cause the motor control circuit **7946** to motivate the motor **7942** to generate at least one rotational motion to selectively move gear elements within the selector gearbox assembly **7938** to selectively position one of the drivetrains **7932**, **7934**, **7936** into engagement with the input drive component of the second motor **7944**. Furthermore, the processor **7954** can be in communication with the motor control circuit **7948**. The memory **7956** also may store program instructions, which when executed by the processor **7954** in response to a user input **7958**, may cause the motor control circuit **7948** to motivate the motor **7944** to generate at least one rotational motion to drive the drivetrain engaged with the input drive component of the second motor **7948**, for example.

The controller **7952** and/or other controllers of the present disclosure may be implemented using integrated and/or discrete hardware elements, software elements, and/or a combination of both. Examples of integrated hardware elements may include processors, microprocessors, microcontrollers, integrated circuits, ASICs, PLDs, DSPs, FPGAs, logic gates, registers, semiconductor devices, chips, microchips, chip sets, microcontrollers, system on a chip (SoC), and/or single in-line package (SIP). Examples of discrete hardware elements may include circuits and/or circuit elements such as logic gates, field effect transistors, bipolar transistors, resistors, capacitors, inductors, and/or relays. In certain instances, the controller **7952** may include a hybrid circuit comprising discrete and integrated circuit elements or components on one or more substrates, for example.

In certain instances, the controller **7952** and/or other controllers of the present disclosure may be an LM4F230H5QR, available from Texas Instruments, for example. In certain instances, the Texas Instruments LM4F230H5QR is an ARM Cortex-M4F Processor Core comprising on-chip memory of 256 KB single-cycle flash memory, or other non-volatile memory, up to 40 MHz, a prefetch buffer to improve performance above 40 MHz, a 32 KB single-cycle SRAM, internal ROM loaded with StellarisWare® software, 2 KB EEPROM, one or more PWM modules, one or more QEI analog, one or more 12-bit ADC with 12 analog input channels, among other features that are readily available. Other microcontrollers may be readily substituted for use with the present disclosure. Accordingly, the present disclosure should not be limited in this context.

In various instances, one or more of the various steps described herein can be performed by a finite state machine comprising either a combinational logic circuit or a sequential logic circuit, where either the combinational logic circuit or the sequential logic circuit is coupled to at least one memory circuit. The at least one memory circuit stores a current state of the finite state machine. The combinational or sequential logic circuit is configured to cause the finite state machine to the steps. The sequential logic circuit may be synchronous or asynchronous. In other instances, one or more of the various steps described herein can be performed by a circuit that includes a combination of the processor **7958** and the finite state machine, for example.

In various instances, it can be advantageous to be able to assess the state of the functionality of a surgical instrument to ensure its proper function. It is possible, for example, for the drive mechanism, as explained above, which is configured to include various motors, drivetrains, and/or gear components in order to perform the various operations of the surgical instrument, to wear out over time. This can occur through normal use, and in some instances the drive mechanism can wear out faster due to abuse conditions. In certain instances, a surgical instrument can be configured to perform self-assessments to determine the state, e.g. health, of the drive mechanism and its various components.

For example, the self-assessment can be used to determine when the surgical instrument is capable of performing its function before a re-sterilization or when some of the components should be replaced and/or repaired. Assessment of the drive mechanism and its components, including but not limited to the rotation drivetrain **7932**, the closure drivetrain **7934**, and/or the firing drivetrain **7936**, can be accomplished in a variety of ways. The magnitude of deviation from a predicted performance can be used to determine the likelihood of a sensed failure and the severity of such failure. Several metrics can be used including: Periodic analysis of repeatably predictable events, Peaks or drops that exceed an expected threshold, and width of the failure.

In various instances, a signature waveform of a properly functioning drive mechanism or one or more of its components can be employed to assess the state of the drive mechanism or the one or more of its components. One or more vibration sensors can be arranged with respect to a properly functioning drive mechanism or one or more of its components to record various vibrations that occur during operation of the properly functioning drive mechanism or the one or more of its components. The recorded vibrations can be employed to create the signature waveform. Future waveforms can be compared against the signature waveform to assess the state of the drive mechanism and its components.

Still referring to FIG. 41, the surgical instrument 7930 includes a drivetrain failure detection module 7962 configured to record and analyze one or more acoustic outputs of one or more of the drivetrains 7932, 7934, 7936. The processor 7954 can be in communication with or otherwise control the module 7962. As described below in greater detail, the module 7962 can be embodied as various means, such as circuitry, hardware, a computer program product comprising a computer readable medium (for example, the memory 7956) storing computer readable program instructions that are executable by a processing device (for example, the processor 7954), or some combination thereof. In some aspects, the processor 36 can include, or otherwise control the module 7962.

FIG. 42 is an alternative system 132000 for controlling the frequency of an ultrasonic electromechanical system 132002 and detecting the impedance thereof, in accordance with at least one aspect of the present disclosure. The system 132000 may be incorporated into a generator. A processor 132004 coupled to a memory 132026 programs a programmable counter 132006 to tune to the output frequency f_o of the ultrasonic electromechanical system 132002. The input frequency is generated by a crystal oscillator 132008 and is input into a fixed counter 132010 to scale the frequency to a suitable value. The outputs of the fixed counter 132010 and the programmable counter 132006 are applied to a phase/frequency detector 132012. The output of the phase/frequency detector 132012 is applied to an amplifier/active filter circuit 132014 to generate a tuning voltage V that is applied to a voltage controlled oscillator 132016 (VCO). The VCO 132016 applies the output frequency f_o to an ultrasonic transducer portion of the ultrasonic electromechanical system 132002, shown here modeled as an equivalent electrical circuit. The voltage and current signals applied to the ultrasonic transducer are monitored by a voltage sensor 132018 and a current sensor 132020.

The outputs of the voltage and current sensors 132018, 132020 are applied to another phase/frequency detector 132022 to determine the phase angle between the voltage and current as measured by the voltage and current sensors 132018, 132020. The output of the phase/frequency detector 132022 is applied to one channel of a high speed analog to digital converter 132024 (ADC) and is provided to the processor 132004 therethrough. Optionally, the outputs of the voltage and current sensors 132018, 132020 may be applied to respective channels of the two-channel ADC 132024 and provided to the processor 132004 for zero crossing, FFT, or other algorithm described herein for determining the phase angle between the voltage and current signals applied to the ultrasonic electromechanical system 132002.

Optionally the tuning voltage V_t , which is proportional to the output frequency f_o , may be fed back to the processor 132004 via the ADC 132024. This provides the processor 132004 with a feedback signal proportional to the output frequency f_o and can use this feedback to adjust and control the output frequency f_o .

Temperature Inference

FIGS. 43A-43B are graphical representations 133000, 133010 of complex impedance spectra of the same ultrasonic device with a cold (room temperature) and hot ultrasonic blade, in accordance with at least one aspect of the present disclosure. As used herein, a cold ultrasonic blade refers to an ultrasonic blade at room temperature and a hot ultrasonic blade refers to an ultrasonic blade after it is frictionally heated in use. FIG. 43A is a graphical representation 133000 of impedance phase angle φ as a function of

resonant frequency f_o of the same ultrasonic device with a cold and hot ultrasonic blade and FIG. 43B is a graphical representation 133010 of impedance magnitude $|Z|$ as a function of resonant frequency f_o of the same ultrasonic device with a cold and hot ultrasonic blade. The impedance phase angle φ and impedance magnitude $|Z|$ are at a minimum at the resonant frequency f_o .

The ultrasonic transducer impedance $Z_g(t)$ can be measured as the ratio of the drive signal generator voltage $V_g(t)$ and current $I_g(t)$ drive signals:

$$Z_g(t) = \frac{V_g(t)}{I_g(t)}$$

As shown in FIG. 43A, when the ultrasonic blade is cold, e.g., at room temperature and not frictionally heated, the electromechanical resonant frequency f_o of the ultrasonic device is approximately 55,500 Hz and the excitation frequency of the ultrasonic transducer is set to 55,500 Hz. Thus, when the ultrasonic transducer is excited at the electromechanical resonant frequency f_o and the ultrasonic blade is cold the phase angle φ is at minimum or approximately 0 Rad as indicated by the cold blade plot 133002. As shown in FIG. 43B, when the ultrasonic blade is cold and the ultrasonic transducer is excited at the electromechanical resonant frequency f_o , the impedance magnitude $|Z|$ is 800 Ω , e.g., the impedance magnitude $|Z|$ is at a minimum impedance, and the drive signal amplitude is at a maximum due to the series resonance equivalent circuit of the ultrasonic electromechanical system as depicted in FIG. 25.

With reference now back to FIGS. 43A and 43B, when the ultrasonic transducer is driven by generator voltage $V_g(t)$ and generator current $I_g(t)$ signals at the electromechanical resonant frequency f_o of 55,500 Hz, the phase angle φ between the generator voltage $V_g(t)$ and generator current $I_g(t)$ signals is zero, the impedance magnitude $|Z|$ is at a minimum impedance, e.g., 800 Ω , and the signal amplitude is at a peak or maximum due to the series resonance equivalent circuit of the ultrasonic electromechanical system. As the temperature of the ultrasonic blade increases, due to frictional heat generated in use, the electromechanical resonant frequency f_o' of the ultrasonic device decreases. Since the ultrasonic transducer is still driven by generator voltage $V_g(t)$ and generator current $I_g(t)$ signals at the previous (cold blade) electromechanical resonant frequency f_o of 55,500 Hz, the ultrasonic device operates off-resonance f_o' causing a shift in the phase angle φ between the generator voltage $V_g(t)$ and generator current $I_g(t)$ signals. There is also an increase in impedance magnitude $|Z|$ and a drop in peak magnitude of the drive signal relative to the previous (cold blade) electromechanical resonant frequency of 55,500 Hz. Accordingly, the temperature of the ultrasonic blade may be inferred by measuring the phase angle φ between the generator voltage $V_g(t)$ and the generator current $I_g(t)$ signals as the electromechanical resonant frequency f_o changes due to the changes in temperature of the ultrasonic blade.

As previously described, an electromechanical ultrasonic system includes an ultrasonic transducer, a waveguide, and an ultrasonic blade. As previously discussed, the ultrasonic transducer may be modeled as an equivalent series resonant circuit (see FIG. 25) comprising first branch having a static capacitance and a second "motional" branch having a serially connected inductance, resistance and capacitance that define the electromechanical properties of a resonator. The electromechanical ultrasonic system has an initial electro-

mechanical resonant frequency defined by the physical properties of the ultrasonic transducer, the waveguide, and the ultrasonic blade. The ultrasonic transducer is excited by an alternating voltage $V_g(t)$ and current $I_g(t)$ signal at a frequency equal to the electromechanical resonant frequency, e.g., the resonant frequency of the electromechanical ultrasonic system. When the electromechanical ultrasonic system is excited at the resonant frequency, the phase angle φ between the voltage $V_g(t)$ and current $I_g(t)$ signals is zero.

Stated in another way, at resonance, the analogous inductive impedance of the electromechanical ultrasonic system is equal to the analogous capacitive impedance of the electromechanical ultrasonic system. As the ultrasonic blade heats up, for example due to frictional engagement with tissue, the compliance of the ultrasonic blade (modeled as an analogous capacitance) causes the resonant frequency of the electromechanical ultrasonic system to shift. In the present example, the resonant frequency of the electromechanical ultrasonic system decreases as the temperature of the ultrasonic blade increases. Thus, the analogous inductive impedance of the electromechanical ultrasonic system is no longer equal to the analogous capacitive impedance of the electromechanical ultrasonic system causing a mismatch between the drive frequency and the new resonant frequency of the electromechanical ultrasonic system. Thus, with a hot ultrasonic blade, the electromechanical ultrasonic system operates "off-resonance." The mismatch between the drive frequency and the resonant frequency is manifested as a phase angle φ between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer.

As previously discussed, the generator electronics can easily monitor the phase angle φ between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer. The phase angle φ may be determined through Fourier analysis, weighted least-squares estimation, Kalman filtering, space-vector-based techniques, zero-crossing method, Lissajous figures, three-voltmeter method, crossed-coil method, vector voltmeter and vector impedance methods, phase standard instruments, phase-locked loops, among other techniques previously described. The generator can continuously monitor the phase angle φ and adjust the drive frequency until the phase angle φ goes to zero. At this point, the new drive frequency is equal to the new resonant frequency of the electromechanical ultrasonic system. The change in phase angle φ and/or generator drive frequency can be used as an indirect or inferred measurement of the temperature of the ultrasonic blade.

A variety of techniques are available to estimate temperature from the data in these spectra. Most notably, a time variant, non-linear set of state space equations can be employed to model the dynamic relationship between the temperature of the ultrasonic blade and the measured impedance

$$Z_g(t) = \frac{V_g(t)}{I_g(t)}$$

across a range of generator drive frequencies, where the range of generator drive frequencies is specific to device model.

Methods of Temperature Estimation

One aspect of estimating or inferring the temperature of an ultrasonic blade may include three steps. First, define a state space model of temperature and frequency that is time

and energy dependent. To model temperature as a function of frequency content, a set of non-linear state space equations are used to model the relationship between the electromechanical resonant frequency and the temperature of the ultrasonic blade. Second, apply a Kalman filter to improve the accuracy of the temperature estimator and state space model over time. Third, a state estimator is provided in the feedback loop of the Kalman filter to control the power applied to the ultrasonic transducer, and hence the ultrasonic blade, to regulate the temperature of the ultrasonic blade. The three steps are described hereinbelow.

Step 1

The first step is to define a state space model of temperature and frequency that is time and energy dependent. To model temperature as a function of frequency content, a set of non-linear state space equations are used to model the relationship between the electromechanical resonant frequency and the temperature of the ultrasonic blade. In one aspect, the state space model is given by:

$$\begin{aligned} \begin{bmatrix} \dot{F}_n \\ \dot{T} \end{bmatrix} &= f(t, T(t), F_n(t), E(t)) \\ \dot{y} &= h(t, T(t), F_n(t), E(t)) \end{aligned}$$

The state space model represents the rate of change of the natural frequency of the electromechanical ultrasonic system \dot{F}_n and the rate of change of the temperature \dot{T} of the ultrasonic blade with respect to natural frequency $F_n(t)$, temperature $T(t)$, energy $E(t)$, and time t . \dot{y} represents the observability of variables that are measurable and observable such as the natural frequency $F_n(t)$ of the electromechanical ultrasonic system, the temperature $T(t)$ of the ultrasonic blade, the energy $E(t)$ applied to the ultrasonic blade, and time t . The temperature $T(t)$ of the ultrasonic blade is observable as an estimate.

Step 2

The second step is to apply a Kalman filter to improve temperature estimator and state space model. FIG. 44 is a diagram of a Kalman filter 133020 to improve the temperature estimator and state space model based on impedance according to the equation:

$$Z_g(t) = \frac{V_g(t)}{I_g(t)}$$

which represents the impedance across an ultrasonic transducer measured at a variety of frequencies, in accordance with at least one aspect of the present disclosure.

The Kalman filter 133020 may be employed to improve the performance of the temperature estimate and allows for the augmentation of external sensors, models, or prior information to improve temperature prediction in the midst of noisy data. The Kalman filter 133020 includes a regulator 133022 and a plant 133024. In control theory a plant 133024 is the combination of process and actuator. A plant 133024 may be referred to with a transfer function which indicates the relation between an input signal and the output signal of a system. The regulator 133022 includes a state estimator

133026 and a controller K 133028. The state regulator 133026 includes a feedback loop 133030. The state regulator 133026 receives y , the output of the plant 133024, as an input and a feedback variable u . The state estimator 133026 is an internal feedback system that converges to the true value of the state of the system. The output of the state estimator 133026 is \hat{x} , the full feedback control variable including $F_n(t)$ of the electromechanical ultrasonic system, the estimate of the temperature $T(t)$ of the ultrasonic blade, the energy $E(t)$ applied to the ultrasonic blade, the phase angle φ , and time t . The input into the controller K 133028 is \hat{x} and the output of the controller K 133028 u is fed back to the state estimator 133026 and t of the plant 133024.

Kalman filtering, also known as linear quadratic estimation (LQE), is an algorithm that uses a series of measurements observed over time, containing statistical noise and other inaccuracies, and produces estimates of unknown variables that tend to be more accurate than those based on a single measurement alone, by estimating a joint probability distribution over the variables for each timeframe and thus calculating the maximum likelihood estimate of actual measurements. The algorithm works in a two-step process. In a prediction step, the Kalman filter 133020 produces estimates of the current state variables, along with their uncertainties. Once the outcome of the next measurement (necessarily corrupted with some amount of error, including random noise) is observed, these estimates are updated using a weighted average, with more weight being given to estimates with higher certainty. The algorithm is recursive and can run in real time, using only the present input measurements and the previously calculated state and its uncertainty matrix; no additional past information is required.

The Kalman filter 133020 uses a dynamics model of the electromechanical ultrasonic system, known control inputs to that system, and multiple sequential measurements (observations) of the natural frequency and phase angle of the applied signals (e.g., magnitude and phase of the electrical impedance of the ultrasonic transducer) to the ultrasonic transducer to form an estimate of the varying quantities of the electromechanical ultrasonic system (its state) to predict the temperature of the ultrasonic blade portion of the electromechanical ultrasonic system that is better than an estimate obtained using only one measurement alone. As such, the Kalman filter 133020 is an algorithm that includes sensor and data fusion to provide the maximum likelihood estimate of the temperature of the ultrasonic blade.

The Kalman filter 133020 deals effectively with uncertainty due to noisy measurements of the applied signals to the ultrasonic transducer to measure the natural frequency and phase shift data and also deals effectively with uncertainty due to random external factors. The Kalman filter 133020 produces an estimate of the state of the electromechanical ultrasonic system as an average of the predicted state of the system and of the new measurement using a weighted average. Weighted values provide better (i.e., smaller) estimated uncertainty and are more "trustworthy" than unweighted values. The weights may be calculated from the covariance, a measure of the estimated uncertainty of the prediction of the system's state. The result of the weighted average is a new state estimate that lies between the predicted and measured state, and has a better estimated uncertainty than either alone. This process is repeated at every time step, with the new estimate and its covariance informing the prediction used in the following iteration. This recursive nature of the Kalman filter 133020 requires only

the last "best guess," rather than the entire history, of the state of the electromechanical ultrasonic system to calculate a new state.

The relative certainty of the measurements and current state estimate is an important consideration, and it is common to discuss the response of the filter in terms of the gain K of the Kalman filter 133020. The Kalman gain K is the relative weight given to the measurements and current state estimate, and can be "tuned" to achieve particular performance. With a high gain K , the Kalman filter 133020 places more weight on the most recent measurements, and thus follows them more responsively. With a low gain K , the Kalman filter 133020 follows the model predictions more closely. At the extremes, a high gain close to one will result in a more jumpy estimated trajectory, while low gain close to zero will smooth out noise but decrease the responsiveness.

When performing the actual calculations for the Kalman filter 133020 (as discussed below), the state estimate and covariances are coded into matrices to handle the multiple dimensions involved in a single set of calculations. This allows for a representation of linear relationships between different state variables (such as position, velocity, and acceleration) in any of the transition models or covariances. Using a Kalman filter 133020 does not assume that the errors are Gaussian. However, the Kalman filter 133020 yields the exact conditional probability estimate in the special case that all errors are Gaussian-distributed.

Step 3

The third step uses a state estimator 133026 in the feedback loop 133032 of the Kalman filter 133020 for control of power applied to the ultrasonic transducer, and hence the ultrasonic blade, to regulate the temperature of the ultrasonic blade.

FIG. 45 is a graphical depiction 133040 of three probability distributions employed by the state estimator 133026 of the Kalman filter 133020 shown in FIG. 44 to maximize estimates, in accordance with at least one aspect of the present disclosure. The probability distributions include the prior probability distribution 133042, the prediction (state) probability distribution 133044, and the observation probability distribution 133046. The three probability distributions 133042, 133044, 133046 are used in feedback control of power applied to an ultrasonic transducer to regulate temperature based on impedance across the ultrasonic transducer measured at a variety of frequencies, in accordance with at least one aspect of the present disclosure. The estimator used in feedback control of power applied to an ultrasonic transducer to regulate temperature based on impedance is given by the expression:

$$Z_g(t) = \frac{V_g(t)}{I_g(t)}$$

which is the impedance across the ultrasonic transducer measured at a variety of frequencies, in accordance with at least one aspect of the present disclosure.

The prior probability distribution 133042 includes a state variance defined by the expression:

$$(\Phi_k^-)^2 = \sigma_{k-1}^2 + \sigma_{P_k}^2$$

The state variance (σ_k^-) is used to predict the next state of the system, which is represented as the prediction (state)

probability distribution **133044**. The observation probability distribution **133046** is the probability distribution of the actual observation of the state of the system where the observation variance σ_m is used to define the gain, which is given by the following expression:

$$K = \frac{(\sigma_k^-)^2}{(\sigma_k^-)^2 + \sigma_m^2}$$

Feedback Control

Power input is decreased to ensure that the temperature (as estimated by the state estimator and of the Kalman filter) is controlled.

In one aspect, the initial proof of concept assumed a static, linear relationship between the natural frequency of the electromechanical ultrasonic system and the temperature of the ultrasonic blade. By reducing the power as a function of the natural frequency of the electromechanical ultrasonic system (i.e., regulating temperature with feedback control), the temperature of the ultrasonic blade tip could be controlled directly. In this example, the temperature of the distal tip of the ultrasonic blade can be controlled to not exceed the melting point of the Teflon pad.

FIG. **46A** is a graphical representation **133050** of temperature versus time of an ultrasonic device without temperature feedback control. Temperature ($^{\circ}$ C.) of the ultrasonic blade is shown along the vertical axis and time (sec) is shown along the horizontal axis. The test was conducted with a chamois located in the jaws of the ultrasonic device. One jaw is the ultrasonic blade and the other jaw is the clamp arm with a TEFLON pad. The ultrasonic blade was excited at the resonant frequency while in frictional engagement with the chamois clamped between the ultrasonic blade and the clamp arm. Over time, the temperature ($^{\circ}$ C.) of the ultrasonic blade increases due to the frictional engagement with the chamois. Over time, the temperature profile **133052** of the ultrasonic blade increases until the chamois sample is cut after about 19.5 seconds at a temperature of 220° C. as indicated at point **133054**. Without temperature feedback control, after cutting the chamois sample, the temperature of the ultrasonic blade increases to a temperature well above the melting point of TEFLON $\sim 380^{\circ}$ C. up to $\sim 490^{\circ}$ C. At point **133056** the temperature of the ultrasonic blade reaches a maximum temperature of 490° C. until the TEFLON pad is completely melted. The temperature of the ultrasonic blade drops slightly from the peak temperature at point **133056** after the pad is completely gone.

FIG. **46B** is a plot of temperature versus time of an ultrasonic device with temperature feedback control, in accordance with at least one aspect of the present disclosure. Temperature ($^{\circ}$ C.) of the ultrasonic blade is shown along the vertical axis and the time (sec) is shown along the horizontal axis. The test was conducted with a chamois sample located in the jaws of the ultrasonic device. One jaw is the ultrasonic blade and the other jaw is the clamp arm with a TEFLON pad. The ultrasonic blade was excited at the resonant frequency while in frictional engagement with the chamois clamped between the ultrasonic blade and the clamp arm pad. Over time, the temperature profile **133062** of the ultrasonic blade increases until the chamois sample is cut after about 23 seconds at a temperature of 220° C. as indicated at point **133064**. With temperature feedback control, the temperature of the ultrasonic blade increases up to a maximum temperature of about 380° C., just below the melting point of TEFLON, as indicated at point **133066** and

then is lowered to an average of about 330° C. as indicated generally at region **133068**, thus preventing the TEFLON pad from melting.

Application of Smart Ultrasonic Blade Technology

When an ultrasonic blade is immersed in a fluid-filled surgical field, the ultrasonic blade cools down during activation rendering less effective for sealing and cutting tissue in contact therewith. The cooling down of the ultrasonic blade may lead to longer activation times and/or hemostasis issues because adequate heat is not delivered to the tissue. In order to overcome the cooling of the ultrasonic blade, more energy delivery may be required to shorten the transection times and achieve suitable hemostasis under these fluid immersion conditions. Using a frequency-temperature feedback control system, if the ultrasonic blade temperature is detected to, either start out below, or remain below a certain temperature for a certain period of time, the output power of the generator can be increased to compensate for cooling due to blood/saline/other fluid present in the surgical field.

Accordingly, the frequency-temperature feedback control system described herein can improve the performance of an ultrasonic device especially when the ultrasonic blade is located or immersed, partially or wholly, in a fluid-filled surgical field. The frequency-temperature feedback control system described herein minimizes long activation times and/or potential issues with ultrasonic device performance in fluid-filled surgical field.

As previously described, the temperature of the ultrasonic blade may be inferred by detecting the impedance of the ultrasonic transducer given by the following expression:

$$Z_g(t) = \frac{V_g(t)}{I_g(t)}$$

or equivalently, detecting the phase angle φ between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer. The phase angle φ information also may be used to infer the conditions of the ultrasonic blade. As discussed with particularity herein, the phase angle φ changes as a function of the temperature of the ultrasonic blade. Therefore, the phase angle φ information may be employed to control the temperature of the ultrasonic blade. This may be done, for example, by reducing the power delivered to the ultrasonic blade when the ultrasonic blade runs too hot and increasing the power delivered to the ultrasonic blade when the ultrasonic blade runs too cold. FIGS. **47A-47B** are graphical representations of temperature feedback control for adjusting ultrasonic power applied to an ultrasonic transducer when a sudden drop in temperature of an ultrasonic blade is detected.

FIG. **47A** is a graphical representation of ultrasonic power output **133070** as a function of time, in accordance with at least one aspect of the present disclosure. Power output of the ultrasonic generator is shown along the vertical axis and time (sec) is shown along the horizontal axis. FIG. **47B** is a graphical representation of ultrasonic blade temperature **133080** as a function of time, in accordance with at least one aspect of the present disclosure. Ultrasonic blade temperature is shown along the vertical axis and time (sec) is shown along the horizontal axis. The temperature of the ultrasonic blade increases with the application of constant power **133072** as shown in FIG. **47A**. During use, the temperature of the ultrasonic blade suddenly drops. This may result from a variety of conditions, however, during use, it may be inferred that the temperature of the ultrasonic blade drops

when it is immersed in a fluid-filled surgical field (e.g., blood, saline, water, etc.). At time t_0 , the temperature of the ultrasonic blade drops below the desired minimum temperature **133082** and the frequency-temperature feedback control algorithm detects the drop in temperature and begins to increase or “ramp up” the power as shown by the power ramp **133074** delivered to the ultrasonic blade to start raising the temperature of the ultrasonic blade above the desired minimum temperature **133082**.

With reference to FIGS. **47A** and **47B**, the ultrasonic generator is outputs substantially constant power **133072** as long the temperature of the ultrasonic blade remains above the desired minimum temperature **133082**. At t_1 , processor or control circuit in the generator or instrument or both detects the drop in temperature of the ultrasonic blade below the desired minimum temperature **133072** and initiates a frequency-temperature feedback control algorithm to raise the temperature of the ultrasonic blade above the minimum desired temperature **133082**. Accordingly, the generator power begins to ramp **133074** at t_1 corresponding to the detection of a sudden drop in the temperature of the ultrasonic blade at t_0 . Under the frequency-temperature feedback control algorithm, the power continues to ramp **133074** until the temperature of the ultrasonic blade is above the desired minimum temperature **133082**.

FIG. **48** is a logic flow diagram **133090** of a process depicting a control program or a logic configuration to control the temperature of an ultrasonic blade, in accordance with at least one aspect of the present disclosure. According to the process, the processor or control circuit of the generator or instrument or both executes one aspect of a frequency-temperature feedback control algorithm discussed in connection with FIGS. **47A** and **47B** to apply **133092** a power level to the ultrasonic transducer to achieve a desired temperature at the ultrasonic blade. The generator monitors **133094** the phase angle ϕ between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to drive the ultrasonic transducer. Based on the phase angle ϕ , the generator infers **133096** the temperature of the ultrasonic blade using the techniques described herein in connection with FIGS. **43A-45**. The generator determines **133098** whether the temperature of the ultrasonic blade is below a desired minimum temperature by comparing the inferred temperature of the ultrasonic blade to a predetermined desired temperature. The generator then adjusts the power level applied to the ultrasonic transducer based on the comparison. For example, the process continues along NO branch when the temperature of the ultrasonic blade is at or above the desired minimum temperature and continues along YES branch when the temperature of the ultrasonic blade is below the desired minimum temperature. When the temperature of the ultrasonic blade is below the desired minimum temperature, the generator increases **133100** the power level to the ultrasonic transducer, e.g., by increasing the voltage $V_g(t)$ and/or current $I_g(t)$ signals, to raise the temperature of the ultrasonic blade and continues increasing the power level applied to the ultrasonic transducer until the temperature of the ultrasonic blade increases above the minimum desired temperature.

Adaptive Advanced Tissue Treatment Pad Saver Mode

FIG. **49** is a graphical representation **133110** of ultrasonic blade temperature as a function of time during a vessel firing, in accordance with at least one aspect of the present disclosure. A plot **133112** of ultrasonic blade temperature is graphed along the vertical axis as a function of time along the horizontal axis. The frequency-temperature feedback control algorithm combines the temperature of the ultrasonic blade feedback control with the jaw sensing ability. The

frequency-temperature feedback control algorithm provides optimal hemostasis balanced with device durability and can deliver energy intelligently for best sealing while protecting the clamp arm pad.

As shown in FIG. **49**, the optimum temperature **133114** for vessel sealing is marked as a first target temperature T_1 and the optimum temperature **133116** for “infinite” clamp arm pad life is marked as a second target temperature T_2 . The frequency-temperature feedback control algorithm infers the temperature of the ultrasonic blade and maintains the temperature of the ultrasonic blade between the first and second target temperature thresholds T_1 and T_2 . The generator power output is thus driven to achieve optimal ultrasonic blade temperatures for sealing vessels and prolonging the life of the clamp arm pad.

Initially, the temperature of the ultrasonic blade increases as the blade heats up and eventually exceeds the first target temperature threshold T_1 . The frequency-temperature feedback control algorithm takes over to control the temperature of the blade to T_1 until the vessel transection is completed **133118** at t_0 and the ultrasonic blade temperature drops below the second target temperature threshold T_2 . A processor or control circuit of the generator or instrument or both detects when the ultrasonic blade contacts the clamp arm pad. Once the vessel transection is completed at t_0 and detected, the frequency-temperature feedback control algorithm switches to controlling the temperature of the ultrasonic blade to the second target threshold T_2 to prolong the life of the clam arm pad. The optimal clamp arm pad life temperature for a TEFLON clamp arm pad is approximately 325°C . In one aspect, the advanced tissue treatment can be announced to the user at a second activation tone.

FIG. **50** is a logic flow diagram **133120** of a process depicting a control program or a logic configuration to control the temperature of an ultrasonic blade between two temperature set points as depicted in FIG. **49**, in accordance with at least one aspect of the present disclosure. According to the process, the generator executes one aspect of the frequency-temperature feedback control algorithm to apply **133122** a first power level to the ultrasonic transducer, e.g., by adjusting the voltage $V_g(t)$ and/or the current $I_g(t)$ signals applied to the ultrasonic transducer, to set the ultrasonic blade temperature to a first target T_1 optimized for vessel sealing. As previously described, the generator monitors **133124** the phase angle ϕ between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer and based on the phase angle ϕ , the generator infers **133126** the temperature of the ultrasonic blade using the techniques described herein in connection with FIGS. **43A-45**. According to the frequency-temperature feedback control algorithm, a processor or control circuit of the generator or instrument or both maintains the ultrasonic blade temperature at the first target temperature T_1 until the transection is completed. The frequency-temperature feedback control algorithm may be employed to detect the completion of the vessel transection process. The processor or control circuit of the generator or instrument or both determines **133128** when the vessel transection is complete. The process continues along NO branch when the vessel transection is not complete and continues along YES branch when the vessel transection is complete.

When the vessel transection is not complete, the processor or control circuit of the generator or instrument or both determines **133130** if the temperature of the ultrasonic blade is set at temperature T_1 optimized for vessel sealing and transecting. If the ultrasonic blade temperature is set at T_1 , the process continues along the YES branch and the pro-

cessor or control circuit of the generator or instrument or both continues to monitor **133124** the phase angle φ between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer and based on the phase angle φ . If the ultrasonic blade temperature is not set at T_1 , the process continues along NO branch and the processor or control circuit of the generator or instrument or both continues to apply **133122** a first power level to the ultrasonic transducer.

When the vessel transection is complete, the processor or control circuit of the generator or instrument or both applies **133132** a second power level to the ultrasonic transducer to set the ultrasonic blade to a second target temperature T_2 optimized for preserving or extending the life of the clamp arm pad. The processor or control circuit of the generator or instrument or both determines **133134** if the temperature of the ultrasonic blade is at set temperature T_2 . If the temperature of the ultrasonic blade is set at T_2 , the process completes **133136** the vessel transection procedure.

Start Temperature of Blade

Knowing the temperature of the ultrasonic blade at the beginning of a transection can enable the generator to deliver the proper quantity of power to heat up the blade for a quick cut or if the blade is already hot add only as much power as would be needed. This technique can achieve more consistent transection times and extend the life of the clam arm pad (e.g., a TEFLON clamp arm pad). Knowing the temperature of the ultrasonic blade at the beginning of the transection can enable the generator to deliver the right amount of power to the ultrasonic transducer to generate a desired amount of displacement of the ultrasonic blade.

FIG. **51** is a logic flow diagram **133140** of a process depicting a control program or a logic configuration to determine the initial temperature of an ultrasonic blade, in accordance with at least one aspect of the present disclosure. To determine the initial temperature of an ultrasonic blade, at the manufacturing plant, the resonant frequencies of ultrasonic blades are measured at room temperature or at a predetermined ambient temperature. The baseline frequency values are recorded and stored in a lookup table of the generator or instrument or both. The baseline values are used to generate a transfer function. At the start of an ultrasonic transducer activation cycle, the generator measures **133142** the resonant frequency of the ultrasonic blade and compares **133144** the measured resonant frequency to the baseline resonant frequency value and determines the difference in frequency (Δf). The Δf is compared to the lookup table or transfer function for corrected ultrasonic blade temperature. The resonant frequency of the ultrasonic blade may be determined by sweeping the frequency of the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer. The resonant frequency is that frequency at which the phase angle φ voltage $V_g(t)$ and current $I_g(t)$ signals is zero as described herein.

Once the resonant frequency of the ultrasonic blade is determined, the processor or control circuit of the generator or instrument or both determines **133146** the initial temperature of the ultrasonic blade based on the difference between the measured resonant frequency and the baseline resonant frequency. The generator sets the power level delivered the ultrasonic transducer, e.g., by adjusting the voltage $V_g(t)$ or current $I_g(t)$ drive signals or both, to one of the following values prior to activating the ultrasonic transducer.

The processor or control circuit of the generator or instrument or both determines **133148** if the initial temperature of the ultrasonic blade is low. If the initial temperature of the ultrasonic blade is low, the process continues along

YES branch and the processor or control circuit of the generator or instrument or both applies **133152** a high power level to the ultrasonic transducer to increase the temperature of the ultrasonic blade and completes **133156** the vessel transection procedure.

If the initial temperature of the ultrasonic blade is not low, the process continues along NO branch and the processor or control circuit of the generator or instrument or both determines **133150** if the initial temperature of the ultrasonic blade is high. If the initial temperature of the ultrasonic blade is high, the process proceeds along YES branch and the processor or control circuit of the generator or instrument or both applies **133154** a low power level to the ultrasonic transducer to decrease the temperature of the ultrasonic blade and completes **133156** the vessel transection procedure. If the initial temperature of the ultrasonic blade is not high, the process continues along NO branch and the processor or control circuit of the generator or instrument or both completes **133156** the vessel transection.

Smart Blade Technology to Control Blade Instability

The temperature of an ultrasonic blade and the contents within the jaws of an ultrasonic end effector can be determined using the frequency-temperature feedback control algorithms described herein. The frequency/temperature relationship of the ultrasonic blade is employed to control ultrasonic blade instability with temperature.

As described herein, there is a well-known relationship between the frequency and temperature in ultrasonic blades. Some ultrasonic blades exhibit displacement instability or modal instability in the presence of increasing temperature. This known relationship may be employed to interpret when an ultrasonic blade is approaching instability and then adjusting the power level driving the ultrasonic transducer (e.g., by adjusting the driving voltage $V_g(t)$ or current $I_g(t)$ signals, or both, applied to the ultrasonic transducer) to modulate the temperature of the ultrasonic blade to prevent instability of the ultrasonic blade.

FIG. **52** is a logic flow diagram **133160** of a process depicting a control program or a logic configuration to determine when an ultrasonic blade is approaching instability and then adjusting the power to the ultrasonic transducer to prevent instability of the ultrasonic transducer, in accordance with at least one aspect of the present disclosure. The frequency/temperature relationship of an ultrasonic blade that exhibits a displacement or modal instability is mapped by sweeping the frequency of the drive voltage $V_g(t)$ or current $I_g(t)$ signals, or both, over the temperature of the ultrasonic blade and recording the results. A function or relationship is developed that can be used/interpreted by a control algorithm executed by the generator. Trigger points can be established using the relationship to notify the generator that an ultrasonic blade is approaching the known blade instability. The generator executes a frequency-temperature feedback control algorithm processing function and closed loop response such that the driving power level is reduced (e.g., by lowering the driving voltage $V_g(t)$ or current $I_g(t)$, or both, applied to the ultrasonic transducer) to modulate the temperature of the ultrasonic blade at or below the trigger point to prevent a given blade from reaching instability.

Advantages include simplification of ultrasonic blade configurations such that the instability characteristics of the ultrasonic blade do not need to be designed out and can be compensated using the present instability control technique. The present instability control technique also enables new ultrasonic blade geometries and can improve stress profile in heated ultrasonic blades. Additionally, ultrasonic blades can

be configured to diminish performance of the ultrasonic blade if used with generators that do not employ this technique.

In accordance with the process depicted by the logic flow diagram **133160**, the processor or control circuit of the generator or instrument or both monitors **133162** the phase angle φ between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer. The processor or control circuit of the generator or instrument or both infers **133164** the temperature of the ultrasonic blade based on the phase angle φ between the voltage $V_g(t)$ and current $I_g(t)$ signals applied to the ultrasonic transducer. The processor or control circuit of the generator or instrument or both compares **133166** the inferred temperature of the ultrasonic blade to an ultrasonic blade instability trigger point threshold. The processor or control circuit of the generator or instrument or both determines **133168** whether the ultrasonic blade is approaching instability. If not, the process proceed along the NO branch and monitors **133162** the phase angle φ , infers **133164** the temperature of the ultrasonic blade, and compares **133166** the inferred temperature of the ultrasonic blade to an ultrasonic blade instability trigger point threshold until the ultrasonic blade approaches instability. The process then proceeds along the YES branch and the processor or control circuit of the generator or instrument or both adjusts **133170** the power level applied to the ultrasonic transducer to modulate the temperature of the ultrasonic blade.

Ultrasonic Sealing Algorithm with Temperature Control

Ultrasonic sealing algorithms for ultrasonic blade temperature control can be employed to improve hemostasis utilizing a frequency-temperature feedback control algorithm described herein to exploit the frequency/temperature relationship of ultrasonic blades.

In one aspect, a frequency-temperature feedback control algorithm may be employed to alter the power level applied to the ultrasonic transducer based on measured resonant frequency (using spectroscopy) which relates to temperature, as described in various aspects of the present disclosure. In one aspect, the frequency-temperature feedback control algorithm may be activated by an energy button on the ultrasonic instrument.

It is known that optimal tissue effects may be obtained by increasing the power level driving the ultrasonic transducer (e.g., by increasing the driving voltage $V_g(t)$ or current $I_g(t)$, or both, applied to the ultrasonic transducer) early on in the sealing cycle to rapidly heat and desiccate the tissue, then lowering the power level driving the ultrasonic transducer (e.g., by lowering the driving voltage $V_g(t)$ or current $I_g(t)$, or both, applied to the ultrasonic transducer) to slowly allow the final seal to form. In one aspect, a frequency-temperature feedback control algorithm according to the present disclosure sets a limit on the temperature threshold that the tissue can reach as the tissue heats up during the higher power level stage and then reduces the power level to control the temperature of the ultrasonic blade based on the melting point of the clamp jaw pad (e.g., TEFLON) to complete the seal. The control algorithm can be implemented by activating an energy button on the instrument for a more responsive/adaptive sealing to reduce more the complexity of the hemostasis algorithm.

FIG. **53** is a logic flow diagram **133180** of a process depicting a control program or a logic configuration to provide ultrasonic sealing with temperature control, in accordance with at least one aspect of the present disclosure. According to the control algorithm, the processor or control circuit of the generator or instrument or both activates

133182 ultrasonic blade sensing using spectroscopy (e.g., smart blade) and measures **133184** the resonant frequency of the ultrasonic blade (e.g., the resonant frequency of the ultrasonic electromechanical system) to determine the temperature of the ultrasonic blade using a frequency-temperature feedback control algorithm (spectroscopy) as described herein. As previously described, the resonant frequency of the ultrasonic electromechanical system is mapped to obtain the temperature of the ultrasonic blade as a function of resonant frequency of the electromechanical ultrasonic system.

A first desired resonant frequency f_x of the ultrasonic electromechanical system corresponds to a first desired temperature Z° of the ultrasonic blade. In one aspect, the first desired ultrasonic blade temperature Z° is an optimal temperature (e.g., 450° C.) for tissue coagulation. A second desired frequency f_y of the ultrasonic electromechanical system corresponds to a second desired temperature ZZ° of the ultrasonic blade. In one aspect, the second desired ultrasonic blade temperature ZZ° is a temperature of 330° C., which is below the melting point of the clamp arm pad, which is approximately 380° C. for TEFLON.

The processor or control circuit of the generator or instrument or both compares **133186** the measured resonant frequency of the ultrasonic electromechanical system to the first desired frequency f_x . In other words, the process determines whether the temperature of the ultrasonic blade is less than the temperature for optimal tissue coagulation. If the measured resonant frequency of the ultrasonic electromechanical system is less than the first desired frequency f , the process continues along the NO branch and the processor or control circuit of the generator or instrument or both increases **133188** the power level applied to the ultrasonic transducer to increase the temperature of the ultrasonic blade until the measured resonant frequency of the ultrasonic electromechanical system exceeds the first desired frequency f_x . In which case, the tissue coagulation process is completed and the process controls the temperature of the ultrasonic blade to the second desired temperature corresponding to the second desired frequency f_y .

The process continues along the YES branch and the processor or control circuit of the generator or instrument or both decreases **133190** the power level applied to the ultrasonic transducer to decrease the temperature of the ultrasonic blade. The processor or control circuit of the generator or instrument or both measures **133192** the resonant frequency of the ultrasonic electromechanical system and compares the measured resonant frequency to the second desired frequency f_y . If the measured resonant frequency is not less than the second desired frequency f_y , the processor or control circuit of the generator or instrument or both decreases **133190** the ultrasonic power level until the measured resonant frequency is less than the second desired frequency f_y . The frequency-temperature feedback control algorithm to maintain the measured resonant frequency of the ultrasonic electromechanical system below the second desired frequency f_y , e.g., the temperature of the ultrasonic blade is less than the temperature of the melting point of the clamp arm pad then, the generator executes the increases the power level applied to the ultrasonic transducer to increase the temperature of the ultrasonic blade until the tissue transection process is complete **133196**.

FIG. **54** is a graphical representation **133200** of ultrasonic transducer current and ultrasonic blade temperature as a function of time, in accordance with at least one aspect of the present disclosure. FIG. **54** illustrates the results of the application of the frequency-temperature feedback control

algorithm described in FIG. 53. The graphical representation **133200** depicts a first plot **133202** of ultrasonic blade temperature as a function of time with respect to a second plot **133204** of ultrasonic transducer current $I_g(t)$ as a function of time. As shown, the transducer $I_g(t)$ is maintained constant until the ultrasonic blade temperature reaches 450° , which is an optimal coagulation temperature. Once the ultrasonic blade temperature reaches 450° , the frequency-temperature feedback control algorithm decrease the transducer current $I_g(t)$ until the temperature of the ultrasonic blade drops to below 330° , which is below the melting point of a TEFLON pad, for example.

Tissue Type Identification or Parameterization

In various aspects, a surgical instrument (e.g., an ultrasonic surgical instrument) is configured to identify or parameterize the tissue grasped by the end effector and adjust various operational parameters of the surgical instrument accordingly. The identification or parameterization of the tissue can include the tissue type (e.g., the physiological tissue type), physical characteristics or properties of the tissue, the composition of the tissue, the location of the tissue within or relative to the end effector, and so on. In one example discussed in greater detail below, the ultrasonic surgical instrument is configured to tune the displacement amplitude of the distal tip of the ultrasonic blade according to the collagen/elastin ratio of the tissue detected in the jaws of the end effector. As previously discussed, an ultrasonic instrument comprises an ultrasonic transducer acoustically coupled to an ultrasonic blade via an ultrasonic waveguide. The displacement of the ultrasonic blade is a function of the electrical power applied to the ultrasonic transducer and, accordingly, the electrical power supplied to the ultrasonic transducer can be modulated according to the detected collagen/elastin ratio of the tissue. In another example discussed in greater detail below, the force exerted by the clamp arm on the tissue can be modulated according to the location of the tissue relative to the end effector. Various techniques for identifying or parameterizing the tissue are described herein and further details can be found in, for example, U.S. Provisional Patent Application No. 62/692,768, titled SMART ENERGY DEVICES, filed on Jun. 30, 2018, the disclosure of which is herein incorporated by reference in its entirety.

Determining Tissue Location Via Impedance Change

Referring back to FIG. 23, there is illustrated an end effector **1122** comprising an ultrasonic blade **1128** and a clamp arm **1140**, in accordance with at least one aspect of the present disclosure. FIG. 55 is a bottom view of an ultrasonic end effector **1122** showing a clamp arm **1140** and ultrasonic blade **1128** and delineating tissue positioning within the ultrasonic end effector **1122**, in accordance with at least one aspect of the present disclosure. The positioning of the tissue between the clamp arm **1140** and ultrasonic blade **1128** can be delineated according to the region or zone in which the tissue is located, such as a distal region **130420** and a proximal region **130422**.

With reference now to FIGS. 23 and 55, as described herein, the ultrasonic end effector **1122** grasps tissue between the ultrasonic blade **1128** and clamp arm **1140**. Once tissue is grasped, the ultrasonic generator (e.g., the generator **1100** described in connection with FIG. 22) may be activated to apply power to the ultrasonic transducer, which is acoustically coupled to the ultrasonic blade **1128** via an ultrasonic waveguide. The power applied to the ultrasonic transducer may be in a therapeutic or non-therapeutic range of energy levels. In a non-therapeutic range of applied power, the resulting displacement of the ultrasonic

blade **1128** does not effect, or minimally effects, the grasped tissue so as not to coagulate or cut the tissue. The non-therapeutic excitation may be particularly useful for determining the impedance of the ultrasonic transducer, which will vary based on a variety of conditions present at the end effector **1122**, including, for example, tissue type, tissue location within the end effector, ratio of different tissue types, and temperature of the ultrasonic blade, among other conditions. A variety of these conditions are described herein. The impedance of the ultrasonic transducer is given by

$$Z_g(t) = \frac{V_g(t)}{I_g(t)},$$

as described herein. Once the conditions at the ultrasonic end effector **1122** are determined using non-therapeutic ultrasonic energy levels, therapeutic ultrasonic energy may be applied based on the determined end effector **1122** conditions to optimize the tissue treatment, effective seal, transection, and duration, among other variables associated with a particular surgical procedure. Therapeutic energy is sufficient to coagulate and cut tissue.

In one aspect, the present disclosure provides a control process, such as an algorithm, to determine the thickness and type of tissue located within the jaws (i.e., between the clamp arm **1140** and the ultrasonic blade **1128**) of an ultrasonic end effector **1122** as shown in FIGS. 23 and 55. Additional detail regarding detecting various states and properties of objects grasped by an end effector **1122** are discussed below under the heading DETERMINING JAW STATE and in U.S. Provisional Patent Application No. 62/692,768, titled SMART ENERGY DEVICES.

FIG. 56 is a graphical representation **130000** depicting change in ultrasonic transducer impedance as a function tissue location within the ultrasonic end effector **1122** over a range of predetermined ultrasonic generator power level increases, in accordance with at least one aspect of the present disclosure. The horizontal axis **130004** represents tissue location and the vertical axis **130002** represents transducer impedance (Ω). Various limits along the horizontal axis **130004**, such as a first or proximal limit **130010** and a second or distal limit **130012**, can delineate or correspond to different positions of the tissue grasped within the ultrasonic end effector **1122**. The delineation of proximal and distal tissue locations is shown schematically in FIG. 55 (i.e., proximal portion **130422** and distal portion **130420**). The plots **130006**, **130008** represent the change in transducer impedance Q as the power applied to the ultrasonic transducer is varied from a minimum or first non-therapeutic power level L_1 to a maximum or second non-therapeutic power level L_2 . The larger the change in transducer impedance Ω is, the closer the resulting plot will be to the distal limit **130012**. Accordingly, the location of the tissue corresponds to the position of the resulting plot relative to the various limits (e.g., the proximal limit **130010** and the distal limit **130012**). In the first plot **130006**, δ_1 represents the change in transducer impedance when tissue is located at the proximal portion **130422** of the end effector **1122**. This can be seen from the fact that the first plot **130006** does not exceed the proximal limit **130010**. In the second plot **130008**, δ_2 represents the change in transducer impedance when tissue is located at the distal end **130012** of the end effector **1122**. This can be seen from the fact that the first plot **130006** exceeds the proximal limit **130010** and/or is located

near to the distal limit **130012**. As indicated by the plots **130006**, **130008**, δ_2 is much greater than δ_1 .

When applying power (voltage and current) to the ultrasonic transducer to activate the ultrasonic blade **1128** in the non-therapeutic range (for example, power that is not sufficient for cutting or coagulating tissue), the resulting measured transducer impedance (Ω) is a useful indicator of the position of tissue within the jaw of the end effector **1122**, whether at the distal end **130420** or proximal end **130422** of the ultrasonic blade **1128** as shown in FIG. **55**. The location of the tissue within the end effector **1122** can be determined based on the change in transducer impedance δ as the non-therapeutic power level applied to the ultrasonic transducer is varied from a minimum power level (e.g., L_1) to a maximum power level (e.g., L_2). In some aspects, the non-therapeutic power level(s) applied to the ultrasonic transducer can cause the ultrasonic blade **1128** to oscillate at a sensing amplitude or below the minimum therapeutic amplitude (e.g., less than or equal to $35\ \mu\text{m}$ at the distal and/or proximal end of the ultrasonic blade **1128**). Calculation of impedance is discussed previously in this disclosure. A measure of the first transducer impedance Z_1 is taken when a first power level L_1 is applied, which provides an initial measurement, and a subsequent measure of impedance Z_2 is taken again when the applied power is increased to a second power level L_2 . In one aspect, the first power level $L_1=0.2\ \text{A}$ and the second power level $L_2=0.4\ \text{A}$ or twice the first power level L_1 , while the voltage is maintained constant. The resulting longitudinal displacement amplitude of the ultrasonic blade **1128**, based on the applied power level, provides an indication of tissue location within the jaws of the end effector **1122**. In one example implementation, the first power level L_1 produces a longitudinal displacement amplitude of $35\ \mu\text{m}$ at the distal end **130420** and $15\ \mu\text{m}$ at the proximal end **130422**. Further in this example, the second power level L_2 produces a longitudinal amplitude of $70\ \mu\text{m}$ at the distal end **130420** and $35\ \mu\text{m}$ at the proximal end **130422**. An algorithm calculates the difference in transducer impedance δ between the first and second measurements to find the change in impedance $\Delta Z_g(t)$. The change in impedance δ is plotted against tissue location and shows that a higher change in impedance represents tissue location distributed at the distal end **130012** and a lower change in impedance represents tissue location distributed at the proximal end **130010** of the end effector **1122**. In sum, if there is a large change in impedance as the power level is increased from L_1 to L_2 , then the tissue is distally positioned only within the end effector **1122**; conversely, if there is only a small change in impedance as the power level is increased from L_1 to L_2 , then the tissue is more distributed within the end effector **1122**.

FIG. **57** is a graphical representation **130050** depicting change in ultrasonic transducer impedance as a function of time relative to the location of tissue within the ultrasonic end effector, in accordance with at least one aspect of the present disclosure. The horizontal axis **130054** represents time (t) and the vertical axis **130052** represents change in transducer impedance (δ) between the first and second measurements. The plots **130060**, **130066** depict the change in transducer impedance (δ) versus time (t), relative to the proximal and distal location of tissue within the bite of the clamp arm **1140**. For proximal and distal tissue location, a clamp arm **1140** force is applied to hold the tissue in the ultrasonic end effector **1122** and a delay period is applied before a first, low-power level is applied and the transducer impedance is measured. Subsequently, the system applies a second, higher power level and measures the impedance

again. It will be appreciated that both the first and second power levels applied to the ultrasonic transducer are non-therapeutic power levels. The algorithm executed by a processor or control circuit portion of the generator, or the surgical instrument, (e.g., the processor **902** in FIG. **21** or the control circuit **760** in FIG. **18**) calculates the difference in transducer impedance (δ) between the first power level and the second power level for the proximal and distal tissue locations. As shown in relation to the first plot **130060**, if the difference in transducer impedance (δ) below a first threshold **130056**, the algorithm determines that the tissue is located in the proximal portion **130422** of the end effector **1122**. In the first plot **130060**, the difference in the transducer impedance between the measurements increases **130062** over time until it plateaus or maintains **130064** below the first threshold **130056**. As shown in relation to the first plot **130066**, if the difference in transducer impedance (δ) is above a second threshold **130058**, the algorithm determines that the tissue is located in the distal portion **130420** of the end effector **1122**. In the second plot **130066**, the difference in the transducer impedance between the measurements increases **130068** over time until it plateaus or maintains **130070** above the second threshold **130058**. If the difference in transducer impedance (δ) is between the first and second thresholds **130056**, **130058**, the algorithm determines that the tissue is located in the middle portion **130424** of the end effector **1122**, e.g., between the proximal and distal portions of the end effector.

FIG. **58** is a logic flow diagram of a process **130100** depicting a control program or a logic configuration to identify operation in the non-therapeutic range of power applied to the instrument in order to determine tissue positioning, in accordance with at least one aspect of the present disclosure. The process **130100** can be executed by a processor or control circuit of a surgical instrument, such as the control circuit **760** of FIG. **18**, or a generator, such as the processor **902** of FIG. **21**. For conciseness, the process **130100** will be described as being executed by a processor, but it should be understood that the following description encompasses the aforementioned variations.

In accordance with one aspect of the process **130100**, a processor applies a control signal to close the clamp arm **1140** to capture the tissue between the clamp arm **1140** and the ultrasonic blade **1128**. After the clamp arm **1140** closes onto the tissue, the processor waits for a predetermined delay period to allow the tissue to relax and give up some moisture content. After the delay period, the processor sets **130102** the power level applied to ultrasonic transducer to a first non-therapeutic power level. Optionally, one aspect of the process **130100** includes a feedback control can be used to verify that the first power is set below a therapeutic power level. In this aspect, the processor determines **130106** whether the first power level is less than a therapeutic power level. If the first power level is not less than a therapeutic power level, the process **130100** continues along the NO branch and the processor decreases **130108** the applied power and loops until the first power level is less than a therapeutic power level. The process **130100** then continues along the YES branch and the processor measures **130110** a first impedance $Z_g(t)$ of the ultrasonic transducer corresponding to the first power level. The processor then sets **130112** the power level applied to ultrasonic transducer to a second non-therapeutic power level, where the second power is greater than the first power level and is below a therapeutic power level. Again, optionally, a feedback control can be used to verify that the second power level is not only greater than the first power level but also is below a

therapeutic power level. In this aspect, the processor determines **130114** whether the second power level is less than a therapeutic power level. If the second power is greater than a therapeutic power level, the process **130100** continues along the NO branch and the processor decreases **130108** the second power level and loops until it is below a therapeutic power level threshold. The process **130100** then continues along the YES branch and the processor measures **130116** a second impedance $Z_{g2}(t)$ of the ultrasonic transducer corresponding to the second power level. The impedance of the ultrasonic transducer can be measured using a variety of techniques as discussed herein. The processor then calculates **130118** the difference in transducer impedance between the applied first and second power levels:

$$\delta = Z_{g2}(t) - Z_{g1}(t).$$

The processor then provides **130120** an indication of the tissue position to the user. The processor can indicate the tissue position via an output device of a surgical instrument (e.g., visual feedback devices, such as the display depicted in FIG. 31, audio feedback devices, and/or tactile feedback devices), a display **135** (FIG. 3), or other output device of a surgical hub **106** communicably connected to the surgical instrument and/or an output device **2140** (FIG. 27B) of a generator **1100** (e.g., visual feedback devices, audio feedback devices, and/or tactile feedback devices).

The processor compares the difference in transducer impedance to a first and second threshold where, as shown in FIG. 57, if the difference in transducer impedance (δ) below a first threshold **130056**, the algorithm determines that the tissue is located in the proximal portion **130422** of the end effector **1122** and if the difference in transducer impedance (δ) is above a second threshold **130058**, the algorithm determines that the tissue is located in the distal portion **130420** of the of the end effector **1122**. If the difference in transducer impedance (δ) is between the first and second thresholds **130056**, **130058**, the algorithm determines that the tissue is located in the middle portion **130424** of the end effector **1122**, e.g., between the proximal and distal portions **130422**, **130420** of the end effector **1122**. According to the described process, the impedance of the ultrasonic transducer can be employed to differentiate what percent of tissue is located in a distal, proximal, or intermediate location of the end effector **1122** and then apply a suitable therapeutic power level.

Switchless Mode Based on Tissue Positioning

In various aspects, the reactions of the ultrasonic instrument may be based on whether tissue is present within the end effector, the type of tissue located in the end effector, or the compressibility or composition of the tissue located in the end effector. Accordingly, the generator or the ultrasonic surgical instrument may contain and/or execute instructions to perform algorithms to control the time between clamping the tissue in the jaws of the end effector and the activation of the ultrasonic transducer to treat the tissue. If tissue is not sensed, the ultrasonic generator activation buttons or pedals may be assigned different meanings to execute different functions. In one aspect, an advanced energy device may employ the detection of the presence of tissue within the jaws of the end effector as the queue for activating the ultrasonic transducer, thereby starting the tissue coagulation cycle. In another aspect, the compressive properties and situational awareness may enable automatic activation of the device to also adjust the parameters of the algorithm for the type of tissue sensed. For example, the advanced generator may ignore activations of buttons or foot pedals unless tissue is sensed in contact with the jaws of the end effector. This

configuration would eliminate inadvertent activation queues that would allow the device to be operated in a simpler manner.

Accordingly, an advanced generator, such as advanced generators described in connection with FIGS. 1-42, and/or surgical instruments, such as ultrasonic surgical instruments described throughout this disclosure, may be configured to operate in a switchless mode. In a switchless mode, the ultrasonic device is automatically activated in coagulation mode upon sensing or detecting the presence of tissue in the jaws of the end effector. In one aspect, when operating in automatic energy activation (or "switchless" mode) mode, control algorithm controlling the activation of the ultrasonic surgical instrument can be configured to apply less energy initially to the ultrasonic instrument than if activated when not operating in the switchless mode. Further, the ultrasonic generator or instrument can be configured to determine both contact with and the type of tissue located in the jaws of the end effector. Based on sensing or detecting the presence of tissue in the jaws of the end effector, control algorithms executed either by a processor or control circuit of the generator or ultrasonic instrument could run the ultrasonic instrument in switchless mode and could adjust the algorithm to achieve the best overall coagulation of the tissue in the jaws of the end effector. In other aspects, in lieu of automatically activating the surgical instrument and/or generator, control algorithms executed by a processor or control circuit of the generator or ultrasonic instrument could prevent activation of the generator or ultrasonic instrument unless the presence of tissue is detected within the end effector.

In one aspect, the present disclosure provides an algorithm executed by a processor or control circuit located either in the generator or handheld ultrasonic instrument to determine the presence of tissue and the type of tissue located within the jaws of the end effector. In one aspect, the control algorithm can be configured to determine that tissue is located within the end effector via techniques described herein for determining tissue location as described below under the heading DETERMINING TISSUE LOCATION VIA ELECTRODE CONTINUITY. For example, a control algorithm can be configured to determine whether tissue is located within the end effector according to whether there is any continuity between electrodes (as described below) and, accordingly, activate the surgical instrument (e.g., by causing the generator to which the surgical instrument is coupled to begin applying power to the surgical instrument) automatically upon detecting tissue or permit activation of the surgical instrument. When the surgical instrument and/or generator is being operated in switchless mode, the control algorithm can further be configured to activate the surgical instrument at a particular power level, which may or may not be different that a standard initial activation power level for the surgical instrument. In some aspects, a control algorithm can be configured to activate or permit activation of the surgical instrument according to the particular type or composition of the tissue, which can be detected via, e.g., techniques described below under the heading DETERMINING TISSUE COLLAGEN-TO-ELASTIN RATIO ACCORDING TO IR SURFACE REFLECTANCE AND EMISSIVITY. For example, the control algorithm can be configured to activate the surgical instrument when tissue having a high collagen composition has been grasped, but not necessarily activate the surgical instrument when tissue having a high elastin composition has been grasped. In some aspects, a control algorithm can be configured to activate or permit activation of the surgical instrument according to

whether the grasped tissue is located at a particular position within the end effector or whether a particular amount of tissue has been grasped by the end effector via, e.g., techniques described below under the heading DETERMINING TISSUE LOCATION VIA ELECTRODE CONTINUITY. For example, the control algorithm can be configured to activate the surgical instrument when the grasped tissue covers a particular percentage of the end effector. As another example, the control algorithm can be configured to activate the surgical instrument when the grasped tissue is located at the distal end of the end effector.

In other aspects, a control algorithm can be configured to determine whether tissue has been grasped by the end effector, the tissue type or composition, and other characteristics of the end effector or the tissue via a situational awareness system, as described in U.S. Provisional Patent Application Ser. No. 62/659,900, titled METHOD OF HUB COMMUNICATION, filed Apr. 19, 2018, which is hereby incorporated by reference in its entirety, and under the heading SITUATIONAL AWARENESS. In these aspects, a surgical hub **106** (FIGS. **1-11**) to which the surgical instrument and/or generator is connected can receive data from the surgical instrument, generator, and/or other medical devices utilized in the operating theater and make inferences about the surgical procedure, or a particular step thereof, being performed. Accordingly, the situational awareness system can infer whether and what type(s) of tissue are being operated on at any given instant or step and then a control algorithm can control the surgical instrument accordingly, including automatically activating the surgical instrument accordingly. For example, the control algorithm could be configured to automatically activate or permit activation of the surgical instrument when the tissue grasped by the end effector corresponds to the tissue type or tissue composition expected by the situational awareness system.

With the ability to detect whether or not the instrument is contacting tissue, and what type of tissue when in contact, the ultrasonic instrument can be operated in a switchless mode of operation where operation is permitted based on the sensing ability of the ultrasonic instrument. In some aspects, a control algorithm can be configured to ignore actuations of activation buttons, foot pedals, and other input devices coupled to the generator and/or the ultrasonic surgical instrument unless tissue is sensed in contact with the jaws/end effector of the surgical instrument, thereby preventing unintended activations of the instrument. In some aspects, a control algorithm can be configured to assign different meanings to the inputs of activation buttons, foot pedals, and other input devices coupled to the generator and/or the ultrasonic surgical instrument according to whether tissue is sensed in contact with the jaws/end effector of the surgical instrument. For example, when tissue is present in the end effector, a control algorithm can be configured to activate the surgical instrument in response to an activation button being actuated; however, when tissue is not present within the end effector, the control algorithm can be configured to execute some different or secondary action when the activation button is actuated.

The ability to determine the lack of tissue present in the jaws of the end effector acts as a permissive to allow the instrument to change to switchless mode and then initiate an automatic coagulation cycle of operation when tissue is then detected, leading to greater uptime use of the instrument and allowing the user to proceed based on its predictive capability. The ability to further detect tissue type in addition to detecting the presence of tissue allows the algorithm to adjust and calculate the best coagulation opportunity.

Tuning an Ultrasonic System According to Tissue Composition

In various aspects, an ultrasonic surgical instrument can include a processor or control circuit executing an adaptive ultrasonic blade control algorithm for detecting the composition of the tissue grasped by or at the end effector and controlling operational parameters of the ultrasonic transducer and/or ultrasonic blade accordingly. The tissue composition can include, for example, the ratio of collagen to elastin in the tissue, the stiffness of the tissue, or the thickness of the tissue. The operational parameters controlled or regulated by the adaptive ultrasonic blade control algorithm can include, for example, the amplitude of the ultrasonic blade, the temperature or heat flux of the ultrasonic blade, and so on. The adaptive ultrasonic blade control algorithm can be executed by a control circuit or processor located either in the generator or the surgical instrument.

In one example described in further detail below, the adaptive ultrasonic blade control algorithm can be configured to control the amplitude of the ultrasonic blade according to the collagen-to-elastin ratio of the tissue. The collagen-to-elastin ratio of the tissue can be determined via a variety of techniques, such as those described below. In another example described in further detail below, the adaptive ultrasonic blade control algorithm can be configured to control the ultrasonic transducer/ultrasonic blade to provide a longer warming time and a lower end temperature of the ultrasonic blade the lower the collagen content of the tissue is.

Determining Tissue Collagen-to-Elastin Ratio According to Frequency Shift

In various aspects, a control algorithm can be configured to determine the collagen-to-elastin ratio of a tissue (e.g., to tune the amplitude of the distal tip of an ultrasonic blade) by detecting the natural frequency of an ultrasonic blade and the shifts in the ultrasonic blade waveform. For example, the techniques described in connection with FIGS. **1-54** may be employed to detect the ratio of collagen to elastin of the tissue located in an end effector of an ultrasonic instrument. In one aspect, the present disclosure provides an adaptive ultrasonic blade control algorithm to detect the natural frequency of the ultrasonic blade and shift in waveform to detect the composition of the tissue in contact with the ultrasonic blade. In another aspect, the adaptive ultrasonic blade control algorithm may be configured to detect the collagen and elastin composition content of the tissue and adjust the therapeutic heat flux of the ultrasonic blade based on the detected collagen content of the tissue. Techniques for monitoring the deviation of the natural frequency of the ultrasonic blade based on the tissue type located in the jaws of the end effector of the ultrasonic instrument are described herein in connection with FIGS. **1-54**. Accordingly for conciseness and clarity of disclosure, such techniques will not be repeated here.

The ratio of elastin to collagen may be determined by monitoring the shift in the natural frequency of the ultrasonic blade and comparing the natural frequency to a look-up table. The look-up table can be stored in memory (e.g., memory **3326** of FIG. **31**) and contains the ratio of elastin to collagen and a corresponding natural frequency shift for a particular ratio as determined empirically. Determining Tissue Collagen-to-Elastin Ratio According to IR Surface Reflectance and Emissivity

In various aspects, a control algorithm can be configured to determine the collagen-to-elastin ratio of a tissue (e.g., to tune the amplitude of the distal tip of an ultrasonic blade) by determining the IR reflectivity of the tissue. For example,

FIG. 59 illustrates an ultrasonic system 130164 comprising an ultrasonic generator 130152 coupled to an ultrasonic instrument 130150. The ultrasonic instrument 130150 is coupled to an ultrasonic end effector 130400 via an ultrasonic waveguide 130154. The ultrasonic generator 130152 may be integral with the ultrasonic instrument 130150 or may be connected to the ultrasonic instrument 130150 using wired or wireless electrical/electronic coupling techniques. The end effector 130400 of the ultrasonic surgical instrument 130150 comprises IR sensors located on the clamp arm 130402 (e.g., jaw member), in accordance with at least one aspect of the present disclosure. The ultrasonic generator 130152 and/or the ultrasonic instrument 130150 may be coupled to the surgical hub 130160 and/or the cloud 130162 over wireless or wired connections, as described in connection with FIGS. 1-11.

FIG. 60 illustrates an IR reflectivity detection sensor circuit 130409 that may be mounted or formed integrally with the clamp arm 130402 of the ultrasonic end effector 130400 to detect tissue composition, in accordance with at least one aspect of the present disclosure. The IR sensor circuit 130409 includes an IR source 130416 (e.g., an IR transmitter) and IR detector 130418 (e.g., an IR receiver). The IR source 130416 is coupled to a voltage source V. A current is generated through R2 when the control circuit 130420 closes the switch SW1. When the switch SW1 is closed, the IR source 130416 emits IR energy towards the tissue 130410 (e.g., tissue clamped or situated between the clamp arm 130402 and the ultrasonic blade 130404). Some of the emitted IR energy is absorbed by the tissue 130410, some of the emitted IR energy is transmitted through the tissue 130410, and some of the emitted IR energy is reflected by the tissue 130410. The IR detector 130418 receives the IR energy reflected by tissue 130410 and generates an output voltage V_o or signal, which is applied to the control circuit 130420 for processing.

With reference to FIGS. 59 and 60, in one aspect, the ultrasonic generator 130152 includes the control circuit 130420 to drive the IR source 130416 and IR detector 130418 located on or in the clamp arm 130402 of the ultrasonic end effector 130400. In other aspects, the ultrasonic instrument 130150 includes the control circuit 130420 to drive the IR source 130416 and IR detector 130418 located on or in the clamp arm 130402 of the ultrasonic end effector 130400. In either aspect, when tissue 130410 is grasped between the ultrasonic blade 130404 and the clamp arm 130402, the IR source 130416 is energized by the control circuit 130420 by closing switch SW1, for example, to illuminate the tissue with IR energy. In one aspect, the IR detector 130418 generates a voltage V_o that is proportional to the IR energy reflected by the tissue 130410. The total IR energy emitted by the IR source 130416 equals the sum of the IR energy reflected by the tissue 130410, the IR energy absorbed by the tissue 130410, and the IR energy that passes through the tissue 130410, plus any losses. Accordingly, the control circuit 130420 or processor may be configured to detect the collagen content of the tissue 130410 by the amount of IR energy detected by the IR detector 130418 relative to the total amount of IR energy emitted by the IR source 130416. An algorithm takes into account the amount of energy absorbed by and/or transmitted through the tissue 130410 to determine the collagen content of the tissue 130410. The IR source 130416 and IR detector 130418 and algorithms are calibrated to provide useful measurements of the collagen content of the tissue 130410 using the principles of IR reflectivity.

The IR reflectivity detection sensor circuit 130409 shown in FIG. 60 provides IR surface reflectance and emissivity to determine the ratio of elastin to collagen. The IR reflectance may be employed to determine tissue composition for tuning the amplitude of an ultrasonic transducer. The refractive index is an optical constant that controls the light-related reflection of IR light. The refractive index may be employed to differentiate tissue types. For example, the refractive index contrast has been shown to differentiate between normal liver tissue and hepatic metastases. The refractive index could be used as an absolute or a comparative measure for tissue differentiation.

A comparative method employs an energy dissection device, such as an ultrasonic blade 130404, for example, to determine the exact ratio (as detailed above) and then to predict the collagen ratio for all further actuation using that index as a baseline. In this manner, an endoscope may update the dissection device (e.g., ultrasonic blade 130404) based on collagen ratio. The dissection device can fine-tune the forecasting each time it is actuated to make an actual collagen denaturing firing. An alternative method may employ an absolute index with a look-up table that can differentiate between surface irregularities and sub-surface collagen concentration. Additional information about IR refractive index properties of tissue can be found in Visible To Near-Infrared Refractive Properties Of Freshly-Excised Human-Liver Tissues: Marking Hepatic Malignancies; Panagiotis Giannios, Konstantinos G. Toutouzas, Maria Matiatou, Konstantinos Stasinou, Manousos M. Konstadoulakis, George C. Zografos, and Konstantinos Moutzouris; Sci. Rep. 2016; 6: 27910, which is hereby incorporated by reference herein.

In other aspects, the ultrasonic dissection device can be configured to change the ideal temperature of the ultrasonic blade control algorithm proportionately with the collagen ratio. For example, the ultrasonic blade temperature control algorithm can be modified based on the collagen ratio received from the control circuit 130420. As one specific example, the ultrasonic blade temperature control algorithm can be configured to lower the set of temperatures at which the ultrasonic blade 130404 is maintained and increase the hold time that the ultrasonic blade 130404 is contacted with the tissue 130410 for higher concentrations of collagen in the grasped tissue 130410. As another example, the wait time for the algorithm to cycle through a full activation could be modified based on the collagen ratio. Various temperature control algorithms for ultrasonic blades are described in connection with FIGS. 43-54.

FIG. 61 is a sectional view of an ultrasonic end effector 130400 comprising a clamp arm 130402 and an ultrasonic blade 130404 according to one aspect of the present disclosure. The clamp arm 130402 comprises IR reflectivity detection sensor circuits 130409a, 130409b that may be mounted or formed integrally with the clamp arm 130402 of the ultrasonic end effector 130400 to detect the composition of the tissue 130410. The IR reflectivity detection sensor circuits 130409a, 130409b may be mounted on a flexible circuit substrate 130412, which is shown in plain view in FIG. 62. The flexible circuit substrate 130412 includes three elongate elements 130408a, 130408b, 130408c on which the IR reflectivity detection sensor circuits 130409a, 130409b and IR sensors 130414a, 130414b are mounted. The IR sensors 130414a, 130414b may include IR sources 130416 and IR detectors 130418 as shown in FIG. 60.

FIG. 63 is a logic flow diagram of a process 130200 depicting a control program or a logic configuration to measure IR reflectance to determine tissue composition to

tune the amplitude of the ultrasonic transducer. The process **130200** can be executed by a processor or control circuit of a surgical instrument, such as the control circuit **760** of FIG. **18**, or a generator, such as the processor **902** of FIG. **21**. For conciseness, the process **130200** will be described as being executed by a control circuit, but it should be understood that the following description encompasses the aforementioned variations.

Accordingly, with reference to FIGS. **1-54** and FIGS. **59-63**, in one aspect, the control circuit energizes **130202** the IR source **130416** to apply IR energy to tissue **130410** clamped in an end effector **13400** of an ultrasonic instrument **130150**. The control circuit then detects **130204**, via an IR detector **130418**, the IR energy reflected by the tissue **130410**. Accordingly, the control circuit determines **130206** the ratio of collagen to elastin of the tissue **130410** based on the detected IR energy reflected by the tissue **130410**. The control circuit adjusts **130208** the ultrasonic blade temperature control algorithm, as discussed in U.S. Provisional Patent Application No. 62/692,768, titled SMART ENERGY DEVICES, based on the determined collagen-to-elastin ratio of the tissue. In one aspect, the collagen content of the tissue **130410** may be detected by according to the reflectivity of an IR light source **130416**. In another aspect, the lower the collagen content of the tissue **130410**, the longer the warming time and the lower the end temperature of the ultrasonic blade **130404**. In yet another aspect, the tissue **130410** composition may be tissue thickness or stiffness and could be used to affect the ultrasonic blade transducer control program.

The ratio of elastin to collagen may be determined by monitoring the IR reflectance of the tissue and comparing the detected IR reflectance to a look-up table. The look-up table can be stored in memory (e.g., memory **3326** of FIG. **31**) and contains the ratio of elastin to collagen and a corresponding IR reflectance for a particular ratio as determined empirically.

Determining Tissue Collagen-to-Elastin Ratio According to Collagen Transformation Point

Different types of tissues are composed of varying amounts of structural proteins, such as collagen and elastin, which provide the different types of tissue with different properties. As heat is applied to the tissue (e.g., by an ultrasonic blade), the structural proteins denature, which affects the tissue integrity and other tissue properties. However, the structural proteins denature at different known temperatures. For example, collagen denatures prior to elastin. Accordingly, by detecting at what temperature the properties of the tissue change, one can infer the tissue composition (e.g., the ratio of collagen to elastin in the tissue). In various aspects, a control algorithm can be configured to determine the collagen-to-elastin ratio of a tissue by determining the collagen transformation point of the tissue. The control algorithm can, in turn, control various operational parameters of the surgical instrument, such as the amplitude of the ultrasonic blade, according to the determined tissue composition. In one aspect, the control algorithm can determine the collagen transformation point of the tissue by measuring the position of the clamp arm actuation member and the rate of change of its displacement while maintaining the load on the clamp arm constant. In another aspect, the control algorithm can determine the collagen transformation point of the tissue by measuring the temperature of the tissue/blade interface directly to identify the collagen/elastin percentage.

FIGS. **16-19** illustrate schematically motorized clamp arm closure mechanisms. FIG. **40** is a system diagram **7400** of a

segmented circuit **7401** comprising a plurality of independently operated circuit segments **7402**, **7414**, **7416**, **7420**, **7424**, **7428**, **7434**, **7440**, according to one aspect of the present disclosure, and FIG. **35** is a circuit diagram of various components of a surgical instrument with motor control functions, according to one aspect of the present disclosure. For example, FIG. **35** illustrates a drive mechanism **7930**, including a closure drivetrain **7934** configured to close a jaw member to grasp tissue with the end effector. FIGS. **38-39** illustrate control systems **12950**, **12970** for controlling the rate of closure of the jaw member, such as a clamp arm portion of an ultrasonic end effector, where FIG. **38** is a diagram of a control system **12950** configured to provide progressive closure of a closure member as it advances distally to close the clamp arm to apply a closure force load at a desired rate according to one aspect of this disclosure, and FIG. **39** illustrates a proportional-integral-derivative (PID) controller feedback control system **12970** according to one aspect of this disclosure. Accordingly, in the following description of an ultrasonic system comprising a motorized clamp arm controller to control the closure rate and/or position of the clamp arm, reference should be made to FIGS. **16-19** and **38-41**.

In one aspect, a control algorithm can be configured to detect the collagen transformation point of a grasped tissue and accordingly control the delivery of ultrasonic energy to the tissue by controlling the phase and/or amplitude of the ultrasonic transducer drive signal or rate of closure of the clamp arm. For example, in one aspect, a control algorithm can be configured to control the force applied to the tissue by the clamp arm according to the collagen transformation point. This may be achieved by measuring the position of the clamp arm actuation member and its rate of change while maintaining the load of the clamp arm constant within coaptation pressure within a set operational range (e.g., 130-180 psi) corresponding to the particular instrument type.

FIG. **64A** is a graphical representation **130250** of the displacement of the clamp arm **1140** (FIG. **23**) versus time as the clamp arm **1140** is closed to identify the collagen transformation point, in accordance with at least one aspect of the present disclosure. FIG. **64B** is a magnified portion **130256** of the graphical representation **130250** shown in FIG. **64A**. The horizontal axis **130254** represents time (e.g., in sec.) and the vertical axis **130252** represents clamp arm displacement δ (e.g., in mm). In one aspect, a control algorithm can be configured to control the load applied to a tissue by the clamp arm **1140** (e.g., by controlling the rate of closure of the clamp arm **1140**) as the ultrasonic blade **1128** (FIG. **23**) heats the tissue according to the collagen transformation point of the tissue. In one such aspect, a control algorithm is configured to close the clamp arm **1140** until the clamp arm load reaches a threshold, which can include a particular value (e.g., 4.5 lbs.) or a range of values (e.g., within a range of 3.5 to 5 lbs.). At that point, the control algorithm sets the clamp arm displacement rate of change threshold θ and monitors the displacement of the clamp arm **1140**. As long as the rate of change of the clamp arm displacement remains within a predefined negative limit (i.e., below the threshold θ), the control algorithm can determine that the tissue is below the transformation temperature. As shown in the graphical representation of FIGS. **64A** and **64B**, when the control algorithm determines that the clamp arm displacement rate of change exceeds the threshold θ , the control algorithm can determine that melt temperature of the collagen has been reached.

In one aspect, once the control algorithm determines that the transition temperature has been reached, the control algorithm can be configured to alter the operation of the ultrasonic instrument accordingly. For example, the control algorithm can switch the surgical instrument from load control (of the clamp arm **1140**) to temperature control. In another aspect, the control algorithm can maintain the load control of the clamp arm after the collagen transformation temperature has been reached and monitor for when a clamp arm displacement rate of change threshold has been reached. The second clamp arm displacement rate of change threshold can correspond to, for example, the transition temperature of elastin. The locations of the collagen and/or elastin transitions temperatures in the plot **130258** of the clamp arm displacement over time can be referred to a "knee" in the plot **130258**. Accordingly, in this aspect, the control algorithm can be configured to alter the operation of the ultrasonic instrument according to whether the second clamp rate displacement rate of change threshold (or elastin "knee") has been reached and alter the operation of the ultrasonic instrument accordingly. For example, the control algorithm can switch the surgical instrument from load control (of the clamp arm **1140**) to temperature control when the elastin knee in the plot **130258** is detected.

Collagen transformation should be constant for a given heat flux between 45° and 50° C. for collagen, where elastin has a different melt temperature. Further, the temperature should flatten as the collagen absorbs the heat. In some aspects, the control algorithm can be configured to sample the position of the clamp arm and/or clamp arm displacement member at a higher rate around particular temperatures or within temperature ranges (e.g., the ranges for expected temperature for collagen and/or elastin transformation) in order to precisely ascertain when the monitored events occur.

In the aspect depicted in FIGS. **64A** and **64B**, the control algorithm acts to change the surgical instrument from load control to temperature control when the collagen transformation point is detected at time t_m . Without changing the surgical instrument to temperature control, the clamp arm displacement would increase geometrically, as shown in the projected plot **130260**. In one aspect, the control algorithm operating in temperature control mode lowers the amplitude of the ultrasonic transducer drive signal to change the heat flux generated by the ultrasonic blade **1128**, as shown by the flat portion of the plot **130258** after the threshold θ is reached. In some aspects, the control algorithm can be configured to increase the amplitude of the ultrasonic transducer drive signal after a particular period of time to, for example, measure the rate of temperature increase to determine when the elastin transformation temperature has been reached. Accordingly, as the clamp arm rate of closure approaches the next knee (i.e., the elastin knee) the clamp arm rate of closure can decrease. Load control of the clamp arm **1140** can be beneficial because, in some cases, it can provide the best sealing of vessels.

FIG. **65** is a logic flow diagram of a process **130300** depicting a control program or a logic configuration to detect the collagen transformation point to control the rate of closure of the of the clamp arm or the amplitude of the ultrasonic transducer, in accordance with at least one aspect of the present disclosure. The process **130300** can be executed by a control circuit or processor located in the surgical instrument or the generator. Accordingly, the control circuit executing the process **130300** measures **130302** a position of the clamp arm actuation member and its rate of change while maintaining the load on the clamp constant. As

previously described, in one aspect, the load on the clamp arm is maintained within coaptation pressure within a suitable range (130-180 psi) set by the ultrasonic surgical instrument. Once the jaws are taken up to a particular clamp arm load (e.g., 4.5 lbs.) or the clamp arm load is within a particular range (e.g., 3.5-5 lbs.), the control circuit sets **130304** the clamp arm displacement rate of change and monitors the position of the clamp arm actuation member for the period of time that the clamp arm displacement rate of change remains within a predefined negative limit (corresponding to the tissue being below the collagen transformation temperature). Accordingly, the control determines **130306** whether the clamp arm displacement rate of change exceeds the set threshold or, in other words, determines whether the tissue has reached the transition temperature. If the transition temperature has been reached, then the process **130300** proceeds along the YES branch and the control circuit switches **130308** the surgical instrument to temperature control (e.g., controls the ultrasonic transducer to decrease or maintain the temperature of the ultrasonic blade). In one aspect, the control circuit continues monitoring for the collagen transformation temperature. Alternatively, in the aspect depicted in FIG. **65**, if the transition temperature has not been reached, then the process **130300** proceeds along the NO branch and the control circuit maintains **130310** the load control of the clamp arm **1140** and monitors the clamp arm displacement rate of change to determine when the next transition point (e.g., the elastin transition point) occurs for the grasped tissue. The control circuit can do this in order to, for example, prevent the temperature of the tissue from increasing beyond the elastin transformation temperature.

It will be appreciated that the collagen transformation should be constant for a given heat flux (45° C.-50° C.). It also will be appreciated that load control of the clamp arm **1140** can, in some cases, provide the best sealing for particular types of tissues (e.g., vessels). During the period of time where collagen transformation is occurring, the temperature of the tissue should flatten while the collagen absorbs the heat. The control circuit can be configured to modulate the rate at which data points are collected around a particular temperature or temperatures of interests (e.g., transformation temperatures). Further, the control circuit can tune the amplitude of the ultrasonic transducer drive signal to control the heat flux generated by the ultrasonic blade **1128** at different points in the surgical procedure. For example, the control circuit can decrease the ultrasonic transducer amplitude during the period of collagen transformation. As another example, the control circuit can increase the ultrasonic transducer amplitude to measure the rate at which the temperature increases when the elastin knee occurs. It will be appreciated that as the elastin knee is approached, the rate of temperature change will decrease.

In another aspect, a control algorithm can be configured to detect the collagen transformation temperature to identify the collagen/elastin percentage of the grasped tissue. As discussed above, the control algorithm can then control various operational parameters of the surgical instrument according to the identified composition of the grasped tissue.

FIG. **66** is a graphical representation **130350** of the identification of the collagen transformation temperature point to identify the collagen/elastin ratio, in accordance with at least one aspect of the present disclosure. The vertical axis **130352** represents ultrasonic transducer impedance and the horizontal axis **130632** represents tissue temperature. The point at which the rate of change of the ultrasonic transducer impedance shifts corresponds to the

collagen/tissue composition of the tissue in an empirically determined manner. For example, if the rate of change of the ultrasonic transducer impedance shifts at the first temperature **130362**, then the tissue composition is 100% collagen. Accordingly, if the rate of change of the ultrasonic transducer impedance shifts at the second temperature **130364**, then the tissue composition is 100% elastin. If the rate of change of the ultrasonic transducer impedance shifts between the first temperature **130362** and the second temperature **130364**, then the tissue composition is a mixture of collagen and elastin.

The collagen transformation temperature can be used to directly identify the collagen/elastin percentage in the tissue and a control algorithm can be configured to adjust the operation of the ultrasonic device accordingly. As shown in FIG. **66**, a plot **130356** represents the empirical relationship between ultrasonic transducer impedance and tissue temperature. As indicated by the plot **130356**, the impedance (Z) of the ultrasonic transducer increases linearly at a first rate of change (slope) as a function of the temperature (T) at the tissue contact area. At the collagen transition temperature shown in the plot at the point **130358**, the rate of change of impedance (Z) as a function of temperature (T) decreases to a second rate of change. At the point **130358** where the slope of the plot **130356** changes, the collagen-to-elastin ratio can correspond to an empirically determined temperature **130360** (e.g., 85%). In one aspect, a control circuit or processor executing the aforementioned algorithm can be configured to determine at what temperature the rate of ultrasonic transducer impedance changes and then retrieve the corresponding tissue composition (e.g., percentage of collagen, percentage of elastin, or collagen/elastin ratio) from a memory (e.g., a look-up table).

FIG. **67** is a logic flow diagram of a process **130450** for identifying the composition of a tissue according to the change in ultrasonic transducer impedance, in accordance with at least one aspect of the present disclosure. The process **130450** can be executed by a control circuit of processor located in, for example, the surgical instrument or generator. Accordingly, the control circuit monitors **130452** the impedance (Z) of the ultrasonic transducer as a function of temperature (T). As previously described, the temperature (T) at the interface of the tissue and ultrasonic blade may be inferred by the algorithms described herein. The control circuit determines **130454** the rate of change of the ultrasonic transducer impedance $\Delta Z/\Delta T$. As the temperature at the ultrasonic blade/tissue interface increases, the impedance (Z) increases linearly at a first rate, as shown in FIG. **66**. Accordingly, the control circuit determines **130456** whether the slope $\Delta Z/\Delta T$ has changed (e.g., has decreased). If the slope $\Delta Z/\Delta T$ has not changed, then the process **130450** proceeds along the NO branch and continues determining **130454** the slope $\Delta Z/\Delta T$. If the slope $\Delta Z/\Delta T$ has changed, the control circuit determines **130458** that the collagen transition temperature has been reached.

The ratio of elastin to collagen may be determined by monitoring the collagen transformation point of the tissue and comparing the detected collagen transformation point to a look-up table. The look-up table can be stored in memory (e.g., memory **3326** of FIG. **31**) and contains the ratio of elastin to collagen and a corresponding collagen transformation point for a particular ratio as determined empirically. Adjusting Clamp Arm Force According to Tissue Location

In various aspects, a control algorithm can be configured to determine the location of a tissue within or relative to an end effector and adjust the clamp arm force accordingly. In one aspect, tissue can be identified or parameterized by

measuring the compression force load on the clamp arm and the position of the tissue within the jaw, e.g., where the tissue is located along the length of the ultrasonic blade. In one aspect, the time to initial measured load on the clamp arm is measured and then the compression rate on the tissue is measured to determine compressibility of the tissue versus the amount of tissue located across the length of the jaw. The rate of change of position of the clamp arm actuator is monitored while in load control as a way to determine tissue compressibility and therefore tissue type/disease state.

FIG. **68** is a graphical representation **130500** of the distribution of compression load across an ultrasonic blade **130404**, in accordance with at least one aspect off the present disclosure. The vertical axis **130502** represents the force applied to the tissue by the clamp arm **1140** and the horizontal axis **130504** represents position. The ultrasonic blade **130404** is dimensioned such that there are periodic nodes and antinodes that occur along the length of the blade. The location of the nodes/antinodes is determined by the wavelength of the ultrasonic displacement induced in the ultrasonic blade **130404** by the ultrasonic transducer. The ultrasonic transducer is driven by an electrical signal of suitable amplitude and frequency. As is known in the art, a node is a point of minimal or zero displacement of the ultrasonic blade **130404** and an antinode is a point of maximum displacement of the ultrasonic blade **130404**.

In the graphical representation **130500**, the ultrasonic blade **130404** is represented such that the nodes and antinodes are aligned with their corresponding position along the horizontal axis **130504**. The graphical representation **130500** includes a first plot **130506** and a second plot **130508**. As represented by either of the plots **130506**, **130508**, the compression force applied to the ultrasonic blade **130404** drops exponentially from a proximal end of the ultrasonic blade **130404** to a distal end of the ultrasonic blade **130404**. Thus, tissue **130410** located at the distal end of the ultrasonic blade **130404** is subjected to much lower compression force as compared to tissue **130410** located closer to the proximal end of the ultrasonic blade **130404**. The first plot **130506** can represent a default closure of the clamp arm **1140** where the resulting force applied to the distal tissue **130410** is F_1 . Generally, the amount of force applied by the clamp arm **1140** to tissue **130410** cannot be broadly increased without consideration because then too much force could be applied to tissue **130410** located proximally along the ultrasonic blade **130404**. However, by monitoring the location of the tissue **130410** along the ultrasonic blade **130404** (e.g., as discussed above under the heading DETERMINING TISSUE LOCATION VIA IMPEDANCE CHANGE and below under the heading DETERMINING TISSUE LOCATION VIA ELECTRODE CONTINUITY), a control algorithm can amplify the force applied to the tissue **130410** by the clamp arm **1140** in situations where tissue **130410** is located only at the distal end of the ultrasonic blade **130404**, as with the situation shown in FIG. **68**. For example, the second plot **130508** can represent a modified closure of the clamp arm **1140** where the control algorithm determines that tissue **130410** is located only at the distal end of the ultrasonic blade **130404** and correspondingly increases the force applied to the distal tissue **130410** by the clamp arm **1140** to F_2 ($F_2 > F_1$).

FIG. **69** is a graphical representation **130520** of pressure applied to tissue versus time, according to one aspect of the present disclosure. The vertical axis **130522** represents the pressure (e.g., in N/mm^2) applied to the tissue and the horizontal axis **130524** represents time. The first plot **130526** represents a normal or default compression force applied to

the distal tissue **130410** without amplification. During a default closure of the clamp arm **1140**, the compression force applied to the tissue **130410** is maintained at a constant value after an initial ramp-up period. The second plot **130528** represents the amplified compression force applied to the distal tissue **130410** to compensate for the presence of only distal tissue **130410**. In the modified closure of the clamp arm **1140**, the pressure is increased **130530** as compared to the default closure until eventually the amplified compression force is returned **130532** to normal compression levels to prevent burning/melting through the clamp arm **1140** pad.

Determining Tissue Location Via Electrode Continuity

In various aspects, a control algorithm can be configured to determine the location of a tissue within or relative to an end effector according to electrical continuity across an array of bipolar (i.e., positive and negative) electrodes positioned along a jaw or jaws of the end effector. The location of the tissue detectable by the bipolar electrode array can correspond to the specific position of the tissue relative to the jaw(s) and/or the percentage of the jaw(s) covered by tissue. In one aspect, the positive and negative electrodes are separated by a physical gap such that electrical continuity is established between the electrodes when tissue bridges the positive and negative electrodes. The positive and negative electrodes are configured in a matrix or array such that a processor or control circuit can be configured to detect where tissue is located between the positive and negative electrodes by monitoring or scanning the electrode array. In one aspect, the bipolar electrode array can be positioned along one jaw of an end effector. Accordingly, a control circuit or processor coupled to the bipolar electrode array can be configured to detect electrical continuity between adjacent electrodes to detect the presence of tissue thereagainst. In another aspect, the bipolar electrode array can be positioned along opposing jaws of an end effector. Accordingly, a control circuit or processor coupled to the bipolar electrode array can be configured to detect electrical continuity between the opposing jaws to detect the presence of tissue therebetween.

Determining what surface area of the jaw(s) is covered with tissue allows a control algorithm to determine the appropriate coaptation pressure for the amount of tissue grasped by the end effector and then calculate the corresponding clamp arm load. The clamp arm load can be expressed in terms of applied pressure (e.g., 130-180 psi) or applied force (e.g., 3.5-5 lbs. or nominally 4.5 lbs.). In some aspects, the bipolar electrode array can be delivered power from a monopolar or bipolar RF electrosurgical generator to the positive and negative electrodes. The generator power output may be a variety of constant, variable, or minimum values (e.g., 45 W, 35 W, or 5 W), a function of various variables associated with the surgical instrument and/or the generator (e.g., amplitude of the ultrasonic blade or clamp arm force), or dictated by an algorithm for controlling the generator according to its power curve (e.g., during ramp up of the generator).

FIG. 70 illustrates an end effector **130400**, including a single-jaw electrode array for detecting tissue location, in accordance with at least one aspect of the present disclosure. In the depicted aspect, the end effector **130400** includes a first jaw **130430** having an electrode array **130431** disposed thereon and a second jaw **130432**. The electrode array **130431** includes electrodes **130429** coupled to an energy source, such as an RF generator. The end effector **130400** can include an end effector for an ultrasonic surgical instrument where the second jaw **130432** is, for example, an

ultrasonic blade **1128** (FIG. 23), an electrosurgical instrument, an end effector for a surgical stapling and cutting instrument, and so on. The second jaw **130432** can include, for example, an ultrasonic blade **1128** (FIG. 23) or a cooperating jaw of an electrosurgical or surgical stapling and cutting instrument. In the depicted aspect, the electrode array **130431** includes 12 electrodes **130429** arranged in a generally herringbone-shaped pattern; however, the number, shape, and arrangement of the electrodes **130429** in the electrode array **130431** are merely for illustrative purposes. Accordingly, the electrode array **130431** can include various numbers, shapes, and/or arrangements of electrodes **130429**. For example, the number of electrodes **130429** can be adjusted according to the desired resolution for detecting tissue position.

In one aspect, the electrode array **130431** can include electrodes **130429** that are separated by a physical gap and alternate in polarity or are coupled to opposing terminals (i.e., a supply terminal and a return terminal) of an energy source. For example, in the depicted aspect, the evenly numbered electrodes **130429** can be a first polarity (e.g., positive polarity or coupled to a supply terminal of a power source) and the odd numbered electrodes **130429** can be a second polarity (e.g., negative or coupled to a return terminal of a power source). Accordingly, when tissue **130410** contacts adjacent electrodes **130429**, the tissue **130410** physically and electrically bridges the bipolar electrodes **130429** and allows current to flow therebetween. The flow of current between bipolar electrodes **130429** can be detected by a control algorithm executed by a control circuit or processor coupled to the electrode array **130431**, thereby allowing the control circuit or processor to detect the presence of tissue **130410**.

The detection of tissue by the electrode array **130431** can be represented graphically by an activation matrix. For example, FIG. 71 illustrates an activation matrix **130550** indicating the position of the tissue **130410** according to the electrode array **130431** depicted in FIG. 70. The vertical axis **130554** and the horizontal axis **130555** both represent the electrodes **130429** of the electrode array **130431**, where the numbers along the axes **130554**, **130555** represent the correspondingly numbered electrodes **130429**. The activation regions **130552** indicate where continuity is present between the corresponding electrodes **130429**, i.e., where tissue **130410** is present. In FIG. 70, a tissue **130410** is present across the first, second, and third electrodes **130429** and, as discussed above, the electrodes **130429** can alternate in polarity in some aspects. Accordingly, there is electrical continuity between the first and second electrodes **130429** and the second and third electrodes **130429**. It should be noted that, in this described aspect, there would be no continuity between the first and third electrodes **130429** because they would be the same polarity. The continuity between these electrodes **130429** is indicated graphically by the activation regions **130552** in the activation matrix **130550**. It should also be noted that the region **130553** bounded by the activation regions **130552** is not indicated as activated because, in this described aspect, an electrode **130429** cannot have continuity with itself. A control algorithm executed by a control circuit or processor coupled to the electrode array **130431** can be configured to infer the position of the tissue **130410** within the end effector **130400** (as the locations of the electrodes **130429** would be known), the proportion of jaws **130430**, **130432** of the end effector **130400** covered by tissue **130410**, and so on because the tissue location corresponds to the particular electrodes **130429** where electrical continuity has been established.

FIG. 72 illustrates an end effector 130400 including a dual-jaw electrode array for detecting tissue location, in accordance with at least one aspect of the present disclosure. In the depicted aspect, the end effector 130400 includes a first jaw 130430 having a first electrode array 130431 disposed thereon and a second jaw 130432 having a second electrode array 130433 disposed thereon. The electrode arrays 130431, 130433 each include electrodes 130429 coupled to an energy source, such as an RF generator. The end effector 130400 can include an end effector for an electrosurgical instrument, an end effector for a surgical stapling and cutting instrument, and so on. As discussed above, the number, shape, and/or arrangement of electrodes 130429 can be varied in various aspects. For example, in FIG. 75 the electrode arrays 130431, 130433 are arranged in overlapping tiled or rectangular patterns.

In one aspect, the opposing electrodes 130429 of the electrode arrays 130431, 130433 are separated by a physical gap, and each electrode array 130431, 130433 is of an opposing polarity or is coupled to an opposing terminal (i.e., a supply terminal and a return terminal) of an energy source. For example, in the depicted aspect, the first electrode array 130431 can be a first polarity (e.g., positive polarity or coupled to a supply terminal of a power source) and the second electrode array 130433 can be a second polarity (e.g., negative or coupled to a return terminal of a power source). Accordingly, when tissue 130410 contacts an electrode 130429 of each of the opposing electrode arrays 130431, 130433, the tissue 130410 physically and electrically bridges the bipolar electrodes 130429 and allows current to flow therebetween. The flow of current between bipolar electrodes 130429 can be detected by a control algorithm executed by a control circuit or processor coupled to the electrode arrays 130431, 130433, thereby allowing the control circuit or processor to detect the presence of tissue 130410.

As discussed above, an activation matrix can graphically represent the presence of tissue. For example, FIG. 73 illustrates an activation matrix 130556 indicating the position of the tissue 130410 as depicted in FIG. 74. The vertical axis 130557 represents the electrodes 130429 of the first electrode array 130431 and the horizontal axis 130558 represents the electrodes 130429 of the second electrode array 130433, where the numbers along the axes 130557, 130558 represent the correspondingly numbered electrodes 130429 for each electrode array 130431, 130433. The activation regions 130552 indicate where continuity is present between the corresponding electrodes 130429, i.e., where tissue 130410 is present. In FIG. 74, a tissue 130410 is positioned against the first, second, and third electrodes 130431a, 130431b, 130431c of the first electrode array 130431 and the first, second, and third electrodes 130433a, 130433b, 130433c of the second electrode array 130433. Accordingly, there is electrical continuity between each of these sets of electrodes of the opposing electrode arrays 130431, 130433 as current can flow between these opposing sets of electrodes. The electrodes between which there is continuity due to the position of the grasped tissue 130410 are indicated graphically by the activation regions 130552 in the activation matrix 130556 of FIG. 73. Further, because the tissue 130410 is not positioned against the fourth, fifth, and sixth electrodes 130431d, 130431e, 130431f of the first electrode array 130431 and the fourth, fifth, and sixth electrodes 130433d, 130433e, 130433f of the second electrode array 130433, there is no electrical continuity between these electrodes. A control algorithm executed by a control circuit or processor coupled to the electrode arrays 130431,

130433 can be configured to infer the position of the tissue 130410 within the end effector 130400 (as the locations of the electrodes 130429 would be known, as indicated in FIG. 74), the proportion of jaws 130430, 130432 of the end effector 130400 covered by tissue 130410, and so on because the tissue location corresponds to the particular electrodes 130429 where electrical continuity has been established. In the depicted example, the activated electrodes of the first and second electrode arrays 130431, 130433 are the electrodes 130429 that overlap and where there is tissue 130410 situated therebetween.

In another aspect, the end effector can be configured to transmit a plurality of signals or pings at varying frequencies and the electrode array can be coupled to a circuit, including corresponding band-pass filters that can each detect a particular frequency signal or signals through a frequency transform. Various portions of the electrode array circuit can include band-pass filters tuned to different frequencies. Therefore, the location of the tissue grasped by the end effector corresponds to the particular detected signals. The signals can be transmitted, for example, at a nontherapeutic frequency (e.g., at frequencies above the therapeutic frequency range for RF electrosurgical instruments). The electrode array circuit can include, for example, a flex circuit.

FIG. 75 illustrates an end effector 130400 including a first jaw 130430 having a first segmented electrode array 130431 and a second jaw 130432 having a second segmented electrode array 130433, according to at least one aspect of the present disclosure. Further, FIG. 76 illustrates a tissue 130410 that is grasped by the end effector 130400 overlaying the second jaw 130432. In one aspect, the first electrode array 130431 is configured to transmit signals at various frequencies (e.g., nontherapeutic frequencies) and the second electrode array 130433 is configured to receive the signals through a tissue 130410 grasped by the end effector 130400 (i.e., when a tissue 130410 is contacting both electrode arrays 130431, 130433). The second electrode array 130433 can include a segmented electrode array circuit 130600, as shown in FIG. 77, wherein each circuit segment includes a band-pass filter 130601 coupled to each electrode 130602 of the second electrode array 130433. Each band-pass filter 130601 can include one or more capacitors 130604 and one or more inductors 130606, wherein the number, arrangement, and values of the capacitors 130604 and inductors 130606 can be selected to tune each band-pass filter 130601 to a particular frequency or frequency band. As the tissue 130410 functions as the signal-conducting medium between the electrode arrays 130431, 130433 and different portions of the second electrode array 130433 are tuned to detect signals of varying frequencies (via differently tuned band-pass filters 130601), a control algorithm executed by a control circuit or processor coupled to the electrode arrays 130431, 130433 can be configured to infer the position of the tissue 130410 according to which signals are detected thereby. In the depicted aspect, the electrode arrays 130431, 130433 include six electrode segments 130602 arranged in a generally tiled pattern with a semicircular end segment; however, the number, shape, and arrangement of the electrodes 130602 in the electrode arrays 130431, 130433 are merely for illustrative purposes. Accordingly, the electrode arrays 130431, 130433 can include various numbers, shapes, and/or arrangements of electrodes 130602. For example, the number of electrodes 130602 can be adjusted according to the desired resolution for detecting tissue position.

FIG. 78 is a graphical representation 130650 of the frequency response corresponding to the tissue 130410

grasped in FIG. 76, in accordance with at least one aspect of the present disclosure. The vertical axis **130652** represents amplitude and the horizontal axis **130654** represents RF frequency. In one aspect, the second electrode array **130433** includes a first electrode circuit segment **130602a** tuned to a frequency band defined by a first center frequency f_{s1} (e.g., 5 MHz), a second electrode circuit segment **130602b** tuned to a frequency band defined by a second center frequency f_{s2} (e.g., 10 MHz), a third electrode circuit segment **130602c** tuned to a frequency band defined by a third center frequency f_{s3} (e.g., 15 MHz), and a fourth electrode circuit segment **130602d** tuned to a frequency band defined by a fourth center frequency f_{s4} (e.g., 20 MHz). As depicted in FIG. 78, the sensing frequency bands define a sensing frequency range **130658** above the therapeutic frequency range **130656** defined by f_{t1} (e.g., 300 kHz) to f_{t2} (e.g., 500 kHz) and/or a preferred therapeutic frequency (e.g., 350 kHz). In one aspect, the center sensing frequencies f_{s1} , f_{s2} , f_{s3} , f_{s4} in the sensing frequency range **130658** are each separated by a defined frequency value (e.g., 5 MHz). Further, although the sensing frequency range **130658** is illustrated as including four sensing frequency bands, this is merely for illustrative purposes. In the depicted example, the grasped tissue **130410** is contacting the second, third, and fourth electrode circuit segments **130602b**, **130602c**, **130602d**. Accordingly, the detected frequency response includes peaks **130655b**, **130655c**, **130655d** at each of the corresponding frequencies. A control algorithm could therefore infer the position of the tissue **130410** from the detected frequency response, i.e., the control circuit can determine that the tissue **130410** is positioned within the end effector **130400** such that it is contacting the second, third, and fourth electrode circuit segments **130602b**, **130602c**, **130602d** and not other circuit segments. The control algorithm can thus infer the position of the tissue **130410** relative to the jaws **130430**, **130432** of the end effector **130400** and/or the percentage of the jaws **130430**, **130432** in contact with the tissue **130410**.

Adaptive Ultrasonic Blade Temperature Monitoring

In one aspect, an adaptive ultrasonic blade control algorithm may be employed to adjust various operational parameters of the ultrasonic system based on the temperature of the ultrasonic blade. The operational parameters controlled or regulated by the adaptive ultrasonic blade control algorithm can include, for example, the amplitude of the ultrasonic blade, the control signal driving the ultrasonic transducer, the pressure applied by the clamp arm, and so on. The adaptive ultrasonic blade control algorithm can be executed by a control circuit or processor located either in the generator or the surgical instrument.

In one example described in further detail below, the adaptive ultrasonic blade control algorithm dynamically monitors the temperature of the ultrasonic blade and adjusts the amplitude of the ultrasonic blade and/or the signal provided to the ultrasonic transducer accordingly. In another example described in further detail below, the adaptive ultrasonic blade control algorithm dynamically monitors the temperature of the ultrasonic blade and adjusts the clamp arm pressure accordingly. The adaptive ultrasonic blade control algorithm can measure the temperature of the ultrasonic blade via various techniques, such as by analyzing the frequency spectrum of the ultrasonic transducer as discussed above under the heading TEMPERATURE INFERENCE. Other techniques for determining the temperature of the ultrasonic blade employ non-contact imaging. These and other techniques are described in detail herein and additional

detail may be found in U.S. Provisional Patent Application No. 62/692,768, titled SMART ENERGY DEVICES. Adjusting Ultrasonic System Parameters According to Temperature

In one aspect, the adaptive ultrasonic blade control algorithm can be configured to adjust operational parameters of the ultrasonic system based on the temperature of the ultrasonic blade. As discussed above under the heading TEMPERATURE INFERENCE, the natural frequency of the ultrasonic blade/transducer moves with temperature, and thus, the temperature of the ultrasonic blade can be inferred from the phase angle between the voltage and current signals applied to drive the ultrasonic transducer. Further, the ultrasonic blade temperature corresponds to the tissue temperature. In some aspects, the adaptive ultrasonic blade control algorithm can be configured to detect the ultrasonic blade temperature and modulate the surgical instrument's operational parameters according to the temperature. The operational parameters can include, for example, the frequency of the ultrasonic transducer drive signal, the amplitude of the ultrasonic blade (which can, for example, correspond to the magnitude or amplitude of the electrical current supplied to the ultrasonic transducer), the pressure applied by the clamp arm, and so on. The adaptive ultrasonic blade control algorithm can be executed by a control circuit or processor located either in the generator or the surgical instrument.

Accordingly, in one aspect, the adaptive ultrasonic blade control algorithm detects the resonant frequency of the ultrasonic blade, as described previously under the heading TEMPERATURE INFERENCE, and then monitors the resonant frequency over time in order to detect a modal shift in the resonant frequency waveform. A shift in the resonant waveform can be correlated to the occurrence of a system change, such as an increase in the temperature of the ultrasonic blade. In some aspects, an adaptive ultrasonic blade control algorithm can be configured to adjust the amplitude of the ultrasonic drive signal, and therefore the amplitude of the ultrasonic blade displacement, in order to measure the temperature of the tissue. In other aspects, an adaptive ultrasonic blade control algorithm can be configured to control the amplitude of the ultrasonic drive signal, and therefore the amplitude of the ultrasonic blade displacement, according to the temperature of the ultrasonic blade and/or tissue in order to hold the temperature of the ultrasonic blade and/or tissue to a predefined temperature or within predefined thresholds (e.g., to allow the ultrasonic blade to cool if its temperature is becoming too high). In still other aspects, an adaptive ultrasonic blade control algorithm can be configured to modulate the RF power and waveform of an electrosurgical instrument in order to, for example, minimize temperature overshoot or change the ultrasonic blade heat flux, according to the tissue impedance, tissue temperature, and/or ultrasonic blade temperature. Further details regarding these and other functions have been described in reference to FIGS. 95-100, for example.

FIG. 79 is a graphical representation **130700** of the frequency of the ultrasonic transducer system as a function of drive frequency and ultrasonic blade temperature drift, in accordance with at least one aspect of the present disclosure. The horizontal axis **130704** represents the drive frequency (e.g., in Hz) applied to the ultrasonic system (e.g., the ultrasonic transducer and/or ultrasonic blade) and the vertical axis **130702** represents the resulting impedance phase angle (e.g., in rads). The first plot **130706** represents a characteristic resonant waveform of the ultrasonic system at normal or ambient temperature. As can be seen in the first plot **130706**, the ultrasonic system is in phase when driven

at the excitation frequency f_e (because the impedance phase angle is at or near 0 rads). Accordingly, f_e represents the resonant frequency of the ultrasonic system at ambient temperature. The second plot **130708** represents a characteristic waveform of the ultrasonic system when the temperature of the ultrasonic system has been elevated. As indicated in FIG. **79**, as the temperature of the ultrasonic system increases, the characteristic waveform (depicted by the first plot **130706**) of the ultrasonic blade and ultrasonic transducer shifts to the left, e.g., to a lower frequency range. Due to the shift in the ultrasonic system frequency waveform, the ultrasonic system is no longer in phase when driven at the excitation frequency f_e . Rather, the resonant frequency has shifted lower to f'_e . Therefore, a control circuit coupled to the ultrasonic system can detect or infer the temperature change in the ultrasonic system by detecting the change in the resonant frequency of the ultrasonic system and/or detecting when the ultrasonic system is out of phase when driven at a previously established resonant frequency.

Correspondingly, in some aspects, a control circuit coupled to the ultrasonic system can be configured to control the drive signal applied to the ultrasonic system by the generator according to the inferred temperature of the ultrasonic system in order to maintain the ultrasonic system in phase. Maintaining the ultrasonic system in phase can be utilized to, for example, control the temperature of the ultrasonic system. As discussed above, the resonant frequency at which the voltage and current signals are in phase shifts from f_e (e.g., 55.5 kHz) at normal temperature to f'_e as the temperature of the ultrasonic blade and/or ultrasonic transducer increases. Therefore, as the temperature of the ultrasonic system increases, a control circuit can control the generator to shift the frequency at which the ultrasonic system is driven from f_e to f'_e to maintain the ultrasonic system in phase with the generator drive signal. For additional description of adaptive ultrasonic blade control algorithms, please refer to the description associated with FIGS. **43A-54** hereinabove.

FIG. **80** is a graphical representation **130750** of temperature of the ultrasonic transducer as a function of time, in accordance with at least one aspect of the present disclosure. The vertical axis **130752** represents the temperature of the ultrasonic transducer, and the horizontal axis **130754** represents time. In one aspect, as the ultrasonic transducer temperature (represented by the plot **130756**) meets or exceeds a temperature threshold T_1 , the adaptive ultrasonic blade control algorithm controls the ultrasonic transducer to maintain the temperature of the ultrasonic transducer at or below the threshold temperature T_1 . The adaptive ultrasonic blade control algorithm can control the temperature of the ultrasonic transducer by, for example, modulating the power and/or drive signal applied to the ultrasonic transducer. Additional description of algorithms and techniques for controlling the temperature of an ultrasonic blade/transducer can be found under the headings FEEDBACK CONTROL and ULTRASONIC SEALING ALGORITHM WITH TEMPERATURE CONTROL and in U.S. Provisional Patent Application No. 62/640,417, titled TEMPERATURE CONTROL IN ULTRASONIC DEVICE AND CONTROL SYSTEM THEREFOR, filed Mar. 8, 2018, the disclosure of which is hereby incorporated by reference herein.

FIG. **81** is a graphical representation of the modal shift of resonant frequency based on the temperature of the ultrasonic blade moving the resonant frequency as a function of the temperature of the ultrasonic blade, in accordance with at least one aspect of the present disclosure. In the first graph

130800, the vertical axis **130802** represents the change in resonant frequency (Δf), and the horizontal axis **130804** represents the ultrasonic transducer drive frequency of the generator. In the second, third, and fourth graphs **130810**, **130820**, **130830**, the vertical axes **130812**, **130822**, **130832** represents frequency (f), current (I), and temperature (T), respectively, and the horizontal axes **130814**, **130824**, **130834** represent time (t). The first graph **130810** represents frequency shift of the ultrasonic system due to temperature change. The second graph **130820** represents current in the ultrasonic transducer, or amplitude adjustment, in order to hold stable frequency and temperature. The third graph **130830** represents temperature change of the tissue and/or ultrasonic system. The set of graphs **130800**, **130810**, **130820**, **130830** in conjunction demonstrate the functioning of a control algorithm configured to control the temperature of an ultrasonic system.

The control algorithm can be configured to control the ultrasonic system (e.g., ultrasonic transducer and/or ultrasonic blade) when the temperature of the ultrasonic system approaches a temperature threshold T_1 . In one aspect, the control algorithm can be configured to determine that the temperature threshold T_1 is being approached or has been reached according to whether the resonant frequency of the ultrasonic system has dropped by a threshold Δf_R . As indicated by the first graph **130800**, the change in frequency threshold Δf_R corresponding to the threshold temperature T_1 can, in turn, be a function of the drive frequency f_o of the ultrasonic system (as represented by the plot **130806**). As indicated by the second graph **130810** and the fourth graph **130830**, as the temperature of the tissue and/or ultrasonic blade increases (represented by the temperature plot **130836**), the resonant frequency correspondingly decreases (represented by the frequency plot **130816**). As the temperature plot **130836** approaches a temperature threshold T_1 (e.g., 130° C.) at time t_1 , the resonant frequency has dropped from f_1 to f_2 , causing the resonant frequency to reach the control algorithm's change in frequency threshold Δf_R and thereby causing the control algorithm act to stabilize the ultrasonic system temperature. By monitoring the change in resonant frequency of the ultrasonic system, the adaptive ultrasonic blade control algorithm can thus monitor the temperature of the ultrasonic system. Further, the adaptive ultrasonic blade control algorithm can be configured to adjust (e.g., decrease) the electrical current applied to the ultrasonic transducer or otherwise adjust the amplitude of the ultrasonic blade (represented by the current plot **130826**) in order to stabilize the tissue and/or ultrasonic blade temperature and/or the resonant frequency when the temperature meets or exceeds the threshold T_1 .

In another aspect, the adaptive ultrasonic blade control algorithm can be configured to adjust (e.g., decrease) the pressure applied by the clamp arm to the tissue when the temperature meets or exceeds the threshold T_1 . In various other aspects, the adaptive ultrasonic blade control algorithm can be configured to adjust a variety of other operational parameters associated with the ultrasonic system according to the temperature. In another aspect, the adaptive ultrasonic blade control algorithm can be configured to monitor multiple temperature thresholds. For example, a second temperature threshold T_2 can represent, for example, a melt or failure temperature of the clamp arm. Accordingly, the adaptive ultrasonic blade control algorithm can be configured to take the same or different actions according to the particular temperature threshold that has been met or exceeded.

In various aspects, non-contact imaging may be employed to determine the temperature of the ultrasonic blade in addition to or in lieu of the aforementioned techniques. For example, short-wave thermography may be employed to measure blade temperature by imaging the blade from the stationary surrounding ground via a CMOS imaging sensor. Thermographic non-contact monitoring of the ultrasonic waveguide or ultrasonic blade temperature may be employed to control tissue temperature. In other aspects, non-contact imaging may be employed to determine surface conditions and finish of the ultrasonic blade to improve the temperature of tissue and/or ultrasonic blade through near IR detection techniques.

Determining Jaw State

A challenge with ultrasonic energy delivery is that ultrasonic acoustics applied on the wrong materials or the wrong tissue can result in device failure, for example, clamp arm pad burn through or ultrasonic blade breakage. It is also desirable to detect what is located in the jaws of an end effector of an ultrasonic device and the state of the jaws without adding additional sensors in the jaws. Locating sensors in the jaws of an ultrasonic end effector poses reliability, cost, and complexity challenges.

Ultrasonic spectroscopy smart blade algorithm techniques may be employed for estimating the state of the jaw (clamp arm pad burn through, staples, broken blade, bone in jaw, tissue in jaw, back-cutting with jaw closed, etc.) based on the impedance

$$Z_g(t) = \frac{V_g(t)}{I_g(t)}$$

of an ultrasonic transducer configured to drive an ultrasonic transducer blade, in accordance with at least one aspect of the present disclosure. The impedance $Z_g(t)$, magnitude $|Z|$, and phase φ are plotted as a function of frequency f .

Dynamic mechanical analysis (DMA), also known as dynamic mechanical spectroscopy or simply mechanical spectroscopy, is a technique used to study and characterize materials. A sinusoidal stress is applied to a material, and the strain in the material is measured, allowing the determination of the complex modulus of the material. The spectroscopy as applied to ultrasonic devices includes exciting the tip of the ultrasonic blade with a sweep of frequencies (compound signals or traditional frequency sweeps) and measuring the resulting complex impedance at each frequency. The complex impedance measurements of the ultrasonic transducer across a range of frequencies are used in a classifier or model to infer the characteristics of the ultrasonic end effector. In one aspect, the present disclosure provides a technique for determining the state of an ultrasonic end effector (clamp arm, jaw) to drive automation in the ultrasonic device (such as disabling power to protect the device, executing adaptive algorithms, retrieving information, identifying tissue, etc.).

FIG. 82 is a spectra 132030 of an ultrasonic device with a variety of different states and conditions of the end effector where the impedance $Z_g(t)$, magnitude $|Z|$, and phase φ are plotted as a function of frequency f , in accordance with at least one aspect of the present disclosure. The spectra 132030 is plotted in three-dimensional space where frequency (Hz) is plotted along the x-axis, phase (Rad) is plotted along the y-axis, and magnitude (Ohms) is plotted along the z-axis.

Spectral analysis of different jaw bites and device states produces different complex impedance characteristic patterns (fingerprints) across a range of frequencies for different conditions and states. Each state or condition has a different characteristic pattern in 3D space when plotted. These characteristic patterns can be used to estimate the condition and state of the end effector. FIG. 82 shows the spectra for air 132032, clamp arm pad 132034, chamois 132036, staple 132038, and broken blade 132040. The chamois 132036 may be used to characterize different types of tissue.

The spectra 132030 can be evaluated by applying a low-power electrical signal across the ultrasonic transducer to produce a non-therapeutic excitation of the ultrasonic blade. The low-power electrical signal can be applied in the form of a sweep or a compound Fourier series

$$Z_g(t) = \frac{V_g(t)}{I_g(t)}$$

to measure the impedance $Z_g(t)$ across the ultrasonic transducer at a range of frequencies in series (sweep) or in parallel (compound signal) using an FFT.

Methods of Classification of New Data

For each characteristic pattern, a parametric line can be fit to the data used for training using a polynomial, a Fourier series, or any other form of parametric equation as may be dictated by convenience. A new data point is then received and is classified by using the Euclidean perpendicular distance from the new data point to the trajectory that has been fitted to the characteristic pattern training data. The perpendicular distance of the new data point to each of the trajectories (each trajectory representing a different state or condition) is used to assign the point to a state or condition.

The probability distribution of distance of each point in the training data to the fitted curve can be used to estimate the probability of a correctly classified new data point. This essentially constructs a two-dimensional probability distribution in a plane perpendicular to the fitted trajectory at each new data point of the fitted trajectory. The new data point can then be included in the training set based on its probability of correct classification to make an adaptive, learning classifier that readily detects high-frequency changes in states but adapts to slow occurring deviations in system performance, such as a device getting dirty or the pad wearing out.

FIG. 83 is a graphical representation of a plot 132042 of a set of 3D training data set (S), where ultrasonic transducer impedance $Z_g(t)$, magnitude $|Z|$, and phase φ are plotted as a function of frequency f , in accordance with at least one aspect of the present disclosure. The 3D training data set (S) plot 132042 is graphically depicted in three-dimensional space where phase (Rad) is plotted along the x-axis, frequency (Hz) is plotted along the y-axis, magnitude (Ohms) is plotted along the z-axis, and a parametric Fourier series is fit to the 3D training data set (S). A methodology for classifying data is based on the 3D training data set (S0) used to generate the plot 132042).

The parametric Fourier series fit to the 3D training data set (S) is given by:

$$\vec{p} = \vec{a}_0 + \sum_{n=1}^{\infty} \left(\vec{a}_n \cos \frac{n\pi t}{L} + \vec{b}_n \sin \frac{n\pi t}{L} \right)$$

For a new point \vec{z} , the perpendicular distance from \vec{p} to \vec{z} is found by:

$$D = \|\vec{p} - \vec{z}\|$$

When:

$$\frac{\partial D}{\partial T} = 0$$

Then:

$$D = D_{\perp}$$

A probability distribution of D can be used to estimate the probability of a data point z belonging to the group S. Control

Based on the classification of data measured before, during, or after activation of the ultrasonic transducer/ultrasonic blade, a variety of automated tasks and safety measures can be implemented. Similarly, the state of the tissue located in the end effector and temperature of the ultrasonic blade also can be inferred to some degree, and used to better inform the user of the state of the ultrasonic device or protect critical structures, etc. Temperature control of an ultrasonic blade is described in commonly owned U.S. Provisional Patent Application No. 62/640,417, filed Mar. 8, 2018, titled TEMPERATURE CONTROL IN ULTRASONIC DEVICE AND CONTROL SYSTEM THEREFOR, which is incorporated herein by reference in its entirety.

Similarly, power delivery can be reduced when there is a high probability that the ultrasonic blade is contacting the clamp arm pad (e.g., without tissue in between) or if there is a probability that the ultrasonic blade has broken or that the ultrasonic blade is touching metal (e.g., a staple). Furthermore, back-cutting can be disallowed if the jaw is closed and no tissue is detected between the ultrasonic blade and the clamp arm pad.

Integration of Other Data to Improve Classification

This system can be used in conjunction with other information provided by sensors, the user, metrics on the patient, environmental factors, etc., by combing the data from this process with the aforementioned data using probability functions and a Kalman filter. The Kalman filter determines the maximum likelihood of a state or condition occurring given a plethora of uncertain measurements of varying confidence. Since this method allows for an assignment of probability to a newly classified data point, this algorithm's information can be implemented with other measures or estimates in a Kalman filter.

FIG. 84 is a logic flow diagram 132044 depicting a control program or a logic configuration to determine jaw conditions based on the complex impedance characteristic pattern (fingerprint), in accordance with at least one aspect of the present disclosure. Prior to determining jaw conditions based on the complex impedance characteristic pattern (fingerprint), a database is populated with reference complex impedance characteristic patterns or a training data sets (S) that characterize various jaw conditions, including, without limitation, air 132032, clamp arm pad 132034, chamois 132036, staple 132038, broken blade 132040, as shown in FIG. 82, and a variety of tissue types and conditions. The chamois dry or wet, full byte or tip, may be used to characterize different types of tissue. The data points used to generate reference complex impedance characteristic pat-

terns or a training data set (S) are obtained by applying a sub-therapeutic drive signal to the ultrasonic transducer, sweeping the driving frequency over a predetermined range of frequencies from below resonance to above resonance, measuring the complex impedance at each of the frequencies, and recording the data points. The data points are then fit to a curve using a variety of numerical methods including polynomial curve fit, Fourier series, and/or parametric equation. A parametric Fourier series fit to the reference complex impedance characteristic patterns or a training data set (S) is described herein.

Once the reference complex impedance characteristic patterns or a training data sets (S) are generated, the ultrasonic instrument measures new data points, classifies the new points, and determines whether the new data points should be added to the reference complex impedance characteristic patterns or a training data sets (S).

Turning now to the logic flow diagram of FIG. 84, in one aspect, the processor or control circuit measures 132046 a complex impedance of an ultrasonic transducer, wherein the complex impedance is defined as

$$Z_g(t) = \frac{V_g(t)}{I_g(t)}$$

The processor or control circuit receives 132048 a complex impedance measurement data point and compares 132050 the complex impedance measurement data point to a data point in a reference complex impedance characteristic pattern. The processor or control circuit classifies 132052 the complex impedance measurement data point based on a result of the comparison analysis and assigns 132054 a state or condition of the end effector based on the result of the comparison analysis.

In one aspect, the processor or control circuit receives the reference complex impedance characteristic pattern from a database or memory coupled to the processor. In one aspect, the processor or control circuit generates the reference complex impedance characteristic pattern as follows. A drive circuit coupled to the processor or control circuit applies a nontherapeutic drive signal to the ultrasonic transducer starting at an initial frequency, ending at a final frequency, and at a plurality of frequencies therebetween. The processor or control circuit measures the impedance of the ultrasonic transducer at each frequency and stores a data point corresponding to each impedance measurement. The processor or control circuit curve fits a plurality of data points to generate a three-dimensional curve of representativity of the reference complex impedance characteristic pattern, wherein the magnitude |Z| and phase φ are plotted as a function of frequency f. The curve fitting includes a polynomial curve fit, a Fourier series, and/or a parametric equation.

In one aspect, the processor or control circuit receives a new impedance measurement data point and classifies the new impedance measurement data point using a Euclidean perpendicular distance from the new impedance measurement data point to a trajectory that has been fitted to the reference complex impedance characteristic pattern. The processor or control circuit estimates a probability that the new impedance measurement data point is correctly classified. The processor or control circuit adds the new impedance measurement data point to the reference complex impedance characteristic pattern based on the probability of the estimated correct classification of the new impedance

measurement data point. In one aspect, the processor or control circuit classifies data based on a training data set (S), where the training data set (S) comprises a plurality of complex impedance measurement data, and curve fits the training data set (S) using a parametric Fourier series, wherein S is defined herein and wherein the probability distribution is used to estimate the probability of the new impedance measurement data point belonging to the group S.

Additional details regarding determining or estimating a state of the jaws or the surgical instrument as a whole can be found in U.S. Provisional Patent Application No. 62/692,768, titled SMART ENERGY DEVICES.

Situational Awareness

Referring now to FIG. 85, a timeline 5200 depicting situational awareness of a hub, such as the surgical hub 106 or 206 (FIGS. 1-11), for example, is depicted. The timeline 5200 is an illustrative surgical procedure and the contextual information that the surgical hub 106, 206 can derive from the data received from the data sources at each step in the surgical procedure. The timeline 5200 depicts the typical steps that would be taken by the nurses, surgeons, and other medical personnel during the course of a lung segmentectomy procedure, beginning with setting up the operating theater and ending with transferring the patient to a post-operative recovery room.

The situationally aware surgical hub 106, 206 receives data from the data sources throughout the course of the surgical procedure, including data generated each time medical personnel utilize a modular device that is paired with the surgical hub 106, 206. The surgical hub 106, 206 can receive this data from the paired modular devices and other data sources and continually derive inferences (i.e., contextual information) about the ongoing procedure as new data is received, such as which step of the procedure is being performed at any given time. The situational awareness system of the surgical hub 106, 206 is able to, for example, record data pertaining to the procedure for generating reports, verify the steps being taken by the medical personnel, provide data or prompts (e.g., via a display screen) that may be pertinent for the particular procedural step, adjust modular devices based on the context (e.g., activate monitors, adjust the field of view (FOV) of the medical imaging device, or change the energy level of an ultrasonic surgical instrument or RF electrosurgical instrument), and take any other such action described above.

As the first step S202 in this illustrative procedure, the hospital staff members retrieve the patient's EMR from the hospital's EMR database. Based on select patient data in the EMR, the surgical hub 106, 206 determines that the procedure to be performed is a thoracic procedure.

Second step S204, the staff members scan the incoming medical supplies for the procedure. The surgical hub 106, 206 cross-references the scanned supplies with a list of supplies that are utilized in various types of procedures and confirms that the mix of supplies corresponds to a thoracic procedure. Further, the surgical hub 106, 206 is also able to determine that the procedure is not a wedge procedure (because the incoming supplies either lack certain supplies that are necessary for a thoracic wedge procedure or do not otherwise correspond to a thoracic wedge procedure).

Third step S206, the medical personnel scan the patient band via a scanner that is communicably connected to the surgical hub 106, 206. The surgical hub 106, 206 can then confirm the patient's identity based on the scanned data.

Fourth step S208, the medical staff turns on the auxiliary equipment. The auxiliary equipment being utilized can vary

according to the type of surgical procedure and the techniques to be used by the surgeon, but in this illustrative case they include a smoke evacuator, insufflator, and medical imaging device. When activated, the auxiliary equipment that are modular devices can automatically pair with the surgical hub 106, 206 that is located within a particular vicinity of the modular devices as part of their initialization process. The surgical hub 106, 206 can then derive contextual information about the surgical procedure by detecting the types of modular devices that pair with it during this pre-operative or initialization phase. In this particular example, the surgical hub 106, 206 determines that the surgical procedure is a VATS procedure based on this particular combination of paired modular devices. Based on the combination of the data from the patient's EMR, the list of medical supplies to be used in the procedure, and the type of modular devices that connect to the hub, the surgical hub 106, 206 can generally infer the specific procedure that the surgical team will be performing. Once the surgical hub 106, 206 knows what specific procedure is being performed, the surgical hub 106, 206 can then retrieve the steps of that procedure from a memory or from the cloud and then cross-reference the data it subsequently receives from the connected data sources (e.g., modular devices and patient monitoring devices) to infer what step of the surgical procedure the surgical team is performing.

Fifth step S210, the staff members attach the EKG electrodes and other patient monitoring devices to the patient. The EKG electrodes and other patient monitoring devices are able to pair with the surgical hub 106, 206. As the surgical hub 106, 206 begins receiving data from the patient monitoring devices, the surgical hub 106, 206 thus confirms that the patient is in the operating theater.

Sixth step S212, the medical personnel induce anesthesia in the patient. The surgical hub 106, 206 can infer that the patient is under anesthesia based on data from the modular devices and/or patient monitoring devices, including EKG data, blood pressure data, ventilator data, or combinations thereof, for example. Upon completion of the sixth step S212, the pre-operative portion of the lung segmentectomy procedure is completed and the operative portion begins.

Seventh step S214, the patient's lung that is being operated on is collapsed (while ventilation is switched to the contralateral lung). The surgical hub 106, 206 can infer from the ventilator data that the patient's lung has been collapsed, for example. The surgical hub 106, 206 can infer that the operative portion of the procedure has commenced as it can compare the detection of the patient's lung collapsing to the expected steps of the procedure (which can be accessed or retrieved previously) and thereby determine that collapsing the lung is the first operative step in this particular procedure.

Eighth step S216, the medical imaging device (e.g., a scope) is inserted and video from the medical imaging device is initiated. The surgical hub 106, 206 receives the medical imaging device data (i.e., video or image data) through its connection to the medical imaging device. Upon receipt of the medical imaging device data, the surgical hub 106, 206 can determine that the laparoscopic portion of the surgical procedure has commenced. Further, the surgical hub 106, 206 can determine that the particular procedure being performed is a segmentectomy, as opposed to a lobectomy (note that a wedge procedure has already been discounted by the surgical hub 106, 206 based on data received at the second step S204 of the procedure). The data from the medical imaging device 124 (FIG. 2) can be utilized to determine contextual information regarding the type of

procedure being performed in a number of different ways, including by determining the angle at which the medical imaging device is oriented with respect to the visualization of the patient's anatomy, monitoring the number or medical imaging devices being utilized (i.e., that are activated and paired with the surgical hub **106, 206**), and monitoring the types of visualization devices utilized. For example, one technique for performing a VATS lobectomy places the camera in the lower anterior corner of the patient's chest cavity above the diaphragm, whereas one technique for performing a VATS segmentectomy places the camera in an anterior intercostal position relative to the segmental fissure. Using pattern recognition or machine learning techniques, for example, the situational awareness system can be trained to recognize the positioning of the medical imaging device according to the visualization of the patient's anatomy. As another example, one technique for performing a VATS lobectomy utilizes a single medical imaging device, whereas another technique for performing a VATS segmentectomy utilizes multiple cameras. As yet another example, one technique for performing a VATS segmentectomy utilizes an infrared light source (which can be communicably coupled to the surgical hub as part of the visualization system) to visualize the segmental fissure, which is not utilized in a VATS lobectomy. By tracking any or all of this data from the medical imaging device, the surgical hub **106, 206** can thereby determine the specific type of surgical procedure being performed and/or the technique being used for a particular type of surgical procedure.

Ninth step **S218**, the surgical team begins the dissection step of the procedure. The surgical hub **106, 206** can infer that the surgeon is in the process of dissecting to mobilize the patient's lung because it receives data from the RF or ultrasonic generator indicating that an energy instrument is being fired. The surgical hub **106, 206** can cross-reference the received data with the retrieved steps of the surgical procedure to determine that an energy instrument being fired at this point in the process (i.e., after the completion of the previously discussed steps of the procedure) corresponds to the dissection step. In certain instances, the energy instrument can be an energy tool mounted to a robotic arm of a robotic surgical system.

Tenth step **S220**, the surgical team proceeds to the ligation step of the procedure. The surgical hub **106, 206** can infer that the surgeon is ligating arteries and veins because it receives data from the surgical stapling and cutting instrument indicating that the instrument is being fired. Similarly to the prior step, the surgical hub **106, 206** can derive this inference by cross-referencing the receipt of data from the surgical stapling and cutting instrument with the retrieved steps in the process. In certain instances, the surgical instrument can be a surgical tool mounted to a robotic arm of a robotic surgical system.

Eleventh step **S222**, the segmentectomy portion of the procedure is performed. The surgical hub **106, 206** can infer that the surgeon is transecting the parenchyma based on data from the surgical stapling and cutting instrument, including data from its cartridge. The cartridge data can correspond to the size or type of staple being fired by the instrument, for example. As different types of staples are utilized for different types of tissues, the cartridge data can thus indicate the type of tissue being stapled and/or transected. In this case, the type of staple being fired is utilized for parenchyma (or other similar tissue types), which allows the surgical hub **106, 206** to infer that the segmentectomy portion of the procedure is being performed.

Twelfth step **S224**, the node dissection step is then performed. The surgical hub **106, 206** can infer that the surgical team is dissecting the node and performing a leak test based on data received from the generator indicating that an RF or ultrasonic instrument is being fired. For this particular procedure, an RF or ultrasonic instrument being utilized after parenchyma was transected corresponds to the node dissection step, which allows the surgical hub **106, 206** to make this inference. It should be noted that surgeons regularly switch back and forth between surgical stapling/cutting instruments and surgical energy (i.e., RF or ultrasonic) instruments depending upon the particular step in the procedure because different instruments are better adapted for particular tasks. Therefore, the particular sequence in which the stapling/cutting instruments and surgical energy instruments are used can indicate what step of the procedure the surgeon is performing. Moreover, in certain instances, robotic tools can be utilized for one or more steps in a surgical procedure and/or handheld surgical instruments can be utilized for one or more steps in the surgical procedure. The surgeon(s) can alternate between robotic tools and handheld surgical instruments and/or can use the devices concurrently, for example. Upon completion of the twelfth step **S224**, the incisions are closed up and the post-operative portion of the procedure begins.

Thirteenth step **S226**, the patient's anesthesia is reversed. The surgical hub **106, 206** can infer that the patient is emerging from the anesthesia based on the ventilator data (i.e., the patient's breathing rate begins increasing), for example.

Lastly, the fourteenth step **S228** is that the medical personnel remove the various patient monitoring devices from the patient. The surgical hub **106, 206** can thus infer that the patient is being transferred to a recovery room when the hub loses EKG, BP, and other data from the patient monitoring devices. As can be seen from the description of this illustrative procedure, the surgical hub **106, 206** can determine or infer when each step of a given surgical procedure is taking place according to data received from the various data sources that are communicably coupled to the surgical hub **106, 206**.

Situational awareness is further described in U.S. Provisional Patent Application Ser. No. 62/659,900, titled METHOD OF HUB COMMUNICATION, filed Apr. 19, 2018, which is herein incorporated by reference in its entirety. In certain instances, operation of a robotic surgical system, including the various robotic surgical systems disclosed herein, for example, can be controlled by the hub **106, 206** based on its situational awareness and/or feedback from the components thereof and/or based on information from the cloud **102**.

While several forms have been illustrated and described, it is not the intention of the applicant to restrict or limit the scope of the appended claims to such detail. Numerous modifications, variations, changes, substitutions, combinations, and equivalents to those forms may be implemented and will occur to those skilled in the art without departing from the scope of the present disclosure. Moreover, the structure of each element associated with the described forms can be alternatively described as a means for providing the function performed by the element. Also, where materials are disclosed for certain components, other materials may be used. It is therefore to be understood that the foregoing description and the appended claims are intended to cover all such modifications, combinations, and variations as falling within the scope of the disclosed forms. The

appended claims are intended to cover all such modifications, variations, changes, substitutions, modifications, and equivalents.

The foregoing detailed description has set forth various forms of the devices and/or processes via the use of block diagrams, flowcharts, and/or examples. Insofar as such block diagrams, flowcharts, and/or examples contain one or more functions and/or operations, it will be understood by those within the art that each function and/or operation within such block diagrams, flowcharts, and/or examples can be implemented, individually and/or collectively, by a wide range of hardware, software, firmware, or virtually any combination thereof. Those skilled in the art will recognize that some aspects of the forms disclosed herein, in whole or in part, can be equivalently implemented in integrated circuits, as one or more computer programs running on one or more computers (e.g., as one or more programs running on one or more computer systems), as one or more programs running on one or more processors (e.g., as one or more programs running on one or more microprocessors), as firmware, or as virtually any combination thereof, and that designing the circuitry and/or writing the code for the software and or firmware would be well within the skill of one of skill in the art in light of this disclosure. In addition, those skilled in the art will appreciate that the mechanisms of the subject matter described herein are capable of being distributed as one or more program products in a variety of forms, and that an illustrative form of the subject matter described herein applies regardless of the particular type of signal bearing medium used to actually carry out the distribution.

Instructions used to program logic to perform various disclosed aspects can be stored within a memory in the system, such as dynamic random access memory (DRAM), cache, flash memory, or other storage. Furthermore, the instructions can be distributed via a network or by way of other computer readable media. Thus a machine-readable medium may include any mechanism for storing or transmitting information in a form readable by a machine (e.g., a computer), but is not limited to, floppy diskettes, optical disks, compact disc, read-only memory (CD-ROMs), and magneto-optical disks, read-only memory (ROMs), random access memory (RAM), erasable programmable read-only memory (EPROM), electrically erasable programmable read-only memory (EEPROM), magnetic or optical cards, flash memory, or a tangible, machine-readable storage used in the transmission of information over the Internet via electrical, optical, acoustical or other forms of propagated signals (e.g., carrier waves, infrared signals, digital signals, etc.). Accordingly, the non-transitory computer-readable medium includes any type of tangible machine-readable medium suitable for storing or transmitting electronic instructions or information in a form readable by a machine (e.g., a computer).

As used in any aspect herein, the term “control circuit” may refer to, for example, hardwired circuitry, programmable circuitry (e.g., a computer processor comprising one or more individual instruction processing cores, processing unit, processor, microcontroller, microcontroller unit, controller, digital signal processor (DSP), programmable logic device (PLD), programmable logic array (PLA), or field programmable gate array (FPGA)), state machine circuitry, firmware that stores instructions executed by programmable circuitry, and any combination thereof. The control circuit may, collectively or individually, be embodied as circuitry that forms part of a larger system, for example, an integrated circuit (IC), an application-specific integrated circuit

(ASIC), a system on-chip (SoC), desktop computers, laptop computers, tablet computers, servers, smart phones, etc. Accordingly, as used herein “control circuit” includes, but is not limited to, electrical circuitry having at least one discrete electrical circuit, electrical circuitry having at least one integrated circuit, electrical circuitry having at least one application specific integrated circuit, electrical circuitry forming a general purpose computing device configured by a computer program (e.g., a general purpose computer configured by a computer program which at least partially carries out processes and/or devices described herein, or a microprocessor configured by a computer program which at least partially carries out processes and/or devices described herein), electrical circuitry forming a memory device (e.g., forms of random access memory), and/or electrical circuitry forming a communications device (e.g., a modem, communications switch, or optical-electrical equipment). Those having skill in the art will recognize that the subject matter described herein may be implemented in an analog or digital fashion or some combination thereof.

As used in any aspect herein, the term “logic” may refer to an app, software, firmware and/or circuitry configured to perform any of the aforementioned operations. Software may be embodied as a software package, code, instructions, instruction sets and/or data recorded on non-transitory computer readable storage medium. Firmware may be embodied as code, instructions or instruction sets and/or data that are hard-coded (e.g., nonvolatile) in memory devices.

As used in any aspect herein, the terms “component,” “system,” “module” and the like can refer to a computer-related entity, either hardware, a combination of hardware and software, software, or software in execution.

As used in any aspect herein, an “algorithm” refers to a self-consistent sequence of steps leading to a desired result, where a “step” refers to a manipulation of physical quantities and/or logic states which may, though need not necessarily, take the form of electrical or magnetic signals capable of being stored, transferred, combined, compared, and otherwise manipulated. It is common usage to refer to these signals as bits, values, elements, symbols, characters, terms, numbers, or the like. These and similar terms may be associated with the appropriate physical quantities and are merely convenient labels applied to these quantities and/or states.

A network may include a packet switched network. The communication devices may be capable of communicating with each other using a selected packet switched network communications protocol. One example communications protocol may include an Ethernet communications protocol which may be capable permitting communication using a Transmission Control Protocol/Internet Protocol (TCP/IP). The Ethernet protocol may comply or be compatible with the Ethernet standard published by the Institute of Electrical and Electronics Engineers (IEEE) titled “IEEE 802.3 Standard”, published in December, 2008 and/or later versions of this standard. Alternatively or additionally, the communication devices may be capable of communicating with each other using an X.25 communications protocol. The X.25 communications protocol may comply or be compatible with a standard promulgated by the International Telecommunication Union-Telecommunication Standardization Sector (ITU-T). Alternatively or additionally, the communication devices may be capable of communicating with each other using a frame relay communications protocol. The frame relay communications protocol may comply or be compatible with a standard promulgated by Consultative Committee for International Telegraph and Telephone

(CCITT) and/or the American National Standards Institute (ANSI). Alternatively or additionally, the transceivers may be capable of communicating with each other using an Asynchronous Transfer Mode (ATM) communications protocol. The ATM communications protocol may comply or be compatible with an ATM standard published by the ATM Forum titled "ATM-MPLS Network Interworking 2.0" published August 2001, and/or later versions of this standard. Of course, different and/or after-developed connection-oriented network communication protocols are equally contemplated herein.

Unless specifically stated otherwise as apparent from the foregoing disclosure, it is appreciated that, throughout the foregoing disclosure, discussions using terms such as "processing," "computing," "calculating," "determining," "displaying," or the like, refer to the action and processes of a computer system, or similar electronic computing device, that manipulates and transforms data represented as physical (electronic) quantities within the computer system's registers and memories into other data similarly represented as physical quantities within the computer system memories or registers or other such information storage, transmission or display devices.

One or more components may be referred to herein as "configured to," "configurable to," "operable/operative to," "adapted/adaptable," "able to," "conformable/conformed to," etc. Those skilled in the art will recognize that "configured to" can generally encompass active-state components and/or inactive-state components and/or standby-state components, unless context requires otherwise.

The terms "proximal" and "distal" are used herein with reference to a clinician manipulating the handle portion of the surgical instrument. The term "proximal" refers to the portion closest to the clinician and the term "distal" refers to the portion located away from the clinician. It will be further appreciated that, for convenience and clarity, spatial terms such as "vertical," "horizontal," "up," and "down" may be used herein with respect to the drawings. However, surgical instruments are used in many orientations and positions, and these terms are not intended to be limiting and/or absolute.

Those skilled in the art will recognize that, in general, terms used herein, and especially in the appended claims (e.g., bodies of the appended claims) are generally intended as "open" terms (e.g., the term "including" should be interpreted as "including but not limited to," the term "having" should be interpreted as "having at least," the term "includes" should be interpreted as "includes but is not limited to," etc.). It will be further understood by those within the art that if a specific number of an introduced claim recitation is intended, such an intent will be explicitly recited in the claim, and in the absence of such recitation no such intent is present. For example, as an aid to understanding, the following appended claims may contain usage of the introductory phrases "at least one" and "one or more" to introduce claim recitations. However, the use of such phrases should not be construed to imply that the introduction of a claim recitation by the indefinite articles "a" or "an" limits any particular claim containing such introduced claim recitation to claims containing only one such recitation, even when the same claim includes the introductory phrases "one or more" or "at least one" and indefinite articles such as "a" or "an" (e.g., "a" and/or "an" should typically be interpreted to mean "at least one" or "one or more"); the same holds true for the use of definite articles used to introduce claim recitations.

In addition, even if a specific number of an introduced claim recitation is explicitly recited, those skilled in the art

will recognize that such recitation should typically be interpreted to mean at least the recited number (e.g., the bare recitation of "two recitations," without other modifiers, typically means at least two recitations, or two or more recitations). Furthermore, in those instances where a convention analogous to "at least one of A, B, and C, etc." is used, in general such a construction is intended in the sense one having skill in the art would understand the convention (e.g., "a system having at least one of A, B, and C" would include but not be limited to systems that have A alone, B alone, C alone, A and B together, A and C together, B and C together, and/or A, B, and C together, etc.). In those instances where a convention analogous to "at least one of A, B, or C, etc." is used, in general such a construction is intended in the sense one having skill in the art would understand the convention (e.g., "a system having at least one of A, B, or C" would include but not be limited to systems that have A alone, B alone, C alone, A and B together, A and C together, B and C together, and/or A, B, and C together, etc.). It will be further understood by those within the art that typically a disjunctive word and/or phrase presenting two or more alternative terms, whether in the description, claims, or drawings, should be understood to contemplate the possibilities of including one of the terms, either of the terms, or both terms unless context dictates otherwise. For example, the phrase "A or B" will be typically understood to include the possibilities of "A" or "B" or "A and B."

With respect to the appended claims, those skilled in the art will appreciate that recited operations therein may generally be performed in any order. Also, although various operational flow diagrams are presented in a sequence(s), it should be understood that the various operations may be performed in other orders than those which are illustrated, or may be performed concurrently. Examples of such alternate orderings may include overlapping, interleaved, interrupted, reordered, incremental, preparatory, supplemental, simultaneous, reverse, or other variant orderings, unless context dictates otherwise. Furthermore, terms like "responsive to," "related to," or other past-tense adjectives are generally not intended to exclude such variants, unless context dictates otherwise.

It is worthy to note that any reference to "one aspect," "an aspect," "an exemplification," "one exemplification," and the like means that a particular feature, structure, or characteristic described in connection with the aspect is included in at least one aspect. Thus, appearances of the phrases "in one aspect," "in an aspect," "in an exemplification," and "in one exemplification" in various places throughout the specification are not necessarily all referring to the same aspect. Furthermore, the particular features, structures or characteristics may be combined in any suitable manner in one or more aspects.

Any patent application, patent, non-patent publication, or other disclosure material referred to in this specification and/or listed in any Application Data Sheet is incorporated by reference herein, to the extent that the incorporated materials is not inconsistent herewith. As such, and to the extent necessary, the disclosure as explicitly set forth herein supersedes any conflicting material incorporated herein by reference. Any material, or portion thereof, that is said to be incorporated by reference herein, but which conflicts with existing definitions, statements, or other disclosure material set forth herein will only be incorporated to the extent that no conflict arises between that incorporated material and the existing disclosure material.

In summary, numerous benefits have been described which result from employing the concepts described herein. The foregoing description of the one or more forms has been presented for purposes of illustration and description. It is not intended to be exhaustive or limiting to the precise form disclosed. Modifications or variations are possible in light of the above teachings. The one or more forms were chosen and described in order to illustrate principles and practical application to thereby enable one of ordinary skill in the art to utilize the various forms and with various modifications as are suited to the particular use contemplated. It is intended that the claims submitted herewith define the overall scope.

Various aspects of the subject matter described herein are set out in the following numbered examples:

Example 1. An ultrasonic electromechanical system for an ultrasonic surgical instrument. The ultrasonic electromechanical system comprises: an ultrasonic blade, an ultrasonic transducer acoustically coupled to the ultrasonic blade, and a control circuit coupled to the ultrasonic transducer. The ultrasonic transducer is configured to ultrasonically oscillate the ultrasonic blade in response to a drive signal. The control circuit is configured to: determine a first resonant frequency of the ultrasonic electromechanical system, determine a second resonant frequency of the ultrasonic electromechanical system as the ultrasonic blade oscillates against a tissue, and determine a state of the ultrasonic surgical instrument according to a comparison between the first resonant frequency and the second resonant frequency.

Example 2. The ultrasonic electromechanical system of Example 1, wherein the state of the ultrasonic surgical instrument comprises a temperature of the ultrasonic blade.

Example 3. The ultrasonic surgical instrument of Example 2, wherein the control circuit is configured to control the ultrasonic transducer to adjust an amplitude of the ultrasonic blade according to the temperature thereof.

Example 4. The ultrasonic electromechanical system of Example 3, wherein the control circuit is configured to adjust the amplitude of the ultrasonic blade to maintain the temperature of the ultrasonic blade at a predefined temperature.

Example 5. The ultrasonic electromechanical system of any of Examples 1-4, wherein the control circuit is configured to: apply the drive signal to the ultrasonic transducer at a plurality of frequencies over a frequency range and determine a voltage signal and a current signal of the ultrasonic transducer corresponding to each frequency of the drive signal within the frequency range. The first resonant frequency corresponds to a frequency where the voltage signal and the current signal are in phase.

Example 6. The ultrasonic electromechanical system of any of Examples 1-5, wherein the control circuit is configured to: monitor a voltage signal and a current signal applied to the ultrasonic transducer as the ultrasonic blade oscillates and adjust the drive signal until the voltage signal and the current signal are in phase. The second resonant frequency corresponds to a frequency where the voltage signal and the current signal are in phase.

Example 7. The ultrasonic electromechanical system of any of Examples 1-6, further comprising a memory coupled to the control circuit. The memory stores a plurality of states of the ultrasonic surgical instrument according to resonant frequency shift. The control circuit is configured to determine the state of the ultrasonic surgical instrument by retrieving from the memory which of the plurality of states of the ultrasonic surgical instrument corresponds to a difference between the first resonant frequency and the second resonant frequency.

Example 8. An ultrasonic generator connectable to an ultrasonic electromechanical system for an ultrasonic surgical instrument. The ultrasonic electromechanical system comprises an ultrasonic blade and an ultrasonic transducer acoustically coupled to the ultrasonic blade. The ultrasonic generator comprises a control circuit couplable to the ultrasonic transducer. The control circuit is configured to: apply a drive signal to the ultrasonic transducer to cause the ultrasonic blade to oscillate, determine a first resonant frequency of the ultrasonic electromechanical system, determine a second resonant frequency of the ultrasonic electromechanical system as the ultrasonic blade oscillates against a tissue, and determine a state of the ultrasonic surgical instrument according to a comparison between the first resonant frequency and the second resonant frequency.

Example 9. The ultrasonic generator of Example 8, wherein the state of the ultrasonic surgical instrument comprises a temperature of the ultrasonic blade.

Example 10. The ultrasonic generator of Example 9, wherein the control circuit is configured to control the ultrasonic transducer to adjust an amplitude of the ultrasonic blade according to the temperature thereof.

Example 11. The ultrasonic generator of Example 10, wherein the control circuit is configured to adjust the amplitude of the ultrasonic blade to maintain the temperature of the ultrasonic blade at a predefined temperature.

Example 12. The ultrasonic generator of any of Examples 8-11, wherein the control circuit is configured to: apply the drive signal to the ultrasonic transducer at a plurality of frequencies over a frequency range and determine a voltage signal and a current signal of the ultrasonic transducer corresponding to each frequency of the drive signal within the frequency range. The first resonant frequency corresponds to a frequency where the voltage signal and the current signal are in phase.

Example 13. The ultrasonic generator of any of Examples 8-12, wherein the control circuit is configured to: monitor a voltage signal and a current signal applied to the ultrasonic transducer as the ultrasonic blade oscillates and adjust the drive signal until the voltage signal and the current signal are in phase. The second resonant frequency corresponds to a frequency where the voltage signal and the current signal are in phase.

Example 14. The ultrasonic generator of any of Examples 8-13, further comprising a memory coupled to the control circuit. The memory stores a plurality of states of the ultrasonic surgical instrument according to resonant frequency shift. The control circuit is configured to determine the state of the ultrasonic surgical instrument by retrieving from the memory which of the plurality of states of the ultrasonic surgical instrument corresponds to a difference between the first resonant frequency and the second resonant frequency.

Example 15. A method of controlling an ultrasonic electromechanical system for an ultrasonic surgical instrument. The ultrasonic electromechanical system comprises an ultrasonic blade and an ultrasonic transducer acoustically coupled to the ultrasonic blade. The method comprises: determining, by a control circuit coupled to the ultrasonic electromechanical system, a natural resonant frequency of the ultrasonic electromechanical system; monitoring, by the control circuit, a resonant frequency of the ultrasonic electromechanical system as the ultrasonic blade oscillates; and determine, by the control circuit, whether a state change of the ultrasonic electromechanical system has occurred

according to whether the resonant frequency of the ultrasonic electromechanical system has deviated from the natural resonant frequency.

Example 16. The method of Example 15, wherein the state change of the ultrasonic surgical instrument comprises a temperature increase of the ultrasonic blade.

Example 17. The method of Example 16, further comprising controlling, by the control circuit, the ultrasonic transducer to adjust an amplitude of the ultrasonic blade according to the temperature increase thereof.

Example 18. The method of Example 17, further comprising controlling, by the control circuit, the ultrasonic transducer to adjust the amplitude of the ultrasonic blade to maintain a temperature of the ultrasonic blade at a pre-defined temperature.

Example 19. The method of any one of Examples 15-18, further comprising: applying, by the control circuit, a drive signal to the ultrasonic electromechanical system at a plurality of frequencies over a frequency range to oscillate the ultrasonic blade and determining, by the control circuit, a voltage signal and a current signal of the ultrasonic transducer corresponding to each frequency of the drive signal within the frequency range. The natural resonant frequency corresponds to a frequency where the voltage signal and the current signal are in phase.

Example 20. The method of any one of Examples 15-19, further comprising: applying, by the control circuit, a drive signal to the ultrasonic electromechanical system to oscillate the ultrasonic blade; monitoring, by the control circuit, a voltage signal and a current signal of the ultrasonic electromechanical system resulting from the drive signal as the ultrasonic blade oscillates; and adjusting, by the control circuit, the drive signal until the voltage signal and the current signal are in phase. The resonant frequency corresponds to a frequency where the voltage signal and the current signal are in phase.

Example 21. The method of any one of Examples 15-20, further comprising: retrieving, from a memory coupled to the control circuit, a state of the ultrasonic surgical instrument corresponding to a difference between the resonant frequency and the natural resonant frequency; and determining whether the state change has occurred according to whether the state of the ultrasonic surgical instrument corresponds to a prior state of the ultrasonic surgical instrument.

The invention claimed is:

1. An ultrasonic electromechanical system for an ultrasonic surgical instrument, the ultrasonic electromechanical system comprising:

an ultrasonic blade;

a clamp arm disposed opposite the ultrasonic blade;

an ultrasonic transducer acoustically coupled to the ultrasonic blade, the ultrasonic transducer configured to ultrasonically oscillate the ultrasonic blade in response to a drive signal; and

a control circuit coupled to the ultrasonic transducer, the control circuit configured to:

determine a temperature of the ultrasonic blade;

increase an amount of power of the drive signal when the temperature of the ultrasonic blade is less than a first predetermined value, wherein the first predetermined value is a value less than a temperature configured to cut a tissue; and

decrease the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than a second predetermined value,

wherein the second predetermined value is greater than the first predetermined value.

2. The ultrasonic electromechanical system of claim 1, wherein the control circuit is configured to determine the temperature of the ultrasonic blade by measuring a phase angle between a voltage of the drive signal and a current of the drive signal.

3. The ultrasonic electromechanical system of claim 1, wherein the temperature configured to cut the tissue is 220° C.

4. The ultrasonic electromechanical system of claim 1, wherein the second predetermined value is a value greater than a temperature configured to melt a pad affixed to the clamp arm.

5. The ultrasonic electromechanical system of claim 4, wherein the temperature to melt the pad affixed to the clamp arm is 380° C.

6. The ultrasonic electromechanical system of claim 1, wherein the control circuit is further configured to maintain the ultrasonic blade at an average temperature over time after the control circuit decreases the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than the second predetermined value,

wherein the average temperature over time is a temperature greater than the first predetermined value and less than the second predetermined value.

7. The ultrasonic electromechanical system of claim 1, wherein the control circuit is configured to increase the amount of power of the drive signal at a first predetermined time after the control circuit determines that the temperature of the ultrasonic blade is less than the first predetermined value.

8. The ultrasonic electromechanical system of claim 1, wherein the control circuit is configured to decrease the amount of power of the drive signal at a second predetermined time after the control circuit determines that the temperature of the ultrasonic blade is greater than a third predetermined value.

9. An ultrasonic generator connectable to an ultrasonic electromechanical system for an ultrasonic surgical instrument, the ultrasonic electromechanical system comprising an ultrasonic blade, a clamp arm disposed opposite the ultrasonic blade, and an ultrasonic transducer acoustically coupled to the ultrasonic blade, the ultrasonic generator comprising:

a control circuit couplable to the ultrasonic transducer, the control circuit configured to:

apply a drive signal to the ultrasonic transducer;

determine a temperature of the ultrasonic blade;

increase an amount of power of the drive signal when the temperature of the ultrasonic blade is less than a first predetermined value, wherein the first predetermined value is a value less than a temperature configured to cut a tissue; and

decrease the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than a second predetermined value,

wherein the second predetermined value is greater than the first predetermined value.

10. A method of controlling an ultrasonic electromechanical system for an ultrasonic surgical instrument, the ultrasonic electromechanical system comprising an ultrasonic blade, a clamp arm disposed opposite the ultrasonic blade, a control circuit, and an ultrasonic transducer acoustically coupled to the ultrasonic blade, the ultrasonic transducer configured to ultrasonically oscillate the ultrasonic blade in response to a drive signal, the method comprising:

159

determining, by the control circuit, a temperature of the ultrasonic blade;

increasing, by the control circuit, an amount of power of the drive signal when the temperature of the ultrasonic blade is less than a first predetermined value, wherein the first predetermined value is a value less than a temperature configured to cut a tissue; and

decreasing, by the control circuit, the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than a second predetermined value, wherein the second predetermined value is greater than the first predetermined value.

11. The method of claim 10, wherein determining, by the control circuit, the temperature of the ultrasonic blade comprises determining, by the control circuit, the temperature of the ultrasonic blade by measuring a phase angle between a voltage of the drive signal and a current of the drive signal.

12. The method of claim 10, wherein the temperature sufficient to cut the tissue is 220° C.

13. The method of claim 10, wherein decreasing, by the control circuit, the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than the second predetermined value comprises decreasing, by the control circuit, the amount of power of the drive signal when

160

the temperature of the ultrasonic blade is greater than a temperature sufficient to melt a pad affixed to the clamp arm.

14. The method of claim 13, wherein the temperature sufficient to melt the pad affixed to the clamp arm is 380° C.

15. The method of claim 10, further comprising maintaining, by the control circuit, the ultrasonic blade at an average temperature over time after the control circuit decreases the amount of power of the drive signal when the temperature of the ultrasonic blade is greater than the second predetermined value,

wherein the average temperature over time is a temperature greater than the first predetermined value and less than the second predetermined value.

16. The method of claim 10, further comprising increasing, by the control circuit, the amount of power of the drive signal at a first predetermined time after determining, by the control circuit, that the temperature of the ultrasonic blade is less than the first predetermined value.

17. The method of claim 10, further comprising decreasing, by the control circuit, the amount of power of the drive signal at a second predetermined time after determining, by the control circuit, that the temperature of the ultrasonic blade is greater than the second predetermined value.

* * * * *