



- (51) International Patent Classification:
A61B 5/1455 (2006.01) *A61B 5/00* (2006.01)
- (21) International Application Number:
PCT/US2018/038764
- (22) International Filing Date:
21 June 2018 (21.06.2018)
- (25) Filing Language: English
- (26) Publication Language: English
- (30) Priority Data:
62/525,807 28 June 2017 (28.06.2017) US
15/654,039 19 July 2017 (19.07.2017) US
- (71) Applicant: WELCH ALLYN, INC. [US/US]; 4341 State Street Road, Skaneateles Falls, New York 13153 (US).
- (72) Inventor: QUINN, David E.; c/o Welch Allyn, Inc., Attn: Legal Department, 4341 State Street Road, Skaneateles Falls, New York 13153 (US).
- (74) Agent: BARAN, Kenneth C.; Hill-Rom Services, Inc., 1069 State Route 46 East, Mail Stop K-03, Batesville, Indiana 47006 (US).
- (81) Designated States (*unless otherwise indicated, for every kind of national protection available*): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BN, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DJ, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IR, IS, JO, JP, KE, KG, KH, KN, KP, KR, KW, KZ, LA, LC, LK, LR, LS, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PA, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SA, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.
- (84) Designated States (*unless otherwise indicated, for every kind of regional protection available*): ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, ST, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM,

(54) Title: PHYSIOLOGICAL MONITORING AND RELATED METHODS




FIG. 1

(57) **Abstract:** A method of estimating oxygen saturation includes illuminating a target site with light of three different wavelengths, for example green, red and infrared, and detecting light returned from the site at each of the wavelengths. The method distinguishes between cardiac pulsatility and nonpulsatility based on a parameter of the return light of the shortest wavelength. The method formulates an estimate of oxygen saturation as a function of a parameter of the light of the two longer wavelengths taken at both a period of pulsatility and at a period of nonpulsatility. An oximeter adapted to carry out the method is also disclosed. Aspects of the method and oximeter may be used to estimate other physiological parameters.

TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW,
KM, ML, MR, NE, SN, TD, TG).

Published:

— *with international search report (Art. 21(3))*

PHYSIOLOGICAL MONITORING AND RELATED METHODS

TECHNICAL FIELD

[0001] The subject matter described herein relates to devices and methods for monitoring physiological parameters including a pulse oximeter and a related methodology which uses a third wavelength of light, in addition to two more conventionally used wavelengths, to improve the accuracy of peripheral blood oxygen saturation (SpO₂) measurements. One example application for the oximeter and method is to determine SpO₂ based on measurements at sites on a subject's body where transmittance methods of oximetry may yield inaccurate readings. Aspects of the disclosed subject matter may be used to estimate other physiological parameters such as heart rate, pulse transit time and blood pressure.

BACKGROUND

[0002] Pulse oximetry is a known technique for determining the oxygen saturation (SpO₂) of the hemoglobin of a subject, such as a medical patient. One type of oximeter is a transmittance oximeter. A typical transmittance oximeter is a clamp-like device having a red light source and an infrared light source on one side of the clamp and a light detector (photodetector) on the opposite side of the clamp. The device is placed at a site on a part of the patient's body, such as an earlobe or finger, so that the body part is between the sources and the detector. The oximeter illuminates the site alternately with red and infrared light from the sources. (As used throughout this specification, "light" is not limited to the visible portion of the electromagnetic spectrum.) Light emitted by each source enters the body part. Light which is not absorbed or otherwise dissipated penetrates through the body part and arrives at the detector. A processor estimates the patient's blood oxygen saturation in a well known manner as a function of the intensity of light received at the detector (which is an indication of the amount of light absorbed by the patient's tissue) in response to the red illumination and the infrared illumination during both pulsatile and nonpulsatile phases of the patient's heart cycle. The oximeter is able to distinguish between the pulsatile and nonpulsatile phases by detecting a characteristic absorption transient that accompanies the pulsatile phase.

[0003] One known way to estimate oxygen saturation is to calculate a value referred to as the modulation ratio, R , the equation for which is shown below.

$$R = \frac{AC_{RED}/DC_{RED}}{AC_{IR} / DC_{IR}} \quad (1)$$

DC_{RED} is a value related to the red light absorbed by the patient's body part, and therefore the intensity of red light detected by the detector during a nonpulsatile portion of a heart cycle. DC_{RED} depends on red light absorption due to venous blood, capillary blood, nonpulsatile arterial blood and other tissues (e.g. skin, fat, bone, muscle). DC_{RED} is substantially time invariant over the period of a heart cycle and is often analogized to the DC component of an alternating electrical current.

[0004] DC_{IR} is a value related to the infrared light absorbed by the patient's body part, and therefore the intensity of infrared light detected by the detector during a nonpulsatile portion of a heart cycle. DC_{IR} depends on infrared light absorption due to venous blood, capillary blood, nonpulsatile arterial blood and other tissues. DC_{IR} is substantially time invariant over the period of a heart cycle and, like DC_{RED} , is often analogized to the DC component of an alternating electrical current.

[0005] AC_{RED} is a value related to the red light absorbed by the patient's body part, and therefore the intensity of red light detected by the detector during a pulsatile portion of a heart cycle. AC_{RED} depends on red light absorption due to venous blood, capillary blood, nonpulsatile arterial blood, pulsatile arterial blood and other tissues. AC_{RED} changes during a heart cycle as the volume of pulsatile arterial blood in the vicinity of the oximeter changes. AC_{RED} is often analogized to the AC component of an alternating electrical current.

[0006] AC_{IR} is a value related to the infrared light absorbed by the patient's body part, and therefore the intensity of infrared light detected by the detector during a pulsatile portion of a heart cycle. AC_{IR} depends on infrared light absorption due to venous blood, capillary blood, nonpulsatile arterial blood, pulsatile arterial blood and other tissues. AC_{RED} changes during a heart cycle as the volume of pulsatile arterial blood in the vicinity of the oximeter changes. AC_{IR} , like AC_{RED} , is often analogized to the AC component of an alternating electrical current.

[0007] Because transmittance oximeters require some light to pass completely through a part of the patient's body, their use is limited to relatively thin parts of the body, for example at a fingertip or an earlobe. However it is sometimes desirable to acquire an SpO₂ reading by taking measurements at a site too thick for a transmittance oximeter, for example on a patient's chest. In such cases a reflectance oximeter is used. The operation of a reflectance oximeter is similar in principle to that of a transmittance oximeter. One difference is that a reflectance oximeter is designed so that the light sources and the photodetector are on the same side of the device rather than on opposite sides as in a transmittance oximeter. As a result, light need not pass all the way through the body part at which the SpO₂ reading is being taken. Instead the photodetector detects light which returns from the body part to the photodetector. Existing literature refers to the returned light as reflected light. As with a transmittance oximeter, a reflectance oximeter estimates SpO₂ as a function of the intensity of light received at the detector in response to red illumination and infrared illumination during both pulsatile and nonpulsatile phases of the patient's heart cycle. However these signals tend to be weak and noisy at the thicker body parts where a reflectance oximeter is typically used, making it difficult for the reflectance oximeter to make the necessary distinction between the pulsatile and nonpulsatile heartbeat phases.

[0008] Accordingly, what is needed is a pulse oximeter suitable for use at sites where the signals of interest are too weak and/or noisy to easily distinguish between pulsatile and nonpulsatile heartbeat phases. Aspects of the oximeter waveforms may be used for monitoring other physiological parameters.

SUMMARY

[0009] A method of estimating oxygen saturation includes illuminating a target site with light of a first wavelength, light of a second wavelength and light of a third wavelength. The second wavelength is greater than the first wavelength, and the third wavelength is greater than the second wavelength. Light returned from the site at each of the wavelengths is detected. The method distinguishes between pulsatility and nonpulsatility based on a first parameter of return light of the first wavelength. The method formulates a first estimate of oxygen saturation as a function of:

a second pulsatile parameter corresponding to the light returned in response to the illumination with the light of the second wavelength during a pulsatile period,

a second nonpulsatile parameter corresponding to the light returned in response to the illumination with the light of the second wavelength during a nonpulsatile period,

a third pulsatile parameter corresponding to the light returned in response to the illumination with the light of the third wavelength during a pulsatile period, and

a third nonpulsatile parameter corresponding to the light returned in response to the illumination with the light of the third wavelength during a nonpulsatile period.

BRIEF DESCRIPTION OF THE DRAWINGS

[0010] The foregoing and other features of the various embodiments of the oximeter described herein will become more apparent from the following detailed description and the accompanying drawings in which:

FIG. 1 is a schematic elevation view of an oximeter as described herein showing, among other things, a photodetector and a processor for formulating an estimate of a patient's oxygen saturation.

FIG. 2 is a plan view in the direction 2--2 of FIG. 1.

FIG. 3 is a block diagram of the methodology carried out by the processor of FIG. 1 acting according to machine readable instructions, the methodology being expressed in terms of light at wavelengths λ_1 , λ_2 and λ_3 .

FIG. 4 is a block diagram similar to that of FIG. 3 expressing the method in terms of green, red and infrared light.

FIG. 5 is a pair of graphs giving an example of a pulse illumination sequence carried out by the oximeter of FIGS. 1-2, and showing return signals which result from the illumination and are detected by the photodetector.

FIG. 6 is a plan view similar to that of FIG. 2 showing an alternate embodiment of the oximeter in which the photodetector is an array of three photodetectors, each sensitive to one of three wavelengths of light.

FIG. 7 is a block diagram of the methodology carried out by the processor of the oximeter of FIG. 6 acting according to machine readable instructions, the methodology being expressed in terms of light at wavelengths λ_1 , λ_2 and λ_3 .

FIGS. 8 and 9 are graphs similar to those of FIG. 5 showing non-pulsed illumination carried out by the oximeter of FIG. 6, and also showing sampling of the return signals resulting from the illumination.

FIG. 10 is a plan view similar to that of FIG. 2 showing a heart rate monitor which employs aspects of the oximeter of FIGS. 1-2.

FIG. 11 is a block diagram showing operation of the heart rate monitor of FIG. 10.

FIG. 12 is a graph of light intensity vs. time showing light intensity exceedance of a threshold used in the method of FIG. 11.

FIG. 13 is a graph of light intensity vs. time showing an example of a spurious light intensity nonexceedance during a time interval generally characterized by exceedance spikes and also showing an example of a spurious light intensity exceedance during a time interval generally characterized by nonexceedance.

FIG. 14 is a block diagram of a methodology which uses aspects of the oximeter described herein and of the oximeter methodology for estimating physiological parameters other than oxygen saturation.

DETAILED DESCRIPTION

[0011] Reference will now be made to embodiments of the invention, examples of which are illustrated in the accompanying drawings. Features similar to or the same as features already described may be identified by the same reference numerals already used. The terms "substantially" and "about" may be used herein to represent the inherent degree of uncertainty that may be attributed to any quantitative comparison, value, measurement or other

representation. These terms are also used herein to represent the degree by which a quantitative representation may vary from a stated reference without resulting in a change in the basic function of the subject matter at issue.

[0012] FIGS. 1-2 show an oximeter **20** which is an instrument for assessing the oxygen saturation of blood hemoglobin. As illustrated, the oximeter is in contact with a patient's chest **C**. The oximeter includes a housing **22** with a light emission aperture **24** and a light return aperture **26**, a light emitter **32** in the form of three light sources **32-1**, **32-2**, **32-3**. One example of a suitable light source is a light emitting diode. When powered, light source **32-1** emits light at a first wavelength λ_1 , (also referred to as λ_1 light). Light source **32-2** emits light at a second wavelength λ_2 (also referred to as λ_2 light) longer than λ_1 . Light source **32-3** emits light at a third wavelength λ_3 (also referred to as λ_3 light) longer than λ_2 . As used herein, "light" is not limited to the visible portion of the electromagnetic spectrum. A reference herein to a specific wavelength (or "color" such as green, red and infrared) means a wavelength band or "color" band narrow enough to approximate the specified wavelength, e.g. within plus or minus five nanometers. The oximeter also includes a driver **36** to drive the light sources in a successive manner.

[0013] The oximeter also includes a photodetector **38**, which is also referred to as a light detector. The photodetector detects a property or parameter of interest of light which returns to the oximeter after having been emitted by the emitter.

[0014] The oximeter also includes an on-board processor **40** and executable instructions **42** stored in an on-board memory **44**. Alternatively the processor and/or memory may be remote components with which the processor communicates by, for example, wireless technology. Either way, the processor and instructions cooperate to formulate an estimate of a patient's oxygen saturation and, as necessary, to control operation of the oximeter. In this specification, reference to the operation or functioning of the processor should be understood to mean operation as directed by the instructions.

[0015] It may be necessary to transduce the detected light signal to produce a signal useable by processor **40**. The illustrated oximeter includes an analogue to digital (A/D) converter **52** for that purpose. The A/D converter produces a digital electrical signal whose magnitude is

proportional to or otherwise depends on a parameter or property of light, such as intensity, received at the photodetector. However, for explanatory simplicity, this specification describes the functioning of the processor as if it were operating directly on the parameter or property of the return light signal itself, for example light intensity, rather than on a transduced signal.

[0016] FIG. 3 is a block diagram of the methodology carried out by processor 40 acting according to instructions 42. The methodology of FIG. 3 is expressed in terms of λ_1 light, λ_2 light, and λ_3 light where $\lambda_3 > \lambda_2 > \lambda_1$. In one specific embodiment first wavelength λ_1 is in the green-blue portion of the electromagnetic spectrum (approx. 480 to 570 nanometers) second wavelength λ_2 is in the red portion of the electromagnetic spectrum (approximately 650-700 nanometers) and third wavelength λ_3 is in the infrared portion of the electromagnetic spectrum (approximately 720 nanometers to 1 millimeter). The block diagram of FIG. 4 is the same as that of FIG. 3 except that the text of selected blocks is reworded to express the method in terms of green light, red light, and infrared light. In addition, the example of FIG. 4 is based on return light intensity as the parameter of interest. The following explanation of the methodology refers to the diagram of FIG. 4 rather than the more general diagram of FIG. 3.

[0017] Referring to FIGS. 4-5, at block 100 the method illuminates the target site S (FIG. 1) with green light from light source 32-1. At block 102 the method illuminates the target site with at least one cycle of red light from light source 32-2 and infrared light from light source 32-3, in either order.

[0018] The top portion of FIG. 5 is a prophetic example of an illumination sequence over a time span of one cardiac cycle. The height of the illumination pulses is not significant except to assist the reader in distinguishing among pulses of different wavelengths. In the illustrated sequence, pulses of λ_1 (green) light are emitted once every 0.250 msec (4000 Hz.). One or more cycles of alternating pulses of λ_2 (red) light and λ_3 (infrared) light, in either order, is emitted and nested between successive pulses of λ_1 light. If the pulses are equally spaced in time, emitter 32 emits one pulse of light, either green light, red light, or infrared light, every .083 msec. Other light pulse rates may be satisfactory provided that the light pulse rate is significantly faster than a typical cardiac pulse rate.

[0019] Although FIG. 5 illustrates one cycle of λ_2 and λ_3 light between each set of λ_1 pulses, two or more cycles can be used if desired.

[0020] The emitted light is directed at a target site **S** on a patient's body (FIG. 1) so that the target site is successively illuminated with green light, red light and infrared light. Some of the incident light may be absorbed by the patient's tissue, however the light not absorbed or otherwise dissipated at each wavelength is returned to the oximeter. The existing literature refers to the returned light as reflected light.

[0021] At block 104 photodetector 38 detects a property or parameter of interest of return light signal **S_R** (FIG. 1) at each wavelength. The parameter of the λ_1 light (green light) which is monitored for and detected, the parameter of the λ_2 light (red light) which is monitored for and detected, and the parameter of the λ_3 light (infrared light) which is monitored for and detected can all be the same parameter, can all be different parameters, or can be a mix of same and different parameters. One useful parameter or property is light intensity, which is power per unit area or, equivalently, energy per unit time per unit area.

[0022] The bottom portion of FIG. 5 shows a set of bars representing a property of the detected return signals resulting from the illumination pulses at the top of the illustration. The heights of the red and infrared return light pulses are not significant except to indicate, by their smaller height in comparison to the heights of the red and infrared emitted signal pulses, that a property of the return signals differs from that property of the emitted light signals (e.g. the return signal is less intense). The heights of the green return light pulses are significant in that they show that the intensity of the return green light pulses varies significantly in response to blood pulsations over time.

[0023] At block 130, the processor, operating as directed by instructions 42, formulates a first estimate of oxygen saturation as a function of:

- 1) a second pulsatile parameter corresponding to the light returned to the detector and detected by the detector in response to the illumination with the light of the second wavelength (λ_2 or red) during a pulsatile period,

2) a second nonpulsatile parameter corresponding to the light returned to the detector and detected by the detector in response to the illumination with the light of the second wavelength during a nonpulsatile period,

3) a third pulsatile parameter corresponding to the light returned to the detector and detected by the detector in response to the illumination with the light of the third wavelength (λ_3 or infrared) during a pulsatile period, and

4) a third nonpulsatile parameter corresponding to the light returned in response to the illumination with the light of the third wavelength during a nonpulsatile period.

[0024] In the previous paragraph, "second" and "third" are used to distinguish between the parameter of returned light which is the result of illumination with λ_2 light and the parameter of returned light which is the result of illumination with λ_3 light, and to distinguish those parameters from a first parameter (described below) which is the result of illumination with λ_1 light. In other words, "first", "second", and "third" are used to associate the parameters with the corresponding wavelength of illumination. "Pulsatile" is used to indicate that the parameter corresponds to a period of time during which cardiac systole has caused a pulse or surge of arterial blood at target site **S**. "Nonpulsatile" is used to indicate that the parameter corresponds to a period of time during which the pulse or surge of arterial blood is not present at target site **S** as a result of cardiac diastole. The second and third pulsatile parameters are the **AC_{RED}** and **AC_{IR}** components of equation **1** in the background section of this specification. The second and third nonpulsatile parameters are the **DC_{RED}** and **DC_{IR}** components of equation **1**.

[0025] Because the illustrated reflectance pulse oximeter is used at a thick or bulky portion of the patient's body, the return signals arising from illumination with red light and infrared light tend to be weak and noisy. This makes it difficult to distinguish between the pulsatile and nonpulsatile heartbeat phases or periods as is required according to substeps 1 and 3 above (pulsatile phase) and 2 and 4 (nonpulsatile phase). This difficulty is overcome by using the λ_1 return light signal to distinguish between pulsatility and nonpulsatility, where λ_1 is a wavelength whose return signal is stronger and less noisy than the λ_2 and λ_3 return light signals. In other words, the amplitude of a parameter or property of the λ_1 light is used to distinguish between pulsatility and nonpulsatility. One parameter on whose basis the pulsatile and nonpulsatile

periods may be identified is the intensity of the return light pulses when λ_1 corresponds to green or blue light. Blood absorbs more light in the green/blue portion of the spectrum than surrounding tissues. Therefore during a period of pulsatility (a period during which a relatively large volume of blood is in the field of view of the photodetector) the intensity of the pulses of reflected green or blue light will be relatively low. Conversely, during a period of nonpulsatility (a period during which a relatively smaller volume of blood is in the field of view of the photodetector) the intensity of the pulses of reflected green or blue light will be relatively high. If the intensity is higher than a nonpulsatility threshold over some period of time, that period is identified as a nonpulsatile period. If the intensity is lower than a pulsatility threshold over some period of time, that period is identified as a pulsatile period. Spurious intensity readings which are lower than the nonpulsatility threshold during a time interval which is generally characterized by exceedance of the nonpulsatility threshold may be disregarded. In other words the time interval is characterized by a preponderance of indications of nonpulsatility. Similarly, spurious intensity readings which are higher than the pulsatility threshold during a time interval which is generally characterized by non-exceedance of the pulsatility threshold may be disregarded. In other words the time interval is characterized by a preponderance of indications of pulsatility.

[0026] In order to distinguish between pulsatility and nonpulsatility the processor uses λ_1 light in the blue to green portion of the electromagnetic spectrum, i.e. in approximately the 480 to 570 nanometer portion of the electromagnetic spectrum. The present example is expressed in terms of green light (approximately 550 - 570 nanometers). As described in more detail below, the processor identifies a period of time during which the λ_1 reflected light is of relatively low intensity as a period of pulsatility and identifies a period during which the λ_1 light is of relatively high intensity as a period of nonpulsatility.

[0027] At block **108** the method assesses whether or not the history of the intensity of the return green light indicates a period of pulsatility. As seen in the graph at the bottom of FIG. 5, return green light pulses of a relatively high intensity (above the nonpulsatility threshold indicated by dashed line **60**) indicates a period of nonpulsatility, while return green light pulses of a relatively low intensity (below the pulsatility threshold indicated by dash-dot line **62**) indicates a period of pulsatility. The example also addresses the possibility that within some range of intensity of the return green light signal (between the dashed line and the dash-dot line)

it may not be possible to assess pulsatility and nonpulsatility with the confidence required by the method.

[0028] At block **108**, if the method assesses that the green light return signal signifies a period of pulsatility, the method proceeds to block **120** and recognizes the red and infrared return signals from that same time period as pulsatile signals. Those signals are therefore eligible to be used as the second pulsatile and third pulsatile parameters (i.e. the **AC_{RED}** and **AC_{IR}** signals) in the formulation of the first estimate of blood oxygen saturation at block **130**.

[0029] At block **108**, if the method assesses that the green light return signal signifies a period of nonpulsatility, the method proceeds to block **124** and recognizes the red and infrared return signals from that same time period as nonpulsatile signals. Those signals are therefore eligible to be used as the second nonpulsatile and third nonpulsatile parameters (i.e. the **DC_{RED}** and **DC_{IR}** signals) in the formulation of a first estimate of oxygen saturation at block **130**. The method as diagrammed reflects the tacit assumption that the period of pulsatility and the period of nonpulsatility are close enough together in time that no meaningful change in oxygen saturation could have occurred. If desired the instructions **42** can be written in a way that enforces a desired limit on the time lapse between the information from blocks **120** and **124** used at block **130**.

[0030] At block **108**, if the method is unable to confidently assess that the green light return signal signifies either a period of pulsatility or a period of nonpulsatility, the method advances to block **130** and declines to recognize those signals as eligible to be used at block **130**.

[0031] The first estimate of oxygen saturation from block **130** may be output at path **132** to a destination, for example a display or an electronic medical record.

[0032] If desired, the method may proceed to block **160** where it formulates a second estimate of oxygen saturation based on the first estimate of oxygen saturation and a calibration relationship **162**. The calibration relationship is based on testing in which the oxygen saturation of a population of volunteer test subjects, as determined by the method up to block **130**, is compared to the actual arterial oxygen saturation, SaO_2 , of those same individuals. The calibration relationship adjusts the first estimate of oxygen saturation for factors that the method

through block **130** is unable to account for. The second estimate of oxygen saturation may be output at path **164** to a destination, for example a display or an electronic medical record.

[0033] The steps of the method, including the first formulating step of blocks **130** and **160**, the assessment step at block **108**, and the illumination steps **100**, **102**, are carried out concurrently and unconditionally. That is, the method carries out the λ_1 , λ_2 , and λ_3 illumination sequence as already described while simultaneously detecting the return signals (block **104**), assessing pulsatility and nonpulsatility (block **108**) and continually updating the formulations of oxygen saturation (blocks **130**, **160**) with updated information from blocks **120** and **124**. Older information from blocks **120** and **124** may be systematically replaced by newer information from blocks **120** and **124** so that the formulations of blocks **130** and **160** are up to date.

[0034] FIG. **6** shows an oximeter similar to that of FIGS. **1-2** but in which photodetector **38** is an array of three photodetectors, **38-1** sensitive to only λ_1 light, **38-2** sensitive to only λ_2 light, and **38-3** sensitive to only λ_3 light.

[0035] FIGS. **7-9** illustrate an alternative method of estimating blood oxygen saturation using the detector of FIG. **6**. As seen in FIG. **8** the alternative method continuously illuminates the target site **S** (FIG. **1**) with λ_1 light, λ_2 light, and λ_3 light rather than using discrete light pulses as seen at the top of FIG. **5**. The photodetector array detects return light at each of the three wavelengths, however detector **38-1** detects only λ_1 return light. Detectors **38-2** and **38-3** similarly detect only λ_2 light and λ_3 light respectively.

[0036] At blocks **200**, **202**, **204**, of FIG. **7** processor **40**, acting according to instructions **42**, samples the readings from photodetectors **38-1**, **38-2** and **38-3** at rates significantly faster than a typical cardiac pulse rate, for example once every 0.25 msec. The sample sequence and timing illustrated in FIG. **9** are the same as the light pulse sequence and timing at the top of FIG. **5**.

[0037] At block **208** the method carried out by the processor according to instructions **42** distinguishes between pulsatility and nonpulsatility based on the λ_1 return light signal. If the method assesses that the λ_1 return light signal indicates a period of pulsatility, it recognizes the parameter or property (e.g. intensity) of the λ_2 return light and the λ_3 return light detected during that same period as pulsatile parameters and, at block **210**, admits those pulsatile parameters to

the formulation step at block 250. If the method instead assesses that the λ_1 return light signal indicates a period of nonpulsatility, it recognizes the parameter or property of the λ_2 return light and the λ_3 return light detected during that same period as nonpulsatile parameters and, at block 212, admits those nonpulsatile parameters to the formulation step at block 250. If the method is unable to distinguish between pulsatility and nonpulsatility it declines to use those parameters in the estimate of oxygen saturation (block 214).

[0038] At block 250 the method formulates the first estimate of oxygen saturation as a function of:

- a) one or more samples of a property of the second wavelength return light signal (λ_2 or red) corresponding to a period of pulsatility,
- b) one or more samples of a property of the second wavelength return light signal corresponding to a period of nonpulsatility,
- c) one or more samples of a property of the third wavelength return light signal (λ_3 or infrared) corresponding to a period of pulsatility and,
- d) one or more samples of a property of the third wavelength return light signal corresponding to a period of nonpulsatility.

[0039] At block **260** the method may formulate a second estimate of oxygen saturation based on the first estimate of oxygen saturation and a calibration relationship **270** as already described in connection with FIG. **4**. The first and second estimates of oxygen saturation may be output at paths **270**, **272** to a destination, for example a display or an electronic medical record.

[0040] The detection of periods of pulsatility based on a property or parameter of the λ_1 return light can be used in other contexts or independently to estimate other physiological attributes. FIG. **10** shows a heart rate sensor similar to the oximeter of FIGS. **1-2** but whose light emitter **32** includes only a single light source **32-1**. The light source emits light at a specified wavelength in the blue-green portion of the electromagnetic spectrum. In a specific embodiment the light source emits light at a wavelength in the green portion of the electromagnetic spectrum.

[0041] Referring additionally to FIGS. **11-12**, the method, carried out by processor **40** in accordance with instructions **42**, illuminates a target site on a patient with light of a defined wavelength, for example green light (block **300**). The illumination may be continually ongoing pulsed illumination as described in connection with the oximeter of FIGS. **1-2** and as seen in the illumination sequence at the top of FIG. **5**. Alternatively the illumination may be continually ongoing continuous illumination as described in connection with the oximeter of FIG. **6** and as seen in FIG. **8**.

[0042] At block **302** the method detects a property of the return light, for example the intensity of the return light pulses. At block **304** the method assess whether or not the measurement of the detected property indicates a period of pulsatility. If so, the method advances to block **306** where it increments a cumulative count of pulsatility periods and records the time of the pulsatility period, for example by recording the time (e.g. time t_2) midway between the first and last return light pulses in a group of pulses. At block **308** the method calculates heart rate as the ratio of a quantity of the pulsatility period counts and the time interval over which those pulsatility period counts occurred. For example the heart rate revealed by FIG. **12** is one beat per $t_4 - t_2$ units of time. Although this example is based on periods of pulsatility, periods of nonpulsatility could be similarly used to determine cardiac pulse rate.

[0043] In FIG. **11** the **NO** path and one branch of the **YES** path from block **304** are dead-ended rather than returning to block **300** or **302** to emphasize that the illumination and detection steps are carried out unconditionally, whether a period of pulsatility has occurred or not.

[0044] FIG. **12** shows that because the pulse rate of the green light (or the sampling rate of steady green light) is much faster than a cardiac cycle (systole to systole or diastole to diastole) a single occurrence of systole will correspond to a group of several consecutive pulses which satisfy the pulsatility criterion (non-exceedance of threshold **62**). Groups **G₂** and **G₄** are examples of such groups. Because there are no intervening return light pulses within each of these groups which fail to satisfy the pulsatility criterion, each group, **G₂**, **G₄**, is counted as a single systole event and is assigned a representative time, e.g. t_2 , t_4 . By contrast, group **G₂** is separated from group **G₄** by a group **G₃** of return light pulses which satisfy the nonpulsatility criterion (exceedance of threshold **60**). Pulse groups **G₂** and **G₄** are therefore recognized as

separate systole events which legitimately corresponds to a heart cycle. Referring to FIGS. 10 and 13, instructions 42 carried out by processor 44 can account for spurious troughs or valleys **V** that accompany the spikes of a genuine diastole event (return light pulse group **G₅**) and for spurious peaks **P** that accompany a genuine systole event (group **G₆**).

[0045] FIG. 14 is a block diagram showing a method for using aspects of the oximeter and of the oximeter methodology for estimating physiological parameters other than oxygen saturation. At block 400-1 the method illuminates a first site on the subject's body with light from the green-blue portion of the electromagnetic spectrum. At block 400-2 the method illuminates a second site on the subject's body with light from the green-blue portion of the electromagnetic spectrum. The wavelength of light used to illuminate the first site need not be the wavelength of light used to illuminate the second site. However, one suitable wavelength is green light at about 550 to 570 nanometers.

[0046] At block 402-1 the method detects a first property of light returned from the sites as a result of the illumination of the first site. At block 402-2 the method detects a second property of light returned from the second site as a result of the illumination of the second site. "First" and "second", are used to distinguish between the property monitored for and detected as a result of illumination of the first site and the property monitored for and detected as a result of illumination of the second site. In other words, "first" and "second" are used to associate the parameters with the site of illumination. One suitable parameter is light intensity. The first property need not be the same as the second property.

[0047] At block 404-1 the method determines whether pulsatility is indicated at the first site based on the detected first property. At block 404-2 the method determines whether pulsatility is indicated at the second site based on the detected second property. The determination can be made with the technique described previously in connection with oxygen saturation, i.e. by using the amplitude of the return light signals as an indicator of pulsatility and nonpulsatility.

[0048] At block 406-1 the method sets a value, TIME1, to the time of pulsatility at the first site. At block 406-2 the method sets a value, TIME2, to the time of pulsatility at the second site.

[0049] At block **412** the method calculates the absolute value of the time difference, $\Delta t = |TIME2 - TIME1|$. This time interval may also be referred to as PTT (pulse transit time). If it is known which time will be later, the absolute value operation can be dispensed with provided the minuend is greater than the subtrahend, so that a positive PTT is calculated. In addition, the method should ascertain that the first time and the second time differ from each other by less than a designated value of a cardiac cycle interval. In one example the designated cardiac cycle interval is the shortest time between heartbeats expected to be encountered in service, for example a three standard deviation interval. The value of PTT may be output on path **414** to a destination such as a display or an electronic medical record.

[0050] If desired, the method may advance to a block **420** and/or a block **430**. At block **420**, arterial compliance, AC, may be determined based on a relationship between arterial compliance and PTT. At block **430**, systolic and/or diastolic blood pressure may be determined based on a relationship between blood pressure and PTT. The values of AC and blood pressure may be output on paths **422**, **432** respectively to a destination such as a display or an electronic medical record.

[0051] Although this disclosure refers to specific embodiments, it will be understood by those skilled in the art that various changes in form and detail may be made without departing from the subject matter set forth in the accompanying claims.

CLAIMS:

1. A method of estimating oxygen saturation comprising:

illuminating a target site with light of a first wavelength, light of a second wavelength and light of a third wavelength, the second wavelength being greater than the first wavelength and the third wavelength being greater than the second wavelength;

detecting light returned from the site at each of the wavelengths;

distinguishing between pulsatility and nonpulsatility based on a first parameter of the return light of the first wavelength; and

formulating a first estimate of oxygen saturation as a function of:

a second pulsatile parameter corresponding to the light returned in response to the illumination with the light of the second wavelength during a pulsatile period,

a second nonpulsatile parameter corresponding to the light returned in response to the illumination with the light of the second wavelength during a nonpulsatile period,

a third pulsatile parameter corresponding to the light returned in response to the illumination with the light of the third wavelength during a pulsatile period, and

a third nonpulsatile parameter corresponding to the light returned in response to the illumination with the light of the third wavelength during a nonpulsatile period.

2. The method of claim **1** comprising formulating a second estimate of oxygen saturation based on the first estimate of oxygen saturation and a calibration relationship.**3.** The method of claim **1** wherein the step of formulating a first estimate of oxygen saturation is carried out concurrently with the illuminating step.

4. The method of claim **1** wherein the first parameter, the second pulsatile parameter, the second nonpulsatile parameter, the third pulsatile parameter and the third nonpulsatile parameter are all the same parameter.
5. The method of claim **4** wherein the same parameter is light intensity.
6. The method of claim **1** wherein the first wavelength is in the blue to green portion of the electromagnetic spectrum, the second wavelength is in the red portion of the electromagnetic spectrum, and the third wavelength is in the infrared portion of the electromagnetic spectrum.
7. The method of claim **1** wherein the first wavelength is in the visible green portion of the electromagnetic spectrum, the second wavelength is in the visible red portion of the spectrum and the third wavelength is in the infrared portion of the spectrum.
8. The method of claim **1** wherein the step of illuminating the target site is carried out with pulses of light of the first, second, and third wavelengths such that the pulses of light of the second and third wavelengths are nested between two successive pulses of light of the first wavelength.

9. A blood oxygenation assessment instrument comprising:

an emitter for emitting light of a first wavelength, a second wavelength, and a third wavelength, the second wavelength being greater than the first wavelength and the third wavelength being greater than the second wavelength;

a light detector for detecting a return light signal at each of the first, second and third wavelengths; and

a processor and executable instructions which, when executed by the processor, cause the processor to:

1) distinguish between pulsatility and nonpulsatility based on the first wavelength return light signal; and

2) formulate a first estimate of oxygen saturation based on:

a) a property of the second wavelength return light signal corresponding to a period of pulsatility;

b) a property of the second wavelength return light signal corresponding to a period of nonpulsatility;

c) a property of the third wavelength return light signal corresponding to a period of pulsatility;

d) a property of the third wavelength return light signal corresponding to a period of nonpulsatility.

10. The instrument of claim **9** wherein the property is light intensity.

11. A method of estimating a physiological attribute of a subject comprising.

illuminating a first site and a second site on the subject's body with light from the green-blue portion of the electromagnetic spectrum;

detecting a first property of light returned from the sites as a result of the illumination of the first site and detecting a second property of light returned from the second site as a result of the illumination of the second site and;

assessing whether or not the detected first property indicates pulsatility at a first site, and assessing whether or not the detected second property indicates pulsatility at the second site; and

if pulsatility is indicated at the first and second sites and occurs at the first site at a first time and occurs at the second site at a second time which differs from the first time by less than a designated cardiac cycle interval, defining the difference between the first and second times as a measure of pulse transit time between the two sites.

12. The method of claim **11** wherein the first property and the second property are the same property.

13. The method of claim **11** wherein the property is light intensity.

14. The method of claim **11** wherein the illumination at the first and second sites is carried out at substantially the same wavelength.

15. The method of claim **11** comprising determining arterial compliance as a function of the pulse transit time.

16. The method of claim **11** comprising determining systolic blood pressure, diastolic blood pressure or both as a function of the pulse transit time.

17. A method of estimating the heart rate of a subject comprising:

illuminating a site on the subject's body with light from the green-blue portion of the electromagnetic spectrum;

detecting a property of light returned from the site as a result of the illumination;

assessing whether the detected property indicates a period of pulsatility;

counting occurrences of periods of pulsatility thereby establishing a pulsatility count over a time interval; and

dividing the pulsatility count by the length of the time interval thereby establishing an estimate of heart rate.

18. The method of claim **17** wherein the property is light intensity;

19. The method of claim **17** wherein the step of illuminating the site is limited to illuminating with light from the green portion of the electromagnetic spectrum.

20. The method of claim **17** including accounting for spurious indications that violate a pulsatility criterion during a time interval characterized by a preponderance of indications that conform to the pulsatility criterion.

21. A method of estimating the heart rate of a subject comprising:

illuminating a site on the subject's body with light from the green-blue portion of the electromagnetic spectrum;

detecting a property of light returned from the site as a result of the illumination;

assessing whether the detected property indicates a period of nonpulsatility;

counting occurrences of periods of nonpulsatility thereby establishing a nonpulsatility count over a time interval; and

dividing the nonpulsatility count by the length of the time interval thereby establishing an estimate of heart rate.

22. A heart rate monitor comprising:

a light emitter which emits light at a defined wavelength in the blue-green portion of the electromagnetic spectrum;

a light detector for detecting a property of a return light signal at the defined wavelength;

a processor and executable instructions which, when executed by the processor, cause the processor to:

assess whether or not the detected property indicates a period of pulsatility and, if so,

increment a cumulative count of pulsatility periods; and

calculate the ratio of the pulsatility counts and the time interval over which the pulsatility counts occurred.

23. A heart rate monitor comprising:

a light emitter which emits light at a defined wavelength in the blue-green portion of the electromagnetic spectrum;

a light detector for detecting a property of a return light signal at the defined wavelength;

a processor and executable instructions which, when executed by the processor, cause the processor to:

assess whether or not the detected property indicates a period of nonpulsatility and, if so,

increment a cumulative count of nonpulsatility periods; and

calculate the ratio of the nonpulsatility counts and the time interval over which the nonpulsatility counts occurred.

1/11




FIG. 1




FIG. 2

2/11




FIG. 3

3/11




FIG. 4




FIG. 5




FIG. 6

6/11






FIG. 7

7/11

**FIG. 8****FIG. 9**





FIG. 10

9/11

**FIG. 11**

10/11




FIG. 12




FIG. 13

11/11




FIG. 14

INTERNATIONAL SEARCH REPORT

International application No.
PCT/US2018/038764**A. CLASSIFICATION OF SUBJECT MATTER****A61B 5/1455(2006.01)i, A61B 5/00(2006.01)i**

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

A61B 5/1455; A61B 5/00; A61B 5/02; A61B 5/026; A61B 7/04

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Korean utility models and applications for utility models

Japanese utility models and applications for utility models

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)

eKOMPASS(KIPO internal) & Keywords: oxygen, saturation, heart rate, light, wavelength, function, pulsatile period, nonpulsatile period

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	KR 10-2017-0064906 A (HYUNDAI MOTOR COMPANY et al.) 12 June 2017 See paragraphs [0007]-[0049], claims 1-14 and figures 1-6.	1-10, 17-23
Y		11-16
Y	EP 0807402 A1 (OHMEDA INC.) 19 November 1997 See page 4, line 8 - page 8, line 27, claims 1-6 and figures 1-5.	11-16
A	US 2014-0275925 A1 (CARDIAC PACEMAKER, INC.) 18 September 2014 See paragraphs [0016]-[0072], claim 1 and figures 1-5.	1-23
A	US 2016-0361004 A1 (CHRONISENSE MEDICAL LTD.) 15 December 2016 See paragraphs [0045]-[0097], claims 1-11 and figures 1-12.	1-23
A	US 2011-0054336 A1 (JORNOD ALAIN) 03 March 2011 See paragraphs [0031]-[0047], claim 1 and figure 1.	1-23



Further documents are listed in the continuation of Box C.



See patent family annex.

* Special categories of cited documents:

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier application or patent but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art

"&" document member of the same patent family

Date of the actual completion of the international search

28 September 2018 (28.09.2018)

Date of mailing of the international search report

28 September 2018 (28.09.2018)

Name and mailing address of the ISA/KR

International Application Division

Korean Intellectual Property Office

189 Cheongsa-ro, Seo-gu, Daejeon, 35208, Republic of Korea



Facsimile No. +82-42-481-8578

Authorized officer

Kim, Yeonkyung

Telephone No. +82-42-481-3325



INTERNATIONAL SEARCH REPORT

Information on patent family members

International application No.

PCT/US2018/038764

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
KR 10-2017-0064906 A	12/06/2017	None	
EP 0807402 A1	19/11/1997	CA 2201577 A1 JP 1052408 A JP 2001-178709 A JP 3350521 B2 US 5766127 A	15/10/1997 24/02/1998 03/07/2001 25/11/2002 16/06/1998
US 2014-0275925 A1	18/09/2014	US 9101275 B2	11/08/2015
US 2016-0361004 A1	15/12/2016	CN 107920786 A EP 3307146 A1 EP 3307150 A1 EP 3307162 A1 US 2016-0360971 A1 US 2016-0360974 A1 US 2016-0360986 A1 US 2016-0361003 A1 WO 2016-199121 A1 WO 2016-199122 A1 WO 2016-199123 A1 WO 2016-199124 A1 WO 2018-025257 A1	17/04/2018 18/04/2018 18/04/2018 18/04/2018 15/12/2016 15/12/2016 15/12/2016 15/12/2016 15/12/2016 15/12/2016 15/12/2016 15/12/2016 15/12/2016 08/02/2018
US 2011-0054336 A1	03/03/2011	CN 102008299 A CN 102008299 B EP 2292141 A1 EP 2292141 B1 HK 1156201 A1 JP 2011-050745 A JP 5466115 B2 KR 10-1225849 B1 KR 10-2011-0025100 A TW 201124113 A TW I538658 B US 8496595 B2	13/04/2011 27/03/2013 09/03/2011 17/06/2015 20/12/2013 17/03/2011 09/04/2014 23/01/2013 09/03/2011 16/07/2011 21/06/2016 30/07/2013