

(12) STANDARD PATENT
(19) AUSTRALIAN PATENT OFFICE

(11) Application No. AU 2012222866 B2

(54) Title
Transcutaneous stimulation method and system

(51) International Patent Classification(s)
A61N 1/24 (2006.01) **A61N 1/375** (2006.01)
A61N 1/36 (2006.01) **G06F 15/00** (2006.01)

(21) Application No: **2012222866** (22) Date of Filing: **2012.03.02**

(87) WIPO No: **WO12/116407**

(30) Priority Data

(31) Number
61/448,378 (32) Date
2011.03.02 (33) Country
US

(43) Publication Date: **2012.09.07**
(44) Accepted Journal Date: **2014.04.24**

(71) Applicant(s)
Murdoch Childrens Research Institute

(72) Inventor(s)
Southwell, Bridget Rae;Hutson, John Medwyn;Tan, Andre Yi Feng

(74) Agent / Attorney
FB Rice, L 14 90 Collins ST, Melbourne, VIC, 3000

(56) Related Art
US 5871534
US 2006/0195153
US 2005/0075678

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property Organization
International Bureau



(10) International Publication Number

WO 2012/116407 A1

(43) International Publication Date
7 September 2012 (07.09.2012)

(51) International Patent Classification:
A61N 1/24 (2006.01) *A61N 1/375 (2006.01)*
A61N 1/36 (2006.01) *G06F 15/00 (2006.01)*

(21) International Application Number:
 PCT/AU2012/000212

(22) International Filing Date:
 2 March 2012 (02.03.2012)

(25) Filing Language: English

(26) Publication Language: English

(30) Priority Data:
 61/448,378 2 March 2011 (02.03.2011) US

(71) Applicant (for all designated States except US): MURDOCH CHILDRENS RESEARCH INSTITUTE [AU/AU]; The Royal Children's Hospital, Flemington Road, Parkville, Victoria 3052 (AU).

(72) Inventors; and

(75) Inventors/Applicants (for US only): SOUTHWELL, Bridget Ray [AU/AU]; 15 Evergreen Way, Gordon, Victoria 3345 (AU). HUTSON, John Medwyn [AU/AU]; 18 Nyora Street, East Malvern, Victoria 3145 (AU). TAN, Andre Yi Feng [AU/AU]; 12 Royal Oak Court, Mulgrave, Victoria 3170 (AU).

(74) Agent: FB RICE; Level 23, 200 Queen Street, Melbourne, Victoria 3000 (AU).

(81) Designated States (unless otherwise indicated, for every kind of national protection available): AF, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IS, JP, KE, KG, KM, KN, KP, KR, KZ, LA, LC, LK, LR, LS, LT, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.

(84) Designated States (unless otherwise indicated, for every kind of regional protection available): ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, ML, MR, NE, SN, TD, TG).

Published:

— with international search report (Art. 21(3))

(54) Title: TRANSCUTANEOUS STIMULATION METHOD AND SYSTEM




Figure 8A

(57) **Abstract:** A method of treating a waste evacuation dysfunction comprising administering transcutaneous electrical stimulation (TES) to at least one lower pelvic and/or sacral region for a specific treatment regimens. Also disclosed is a system for configuring a stimulation device to deliver transcutaneous electrical stimulation (TES) the system comprising: a computing device storing or having access to a plurality of TES settings and comprising a user interface to enable authorised selection of at least one of the TES settings for provision of TES by the stimulation device according to the at least one selected TES setting and the stimulation device communicatively coupled to the computing device to receive and store the selected at least one TES setting the stimulation device being of a size to be readily carried on a body and configured to selectively provide current to external electrode according to the one TES setting.

TRANSCUTANEOUS STIMULATION METHOD AND SYSTEM**FIELD**

Described embodiments relate generally to methods and systems for transcutaneous stimulation. More particularly, methods, systems, devices and apparatus for

5 transcutaneous stimulation in one or more of the lumbar, abdominal, lower pelvic and sacral regions are described to treat one or more dysfunctions associated with a body's ability to evacuate waste.

BACKGROUND

Reference to any prior art in this specification is not, and should not be taken as, an

10 acknowledgment or any form of suggestion that this prior art forms part of the common general knowledge in any country.

Reference in this specification to any prior publication (or information derived from it), or to any matter which is known, is not, and should not be taken as, an acknowledgment or admission or any form of suggestion that that prior publication (or information

15 derived from it) or known matter forms part of the common general knowledge in the field of endeavour to which this specification relates.

Bibliographic details of the publications referred to by author in this specification are collected alphabetically at the end of the description.

Waste elimination dysfunction can take many forms. For example, urinary incontinence,

20 intestinal incontinence or constipation can occur.

Treatment systems exist for treating constipation by providing electrical stimulus *via* subcutaneously implanted electrodes positioned around the lower bowel. Electrical stimulation provided using such electrodes can be used to sequentially activate muscle fibres around the bowel to force a peristaltic action to occur. However, such treatment

25 systems are undesirably invasive. Further, while such systems may have an immediate effect in assisting to evacuate the bowel, they do not necessarily address the cause of the

- 2 -

constipation. Importantly, this effect has not been described as long lasting or having an effect beyond the immediate time of electrical stimulation.

Intractable constipation and soiling are extremely common in the community, in both the young and old, and available treatments are generally uncomfortable, can cause social

5 distress for sufferers and are a significant drain on the health care system. Individuals that suffer from constipation and soiling who are young or old may also have psychological issues. In addition, constipation may be a side effect of some kinds of medication, such as opiates. Most laxative therapies are designed to either soften the stool or stimulate the bowel by chemicals in the lumen. Patients with chronic

10 constipation or intractable constipation may have failed other treatment methods including pharmaceutical treatment. Patients on therapies for other diseases in which constipation is a side effect of the medication may not be able to be co-administered pharmaceutical treatments for constipation. Non-invasive, non-drug-based treatment methodologies may be desired in such cases.

15 Sometimes constipation may be unrelated to diet or medications, and can be due to poor motility in the whole colon (Benninga *et al*, *J Pediatr Gastroenterol Nutr.*, 23:241-51, 1996; Hutson *et al*, *J Pediatr Surg.*, 31:580-583, 1996). A newly identified disorder, which is known as slow-transit constipation (STC), is not uncommon amongst children who fail standard medical therapy, and such children often have signs of colonic

20 dysfunction even at birth. (Shin *et al*, *J Pediatr Surg.*, 37:1762-1765, 2002).

Previous electrical stimulation involved non-daily stimulation (i.e. 12 sessions in 4 week period) of short periods (i.e. 20 min of treatment) using transcutaneous devices that have been developed for physiotherapy treatment of muscular injuries (Clarke *et al*, *J. Pediatr. Surg.*, 44:408-412, 2009.)

25 It is desired to address or ameliorate one or more disadvantages or shortcomings associated with existing treatment systems, methods or regimes, or to at least provide a useful alternative thereto.

Any discussion of documents, acts, materials, devices, articles or the like which has been included in the present specification is not to be taken as an admission that any or all of these matters form part of the prior art base or were common general knowledge in the field relevant to the present disclosure as it existed before the priority date of each claim 5 of this application.

Throughout this specification the word "comprise", or variations such as "comprises" or "comprising", will be understood to imply the inclusion of a stated element, integer or step, or group of elements, integers or steps, but not the exclusion of any other element, 10 integer or step, or group of elements, integers or steps.

SUMMARY

Non-invasive electrical stimulation apparatus, devices, methods and systems are described herein to be used specifically for the treatment of waste elevation dysfunctions 15 constipation and improved treatment regimens. This treatment may have long term benefits in some patients with continued improvement after the period of treatment by the non-invasive electrical stimulation.

Throughout this specification and the claims which follow, unless the context requires otherwise, the word "comprise", and variations such as "comprises" and "comprising", 20 will be understood to imply the inclusion of a stated integer or step or group of integers or steps but not the exclusion of any other integer or step or group of integers or steps.

Some embodiments relate to a method of treating a waste evacuation dysfunction, comprising administering transcutaneous electrical stimulation (TES) to at least one lower pelvic region and at least one sacral region for at least one treatment period per 25 day over a treatment term of at least one week, wherein the TES comprises interferential electrical stimulation. Some embodiments relate generally to treatment regimes for transcutaneous stimulation. More particularly, methods, devices, apparatus and systems are provided for treating constipation or another waste evacuation dysfunction by the administration of transcutaneous electrical stimulation to at least one lumbar and/or 30 abdominal region for a period of time on a daily basis or greater than 12 sessions in a 4 week period. The electrical stimulation may alternatively or additionally be administered to a lower pelvic and/or sacral region. Reference to "greater than 12 sessions" includes from about 12 to about 100 sessions such as about 13, 14, 15, 16, 17,

28 Mar 2014

2012222866

18, 19, 20, 21, 22, 23, 24, 25, 26, 27, 28, 29, 30, 31, 32, 33, 34, 35, 36, 37, 38, 39, 40, 41, 42, 43, 44, 45, 46, 47, 48, 49, 50, 51, 52, 53, 54, 55, 56, 57, 58, 59, 60, 61, 62, 63, 64, 65, 66, 67, 68, 69, 70, 71, 72, 73, 74, 75, 76, 77, 78, 79, 80, 81, 82, 83, 84, 85, 86, 87, 88, 89, 90, 91, 92, 93, 94, 95, 96, 97, 98, 99 or 100 sessions, for example, or even

5 more.

In some embodiments, the treatment may involve a single treatment session each day or multiple (e.g. 2 or 3) treatment sessions per day. The treatment sessions may be for periods of between about 10 and about 90 minutes or from about 20 and about 60 minutes. Other time periods include about 11, 12, 13, 14, 15, 16, 17, 18, 19, 21, 22, 23, 24, 25, 26, 27, 28, 29, 30, 31, 32, 33, 34, 35, 36, 37, 38, 39, 40, 41, 42, 43, 44, 45, 46, 47, 48, 49, 50, 51, 52, 53, 54, 55, 56, 57, 58, 59, 61, 62, 63, 64, 65, 66, 67, 68, 69, 70, 71, 72, 73, 74, 75, 76, 77, 78, 79, 80, 81, 82, 83, 84, 85, 86, 87, 88 and 89 minutes, for example.

In some embodiments, the daily (or greater than 12 sessions in a 4 week period) electrical stimulation regime may be performed as part of a longer-term treatment plan, in which the stimulation is performed daily or greater than 12 sessions in a 4 week period for between about 2 weeks and 2 to 3 months. This includes periods of about 3, 4, 5, 6, 7, 8, 9, 10, 11, 12 and 13 weeks, for example. The regime may also involve repeating the longer-term treatment within periods of about 4 months to two years. Such longer-term periods include about 4, 5, 6, 7, 8, 9, 10, 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 23 and 24 months, for example. The transcutaneous electrical treatment may provide a longer term effect of greater than 1 day, for example up to 1 month to 1 year or more. Hence, the effect of treatment lasts beyond the immediate time of electrical stimulation.

25 In some embodiments, where daily treatment (or greater than 12 treatment sessions in a 4 week period) is anticipated, for ease of administration, the treatment device is such that it can be used at home without the supervision of a trained professional during the daily (or greater than 12 sessions in a 4 week period) treatment regimen.

The stimulation may be performed using one to ten electrodes, for example, placed on the lumbar and/or lower front abdominal areas. Electrodes may be additionally or alternatively placed over the lower pelvic and/or sacral regions, including the buttocks or other tissue overlying or adjacent the ilium, to excite or modulate the nerves and other

tissues around the rectum, including the recto-sigmoid colon. In particular embodiments, an even number of electrodes are employed in a spaced pair arrangement. Alternatively, one, three, five or more electrodes may be employed. In some embodiments, the one to ten or more electrodes may be fixed in an apparatus such as a belt. The fixed positioning 5 of the electrodes in a device such as a belt may aid in the positioning of the electrodes to the lumbar and/or lower front abdominal areas and the spacing of the electrodes. The electrodes used to provide the electrical stimulation may be provided on a carrier that can be removably affixed, for example by adhesion, to a desired skin surface area to facilitate appropriate spacing of the electrodes from each other.

10 Some embodiments relate to a system for transcutaneously treating a waste evacuation dysfunction, comprising:

15 at least four electrodes comprising a first pair of electrodes positioned one on each lateral side of the spine or sacrum in the sacral region and a second pair of electrodes positioned laterally across the lower pelvic region;

a stimulation device configured to provide electrical stimulation signals to the at least four electrodes to perform the method according to some embodiments.

20 The electrodes may be coupled to and receive a stimulation current from a stimulation device having a low voltage, low current power source. The stimulation device may comprise a handheld portable device, for example, that can be operated without needing to be coupled to an external power supply. The stimulation device may be configured to be powered by a disposable and/or rechargeable battery or other small self-contained power source, for example. The power source of the stimulation device may be 25 rechargeable, for example by coupling it *via* a transformer to a mains power supply. The

stimulation device may be configured to prevent or minimise the current supply to the electrodes while the stimulation device is having its portable power source recharged.

The stimulation device may provide or consist of a primitive display, for example to indicate its on/off state, whether it is in use (i.e. providing stimulation signals), a display

- 5 of an accrued time of use and/or a remaining time for use in a particular session. The stimulation device may also be configured to prevent electrical stimulation being provided for more than a predetermined total amount of time for a day or 24 hour period or for more than a predetermined amount of time (e.g. 60, 70, 80 or 90 minutes) in any one usage session. Alternatively or in addition, the device may be configured to restrict
- 10 the total amount of electrical energy delivered to the stimulation electrodes over a usage session or a particular period of time.

The stimulation device may be free of external manually operable mechanisms but for an on/off button or switch and a stop/start button or switch. In some alternative embodiments, the stimulation device may have external manually operable mechanisms,

- 15 for example to interface with a device display, but may be free of any manually operable mechanisms to provide input to the stimulation device that would alter device settings or parameters according to which the electrical stimulation signals are provided.

In some embodiments, the stimulation device is configured to only operate according to a single set of operating parameters at a given time. In some embodiments, this single

- 20 set of operating parameters may only be replaced with another single set of operating

- 6 -

parameters using an electronic configuration interface separate from but communicably coupleable to the stimulation device. For example, the stimulation device may be provided with a default set of operating parameters by a manufacturer of the device, and this set of parameters may be subsequently modified by a therapist using software

- 5 authorized to reconfigure the settings of the stimulation device *via* a wired or wireless connection. In other embodiments, the stimulation device may be configurable (e.g. by a trained professional using authorised software) with multiple stored sets of stimulation settings for separate use by multiple users. Thus, the user interface of the stimulation device may be configured to be quite simple and to disallow user modification of the
- 10 settings, in order to facilitate ease and appropriateness of use of the device. However, in alternative embodiments, the stimulation device may have greater user interface functionality and may allow for one of two, three, four or more stimulation settings to be selected by a user.

In some embodiments, the stimulation device may be pre-configured or configurable to

- 15 provide output signals to stimulation electrodes having an approximately sinusoidal form, with an approximately 4 kilohertz carrier frequency, roughly 80 to 150 hertz modulated frequency and a current intensity of around 5 to around 33 millamps. Such stimulation signals may be applied to two or four or more electrodes, for example including two electrodes spaced across the lower front abdominal area and two
- 20 electrodes spaced across the lumbar area, applying interferential current stimulation from left front to right back and/or right front to left back.

In some embodiments, the stimulation frequency may be selected or configured to take account of the Body Mass Index (BMI) of the patient, which can be different for a child *versus* an adult, and also whether the patient is of normal weight, overweight, obese or

- 25 underweight.

Some embodiments relate to a system for configuring a stimulation device to deliver transcutaneous electrical stimulation (TES), the system comprising:

- 30 a computing device storing or having access to a plurality of TES settings and comprising a user interface to enable authorised selection of one of the TES settings for provision of TES by the stimulation device according to the one setting; and

- 7 -

the stimulation device communicatively coupled to the computing device to receive and store the selected one TES setting, the stimulation device being of a size to be readily carried on a body and configured to selectively provide current to external electrodes according to the one TES setting.

5

In some embodiments, the stimulation device may receive and store multiple selected TES settings for delivering TES to multiple users.

BRIEF DESCRIPTION OF THE DRAWINGS

Embodiments will now be described in further detail, by way of example, with reference

10 to the accompanying drawings, in which:

Figure 1 is a schematic diagram illustrating use of electrical stimulation to treat intestinal dysfunction in a child;

Figure 2 is a schematic diagram illustrating use of electrical stimulation to treat intestinal dysfunction in an adult;

15 **Figures 3A, 3B, 3C and 3D** are schematic diagrams of an example stimulation device in different views;

Figure 4 is a block diagram of the stimulation device;

Figure 5 is a schematic illustration of software controls applicable to use of the stimulation device;

20 **Figure 6** is a block diagram of a system for configuration of the stimulation device;

Figure 7 is a graphical representation showing before and after daily stimulation treatment for 1 month (Rx1) and 2 months (Rx2) in 11 patients:

- 25 A) Total number of defecations/week,
- B) Number of days with soiling, and
- C) Number of days with abdominal pain;

- 8 -

Figure 8A is a schematic illustration showing placement of electrodes in a lower pelvic region on a person's front side;

Figure 8B is a schematic representation of placement of electrodes in a sacral or lower lumbar region on a backside of the person;

5 **Figure 8C** is a schematic plan view illustrating crossing interferential currents between the lower pelvic and sacral or lower lumbar region electrodes;

Figure 9A is a schematic illustration showing placement of multiple frontal electrode pairs;

10 **Figure 9B** is a schematic illustration showing placement of multiple posterior electrode pairs; and

Figure 10 is a schematic diagram of a belt for assisting electrode placement of single or multiple frontal and posterior electrode pairs.

DETAILED DESCRIPTION

Embodiments herein relate generally to methods, systems, devices and treatment regimes 15 for treating or enabling the treatment of a waste elimination dysfunction, such as, for example and without limitation, constipation, ileus, urinary incontinence or intestinal incontinence. Such embodiments generally involve the application of transcutaneous electrical stimulation (TES) to a front or back (lumbar) abdominal region for at least one treatment session per day over consecutive days of a treatment period of at least one 20 week. TES may alternatively or additionally be applied to lower pelvic and/or sacral regions including the buttocks, to excite or modulate the nerves and other tissues around the rectum, including the recto-sigmoid colon.

The term "waste evacuation dysfunction" or "waste elimination dysfunction" used herein is intended to include dysfunction associated with a body's gastrointestinal and urinary 25 waste processing and/or elimination functions, whether or not the processing actually immediately results in evacuation or elimination of the waste from the body. For example, ileus or slow transit of waste through parts of the intestinal tract to the rectum

- 9 -

are considered to be waste evacuation dysfunctions in this sense because they affect the body's process of eliminating that waste from the body, even though the ileus or slow transit may be accurately described as a dysfunction preceding the actual elimination of that waste.

- 5 The treatment session may be performed multiple times per day or just once and may be performed for a time between about 10 and about 90 minutes for each session. In some embodiments, the treatment session may be between about 20 minutes and about 60 minutes, preferably closer to 60 minutes, such as 25, 30, 35, 40, 45, 50, 55, 65 or 70 minutes or other times in between.
- 10 In some embodiments, daily treatments are anticipated of greater than 12 sessions in a 4 week period (i.e. three sessions per week). For ease of administration, the treatment device is such that it can be used at home without the supervision of a trained healthcare professional during the daily (or greater than 3 sessions a week) treatment regimen.

The treatment term of at least one week may be, for example, between about 2 weeks 15 and about 3 months. In some embodiments, the treatment term may be between about 1 month and about 2 months.

The treatment term may be repeated over an extended term of from about 4 months to about 2 years or possibly more, in order to have the treatment suitably program, teach or train the various muscles and/or nerves responsible for proper function of the affected 20 organs or tissues. Thus, the treatment period may be repeated multiple times over the longer term, with the degree of repetition depending on physiological response to one or more initial treatment terms.

The treatment may have an effect beyond the immediate time of electrical stimulation which may last 1 day, for example up to 1 week, to 1 month to one year or more beyond 25 the last time of electrical stimulation.

Generally, as illustrated in Figures 1 and 2 in relation to a child 10 or adult 60, the electrical stimulation may be provided to electrodes 30 positioned over a front abdomen region 12 (on either side of the umbilicus 11) and/or on a back (lumbar) abdominal

region 14. The electrodes 30 receive electrical stimulation signals *via* conductors 32 to which they are coupled and convey these to the skin surface of the child 10 or adult 60 to which they are affixed or otherwise conductively positioned against. A suitable conductive gel may be used to increase conductivity of the electrical signals from

5 electrodes 30 into the body *via* the skin.

In some embodiments, four surface electrodes 30 may be used, two electrodes 30 being positioned one to either side of the umbilicus 11 on the anterior abdominal wall beneath the costal margin, and two electrodes 30 being positioned on the para-spinal area of T9-10 to L2. Positioning of the electrodes 30, whether four or more than four electrodes 30

10 are used, is intended to stimulate the proximal colon (including at least part of the ascending colon and the transverse colon) and at least an upper part of the descending colon, which generally correlates to the abdominal vicinity of the umbilicus 11.

Positioning of the electrodes 30 in the manner shown in Figures 1 and 2 is not intended to particularly affect the sigmoid colon or distal parts of the descending colon or rectum.

15 Thus, although the positioning of the electrodes 30 is applied to provide stimulation to parts of the large bowel closer to the costal margin, not necessarily all of the large bowel is to be stimulated in this way.

Lateral spacing of the electrode positions from the umbilicus 11 may be in the vicinity of 1, 2, 5 or 8 to 20 cm, for example, thereby providing a lateral separation between the

20 electrodes 30 of about 2, 4, 10 or 15 to 40 cm. Other lateral spacings within such ranges may be employed, as appropriate. The electrodes 30 may be positioned approximately level with the umbilicus 11, although some small variation of locations, for example slightly closer to or further from the costal margin, may be employed. The electrodes 30 positioned in the para-spinal area may be located substantially directly across the

25 abdomen from the frontal electrodes 30. In some embodiments, the electrodes 30 may be slightly offset from each other vertically or laterally across the pelvis and/or abdomen.

Electrodes 30 may be provided on a carrier 20 that comprises a flexible substrate conveniently positioning the electrodes 30 a fixed distance apart from each other to

30 assist in proper positioning of the electrodes in one or more regions 12, 14. The flexible substrate 20 may comprise adhesive substances on one or more portions thereof in order to facilitate removable application of electrodes 30 to the skin and retention of the electrodes 30 in a specific selected location. Each carrier 20 may comprise 1, 2, 3, 4, 5,

6, 7, 8, 9, 10 or more electrodes 30 in specific spaced relation. Once the electrodes 30 are appropriately positioned, either with or without the aid of a carrier 20, electrode conductor leads 32 are used to couple the conductors 30 to respective channels 139a, 139b (Figure 4) of a stimulation device 100.

- 5 A stimulation device 100 and its components and features are described in further detail below, with reference to Figures 3A, 3B, 3C, 3D and 4. Figure 3A is a schematic plan view of the device 100. Figure 3B is one end view of the device 100. Figure 3C is an opposed end view of the device 100. Figure 3D is a schematic perspective view of the device 100. Figure 4 is a block diagram of the device 100, showing components and
- 10 circuitry housed within a casing 105 of device 100. Stimulation device 100 is designed to be simple, portable and light, so that it can be worn by a user, possibly on or within a carrier belt, while it provides the stimulation signals to the electrodes 30 as the user performs normal activities.

The portable stimulator device 100 is designed to provide therapeutic electrical stimulation for individuals suffering dysfunction. Device 100 delivers a specified electrical signal to the patient through a set of electrodes that are placed on the skin surface of the abdomen, lower back, lower pelvis and/or sacral region. The device is designed to be as simple as possible, optionally in combination with a belt to assist electrode positioning (Fig. 10), so that those with relatively low technological sophistication, including some children and the elderly, may operate the device without clinical supervision (e.g. at home) and with minimal complication. It has been found that devices employing too much user interface sophistication are liable to be misused, with the result being that the patient may experience pain from improper treatment or may not receive the intended therapeutic benefit of the treatment. Some embodiments of device 100 thus provide a simplified on/off type of user interface, leaving any selection or reconfiguration of stimulation settings to be performed by a trained clinician or therapist using external software.

Device 100 may be provided with an accompanying software package for use by a patient or the patient's carer on a separate computer system 605 (Figure 6) to facilitate user education and instruction. Additionally, separate software may be provided for use by a clinician to allow the clinician to set or modify settings or functions of device 100 to further benefit patients.

The device 100 is configured to allow its internal software (firmware) to be easily updated. Should more effective treatment settings be determined, the device 100 can be updated through a firmware update so that patients can have access to different treatment regimes or settings.

- 5 The device 100 is small enough for a small child to carry, and variations of the design may feature child-friendly shapes and colours through changeable faceplates and covers, and also larger, Braille-studded or other tactile buttons and/or displays for vision-impaired or geriatric users.

Measures to provide device safety may include short-circuit protections, isolated

- 10 charging circuits to prevent mains power being delivered to a user in the event of device malfunction and conditional operation failsafe mechanisms (i.e. the device cannot be operated if electrodes 30 and/or electrode leads 32 are not connected or not correctly connected).

The device 100 may feature one or more of:

- 15 1. Two (or in some embodiments, more) independent electrical channels 139a, 139b, each of which is capable of providing a specified current, voltage and waveform characteristic *via* electrode leads 32 to electrodes 30.
- 20 2. A display 120 including a liquid crystal display (LCD) and/or an LED-based display or other form of display, responsive to signals received at display circuitry 138 from microprocessor 130 to display some or all of the following information:
 - a. ON/OFF status of the device 100.
 - b. The remaining battery life of the device 100.

- 13 -

- c. START/STOP status of the electrical stimulation.
- d. A timer to indicate elapsed and/or remaining time of stimulation for a treatment session.

3. An ON/OFF switch or button 122 to turn on or off the device 100.

5 4. A START/STOP switch or button 124 to turn on or off the electrical stimulation signals.

Device 100 may comprise various communication and power supply inputs, including (but not limited to):

- a. Stimulator electrode sockets 112 - these allow the electrode leads and wires to plug into the device. As a failsafe, the device 100 may be configured to not be operable unless the electrode leads and electrodes are correctly attached and connected. In some embodiments, more than two (e.g. four) sockets 112 may be provided and/or more than two (e.g. four) interferential current channels may be provided in order to be able to deliver interferential TES to multiple frontal and posterior electrode pairs.
- b. DC power supply socket 114 to supply the device and internal battery 142 with power. The power may be converted from mains power (110V/240V 60/50Hz) to a suitable DC voltage *via* a switched-mode power supply (SMPS) or other suitable electrical power converter.
- c. A computer interface port 116 (Universal Serial Bus, or other industry standard computer interface, wired or wireless) to allow authorized users to add, modify or change the function of the device, based on the level of authority.
- d. A reset button (not shown) that cannot be easily accessed, but be accessible, for example through battery compartment 140, should a device reset be required.

- 14 -

- e. An in-built speaker (not shown) to provide audible messages, beeps, alerts or other signals to vision-impaired users and younger users.

Device 100 may comprise a rigid casing 105 to house the electronics (e.g. on PCB 108) and may comprise environmental sealing to industry standards (i.e. rubber gaskets on 5 exposed connectors, rubber sealing within the device to stop liquids and other foreign material from breaching the device casing).

Device 100 may further comprise one or more accelerometers to detect and record movement and/or orientation of device 100, to thereby infer one or more states or situations of use. Additionally, device 100 may have means for sensing and recording 10 temperature, for example via the electrodes 30, in order to infer additional information concerning use of device 100. This patient usage information can then be used to assess patient compliance with the treatment regime.

Device 100 may, in some embodiments, comprise selection means, such as a button or buttons, dial, touch-pad or touch-screen, in cooperation with display 120, to vary the 15 current intensity of TES to be delivered via electrodes 30. This permissible variation may be limited to within a predetermined or pre-set range, for example. In some further embodiments, the selection means may be employed in combination with multiple personalised stimulation settings (for multiple different people) stored into the device, possibly by a clinician in consultation with the people for whom the TES is intended. 20 This may allow a single device 100 to be used in a setting where multiple users are present, such as a nursing home or other shared accommodation or treatment facility.

The physical size or external appearance of the device 100 may vary depending on the target market. It may:

- a. Be small enough to fit in a small bag that a child can easily carry, may 25 feature various case styles and designs to give it a child-friendly appearance, and make it more appealing to young users.
- b. Be large enough to be handled by geriatric users, and feature:

- 15 -

- i. Extra-large LCD (or other display) with larger symbols (if a display is provided).
- ii. Large buttons for easy operation.
- iii. Braille embedded on various parts of the device and tactile buttons to allow users with poor vision to operate it.

5

- c. Feature an exterior casing/faceplate that may be changeable to modify the physical appearance of the device so that the user may customise the "look" of the device to taste (i.e. a case that looks like a rabbit for young children, a case that looks like a car for a 10-yr old male user, a doll for a 10-yr old female user). This case is completely separate from the internal case 105 that houses the device's electronics.

The device 100 may comprise a battery compartment 140 defined by casing 105 and housing a (optionally rechargeable) battery 142 that can be coupled to an external power supply *via* socket 114. Battery 142 supplies a DC voltage, such as 9V, to power supply circuitry 134, which supplies power to the various electrical/electronic components of device 100.

The device 100 comprises a combination of programmable and non-programmable circuitry, digital or analog, embedded onto or coupled to at least one printed circuit board (PCB) 108. The circuitry includes, but is not limited to:

20 1. Signal generator circuitry 136 to produce the electrical waveforms provided to channels 139a, 139b. As a failsafe, signal generator circuitry 136 may be current-limited to prevent oversupply of current.

2. An on-board microprocessor system 130, which may comprise a suitable microcontroller, an Application-Specific Integrated Circuit (ASIC) and/or Field-
25 Programmable Gate Array (FPGA). The microprocessor system 130 has access to sufficient Read-Only Memory (ROM) 131 and Random-Access Memory (RAM) 132 to facilitate device operation, communication between external devices, firmware update functions and service/maintenance functions. Device 100 operates according to control

- 16 -

software (firmware) pre-programmed into the ROM 132 to facilitate operation and control of the device.

3. On-board load-testing circuitry (as part of, or controlled by and responsive to, microprocessor 130) to check that each and all electrodes are connected and in proper contact to an electrical load that is representative of human tissue.
4. In-built safety features to prevent or minimise unintended current being delivered to the patient. This may include (but not be limited to):
 - a. Charging circuit isolation *via* power supply circuitry 134 to prevent provision of electrical stimulation during a battery charge operation.
 - 10 b. Failsafe measures provided by the load-testing circuitry to prevent the provision of stimulation current from the device when improper electrical connection is detected, for example where electrodes are not connected to electrode leads, when electrode leads are not connected to the device 100, or when electrodes are incorrectly attached to the device 100 or when the electrodes are not connected to a human body.
 - 15 c. Short-circuit protections to prevent or mitigate delivery of unintended current to the patient.

The device software (firmware) and separate computer software (to be executed by computer system 605) may feature three modes of access by which three different classes of users can interact with it. These modes are described below and schematically illustrated in Figure 5.

Patient Mode: The user may operate the device for therapeutic purposes, but may not modify, change or delete device functions when interacting with the physical device. An exception to this is that the user may change device function only with an authorized software (firmware) update provided by the manufacturer. This update should be sufficiently tamper-proof to prevent user errors and device corruption.

- 17 -

The user may operate the personal computer software to access informational or help files to learn how to operate the device 100 or optionally to view usage statistics, for example, but may not modify, change or delete device functions when interacting with the device 100 through the personal computer interface 620 (Figure 6). This user-

5 focussed software may be provided when the device is purchased, either as a software CD (or other computer-readable medium) or electronic download from the manufacturer.

Clinician Mode: The clinician user may alter the device's function (in a restricted manner), for example to select different stimulation settings for device 100 to improve the therapeutic benefit of the device 100 to the patient by interacting only with the

10 clinician-user interface module executing on computer system 605. This may include authorized firmware updates as listed in the section relating to Patient Mode.

Personal Computer software for use by the clinician may not be supplied with the device, but instead may only be obtained once a clinician has registered with the manufacturer and been certified to make limited function modifications to the device for

15 therapeutic benefits only.

Technician/Service Mode: An authorized repair agent or manufacturer technician/engineer may access the core program of the device in order to facilitate diagnostics and repair functions. This mode allows for full/authorized modification to device function.

20 The software/firmware is split into two separate code modules that interact with each other:

1. Software programmed into the device 100 (firmware):

a. The firmware may be programmed into the device 100 at the time of its manufacture. The firmware has all three levels of functionality pre-programmed, but specific functions may only be accessed by licence holders or authorized persons as specified above.

2. Software programmed for use on the computer 605:

- a. The software may only interact with its authorized level, and any levels that it is authorized to interact with. For example, the Patient Mode on the device may only interact with the Patient Mode on the computer software, and the Technician Mode may interact with both the Patient Mode and Clinician Mode.
5
 - b. The software on the computer 605 facilitates the firmware update in an automated fashion to minimise complications in the upgrade process. This will also provide security verification so that the device cannot be tampered with through this access method.
- 10 Software and firmware updates may be provided from time to time by the manufacturer as required.

The device is intended to be operated in the following manner:

- The user attaches the electrode carrier 20 and electrode pads 30 (or just the electrode pads 30 if they're not provided on a carrier 20) to the lower front abdomen and lower back as shown in Figures 1 (child) and 2 (adult). The electrode carrier and/or electrode pads 30 may additionally or alternatively be attached in the manner shown in Figures 8A to 8C, 9A, 9B and 10, in which the electrode carrier comprises a wearable structure 910.
15

- The electrode leads 32 are connected to all of the electrode pads 30, and then to the correct sockets 112 on the device 100. Alternatively, leads 32 may be provided to a patient pre-connected to pads 30 and/or sockets 112.
20

- The user then switches on the device 100 using button 122, in response to which the device 100 performs a back-end function check to ensure that all systems are normal and that electrodes 30 and electrode leads 32 are connected correctly. During this time, the
25 LCD 120 may display a short message to the user that the device is starting up.

Once the device 100 is ready, the LCD 120 may indicate a message or signal (e.g. green LED lights up) to inform the user that stimulation is ready to start. A backlight of the START/STOP button 124 may light up and flash, and the user can then press the START/STOP button 124 to commence stimulation.

During current delivery, a counter or timer function executed by microprocessor 130 may cause display circuitry 138 coupled to display 120 to indicate the remaining stimulation time. Should the user wish to terminate current delivery, the user presses the START/STOP button 124. Pressing the ON/OFF button 122 in the ON state will also

5 terminate current delivery and switch the device 100 off.

Once current delivery has ceased, the device 100 may indicate on the LCD 120 that stimulation for the current treatment session is complete. Should the user not press the ON/OFF button 122, the device 100 may be configured to automatically switch itself off after a manufacturer-specified time to prevent accidental operation and minimise battery

10 consumption. The microprocessor system 130 may control device 100 to disallow further administration of stimulation for a period of several (e.g. up to 24) hours, regardless of whether device 100 is turned off. For this purpose, microprocessor 130 may comprise a suitable timer function (with possibly a long-term back-up power supply) that cannot be disrupted, even with the device 100 being switched off or the

15 battery 142 being removed or drained. This feature may be disabled or not present in embodiments of device 100 configured for use by multiple different users.

Device 100 may, in alternative embodiments, comprise greater user interface functionality than is described above, for example in order to enable a user to select from a number of stimulation settings, including carrier frequency, modulated frequency,

20 current intensity, duration of the treatment session, etc. However, regardless of the user interface functionality of device 100, it should be operable to provide a stimulation current of magnitude less than about 40 mA at a carrier frequency of between about 1 kHz and about 10 kHz, with a modulated frequency of about 20 to about 300 Hz. Preferably, the device 100 is configured to provide stimulation currents having a

25 magnitude of 33 to 40 mA or less (but more than zero) at a carrier frequency of about 4 kHz, with a modulated frequency of about 80 Hz to 150 Hz. The electrical stimulation may be provided as interferential electrical stimulation, for example from left front to right back and/or right front to left back.

In some embodiments, because the electric current is delivered transcutaneously to the

30 individual, the amount of electrical stimulation to be delivered to an individual may vary

- 20 -

from individual to individual based on either Body Mass Index (BMI) and /or circumference at waist and/or weight. For example, the optimal energy for electrical stimulation required for a normal weight child would be less than for an obese child. An algorithm for determining the electrical stimulation energy required, for example based 5 on age, BMI, weight, or circumference around the waist, may be included in the device firmware, together with operating parameters to allow specific electrical stimulation parameters to be initially set by the treating clinician or other medically trained professional.

One specific dysfunction that some embodiments are considered suitable for treating is 10 slow transit constipation (STC). However, described embodiments may be applied to treat various other waste evacuation dysfunctions, such as other types of constipation, incontinence, irritable bowel syndrome (IBS) and ileus, for example. In some embodiments, stimulation device 100 may be coupled to computer system 605, as shown in Figure 6. In the illustrated arrangement, computer system 605 and stimulation device 15 100 form part of a system 100 for facilitating configuration of device 100 and/or facilitating communication between stimulation device 100 and computer system 605 or a network 630 to which computer system 605 is coupled. Computer system 605 may comprise a desktop, laptop or handheld computing device having a processor 610, memory 615 and user interface 620. Processor 610 may comprise more than one 20 processing device and has access to memory 615 which comprises volatile and non-volatile storage for executing software functions as described herein. User interface 620 comprises normal peripheral devices and/or user interface functionality for facilitating user interaction with computer system 605 and may include suitable display-related software in addition to the normal display screen, keyboard, mouse, touch screen and/or 25 stylus, etc.

Computer system 605 comprises an input/output port 655 for communicating *via* a wired or wireless connection 660 with port 116 of stimulation device 100, thus enabling processor 610 to reconfigure (to the extent permitted) or otherwise interface with stimulation device 100. Ports 116 and 655 may adopt a common configuration, for 30 example according to the universal serial bus (USB) standard or may have a non-standard proprietary port/communication configurations. Computer system 605 may be

a computer system used by a patient to interface with stimulation device 100, for example in order to communicate with (or even power up using USB power, for example) stimulation device 100.

Alternatively, computer system 605 may be a computing device used by a clinician, such as a specialist therapist, to select one or a few configuration settings stored locally on system 605 or accessible to system 605 (for example over network 630), from among multiple possible selections to be used to configure or reconfigure the stimulation settings of device 100. The software operable on system 605 in order to effect such reconfiguration of device 100 may be stored in memory 615 and executable by processor 610. Such software may be downloaded from a remote location over network 630 in response to suitable accreditation or authentication of the clinician for use of such reconfiguration software.

As a further alternative, computer system 605 may be a computer used by a technician or manufacturer to interface with device 100 as needed, for example in order to provide initial or default stimulation settings.

In some embodiments, the electrodes 30 may be positioned in regions closer to the anus, for example, in an anterior lower pelvic area and in a posterior sacral or lower lumbar region, as illustrated in Figures 8A and 8B. Such electrode positioning is believed to assist with treatment of anorectal retention (AR) as evidenced by the information described in Example 3 below. Other than the lower positioning of the electrodes 30, such embodiments may use device 100 in combination with conductors 32 and electrodes 30 in the same way as described above.

As is illustrated in Figures 8A to 8C, the lower positioning of the electrodes 30 is intended to excite or modulate nerves and tissues around the recto-sigmoid colon to assist in evacuation of faecal matter accumulating in the large bowel distal of the proximal colon (i.e. through the descending colon, the sigmoid colon and the rectum). Figure 8C illustrates schematically how the crossing interferential currents are arranged between front left and back right (X to X) and between front right and back left (Y to Y).

The anterior electrodes 30 are positioned between about 1 and about 10 cm laterally apart relative to the vertical centre line of the umbilicus 11 in a lower pelvic area not extending down to the genitalia and not extending to the groin. The lateral separation of the electrodes 30 in some embodiments may be less than the embodiments described

- 5 above in relation to Figures 1 and 2. On the posterior side illustrated in Figure 8B, electrodes 30 may be approximately placed over the dimples that are commonly visible at the top inside part of the buttocks, generally corresponding with a lower lumbar or sacral region overlying or adjacent parts of the sacrum or ilium. The lateral spacing of the posterior electrodes may thus be between about 1 cm and about 10 cm, for example.
- 10 Optionally, the posterior electrodes 30 may be positioned slightly higher than the upper extent of the sacrum or ilium so as to reduce the likelihood of electrical current being directed through the bones of the sacrum and ilium as interferential current is passed between the anterior and posterior electrodes 30.

Some embodiments may employ more than two pairs of electrodes, such as the four

- 15 pairs illustrated in Figures 9A and 9B. In such embodiments, the electrode positions illustrated in Figures 1, 2, 8A and 8B may be combined, with two electrode pairs 810, 812 laterally and vertically spaced in anterior abdominal (810) and lower pelvic (812) areas. Additionally, on the posterior side, two pairs of electrodes 816, 818 are laterally and vertically spaced relative to the spine. The lower pair of posterior electrodes (818)
- 20 may be positioned just above the buttocks on either side of the spine or in a lower position that overlies an upper inner part of the buttocks overlying or adjacent the sacrum or ilium. The upper pair (816) of electrodes 30 may be positioned generally opposite the corresponding anterior upper pair (810) of electrodes 30 so as to be located in a para-spinal area on each lateral side of lumbar vertebrae T9-10 to L2.

- 25 In the embodiments as illustrated in Figures 9A and 9B, the electrodes 30 may be operated in two upper pairs and two lower pairs to deliver interferential current stimulation in sequence with one another or simultaneously. In some embodiments, the interferential current may be applied between opposed upper and lower electrodes. For example, stimulation current may be applied between one lower posterior electrode 30
- 30 and one diagonally opposite upper anterior electrode 30 and optionally also one diagonally opposite lower anterior electrode 30.

- 23 -

The upper electrode pairs should be located below the costal margin 802 and generally be positioned to excite or modulate parts of the ascending, transverse and descending colon, while the lower pairs of electrodes 30 should be generally positioned above the anus and genitals so as to excite or modulate the rectum and thereby assist in treatment 5 of anorectal retention. The two upper and two lower pairs of stimulation electrodes 30 are believed to be likely to have a combined positive treatment effect for evacuation dysfunction affecting parts of the large bowel, including the proximal colon as well as parts of the large bowel distal of the proximal colon, such as the sigmoid colon and rectum.

10 Referring now to Figure 10, a wearable electrode carrying structure in the exemplary form of a belt arrangement 910 is schematically illustrated. The belt 910 carries two pairs of anterior electrodes 30 and a further two pairs of posterior electrodes 30 (not shown) to provide TES using interferential current delivered from device 100 using conductors 32. Conductors 32 are at least partially supported by the belt 910 and are 15 preferably threaded or passed through at least part of the belt 910 or portions thereof. Belt 910 may comprise carrying means, such as a pocket, pouch, cradle or attachment mechanism, to support and carry the device 100 as the patient moves around.

20 Although the belt 910 is illustrated as having electrodes 30 positioned to provide TES in the manner described in relation to Figures 9A and 9B, belt 910 may instead carry a single anterior pair (810/812) and a single posterior pair (816/818) of electrodes for placement in the regions as illustrated in Figures 1, 2, 8A and 8B.

In further embodiments, the belt 910 may comprise a selectable array of interior 25 electrode connection positions by which the electrode 30 is fixed relative to the belt and is in electrical communication with conductors 32 in any of the array of positions. In this way, one or more than one pair of electrodes 30 may be provided anteriorly and one or more than one pair of electrodes 30 can be provided at a selected position or positions posteriorly according to the desired treatment regime prescribed by a medical professional. Once suitable positions of the electrodes 30 on the belt 910 are selected by the medical professional, the patient 10, 60 can simply place the electrodes 30 in the

correct position on their skin for each treatment session by wearing the belt 910 in the same position with respect to the patient's own anatomy.

Belt 910 may be suitably flexible and may be fitted and removed by suitable coupling means, for example such as a side flap 920 having fastening means 925 thereon, such as

- 5 hook and loop fasteners, buttons or clasps. Belt 910 may be formed of one or more individual or composite layers of flexible (optionally at least partially stretchable) fabric, including for example leather, Lycra, Spandex, cotton, nylon, plastic or other suitable fabric, to provide a wearable garment structure to support the device 100, conductors 32 and electrodes 30. Belt 910 may be generally fluid permeable or impermeable.
- 10 Preferably, belt 910 is made of one or more machine-washable materials. Belt 910 is preferably sized to be worn with reasonable comfort underneath normal clothing so that a patient wearing the belt 910 can perambulate normally while undergoing the TES.

In self-administering the treatment prescribed by the medical professional, each patient may be instructed to follow particular instructions for care of the electrodes and their

- 15 placement in order to maximise effective delivery of the TES.

Modifications of the described embodiments may be apparent to those skilled in the art, without departing from the spirit and scope of the described embodiments. The described embodiments are therefore intended to be exemplary and non-limiting when considered in the context of the appended claims.

- 20 Studies involving some described embodiments are described by the following non-limiting Examples.

- 25 -

EXAMPLE 1

Daily transcutaneous electrical stimulation increases defecation in children with slow-transit constipation.

Patient Groups Halfway through a randomized control trial (Clark *et al*, 2009 *supra*)

5 (RCT) testing transcutaneous electrical stimulation (TES), battery-operated machines became available, thus allowing home-based stimulation by parents. Eleven children (6M/5F, mean 14yr, range 12-18yrs) with slow-transit constipation who completed the RCT study, but relapsed or without an increase in defecation, were offered the opportunity to try the machine by a paediatrician (11±5 months after TES trial). These 10 children all had chronic constipation and soiling for a minimum of two years prior to the RCT and had failed to respond significantly to the TES in the trial, as well as medical treatments such as dietary modifications, oral and rectal laxatives.

All eleven children had undergone a radio-nuclear transit study to show slow-transit in the transverse colon. However, one child was found to have normal colonic transit and 15 because of that he was excluded from the TES Trial. He still had the TES treatment but did not have any improvement, so he was included in this study. Another child had more severe symptoms requiring an appendix stoma and his constipation and soiling was managed with antegrade washouts every 2-3 days (King *et al*, *J. Pediatr. Surg.*, 40:1935-1940, 2005).

20 ***Stimulation Regime*** Parents of the children were trained to use the battery-operated interferential stimulating machine (EPM IF 4160, Fuji Dynamics, Hong Kong) by the trial physiotherapist. Stimulation was performed and monitored by the parents at home (1 hour daily for a minimum of 2 months). Interferential treatments delivered a 4 kHz carrier frequency, a beat frequency of 80-150 Hz with an intensity of <33mAmp. Two 25 adhesive 3cm² electrodes were placed on the anterior abdominal wall below the costal margin of the children and two other electrodes placed on the posterior abdominal wall between T9 and L2 on either side. The current from the electrodes was crossed diagonally from front to back to ensure that the stimulation current from each electrode crossed over within the abdomen of the child.

Outcome Measures The number of spontaneous and “sit” defecations, number of antegrade enema washouts or medication and number of incidents of soiling were recorded daily in a specially designed continence diary. The daily diary was kept one month prior and for two months during electrical stimulation. The (a) frequency of 5 defecation, (b) frequency of soiling and (c) frequency of episodes of abdominal pain per month were compared using ANOVA with post-test analysis and paired t-tests (two tailed). P values <0.05 were considered statistically significant.

Results

Figure 7 is a graphical representation showing:

10 A) Total number of defecations/week,
 B) Number of days with soiling, and
 C) Number of days with abdominal pain.

Daily diaries were assessed for 1 month before stimulation (Pre Rx), and during the first month (Rx 1) and second month (Rx 2) of stimulation performed daily at home.

15 Statistically analysis by paired t-test.

There were no adverse events or complaints of discomfort reported from the children. All patients tolerated home stimulation well.

a) Defecation From Table 1, defecation increased in nine out of eleven children. There was a significant increase ($p=0.008$) in total defecation per week (mean \pm SD, 2.5 ± 2.1 vs 20 6.7 ± 4.4) (Figure 7A). Five of the eleven patients had less than three defecations/week in the baseline period and all of them had normal episodes of defecation (>3/week) during stimulation. Six children had an increase in spontaneous defecation (Table 1) and three experienced increase in defecation during trained sits on the toilet. One child had no increase in total defecation but there was a marked increase in spontaneous defecation 25 i.e. changes from timed sits to spontaneous defecation (urge to defecate followed by the child performing controlled defecation in the toilet).

b) Soiling There was a qualitative decrease in four out of eleven children, however this was not statistically significant (3.8 ± 1.6 vs 1.1 ± 0.5 episodes/week) ($p=0.1$) (Figure 7B).

- 27 -

In the baseline period, soiling was frequent in 3/11 children (Table 1). This was reduced to low episodes during stimulation in all of the three children. The patient with an appendix stoma had his soiling managed by antegrade enemas and therefore recorded no soiling.

5 **c) Pain** Daily stimulation did not affect abdominal pain (0.97 ± 1.8 vs 1.03 ± 2.0 episodes/week, $p=0.7$) (Figure 7C).

This study found that the use of a battery-operated, portable machine was safe to deliver interferential TES in the home environment. Moreover, daily stimulation of one hour per session for ≥ 1 month caused a significant increase in total episodes of defecation/week, 10 with no adverse effects. Importantly, this improvement in bowel function occurred in patients who had had only marginal or temporary improvement following stimulation three times per week in our recently conducted randomized controlled trial (Clark *et al*, *J. Pediatr. Surg.*, 43:320-324, 2008).

In the formal trial, the TES was given for 20 minutes in each treatment session, three 15 times per week, so that the total duration of therapy was one hour per week. When the trial was designed, the optimal parameters (if any) were unknown, and hence were selected on arbitrary criteria. Having demonstrated a statistical improvement in transit times (Clark *et al*, 2009 *supra*), 24-hour colonic manometry (King *et al*, *Am. J. Gastroenterol.*, 103:2083-2091, 2008), and quality of life (Clark *et al*, 2008 *supra*) in the 20 trial, the current study aimed to find out whether more frequent TES treatment might further improve function.

Daily TES at home enabled the children in this pilot study to receive 7 hours of treatment per week (compared to 1 hour per week in the formal trial), which improved their overall bowel function in frequency of defecation. Interestingly, this was one of the 25 few parameters which was not improved in the randomized placebo-controlled trial, despite increased colonic transit and peristaltic activity. The reason for this difference is unknown but it may be that TES three times per week caused increased proximal colonic transit, but without altering rectal evacuation. The children then needed to learn how to defecate normally to take advantage of this increased proximal colonic contractility. In

this pilot study, improved defecation may have occurred because more time had passed since the trial and the patients were better able to learn how to empty their colon under conscious control. Another possibility is that the increased time length and frequency of treatment from 1 hour to 7 hours per week provided a much more substantial stimulus to 5 colonic function, enabling not only more rapid proximal transit but also more effective evacuation.

There was concern that the electrical stimulation might increase soiling but this did not occur. Daily stimulation showed a qualitative reduction in frequency of soiling but this was not statistically significant. Although more patients need to be tested, it is 10 considered that this qualitative improvement is likely to be important. The frequency of abdominal pains, assumed to be caused by colonic peristalsis, was also measured and this is unchanged by the daily TES.

TES has only been tested in very limited studies in children, outside the controlled trial described herein, so its place in the armamentarium of constipation therapies is 15 unknown. The initial pilot study with TES used sophisticated machines connected to the power grid and operated by a licensed physiotherapist (Chase *et al*, *J. Gastroenterol. Hepatol.*, 20:1054-1061, 2005). This study extends that and shows that a simpler machine with a 9-volt rechargeable battery is adequate to provide the requisite stimulation, and may even provide superior effects on the bowel. 20 The RCT showed that TES is significantly more effective than a placebo. Hence, an increased duration and frequency of TES treatment times is an advantage.

In the cohort of children being treated for slow transit constipation by transcutaneous electrical stimulation using the battery-operated interferential stimulating device (EPM IF 4160, Fuji Dynamics, Hong Kong), it was found during patient follow-up that there 25 were some treatment efficacy problems arising from misuse or improper use of the device. The Fuji Dynamics device has a relatively large number of buttons to push and many stimulation settings and options to choose from.

In one case, the parent was confused by which buttons to push to get the proper settings for the treatment (needing 6 buttons to be pushed). This parent was also illiterate and the

- 29 -

written instructions for use were essentially useless, leading to a hands-on demonstration of use of the device being needed in order to make sure the parent was getting the stimulation settings right. It was also found that the patient or the parent would commonly forget to set the buttons to achieve the essential settings for treatment. In one 5 instance, a patient had gone through the treatment period without setting the stimulation frequency, resulting in inadequate stimulation being applied and the patient not experiencing any improvement in symptoms.

In younger children, it has been found that they tend to play with the buttons of the device. The Fuji Dynamics device used in the trial did not have any locking features to 10 prevent inappropriate user modification of the device settings. One child experienced shooting pain in both legs after changing the stimulation settings on the device.

EXAMPLE 2

TES treatment long term benefits

15 In a follow-up study of patients that have previously been enrolled in a TES trial it was noted that there was continued improvement in patients after the time of usage of TES.

There are total of 105 patients which have been included into the transcutaneous electrical stimulation (interferential) therapy (Table 2). They comprise of 66 male patients and 39 female patients. The age ranges of these patients are 6 – 18years, with 20 the mean age of 11.5 years. All these patients have been selected for the treatment after being diagnosed with slow transit constipation using nuclear transit study. There are different phases of development of the transcutaneous electrical stimulation, from TIC TOC to TENS trial and the home stimulation therapy. All patients were required to fill out the continence diaries and PedsQL questionnaires appropriate to age upon enrolment 25 into the trial and as the scheme for follow-up during and after treatment.

In the TIC TOC trial, at 6 months follow-up, out of the 39 patients with data analysed at the completion of the trial, 20 patients responded to the questionnaires survey. Sixteen out of 20 patients have ≥ 3 episodes of defecations per week at 6 months after treatment.

- 30 -

At 12 months follow-up, 13 responded to the questionnaire survey. Eleven out of 13 patients have ≥ 3 episodes of defecations per week. As for soiling, at 6 months follow-up, 14 out of 23 respondents have no soiling. At 12 months follow-up, 6 out of 11 respondents have no soiling. There was overall improvement in patients with slow-

5 transit constipation treated by transcutaneous interferential therapy.

In the latest home stimulation therapy, one patient has complete resolution of abdominal pain, twofold increase in defecation, and 3.7-fold improvement of soiling at one month follow-up.

The described stimulation regimes in Examples 1, 2 and 3 have been used without also

10 directly stimulating the pelvic floor or area around the perineum or perianum.

EXAMPLE 3

There are 4 children currently using the Anorectal retention (AR) protocol to treat their chronic constipation, the results of which are shown in Table 3. The position of the electrodes 30 is shown and described above in relation to Figures 8A to 8C. The

15 stipulated TES treatment regime and settings were the same as the other Examples, including the treatments being for about 60 minutes each day over a period of several weeks or months.

Table 3

Children with chronic constipation using AR protocol.

	Gender	Age	Clinical diagnosis of chronic constipation	Elecrode positions used	Current status (25 Feb 2011)
1	Male	9	Slow-transit constipation (STC)	STC protocol, developed AR (palpable faecaloma) 6 months later. AR protocol used.	Clinically improved after 6 months and remained well (phone interview done 22 Feb 2011). No more faecaloma.
2	Male	4	STC	STC protocol for 6 months. Repeat transit study showed AR. Started AR protocol Dec 2010.	Feeling stronger urge at bottom with reduced laxative used (phone interview done on 22 Feb 2011)
3	Male	6	STC+Anorectal retention (AR)	Used STC protocol for 4 months before switched over to AR protocol.	Clinically improved with more regular bowel actions. Remained same (phone interview done on 22 Feb 2011).
4	Female	6	AR	Palpable faecaloma. Started on AR protocol on 17 Sep 2010.	Still awaiting phone reply from parent.

As these preliminary data in Table 3 indicate, TES treatment using interferential current applied at lower pelvic and sacral regions is effective to treat evacuation dysfunction in the recto-sigmoid colon. This has been shown to be effective as a procedure performed after treatment at higher electrode positions for STC. The indication is that this lower electrode positioning can be used as a supplemental or independent procedure to the higher electrode positioning. If the treatment procedure with the lower electrodes is

- 32 -

supplemental to the treatment with higher electrodes, it may be performed either simultaneously or in sequence.

Those skilled in the art will appreciate that the invention described herein is susceptible to variations and modifications other than those specifically described. It is to be
5 understood that the invention includes all such variations and modifications. The invention also includes all of the steps, features, compositions and compounds referred to or indicated in this specification, individually or collectively, and any and all combinations of any two or more of said steps or features.

Table 1

Sex	Appendix	Stoma	Transit	Defecation/week				Days with soiling/week				Days with pain/week			
				Study	Before	IFT 1	IFT 2	Change	Before	IFT 1	IFT 2	Before	IFT 1	IFT 2	Before
<i>With slow transit and no appendix stoma</i>															
F	No	Slow	0	12	14	1/2 sits & spontaneous	0	3	0	0	1	0			
F	No	Slow	0	4	ND	↑ sits	0	1	ND	3	0	ND			
F	No	Slow	1	3	ND	↑ spontaneous	6	1	ND	5	5	ND			
F	No	Slow	3	11	11	↑ sits	7	3	3	0	1	1			
F	No	Slow	7	10	12	↑ spontaneous	2	0	0	0	2	0			
M	No	Slow	1	4	3	↑ spontaneous	0	0	0	1	3	4			
M	No	Slow	2	6	6	↑ spontaneous	3	5	3	0	1	1			
M	No	Slow	3	2	4	↑ spontaneous	0	0	0	0	0	0			
M	No	Slow	4	2	3	1/2 sits & spontaneous	4	4	3	0	2	1			
<i>With appendix stoma</i>															
M	Yes	Slow	3	5	9	↑ spontaneous	0	0	0	0	0	0			
<i>With Normal Transit</i>															
M	No	Normal	4	12	ND	↑ sits	7	2	ND	2	0	ND			

ND: No Data

Table 2**TIC TOC**

Started with 60 patients; 39 patients available for analysis and on follow-up. Follow-up periods of 1 year to 2 years. 21 patients were excluded from study.

GENDER	NUMBER (N)
MALE	21
FEMALE	18

AGE RANGES FROM 7 – 18 YEARS, MEAN AGE OF 11.8 YEARS (OF THE 39 PATIENTS INCLUDED IN THE FINAL ANALYSIS).

GENDER	NUMBER (N)
MALE	39
FEMALE	21

AGE RANGES FROM 6 – 18 YEARS, MEAN AGE OF 12.3 YEARS (OF THE 60 PATIENTS AT THE BEGINNING OF THE STUDY)

TENS

GENDER	NUMBER (N)
MALE	5
FEMALE	5

AGE RANGES FROM 5 – 16 YEARS, MEAN AGE OF 9.3 YEARS.

FOLLOW-UP OF PATIENTS FROM 6 MONTHS TO 1YEAR

1ST HOME STIMULATION THERAPY

GENDER	NUMBER (N)
MALE	6
FEMALE	5

AGE RANGES FROM 9 – 15 YEARS, MEAN AGE OF 12 YEARS.

FOLLOW-UP OF PATIENTS FROM 6 MONTHS TO 1YEAR.

- 35 -

2ND HOME STIMULATION THERAPY

GENDER	NUMBER (N)
MALE	16
FEMALE	8

AGE RANGES FROM 4 -16 YEARS. MEAN AGE OF 10.3 YEARS

FOLLOW UP OF PATIENTS FROM 0-5 MONTHS.

BIBLIOGRAPHY

Benninga *et al*, *J Pediatr Gastroenterol Nutr.*, 23:241-51, 1996

Chase *et al*, *J. Gastroenterol. Hepatol.*, 20:1054-1061, 2005

Clark *et al*, *J. Pediatr. Surg.*, 43:320-324, 2008

Clark *et al*, *J. Pediatr. Surg.*, 44:408-412, 2009

Hutson *et al*, *J Pediatr Surg.*, 31:580-583, 1996

King *et al*, *J. Pediatr. Surg.*, 40:1935-1940, 2005

King *et al*, *Am. J. Gastroenterol.*, 103:2083-2091, 2008

Shin *et al*, *J Pediatr Surg.*, 37:1762-1765, 2002


CLAIMS:

1. A method of treating a waste evacuation dysfunction, comprising administering transcutaneous electrical stimulation (TES) to at least one lower pelvic region and at least one sacral region for at least one treatment period per day over a treatment term of at least one week, wherein the TES comprises interferential electrical stimulation.
2. The method of Claim 1, wherein the stimulation is administered for two or three treatment periods per day.
3. The method of Claim 1 or Claim 2, wherein a stimulation time of the treatment period is between about 10 minutes and about 90 minutes.
4. The method of any one of Claims 1 to 3, wherein the treatment term is between two weeks and three months.
5. The method of Claim 4, wherein the treatment term is between one month and two months.
6. The method of any one of Claims 1 to 5, further comprising repeating the administering for more than one treatment term.
7. The method of any one of Claims 1 to 6, wherein the administering is repeated over a greater term that is between about four months and about two years.
8. The method of any one of Claims 1 to 7 , wherein the method comprises providing a portable stimulation device and the stimulation is administered using the portable stimulation device.
9. The method of Claim 8 , wherein the stimulation device is battery powered.


10. The method of any one of Claim 8 or 9, wherein the stimulation device is configured to provide the stimulation according to a selected set of fixed stimulation parameters stored in the stimulation device.
11. The method of Claim 8 or Claim 9, wherein the stimulation device is configured to provide the stimulation according to a single set of fixed stimulation parameters at any time, and wherein the stimulation device is configured so that the single set of stimulation parameters may only be replaced with another single set of operating parameters using an electronic configuration interface separate from, but communicably coupleable to, the stimulation device.
12. The method of any one of Claims 1 to 11, wherein the stimulation is provided using first and second stimulation channels, the first stimulation channel comprising conductors electrically coupled to two spaced electrodes positioned on skin surfaces of an anterior lower pelvic region, the second stimulation channel comprising conductors electrically coupled to two spaced electrodes positioned on skin surfaces of a sacral region.
13. The method of Claim 12, wherein the electrodes are provided on a carrier configured to be secured to the lower pelvic or sacral region.
14. The method of Claim 13, further comprising securing the carrier to the lower pelvic or sacral region so that the electrodes are substantially evenly laterally spaced relative to the spine.
15. The method of any one of Claims 1 to 14, wherein the stimulation comprises providing stimulation current of magnitude less than about 40 mA and greater than zero.
16. The method of any one of Claims 1 to 15, wherein the current is provided at a carrier frequency of between about 1 kHz and about 10 kHz, with a modulated frequency of about 20 to about 300 Hz.

17. The method of any one of claims 1 to 16, further comprising administering the TES to at least one lumbar and/or abdominal region simultaneously with administering the TES to the at least one lower pelvic and sacral region.
18. The method of claim 17, wherein the TES is administered to lower pelvic, sacral, lumbar and abdominal regions in a same treatment period.
19. The method of claim 18, wherein the administering TES includes administering the TES using four pairs of spaced electrodes positioned one in each of the lower pelvic, sacral, lumbar and abdominal regions.
20. A system for transcutaneously treating a waste evacuation dysfunction, comprising:
 - at least four electrodes comprising a first pair of electrodes positioned one on each lateral side of the spine or sacrum in the sacral region and a second pair of electrodes positioned laterally across the lower pelvic region;
 - a stimulation device configured to provide electrical stimulation signals to the at least four electrodes to perform the method of any one of Claims 1 to 19.
21. The system of Claim 20, wherein the stimulation device is configured to have a single set of stored stimulation settings that control the provision of the electrical stimulation signals.
22. The system of Claim 21, wherein the stimulation device is free of user-operable controls that would allow modification of the stimulation settings.
23. The system of Claim 21 or Claim 22, wherein the stimulation device comprises a port coupleable to an external computing device to allow modification of the stimulation settings.
24. The system of any one of Claims 20 to 23, wherein the stimulation device stores a limited set of selectable stimulation settings that control the provision of the electrical stimulation signals.



25. The system of any one of Claims 20 to 21, wherein the at least four electrodes comprise a first pair of electrodes positioned one on each side of the umbilicus and a second pair of electrodes positioned in a para-spinal area on each side of lumbar vertebrae T9-10 to L2.
26. The system of any one of claims 20 to 25, further comprising a wearable structure to carry the stimulation device and to assist in positioning the at least four electrodes on the skin.
27. The system of claim 37, wherein conductors electrically coupling the stimulation device to the at least four electrodes are supported by the wearable structure.
28. The system of claim 26 or claim 27, wherein the wearable structure comprises a belt.
29. The system of any one of claims 20 to 28, wherein the stimulation device comprises at least one accelerometer to detect at least one of movement and orientation of the stimulation device.
30. The system of any one of claims 20 to 29, wherein the stimulation device comprises means to assess patient compliance with the treatment regime based on information collected and stored on the stimulation device regarding patient usage of the stimulation device.
31. The system of any one of claims 20 to 30, wherein the stimulation device comprises a microcontroller having a timing function and is configured to disallow further administration of electrical stimulation for a period following provision of electrical stimulation.
32. Use of the method of any one of claims 1 to 19 or the system of any one of claims 20 to 31 to treat at least one of: constipation; slow transit constipation; irritable bowel syndrome; ileus; incontinence; and anorectal retention.

**FIGURE 1**

2/10

**FIGURE 2**

3/10



4/10




FIGURE 4

5/10






FIGURE 5

6/10

**FIGURE 6**

7/10

**FIGURE 7**

8/10





Figure 8A

Figure 8B

Figure 8C



10/10




Figure 10